APCL

Applied Physics & Chemistry Laboratory

13760 Magnolia Ave. Chino CA 91710
Tel. (909) 590-1828 Fax (909) 590-1498

November 21, 2002

SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr. Suite 212
San Diego CA 92127

Dear Yu,

This package contains samples in our Service ID 02-5485 and your project is 00HW019 JPL. Enclosed please find:

(1) One original report.
(2) One original Chain of Custody.
(3) One diskette containing EDD Deliverable.
(4) One original of Level D Data Package Deliverable.

If anything is missing or you have any questions, please feel free to contact me.

Respectfully submitted,

Regina Kirakozova
Associate QA/QC Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Submitted to:
SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr, Ste. 212
San Diego CA 92127
Tel: (858)485-8100 Fax: (858)485-0812

Service ID #: 801-025485A
Collected by: TAM/DJL/MES.
Collected on: 10/15/02

Received: 10/15/02
Extracted: N/A
Tested: 10/16-21/02
Reported: 10/23/02

Sample Description: Water from Pasadena, CA.
Project Description: 00HW019 JPL

Analysis of Water Samples

Component Analyzed	Method	Unit	PQL	Analysis Result ER-12 02-05485-1	MW-12-1 02-05485-2	MW-12-2 02-05485-3	MW-12-3 02-05485-4
Dilution Factor				1	1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	<4	<4
VOLATILE ORGANIC COMPOUNDS							
Dilution Factor				1	1	1	1
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	0.8	2.4
CHLOROFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5	2.2
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TOLUENE	524.2	µg/L	0.5	0.5J	-	-	-
TRICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5

Component Analyzed	Method	Unit	PQL	Analysis Result MW-12-4 02-05485-5	MW-12-5 02-05485-6	SB-12 02-05485-7	TB-12 02-05485-8
Dilution Factor				1	1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	-	-
VOLATILE ORGANIC COMPOUNDS							
Dilution Factor				1	1	1	1
CARBON TETRACHLORIDE	524.2	µg/L	0.5	1.6	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	0.7	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	-	-	-	0.7J
METHYL-T-BUTYL ETHER (MTBE)	524.2	µg/L	1	-	-	0.4J	-
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TOLUENE	524.2	µg/L	0.5	-	-	0.5J	-
TRICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	-	-	0.3J	-

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result ER-12 02-05485-1	MW-12-1 02-05485-2	MW-12-2 02-05485-3	MW-12-3 02-05485-4
CHROMIUM (VI)	7196	mg/L	0.01	< 0.01	< 0.01	< 0.01	< 0.01

PQL: Practical Quantitation Limit. MDL: Method Detection Limit CRDL: Contract Required Detection Limit
N.D.: Not Detected or less than the practical quantitation limit. "-": Analysis is not required.
J: Reported between PQL and MDL.
Listed Dilution Factors (DF) are relative to the method default DF. All unlisted DFs are 1.0

Respectfully submitted,

Dominic Lau
Laboratory Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Submitted to:
SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr. Ste. 212
San Diego CA 92127
Tel: (858)485-8100 Fax: (858)485-0812

Service ID #: 801-025485
Collected by: TAM/DJL/MES.
Collected on: 10/15/02

Received: 10/15/02
Extracted: N/A
Tested: 10/16-21/02
Reported: 10/23/02

Sample Description: Water from Pasadena, CA.
Project Description: 00HW019 JPL

Analysis of Water Samples

Component Analyzed	Method	Unit	PQL	Analysis Result ER-12 02-05485-1	MW-12-1 02-05485-2	MW-12-2 02-05485-3	MW-12-3 02-05485-4
Dilution Factor				1	1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	<4	<4
VOLATILE ORGANIC COMPOUNDS							
Dilution Factor				1	1	1	1
BENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOCHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMODICHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
SEC-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TERT-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	0.8	2.4
CHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLORODIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5	2.2
CHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
4-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMO-3-CHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMOETHANE (EDB)	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
DIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,4-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
DICHLORODIFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRANS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result ER-12 02-05485-1	MW-12-1 02-05485-2	MW-12-2 02-05485-3	MW-12-3 02-05485-4
1,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRANS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
ETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
HEXACHLOROBUTADIENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
ISOPROPYLBENZENE (CUMENE)	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
P-ISOPROPYLTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	<1	<1	<1	<1
METHYL-T-BUTYL ETHER (MTBE)	524.2	µg/L	1	<1	<1	<1	<1
NAPHTHALENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-PROPYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
STYRENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TOLUENE	524.2	µg/L	0.5	0.5J	<0.5	<0.5	<0.5
1,2,3-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRICHLOROFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3,5-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
VINYL CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
O-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-12-4 02-05485-5	MW-12-5 02-05485-6	SB-12 02-05485-7	TB-12 02-05485-8
Dilution Factor				1	1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	-	-
VOLATILE ORGANIC COMPOUNDS							
Dilution Factor				1	1	1	1
BENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOCHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMODICHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
SEC-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TERT-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CARBON TETRACHLORIDE	524.2	µg/L	0.5	1.6	<0.5	<0.5	<0.5
CHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLORODIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	0.7	<0.5	<0.5	<0.5
CHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
4-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMO-3-CHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMOETHANE (EDB)	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
DIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,4-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
DICHLORODIFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRANS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2,2 DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRANS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
ETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
HEXACHLOROBUTADIENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
ISOPROPYLBENZENE (CUMENE)	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-12-4 02-05485-5	MW-12-5 02-05485-6	SB-12 02-05485-7	TB-12 02-05485-8
P-ISOPROPYLTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	<1	<1	<1	0.7J
METHYL-T-BUTYL ETHER (MTBE)	524.2	µg/L	1	<1	<1	0.4J	<1
NAPHTHALENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-PROPYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
STYRENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TOLUENE	524.2	µg/L	0.5	<0.5	<0.5	0.5J	<0.5
1,2,3-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRICHLOROFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3,5-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
VINYL CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
O-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	0.3J	<0.5

Component Analyzed	Method	Unit	PQL	Analysis Result ER-12 02-05485-1	MW-12-1 02-05485-2	MW-12-2 02-05485-3	MW-12-3 02-05485-4
CHROMIUM (VI)	7196	mg/L	0.01	<0.01	<0.01	<0.01	<0.01

PQL: Practical Quantitation Limit. MDL: Method Detection Limit. CRDL: Contract Required Detection Limit
N.D.: Not Detected or less than the practical quantitation limit. "-": Analysis is not required.
J: Reported between PQL and MDL.
Listed Dilution Factors (DF) are relative to the method default DF. All unlisted DFs are 1.0

Respectfully submitted,

Dominic Lau
Laboratory Director
Applied P & Ch Laboratory

Level D Data Package Deliverables

General Information

Project: 00HW019 JPL

APCL Service ID: 02-5485

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino, CA 91710
Telephone (909)590-1828
Fax (909)590-1498

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Case Narrative

Project: JPL/00HW019

For SOTA Environmental

APCL Service No: 02-5485

1. Sample Identification

The sample identifications are listed in the following table:

SOTA Environmental Sample ID	APCL Sample ID
TB-12	02-05485-8
MW-12-5	02-05485-6
MW-12-4	02-05485-5
MW-12-3	02-05485-4
MW-12-2	02-05485-3
MW-12-1	02-05485-2
SB-12	02-05485-7
ER-12	02-05485-1

2. Analytical Methodology

Samples are analyzed by EPA methods

524.2 (Volatile Organic Compounds),
7196 (Chromium (VI)),
314.0 (Perchlorate, low level),

3. Holding Time

All samples were extracted, digested and analyzed within the holding times defined by the appropriate EPA methods of the analyses.

4. Preservation

All samples were preserved and stored according to the appropriate EPA methods.

5. Tele-log

None

6. Anomaly

None

0 5300

"I certify that these data are technically accurate, complete, and in compliance with the terms and conditions of the contract, for other than the conditions detailed above. Release of the data contained in the hardcopy data package and its electronic data deliverable submitted on diskette had been authorized by the Laboratory Manager or her/his designee, as verified by the following signature."

Respectfully submitted,

Regina Kirakozova
Associate QA/QC Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory

APCL

13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Chain of Custody

Please Print in pen Page 1 of 2

Client: SOTA ENVIR. TECH., INC. Contact: MIKE SAYRE Tel #: 858-485-8100 Fax #: 858-485-0812

Address: 16835 W. BERNARDO DR. #212 City: SAN DIEGO State: CA Zip code: 92127

Bill to: SOTA

Project Name/Code JPL Job #00HW019 P.O. #

Project Address PASADENA, CALIFORNIA APCL Quotation #

Due Date: ☐regular ☐rush: ___ days ___ hours Sampled by: TAM/DJL/MES

Field Sample ID No.	Sample Description	Date Collected	Time Collected	Sample Matrix	Preservation	# of Containers	VOCs (524.2)	PERCHLORATE (314.0)	TOTAL Cr (200.8)	CrVI (7196)	Remarks
TB-12	TRIP BLANK	10/15/02	0915	WATER	HCl	2	X				EPA LEVEL IV
MW-12-5	MW-12-5		1004		HCl	3	X				
MW-12-5	MW-12-5		1004		—	1		X			QA/QC
MW-12-4	MW-12-4		~~0946~~ 1046		HCl	3	X				
MW-12-4	MW-12-4		1046		—	1		X			
MW-12-3	MW-12-3		1104		HCl	3	X				
					—	1		X	X		
					HNO_3	1			X		
MW-12-2	MW-12-2		1132		HCl	3	X				
					—	1		X	X		
					HNO_3	1			X		
MW-12-1	MW-12-1		1200		HCl	3	X				
					—	1		X	X		
					HNO_3	1			X		
SB-12	SOURCE BLANK		0955		HCl	2	X				

5485

White – With report
Yellow – Lab copy
Pink – Originator

QC Requirement: ☐Regular; ☐QA/QC Report; ☐WIP; ☐Raw Data; ☐Extended Raw Data ☐CLP; ☐ACE ☐AFCEE ☐NEESA ___ (E, C or D); ☐Other _____ (Please specify)

Sample Disposal: ☐Return ☐Disposal by APCL ☐Hold for _____ days after receiving date. **If not specified, samples will be discarded 45 days after samples are received.**

Sample Conditions: ☐Intact; ☐Broken. Cooler Seal: ☐Intact; ☐Broken; ☐None. Tag # _____ Temperature: ☐Room ☐Cold (_____ °C).

Relinquished by M. SAYRE Date/Time 10/15/02 1838 Received by Ray Aguilera Date/Time 10/15/02 1838

Relinquished by Ray Aguilera Date/Time 10/15/02 1930 Received by [signature] Date/Time 10/15/02 1930

APCL USE ONLY Service # Note:

Clients understand that all terms described in the proposals, quotations for this project, and/or the general terms provided in the current APCL price schedules will be followed. APCL reserves the right to terminate its service or withhold delivery of any reports, if in APCL's sole discretion the terms of the project have been broken.

Applied P & Ch Laboratory

APCL

13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Chain of Custody

Please Print in pen Page 2 of 2

Client: SOTA Environ. Tech. Inc. Contact: Mike Sayre Tel #: 858-485-8100 Fax #: 858-485-0812

Address: 16835 W. Bernardo Dr # 212 City: San Diego State: CA Zip code: 92127

Bill to: SOTA

Project Name/Code JPL Job # 00HW019 P.O. #

Project Address Pasadena, CA APCL Quotation #

Due Date: ☑ regular ☐ rush: ___ days ___ hours Sampled by: MES/TAM/DJL

White – With report
Yellow – Lab copy
Pink – Originator

Field Sample ID No.	Sample Description	Date Collected	Time	Sample Matrix	Preservation	# of Containers	Analysis Items: VOCs (5242)	Total Cr (200.8)	Cr VI (7196)	Perchlorate (314.0)	Remarks
ER-12-	Equipment Rinsate	10/15/02	1105	Water	HCl	3	X				EPA Level IV QA/QC
		10/15/02		water	HNO3	1		X			
		10/15/02		Water	—	1			X	X	

5485

QC Requirement: ☐ Regular; ☐ QA/QC Report; ☐ WIP; ☐ Raw Data; ☐ Extended Raw Data ☐ CLP; ☐ ACE ☐ AFCEE ☐ NEESA ___ (E, C or D); ☐ Other _____ (Please specify)

Sample Disposal: ☐ Return ☐ Disposal by APCL ☐ Hold for _____ days after receiving date. **If not specified, samples will be discarded 45 days after samples are received.**

Sample Conditions: ☐ Intact; ☐ Broken. Cooler Seal: ☐ Intact; ☐ Broken; ☐ None . Tag # _____ Temperature: ☐ Room ☐ Cold (_____ °C).

Relinquished by M. SAYRE Date/Time 10/15/02 1838 Received by Roy Gillette Date/Time 10/15/02 18:58

Relinquished by Roy Gillette Date/Time / Received by Date/Time /

APCL USE ONLY Service # Note:

Clients understand that all terms described in the proposals, quotations for this project, and/or the general terms provided in the current APCL price schedules will be followed. APCL reserves the right to terminate its service or withhold delivery of any reports, if in APCL's sole discretion the terms of the project have been broken.

Applied P & Ch Laboratory
13760 Magnolia Ave., Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Sample Receiving Checklist

APCL ServiceID: 5485 **Client Name/Project:** Sota Environmental

1. Sample Arrival

Date/Time Received 10/15/02 1930 Date/Time Opened 10/15/02 1930 By (name): Kenny Chan

Custody Transfer: ☐ Client ☐ Golden State ☐ UPS ☐ US Mail ☐ FedEx ☑ APCL Empl: Rory A.

2. Chain-of-Custody (CoC)

☑ With Samples? ☐ Faxed? ☑ Client has Copy? ☑ Signed, dated? By: Mike S.
☐ Project ID? ☑ Analyses Clear? ☐ Hold Samples? #on Hold _____ # Received 8
☑ CoC/Docs Zip-Locked under lid? ☐ Compos.#: _ ☑ #Samples OK? _____
☐ Discrepancies? ☐ Client notified? ☐ Response (attach docs): _____

3. Shipping Container/Cooler

☑ Cooler Used? # of 1 Cooled by: ☑ Ice ☐ Blue Ice ☐ Dry Ice ☐ None
Temp °C 4.0 _____ _____ _____ _____ _____
(Cooler temperature measured from temp blank if present, otherwise measured from the cooler).
Cooler Custody Seal? ☐ Absent ☐ Intact ☐ Tampered?

4. Sample Preservation

☐ **pH <2** ☐ pH >12
If Not, pH = _____ Preserved by: ☐ Client ☐ APCL ☐ Third Party _____

5. Holding-time Requirements

☐ pH 24hr ☐ BACT 6/24hr ☑ Cr^{VI} 24hr ☐ NO_3^- 48hr ☐ BOD 48hr
☐ Cl_2 ASAP ☐ Turbidity 48hr ☐ DO ASAP ☐ Fe(II) ASAP
☐ HT Expired? ☐ Client notified?

6. Sample Container Condition

☑ Intact? ☐ Broken? ☐ Documented? Number: _____
Type: ☑ plastic ☑ glass ☐ Tube: brass/SS ☐ Tedlar Bag
☑ Quantity OK? ☐ Leaking? ☐ Anomaly?
☑ Caps tight? ☐ Air Bubbles? ☐ Anomaly?
Labels: ☑ Unique ID? ☑ Date/Time ☐ Preserved?

7. Turn Around Time

☐ RUSH TAT: _____ ☑ Std (7-10 days) ☐ Not Marked

8. Sample Matrix

☐ Drinking H_2O ☑ Other Liq ☐ Soil ☐ Wipe ☐ Polymer ☐ Air ☐ Other: _____
☐ Ground H_2O ☐ Sludge ☐ Filter ☐ Oil/Petro ☐ Paint ☐ W. Water ☐ Extract ☐ Unknown

9. Pre-Login Check List Completed & OK?

☑ ALL OK? (if not, attach docs) ☐ Client Contact? (Name:_____)Date/Time:_____

Received/Checked by: Ky Le Date: 15 Oct 2002 Time: 8:21 a.m.

*HT: Samples must be analyzed for results to reflect total concentrations. Results generated outside required of holding times are considered minimal values and may be used to define waste as hazardous but not as non-hazardous

DocumentFile: [neal.texfiles]smprcl.tex.

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Sample Login: Check List

02-05485 (1288_ 397) (4858100_ 397)

10/16/02

Part 1: General Information

☐	Company Information	Name:	*SCTA Environmental*
		Address:	*16835 W. Bernardo Dr, Ste. 212 ,San Diego ,CA 92127*
☐	Project Information	Project Description:	*JPL*
		Project #:	*00HW019*
☐	Billing Information	P.O. #:	
		Bill Address:	*16835 W. Bernardo Dr, Ste. 212 ,San Diego ,CA 92127*
		Lab Project ID:	*2002-0002*
		Client Database #:	*01*
☐	Receiving Information	Who Received Sample?	*Kenny Chan*
		Receiving Date/Time:	*10/15/02 1930*
		COC No.	
☐	Shipping Information	Shipping Company	*APCL pick up*
		Packing Information:	*Cooler/Ice Chester*
		Cooler Temperature:	*4.0 °C*
☐	Container Information	Container Provider:	*Client*
☐	Sampling Information	Sampling Person:	
		Sampling Company:	*Client*
☐	Turn-Around-Time Option:		*Rush 5 working day(s)*
☐	QC Option:		*NELESA D*
☐	Disposal Option:		*Not specify*

0 5003

Part 2: Sample Information

Seq. #	Sample ID (on COC)	Sample Sub-ID	APCL Sample ID	Matrix	Cont-tainer	Preser-vative	Vol, ml Am	g	# of Replica	Condition G, L, B	Collected mmddyy	Hold ?	Composite Group	TAT Days	
1	TB-12	VOC	02-05485-8	W	V	C	40		2	G	101502	N	0	7	☐
2	MW-12-5	VOC	02-05485-6-α	W	V	C	40		3	G	101502	N	0	7	☐
	MW-12-5	Perch	02-05485-6-β	W	P		500		1	G	101502	N	0	7	☐
3	MW-12-4	VOC	02-05485-5-α	W	V	C	40		3	G	101502	N	0	7	☐
	MW-12-4	Perch	02-05485-5-β	W	P		500		1	G	101502	N	0	7	☐
4	MW-12-3	VOC	02-05485-4-α	W	V	C	40		3	G	101502	N	0	7	☐
	MW-12-3	CRVI/Perch	02-05485-4-β	W	P		500		1	G	101502	N	0	7	☐
5	MW-12-2	VOC	02-05485-3-α	W	V	C	40		3	G	101502	N	0	7	☐
	MW-12-2	CRVI/Perch	02-05485-3-β	W	P		500		1	G	101502	N	0	7	☐
6	MW-12-1	VOC	02-05485-2-α	W	V	C	40		3	G	101502	N	0	7	☐
	MW-12-1	CRVI/Perch	02-05485-2-β	W	P		500		1	G	101502	N	0	7	☐
7	SB-12	VOC	02-05485-7	W	V	C	40		3	G	101502	N	0	7	☐
8	ER-12	VOC	02-05485-1-α	W	V	C	40		3	G	101502	N	0	7	☐
	ER-12	CRVI/Perch	02-05485-1-β	W	P		500		1	G	101502	N	0	7	☐

Part 3: Analysis Information

Test Items:

☑	*524.2*	Volatile Organic Compounds
☐	*7196*	Chromium (VI)
☐	*300.0*	Perchlorate, low level
☐	*200.7/6010*	Sodium, Na, by ICP
☐	*200.7/6010*	Potassium, K, by ICP
☐	*200.7/6010*	Calcium, Ca, by ICP
☐	*200.7/6010*	Magnesium, Mg, by ICP
☐	*200.7/6010*	Iron, Fe, by ICP
☐	*300.0*	Sulfate (SO_4^{--}), by IC
☐	*300.0/SM4500NO3C*	Nitrate (NO_3^-) as N by IC
☐	*300.0*	Chloride Cl^- by IC
☐	*SM2320B*	Carbonate
☐	*SM2320B*	Bicarbonate
☐	*9040/150.1*	pH
☐	*160.1*	Solids, Total Dissolved (TDS)
☐	*206.2/7060*	Arsenic, As, by GFAA
☐	*310.1*	Alkalinity
☐	*PAH-SIM*	PAH (NOAA)

0 5004

APCL

Applied Physics & Chemistry Laboratory

13760 Magnolia Ave. Chino CA 91710
Tel. (909) 590-1828 Fax (909) 590-1498

December 09, 2002

SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr. Suite 212
San Diego CA 92127

Dear Yu,

This package contains samples in our Service ID 02-5826 and your project is 00HW019 JPL. Enclosed please find:

(1) One original report.
(2) One original Chain of Custody.
(3) One diskette containing EDD Deliverable.
(4) One original of Level D Data Package Deliverable.

If anything is missing or you have any questions, please feel free to contact me.

Respectfully submitted,

Regina Kirakozova
Associate QA/QC Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Submitted to:
SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr, Ste. 212
San Diego CA 92127
Tel: (858)485-8100 Fax: (858)485-0812

Service ID #: 801-025826A
Collected by: MES/TAM/DJL
Collected on: 10/31/02

Received: 10/31/02
Extracted: N/A
Tested: 11/01-05/02
Reported: 11/11/02

Sample Description: Water
Project Description: 00HW019 JPL

Analysis of Water Samples

Component Analyzed	Method	Unit	PQL	Analysis Result ER-14 02-05826-1	MW-14-1 02-05826-2	MW-14-2 02-05826-3
Dilution Factor				1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	<4
VOLATILE ORGANIC COMPOUNDS						
Dilution Factor				1	1	1
BENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOCHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMODICHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
N-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
SEC-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TERT-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLORODIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	<0.5	0.4J	0.5
CHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
2-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
4-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DIBROMO-3-CHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DIBROMOETHANE (EDB)	524.2	µg/L	0.5	<0.5	<0.5	<0.5
DIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,4-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
DICHLORODIFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	0.4J
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CIS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRANS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result ER-14 02-05826-1	MW-14-1 02-05826-2	MW-14-2 02-05826-3
1,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
2,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CIS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRANS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
ETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
HEXACHLOROBUTADIENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
ISOPROPYLBENZENE (CUMENE)	524.2	µg/L	0.5	<0.5	<0.5	<0.5
P-ISOPROPYLTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	<1	<1	<1
METHYL-T-BUTYL ETHER (MTBE)	524.2	µg/L	1	<1	<1	<1
NAPHTHALENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
N-PROPYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
STYRENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,1,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,2,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	0.5	0.5
TOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,4-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,1-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,2-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	5.3
TRICHLOROFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,4-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3,5-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
VINYL CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
O-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-14-3 02-05826-4	MW-14-4 02-05826-5	MW-14-5 02-05826-6	TB-14 02-05826-7
Dilution Factor				1	1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	<4	-
VOLATILE ORGANIC COMPOUNDS							
Dilution Factor				1	1	1	1
BENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOCHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMODICHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
SEC-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TERT-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLORODIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	0.5	<0.5	<0.5	<0.5
CHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
4-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMO-3-CHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMOETHANE (EDB)	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
DIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,4-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
DICHLORODIFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRANS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRANS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
ETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
HEXACHLOROBUTADIENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-14-3 02-05826-4	MW-14-4 02-05826-5	MW-14-5 02-05826-6	TB-14 02-05826-7
ISOPROPYLBENZENE (CUMENE)	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
P-ISOPROPYLTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	<1	<1	0.3J	<1
METHYL-T-BUTYL ETHER (MTBE)	524.2	µg/L	1	<1	<1	<1	<1
NAPHTHALENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-PROPYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
STYRENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	0.3J	<0.5	<0.5	<0.5
TOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	0.8	<0.5	<0.5	<0.5
TRICHLOROFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3,5-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
VINYL CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
O-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5

Component Analyzed	Method	Unit	PQL	Analysis Result ER-14 02-05826-1	MW-14-1 02-05826-2
CHROMIUM (VI)	7196	mg/L	0.01	<0.01	<0.01

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-14-2 02-05826-3	MW-14-3 02-05826-4	MW-14-4 02-05826-5
CHROMIUM (VI)	7196	mg/L	0.01	<0.01	<0.01	<0.01

PQL: Practical Quantitation Limit. MDL: Method Detection Limit CRDL: Contract Required Detection Limit

N.D.: Not Detected or less than the practical quantitation limit. "-": Analysis is not required.

J: Reported between PQL and MDL.

Listed Dilution Factors (DF) are relative to the method default DF. All unlisted DFs are 1.0

Respectfully submitted,

Dominic Lau
Laboratory Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Submitted to:
SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr, Ste. 212
San Diego CA 92127
Tel: (858)485-8100 Fax: (858)485-0812

Service ID #: 801-025826B
Collected by: MES/TAM/DJL
Collected on: 10/31/02

Received: 10/31/02
Extracted: N/A
Tested: 11/01-05/02
Reported: 11/11/02

Sample Description: Water
Project Description: 00HW019 JPL

Analysis of Water Samples

Component Analyzed	Method	Unit	PQL	Analysis Result ER-14 02-05826-1	MW-14-1 02-05826-2	MW-14-2 02-05826-3
Dilution Factor				1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	<4
VOLATILE ORGANIC COMPOUNDS						
Dilution Factor				1	1	1
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	<0.5	0.4J	0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	0.4J
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	0.5	0.5
TRICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	5.3
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5

Component Analyzed	Method	Unit	PQL	Analysis Result MW-14-3 02-05826-4	MW-14-4 02-05826-5	MW-14-5 02-05826-6	TB-14 02-05826-7
Dilution Factor				1	1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	<4	-
VOLATILE ORGANIC COMPOUNDS							
Dilution Factor				1	1	1	1
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	-	-	0.3J	-
TETRACHLOROETHENE	524.2	µg/L	0.5	0.3J	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	0.8	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result ER-14 02-05826-1	MW-14-1 02-05826-2
CHROMIUM (VI)	7196	mg/L	0.01	<0.01	<0.01

Component Analyzed	Method	Unit	PQL	Analysis Result MW-14-2 02-05826-3	MW-14-3 02-05826-4	MW-14-4 02-05826-5
CHROMIUM (VI)	7196	mg/L	0.01	<0.01	<0.01	<0.01

PQL: Practical Quantitation Limit. MDL: Method Detection Limit CRDL: Contract Required Detection Limit

N.D.: Not Detected or less than the practical quantitation limit. "-": Analysis is not required.

J: Reported between PQL and MDL.

Listed Dilution Factors (DF) are relative to the method default DF. All unlisted DFs are 1.0

Respectfully submitted,

Dominic Lau
Laboratory Director
Applied P & Ch Laboratory

Level D Data Package Deliverables

General Information

Project: 00HW019 JPL

APCL Service ID: 02-5826

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino, CA 91710
Telephone (909)590-1828
Fax (909)590-1498

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Case Narrative

Project: JPL/00HW019

For SOTA Environmental

APCL Service No: 02-5826

1. Sample Identification

The sample identifications are listed in the following table:

SOTA Environmental Sample ID	APCL Sample ID
TB-14	02-05826-7
ER-14	02-05826-1
MW-14-5	02-05826-6
MW-14-4	02-05826-5
MW-14-3	02-05826-4
MW-14-2	02-05826-3
MW-14-1	02-05826-2

2. Analytical Methodology

Samples are analyzed by EPA methods

524.2 (Volatile Organic Compounds),
7196 (Chromium (VI)),
314.0 (Perchlorate, low level),

3. Holding Time

All samples were extracted, digested and analyzed within the holding times defined by the appropriate EPA methods of the analyses.

4. Preservation

All samples were preserved and stored according to the appropriate EPA methods.

5. Tele-log

None

6. Anomaly

None

"I certify that these data are technically accurate, complete, and in compliance with the terms and conditions of the contract, for other than the conditions detailed above. Release of the data contained in the hardcopy data package and its electronic data deliverable submitted on diskette had been authorized by the Laboratory Manager or her/his designee, as verified by the following signature."

Respectfully submitted,

Regina Kirakozova
Associate QA/QC Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory

APCL

13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Chain of Custody

Please Print in pen Page 1 of 2

Client: SOTA Env. Tech. Inc. Contact: Mike Sayre Tel #: 858-485-8100 Fax #: 858-485-0812

Address: 16835 W. Bernardo Dr. #212 City: San Diego State: CA Zip code: 92127

Bill to: SOTA

Project Name/Code JPL Job #00HW019 P.O. #

Project Address Pasadena, CA APCL Quotation #

Due Date: ☒ regular ☐ rush: ___ days ___ hours Sampled by: MES/TAM/DJL

Field Sample ID No.	Sample Description	Date Collected	Time Collected	Sample Matrix	Preservation	# of Containers	VOCs (524.2)	Tot Cr (2008)	Cr VI (7196)	Perchlorate (314.0)	Remarks
TB-14	Trip Blank	10-31-02	1222	Water	HCl	2	X				EPA Level IV QA/QC
ER-14	Equipment Rinsate		1215		HCl	3	X				
					HNO3	1		X			
					—	1			X	X	
MW-BH-5	MW-14-5		0921		HCl	3	X				
MW-14-5	MW-14-5		0921		—	1				X	
MW-14-4	MW-14-4		1001		HCl	3	X				
					HNO3	1		X			
					—	1			X	X	
MW-14-3	MW-14-3		1108		HCl	3	X			DJL	
					HNO3	1		X			
					—	1			X	X	
MW-14-2	MW-14-2		1141		HCl	3	X				
					HNO3	1		X			
					—	1			X	X	

5826

White – With report
Yellow – Lab copy
Pink – Originator

QC Requirement: ☐ Regular; ☐ QA/QC Report; ☐ WIP; ☐ Raw Data; ☐ Extended Raw Data ☐ CLP; ☐ ACE ☐ AFCEE ☐ NEESA ___ (E, C or D); ☐ Other _____ (Please specify)

Sample Disposal: ☐ Return ☐ Disposal by APCL ☐ Hold for _____ days after receiving date. **If not specified, samples will be discarded 45 days after samples are received.**

Sample Conditions: ☐ Intact; ☐ Broken. Cooler Seal: ☐ Intact; ☐ Broken; ☐ None . Tag # _____ Temperature: ☐ Room ☐ Cold (_____ °C).

Relinquished by [signature] Date/Time 10/31/02 1825 Received by [signature illegible] Date/Time [illegible]

Relinquished by [signature] Date/Time 10/31/02 1928 Received by [signature] Date/Time 10/31/02 1928

APCL USE ONLY / Service # Note:

Clients understand that all terms described in the proposals, quotations for this project, and/or the general terms provided in the current APCL price schedules will be followed. APCL reserves the right to terminate its service or withhold delivery of any reports, if in APCL's sole discretion the terms of the project have been broken.

Applied P & Ch Laboratory

APCL

13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Chain of Custody

Please Print in pen Page 2 of 2

Client: SOTA Env. Tech. Inc. Contact: Mike Sayre Tel #: 858-485-8100 Fax #: 858-485-0812

Address: 16835 W. Bernardo Dr. #212 City: San Diego State: CA Zip code: 92127

Bill to: SOTA

Project Name/Code JPL Job # 00HW019 P.O. #

Project Address Pasadena, CA APCL Quotation #

Due Date: ☒ regular ☐ rush: ___ days ___ hours Sampled by: MES/TAM/DJL

White – With report
Yellow – Lab copy
Pink – Originator

Field Sample ID No.	Sample Description	Date Collected	Time Collected	Sample Matrix	Preservation	# of Containers	VOCs (524.2)	Total Cr (200.8)	Cr VI (7196)	Perchlorate (314.0)	Remarks
MW-14-1	MW-14-1	10-31-02	1215	Water	HCl	3					EPA Level III
"	"	"	"	"	HNO3	1					QA/QC
"	"	"	"	"	—	1					"

Analysis Items

5826

QC Requirement: ☐ Regular; ☐ QA/QC Report; ☐ WIP; ☐ Raw Data; ☐ Extended Raw Data ☐ CLP; ☐ ACE ☐ AFCEE ☐ NEESA ___ (E, C or D); ☐ Other _____ (Please specify)

Sample Disposal: ☐ Return ☐ Disposal by APCL ☐ Hold for _____ days after receiving date. **If not specified, samples will be discarded 45 days after samples are received.**

Sample Conditions: ☐ Intact; ☐ Broken. Cooler Seal: ☐ Intact; ☐ Broken; ☐ None. Tag # _____ Temperature: ☐ Room ☐ Cold (_____ °C).

Relinquished by [signature] Date/Time 10/31/02 1825 Received by [signature] Date/Time [illegible]

Relinquished by [signature] Date/Time 10/31/02 1928 Received by [signature] Date/Time 10/31/02 1928

APCL USE ONLY Service # Note:

Clients understand that all terms described in the proposals, quotations for this project, and/or the general terms provided in the current APCL price schedules will be followed. APCL reserves the right to terminate its service or withhold delivery of any reports, if in APCL's sole discretion the terms of the project have been broken.

Applied P & Ch Laboratory
13760 Magnolia Ave., Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Sample Receiving Checklist

APCL ServiceID: 5826 **Client Name/Project:** Sota Environmental

1. Sample Arrival

Date/Time Received 10/31/02 1928 Date/Time Opened 10/31/02 1928 By (name): Kenny Chan

Custody Transfer: ☐ Client ☐ Golden State ☐ UPS ☐ US Mail ☐ FedEx ☑ APCL Empl: Richard

2. Chain-of-Custody (CoC)

☑ With Samples? ☐ Faxed? ☐ Client has Copy? ☐ Signed, dated? By: _____

☐ Project ID? ☑ Analyses Clear? ☐ Hold Samples? #on Hold _____ # Received 6

☐ CoC/Docs Zip-Locked under lid? ☐ Compos #: __ ☑ #Samples OK? _____

☐ Discrepancies? ☐ Client notified? ☐ Response (attach docs): _____

3. Shipping Container/Cooler

☑ Cooler Used? # of 1 Cooled by: ☑ Ice ☐ Blue Ice ☐ Dry Ice ☐ None

Temp °C 3.7 _____ _____ _____ _____ _____

(Cooler temperature measured from temp blank if present, otherwise measured from the cooler).

Cooler Custody Seal? ☐ Absent ☐ Intact ☐ Tampered?

4. Sample Preservation

☐ **pH <2** ☐ pH >12

If Not, pH = _____ Preserved by: ☐ Client ☐ APCL ☐ Third Party _____

5. Holding-time Requirements

☐ pH 24hr ☐ BACT 6/24hr ☑ Cr^{VI} 24hr ☐ NO_3^- 48hr ☐ BOD 48hr

☐ Cl_2 ASAP ☐ Turbidity 48hr ☐ DO ASAP ☐ Fe(II) ASAP

☐ HT Expired? ☐ Client notified?

6. Sample Container Condition

☑ Intact? ☐ Broken? ☐ Documented? Number: _____

Type: ☑ plastic ☑ glass ☐ Tube: brass/SS ☐ Tedlar Bag

☑ Quantity OK? ☐ Leaking? ☐ Anomaly?

☑ Caps tight? ☐ Air Bubbles? ☐ Anomaly?

Labels: ☐ Unique ID? ☑ Date/Time ☐ Preserved?

7. Turn Around Time

☐ RUSH TAT: _____ ☑ Std (7-10 days) ☐ Not Marked

8. Sample Matrix

☐ Drinking H_2O ☑ Other Liq ☐ Soil ☐ Wipe ☐ Polymer ☐ Air ☐ Other: _____

☐ Ground H_2O ☐ Sludge ☐ Filter ☐ Oil/Petro ☐ Paint ☐ W. Water ☐ Extract ☐ Unknown

9. Pre-Login Check List Completed & OK?

☑ ALL OK? (if not, attach docs) ☐ Client Contact? (Name:_____)Date/Time:_____

Received/Checked by: _____ Date: 31 Oct 2002 Time: 8:48 a.m.

***HT:** Samples must be analyzed for results to reflect total concentrations. Results generated outside required of holding times are considered minimal values and may be used to define waste as hazardous but not as non-hazardous

DocumentFile: {neal.texfiles}smprcl.tex

017062

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Sample Login: Check List

02-05826 (1288_ 412) (4858100_ 412)

11/01/02

Part 1: General Information

☐	Company Information	Name:	*SCTA Environmental*
		Address:	*16835 W. Bernardo Dr, Ste. 212 ,San Diego ,CA 92127*
☐	Project Information	Project Description:	*JIL*
		Project #:	*06HW019*
☐	Billing Information	P.O. #:	
		Bill Address:	*16835 W. Bernardo Dr, Ste. 212 ,San Diego ,CA 92127*
		Lab Project ID:	*2002_0002*
		Client Database #:	*01*
☐	Receiving Information	Who Received Sample?	*Kenny Chan*
		Receiving Date/Time:	*10/31/02 1928*
		COC No.	
☐	Shipping Information	Shipping Company	*APCL pick up*
		Packing Information:	*Cooler/Ice Chester*
		Cooler Temperature:	*3.7 °C*
☐	Container Information	Container Provider:	*Client*
☐	Sampling Information	Sampling Person:	
		Sampling Company:	*Client*
☐	Turn-Around-Time Option:		*Normal*
☐	QC Option:		*NEESA D*
☐	Disposal Option:		*Not specify*

017003

Part 2: Sample Information

Seq. #	Sample ID (on COC)	Sample Sub-ID	APCL Sample ID	Matrix	Container	Preservative	Vol, ml Am	g	# of Replica	Condition G, L, B	Collected mmddyy	Hold ?	Composite Group	TAT Days	
1	TB-14	VOC	02-05826-7	W	V	C	40		2	G	103102	N	0	9	☐
2	ER-14	VOC	02-05826-1-α	W	V	C	40		3	G	103102	N	0	9	☐
	ER-14	CRVI/Perch	02-05826-1-β	W	P		500		1	G	103102	N	0	9	☐
3	MW-14-5	VOC	02-05826-6-α	W	V	C	40		3	G	103102	N	0	9	☐
	MW-14-5	Perch	02-05826-6-β	W	P		500		1	G	103102	N	0	9	☐
4	MW-14-4	VOC	02-05826-5-α	W	V	C	40		3	G	103102	N	0	9	☐
	MW-14-4	CRVI/Perch	02-05826-5-β	W	P		500		1	G	103102	N	0	9	☐
5	MW-14-3	VOC	02-05826-4-α	W	V	C	40		3	G	103102	N	0	9	☐
	MW-14-3	CRVI/Perch	02-05826-4-β	W	P		500		1	G	103102	N	0	9	☐
6	MW-14-2	VOC	02-05826-3-α	W	V	C	40		3	G	103102	N	0	9	☐
	MW-14-2	CRVI/Perch	02-05826-3-β	W	P		500		1	G	103102	N	0	9	☐
7	MW-14-1	VOC	02-05826-2-α	W	V	C	40		3	G	103102	N	0	9	☐
	MW-14-1	CRVI/Perch	02-05826-2-β	W	P		500		1	G	103102	N	0	9	☐

Part 3: Analysis Information

Test Items:

☐	*524.2*	Volatile Organic Compounds
☐	*7196*	Chromium (VI)
☐	*300.0*	Perchlorate, low level
☐	*200.7/6010*	Sodium, Na, by ICP
☐	*200.7/6010*	Potassium, K, by ICP
☐	*200.7/6010*	Calcium, Ca, by ICP
☐	*200.7/6010*	Magnesium, Mg, by ICP
☐	*200.7/6010*	Iron, Fe, by ICP
☐	*300.0*	Sulfate (SO_4^{--}), by IC
☐	*300.0/SM4500NO3*	Nitrate (NO_3^-) as N by IC
☐	*300.0*	Chloride Cl^- by IC
☐	*SM2320B*	Carbonate
☐	*SM2320B*	Bicarbonate
☐	*9040/150.1*	pH
☐	*160.1*	Solids, Total Dissolved (TDS)
☐	*206.2/7060*	Arsenic, As, by GFAA
☐	*310.1*	Alkalinity
☐	*PAH-SIM*	PAH (NOAA)

017004

APCL

Applied Physics & Chemistry Laboratory

13760 Magnolia Ave. Chino CA 91710
Tel. (909) 590-1828 Fax (909) 590-1498

December 10, 2002

SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr. Suite 212
San Diego CA 92127

Dear Yu,

This package contains samples in our Service ID 02-5832 and your project is 00HW019 JPL. Enclosed please find:

(1) One original report.
(2) One original Chain of Custody.
(3) One diskette containing EDD Deliverable.
(4) One original of Level D Data Package Deliverable.

If anything is missing or you have any questions, please feel free to contact me.

Respectfully submitted,

Regina Kirakozova
Associate QA/QC Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Submitted to:
SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr. Ste. 212
San Diego CA 92127
Tel: (858)485-8100 Fax: (858)485-0812

APCL Analytical Report

Service ID #: 801-025832A
Collected by: TAM/DJL
Collected on: 11/01/02

Received: 11/01/02
Extracted: N/A
Tested: 11/01-13/02
Reported: 11/13/02

Sample Description: Water from Pasadena, CA.
Project Description: 00HW019 JPL

Analysis of Water Samples

Component Analyzed	Method	Unit	PQL	Analysis Result ER-17 02-05832-1	MW-17-2 02-05832-2	MW-17-3 02-05832-3
Dilution Factor				1	1	4
PERCHLORATE	314.0	µg/L	4	<4	<4	115
VOLATILE ORGANIC COMPOUNDS						
Dilution Factor				1	1	1
BENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOCHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMODICHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
N-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
SEC-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TERT-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	7.1
CHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLORODIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	<0.5	<0.5	2.7
CHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
2-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
4-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DIBROMO-3-CHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DIBROMOETHANE (EDB)	524.2	µg/L	0.5	<0.5	<0.5	<0.5
DIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,4-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
DICHLORODIFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CIS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRANS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result ER-17 02-05832-1	MW-17-2 02-05832-2	MW-17-3 02-05832-3
1,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
2,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CIS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRANS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
ETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
HEXACHLOROBUTADIENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
ISOPROPYLBENZENE (CUMENE)	524.2	µg/L	0.5	<0.5	<0.5	<0.5
P-ISOPROPYLTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	<1	<1	<1
METHYL-T-BUTYL ETHER (MTBE)	524.2	µg/L	1	<1	<1	<1
NAPHTHALENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
N-PROPYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
STYRENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,1,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,2,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,4-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,1-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,2-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	<0.5	0.8	2.6
TRICHLOROFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,4-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3,5-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
VINYL CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
O-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	0.4J	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-17-4 02-05832-4	MW-17-5 02-05832-5	TB-17 02-05832-6
Dilution Factor				1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	-
VOLATILE ORGANIC COMPOUNDS						
Dilution Factor				1	1	1
BENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOCHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMODICHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
N-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
SEC-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TERT-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLORODIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	0.4J	<0.5	<0.5
CHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
2-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
4-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DIBROMO-3-CHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DIBROMOETHANE (EDB)	524.2	µg/L	0.5	<0.5	<0.5	<0.5
DIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,4-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
DICHLORODIFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CIS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRANS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
2,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CIS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRANS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
ETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
HEXACHLOROBUTADIENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
ISOPROPYLBENZENE (CUMENE)	524.2	µg/L	0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-17-4 02-05832-4	MW-17-5 02-05832-5	TB-17 02-05832-6
P-ISOPROPYLTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	<1	<1	0.4J
METHYL-T-BUTYL ETHER (MTBE)	524.2	µg/L	1	<1	<1	<1
NAPHTHALENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
N-PROPYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
STYRENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,1,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,2,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,4-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,1-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,2-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	3.0	1.0	<0.5
TRICHLOROFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,4-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3,5-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
VINYL CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
O-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5

Component Analyzed	Method	Unit	PQL	Analysis Result ER-17 02-05832-1	MW-17-2 02-05832-2	MW-17-3 02-05832-3	MW-17-4 02-05832-4
CHROMIUM (VI)	7196	mg/L	0.01	<0.01	<0.01	<0.01	0.045

PQL: Practical Quantitation Limit. MDL: Method Detection Limit. CRDL: Contract Required Detection Limit
N.D.: Not Detected or less than the practical quantitation limit. "-": Analysis is not required.
J: Reported between PQL and MDL.
Listed Dilution Factors (DF) are relative to the method default DF. All unlisted DFs are 1.0

Respectfully submitted,

Dominic Lau
Laboratory Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Submitted to:
SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr, Ste. 212
San Diego CA 92127
Tel: (858)485-8100 Fax: (858)485-0812

Service ID #: 801-025832B
Collected by: TAM/DJL
Collected on: 11/01/02

Received: 11/01/02
Extracted: N/A
Tested: 11/01-13/02
Reported: 11/13/02

Sample Description: Water from Pasadena, CA.
Project Description: 00HW019 JPL

Analysis of Water Samples

Component Analyzed	Method	Unit	PQL	Analysis Result ER-17 02-05832-1	MW-17-2 02-05832-2	MW-17-3 02-05832-3
Dilution Factor				1	1	4
PERCHLORATE	314.0	µg/L	4	<4	<4	115
VOLATILE ORGANIC COMPOUNDS						
Dilution Factor				1	1	1
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	7.1
CHLOROFORM	524.2	µg/L	0.5	<0.5	<0.5	2.7
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	<0.5	0.8	2.6
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	0.4J	-	-

Component Analyzed	Method	Unit	PQL	Analysis Result MW-17-4 02-05832-4	MW-17-5 02-05832-5	TB-17 02-05832-6
Dilution Factor				1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	-
VOLATILE ORGANIC COMPOUNDS						
Dilution Factor				1	1	1
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	0.4J	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	-	-	0.4J
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	3.0	1.0	-
TRICHLOROFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result ER-17 02-05832-1	MW-17-2 02-05832-2	MW-17-3 02-05832-3	MW-17-4 02-05832-4
CHROMIUM (VI)	7196	mg/L	0.01	<0.01	<0.01	<0.01	0.045

PQL: Practical Quantitation Limit. MDL: Method Detection Limit. CRDL: Contract Required Detection Limit
N.D.: Not Detected or less than the practical quantitation limit. "-": Analysis is not required.
J: Reported between PQL and MDL.
Listed Dilution Factors (DF) are relative to the method default DF. All unlisted DFs are 1.0

Respectfully submitted,

Dominic Lau
Laboratory Director
Applied P & Ch Laboratory

Level D Data Package Deliverables

General Information

Project: 00HW019 JPL

APCL Service ID: 02-5832

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino, CA 91710
Telephone (909)590-1828
Fax (909)590-1498

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Case Narrative

Project: JPL/Pasadena, CA./00HW019

For SOTA Environmental

APCL Service No: 02-5832

1. Sample Identification

The sample identifications are listed in the following table:

SOTA Environmental Sample ID	APCL Sample ID
TB-17	02-05832-6
MW-17-5	02-05832-5
MW-17-4	02-05832-4
MW-17-3	02-05832-3
MW-17-2	02-05832-2
ER-17	02-05832-1

2. Analytical Methodology

Samples are analyzed by EPA methods

524.2 (Volatile Organic Compounds),
7196 (Chromium (VI)),
314.0 (Perchlorate, low level),

3. Holding Time

All samples were extracted, digested and analyzed within the holding times defined by the appropriate EPA methods of the analyses.

4. Preservation

All samples were preserved and stored according to the appropriate EPA methods.

5. Tele-log

None

6. Anomaly

None

"I certify that these data are technically accurate, complete, and in compliance with the terms and conditions of the contract, for other than the conditions detailed above. Release of the data contained in the hardcopy data package and its electronic data deliverable submitted on diskette had been authorized by the Laboratory Manager or her/his designee, as verified by the following signature."

Respectfully submitted,

Regina Kirakozova
Associate QA/QC Director
Applied P & Ch Laboratory

019301

Applied P & Ch Laboratory

APCL

13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Chain of Custody

Please Print in pen Page 1 of 1

Client: SOTA Env. Tech., Inc Contact: Mike Sayre Tel #: 858-485-8100 Fax #: 858-485-0812

Address: 16835 W. Bernardo Dr. #212 City: San Diego State: CA Zip code: 92127

Bill to: SOTA

Project Name/Code JPL Job #00HW019 P.O. #

Project Address Pasadena, CA APCL Quotation #

Due Date: ☒regular ☐rush: ___ days ___ hours Sampled by: TAM / DSL

Field Sample ID No.	Sample Description	Date Collected	Time	Sample Matrix	Preservation	# of Containers	VOCs (524.2)	Tot.Cr (200.8)	CrIII (7196)	Perchlorate (314.0)	Remarks
TB-17	Trip Blank	11-1-02	1152	Water	HCl	2	X				EPA Level IV
MW-17-5	MW-17-5		1052		HCl	3	X				
MW-17-5	MW-17-5		1052		—	1				X	QA/QC
MW-17-4	MW-17-4		1124		HCl	9	X				MS/MSD
					HNO3	2		X			
					—	2			X	X	
MW-17-3	MW-17-3		1224		HCl	3	X				
					HNO3	1		X			
					—	1			X	X	
MW-17-2	MW-17-2		1251		HCl	3	X				
					HNO3	1		X			
					—	1			X	X	
ER-17 ~~MW-1 DSL~~	Equip. Rinsate		1150		HCl	3	X				
ER-17					HNO3	1		X			
ER-17					—	1			X	X	

5832

White – With report
Yellow – Lab copy
Pink – Originator

QC Requirement: ☐Regular; ☐QA/QC Report; ☐WIP; ☐Raw Data; ☐Extended Raw Data ☐CLP; ☐ACE ☐AFCEE ☐NEESA ___ (E, C or D); ☐Other _____ (Please specify)

Sample Disposal: ☐Return ☐Disposal by APCL ☐Hold for _____ days after receiving date. **If not specified, samples will be discarded 45 days after samples are received.**

Sample Conditions: ☐Intact; ☐Broken. Cooler Seal: ☐Intact; ☐Broken; ☐None. Tag # _____ Temperature: ☐Room ☐Cold (_____ °C).

Relinquished by Dan Cohn Date/Time 11-1-02 / 1425 Received by [signature] Date/Time 11/1/02 / 1425

Relinquished by Date/Time / Received by Date/Time /

APCL USE ONLY Service # Note:

Clients understand that all terms described in the proposals, quotations for this project, and/or the general terms provided in the current APCL price schedules will be followed. APCL reserves the right to terminate its service or withhold delivery of any reports, if in APCL's sole discretion the terms of the project have been broken.

Applied P & Ch Laboratory
13760 Magnolia Ave., Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Sample Receiving Checklist

APCL ServiceID: 5832 **Client Name/Project:** Soto Environmental

1. Sample Arrival

Date/Time Received 11/1/02 1425 Date/Time Opened 11/1/02 1425 By (name): Henry Chan
Custody Transfer: ☒ Client ☐ Golden State ☐ UPS ☐ US Mail ☐ FedEx ☐ APCL Empl: _____

2. Chain-of-Custody (CoC)

☒ With Samples? ☐ Faxed? ☒ Client has Copy? ☐ Signed, dated? By: _____
☒ Project ID? ☒ Analyses Clear? ☐ Hold Samples? #on Hold _____ # Received 6
☒ CoC/Docs Zip-Locked under lid? ☐ Compos #: _ ☐ #Samples OK? _____
☐ Discrepancies? ☐ Client notified? ☐ Response (attach docs): _____

3. Shipping Container/Cooler

☒ Cooler Used? # of 1 Cooled by: ☒ Ice ☐ Blue Ice ☐ Dry Ice ☐ None
Temp °C 3.6 _____ _____ _____ _____ _____
(Cooler temperature measured from temp blank if present, otherwise measured from the cooler).
Cooler Custody Seal? ☐ Absent ☐ Intact ☐ Tampered?

4. Sample Preservation

☐ **pH <2** ☐ pH >12
If Not, pH = _____ Preserved by: ☐ Client ☐ APCL ☐ Third Party _____

5. Holding-time Requirements

☐ pH 24hr ☐ BACT 6/24hr ☐ Cr^{VI} 24hr ☐ NO_3^- 48hr ☐ BOD 48hr
☐ Cl_2 ASAP ☐ Turbidity 48hr ☐ DO ASAP ☐ Fe(II) ASAP
☐ HT Expired? ☐ Client notified?

6. Sample Container Condition

☒ Intact? ☐ Broken? ☐ Documented? Number: _____
Type: ☒ plastic ☒ glass ☐ Tube: brass/SS ☐ Tedlar Bag
☒ Quantity OK? ☐ Leaking? ☐ Anomaly?
☒ Caps tight? ☐ Air Bubbles? ☐ Anomaly?
Labels: ☒ Unique ID? ☒ Date/Time ☐ Preserved?

7. Turn Around Time

☐ RUSH TAT: _____ ☒ Std (7-10 days) ☐ Not Marked

8. Sample Matrix

☐ Drinking H_2O ☒ Other Liq ☐ Soil ☐ Wipe ☐ Polymer ☐ Air ☐ Other: _____
☐ Ground H_2O ☐ Sludge ☐ Filter ☐ Oil/Petro ☐ Paint ☐ W. Water ☐ Extract ☐ Unknown

9. Pre-Login Check List Completed & OK?

☒ ALL OK? (if not, attach docs) ☐ Client Contact? (Name: _____)Date/Time: _____

Received/Checked by: [signature] Date: 1 Nov 2002 Time: 10:02 a.m.

*HT: Samples must be analyzed for results to reflect total concentrations. Results generated outside required of holding times are considered minimal values and may be used to define waste as hazardous but not as non-hazardous.

DocumentFile: [neal.texfiles]smprcl.tex.

019302

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Sample Login: Check List

02-05832 (1288_ 414) (4858100_ 414)

11/01/02

Part 1: General Information

☐	Company Information	Name:	*SOTA Environmental*
		Address:	*16835 W. Bernardo Dr, Ste. 212 ,San Diego ,CA 92127*
☐	Project Information	Project Description:	*JPL*
		Project #:	*00HW019*
☐	Billing Information	P.O. #:	
		Bill Address:	*16835 W. Bernardo Dr, Ste. 212 ,San Diego ,CA 92127*
		Lab Project ID:	*2002_0002*
		Client Database #:	*01*
☐	Receiving Information	Who Received Sample?	*Kenny Chan*
		Receiving Date/Time:	*11/01/02 1425*
		COC No.	
☐	Shipping Information	Shipping Company	*by Client*
		Packing Information:	*Cooler/Ice Chester*
		Cooler Temperature:	*3.6 °C*
☐	Container Information	Container Provider:	*Client*
☐	Sampling Information	Sampling Person:	
		Sampling Company:	*Client*
☐	Turn-Around-Time Option:		*Rush 5 working day(s)*
☐	QC Option:		*NEESA D*
☐	Disposal Option:		*Not specify*

019303

Part 2: Sample Information

Seq. #	Sample ID (on COC)	Sample Sub-ID	APCL Sample ID	Matrix	Cont-tainer	Preser-vative	Vol, ml Am. g	# of Replica	Condition G, L, B	Collected mmddyy	Hold ?	Composite Group	TAT Days	
1	TB-17	VOC	02-05832-6	W	V	C	40	2	G	110102	N	0	7	☐
2	MW-17-5	VOC	02-05832-5-α	W	V	C	40	3	G	110102	N	0	7	☐
	MW-17-5	Perchl	02-05832-5-β	W	P		500	1	G	110102	N	0	7	☐
3	MW-17-4	VOC	02-05832-4-α	W	V	C	40	9	G	110102	N	0	7	☐
	MW-17-4	CRVI/Perch	02-05832-4-β	W	P		500	2	G	110102	N	0	7	☐
4	MW-17-3	VOC	02-05832-3-α	W	V	C	40	3	G	110102	N	0	7	☐
	MW-17-3	CRVI/Perch	02-05832-3-β	W	P		500	1	G	110102	N	0	7	☐
5	MW-17-2	VOC	02-05832-2-α	W	V	C	40	3	G	110102	N	0	7	☐
	MW-17-2	CRVI/Perch	02-05832-2-β	W	P		500	1	G	110102	N	0	7	☐
6	ER-17	VOC	02-05832-1-α	W	V	C	40	3	G	110102	N	0	7	☐
	ER-17	CRVI/Perch	02-05832-1-β	W	P		500	1	G	110102	N	0	7	☐

Part 3: Analysis Information

Test Items:

☐	*524.2*	Volatile Organic Compounds
☐	*7196*	Chromium (VI)
☐	*300.0*	Perchlorate, low level
☐	*200.7/6010*	Sodium, Na, by ICP
☐	*200.7/6010*	Potassium, K, by ICP
☐	*200.7/6010*	Calcium, Ca, by ICP
☐	*200.7/6010*	Magnesium, Mg, by ICP
☐	*200.7/6010*	Iron, Fe, by ICP
☐	*300.0*	Sulfate (SO_4^{--}), by IC
☐	*300.0/SM4500NO3*	Nitrate (NO_3^-) as N by IC
☐	*300.0*	Chloride Cl^- by IC
☐	*SM2320B*	Carbonate
☐	*SM2320B*	Bicarbonate
☐	*9040/150.1*	pH
☐	*160.1*	Solids, Total Dissolved (TDS)
☐	*206.2/7060*	Arsenic, As, by GFAA
☐	*310.1*	Alkalinity
☐	*PAH-SIM*	PAH (NOAA)

Seq. #	Client"s Sample ID (as given on COC)	Sample Sub-ID	APCL Sample ID	Matrix	524.2	CHROMIUM	PERCHL	NA	K	CA	MG	FE

019304

APCL

Applied Physics & Chemistry Laboratory

13760 Magnolia Ave. Chino CA 91710
Tel. (909) 590-1828 Fax (909) 590-1498

December 10, 2002

SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr. Suite 212
San Diego CA 92127

Dear Yu,

This package contains samples in our Service ID 02-5853 and your project is 00HW019 JPL. Enclosed please find:

(1) One original report.
(2) One original Chain of Custody.
(3) One diskette containing EDD Deliverable.
(4) One original of Level D Data Package Deliverable.

If anything is missing or you have any questions, please feel free to contact me.

Respectfully submitted,

Regina Kirakozova
Associate QA/QC Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Submitted to:
SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr, Ste. 212
San Diego CA 92127
Tel: (858)485-8100 Fax: (858)485-0812

Service ID #: 801-025853B
Collected by: TAM/DJL
Collected on: 11/04/02

Received: 11/04/02
Extracted: N/A
Tested: 11/04-12/02
Reported: 11/13/02

Sample Description: Water from Pasadena, CA.
Project Description: 00HW019 JPL

Analysis of Water Samples

Component Analyzed	Method	Unit	PQL	Analysis Result ER-18 02-05853-1	MW-18-2 02-05853-2	MW-18-3 02-05853-3
Dilution Factor				1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	<4
VOLATILE ORGANIC COMPOUNDS						
Dilution Factor				1	1	1
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	<0.5	<0.5	1.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	0.4J
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	0.3J	-	-

Component Analyzed	Method	Unit	PQL	Analysis Result MW-18-4 02-05853-4	MW-18-5 02-05853-5	TB-18 02-05853-6
Dilution Factor				1	1	1
PERCHLORATE	314.0	µg/L	4	28.2	<4	-
VOLATILE ORGANIC COMPOUNDS						
Dilution Factor				1	1	1
CARBON TETRACHLORIDE	524.2	µg/L	0.5	4.9	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	1.3	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	-	-	0.6J
TETRACHLOROETHENE	524.2	µg/L	0.5	3.7	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	2.4	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result ER-18 02-05853-1	MW-18-2 02-05853-2	MW-18-3 02-05853-3	MW-18-4 02-05853-4
CHROMIUM (VI)	7196	mg/L	0.01	< 0.01	< 0.01	< 0.01	< 0.01

PQL: Practical Quantitation Limit. MDL: Method Detection Limit CRDL: Contract Required Detection Limit
N.D.: Not Detected or less than the practical quantitation limit. "-": Analysis is not required.
J: Reported between PQL and MDL.
Listed Dilution Factors (DF) are relative to the method default DF. All unlisted DFs are 1.0

Respectfully submitted,

Dominic Lau
Laboratory Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Submitted to:
SOTA Environmental
Attention: Yu Zeng
16835 W. Bernardo Dr, Ste. 212
San Diego CA 92127
Tel: (858)485-8100 Fax: (858)485-0812

Service ID #: 801-025853A
Collected by: TAM/DJL
Collected on: 11/04/02

Received: 11/04/02
Extracted: N/A
Tested: 11/04-12/02
Reported: 11/13/02

Sample Description: Water from Pasadena, CA.
Project Description: 00HW019 JPL

Analysis of Water Samples

Component Analyzed	Method	Unit	PQL	Analysis Result ER-18 02-05853-1	MW-18-2 02-05853-2	MW-18-3 02-05853-3
Dilution Factor				1	1	1
PERCHLORATE	314.0	µg/L	4	<4	<4	<4
VOLATILE ORGANIC COMPOUNDS						
Dilution Factor				1	1	1
BENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOCHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMODICHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
N-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
SEC-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TERT-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLORODIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	<0.5	<0.5	1.5
CHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
2-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
4-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DIBROMO-3-CHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DIBROMOETHANE (EDB)	524.2	µg/L	0.5	<0.5	<0.5	<0.5
DIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,4-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
DICHLORODIFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CIS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRANS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result ER-18 02-05853-1	MW-18-2 02-05853-2	MW-18-3 02-05853-3
1,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3-DICHLOROPROPANE	524.2	µg/L	0.5	0.6	<0.5	<0.5
2,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CIS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRANS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
ETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
HEXACHLOROBUTADIENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
ISOPROPYLBENZENE (CUMENE)	524.2	µg/L	0.5	<0.5	<0.5	<0.5
P-ISOPROPYLTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	<1	<1	<1
METHYL-T-BUTYL ETHER (MTBE)	524.2	µg/L	1	<1	<1	<1
NAPHTHALENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
N-PROPYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
STYRENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,1,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,2,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,4-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,1-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,2-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	0.4J
TRICHLOROFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,4-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3,5-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
VINYL CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
O-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	0.3J	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-18-4 02-05853-4	MW-18-5 02-05853-5	TB-18 02-05853-6
Dilution Factor				1	1	1
PERCHLORATE	314.0	µg/L	4	28.2	<4	-
VOLATILE ORGANIC COMPOUNDS						
Dilution Factor				1	1	1
BENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOCHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMODICHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5
BROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
N-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
SEC-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TERT-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CARBON TETRACHLORIDE	524.2	µg/L	0.5	4.9	<0.5	<0.5
CHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLORODIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	1.3	<0.5	<0.5
CHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
2-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
4-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DIBROMO-3-CHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DIBROMOETHANE (EDB)	524.2	µg/L	0.5	<0.5	<0.5	<0.5
DIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,4-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
DICHLORODIFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CIS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRANS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
2,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
CIS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRANS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
ETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
HEXACHLOROBUTADIENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
ISOPROPYLBENZENE (CUMENE)	524.2	µg/L	0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-18-4 02-05853-4	MW-18-5 02-05853-5	TB-18 02-05853-6
P-ISOPROPYLTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	1	<1	<1	0.6J
METHYL-T-BUTYL ETHER (MTBE)	524.2	µg/L	1	<1	<1	<1
NAPHTHALENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
N-PROPYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
STYRENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,1,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,2,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	3.7	<0.5	<0.5
TOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,4-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,1-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,1,2-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	2.4	<0.5	<0.5
TRICHLOROFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,2,4-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
1,3,5-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
VINYL CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
O-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5

Component Analyzed	Method	Unit	PQL	Analysis Result ER-18 02-05853-1	MW-18-2 02-05853-2	MW-18-3 02-05853-3	MW-18-4 02-05853-4
CHROMIUM (VI)	7196	mg/L	0.01	<0.01	<0.01	<0.01	<0.01

PQL: Practical Quantitation Limit. MDL: Method Detection Limit CRDL: Contract Required Detection Limit
N.D.: Not Detected or less than the practical quantitation limit. "-": Analysis is not required.
J: Reported between PQL and MDL.
Listed Dilution Factors (DF) are relative to the method default DF. All unlisted DFs are 1.0

Respectfully submitted,

Dominic Lau
Laboratory Director
Applied P & Ch Laboratory

Level D Data Package Deliverables

General Information

Project: 00HW019 JPL

APCL Service ID: 02-5853

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino, CA 91710
Telephone (909)590-1828
Fax (909)590-1498

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Case Narrative

Project: JPL/00HW019

For SOTA Environmental

APCL Service No: 02-5853

1. Sample Identification

The sample identifications are listed in the following table:

SOTA Environmental Sample ID	APCL Sample ID
TB-18	02-05853-6
ER-18	02-05853-1
MW-18-5	02-05853-5
MW-18-4	02-05853-4
MW-18-3	02-05853-3
MW-18-2	02-05853-2

2. Analytical Methodology

Samples are analyzed by EPA methods

524.2 (Volatile Organic Compounds),
7196 (Chromium (VI)),
314.0 (Perchlorate, low level),

3. Holding Time

All samples were extracted, digested and analyzed within the holding times defined by the appropriate EPA methods of the analyses.

4. Preservation

All samples were preserved and stored according to the appropriate EPA methods.

5. Tele-log

None

6. Anomaly

None

019000

"I certify that these data are technically accurate, complete, and in compliance with the terms and conditions of the contract, for other than the conditions detailed above. Release of the data contained in the hardcopy data package and its electronic data deliverable submitted on diskette had been authorized by the Laboratory Manager or her/his designee, as verified by the following signature."

Respectfully submitted,

Regina Kirakozova
Associate QA/QC Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory

APCL

13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Chain of Custody

DJL

Please Print in pen Page 1 of 2

Client: SOTA Env. Tech., Inc Contact: Mike Sayre Tel #: 858-485-8100 Fax #: 858-485-0812

Address: 16835 W. Bernardo Dr. #212 City: San Diego State: CA, 92127 Zip code:

Bill to: SOTA

Project Name/Code JPL Job #00HW019 P.O. #

Project Address Pasadena, CA APCL Quotation #

Due Date: ☒regular ☐rush: ___ days ___ hours Sampled by: TAM/DJL

Field Sample ID No.	Sample Description	Date Collected	Time Collected	Sample Matrix	Preservation	# of Containers	VOCs (524.2)	Tot. Cr (200.8)	Cr VI (7196)	Perchlorate (314.0)	Remarks
TB-18	Trip Blank	11/4/02	1034	Water	HCl	2	X				EPA Level IV QA/QC
EA-18	Equip. Rinsate		1120		HCl	3	X				
"	"		1120		HNO3	1		X			
"	"		1120		—	1			X	X	
MW-18-5	MW-18-5		1015		HCl	3	X				
MW-18-5	MW-18-5		1015		—	1				X	
MW-18-4	MW-18-4		1047		HCl	3	X				
MW-18-4	MW-18-4		"		HNO3	1		X			
MW-18-4	MW-18-4		"		—	1			X	X	
MW-18-3	MW-18-3		1037		HCl	3	X				
"	"		"		HNO3	1		X			
"	"		"		—	1			X	X	
MW-18-2	MW-18-2		1208		HCl	3	X				
"	"		"		HNO3	1		X			
"	"		"		—	1			X	X	

Analysis Items

White – With report
Yellow – Lab copy
Pink – Originator

5853

QC Requirement: ☐Regular; ☐QA/QC Report; ☐WIP; ☐Raw Data; ☐Extended Raw Data ☐CLP; ☐ACE ☐AFCEE ☐NEESA ___ (E, C or D); ☐Other _____ (Please specify)

Sample Disposal: ☐Return ☐Disposal by APCL ☐Hold for _____ days after receiving date. **If not specified, samples will be discarded 45 days after samples are received.**

Sample Conditions: ☐Intact; ☐Broken. Cooler Seal: ☐Intact; ☐Broken; ☐None. Tag # _____ Temperature: ☐Room ☐Cold (_____ °C)

Relinquished by [signature] Date/Time 11-4-02/1615 Received by [signature] Date/Time 11-4-02/1615

Relinquished by [signature] Date/Time 11-4-02/1715 Received by [signature] Date/Time 11/4/02, 1715

APCL USE ONLY Service # Note:

Clients understand that all terms described in the proposals, quotations for this project, and/or the general terms provided in the current APCL price schedules will be followed. APCL reserves the right to terminate its service or withhold delivery of any reports, if in APCL's sole discretion the terms of the project have been broken.

Applied P & Ch Laboratory
13760 Magnolia Ave., Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Sample Receiving Checklist

APCL ServiceID: 5853 **Client Name/Project:** Zota Environmental

1. Sample Arrival

Date/Time Received 11/4/02 1715 Date/Time Opened 11/4/02 1715 By (name): Kenny Chan

Custody Transfer: ☐ Client ☐ Golden State ☐ UPS ☐ US Mail ☐ FedEx ☒ APCL Empl: Adam Wood

2. Chain-of-Custody (CoC)

☒ With Samples? ☐ Faxed? ☐ Client has Copy? ☐ Signed, dated? By: _____

☒ Project ID? ☒ Analyses Clear? ☐ Hold Samples? #on Hold _____ # Received 6

☒ CoC/Docs Zip-Locked under lid? ☐ Compos #: _ ☒ #Samples OK? _____

☐ Discrepancies? ☐ Client notified? ☐ Response (attach docs): _____

3. Shipping Container/Cooler

☒ Cooler Used? # of 1 Cooled by: ☒ Ice ☐ Blue Ice ☐ Dry Ice ☐ None

Temp °C 3.7 _____ _____ _____ _____ _____

(Cooler temperature measured from temp blank if present, otherwise measured from the cooler).

Cooler Custody Seal? ☐ Absent ☐ Intact ☐ Tampered?

4. Sample Preservation

☐ **pH <2** ☐ pH >12

If Not, pH = _____ Preserved by: ☐ Client ☐ APCL ☐ Third Party _____

5. Holding-time Requirements

☐ pH 24hr ☐ BACT 6/24hr ☒ Cr^{VI} 24hr ☐ NO_3^- 48hr ☐ BOD 48hr

☐ Cl_2 ASAP ☐ Turbidity 48hr ☐ DO ASAP ☐ Fe(II) ASAP

☐ HT Expired? ☐ Client notified?

6. Sample Container Condition

☒ Intact? ☐ Broken? ☐ Documented? Number: _____

Type: ☒ plastic ☒ glass ☐ Tube: brass/SS ☐ Tedlar Bag

☒ Quantity OK? ☐ Leaking? ☐ Anomaly?

☒ Caps tight? ☐ Air Bubbles? ☐ Anomaly?

Labels: ☒ Unique ID? ☒ Date/Time ☐ Preserved?

7. Turn Around Time

☐ RUSH TAT: _____ ☒ Std (7-10 days) ☐ Not Marked

8. Sample Matrix

☐ Drinking H_2O ☒ Other Liq ☐ Soil ☐ Wipe ☐ Polymer ☐ Air ☐ Other: _____

☐ Ground H_2O ☐ Sludge ☐ Filter ☐ Oil/Petro ☐ Paint ☐ W. Water ☐ Extract ☐ Unknown

9. Pre-Login Check List Completed & OK?

☒ ALL OK? (if not, attach docs) ☐ Client Contact? (Name:_____)Date/Time:_____

Received/Checked by: Kenny Chan Date: 4 Nov 2002 Time: 8:59 a.m.

*HT: Samples must be analyzed for results to reflect total concentrations. Results generated outside required of holding times are considered minimal values and may be used to define waste as hazardous but not as non-hazardous

DocumentFile: [neal.texfiles]smprcl.tex.

019002

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Sample Login: Check List

02-05853 (1288_ 415) (4858100_ 415)

11/05/02

Part 1: General Information

☐	Company Information	Name:	*SCTA Environmental*
		Address:	*16835 W. Bernardo Dr, Ste. 212 ,San Diego ,CA 92127*
☐	Project Information	Project Description:	*JPL*
		Project #:	*00HW019*
☐	Billing Information	P.O. #:	
		Bill Address:	*16835 W. Bernardo Dr, Ste. 212 ,San Diego ,CA 92127*
		Lab Project ID:	*2002_0002*
		Client Database #:	*01*
☐	Receiving Information	Who Received Sample?	*Kenny Chan*
		Receiving Date/Time:	*11/04/02 1715*
		COC No.	
☐	Shipping Information	Shipping Company	*APCL pick up*
		Packing Information:	*Cooler/Ice Chester*
		Cooler Temperature:	*3.7 °C*
☐	Container Information	Container Provider:	*Client*
☐	Sampling Information	Sampling Person:	
		Sampling Company:	*Client*
☐	Turn-Around-Time Option:		*Rush 5 working day(s)*
☐	QC Option:		*NEESA D*
☐	Disposal Option:		*Not specify*

019003

Part 2: Sample Information

Seq. #	Sample ID (on COC)	Sample Sub-ID	APCL Sample ID	Matrix	Cont-tainer	Preser-vative	Vol, ml Am g	# of Replica	Condition G, L, B	Collected mmddyy	Hold ?	Composite Group	TAT Days	
1	TB-18	VOC	02-05853-6	W	V	C	40	2	G	110402	N	0	6	☐
2	ER-18	VOC	02-05853-1-α	W	V	C	40	3	G	110402	N	0	6	☐
	ER-18	CRVI/Perch	02-05853-1-β	W	P		500	1	G	110402	N	0	6	☐
3	MW-18-5	VOC	02-05853-5-α	W	V	C	40	3	G	110402	N	0	6	☐
	MW-18-5	Perch	02-05853-5-β	W	P		500	1	G	110402	N	0	6	☐
4	MW-18-4	VOC	02-05853-4-α	W	V	C	40	3	G	110402	N	0	6	☐
	MW-18-4	CRVI/Perch	02-05853-4-β	W	P		500	1	G	110402	N	0	6	☐
5	MW-18-3	VOC	02-05853-3-α	W	V	C	40	3	G	110402	N	0	6	☐
	MW-18-3	CRVI/Perch	02-05853-3-β	W	P		500	1	G	110402	N	0	6	☐
6	MW-18-2	VOC	02-05853-2-α	W	V	C	40	3	G	110402	N	0	6	☐
	MW-18-2	CRVI/Perch	02-05853-2-β	W	P		500	1	G	110402	N	0	6	☐

Part 3: Analysis Information

Test Items:

☐	*524.2*	Volatile Organic Compounds
☐	*7196*	Chromium (VI)
☐	*300.0*	Perchlorate, low level
☐	*200.7/6010*	Sodium, Na, by ICP
☐	*200.7/6010*	Potassium, K, by ICP
☐	*200.7/6010*	Calcium, Ca, by ICP
☐	*200.7/6010*	Magnesium, Mg, by ICP
☐	*200.7/6010*	Iron, Fe, by ICP
☐	*300.0*	Sulfate (SO_4^{--}), by IC
☐	*300.0/SM4500NO3C*	Nitrate (NO_3^-) as N by IC
☐	*300.0*	Chloride Cl^- by IC
☐	*SM2320B*	Carbonate
☐	*SM2320B*	Bicarbonate
☐	*9040/150.1*	pH
☐	*160.1*	Solids, Total Dissolved (TDS)
☐	*206.2/7060*	Arsenic, As, by GFAA
☐	*310.1*	Alkalinity
☐	*PAH-SIM*	PAH (NOAA)

Seq. #	Client"s Sample ID (as given on COC)	Sample Sub-ID	APCL Sample ID	Matrix	524.2	CHROMIUM	PERCHL	NA	K	CA	MG	FE

019004