

**LABORATORY DATA CONSULTANTS, INC.**

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

SOTA Environmental  
16835 W. Bernardo, Drive, Suite 212  
San Diego, CA 92127-1813  
ATTN: Ms. Yu Zeng

December 3, 2002

SUBJECT: JPL, 00HW019, Data Validation

Dear Ms. Zeng,

Enclosed are the final validation reports for the fractions listed below. This SDG was received on November 20, 2002. Attachment 1 is a summary of the samples that were reviewed for each analysis.

**LDC Project # 9412:**

<b><u>SDG #</u></b>	<b><u>Fraction</u></b>
02-5527	Volatiles, Wet Chemistry

The data validation was performed under EPA Level IV guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, February 1994
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996

Please feel free to contact us if you have any questions.

Sincerely,

Stacey A. Mavrakos  
Operations Manager/Senior Chemist



**JPL, 00HW019**  
**Data Validation Reports**  
**LDC# 9412**

Volatiles

*LDC*

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** JPL, 00HW019  
**Collection Date:** October 16, 2002  
**LDC Report Date:** November 26, 2002  
**Matrix:** Water  
**Parameters:** Volatiles  
**Validation Level:** EPA Level IV  
**Laboratory:** Applied P & Ch Laboratory  
**Sample Delivery Group (SDG):** 02-5527

**Sample Identification**

ER-11  
MW-11-1  
MW-11-2  
MW-11-3  
MW-11-4  
MW-11-3D  
TB-11  
MW-11-4MS  
MW-11-4MSD

## Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
  - J Indicates an estimated value.
  - R Quality control indicates the data is not usable.
  - N Presumptive evidence of presence of the constituent.
  - UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
  - A Indicates the finding is based upon technical validation criteria.
  - P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

## III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination ( $r^2$ ) was greater than or equal to 0.990 .

## IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

## V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
02G4259MB01	10/17/02	Methylene chloride	0.4 ug/L	All samples in SDG 02-5527

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks.

Sample TB-11 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample ER-11 was identified as an equipment rinsate. No volatile contaminants were found in this blank with the following exceptions:

Equipment Rinsate ID	Sampling Date	Compound	Concentration	Associated Samples
ER-11	10/16/02	Toluene	0.4 ug/L	MW-11-1 MW-11-2 MW-11-3 MW-11-4 MW-11-3D

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks.

#### **VI. Surrogate Spikes**

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

#### **VII. Matrix Spike/Matrix Spike Duplicates**

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

#### **VIII. Laboratory Control Samples (LCS)**

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

#### **IX. Regional Quality Assurance and Quality Control**

Not applicable.

#### **X. Internal Standards**

All internal standard areas and retention times were within QC limits.

#### **XI. Target Compound Identifications**

All target compound identifications were within validation criteria.

#### **XII. Compound Quantitation and CRQLs**

All compound quantitation and CRQLs were within validation criteria.

### **XIII. Tentatively Identified Compounds (TICs)**

Tentatively identified compounds were not reported by the laboratory.

### **XIV. System Performance**

The system performance was acceptable.

### **XV. Overall Assessment of Data**

Data flags have been summarized at the end of the report.

### **XVI. Field Duplicates**

Samples MW-11-3 and MW-11-3D were identified as field duplicates. No volatiles were detected in any of the samples.

**JPL, 00HW019**  
**Volatiles - Data Qualification Summary - SDG 02-5527**

No Sample Data Qualified in this SDG

**JPL, 00HW019**  
**Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5527**

No Sample Data Qualified in this SDG

**JPL, 00HW019**  
**Volatiles - Field Blank Data Qualification Summary - SDG 02-5527**

No Sample Data Qualified in this SDG

**JPL, 00HW019**  
**Data Validation Reports**  
**LDC# 9412**

Wet Chemistry

*LDC*

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** JPL, 00HW019  
**Collection Date:** October 16, 2002  
**LDC Report Date:** November 26, 2002  
**Matrix:** Water  
**Parameters:** Wet Chemistry  
**Validation Level:** EPA Level IV  
**Laboratory:** Applied P & Ch Laboratory

**Sample Delivery Group (SDG):** 02-5527

**Sample Identification**

ER-11  
MW-11-1  
MW-11-2  
MW-11-3  
MW-11-4  
MW-11-3D  
ER-11MS  
ER-11MSD  
MW-11-4MS  
MW-11-4MSD

## Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 7196A for Hexavalent Chromium and EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## **I. Technical Holding Times**

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## **II. Calibration**

### **a. Initial Calibration**

All criteria for the initial calibration were met.

### **b. Calibration Verification**

Calibration verification frequency and analysis criteria were met for each method when applicable.

## **III. Blanks**

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

Sample ER-11 was identified as an equipment rinsate. No contaminant concentrations were found in this blank.

## **IV. Matrix Spike/Matrix Spike Duplicates**

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## **V. Duplicates**

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

## **VI. Laboratory Control Samples**

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## **VII. Sample Result Verification**

All sample result verifications were within validation criteria.

## **VIII. Overall Assessment of Data**

Data flags are summarized at the end of this report.

## **IX. Field Duplicates**

Samples MW-11-3 and MW-11-3D were identified as field duplicates. No contaminant concentrations were detected in any of the samples.

**JPL, 00HW019**  
**Wet Chemistry - Data Qualification Summary - SDG 02-5527**


No Sample Data Qualified in this SDG

**JPL, 00HW019**  
**Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 02-5527**

No Sample Data Qualified in this SDG

**JPL, 00HW019**  
**Wet Chemistry - Field Blank Data Qualification Summary - SDG 02-5527**

No Sample Data Qualified in this SDG



**LABORATORY DATA CONSULTANTS, INC.**

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

SOTA Environmental  
16835 W. Bernardo, Drive, Suite 212  
San Diego, CA 92127-1813  
ATTN: Ms. Yu Zeng

December 3, 2002

SUBJECT: JPL, 00HW019, Data Validation

Dear Ms. Zeng,

Enclosed are the final validation reports for the fractions listed below. This SDG was received on November 22, 2002. Attachment 1 is a summary of the samples that were reviewed for each analysis.

**LDC Project # 9417:**

<b><u>SDG #</u></b>	<b><u>Fraction</u></b>
02-5555	Volatiles, Wet Chemistry

The data validation was performed under EPA Level IV guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, February 1994
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996

Please feel free to contact us if you have any questions.

Sincerely,

Stacey A. Mavrakos  
Operations Manager/Senior Chemist



**JPL, 00HW019**  
**Data Validation Reports**  
**LDC# 9417**

Volatiles

*LDC*

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** JPL, 00HW019  
**Collection Date:** October 17, 2002  
**LDC Report Date:** November 27, 2002  
**Matrix:** Water  
**Parameters:** Volatiles  
**Validation Level:** EPA Level IV  
**Laboratory:** Applied P & Ch Laboratory  
**Sample Delivery Group (SDG):** 02-5555

**Sample Identification**

ER-20  
MW-20-1  
MW-20-2  
MW-20-3  
MW-20-4  
MW-20-5  
TB-20  
MW-20-1MS  
MW-20-1MSD

## Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

## III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination ( $r^2$ ) was greater than or equal to 0.990 .

## IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

## V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample TB-20 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Sampling Date	Compound	Concentration	Associated Samples
TB-20	10/17/02	Methylene chloride	0.7 ug/L	ER-20 MW-20-1 MW-20-2 MW-20-3 MW-20-4 MW-20-5

Sample ER-20 was identified as an equipment rinsate. No volatile contaminants were found in this blank.

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks.

#### **VI. Surrogate Spikes**

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

#### **VII. Matrix Spike/Matrix Spike Duplicates**

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

#### **VIII. Laboratory Control Samples (LCS)**

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

#### **IX. Regional Quality Assurance and Quality Control**

Not applicable.

#### **X. Internal Standards**

All internal standard areas and retention times were within QC limits.

#### **XI. Target Compound Identifications**

All target compound identifications were within validation criteria.

#### **XII. Compound Quantitation and CRQLs**

All compound quantitation and CRQLs were within validation criteria.

#### **XIII. Tentatively Identified Compounds (TICs)**

Tentatively identified compounds were not reported by the laboratory.

#### **XIV. System Performance**

The system performance was acceptable.

## **XV. Overall Assessment of Data**

Data flags have been summarized at the end of the report.

## **XVI. Field Duplicates**

No field duplicates were identified in this SDG.

**JPL, 00HW019**

**Volatiles - Data Qualification Summary - SDG 02-5555**

No Sample Data Qualified in this SDG

**JPL, 00HW019**

**Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5555**

No Sample Data Qualified in this SDG

**JPL, 00HW019**

**Volatiles - Field Blank Data Qualification Summary - SDG 02-5555**

No Sample Data Qualified in this SDG

**JPL, 00HW019**  
**Data Validation Reports**  
**LDC# 9417**

Wet Chemistry

*LDC*

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** JPL, 00HW019  
**Collection Date:** October 17, 2002  
**LDC Report Date:** November 26, 2002  
**Matrix:** Water  
**Parameters:** Wet Chemistry  
**Validation Level:** EPA Level IV  
**Laboratory:** Applied P & Ch Laboratory

**Sample Delivery Group (SDG):** 02-5555

**Sample Identification**

ER-20  
MW-20-1  
MW-20-2  
MW-20-3  
MW-20-4  
MW-20-5  
MW-20-5MS  
MW-20-5MSD

## Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA SW 846 Method 7196A for Hexavalent Chromium and EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## **I. Technical Holding Times**

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## **II. Calibration**

### **a. Initial Calibration**

All criteria for the initial calibration were met.

### **b. Calibration Verification**

Calibration verification frequency and analysis criteria were met for each method when applicable.

## **III. Blanks**

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

Sample ER-20 was identified as an equipment rinsate. No contaminant concentrations were found in this blank.

## **IV. Matrix Spike/Matrix Spike Duplicates**

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## **V. Duplicates**

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

## **VI. Laboratory Control Samples**

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## **VII. Sample Result Verification**

All sample result verifications were within validation criteria.

## **VIII. Overall Assessment of Data**

Data flags are summarized at the end of this report.

## **IX. Field Duplicates**

No field duplicates were identified in this SDG.

**JPL, 00HW019**

**Wet Chemistry - Data Qualification Summary - SDG 02-5555**

No Sample Data Qualified in this SDG

**JPL, 00HW019**


**Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 02-5555**

No Sample Data Qualified in this SDG

**JPL, 00HW019**

**Wet Chemistry - Field Blank Data Qualification Summary - SDG 02-5555**

No Sample Data Qualified in this SDG



**LABORATORY DATA CONSULTANTS, INC.**

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

SOTA Environmental  
16835 W. Bernardo, Drive, Suite 212  
San Diego, CA 92127-1813  
ATTN: Ms. Yu Zeng

December 23, 2002

SUBJECT: JPL, 00HW019, Data Validation

Dear Ms. Zeng,

Enclosed are the final validation reports for the fractions listed below. These SDGs were received on December 12, 2002. Attachment 1 is a summary of the samples that were reviewed for each analysis.

**LDC Project # 9513:**

<u>SDG #</u>	<u>Fraction</u>
02-5832, 02-5853, 02-5889, 02-5923	Volatiles, Wet Chemistry

The data validation was performed under EPA Level IV guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, February 1994
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996

Please feel free to contact us if you have any questions.

Sincerely,



Stacey A. Mavrakos  
Operations Manager/Senior Chemist



**JPL, 00HW019**  
**Data Validation Reports**  
**LDC# 9513**

**Volatiles**

*LDC*

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** JPL, 00HW019  
**Collection Date:** November 1, 2002  
**LDC Report Date:** December 17, 2002  
**Matrix:** Water  
**Parameters:** Volatiles  
**Validation Level:** EPA Level IV  
**Laboratory:** Applied P & Ch Laboratory

**Sample Delivery Group (SDG):** 02-5832

**Sample Identification**

ER-17  
MW-17-2  
MW-17-3  
MW-17-4  
MW-17-5  
TB-17  
MW-17-4MS  
MW-17-4MSD

## Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

## III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination ( $r^2$ ) was greater than or equal to 0.990 .

## IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

## V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample TB-17 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Sampling Date	Compound	Concentration	Associated Samples
TB-17	11/1/02	Methylene chloride	0.4 ug/L	ER-17 MW-17-2 MW-17-3 MW-17-4 MW-17-5

Sample ER-17 was identified as an equipment rinsate. No volatile contaminants were found in this blank with the following exceptions:

Equipment Rinsate ID	Sampling Date	Compound	Concentration	Associated Samples
ER-17	11/1/02	m,p-Xylenes	0.4 ug/L	MW-17-2 MW-17-3 MW-17-4 MW-17-5

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks.

#### **VI. Surrogate Spikes**

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

#### **VII. Matrix Spike/Matrix Spike Duplicates**

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

#### **VIII. Laboratory Control Samples (LCS)**

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

#### **IX. Regional Quality Assurance and Quality Control**

Not applicable.

#### **X. Internal Standards**

All internal standard areas and retention times were within QC limits.

#### **XI. Target Compound Identifications**

All target compound identifications were within validation criteria.

#### **XII. Compound Quantitation and CRQLs**

All compound quantitation and CRQLs were within validation criteria.

#### **XIII. Tentatively Identified Compounds (TICs)**

Tentatively identified compounds were not reported by the laboratory.

#### **XIV. System Performance**

The system performance was acceptable.

#### **XV. Overall Assessment of Data**

Data flags have been summarized at the end of the report.

#### **XVI. Field Duplicates**

No field duplicates were identified in this SDG.

**JPL, 00HW019**

**Volatiles - Data Qualification Summary - SDG 02-5832**

No Sample Data Qualified in this SDG

**JPL, 00HW019**

**Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5832**

No Sample Data Qualified in this SDG

**JPL, 00HW019**

**Volatiles - Field Blank Data Qualification Summary - SDG 02-5832**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** JPL, 00HW019  
**Collection Date:** November 4, 2002  
**LDC Report Date:** December 17, 2002  
**Matrix:** Water  
**Parameters:** Volatiles  
**Validation Level:** EPA Level IV  
**Laboratory:** Applied P & Ch Laboratory

**Sample Delivery Group (SDG):** 02-5853

**Sample Identification**

ER-18  
MW-18-2  
MW-18-3  
MW-18-4  
MW-18-5  
TB-18

## Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

## III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination ( $r^2$ ) was greater than or equal to 0.990 .

## IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

## V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample TB-18 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Sampling Date	Compound	Concentration	Associated Samples
TB-18	11/4/02	Methylene chloride	0.6 ug/L	ER-18 MW-18-2 MW-18-3 MW-18-4 MW-18-5

Sample ER-18 was identified as an equipment rinsate. No volatile contaminants were found in this blank with the following exceptions:

Equipment Rinsate ID	Sampling Date	Compound	Concentration	Associated Samples
ER-18	11/4/02	m,p-Xylenes	0.3 ug/L	MW-18-2 MW-18-3 MW-18-4 MW-18-5

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks.

#### **VI. Surrogate Spikes**

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

#### **VII. Matrix Spike/Matrix Spike Duplicates**

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

#### **VIII. Laboratory Control Samples (LCS)**

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

#### **IX. Regional Quality Assurance and Quality Control**

Not applicable.

#### **X. Internal Standards**

All internal standard areas and retention times were within QC limits.

#### **XI. Target Compound Identifications**

All target compound identifications were within validation criteria.

#### **XII. Compound Quantitation and CRQLs**

All compound quantitation and CRQLs were within validation criteria.

#### **XIII. Tentatively Identified Compounds (TICs)**

Tentatively identified compounds were not reported by the laboratory.

#### **XIV. System Performance**

The system performance was acceptable.

#### **XV. Overall Assessment of Data**

Data flags have been summarized at the end of the report.

#### **XVI. Field Duplicates**

No field duplicates were identified in this SDG.

**JPL, 00HW019**

**Volatiles - Data Qualification Summary - SDG 02-5853**

No Sample Data Qualified in this SDG

**JPL, 00HW019**

**Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5853**

No Sample Data Qualified in this SDG

**JPL, 00HW019**

**Volatiles - Field Blank Data Qualification Summary - SDG 02-5853**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** JPL, 00HW019  
**Collection Date:** November 5, 2002  
**LDC Report Date:** December 17, 2002  
**Matrix:** Water  
**Parameters:** Volatiles  
**Validation Level:** EPA Level IV  
**Laboratory:** Applied P & Ch Laboratory

**Sample Delivery Group (SDG):** 02-5889

**Sample Identification**

MW-10  
MW-16  
MW-16D  
TB-10  
MW-10MS  
MW-10MSD

## Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

## III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination ( $r^2$ ) was greater than or equal to 0.990 .

## IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

## V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
02G4629MB01	11/11/02	Methylene chloride	3.1 ug/L	All samples in SDG 02-5889

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
MW-10	Methylene chloride	1.2 ug/L	1.2U ug/L
MW-16	Methylene chloride	1.2 ug/L	1.2U ug/L
MW-16D	Methylene chloride	0.4 ug/L	0.4U ug/L
TB-10	Methylene chloride	1.7 ug/L	1.7U ug/L

Sample TB-10 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Sampling Date	Compound	Concentration	Associated Samples
TB-10	11/5/02	Methylene chloride	1.7 ug/L	MW-10 MW-16 MW-16D

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks with the following exceptions:

Sample	Compound	Reported Concentration	Modified Final Concentration
MW-10	Methylene chloride	1.2 ug/L	1.2U ug/L
MW-16	Methylene chloride	1.2 ug/L	1.2U ug/L
MW-16D	Methylene chloride	0.4 ug/L	1U ug/L

## VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

## VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

### VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

### IX. Regional Quality Assurance and Quality Control

Not applicable.

### X. Internal Standards

All internal standard areas and retention times were within QC limits.

### XI. Target Compound Identifications

All target compound identifications were within validation criteria.

### XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

### XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

### XIV. System Performance

The system performance was acceptable.

### XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

### XVI. Field Duplicates

Samples MW-16 and MW-16D were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/L)		RPD
	MW-16	MW-16D	
Carbon tetrachloride	2.0	2.1	5
Chloroform	4.3	4.1	5
Methylene chloride	1.2	0.4	100

Compound	Concentration (ug/L)		RPD
	MW-16	MW-16D	
Trichloroethene	0.6	0.8	29
Trichlorofluoromethane	1.5	1.5	0

**JPL, 00HW019**

**Volatiles - Data Qualification Summary - SDG 02-5889**

No Sample Data Qualified in this SDG

**JPL, 00HW019**

**Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5889**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P
02-5889	MW-10	Methylene chloride	1.2U ug/L	A
02-5889	MW-16	Methylene chloride	1.2U ug/L	A
02-5889	MW-16D	Methylene chloride	0.4U ug/L	A
02-5889	TB-10	Methylene chloride	1.7U ug/L	A

**JPL, 00HW019**

**Volatiles - Field Blank Data Qualification Summary - SDG 02-5889**

SDG	Sample	Compound	Modified Final Concentration	A or P
02-5889	MW-10	Methylene chloride	1.2U ug/L	A
02-5889	MW-16	Methylene chloride	1.2U ug/L	A
02-5889	MW-16D	Methylene chloride	1U ug/L	A

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** JPL, 00HW019  
**Collection Date:** November 6, 2002  
**LDC Report Date:** December 17, 2002  
**Matrix:** Water  
**Parameters:** Volatiles  
**Validation Level:** EPA Level IV  
**Laboratory:** Applied P & Ch Laboratory

**Sample Delivery Group (SDG):** 02-5923

**Sample Identification**

MW-7  
MW-8  
TB-7

## Introduction

This data review covers 3 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

## III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination ( $r^2$ ) was greater than or equal to 0.990 .

## IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

## V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
02G4629MB01	11/11/02	Methylene chloride	3.1 ug/L	All samples in SDG 02-5923

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
MW-7	Methylene chloride	1.0 ug/L	1.0U ug/L
TB-7	Methylene chloride	1.0 ug/L	1.0U ug/L

Sample TB-7 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Sampling Date	Compound	Concentration	Associated Samples
TB-7	11/6/02	Methylene chloride	1.0 ug/L	MW-7 MW 8

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks with the following exceptions:

Sample	Compound	Reported Concentration	Modified Final Concentration
MW-7	Methylene chloride	1.0 ug/L	1.0U ug/L

## VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

## VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

## IX. Regional Quality Assurance and Quality Control

Not applicable.

## **X. Internal Standards**

All internal standard areas and retention times were within QC limits.

## **XI. Target Compound Identifications**

All target compound identifications were within validation criteria.

## **XII. Compound Quantitation and CRQLs**

All compound quantitation and CRQLs were within validation criteria.

## **XIII. Tentatively Identified Compounds (TICs)**

Tentatively identified compounds were not reported by the laboratory.

## **XIV. System Performance**

The system performance was acceptable.

## **XV. Overall Assessment of Data**

Data flags have been summarized at the end of the report.

## **XVI. Field Duplicates**

No field duplicates were identified in this SDG.

**JPL, 00HW019**

**Volatiles - Data Qualification Summary - SDG 02-5923**

No Sample Data Qualified in this SDG

**JPL, 00HW019**

**Volatiles - Laboratory Blank Data Qualification Summary - SDG 02-5923**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P
02-5923	MW-7	Methylene chloride	1.0U ug/L	A
02-5923	TB-7	Methylene chloride	1.0U ug/L	A

**JPL, 00HW019**

**Volatiles - Field Blank Data Qualification Summary - SDG 02-5923**

SDG	Sample	Compound	Modified Final Concentration	A or P
02-5923	MW-7	Methylene chloride	1.0U ug/L	A