Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL

Collection Date: October 28, 2003

LDC Report Date: December 12, 2003

Matrix: Water

Parameters: Volatiles

Validation Level: EPA Level III & IV

Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-5852

Sample Identification

Dupe-1-4Q03**
EB-5-10-28-03
MW-24-1
MW-24-2
MW-24-3
MW-24-4
MW-24-5
TB-5-10-28-03

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

J Indicates an estimated value.

R Quality control indicates the data is not usable.

N Presumptive evidence of presence of the constituent.

UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
11/1/03	Chloromethane Bromomethane	34.05 59.05	All samples in SDG 03-5852	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
03G4668MB01	11/1/03	2-Butanone	0.61 ug/L	All samples in SDG 03-5852

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples Dupe-1-4Q03** and MW-24-4 were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-5-10-28-03 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-5-10-28-03 was identified as an equipment blank. No volatile contaminants were found in this blank.

NASA JPL
Volatiles - Data Qualification Summary - SDG 03-5852

SDG	Sample	Compound	Flag	A or P	Reason
03-5852	Dupe-1-4Q03** EB-5-10-28-03 MW-24-1 MW-24-2 MW-24-3 MW-24-4 MW-24-5 TB-5-10-28-03	Chloromethane Bromomethane	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-5852

No Sample Data Qualified in this SDG

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/28/2003
Project ID:	JPL	Service ID:	35852	Collected by:	Jr
		Lab Sample ID:	03-5852-1	Received Date:	10/28/2003
Sample ID:	**DUPE-1-4Q03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	11/01/03	Anal. Date:	11/01/03
Batch No:	03G4668	Prep. No:	–	Anal. Time:	07:35
Data File Name:	5852-01	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	UUJ
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	UUJ
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROETHENE (TOTAL)	540-59-0	μg/L	0.5	<0.5	U
32	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
33	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
34	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
35	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
36	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
37	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
38	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
39	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U
41	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
42	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
45	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
46	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
47	STYRENE	100-42-5	µg/L	0.5	<0.5	U
48	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
49	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
50	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
51	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
52	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
53	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
54	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
55	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
56	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
57	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
60	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
61	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
62	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
63	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
64	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	98
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	97
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	99
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	101
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	107
3	FLUOROBENZENE	462-06-6	50-200	108
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/28/2003
Project ID:	JPL	Service ID:	35852	Collected by:	Jr
		Lab Sample ID:	03-5852-2	Received Date:	10/28/2003
Sample ID:	**EB-5-10-28-03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	11/01/03	Anal. Date:	11/01/03
Batch No:	03G4668	Prep. No:	–	Anal. Time:	08:01
Data File Name:	5852-02	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U UJ
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U UJ
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROETHENE (TOTAL)	540-59-0	μg/L	0.5	<0.5	U
32	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
33	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
34	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
35	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
36	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
37	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
38	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
39	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U

11/28

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U
41	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
42	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
45	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
46	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
47	STYRENE	100-42-5	µg/L	0.5	<0.5	U
48	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
49	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
50	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
51	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
52	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
53	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
54	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
55	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
56	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
57	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
60	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
61	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
62	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
63	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
64	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	97
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	99
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	99
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	101
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	108
3	FLUOROBENZENE	462-06-6	50-200	108
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

M2482

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/28/2003
Project ID:	JPL	Service ID:	35852	Collected by:	Jr
		Lab Sample ID:	03-5852-3	Received Date:	10/28/2003
Sample ID:	**MW-24-1**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	11/01/03	Anal. Date:	11/01/03
Batch No:	03G4668	Prep. No:	–	Anal. Time:	08:27
Data File Name:	5852-03	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	UJ
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	19.1	
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	6.8	
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U UJ
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	0.7	
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROETHENE (TOTAL)	540-59-0	μg/L	0.5	<0.5	U
32	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
33	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
34	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
35	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
36	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
37	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
38	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
39	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U
41	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
42	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
45	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
46	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
47	STYRENE	100-42-5	µg/L	0.5	<0.5	U
48	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
49	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
50	TETRACHLOROETHENE	127-18-4	µg/L	0.5	1.6	
51	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
52	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
53	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
54	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
55	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
56	TRICHLOROETHENE	79-01-6	µg/L	0.5	3.7	
57	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
60	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
61	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
62	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
63	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
64	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	97
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	98
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	99
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	100
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	105
3	FLUOROBENZENE	462-06-6	50-200	107
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/28/2003
Project ID:	JPL	Service ID:	35852	Collected by:	Jr
		Lab Sample ID:	03-5852-4	Received Date:	10/28/2003
Sample ID:	**MW-24-2**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	11/01/03	Anal. Date:	11/01/03
Batch No:	03G4668	Prep. No:	–	Anal. Time:	08:54
Data File Name:	5852-04	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U UJ
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	3.4	
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	1.4	
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U UJ
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROETHENE (TOTAL)	540-59-0	μg/L	0.5	<0.5	U
32	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
33	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
34	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
35	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
36	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
37	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
38	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
39	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U
41	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
42	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
45	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
46	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
47	STYRENE	100-42-5	µg/L	0.5	<0.5	U
48	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
49	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
50	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
51	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
52	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
53	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
54	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
55	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
56	TRICHLOROETHENE	79-01-6	µg/L	0.5	0.6	
57	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
60	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
61	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
62	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
63	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
64	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	98
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	97
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	99
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	100
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	106
3	FLUOROBENZENE	462-06-6	50-200	108
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/28/2003
Project ID:	JPL	Service ID:	35852	Collected by:	Jr
		Lab Sample ID:	03-5852-5	Received Date:	10/28/2003
Sample ID:	**MW-24-3**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	11/01/03	Anal. Date:	11/01/03
Batch No:	03G4668	Prep. No:	–	Anal. Time:	09:20
Data File Name:	5852-05	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U UJ
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U UJ
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	98
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	97
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	99
4	TOLUENE-D8	2037-26-5	73-129	104
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	100
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	105
3	FLUOROBENZENE	462-06-6	50-200	107
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/28/2003
Project ID:	JPL	Service ID:	35852	Collected by:	Jr
		Lab Sample ID:	03-5852-6	Received Date:	10/28/2003
Sample ID:	**MW-24-4**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	11/01/03	Anal. Date:	11/01/03
Batch No:	03G4668	Prep. No:	–	Anal. Time:	09:46
Data File Name:	5852-06	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U UJ
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U UJ
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	μg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	98
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	99
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	100
4	TOLUENE-D8	2037-26-5	73-129	104
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	101
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	107
3	FLUOROBENZENE	462-06-6	50-200	107
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/28/2003
Project ID:	JPL	Service ID:	35852	Collected by:	Jr
		Lab Sample ID:	03-5852-7	Received Date:	10/28/2003
Sample ID:	**MW-24-5**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	11/01/03	Anal. Date:	11/01/03
Batch No:	03G4668	Prep. No:	–	Anal. Time:	10:12
Data File Name:	5852-07	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U UJ
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U UJ
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	97
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	98
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	98
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	100
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	107
3	FLUOROBENZENE	462-06-6	50-200	107
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/28/2003
Project ID:	JPL	Service ID:	35852	Collected by:	Jr
		Lab Sample ID:	03-5852-8	Received Date:	10/28/2003
Sample ID:	**TB-5-10-28-03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	11/01/03	Anal. Date:	11/01/03
Batch No:	03G4668	Prep. No:	–	Anal. Time:	04:59
Data File Name:	5852-08	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	UUJ
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	UUJ
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	μg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	96
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	99
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	98
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	104
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	111
3	FLUOROBENZENE	462-06-6	50-200	112
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

LDC #: 11272D1
SDG #: 03-5852
Laboratory: Applied P & Ch Laboratory

VALIDATION COMPLETENESS WORKSHEET
Level III/IV

Date: 12/12/03
Page: 1 of 1
Reviewer:
2nd Reviewer:

METHOD: GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 10/28/03
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	%RSD. r^2 no RRFs
IV.	Continuing calibration	SW	✓
V.	Blanks	SW	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	MW-21-4 (03-5892)
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	A	Not reviewed for Level III validation.
XII.	Compound quantitation/CRQLs	A	Not reviewed for Level III validation.
XIII.	Tentitatively identified compounds (TICs)	N	Not reviewed for Level III validation.
XIV.	System performance	A	Not reviewed for Level III validation.
XV.	Overall assessment of data	A	
XVI.	Field duplicates	ND	D=1+6
XVII.	Field blanks	ND	EB=2. TB=8

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet
ND = No compounds detected
R = Rinsate
FB = Field blank
D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: ** Indicates sample underwent Level IV validation.
MI 1-20s

1	Dupe-1-4Q03**	11	03G4668MB01	21		31	
2	EB-5-10-28-03	12		22		32	
3	MW-24-1	13		23		33	
4	MW-24-2	14		24		34	
5	MW-24-3	15		25		35	
6	MW-24-4	16		26		36	
7	MW-24-5	17		27		37	
8	TB-5-10-28-03	18		28		38	
9		19		29		39	
10		20		30		40	

LDC #: 112701
SDG #: 03-5852

VALIDATION FINDINGS CHECKLIST

Reviewer: 9
2nd Reviewer: 2

Method: Volatiles (EPA Method 524.2)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical holding times				
All technical holding times were met.	✓			
Cooler temperature criteria was met.	✓			
II. GC/MS Instrument performance check				
Were the BFB performance results reviewed and found to be within the specified criteria?	✓			
Were all samples analyzed within the 12 hour clock criteria?	✓			
III. Initial calibration				
Did the laboratory perform a 5 point calibration prior to sample analysis?	✓			
Were all percent relative standard deviations (%RSD) $\leq$ 20%?	✓			
IV. Continuing calibration				
Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?	✓			
Were all percent differences (%D) $\leq$ 30%?		✓		
V. Blanks				
Was a method blank associated with every sample in this SDG?	✓			
Was a method blank analyzed at least once every 12 hours for each matrix and concentration?	✓			
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.	✓			
VI. Surrogate spikes				
Were all surrogate %R within QC limits?	✓			
If the percent recovery (%R) for one or more surrogates was out of QC limits, was a reanalysis performed to confirm samples with %R outside of criteria?			✓	
VII. Matrix spike/Matrix spike duplicates				
Was a matrix spike (MS) and matrix spike duplicate (MSD) analyzed for this SDG?	✓			
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the QC limits?	✓			
VIII. Laboratory control samples				
Was an LCS analyzed for this SDG?	✓			
Was an LCS analyzed per analytical batch?	✓			
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the QC limits?	✓			

LDC #: 11272D1
SDG #: 03-5852

VALIDATION FINDINGS CHECKLIST

Reviewer: 2
2nd Reviewer:

Validation Area	Yes	No	NA	Findings/Comments
IX. Regional Quality Assurance and Quality Control				
Were performance evaluation (PE) samples performed?		✓		
Were the performance evaluation (PE) samples within the acceptance limits?			✓	
X. Internal standards				
Were internal standard area counts within +/-40% from the associated calibration standard?	✓			
Were retention times within - 30% of the last continuing calibration or +/- 50% of the initial calibration?	✓			
XI. Target compound identification				
Were relative retention times (RRT's) within ± 0.06 RRT units of the standard?	✓			
Did compound spectra meet specified EPA "Functional Guidelines" criteria?	✓			
Were chromatogram peaks verified and accounted for?	✓			
XII. Compound quantitation/CRQLs				
Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?	✓			
Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	✓			
XIII. Tentatively identified compounds (TICs)				
Were the major ions (> 25 percent relative intensity) in the reference spectrum evaluated in sample spectrum?			✓	
Were relative intensities of the major ions within ± 20% between the sample and the reference spectra?			✓	
Did the raw data indicate that the laboratory performed a library search for all required peaks in the chromatograms (samples and blanks)?		✓		
XIV. System performance				
System performance was found to be acceptable.	✓			
XV. Overall assessment of data				
Overall assessment of data was found to be acceptable.	✓			
XVI. Field duplicates				
Field duplicate pairs were identified in this SDG.	✓			
Target compounds were detected in the field duplicates.		✓		
XVII. Field blanks				
Field blanks were identified in this SDG.	✓			
Target compounds were detected in the field blanks.		✓		

TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method 524.2)

A. Chloromethane	Q. 1,2-Dichloropropane	GG. Xylenes, total	WW. Bromobenzene	MMM. Naphthalene
B. Bromomethane	R. cis-1,3-Dichloropropene	HH. Vinyl acetate	XX. 1,2,3-Trichloropropane	NNN. 1,2,3-Trichlorobenzene
C. Vinyl choride	S. Trichloroethene	II. 2-Chloroethylvinyl ether	YY. n-Propylbenzene	OOO. 1,3,5-Trichlorobenzene
D. Chloroethane	T. Dibromochloromethane	JJ. Dichlorodifluoromethane	ZZ. 2-Chlorotoluene	PPP. trans-1,2-Dichloroethene
E. Methylene chloride	U. 1,1,2-Trichloroethane	KK. Trichlorofluoromethane	AAA. 1,3,5-Trimethylbenzene	QQQ. cis-1,2-Dichloroethene
F. Acetone	V. Benzene	LL. Methyl-tert-butyl ether	BBB. 4-Chlorotoluene	RRR. m,p-Xylenes
G. Carbon disulfide	W. trans-1,3-Dichloropropene	MM. 1,2-Dibromo-3-chloropropane	CCC. tert-Butylbenzene	SSS. o-Xylene
H. 1,1-Dichloroethene	X. Bromoform	NN. Diethyl ether	DDD. 1,2,4-Trimethylbenzene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane
I. 1,1-Dichloroethane	Y. 4-Methyl-2-pentanone	OO. 2,2-Dichloropropane	EEE. sec-Butylbenzene	UUU. Benzyl chloride
J. 1,2-Dichloroethene, total	Z. 2-Hexanone	PP. Bromochloromethane	FFF. 1,3-Dichlorobenzene	VVV. 4-Ethyltoluene
K. Chloroform	AA. Tetrachloroethene	QQ. 1,1-Dichloropropene	GGG. p-Isopropyltoluene	WWW. Ethanol
L. 1,2-Dichloroethane	BB. 1,1,2,2-Tetrachloroethane	RR. Dibromomethane	HHH. 1,4-Dichlorobenzene	XXX. Ethyl ether
M. 2-Butanone	CC. Toluene	SS. 1,3-Dichloropropane	III. n-Butylbenzene	
N. 1,1,1-Trichloroethane	DD. Chlorobenzene	TT. 1,2-Dibromoethane	JJJ. 1,2-Dichlorobenzene	
O. Carbon tetrachloride	EE. Ethylbenzene	UU. 1,1,1,2-Tetrachloroethane	KKK. 1,2,4-Trichlorobenzene	
P. Bromodichloromethane	FF. Styrene	VV. Isopropylbenzene	LLL. Hexachlorobutadiene	

Notes:

LDC #: 11421

SDG #: 03-5852

VALIDATION FINDINGS WORKSHEET
Continuing Calibration

Page: 1 of 1
Reviewer: 9
2nd Reviewer: [illegible]

METHOD: GC/MS VOA (EPA Method 524.2)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y N N/A Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?

Y N N/A Were all percent differences (%D) ≤ 30% ?

#	Date	Standard ID	Compound	Finding %D (Limit: ≤30.0%)	Associated Samples	Qualifications
	11/1/03	G46680/	A	34.05	All + Blk	J/UJ/P
			B	59.05		↓

CONCAL1S5

LDC #: 12720D1
SDG #: 03-5852

VALIDATION FINDINGS WORKSHEET
Blanks

Page: 1 of 1
Reviewer:
2nd Reviewer:

METHOD: GC/MS VOA (EPA Method 524.2)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y N N/A Was a method blank associated with every sample in this SDG?
Y N N/A Was a method blank analyzed at least once every 12 hours for each matrix and concentration?
Y N N/A Was there contamination in the method blanks? If yes, please see the qualifications below.

Blank analysis date: 11/1/03
Conc. units: µg/l
Associated Samples: All

Compound	Blank ID	Sample Identification								
	03G1668MB01									
Methylene chloride										
Acetone										
M	0.61									
CRQL										

Blank analysis date:
Conc. units:
Associated Samples:

Compound	Blank ID	Sample Identification								
Methylene chloride										
Acetone										
CRQL										

CIRCLED RESULTS WERE NOT QUALIFIED. ALL RESULTS NOT CIRCLED WERE QUALIFIED BY THE FOLLOWING STATEMENT:

Note: Common contaminants such as Methylene chloride, Acetone, 2-Butanone, Carbon disulfide and TICs that were detected in samples within ten times the associated method blank concentration were qualified as not detected, "U". Other contaminants within five times the method blank concentration were also qualified as not detected, "U".

LDC #: 112701
SDG #: 03-5852

VALIDATION FINDINGS WORKSHEET
Initial Calibration Calculation Verification

Page: 1 of 1
Reviewer:
2nd Reviewer:

METHOD: GC/MS VOA (EPA Method 524.2)

The Relative Response Factor (RRF), average RRF, and percent relative standard deviation (%RSD) were recalculated for the compounds identified below using the following calculations:

RRF = $(A_x)(C_{is})/(A_{is})(C_x)$
average RRF = sum of the RRFs/number of standards
%RSD = 100 * (S/X)

A_x = Area of compound,
C_x = Concentration of compound,
S = Standard deviation of the RRFs
X = Mean of the RRFs

A_{is} = Area of associated internal standard
C_{is} = Concentration of internal standard

#	Standard ID	Calibration Date	Compound (Reference Internal Standard)	Reported RRF (10 std)	Recalculated RRF (10 std)	Reported Average RRF (initial)	Recalculated Average RRF (initial)	Reported %RSD	Recalculated %RSD
1	ICAL	10/21/03	~~Methylene chloride~~ S (1st Internal Standard)	0.278	0.278	0.271	0.271	4.23	4.18
			~~Trichlorethene~~ AA (2nd internal standard)	0.389	0.389	0.373	0.373	7.19	7.23
			~~Bromoform~~ EE (3rd internal standard)	2.614	2.614	2.542	2.542	4.04	4.04
2			Methylene chloride (1st Internal Standard)						
			Trichlorethene (2nd internal standard)						
			Bromoform (3rd internal standard)						
3			Methylene chloride (1st Internal Standard)						
			Trichlorethene (2nd internal standard)						
			Bromoform (3rd internal standard)						
4			Methylene chloride (1st Internal Standard)						
			Trichlorethene (2nd internal standard)						
			Bromoform (3rd internal standard)						

Comments: Refer to Initial Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 112201
SDG #: 03-5852

VALIDATION FINDINGS WORKSHEET
Continuing Calibration Results Verification

Page: 1 of 1
Reviewer: 9
2nd Reviewer: [illegible]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent difference (%D) of the initial calibration average Relative Response Factors (RRFs) and the continuing calibration RRFs were recalculated for the compounds identified below using the following calculation:

% Difference = 100 * (ave. RRF - RRF)/ave. RRF
RRF = $(A_x)(C_{is})/(A_{is})(C_x)$

Where: ave. RRF = initial calibration average RRF
RRF = continuing calibration RRF
A_x = Area of compound, A_{is} = Area of associated internal standard
C_x = Concentration of compound, C_{is} = Concentration of internal standard

#	Standard ID	Calibration Date	Compound (Reference internal Standard)	Average RRF (initial)	Reported RRF (CC)	Recalculated RRF (CC)	Reported %D	Recalculated %D
1	GH668801	11/1/03	S ~~Methylene chloride~~ (1st Internal Standard)	0.271	0.266	0.266	1.8	1.7
			AA ~~Trichlorethene~~ (2nd internal standard)	0.373	0.384	0.384	2.9	3.0
			EE ~~Bromoform~~ (3rd internal standard)	2.542	2.421	2.421	4.8	4.8
2			Methylene chloride (1st Internal Standard)					
			Trichlorethene (2nd internal standard)					
			Bromoform (3rd internal standard)					
3			Methylene chloride (1st Internal Standard)					
			Trichlorethene (2nd internal standard)					
			Bromoform (3rd internal standard)					
4			Methylene chloride (1st Internal Standard)					
			Trichlorethene (2nd internal standard)					
			Bromoform (3rd internal standard)					

Comments: Refer to Continuing Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12701
SDG #: 03-5852

VALIDATION FINDINGS WORKSHEET
Surrogate Results Verification

Page: 1 of 1
Reviewer: [illegible]
2nd reviewer: [illegible]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) of surrogates were recalculated for the compounds identified below using the following calculation:

% Recovery: SF/SS * 100

Where: SF = Surrogate Found
SS = Surrogate Spiked

Sample ID: 1

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8	20	20.97	105	105	0
Bromofluorobenzene	↓	19.49	98	98	↓
1,2-Dichlorobenzene-d4		19.38	97	97	

Sample ID:

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID:

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID:

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID:

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

LDC #: 112701
SDG #: 03-5852

VALIDATION FINDINGS WORKSHEET
Laboratory Control Sample Results Verification

Page: 1 of 1
Reviewer: [illegible]
2nd Reviewer: [illegible]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the laboratoy control sample and laboratory control sample duplicate (if applicable) were recalculated for the compounds identified below using the following calculation:

% Recovery = 100 * SSC/SA Where: SSC = Spiked sample concentration
SA = Spike added

RPD = I LCS - LCSD I * 2/(LCS + LCSD) LCS = Laboraotry control sample percent recovery LCSD = Laboratory control sample duplicate percent recovery

LCS ID: G4668L01

Compound	Spike Added (µg/L)		Spiked Sample Concentration (µg/L)		LCS Percent Recovery		LCSD Percent Recovery		LCS/LCSD RPD	
	LCS	LCSD	LCS	LCSD	Reported	Recalc.	Reported	Recalc.	Reported	Recalculated
1,1-Dichloroethene	20	NA	19.3	NA	97	97				
Trichloroethene	↓	↓	19.1	↓	96	96				
Benzene	↓	↓	17.9	↓	90	90				
Toluene	↓	↓	17.7	↓	89	89				
Chlorobenzene	↓	↓	18.4	↓	92	92				

Comments: Refer to Laboratory Control Sample findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 112T2D1
SDG #: 03-5852

VALIDATION FINDINGS WORKSHEET
Sample Calculation Verification

Page: 1 of 1
Reviewer: [illegible]
2nd reviewer: [illegible]

METHOD: GC/MS VOA (EPA Method 524.2)

Y N N/A Were all reported results recalculated and verified for all level IV samples?
Y N N/A Were all recalculated results for detected target compounds agree within 10.0% of the reported results?

$$\text{Concentration} = \frac{(A_x)(I_s)(DF)}{(A_{is})(RRF)(V_o)(\%S)}$$

A_x = Area of the characteristic ion (EICP) for the compound to be measured

A_{is} = Area of the characteristic ion (EICP) for the specific internal standard

I_s = Amount of internal standard added in nanograms (ng)

RRF = Relative response factor of the calibration standard.

V_o = Volume or weight of sample pruged in milliliters (ml) or grams (g).

Df = Dilution factor.

%S = Percent solids, applicable to soils and solid matrices only.

Example:

Sample I.D. 1, ND:

Conc. = ( ) ( ) ( )

( ) ( ) ( ) ( )

=

#	Sample ID	Compound	Reported Concentration ( )	Calculated Concentration ( )	Qualification

NASA JPL
Data Validation Reports
LDC# 11272

Chromium

LDC

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL

Collection Date: October 22 through October 29, 2003

LDC Report Date: December 18, 2003

Matrix: Water

Parameters: Chromium

Validation Level: EPA Level III & IV

Laboratory: Advanced Technology Laboratories & Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-5876/065632

Sample Identification

MW-20-5	MW-4-3	MW-19-3
MW-20-4	MW-4-2	MW-19-2
MW-20-3	MW-4-1	MW-19-1
MW-20-2	Dupe-3-4-Q03**	MW-4-3MS
MW-20-1	EB-4-10-27-03	MW-4-3MSD
EB-3-10-24-03	MW-24-5	MW-4-3DUP
Dupe-6-4-Q03**	MW-24-4	MW-11-3DUP
MW-11-5	Dupe-1-4-Q03**	MW-18-2MS
MW-11-4	MW-24-3	MW-18-2MSD
MW-11-3	MW-24-2	MW-18-2DUP
MW-11-2	MW-24-1	MW-19-1MS
MW-11-1	EB-5-10-28-03	MW-19-1MSD
EB-1-10-22-03	MW-18-5	MW-19-1DUP
MW-22-5	MW-18-4	MW-19-3MS
MW-22-4	MW-18-3	MW-19-3MSD
MW-22-3	MW-18-2	MW-19-3DUP
MW-22-2	MW-18-1	MW-24-2DUP
MW-22-1	EB-6-10-29-03	
MW-4-5	MW-19-5	
MW-4-4	MW-19-4	

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 57 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

J Indicates an estimated value.

R Quality control indicates the data is not usable.

N Presumptive evidence of presence of the constituent.

UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

CRDL standards for ICP and AA were analyzed and reported as required.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
ICB/CCB	Chromium	0.284 ug/L	MW-20-5 MW-20-4 MW-20-3 MW-20-2 MW-20-1 EB-3-10-24-03 Dupe-6-4-Q03** MW-11-5 MW-11-4 MW-11-3 MW-11-2 MW-11-1 EB-1-10-22-03 MW-22-5 MW-22-4 MW-22-3 MW-22-2 MW-22-1 MW-4-5 MW-19-3
PB (prep blank)	Chromium	0.461 ug/L	MW-4-4 MW-19-2 MW-19-1

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
ICB/CCB	Chromium	0.442 ug/L	MW-4-4 MW-4-2 MW-4-1 Dupe-3-4-Q03** EB-4-10-27-03 MW-24-5 MW-24-4 Dupe-1-4-Q03** MW-24-3 MW-24-2 MW-24-1 EB-5-10-28-03 MW-18-5 MW-18-4 MW-18-3 MW-18-2 MW-18-1 EB-6-10-29-03 MW-19-5 MW-19-4 MW-19-2 MW-19-1
PB (prep blank)	Chromium	0.1825 ug/L	MW-4-3
ICB/CCB	Chromium	0.400 ug/L	MW-4-3

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
MW-20-5	Chromium	1.3 ug/L	1.3U ug/L
MW-20-2	Chromium	1.3 ug/L	1.3U ug/L
Dupe-6-4-Q03**	Chromium	1.4 ug/L	1.4U ug/L
MW-11-4	Chromium	0.8 ug/L	0.8U ug/L
MW-11-2	Chromium	1 ug/L	1U ug/L
MW-22-5	Chromium	0.7 ug/L	0.7U ug/L
MW-22-2	Chromium	0.9 ug/L	0.9U ug/L

Sample	Analyte	Reported Concentration	Modified Final Concentration
MW-20-4	Chromium	1.6 ug/L	1.6U ug/L
MW-11-5	Chromium	1.5 ug/L	1.5U ug/L
MW-24-4	Chromium	1.2 ug/L	1.2U ug/L
Dupe-1-4-Q03**	Chromium	1.1 ug/L	1.1U ug/L
MW-24-3	Chromium	1.7 ug/L	1.7U ug/L
MW-18-5	Chromium	1.0 ug/L	1.0U ug/L
MW-18-2	Chromium	1.9 ug/L	1.9U ug/L
MW-18-1	Chromium	1.5 ug/L	1.5U ug/L
MW-19-5	Chromium	1.8 ug/L	1.8U ug/L
MW-19-1	Chromium	1.2 ug/L	1.2U ug/L
MW-20-4	Chromium	2.4 ug/L	2.4U ug/L
MW-4-1	Chromium	2.6 ug/L	2.6U ug/L
MW-24-2	Chromium	2.7 ug/L	2.7U ug/L
MW-18-4	Chromium	2.6 ug/L	2.6U ug/L
MW-19-4	Chromium	2.4 ug/L	2.4U ug/L
MW-4-3	Chromium	0.4 ug/L	0.4U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

Spike ID (Associated Samples)	Analyte	MS (%R) (Limits)	MSD (%R) (Limits)	RPD (Limits)	Flag	A or P
MW-19-3MS/MSD (MW-20-5 MW-20-4 MW-20-3 MW-20-2 MW-20-1 EB-3-10-24-03 Dupe-6-4-Q03** MW-11-5 MW-11-4 MW-11-3 MW-11-2 MW-11-1 EB-1-10-22-03 MW-22-5 MW-22-4 MW-22-3 MW-22-2 MW-22-1 MW-4-5 MW-19-3)	Chromium	68.4 (80-120)	72.8 (80-120)	-	J (all detects) UJ (all non-detects)	A

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

ICP serial dilution was not required by the method.

XI. Sample Result Verification

All sample result verifications met validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

Samples MW-20-2 and Dupe-6-4-Q03**, samples MW-4-5 and Dupe-3-4-Q03**, and samples MW-24-4 and Dupe-1-4-Q03** were identified as field duplicates. No chromium was detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	MW-20-2	Dupe-6-4-Q03**	
Chromium	1.3	1.4	7

Analyte	Concentration (ug/L)		RPD
	MW-4-5	Dupe-3-4-Q03**	
Chromium	3.5	5.6	46

Analyte	Concentration (ug/L)		RPD
	MW-24-4	Dupe-1-4-Q03**	
Chromium	1.2	1.1	9

XIV. Field Blanks

Samples EB-3-10-24-03, EB-1-10-22-03, EB-4-10-27-03, EB-5-10-28-03, and EB-6-10-29-03 were identified as equipment blanks. No chromium was found in these blanks.

NASA JPL
Chromium - Data Qualification Summary - SDG 03-5876/065632

SDG	Sample	Analyte	Flag	A or P	Reason
03-5876/ 065632	MW-20-5 MW-20-4 MW-20-3 MW-20-2 MW-20-1 EB-3-10-24-03 Dupe-6-4-Q03** MW-11-5 MW-11-4 MW-11-3 MW-11-2 MW-11-1 EB-1-10-22-03 MW-22-5 MW-22-4 MW-22-3 MW-22-2 MW-22-1 MW-4-5 MW-19-3	Chromium	J (all detects) UJ (all non-detects)	A	Matrix spike/Matrix spike duplicates (%R)

NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 03-5876/065632

SDG	Sample	Analyte	Modified Final Concentration	A or P
03-5876/ 065632	MW-20-5	Chromium	1.3U ug/L	A
03-5876/ 065632	MW-20-2	Chromium	1.3U ug/L	A
03-5876/ 065632	Dupe-6-4-Q03**	Chromium	1.4U ug/L	A
03-5876/ 065632	MW-11-4	Chromium	0.8U ug/L	A
03-5876/ 065632	MW-11-2	Chromium	1U ug/L	A
03-5876/ 065632	MW-22-5	Chromium	0.7U ug/L	A
03-5876/ 065632	MW-22-2	Chromium	0.9U ug/L	A

SDG	Sample	Analyte	Modified Final Concentration	A or P
03-5876/ 065632	MW-20-4	Chromium	1.6U ug/L	A
03-5876/ 065632	MW-11-5	Chromium	1.5U ug/L	A
03-5876/ 065632	MW-24-4	Chromium	1.2U ug/L	A
03-5876/ 065632	Dupe-1-4-Q03**	Chromium	1.1U ug/L	A
03-5876/ 065632	MW-24-3	Chromium	1.7U ug/L	A
03-5876/ 065632	MW-18-5	Chromium	1.0U ug/L	A
03-5876/ 065632	MW-18-2	Chromium	1.9U ug/L	A
03-5876/ 065632	MW-18-1	Chromium	1.5U ug/L	A
03-5876/ 065632	MW-19-5	Chromium	1.8U ug/L	A
03-5876/ 065632	MW-19-1	Chromium	1.2U ug/L	A
03-5876/ 065632	MW-20-4	Chromium	2.4U ug/L	A
03-5876/ 065632	MW-4-1	Chromium	2.6U ug/L	A
03-5876/ 065632	MW-24-2	Chromium	2.7U ug/L	A
03-5876/ 065632	MW-18-4	Chromium	2.6U ug/L	A
03-5876/ 065632	MW-19-4	Chromium	2.4U ug/L	A

SDG	Sample	Analyte	Modified Final Concentration	A or P
03-5876/ 065632	MW-4-3	Chromium	0.4U ug/L	A

NASA JPL
Chromium - Field Blank Data Qualification Summary - SDG 03-5876/065632

No Sample Data Qualified in this SDG

E

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-001A

Client Sample ID: MW-20-5
Collection Date: 10/24/2003 7:33:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	1.3	UJ	0.11	1.0	µg/L	1	11/4/2003 3:18:39 PM

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories
3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Client Sample ID: MW-20-4
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Collection Date: 10/24/2003 8:00:00 AM
Lab ID: 065632-002A
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				**Analyst: NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	1.6	J	0.11	1.0	µg/L	1	11/4/2003 3:21:08 PM

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories
3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-003A

Client Sample ID: MW-20-3
Collection Date: 10/24/2003 8:49:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	2.9	J	0.11	1.0	µg/L	1	11/4/2003 3:23:38 PM

11/24/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-20-2
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/24/2003 9:11:00 AM
Lab ID:	065632-004A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B	QC Bach: R32245			PrepDate:			
Chromium	1.3	UJ	0.11	1.0	µg/L	1	11/4/2003 3:26:08 PM

11/28

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-20-1
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/24/2003 9:49:00 AM
Lab ID:	065632-005A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	1.9	J	0.11	1.0	µg/L	1	11/4/2003 3:28:38 PM

Qualifiers:

- ND - Not Detected at the Reporting Limit
- J - Analyte detected below quantitation limits
- B - Analyte detected in the associated Method Blank
- * - Value exceeds Maximum Contaminant Level
- S - Spike Recovery outside accepted recovery limits
- R - RPD outside accepted recovery limits
- E - Value above quantitation range
- H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	EB-3-10-24-03
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/24/2003 8:41:00 AM
Lab ID:	065632-006A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	ND	UJ	0.11	1.0	µg/L	1	11/4/2003 3:31:09 PM

Qualifiers:

- ND - Not Detected at the Reporting Limit
- J - Analyte detected below quantitation limits
- B - Analyte detected in the associated Method Blank
- * - Value exceeds Maximum Contaminant Level
- S - Spike Recovery outside accepted recovery limits
- R - RPD outside accepted recovery limits
- E - Value above quantitation range
- H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	Dupe-6-4-Q03
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/24/2003 9:29:00 AM
Lab ID:	065632-007A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B	QC Bach: R32245		PrepDate:				
Chromium	1.4	UJ	0.11	1.0	µg/L	1	11/4/2003 3:33:41 PM

UPR

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-008A

Client Sample ID: MW-11-5
Collection Date: 10/22/2003 7:48:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	1.5	J	0.11	1.0	µg/L	1	11/4/2003 3:36:12 PM

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories
3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-11-4
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/22/2003 8:10:00 AM
Lab ID:	065632-009A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				**Analyst: NS**
RunID: ICP4_031104B QC Bach: R32245				PrepDate:			
Chromium	0.8	J UJ	0.11	1.0	µg/L	1	11/4/2003 3:43:54 PM

11/18/

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-010A

Client Sample ID: MW-11-3
Collection Date: 10/22/2003 8:47:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	3.4	J	0.11	1.0	µg/L	1	11/4/2003 3:46:27 PM

MDL

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories
3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-011A

Client Sample ID: MW-11-2
Collection Date: 10/22/2003 9:10:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B	QC Bach: R32245			PrepDate:			
Chromium	1	J UJ	0.22	2.0	µg/L	2	11/4/2003 4:38:59 PM

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories
3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-012A

Client Sample ID: MW-11-1
Collection Date: 10/22/2003 9:32:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				**Analyst: NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	2	J	0.22	2.0	µg/L	2	11/4/2003 4:41:32 PM

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-013A

Client Sample ID: EB-1-10-22-03
Collection Date: 10/22/2003 9:00:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B	QC Bach: R32245		PrepDate:				
Chromium	ND	UJ	0.11	1.0	µg/L	1	11/4/2003 3:56:39 PM

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories
3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Client Sample ID: MW-22-5
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Collection Date: 10/22/2003 10:21:00 AM
Lab ID: 065632-014A
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	0.7	J UJ	0.11	1.0	µg/L	1	11/4/2003 3:59:06 PM

MDR

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-22-4
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/22/2003 10:41:00 AM
Lab ID:	065632-015A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	3.1	J	0.22	2.0	µg/L	2	11/4/2003 4:44:05 PM

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-22-3
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/22/2003 11:00:00 AM
Lab ID:	065632-016A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				**Analyst: NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	3.2	J	0.22	2.0	µg/L	2	11/4/2003 4:46:39 PM

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-017A

Client Sample ID: MW-22-2
Collection Date: 10/22/2003 11:20:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	0.9	J NJ	0.56	5.0	µg/L	5	11/4/2003 5:09:44 PM

NJB

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-018A

Client Sample ID: MW-22-1
Collection Date: 10/22/2003 11:40:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				**Analyst: NS**
RunID: ICP4_031104B QC Bach: R32245			PrepDate:				
Chromium	3	J J	0.56	5.0	µg/L	5	11/4/2003 5:12:14 PM

UDK

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *17-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-019A

Client Sample ID: MW-4-5
Collection Date: 10/27/2003 7:44:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031104B	QC Bach: R32245		PrepDate:				
Chromium	3.5	J	0.11	1.0	µg/L	1	11/4/2003 4:16:39 PM

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-4-4
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/27/2003 8:39:00 AM
Lab ID:	065632-020A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105C	QC Bach: R32300		PrepDate:				
Chromium	2.4		0.11	1.0	µg/L	1	11/5/2003 11:58:43 A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-4-3
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/27/2003 9:01:00 AM
Lab ID:	065632-021A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**	**(EPA 200.8)**			Analyst: **NS**
RunID: ICP4_031105D	QC Bach: 15669			PrepDate: 11/5/2003			
Chromium	0.4	J U	0.11	1.0	µg/L	1	11/5/2003 2:28:26 PM

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-4-2
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/27/2003 9:20:00 AM
Lab ID:	065632-022A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B	QC Bach: R32298			PrepDate:			
Chromium	3.7		0.11	1.0	µg/L	1	11/5/2003 10:38:18 A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT: Applied P & Ch Laboratories
Client Sample ID: MW-4-1
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Collection Date: 10/27/2003 9:46:00 AM
Lab ID: 065632-023A
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	2.6		0.11	1.0	µg/L	1	11/5/2003 10:40:47 A

NPB

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories
3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	Dupe-3-4-Q03
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/27/2003 8:12:00 AM
Lab ID:	065632-024A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B	QC Bach: R32298			PrepDate:			
Chromium	5.6		0.11	1.0	µg/L	1	11/5/2003 10:43:17 A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	EB-4-10-27-03
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/27/2003 9:59:00 AM
Lab ID:	065632-025A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	ND		0.11	1.0	µg/L	1	11/5/2003 10:45:47 A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-24-5
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/28/2003 7:58:00 AM
Lab ID:	065632-026A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	3.7		0.11	1.0	µg/L	1	11/5/2003 10:48:17 A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-24-4
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/28/2003 8:29:00 AM
Lab ID:	065632-027A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	1.2	U	0.11	1.0	µg/L	1	11/5/2003 10:50:48 A

N/A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-028A

Client Sample ID: Dupe-1-4Q03
Collection Date: 10/28/2003 9:21:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				**Analyst: NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	1.1	U	0.11	1.0	µg/L	1	11/5/2003 10:53:19 A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories
3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-24-3
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/28/2003 9:46:00 AM
Lab ID:	065632-029A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	1.7	U	0.11	1.0	µg/L	1	11/5/2003 10:55:51 A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-030A

Client Sample ID: MW-24-2
Collection Date: 10/28/2003 10:11:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	2.7	U	0.11	1.0	µg/L	1	11/5/2003 11:03:33 A

a/12/8

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-24-1
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/28/2003 10:41:00 AM
Lab ID:	065632-031A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	4.0		0.11	1.0	µg/L	1	11/5/2003 11:08:39 A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-032A

Client Sample ID: EB-5-10-28-03
Collection Date: 10/28/2003 10:30:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	ND		0.11	1.0	µg/L	1	11/5/2003 11:11:13 A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories
3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-18-5
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/29/2003 7:46:00 AM
Lab ID:	065632-033A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B	QC Bach: R32298		PrepDate:				
Chromium	1.0	U	0.11	1.0	µg/L	1	11/5/2003 11:13:47 A

ND

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-18-4
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/29/2003 8:14:00 AM
Lab ID:	065632-034A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	2.6	U	0.11	1.0	µg/L	1	11/5/2003 11:16:18 A

N 12/8

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-18-3
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/29/2003 8:41:00 AM
Lab ID:	065632-035A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				**Analyst: NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	5.9		0.11	1.0	µg/L	1	11/5/2003 11:18:45 A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT: Applied P & Ch Laboratories
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Lab ID: 065632-036A

Client Sample ID: MW-18-2
Collection Date: 10/29/2003 9:06:00 AM
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	1.9	U	0.11	1.0	µg/L	1	11/5/2003 11:21:13 A

11/12/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories *3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040*

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT: Applied P & Ch Laboratories
Client Sample ID: MW-18-1
Lab Order: 065632
Project: JPL GW Mon, 04-4428.10
Collection Date: 10/29/2003 9:55:00 AM
Lab ID: 065632-037A
Matrix: WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	1.5	U	0.11	1.0	µg/L	1	11/5/2003 11:26:10 A

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	EB-6-10-29-03
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/29/2003 9:45:00 AM
Lab ID:	065632-038A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	ND		0.11	1.0	µg/L	1	11/5/2003 11:33:48 A

n/2K

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-19-5
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/29/2003 11:17:00 AM
Lab ID:	065632-039A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B	QC Bach: R32298		PrepDate:				
Chromium	1.8	U	0.11	1.0	µg/L	1	11/5/2003 11:36:18 A

Qualifiers:

- ND - Not Detected at the Reporting Limit
- J - Analyte detected below quantitation limits
- B - Analyte detected in the associated Method Blank
- * - Value exceeds Maximum Contaminant Level
- S - Spike Recovery outside accepted recovery limits
- R - RPD outside accepted recovery limits
- E - Value above quantitation range
- H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-19-4
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/29/2003 11:40:00 AM
Lab ID:	065632-040A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105B QC Bach: R32298			PrepDate:				
Chromium	2.4	U	0.11	1.0	µg/L	1	11/5/2003 11:38:48 A

11/12/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-19-3
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/29/2003 12:10:00 PM
Lab ID:	065632-041A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031103A QC Bach: R32245			PrepDate:				
Chromium	4.3	J	0.22	2.0	µg/L	2	11/3/2003 4:54:04 PM

11/21/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-19-2
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/29/2003 12:50:00 PM
Lab ID:	065632-042A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105C QC Bach: R32300			PrepDate:				
Chromium	4.0		0.11	1.0	µg/L	1	11/5/2003 12:01:16 P

11/28

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *10-Nov-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-19-1
Lab Order:	065632		
Project:	JPL GW Mon, 04-4428.10	**Collection Date:**	10/29/2003 1:13:00 PM
Lab ID:	065632-043A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_031105C	QC Bach: R32300			PrepDate:			
Chromium	1.2	U	0.11	1.0	µg/L	1	11/5/2003 12:03:50 P

n/2 KB

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

LDC #: 11272E4

VALIDATION COMPLETENESS WORKSHEET

Level III/IV

SDG #: 03-5876/065632

Laboratory: Advanced Technology Laboratories/Applied P & Ch Laboratory

Date: 12/15/03

Page: 1 of 2

Reviewer: MR

2nd Reviewer: [illegible]

METHOD: Chromium (EPA Method 200.8)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 10/22 - 29/03
II.	Calibration	A	
III.	Blanks	SW	
IV.	ICP Interference Check Sample (ICS) Analysis	N	Not required
V.	Matrix Spike Analysis	SW	3 MS/MSD/pair
VI.	Duplicate Sample Analysis	A	
VII.	Laboratory Control Samples (LCS)	A	LCS
VIII.	Internal Standard (ICP-MS)	A	
IX.	Furnace Atomic Absorption QC	N	Not utilized
X.	ICP Serial Dilution	N	Not required
XI.	Sample Result Verification	A	Not reviewed for Level III validation.
XII.	Overall Assessment of Data	A	
XIII.	Field Duplicates	SW	(4,7), (19,24), (27,28)
XIV.	Field Blanks	ND	EB=6, 13, 25, 32, 38.

Note: A = Acceptable; N = Not provided/applicable; SW = See worksheet; ND = No compounds detected; R = Rinsate; FB = Field blank; D = Duplicate; TB = Trip blank; EB = Equipment blank

Validated Samples: ** Indicates sample underwent Level IV validation

1	MW-20-5	11	MW-11-2	21	MW-4-3	31	MW-24-1
2	MW-20-4	12	MW-11-1	22	MW-4-2	32	EB-5-10-28-03
3	MW-20-3	13	EB-1-10-22-03	23	MW-4-1	33	MW-18-5
4	MW-20-2	14	MW-22-5	24	Dupe-3-4-Q03**	34	MW-18-4
5	MW-20-1	15	MW-22-4	25	EB-4-10-27-03	35	MW-18-3
6	EB-3-10-24-03	16	MW-22-3	26	MW-24-5	36	MW-18-2
7	Dupe-6-4-Q03**	17	MW-22-2	27	MW-24-4	37	MW-18-1
8	MW-11-5	18	MW-22-1	28	Dupe-1-4-Q03**	38	EB-6-10-29-03
9	MW-11-4	19	MW-4-5	29	MW-24-3	39	MW-19-5
10	MW-11-3	20	MW-4-4	30	MW-24-2	40	MW-19-4

Notes:

LDC #: 11272E4

VALIDATION COMPLETENESS WORKSHEET

Level III/IV

SDG #: 03-5876/065632

Laboratory: Advanced Technology Laboratories/Applied P & Ch Laboratory

Date: 12/15/03

Page: 2 of 2

Reviewer: MW

2nd Reviewer: ✓

METHOD: Chromium (EPA Method 200.8)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times		Sampling dates:
II.	Calibration		
III.	Blanks		
IV.	ICP Interference Check Sample (ICS) Analysis		
V.	Matrix Spike Analysis	See	page 1
VI.	Duplicate Sample Analysis		
VII.	Laboratory Control Samples (LCS)		
VIII.	Internal Standard (ICP-MS)		
IX.	Furnace Atomic Absorption QC		
X.	ICP Serial Dilution		
XI.	Sample Result Verification		Not reviewed for Level III validation.
XII.	Overall Assessment of Data		
XIII.	Field Duplicates		
XIV.	Field Blanks		

Note:
A = Acceptable
N = Not provided/applicable
SW = See worksheet
ND = No compounds detected
R = Rinsate
FB = Field blank
D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: ** Indicates sample underwent Level IV validation

41	MW-19-3	51	MW-19-1MS				
42	MW-19-2	52	MW-19-1MSD				
43	MW-19-1	53	MW-19-1DUP				
44	MW-4-3MS	54	MW-19-3MS				
45	MW-4-3MSD	55	MW-19-3MSD				
46	MW-4-3DUP	56	MW-19-3DUP				
47	MW-11-3DUP	57	MW-24-2DUP				
48	MW-18-2MS	58	PB				
49	MW-18-2MSD	59					
50	MW-18-2DUP	60					

Notes:

LDC #: 11272E4
SDG #: See cover

VALIDATION FINDINGS CHECKLIST

Page: 1 of 1
Reviewer: MB
2nd Reviewer: [illegible]

Method: Metals (EPA SW 846 Method 6010/7000/6020)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical holding times				
All technical holding times were met.	✓			
Cooler temperature criteria was met.	✓			
II. Calibration				
Were all instruments calibrated daily, each set-up time?	✓			
Were the proper number of standards used?	✓			
Were all initial and continuing calibration verification %Rs within the 90-110% (80-120% for mercury and 85-115% for cyanide) QC limits?	✓			
Were all initial calibration correlation coefficients ≥ 0.995?	✓			
Was a midrange cyanide standard distilled?			✓	
III. Blanks				
Was a method blank associated with every sample in this SDG?	✓			
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.	✓			
IV. ICP Interference Check Sample				
Were ICP interference check samples performed daily?			✓	
Were the AB solution percent recoveries (%R) with the 80-120% QC limits?			✓	
IV. Matrix spike/Matrix spike duplicates				
Were a matrix spike (MS) and duplicate (DUP) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD or MS/DUP. Soil / Water.	✓			
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the 75-125 QC limits? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken.		✓		
Were the MS/MSD or duplicate relative percent differences (RPD) ≤ 20% for waters and ≤ 35% for soil samples? A control limit of +/- RL(+/-2X RL for soil) was used for samples that were ≤ 5X the RL, including when only one of the duplicate sample values were ≤ 5X the RL.	✓			
V. Laboratory control samples				
Was an LCS anaylzed for this SDG?	✓			
Was an LCS analyzed per extraction batch?	✓			
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the 80-120% QC limits for water samples and laboratory established QC limits for soils?	✓			
VI. Furnace Atomic Absorption QC				
If MSA was performed, was the correlation coefficients ≥ 0.995?			✓	
Do all applicable analysies have duplicate injections?			✓	

LDC #: 11272E4
SDG #: See cover

VALIDATION FINDINGS CHECKLIST

Page: 1 of 1
Reviewer: MB
2nd Reviewer: ✓

Validation Area	Yes	No	NA	Findings/Comments
For sample concentrations > RL, are applicable duplicate injection RSD values < 20%?			✓	
Were analytical spike recoveries within the 85-115% QC limits?			✓	
VII. ICP Serial Dilution				
Was an ICP serial dilution analyzed if analyte concentrations were > 50X the IDL?			✓	
Were all percent differences (%Ds) ≤ 10%?			✓	
Was there evidence of negative interference? If yes, professional judgement will be used to qualify the data.			✓	
VIII. Internal Standards (EPA SW 846 Method ~~6020~~) 200.8 6020				
Were all the percent recoveries (%R) within the ~~30-120%~~ of the intensity of the internal standard in the associated initial calibration?	✓			
If the %Rs were outside the criteria, was a reanalysis performed?			✓	
IX. Regional Quality Assurance and Quality Control				
Were performance evaluation (PE) samples performed?			✓	
Were the performance evaluation (PE) samples within the acceptance limits?			✓	
X. Sample Result Verification				
Were RLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	✓			
XI. Overall assessment of data				
Overall assessment of data was found to be acceptable.	✓			
XII. Field duplicates				
Field duplicate pairs were identified in this SDG.	✓			
Target analytes were detected in the field duplicates.	✓			
XIII. Field blanks				
Field blanks were identified in this SDG.	✓			
Target analytes were detected in the field blanks.		✓		

LDC #: 11272E4
SDG #: See cover

VALIDATION FINDINGS WORKSHEET
PB/ICB/CCB QUALIFIED SAMPLES

Page: 1 of 84
Reviewer: MA
2nd Reviewer: D

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: 1.75X, 4ome → 50ml
Sample Concentration units, unless otherwise noted: ug/L Associated Samples: 1-19, 41

					Sample Identification										
Analyte	Maximum PB[a] (mg/Kg)	Maximum PB[a] (ug/L)	Maximum ICB/CCB[a] (ug/L)	Blank Action Limit	1	4	7	9	11	14	17	2	8		
Al															Al
Sb															Sb
As															As
Ba															Ba
Be															Be
Cd															Cd
Ca															Ca
Cr			0.284	1.775 ~~1.42~~	1.3	1.3	1.4	0.8	1	0.7	0.9	1.6	1.5		Cr
Co				HW											Co
Cu															Cu
Fe															Fe
Pb															Pb
Mg															Mg
Mn															Mn
Hg															Hg
Ni															Ni
K															K
Se															Se
Ag															Ag
Na															Na
Tl															Tl
V															V
Zn															Zn
B															B
Mo															Mo
Sr															Sr

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".
Note : a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

BLNKSMP.4S2

LDC #: 11272E4
SDG #: See cover

VALIDATION FINDINGS WORKSHEET
PB/ICB/CCB QUALIFIED SAMPLES

Page: 2 of 4
Reviewer: MA
2nd Reviewer: [illegible]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: 1.25, 40ml → 50ml
Sample Concentration units, unless otherwise noted: ug/L Associated Samples: PB: 20, 42, 43, ICB/CCB: 20, 22-40, 42, 43

					Sample Identification										
Analyte	Maximum PB[a] (mg/Kg)	Maximum PB[a] (ug/L)	Maximum ICB/CCB[a] (ug/L)	Blank Action Limit	27	28	29	33	36	37	39	43	20	23	
Al															Al
Sb															Sb
As															As
Ba															Ba
Be															Be
Cd															Cd
Ca															Ca
Cr		0.461	0.442	2.76 ~~2.305~~ 2.21	1.2	1.1	1.7	1.0	1.9	1.5	1.8	1.2	2.4	2.6	Cr
Co															Co
Cu															Cu
Fe															Fe
Pb															Pb
Mg															Mg
Mn															Mn
Hg															Hg
Ni															Ni
K															K
Se															Se
Ag															Ag
Na															Na
Tl															Tl
V															V
Zn															Zn
B															B
Mo															Mo
Sr															Sr

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".
Note : a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

BLNKSMP.4S2

LDC #: 11272E4
SDG #: See cover

VALIDATION FINDINGS WORKSHEET
PB/ICB/CCB QUALIFIED SAMPLES

Page: 3 of 4
Reviewer: MH
2nd Reviewer:

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied:
Sample Concentration units, unless otherwise noted: ug/L Associated Samples: PB: 20, 42, 43, ICB/CCB: 20, 22-40, 42, 43

					Sample Identification										
Analyte	Maximum PB[a] (mg/Kg)	Maximum PB[a] (ug/L)	Maximum ICB/CCB[a] (ug/L)	Blank Action Limit	30	34	40								
Al															Al
Sb															Sb
As															As
Ba															Ba
Be															Be
Cd															Cd
Ca															Ca
Cr		0.461	0.442	2.7	2.7	2.6	2.4								Cr
Co															Co
Cu															Cu
Fe															Fe
Pb															Pb
Mg															Mg
Mn															Mn
Hg															Hg
Ni															Ni
K															K
Se															Se
Ag															Ag
Na															Na
Tl															Tl
V															V
Zn															Zn
B															B
Mo															Mo
Sr															Sr

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".

Note : a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

BLNKSMP.4S2

LDC #: 11272E4
SDG #: see cover

VALIDATION FINDINGS WORKSHEET
PB/ICB/CCB QUALIFIED SAMPLES

Page: 4 of 4
Reviewer: MH
2nd Reviewer:

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: 1.25X
Sample Concentration units, unless otherwise noted: ug/L Associated Samples: 2

					Sample Identification										
Analyte	Maximum PB[a] (mg/Kg)	Maximum PB[a] (ug/L)	Maximum ICB/CCB[a] (ug/L)	Blank Action Limit	2										
Al															Al
Sb															Sb
As															As
Ba															Ba
Be															Be
Cd															Cd
Ca															Ca
Cr		0.125 ~~0.146~~	0.400	2.0	0.4										Cr
Co															Co
Cu															Cu
Fe															Fe
Pb															Pb
Mg															Mg
Mn															Mn
Hg															Hg
Ni															Ni
K															K
Se															Se
Ag															Ag
Na															Na
Tl															Tl
V															V
Zn															Zn
B															B
Mo															Mo
Sr															Sr

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".
Note : a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

BLNKSMP.4S2

LDC #: 11272E4
SDG #: See cover

VALIDATION FINDINGS WORKSHEET
Matrix Spike/Matrix Spike Duplicates

Page: 1 of 1
Reviewer: MD
2nd Reviewer: A

METHOD: Trace metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y N N/A Was a matrix spike analyzed for each matrix in this SDG?

Y N N/A Were matrix spike percent recoveries (%R) within the control limits of 75-125? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken. Lab limit

Y N N/A Were all duplicate sample relative percent differences (RPD) ≤ 20% for water samples and ≤35% for soil samples?

LEVEL IV ONLY:

Y N N/A Were recalculated results acceptable? See Level IV Recalculation Worksheet for recalculations.

#	MS/MSD ID	Matrix	Analyte	MS %Recovery	MSD %Recovery	RPD (Limits)	Associated Samples	Qualifications
1	E4155	AQ	Cr	68.4 (80-120)	72.8 (80-120)		1-19, 41	J/UJ/A

Comments:

MSD.4S2

LDC #: 11272E4
SDG #: See cover

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
Reviewer: MH
2nd reviewer: A

METHOD: ~~Inorganics, Method~~ Cr 200.8

(Y) N N/A Were field duplicate pairs identified in this SDG?
(Y) N N/A Were target analytes detected in the field duplicate pairs?

Analyte	Concentration (ug/l)		RPD (Limits)	Qualifier
	4	7		
Cr	1.3	1.4	7	

Analyte	Concentration (ug/l)		RPD (Limits)	Qualifier
	19	24		
Cr	3.5	5.6	46	

Analyte	Concentration (ug/l)		RPD (Limits)	Qualifier
	27	28		
Cr	1.2	1.1	9	

Analyte	Concentration ( )		RPD (Limits)	Qualifier

LDC #: 11272E4
SDG #: See cover

VALIDATION FINDINGS WORKSHEET
Initial and Continuing Calibration Calculation Verification

Page: 1 of 1
Reviewer: MB
2nd Reviewer: A

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

An initial and continuing calibration verification percent recovery (%R) was recalculated for each type of analysis using the following formula:

$$\%R = \frac{\text{Found}}{\text{True}} \times 100$$

Where, Found = concentration (in ug/L) of each analyte measured in the analysis of the ICV or CCV solution
True = concentration (in ug/L) of each analyte in the ICV or CCV source

Standard ID	Type of Analysis	Element	Found (ug/L)	True (ug/L)	Recalculated %R	Reported %R	Acceptable (Y/N)
ICV	ICP/MS (Initial calibration)	Cr	10.19	10	102	102	Y
	GFAA (Initial calibration)						
	CVAA (Initial calibration)						
CCV	ICP/MS (Continuing calibration)	Cr	9.958	10	100	100	Y
	GFAA (Continuing calibration)						
	CVAA (Continuing calibration)						
	Cyanide (Initial calibration)						
	Cyanide (Continuing calibation)						

Comments: Refer to Calibration Verification findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

CALCLC.4SW

LDC #: 11272 G4
SDG #: See cover

VALIDATION FINDINGS WORKSHEET
Level IV Recalculation Worksheet

Page: 1 of 1
Reviewer: MM
2nd Reviewer: [illegible]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Percent recoveries (%R) for an ICP interference check sample, a laboratory control sample and a matrix spike sample were recaluculated using the following formula:

$$\%R = \frac{Found}{True} \times 100$$

Where, Found = Concentration of each analyte measured in the analysis of the sample. For the matrix spike calculation, Found = SSR (spiked sample result) - SR (sample result).
True = Concentration of each analyte in the source.

A sample and duplicate relative percent difference (RPD) was recalculated using the following formula:

$$RPD = \frac{|S-D|}{(S+D)/2} \times 100$$

Where, S = Original sample concentration
D = Duplicate sample concentration

An ICP serial dilution percent difference (%D) was recalculated using the following formula:

$$\%D = \frac{|I-SDR|}{I} \times 100$$

Where, I = Initial Sample Result (mg/L)
SDR = Serial Dilution Result (mg/L) (Instrument Reading x 5)

Sample ID	Type of Analysis	Element	Found / S / I (units)	True / D / SDR (units)	Recalculated %R / RPD / %D	Reported %R / RPD / %D	Acceptable (Y/N)
MB	ICP interference check	1					
LCS	Laboratory control sample	Cr	10.153	10	102	102	Y
S4	Matrix spike	Cr	(SSR-SR) 6.838	10	68.4	68.4	↓
S6	Duplicate	Cr	4.362	4.286	1.76	1.76	↓
MD	ICP serial dilution						

Comments: Refer to appropriate worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

TOTCLC.4SW

LDC #: 11272E4
SDG #: See cover

VALIDATION FINDINGS WORKSHEET
Sample Calculation Verification

Page: 1 of 1
Reviewer: MV
2nd reviewer: [illegible]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y N N/A Have results been reported and calculated correctly?
Y N N/A Are results within the calibrated range of the instruments and within the linear range of the ICP?
Y N N/A Are all detection limits below the CRDL?

Detected analyte results for 7 were recalculated and verified using the following equation:

Concentration = $\frac{(RD)(FV)(Dil)}{(In.\ Vol.)(\%S)}$

RD = Raw data concentration
FV = Final volume (ml)
In. Vol. = Initial volume (ml) or weight (G)
Dil = Dilution factor
%S = Decimal percent solids

Recalculation:

From the raw data

Cr = 1.370 ug/l

Sample ID	Analyte	Reported Concentration (ug/l)	Calculated Concentration (ug/l)	Acceptable (Y/N)
7	Cr	1.4	1.4	Y

RECALC.4S2