Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL

Collection Date: August 25, 2003

LDC Report Date: October 15, 2003

Matrix: Water

Parameters: Volatiles

Validation Level: EPA Level III

Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-4842

Sample Identification

MW-5
MW-8
MW-10
TB-8-8-25-03
MW-8MS
MW-8MSD

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

J Indicates an estimated value.

R Quality control indicates the data is not usable.

N Presumptive evidence of presence of the constituent.

UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-8-8-25-03 was identified as a trip blank. No volatile contaminants were found in this blank.

NASA JPL
Volatiles - Data Qualification Summary - SDG 03-4842

No Sample Data Qualified in this SDG

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-4842

No Sample Data Qualified in this SDG

1.962-D

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	08/25/2003
Project ID:	JPL	Service ID:	34842	Collected by:	J. Robinson
		Lab Sample ID:	03-4842-1	Received Date:	08/25/2003
Sample ID:	MW-5	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	09/03/03	Anal. Date:	09/03/03
Batch No:	03G4023	Prep. No:	–	Anal. Time:	14:44
Data File Name:	4842-01	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

10/18/03

Continued

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	< 0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	< 0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	97
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	107
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	102
4	TOLUENE-D8	2037-26-5	73-129	94
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	89
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	89
3	FLUOROBENZENE	462-06-6	50-200	87
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

10/8/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	08/25/2003
Project ID:	JPL	Service ID:	34842	Collected by:	J. Robinson
		Lab Sample ID:	03-4842-2	Received Date:	08/25/2003
Sample ID:	MW-8	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	09/03/03	Anal. Date:	09/03/03
Batch No:	03G4023	Prep. No:	–	Anal. Time:	15:11
Data File Name:	4842-02	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

10/18/03

Continued

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	< 0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	< 0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	96
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	107
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	103
4	TOLUENE-D8	2037-26-5	73-129	95
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	88
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	89
3	FLUOROBENZENE	462-06-6	50-200	86
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

10/18/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	08/25/2003
Project ID:	JPL	Service ID:	34842	Collected by:	J. Robinson
		Lab Sample ID:	03-4842-3	Received Date:	08/25/2003
Sample ID:	MW-10	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	09/03/03	Anal. Date:	09/03/03
Batch No:	03G4023	Prep. No:	–	Anal. Time:	15:39
Data File Name:	4842-03	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	0.3	J
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	1.3	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	0.6	
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

10/18/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	0.9	
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	12.3	
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	95
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	110
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	104
4	TOLUENE-D8	2037-26-5	73-129	93
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	90
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	91
3	FLUOROBENZENE	462-06-6	50-200	85
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

10/18/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	08/25/2003
Project ID:	JPL	Service ID:	34842	Collected by:	J. Robinson
		Lab Sample ID:	03-4842-4	Received Date:	08/25/2003
Sample ID:	**TB-8-8-25-03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	09/03/03	Anal. Date:	09/03/03
Batch No:	03G4023	Prep. No:	–	Anal. Time:	13:21
Data File Name:	4842-04	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

Continued

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	< 0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	< 0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	96
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	106
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	102
4	TOLUENE-D8	2037-26-5	73-129	95
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	90
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	91
3	FLUOROBENZENE	462-06-6	50-200	87
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

10/18/03

LDC #: 10962D1

SDG #: 03-4842

Laboratory: Applied Physics & Chemistry Laboratory

VALIDATION COMPLETENESS WORKSHEET

Level III

Date: 10/9/03

Page: 1 of 1

Reviewer:

2nd Reviewer:

METHOD: GC/MS Volatiles (EPA ~~SW 846~~ Method ~~8260B~~ SOM1.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 8/25/03
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	%RSD . r² no RRF s
IV.	Continuing calibration	A	
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	ND	TB = 4

Note: A = Acceptable; N = Not provided/applicable; SW = See worksheet; ND = No compounds detected; R = Rinsate; FB = Field blank; D = Duplicate; TB = Trip blank; EB = Equipment blank

Validated Samples:

1	MW-5 W	11	0364023HB01	21		31	
2	MW-8	12		22		32	
3	MW-10	13		23		33	
4	TB-8-8-25-03	14		24		34	
5	MW-8MS	15		25		35	
6	MW-8MSD	16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL

Collection Date: August 26, 2003

LDC Report Date: October 15, 2003

Matrix: Water

Parameters: Volatiles

Validation Level: EPA Level III

Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-4854

Sample Identification

MW-7
MW-13
MW-16
TB-9-8-26-03

Introduction

This data review covers 4 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

J Indicates an estimated value.

R Quality control indicates the data is not usable.

N Presumptive evidence of presence of the constituent.

UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-9-8-26-03 was identified as a trip blank. No volatile contaminants were found in this blank.

NASA JPL
Volatiles - Data Qualification Summary - SDG 03-4854

No Sample Data Qualified in this SDG

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-4854

No Sample Data Qualified in this SDG

1.962E

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	08/26/2003
Project ID:		Service ID:	34854	Collected by:	
		Lab Sample ID:	03-4854-1	Received Date:	08/26/2003
Sample ID:	MW-7	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	09/03/03	Anal. Date:	09/03/03
Batch No:	03G4023	Prep. No:	–	Anal. Time:	16:07
Data File Name:	4854-01	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	40.4	
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	6.8	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	2.2	
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

10/18/05

Continued

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	4.9	
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	4.5	
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	2.2	
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	97
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	106
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	104
4	TOLUENE-D8	2037-26-5	73-129	94
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	91
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	90
3	FLUOROBENZENE	462-06-6	50-200	86
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

10/18/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	08/26/2003
Project ID:		Service ID:	34854	Collected by:	
		Lab Sample ID:	03-4854-2	Received Date:	08/26/2003
Sample ID:	**MW-13**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	09/03/03	Anal. Date:	09/03/03
Batch No:	03G4023	Prep. No:	–	Anal. Time:	16:34
Data File Name:	4854-02	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	< 0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	< 0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	< 0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	0.4	J
5	BROMOFORM	75-25-2	µg/L	0.5	< 0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	< 0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	< 0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	< 0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	< 0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	< 10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	1.0	
12	CHLOROBENZENE	108-90-7	µg/L	0.5	< 0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	0.8	
14	CHLOROETHANE	75-00-3	µg/L	0.5	< 0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	3.3	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	< 0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	< 0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	< 0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	< 0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	< 0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	< 0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	< 0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	< 0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	< 0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	< 0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	< 0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	< 0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	< 0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	< 0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	< 0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	< 0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	< 0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	< 0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	< 0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	< 0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	< 0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	< 0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	< 0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	< 0.5	U

10/18/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	0.8	
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	20.0	
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	95
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	107
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	103
4	TOLUENE-D8	2037-26-5	73-129	94
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	89
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	90
3	FLUOROBENZENE	462-06-6	50-200	86
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

10/18/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	08/26/2003
Project ID:		Service ID:	34854	Collected by:	
		Lab Sample ID:	03-4854-3	Received Date:	08/26/2003
Sample ID:	**MW-16**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	09/03/03	Anal. Date:	09/03/03
Batch No:	03G4023	Prep. No:	–	Anal. Time:	17:02
Data File Name:	4854-03	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	< 0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	< 0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	< 0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	< 0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	< 0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	< 0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	< 0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	< 0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	< 0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	< 10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	1.9	
12	CHLOROBENZENE	108-90-7	μg/L	0.5	< 0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	0.4	J
14	CHLOROETHANE	75-00-3	μg/L	0.5	< 0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	3.5	
16	CHLOROMETHANE	74-87-3	μg/L	0.5	< 0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	< 0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	< 0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	< 0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	< 0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	< 0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	< 0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	< 0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	< 0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	< 0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	< 0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	< 0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	< 0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	< 0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	< 0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	< 0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	< 0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	< 0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	< 0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	< 0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	< 0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	< 0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	< 0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	< 0.5	U

10/18/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	μg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	μg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	< 0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	3.7	
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	< 0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	μg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	< 0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	97
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	108
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	103
4	TOLUENE-D8	2037-26-5	73-129	94
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	90
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	90
3	FLUOROBENZENE	462-06-6	50-200	87
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

10/18/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	08/26/2003
Project ID:		Service ID:	34854	Collected by:	
		Lab Sample ID:	03-4854-4	Received Date:	08/26/2003
Sample ID:	**TB-9-8-26-03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	09/03/03	Anal. Date:	09/03/03
Batch No:	03G4023	Prep. No:	–	Anal. Time:	12:54
Data File Name:	4854-04	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	< 0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	< 0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	< 0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	< 0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	< 0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	< 0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	< 0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	< 0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	< 0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	< 10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	< 0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	< 0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	< 0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	< 0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	< 0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	< 0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	< 0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	< 0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	< 0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	< 0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	< 0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	< 0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	< 0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	< 0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	< 0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	< 0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	< 0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	< 0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	< 0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	< 0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	< 0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	< 0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	< 0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	< 0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	< 0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	< 0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	< 0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	< 0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	< 0.5	U

9/10/18/05

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	96
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	101
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	100
4	TOLUENE-D8	2037-26-5	73-129	95
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	90
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	90
3	FLUOROBENZENE	462-06-6	50-200	88
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

10/18/03

LDC #: 10962E1

SDG #: 03-4854

Laboratory: Applied Physics & Chemistry Laboratory

VALIDATION COMPLETENESS WORKSHEET

Level III

Date: 10/9/03

Page: 1 of 1

Reviewer:

2nd Reviewer:

METHOD: GC/MS Volatiles (EPA ~~SW 846~~ Method ~~8260B~~ 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 8/26/03
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	%RSD . r², NO RRFS
IV.	Continuing calibration	A	↓
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	MW-8 (03-4842)
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	ND	TB = 4

Note: A = Acceptable; N = Not provided/applicable; SW = See worksheet; ND = No compounds detected; R = Rinsate; FB = Field blank; D = Duplicate; TB = Trip blank; EB = Equipment blank

Validated Samples:

1	MW-7 W	11	0344023MB01	21		31	
2	MW-13 ↓	12		22		32	
3	MW-16 ↓	13		23		33	
4	TB-9-8-26-03 ↓	14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL

Collection Date: August 27, 2003

LDC Report Date: October 15, 2003

Matrix: Water

Parameters: Volatiles

Validation Level: EPA Level III

Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-4878

Sample Identification

MW-6
TB-10-8-27-03

Introduction

This data review covers 2 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

J Indicates an estimated value.

R Quality control indicates the data is not usable.

N Presumptive evidence of presence of the constituent.

UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-10-8-27-03 was identified as a trip blank. No volatile contaminants were found in this blank.

NASA JPL
Volatiles - Data Qualification Summary - SDG 03-4878

No Sample Data Qualified in this SDG

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-4878

No Sample Data Qualified in this SDG

10962F

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	08/27/2003
Project ID:	JPL	Service ID:	34878	Collected by:	
		Lab Sample ID:	03-4878-2	Received Date:	08/27/2003
Sample ID:	MW-6	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	09/03/03	Anal. Date:	09/03/03
Batch No:	03G4023	Prep. No:	–	Anal. Time:	14:16
Data File Name:	4878-02	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	0.3	J
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	0.7	
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

10/16/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	2.3	
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	97
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	106
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	102
4	TOLUENE-D8	2037-26-5	73-129	93
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	90
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	90
3	FLUOROBENZENE	462-06-6	50-200	87
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

10/18/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	08/27/2003
Project ID:	JPL	Service ID:	34878	Collected by:	
		Lab Sample ID:	03-4878-4	Received Date:	08/27/2003
Sample ID:	**TB-10-8-27-03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	09/03/03	Anal. Date:	09/03/03
Batch No:	03G4023	Prep. No:	–	Anal. Time:	13:49
Data File Name:	4878-04	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

10/15/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	95
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	107
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	103
4	TOLUENE-D8	2037-26-5	73-129	94
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	90
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	91
3	FLUOROBENZENE	462-06-6	50-200	85
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

10/18/07

LDC #: 10962F1

SDG #: 03-4878

Laboratory: Applied Physics & Chemistry Laboratory

VALIDATION COMPLETENESS WORKSHEET

Level III

Date: 10/9/03

Page: 1 of 1

Reviewer: [signature]

2nd Reviewer: [signature]

METHOD: GC/MS Volatiles (EPA ~~SW 846~~ Method ~~8260B~~ 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 8/27/03
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	%RSD. r^2 no RFs
IV.	Continuing calibration	A	↓
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	MW-8(03-4842
VIII.	Laboratory control samples	A	
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	ND	TB = 2

Note: A = Acceptable; N = Not provided/applicable; SW = See worksheet; ND = No compounds detected; R = Rinsate; FB = Field blank; D = Duplicate; TB = Trip blank; EB = Equipment blank

Validated Samples:

1	MW-6 W	11	0344023MB01	21		31	
2	TB-10-8-27-03 ↓	12		22		32	
3		13		23		33	
4		14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

NASA JPL
Data Validation Reports
LDC# 10962

Chromium

LDC Report# 10962A4

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL

Collection Date: July 29 through July 31, 2003

LDC Report Date: October 17, 2003

Matrix: Water

Parameters: Chromium

Validation Level: EPA Level III

Laboratory: Applied P & Ch Laboratory/Advanced Technology Laboratories

Sample Delivery Group (SDG): 03-4445/064234

Sample Identification

MW-21-5	MW-20-4MS
MW-21-4	MW-20-4MSD
MW-21-3	MW-20-4DUP
MW-21-2	
MW-21-1	
EB-1-7-29-03	
MW-3-4	
MW-3-3	
MW-3-2	
MW-17-4	
MW-17-3	
MW-17-2	
MW-20-5	
MW-20-4	
MW-20-3	
MW-20-2	
MW-20-1	
DUPE-2-3-Q03	
EB-3-7-31-03	
MW-17-4DUP	

Introduction

This data review covers 23 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

J Indicates an estimated value.

R Quality control indicates the data is not usable.

N Presumptive evidence of presence of the constituent.

UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Chromium	0.115 ug/L	All samples in SDG 03-4445/064234

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-3-7-31-03	Chromium	0.2 ug/L	0.2U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check was not required by the method.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

Spike ID (Associated Samples)	Analyte	MS (%R) (Limits)	MSD (%R) (Limits)	RPD (Limits)	Flag	A or P
MW-20-4MS/MSD (All samples in SDG 03-4445/064234)	Chromium	79.9 (80-120)	-	-	J (all detects) UJ (all non-detects)	A

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

VIII. Internal Standards (ICP-MS)

All internal standard percent recoveries (%R) were within QC limits.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

ICP serial dilution was not required by the method.

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

Samples MW-20-3 and DUPE-2-3-Q03 were identified as field duplicates. No chromium was detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	MW-20-3	DUPE-2-3-Q03	
Chromium	4.0	4.0	0

XIV. Field Blanks

Samples EB-1-7-29-03 and EB-3-7-31-03 were identified as equipment blanks. No chromium contaminants were found in these blanks with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-3-7-31-03	Chromium	0.2

NASA JPL
Chromium - Data Qualification Summary - SDG 03-4445/064234

SDG	Sample	Analyte	Flag	A or P	Reason
03-4445/064234	MW-21-5 MW-21-4 MW-21-3 MW-21-2 MW-21-1 EB-1-7-29-03 MW-3-4 MW-3-3 MW-3-2 MW-17-4 MW-17-3 MW-17-2 MW-20-5 MW-20-4 MW-20-3 MW-20-2 MW-20-1 DUPE-2-3-Q03 EB-3-7-31-03	Chromium	J (all detects) UJ (all non-detects)	A	Matrix spike/Matrix spike duplicates (%R)

NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 03-4445/064234

SDG	Sample	Analyte	Modified Final Concentration	A or P
03-4445/064234	EB-3-7-31-03	Chromium	0.2U ug/L	A

1096~~7~~A

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-21-5
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/29/2003 11:49:00 AM
Lab ID:	064234-001A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813A	QC Batch: R30051			PrepDate:			
Chromium	2.9	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-21-4
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/29/2003 12:21:00 PM
Lab ID:	064234-002A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813A QC Batch: R30051			PrepDate:				
Chromium	4.0	J	0.11	1.0	µg/L	1	8/13/2003

10/18/10

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-21-3
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/29/2003 12:50:00 PM
Lab ID:	064234-003A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A	QC Batch: R30051		PrepDate:				
Chromium	3.7	J	0.11	1.0	µg/L	1	8/13/2003

9/10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-21-2
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/29/2003 1:27:00 PM
Lab ID:	064234-004A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A	QC Batch: R30051		PrepDate:				
Chromium	4.2	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-21-1
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/29/2003 1:55:00 PM
Lab ID:	064234-005A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A QC Batch: R30051			PrepDate:				
Chromium	3.8	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	EB-1-7-29-03
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/29/2003 12:30:00 PM
Lab ID:	064234-006A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A	QC Batch: R30051		PrepDate:				
Chromium	ND	UJ	0.11	1.0	µg/L	1	8/13/2003

10/12/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-3-4
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/30/2003 10:05:00 AM
Lab ID:	064234-007A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			**Analyst: NS**	
RunID: ICP4_030813A	QC Batch: R30051		PrepDate:				
Chromium	1.8	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-3-3
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/30/2003 10:26:00 AM
Lab ID:	064234-008A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A	QC Batch: R30051		PrepDate:				
Chromium	2.0	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-3-2
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/30/2003 10:40:00 AM
Lab ID:	064234-009A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A QC Batch: R30051			PrepDate:				
Chromium	2.4	J	0.11	1.0	µg/L	1	8/13/2003

9/10/18/07

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories *3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040*

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-17-4
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/30/2003 11:28:00 AM
Lab ID:	064234-010A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A	QC Batch: R30051			PrepDate:			
Chromium	1.9	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-17-3
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/30/2003 11:54:00 AM
Lab ID:	064234-011A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A	QC Batch: R30051		PrepDate:				
Chromium	4.0	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-17-2
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/30/2003 12:15:00 PM
Lab ID:	064234-012A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A QC Batch: R30051			PrepDate:				
Chromium	2.6	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-20-5
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/31/2003 8:32:00 AM
Lab ID:	064234-013A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A QC Batch: R30051			PrepDate:				
Chromium	1.6	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-20-4
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/31/2003 9:35:00 AM
Lab ID:	064234-014A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A	QC Batch: R30051		PrepDate:				
Chromium	1.9	J	0.11	1.0	µg/L	1	8/13/2003

8/10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-20-3
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/31/2003 11:13:00 AM
Lab ID:	064234-015A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A	QC Batch: R30051			PrepDate:			
Chromium	4.0	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-20-2
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/31/2003 12:03:00 PM
Lab ID:	064234-016A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A	QC Batch: R30051		PrepDate:				
Chromium	1.5	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-20-1
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/31/2003 12:30:00 PM
Lab ID:	064234-017A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A QC Batch: R30051			PrepDate:				
Chromium	1.8	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	Dupe-2-3-Q03
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/31/2003 11:38:00 AM
Lab ID:	064234-018A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A QC Batch: R30051			PrepDate:				
Chromium	4.0	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories *3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040*

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	EB-3-7-31-03
Lab Order:	064234	**Tag Number:**	
Project:	#4445, JPL	**Collection Date:**	7/31/2003 8:49:00 AM
Lab ID:	064234-019A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813A QC Batch: R30051			PrepDate:				
Chromium	0.2	J, U	0.11	1.0	µg/L	1	8/13/2003

10/8/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

LDC #: 10962A4

VALIDATION COMPLETENESS WORKSHEET

Level III

SDG #: 03-4445/064234

Laboratory: Applied Physics & Chemistry Laboratory/Advanced Technology Laboratories

Date: 10/15/03

Page: 1 of 1

Reviewer: HO

2nd Reviewer: [illegible]

METHOD: Chromium (EPA Method 200.8)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 7/29 - 31/03
II.	Calibration	A	
III.	Blanks	SW	
IV.	ICP Interference Check Sample (ICS) Analysis	N	Not required
V.	Matrix Spike Analysis	SW	
VI.	Duplicate Sample Analysis	A	
VII.	Laboratory Control Samples (LCS)	A	LCS
VIII.	Internal Standard (ICP-MS)	~~N~~ A	~~Not reviewed~~ HO
IX.	Furnace Atomic Absorption QC	N	Not utilized
X.	ICP Serial Dilution	N	Not required
XI.	Sample Result Verification	N	
XII.	Overall Assessment of Data	A	
XIII.	Field Duplicates	SW	(15, 18)
XIV.	Field Blanks	SW	EB = 6, 19

Note: A = Acceptable; N = Not provided/applicable; SW = See worksheet; ND = No compounds detected; R = Rinsate; FB = Field blank; D = Duplicate; TB = Trip blank; EB = Equipment blank

Validated Samples: All

1	MW-21-5	11	MW-17-3	21	MW-20-4MS	31	
2	MW-21-4	12	MW-17-2	22	MW-20-4MSD	32	
3	MW-21-3	13	MW-20-5	23	MW-20-4DUP	33	
4	MW-21-2	14	MW-20-4	24		34	
5	MW-21-1	15 ✓	MW-20-3	25		35	
6	EB-1-7-29-03	16	MW-20-2	26		36	
7	MW-3-4	17	MW-20-1	27		37	
8	MW-3-3	18 ✓	DUPE-2-3-Q03	28		38	
9	MW-3-2	19	EB-3-7-31-03	29		39	
10	MW-17-4	20	MW-17-4DUP	30		40	

Notes:

LDC #: 10962A4
SDG #: 03-4445/064443

VALIDATION FINDINGS WORKSHEET
PB/ICB/CCB QUALIFIED SAMPLES

Page: 1 of 1
Reviewer: MH
2nd Reviewer: A

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: _____
Sample Concentration units, unless otherwise noted: ug/l Associated Samples: All

					Sample Identification										
Analyte	Maximum PB* (mg/Kg)	Maximum PB* (ug/L)	Maximum ICB/CCB* (ug/L)	Blank Action Limit	19										
Al															Al
Sb															Sb
As															As
Ba															Ba
Be															Be
Cd															Cd
Ca															Ca
Cr		0.115		0.575	0.2										Cr
Cc															Co
Cu															Cu
Fe															Fe
Pb															Pb
Mg															Mg
Mn															Mn
Hg															Hg
Ni															Ni
K															K
Se															Se
Ag															Ag
Na															Na
Tl															Tl
V															V
Zn															Zn
B															B
Mo															Mo
Sr															Sr

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".
Note : a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

BLNKSMP.4S2

LDC #: 10962A21
SDG #: 03-4445/064234

VALIDATION FINDINGS WORKSHEET
Matrix Spike/Matrix Spike Duplicates

Page: 1 of 1
Reviewer: [illegible]
2nd Reviewer: [illegible]

METHOD: Trace metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y N N/A Was a matrix spike analyzed for each matrix in this SDG?

Y N N/A Were matrix spike percent recoveries (%R) within the control limits of 75-125? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken. Lab limits

Y N N/A Were all duplicate sample relative percent differences (RPD) ≤ 20% for water samples and ≤35% for soil samples?

LEVEL IV ONLY:

Y N N/A Were recalculated results acceptable? See Level IV Recalculation Worksheet for recalculations.

#	MS/MSD ID	Matrix	Analyte	MS %Recovery	MSD %Recovery	RPD (Limits)	Associated Samples	Qualifications
1	21/22	AQ	Cr	79.9 (80-120)			All	J/UJ/A

Comments:

LDC #: 10962A4
SDG #: 03-4445-106423-4

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
Reviewer: MH
2nd reviewer: A

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Y N N/A Were field duplicate pairs identified in this SDG?
Y N N/A Were target analytes detected in the field duplicate pairs?

Analyte	Concentration (ug/l)		RPD (Limits)	Difference (Limits)	Qualifications
	15	18			
Aluminum					
Antimony					
Arsenic					
Barium					
Beryllium					
Cadmium					
Calcium					
Chromium	4.0	4.0	0		
Cobalt					
Copper					
Iron					
Lead					
Magnesium					
Manganese					
Mercury					
Nickel					
Potassium					
Selenium					
Silver					
Sodium					
Thallium					
Vanadium					
Zinc					
Cyanide					
Boron					
Molybdenum					
Strontium					
Silicon					

Notes:

FLDUP.4S2

LDC #: 10962A4
SDG #: 03-4440/0 64234

VALIDATION FINDINGS WORKSHEET
Field Blanks

Page: 1 of 1
Reviewer: MH
2nd reviewer: [illegible]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Y N N/A Were field blanks identified in this SDG?
Y N N/A Were target analytes detected in the field blanks?

Sample: 19 Field Blank / Trip Blank / Rinsate / Other ZB (circle one)

Analyte	Concentration Units (ug/L)
Cr	0.2

Sample: _____ Field Blank / Trip Blank / Rinsate / Other _____ (circle one)

Analyte	Concentration Units ( )

FLDBLK2.4SW

LDC Report# 10962B4

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL

Collection Date: August 4 through August 7, 2003

LDC Report Date: October 17, 2003

Matrix: Water

Parameters: Chromium

Validation Level: EPA Level III

Laboratory: Applied P & Ch Laboratory/Advanced Technology Laboratories

Sample Delivery Group (SDG): 03-4558/064235

Sample Identification

MW-18-4	MW-11-1
MW-18-3	MW-24-4
MW-18-2	MW-24-3
EB-4-8-03	MW-24-2
MW-4-3	MW-24-1
MW-4-2	MW-4-3MS
MW-4-1	MW-4-3MSD
DUPE-3-3Q03	MW-4-3DUP
MW-23-4	MW-4-2MS
MW-23-3	MW-4-2MSD
MW-23-2	MW-4-2DUP
MW-23-1	MW-14-2MS
EB-5-8-6-03	MW-14-2MSD
MW-14-4	MW-14-2DUP
MW-14-3	MW-24-4DUP
MW-14-2	
MW-14-1	
DUPE-4-3-Q03	
MW-11-3	
MW-11-2	

Introduction

This data review covers 35 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

J Indicates an estimated value.

R Quality control indicates the data is not usable.

N Presumptive evidence of presence of the constituent.

UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Chromium	0.1512 ug/L	MW-4-3

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
MW-4-3	Chromium	0.4 ug/L	0.4U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check was not required by the method.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

Spike ID (Associated Samples)	Analyte	MS (%R) (Limits)	MSD (%R) (Limits)	RPD (Limits)	Flag	A or P
MW-4-2MS/MSD (MW-18-4 MW-18-3 MW-18-2 EB-4-8-03 MW-4-2 MW-4-1 DUPE-3-3Q03 MW-23-4 MW-23-3 MW-23-2 MW-23-1 EB-5-8-6-03 MW-14-4 MW-14-3 MW-14-1 DUPE-4-3-Q03 MW-11-3 MW-11-2 MW-11-1 MW-24-4)	Chromium	71.7 (80-120)	72 (80-120)	-	J (all detects) UJ (all non-detects)	A

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

VIII. Internal Standards (ICP-MS)

All internal standard percent recoveries (%R) were within QC limits.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

ICP serial dilution was not required by the method.

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

Samples MW-4-1 and DUPE-3-3Q03 and samples MW-14-3 and DUPE-4-3-Q03 were identified as field duplicates. No chromium was detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	MW-4-1	DUPE-3-3Q03	
Chromium	2.7	2.5	8

Analyte	Concentration (ug/L)		RPD
	MW-14-3	DUPE-4-3-Q03	
Chromium	3.6	3.4	6

XIV. Field Blanks

Samples EB-4-8-03 and EB-5-8-6-03 were identified as equipment blanks. No chromium contaminants were found in these blanks.

NASA JPL
Chromium - Data Qualification Summary - SDG 03-4558/064235

SDG	Sample	Analyte	Flag	A or P	Reason
03-4558/064235	MW-18-4 MW-18-3 MW-18-2 EB-4-8-03 MW-4-2 MW-4-1 DUPE-3-3Q03 MW-23-4 MW-23-3 MW-23-2 MW-23-1 EB-5-8-6-03 MW-14-4 MW-14-3 MW-14-1 DUPE-4-3-Q03 MW-11-3 MW-11-2 MW-11-1 MW-24-4	Chromium	J (all detects) UJ (all non-detects)	A	Matrix spike/Matrix spike duplicates (%R)

NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 03-4558/064235

SDG	Sample	Analyte	Modified Final Concentration	A or P
03-4558/064235	MW-4-3	Chromium	0.4U ug/L	A

10962B

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-18-4
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/4/2003 8:30:00 AM
Lab ID:	064235-001A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813B QC Batch: R30052			PrepDate:				
Chromium	2.7	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-18-3
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/4/2003 8:51:00 AM
Lab ID:	064235-002A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analystt **NS**	
RunID: ICP4_030813B QC Batch: R30052			PrepDate:				
Chromium	5.9	J	0.11	1.0	µg/L	1	8/13/2003

10/18/05

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-18-2
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/4/2003 9:12:00 AM
Lab ID:	064235-003A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813B	QC Batch: R30052		PrepDate:				
Chromium	2.1	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	EB-4-8-03
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/4/2003 8:40:00 AM
Lab ID:	064235-004A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813B	QC Batch: R30052		PrepDate:				
Chromium	ND	U J.	0.11	1.0	µg/L	1	8/13/2003

10/18/05

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-4-3
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/4/2003 10:10:00 AM
Lab ID:	064235-005A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**	**(EPA 200.8)**		Analyst: **NS**	
RunID: ICP4_030814B QC Batch: 14733				PrepDate: 8/13/2003			
Chromium	0.4	J U	0.11	1.0	µg/L	1	8/14/2003

10/18/02

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-4-2
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/4/2003 10:31:00 AM
Lab ID:	064235-006A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813B	QC Batch: R30052		PrepDate:				
Chromium	5.2	J	0.11	1.0	µg/L	1	8/13/2003

10/18/07

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-4-1
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/4/2003 11:04:00 AM
Lab ID:	064235-007A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813B QC Batch: R30052			PrepDate:				
Chromium	2.7	J	0.11	1.0	µg/L	1	8/13/2003

10/18/05

Qualifiers:

ND - Not Detected at the Reporting Limit.
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories *3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040*

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	Dupe-3-3Q03
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/4/2003 11:29:00 AM
Lab ID:	064235-008A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813B	QC Batch: R30052		PrepDate:				
Chromium	2.5	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-23-4
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/6/2003 7:26:00 AM
Lab ID:	064235-009A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813B QC Batch: R30052			PrepDate:				
Chromium	2.6	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-23-3
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/6/2003 7:48:00 AM
Lab ID:	064235-010A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813B	QC Batch: R30052			PrepDate:			
Chromium	3.5	J	0.11	1.0	µg/L	1	8/13/2003

10/18/02

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-23-2
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/6/2003 8:09:00 AM
Lab ID:	064235-011A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813B	QC Batch: R30052			PrepDate:			
Chromium	3.9	J	0.11	1.0	µg/L	1	8/13/2003

10/18/05

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories *3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040*

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-23-1
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/6/2003 8:24:00 AM
Lab ID:	064235-012A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813B	QC Batch: R30052			PrepDate:			
Chromium	4.2	J	0.11	1.0	µg/L	1	8/13/2003

10/18/05

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	EB-5-8-6-03
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/6/2003 7:41:00 AM
Lab ID:	064235-013A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813B QC Batch: R30052			PrepDate:				
Chromium	ND	UJ	0.11	1.0	µg/L	1	8/13/2003

10/18/05

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-14-4
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/6/2003 9:26:00 AM
Lab ID:	064235-014A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813B	QC Batch: R30052			PrepDate:			
Chromium	1.6	J	0.11	1.0	µg/L	1	8/13/2003

10/18/05

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-14-3
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/6/2003 9:50:00 AM
Lab ID:	064235-015A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813B	QC Batch: R30052		PrepDate:				
Chromium	3.6	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-14-2
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/6/2003 10:41:00 AM
Lab ID:	064235-016A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813C	QC Batch: R30053			PrepDate:			
Chromium	1.9		0.11	1.0	µg/L	1	8/13/2003

10/18/05

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology 3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-14-1
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/6/2003 11:23:00 AM
Lab ID:	064235-017A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813B	QC Batch: R30052			PrepDate:			
Chromium	3.9	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	Dupe-4-3-Q03
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/6/2003 10:18:00 AM
Lab ID:	064235-018A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813B	QC Batch: R30052			PrepDate:			
Chromium	3.4	J	0.11	1.0	µg/L	1	8/13/2003

10/18/05

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-11-3
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/7/2003 8:06:00 AM
Lab ID:	064235-019A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813B	QC Batch: R30052			PrepDate:			
Chromium	2.3	J	0.11	1.0	µg/L	1	8/13/2003

10/18/05

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-11-2
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/7/2003 8:24:00 AM
Lab ID:	064235-020A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813B QC Batch: R30052				PrepDate:			
Chromium	1.5	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories *3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040*

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-11-1
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/7/2003 8:44:00 AM
Lab ID:	064235-021A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813B	QC Batch: R30052			PrepDate:			
Chromium	2.0	J	0.11	1.0	µg/L	1	8/13/2003

10/18/05

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-24-4
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/7/2003 9:47:00 AM
Lab ID:	064235-022A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813B QC Batch: R30052				PrepDate:			
Chromium	0.7	J	J	0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue · Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-24-3
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/7/2003 10:03:00 AM
Lab ID:	064235-023A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813C	QC Batch: R30053			PrepDate:			
Chromium	1.3		0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories *3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040*

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-24-2
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/7/2003 10:47:00 AM
Lab ID:	064235-024A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**				Analyst: **NS**
RunID: ICP4_030813C	QC Batch: R30053			PrepDate:			
Chromium	2.0		0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories *3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040*

Advanced Technology Laboratories

Date: *19-Aug-03*

CLIENT:	Applied P & Ch Laboratories	**Client Sample ID:**	MW-24-1
Lab Order:	064235	**Tag Number:**	
Project:	#4558, JPL	**Collection Date:**	8/7/2003 10:52:00 AM
Lab ID:	064235-025A	**Matrix:**	WATER

Analyte	Result	Qual	MDL	PQL	Units	DF	Date Analyzed
ICP-MS METALS			**EPA 200.8**			Analyst: **NS**	
RunID: ICP4_030813C QC Batch: R30053			PrepDate:				
Chromium	3.0		0.11	1.0	µg/L	1	8/13/2003

10/18/03

Qualifiers:

ND - Not Detected at the Reporting Limit
J - Analyte detected below quantitation limits
B - Analyte detected in the associated Method Blank
* - Value exceeds Maximum Contaminant Level
S - Spike Recovery outside accepted recovery limits
R - RPD outside accepted recovery limits
E - Value above quantitation range
H-Samples exceed holding time

Results are wet unless otherwise specified

Advanced Technology Laboratories

3275 Walnut Avenue Signal Hill, CA 90755 Tel: 562 989-4045 Fax: 562 989-4040

LDC #: 10962B4

VALIDATION COMPLETENESS WORKSHEET

Date: 10/15/03

SDG #: 03-4558/064235

Level III

Page: 1 of 1

Laboratory: Applied Physics & Chemistry Laboratory/Advanced Technology Laboratories

Reviewer: MW

2nd Reviewer: /

METHOD: Chromium (EPA Method 200.8)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 8/4-7/03
II.	Calibration	A	
III.	Blanks	SW	
IV.	ICP Interference Check Sample (ICS) Analysis	N	Not required
V.	Matrix Spike Analysis	SW	
VI.	Duplicate Sample Analysis	A	
VII.	Laboratory Control Samples (LCS)	A	LCS
VIII.	Internal Standard (ICP-MS)	A	
IX.	Furnace Atomic Absorption QC	N	Not utilized
X.	ICP Serial Dilution	N	Not required.
XI.	Sample Result Verification	N	
XII.	Overall Assessment of Data	A	
XIII.	Field Duplicates	SW	(7,8), (15,18)
XIV.	Field Blanks	ND	EB=4, 13.

Note: A = Acceptable; N = Not provided/applicable; SW = See worksheet; ND = No compounds detected; R = Rinsate; FB = Field blank; D = Duplicate; TB = Trip blank; EB = Equipment blank

Validated Samples: 18

1	MW-18-4	11	MW-23-2	21	MW-11-1	31	MW-4-2DUP
2	MW-18-3	12	MW-23-1	22	MW-24-4	32	MW-14-2MS
3	MW-18-2	13	EB-5-8-6-03	23 3	MW-24-3	33	MW-14-2MSD
4	EB-4-8-03	14	MW-14-4	24 3	MW-24-2	34	MW-14-2DUP
5 3	MW-4-3	15 ✓	MW-14-3	25 3	MW-24-1	35	MW-24-4DUP
6	MW-4-2	3 16	MW-14-2	26	MW-4-3MS	36	FB
7 ✓	MW-4-1	17	MW-14-1	27	MW-4-3MSD	37	
8 ✓	DUPE-3-3Q03	18 ✓	DUPE-4-3-Q03	28	MW-4-3DUP	38	
9	MW-23-4	19	MW-11-3	29	MW-4-2MS	39	
10	MW-23-3	20	MW-11-2	30	MW-4-2MSD	40	

Notes:

LDC #: 16962B4
SDG #: 03-4558/064235

VALIDATION FINDINGS WORKSHEET
PB/ICB/CCB QUALIFIED SAMPLES

Page: 1 of 1
Reviewer: MH
2nd Reviewer: [illegible]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: _____
Sample Concentration units, unless otherwise noted: ug/L Associated Samples: 5

					Sample Identification										
Analyte	Maximum PB[a] (mg/Kg)	Maximum PB[a] (ug/L)	Maximum ICB/CCB[a] (ug/L)	Blank Action Limit	5										
Al															Al
Sb															Sb
As															As
Ba															Ba
Be															Be
Cd															Cd
Ca															Ca
Cr		0.1512		0.756	0.4										Cr
Cc															Co
Cu															Cu
Fe															Fe
Pb															Pb
Mg															Mg
Mn															Mn
Hg															Hg
Ni															Ni
K															K
Se															Se
Ag															Ag
Na															Na
Tl															Tl
V															V
Zn															Zn
B															B
Mo															Mo
Sr															Sr

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".
Note : a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

BLNKSMP.4S2

LDC #: 10962 B4
SDG #: 03-4558/064235

VALIDATION FINDINGS WORKSHEET
Matrix Spike/Matrix Spike Duplicates

Page: 1 of 1
Reviewer: MH
2nd Reviewer: [illegible]

METHOD: Trace metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y N N/A Was a matrix spike analyzed for each matrix in this SDG?

Y N N/A Were matrix spike percent recoveries (%R) within the control limits of 75-125? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken. lab limit 80-120

Y N N/A Were all duplicate sample relative percent differences (RPD) ≤ 20% for water samples and ≤35% for soil samples?

LEVEL IV ONLY:

Y N N/A Were recalculated results acceptable? See Level IV Recalculation Worksheet for recalculations.

#	MS/MSD ID	Matrix	Analyte	MS %Recovery	MSD %Recovery	RPD (Limits)	Associated Samples	Qualifications
1	29/30	Aq	Cr	71.7 (80-120)	72 (80-120)		1-4, 6-15, 17-22	J/UJ/A

Comments:

MSD.4S2

LDC #: 10962 B4
SDG #: 03-4558/064235

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
Reviewer: WW
2nd reviewer: N

METHOD: ~~Inorganics~~ Cr, Method 200.8

Y N N/A Were field duplicate pairs identified in this SDG?
Y N N/A Were target analytes detected in the field duplicate pairs?

Analyte	Concentration (ug/l)		RPD (Limits)	Qualifier
	7	8		
Cr	2.7	2.5	8	

Analyte	Concentration (ug/l)		RPD (Limits)	Qualifier
	15	18		
Cr	3.6	3.4	6	

Analyte	Concentration ( )		RPD (Limits)	Qualifier

Analyte	Concentration ( )		RPD (Limits)	Qualifier

NASA JPL
Data Validation Reports
LDC# 10962

Wet Chemistry

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL

Collection Date: August 11, 2003

LDC Report Date: October 17, 2003

Matrix: Water

Parameters: Wet Chemistry

Validation Level: EPA Level III

Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-4572

Sample Identification

DUPE-5-3-Q03
EB-7-8-11-03
MW-12-1
MW-12-2
MW-12-3
MW-12-4
MW-12-5
MW-22-1
MW-22-2
MW-22-3
MW-12-4MS
MW-12-4MSD
MW-22-3MS
MW-22-3MSD

Introduction

This data review covers 14 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

J Indicates an estimated value.

R Quality control indicates the data is not usable.

N Presumptive evidence of presence of the constituent.

UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

Samples DUPE-5-3-Q03 and MW-22-2 were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	DUPE-5-3-Q03	MW-22-2	
Perchlorate	2.1	2.4	13

X. Field Blanks

Sample EB-7-8-11-03 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 03-4572

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-4572

No Sample Data Qualified in this SDG

Applied P & Ch Laboratory

Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc. Project No: 04-4428.10 Anal. Method 314.0
Project ID: JPL Service ID: 34572 Collected by:

Component Name: Perchlorate
CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-4572-1	DUPE-5-3-Q03	Water	08/11/03	08/11/03	08/12/03	03W4044	μg/L	4	2.1	B
03-4572-2	EB-7-8-11-03	Water	08/11/03	08/11/03	08/12/03	03W4044	μg/L	4	<4	U
03-4572-3	MW-12-1	Water	08/11/03	08/11/03	08/12/03	03W4044	μg/L	4	<4	U
03-4572-4	MW-12-2	Water	08/11/03	08/11/03	08/12/03	03W4044	μg/L	4	3.4	B
03-4572-5	MW-12-3	Water	08/11/03	08/11/03	08/12/03	03W4044	μg/L	4	2.8	B
03-4572-6	MW-12-4	Water	08/11/03	08/11/03	08/12/03	03W4044	μg/L	4	5.6	
03-4572-7	MW-12-5	Water	08/11/03	08/11/03	08/12/03	03W4044	μg/L	4	1.9	B
03-4572-8	MW-22-1	Water	08/11/03	08/11/03	08/12/03	03W4044	μg/L	4	2.7	B
03-4572-9	MW-22-2	Water	08/11/03	08/11/03	08/12/03	03W4044	μg/L	4	2.4	B
03-4572-10	MW-22-3	Water	08/11/03	08/11/03	08/12/03	03W4044	μg/L	4	2.2	B
03W4044-MB-01	03W4044-MB-01	Water	08/12/03	08/12/03	08/12/03	03W4044	μg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.
Qualifier: U - Not Detected or less than MDL
B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

10/18/03

Applied P & Ch Laboratory

Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc. Project No: 04-4428.10 Anal. Method 7196
Project ID: JPL Service ID: 34572 Collected by:

Component Name: Chromium (VI)
CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-4572-1	DUPE-5-3-Q03	Water	08/11/03	08/11/03	08/11/03	03W4039	mg/L	0.01	<0.01	U
03-4572-2	EB-7-8-11-03	Water	08/11/03	08/11/03	08/11/03	03W4039	mg/L	0.01	<0.01	U
03-4572-3	MW-12-1	Water	08/11/03	08/11/03	08/11/03	03W4039	mg/L	0.01	<0.01	U
03-4572-4	MW-12-2	Water	08/11/03	08/11/03	08/11/03	03W4039	mg/L	0.01	<0.01	U
03-4572-5	MW-12-3	Water	08/11/03	08/11/03	08/11/03	03W4039	mg/L	0.01	<0.01	U
03-4572-8	MW-22-1	Water	08/11/03	08/11/03	08/11/03	03W4039	mg/L	0.01	<0.01	U
03-4572-9	MW-22-2	Water	08/11/03	08/11/03	08/11/03	03W4039	mg/L	0.01	<0.01	U
03-4572-10	MW-22-3	Water	08/11/03	08/11/03	08/11/03	03W4039	mg/L	0.01	<0.01	U
03W4039-MB-01	03W4039-MB-01	Water	08/11/03	08/11/03	08/11/03	03W4039	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.
Qualifier: U - Not Detected or less than MDL
B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

10/18/03

LDC #: 10962C6

SDG #: 03-4572

Laboratory: Applied P & Ch Laboratory

VALIDATION COMPLETENESS WORKSHEET

Level III

Date: 10/14/03

Page: 1 of 1

Reviewer: MH

2nd Reviewer: [illegible]

METHOD: Hexavalent chromium (EPA SW 846 Method 7196) Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 8/11/03
IIa.	Initial calibration	A	
IIb.	Calibration verification	A	
III.	Blanks	A	
IVa.	Matrix Spike/(Matrix Spike) Duplicates	A	MS/MSD
IVb.	Laboratory control samples	A	LCS/LCSD
V.	Sample result verification	N	
VI.	Overall assessment of data	A	
VII.	Field duplicates	SW	(1, 9)
VIII.	Field blanks	ND	FB=2

Note: A = Acceptable; N = Not provided/applicable; SW = See worksheet; ND = No compounds detected; R = Rinsate; FB = Field blank; D = Duplicate; TB = Trip blank; EB = Equipment blank

Validated Samples: All

1 ✓	DUPE-5-3-Q03	11	MW-12-4MS	21		31	
2	EB-7-8-11-03	12	MW-12-4MSD	22		32	
3	MW-12-1	13	MW-22-3MS	23		33	
4 ✓	MW-12-2	14	MW-22-3MSD	24		34	
5	MW-12-3	15	MB	25		35	
6	MW-12-4	16		26		36	
7	MW-12-5	17		27		37	
8	MW-22-1	18		28		38	
9 ✓	MW-22-2	19		29		39	
10	MW-22-3	20		30		40	

Notes:

LDC #: 09b2c6
SDG #: 03-4572

VALIDATION FINDINGS WORKSHEET
Sample Specific Analysis Reference

Page: 1 of 1
Reviewer: MH
2nd reviewer: [illegible]

All circled methods are applicable to each sample.

Sample ID	Parameter
1-5, 8-10	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+} (ce04)
6, 7	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+} (ce04)
12, 13	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+} (ce04)
14, 15	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}
	pH TDS Cl F NO_3 NO_2 SO_4 PO_4 ALK CN^- NH_3 TKN TOC CR^{6+}

Comments:

LDC #: 10962C6
SDG #: 03-4572

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
Reviewer: MK
2nd reviewer: [illegible]

METHOD: Inorganics, Method See cover

(Y) N N/A Were field duplicate pairs identified in this SDG?
(Y) N N/A Were target analytes detected in the field duplicate pairs?

Analyte	Concentration (ug/L)		RPD (Limits)	Qualifier
	1	9		
U04	2.1	2.4	13	

Analyte	Concentration ( )		RPD (Limits)	Qualifier

Analyte	Concentration ( )		RPD (Limits)	Qualifier

Analyte	Concentration ( )		RPD (Limits)	Qualifier