APCL

Applied Physics & Chemistry Laboratory

13760 Magnolia Ave. Chino CA 91710
Tel. (909) 590-1828 Fax (909) 590-1498

November 25, 2003

GEOFON, Inc.
Attention: Brad Shojaee
22632 Golden Spring Dr Ste 270
Diamond Bar CA 91765

Dear Brad,

This package contains samples in our Service ID 03-5874 and your project : 04-4428.10 JPL
Enclosed please find:

(1) Original analytical report.
(2) Original Chain of Custody.
(3) One diskette containing EDD deliverable.
(4) One original Level C Data Package Deliverable.

If anything is missing or you have any questions, please feel free to contact me.

Respectfully submitted,

Regina Kirakozova
Associate QA/QC Director
Applied P & Ch Laboratory

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Submitted to:
GEOFON, Inc.
Attention: Brad Shojaee
22632 Golden Spring Dr Ste 270
Diamond Bar CA 91765
Tel: (909)396-7662 Fax: (909)396-1455

Service ID #: 801-035874
Collected by: Jr
Collected on: 10/29/03

Received: 10/29/03
Extracted: N/A
Tested: 10/29-31/03
Reported: 11/05/03

Sample Description: Water from MW-18,19.
Project Description: 04.4428.10 JPL

Analysis of Water Samples

Component Analyzed	Method	Unit	PQL	Analysis Result EB-6-10-29-03 03-05874-1	MW-18-1 03-05874-2	MW-18-2 03-05874-3	MW-18-3 03-05874-4
CHROMIUM (VI)	7196	mg/L	0.01	<0.01	<0.01	<0.01	<0.01
Dilution Factor				1	1	1	1
PERCHLORATE	314.0	μg/L	4	<4	<4	<4	<4
VOLATILE ORGANIC COMPOUNDS							
Dilution Factor				1	1	1	1
BENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOCHLOROMETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMODICHLOROMETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOFORM	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOMETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-BUTYLBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
SEC-BUTYLBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
TERT-BUTYLBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
2-BUTANONE	524.2	μg/L	10	<10	<10	<10	<10
CARBON TETRACHLORIDE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLORODIBROMOMETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	μg/L	0.5	<0.5	<0.5	<0.5	1.1
CHLOROMETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
2-CHLOROTOLUENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
4-CHLOROTOLUENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMO-3-CHLOROPROPANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMOETHANE (EDB)	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
DIBROMOMETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,4-DICHLOROBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
DICHLORODIFLUOROMETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,2-DICHLOROETHENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result EB-6-10-29-03 03-05874-1	MW-18-1 03-05874-2	MW-18-2 03-05874-3	MW-18-3 03-05874-4
TRANS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRANS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
ETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
HEXACHLOROBUTADIENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
ISOPROPYLBENZENE (CUMENE)	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
P-ISOPROPYLTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
METHYL-T-BUTYL ETHER (MTBE)	524.2	µg/L	1	<1	<1	<1	<1
4-METHYL-2-PENTANONE (MIBK)	524.2	µg/L	10	<10	<10	<10	<10
NAPHTHALENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-PROPYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
STYRENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	0.4J
TOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	0.4J
TRICHLOROFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3,5-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
VINYL CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
O-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-18-4 03-05874-5	MW-18-5 03-05874-6	MW-19-1 03-05874-7	MW-19-2 03-05874-8
CHROMIUM (VI)	7196	mg/L	0.01	<0.01	<0.01	<0.01	<0.01
Dilution Factor				1	1	1	1
PERCHLORATE	314.0	µg/L	4	17.2	<4	<4	4.4
VOLATILE ORGANIC COMPOUNDS							
Dilution Factor				1	1	1	1
BENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOCHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMODICHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	0.5
BROMOFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
SEC-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TERT-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2-BUTANONE	524.2	µg/L	10	<10	<10	<10	<10
CARBON TETRACHLORIDE	524.2	µg/L	0.5	3.4	<0.5	<0.5	<0.5
CHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLORODIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	0.4J
CHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	0.8	<0.5	<0.5	0.8
CHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
4-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMO-3-CHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMOETHANE (EDB)	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
DIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,4-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
DICHLORODIFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRANS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRANS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-18-4 03-05874-5	MW-18-5 03-05874-6	MW-19-1 03-05874-7	MW-19-2 03-05874-8
ETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
HEXACHLOROBUTADIENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
ISOPROPYLBENZENE (CUMENE)	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
P-ISOPROPYLTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
METHYL-T-BUTYL ETHER (MTBE)	524.2	µg/L	1	<1	<1	<1	<1
4-METHYL-2-PENTANONE (MIBK)	524.2	µg/L	10	<10	<10	<10	<10
NAPHTHALENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-PROPYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
STYRENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2,2-TETRACHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	µg/L	0.5	1.5	<0.5	<0.5	1.5
TOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2-TRICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	µg/L	0.5	1.0	<0.5	<0.5	0.3J
TRICHLOROFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3,5-TRIMETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
VINYL CHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
O-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5

Component Analyzed	Method	Unit	PQL	Analysis Result MW-19-3 03-05874-9	MW-19-4 03-05874-10	MW-19-5 03-05874-11	TB-6-10-29-03 03-05874-12
CHROMIUM (VI)	7196	mg/L	0.01	<0.01	<0.01	<0.01	-
Dilution Factor				1	1	1	1
PERCHLORATE	314.0	µg/L	4	5.1	<4	<4	-

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-19-3 03-05874-9	MW-19-4 03-05874-10	MW-19-5 03-05874-11	TB-6-10-29-03 03-05874-12
VOLATILE ORGANIC COMPOUNDS							
Dilution Factor				1	1	1	1
BENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOCHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMODICHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOFORM	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
BROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
SEC-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TERT-BUTYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2-BUTANONE	524.2	µg/L	10	<10	<10	<10	<10
CARBON TETRACHLORIDE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLORODIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CHLOROFORM	524.2	µg/L	0.5	0.4J	1.3	0.3J	<0.5
CHLOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
4-CHLOROTOLUENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMO-3-CHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DIBROMOETHANE (EDB)	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
DIBROMOMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,4-DICHLOROBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
DICHLORODIFLUOROMETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRANS-1,2-DICHLOROETHENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROETHENE (TOTAL)	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
2,2-DICHLOROPROPANE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
CIS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRANS-1,3-DICHLOROPROPENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
ETHYLBENZENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5
HEXACHLOROBUTADIENE	524.2	µg/L	0.5	<0.5	<0.5	<0.5	<0.5

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

APCL Analytical Report

Component Analyzed	Method	Unit	PQL	Analysis Result MW-19-3 03-05874-9	MW-19-4 03-05874-10	MW-19-5 03-05874-11	TB-6-10-29-03 03-05874-12
ISOPROPYLBENZENE (CUMENE)	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
P-ISOPROPYLTOLUENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
METHYLENE CHLORIDE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	0.4J
METHYL-T-BUTYL ETHER (MTBE)	524.2	μg/L	1	<1	<1	<1	<1
4-METHYL-2-PENTANONE (MIBK)	524.2	μg/L	10	<10	<10	<10	<10
NAPHTHALENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
N-PROPYLBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
STYRENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1,2-TETRACHLOROETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2,2-TETRACHLOROETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
TETRACHLOROETHENE	524.2	μg/L	0.5	1.4	0.7	3.9	<0.5
TOLUENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRICHLOROBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,1-TRICHLOROETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,1,2-TRICHLOROETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
TRICHLOROETHENE	524.2	μg/L	0.5	0.3J	<0.5	0.3J	<0.5
TRICHLOROFLUOROMETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,3-TRICHLOROPROPANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
112TRICHLORO-122TRIFLUOROETHANE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,2,4-TRIMETHYLBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
1,3,5-TRIMETHYLBENZENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
VINYL CHLORIDE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
O-XYLENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5
M/P-XYLENE	524.2	μg/L	0.5	<0.5	<0.5	<0.5	<0.5

PQL: Practical Quantitation Limit. MDL: Method Detection Limit. CRDL: Contract Required Detection Limit

N.D.: Not Detected or less than the practical quantitation limit. "-": Analysis is not required.

J: Reported between PQL and MDL.

Listed Dilution Factors (DF) are relative to the method default DF. All unlisted DFs are 1.0

Respectfully submitted,

Dominic Lau
Laboratory Director
Applied P & Ch Laboratory

Level C Data Package Deliverables

General Information

Project: 04-4428.10 JPL

APCL Service ID: 03-5874

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino, CA 91710
Telephone (909)590-1828
Fax (909)590-1498

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Case Narrative

Project: JPL/MW-18,19./04.4428.10

For GEOFON, Inc.

APCL Service No: 03-5874

1. Sample Identification

The sample identifications are listed in the following table:

GEOFON, Inc. Sample ID	APCL Sample ID
MW-18-5	03-05874-6
MW-18-4	03-05874-5
MW-18-3	03-05874-4
MW-18-2	03-05874-3
MW-18-1	03-05874-2
EB-6-10-29-03	03-05874-1
TB-6-10-29-03	03-05874-12
MW-19-5	03-05874-11
MW-19-4	03-05874-10
MW-19-3	03-05874-9
MW-19-2	03-05874-8
MW-19-1	03-05874-7

2. Analytical Methodology

Samples are analyzed by EPA methods

524.2 (Volatile Organic Compounds),
7196 (Chromium (VI)),
314.0 (Perchlorate, low level),

3. Holding Time

All samples were extracted, digested and analyzed within the holding times defined by the appropriate EPA methods of the analyses.

4. Preservation

All samples were preserved and stored according to the appropriate EPA methods.

5. Tele-log

None

6. Anomaly

None

"I certify that these data are technically accurate, complete, and in compliance with the terms and conditions of the contract, for other than the conditions detailed above. Release of the data contained in the hardcopy data package and its electronic data deliverable submitted on diskette had been authorized by the Laboratory Manager or her/his designee, as verified by the following signature."

Respectfully submitted,

Regina Kirakozova
Associate QA/QC Director
Applied P & Ch Laboratory

0064

GEOFON INCORPORATED

22632 GOLDEN SPRINGS DR., SUITE 270
DIAMOND BAR, CA 91765 • (909) 396-7662 • FAX (909) 396-1455

CHAIN-OF-CUSTODY RECORD

LABORATORY COPY

MW-18

GEOFON'S LAB COORDINATOR: Scott Buhmer	LAB COORDINATOR'S PHONE: 909 396 7662	LAB COORDINATOR'S FAX: 909 396 1455	LABORATORY SERVICE ID:	LABORATORY CONTACT: Kenny Chan	MAIL REPORT (COMPANY NAME): Geofon
PROJECT NAME: SPL 660 Mon-4003	PROJECT LOCATION: MW-18	PROJECT NUMBER: 04.4428.10	LABORATORY PHONE: 909 590 1828	LABORATORY FAX: 909 590 1498	RECIPIENT NAME: Tony Ford
PROJECT CONTACT: Brad Shipee	PROJECT PHONE NUMBER: 909 396 7662	PROJECT FAX: 909 396 1455	LABORATORY ADDRESS: 13760 Magnolia Ave.		ADDRESS: 22632 Golden Springs Dr. Suite 27
PROJECT ADDRESS: 4800 Oak Grove Dr.	CITY, STATE AND ZIPCODE: Pasadena, CA	CLIENT: US NAVY SWDIV	CITY, STATE AND ZIPCODE: Chino, CA 91710		CITY, STATE AND ZIPCODE: Diamond Bar, CA 91765
PROJECT MANAGER: Tony Ford	PROJECT MANAGER'S PHONE: 909 396 7662	PROJECT MANAGER'S FAX: 909 396 1455			

5874

Item	Sample Identifier	Matrix	Date	Time	Preserved	# of Cont.	QC Level	T.A.T	Analyses: 524.2 (VOCs)	314.0 (Clorox)	200.8 (Metals)	7196 (Cr (VI))					Comments
1	MW-18-5	W	10.29.03	0746	HNO3 NaOH	10.29.03 25	III	Normal	X	X	X	X					
2	MW-18-4	↓	↓	0814	HCl HNO3 NaOH	5	↓	↓	X	X	X	X					
3	MW-18-3			0841	↓	↓			X	X	X	X					
4	MW-18-2			0906					X	X	X	X					MS/MSD
5	MW-18-1			0955	↓	↓			X	X	X	X					
6	~~Trip blank~~ TB-6-10-29-03			—	HCl	2			X								
7	EB-6-10-29-03	↓	↓	0945	HCl HNO3 NaOH	5	↓	↓	X	X	X	X					
8																	
9							✓	10.29.03									
10																	

SAMPLES COLLECTED BY: SR

COURIER AND AIR BILL NUMBER:

COOLER TEMPERATURE UPON RECEIPT

RELINQUISHED BY	RECEIVED BY	DATE	TIME	SAMPLE'S CONDITION UPON RECEIPT
[signature]	[signature]	10/29/03	1341	
[signature]	[signature]	10/29/03	1341	
[signature]	[signature]	10/29/03	1428	

Distribution: White - Laboratory (To be returned with Analytical Report); Goldenrod - Project File; Yellow - Project Data Manager

0065

GEOFON INCORPORATED
22632 GOLDEN SPRINGS DR., SUITE 270
DIAMOND BAR, CA 91765 • (909) 396-7662 • FAX (909) 396-1455

CHAIN-OF-CUSTODY RECORD

LABORATORY COPY

MW-19

GEOFON's LAB COORDINATOR: Scott Buhmer	LAB COORDINATOR'S PHONE: 909 396 7662	LAB COORDINATOR'S FAX: 909 396 1455	LABORATORY SERVICE ID:	LABORATORY CONTACT: Kenny Chan	MAIL REPORT (COMPANY NAME): Geofon
PROJECT NAME: JPL GW Mon-4Q03	PROJECT LOCATION: MW-19	PROJECT NUMBER: 04-4428.10	LABORATORY PHONE: 909 590 1828	LABORATORY FAX: 909 590 1498	RECIPIENT NAME: Tony Ford
PROJECT CONTACT: Brad Shojaee	PROJECT PHONE NUMBER: 909 396 7662	PROJECT FAX: 909 396 1455	LABORATORY ADDRESS: 13710 Magnolia Ave.		ADDRESS: 22632 Golden Springs Dr. Ste 270
PROJECT ADDRESS: 4800 Oak Grove Dr.	CITY, STATE AND ZIPCODE: Pasadena, CA	CLIENT: US Navy SWDIV	CITY, STATE AND ZIPCODE: Chino, CA 91710		CITY, STATE AND ZIPCODE: Diamond Bar, CA 9176
PROJECT MANAGER: Tony Ford	PROJECT MANAGER'S PHONE: 909 396 7662	PROJECT MANAGER'S FAX: 909 396 1455			

5874

Item	Sample Identifier	Matrix	Date	Time	Preserved	# of Cont	QC Level	T.A.T	Analyses: 200.8 (W) 524.2 (VOCs)	7196 (Cr(VI))	524.2 (VOCs)	314.0 (ClO4)					Comments
1	MW-19-5	W	10/29/03	1117	HNO3 HCl NONE	25	III	Normal	X	X	X	X					
2	MW-19-4			1140					X	X	X	X					
3	MW-19-3			1210					X	X	X	X					MS/MSD
4	MW-19-2			1250					X	X	X	X					
5	MW-19-1			1313					X	X	X	X					
6																	
7																	
8																	
9																	
10																	

10.29.03

SAMPLES COLLECTED BY: JL

COURIER AND AIR BILL NUMBER:

COOLER TEMPERATURE UPON RECEIPT

RELINQUISHED BY	RECEIVED BY	DATE	TIME	SAMPLE'S CONDITION UPON RECEIPT
[signature]	[signature]	10/29/03	1341	
[signature]	[signature]	10/29/03	1341	
[signature]	[signature]	10/29/03	1428	

Distribution: White - Laboratory (To be returned with Analytical Report); Goldenrod - Project File; Yellow - Project Data Manager

Applied P & Ch Laboratory
13760 Magnolia Ave., Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Sample Receiving Checklist

APCL ServiceID: 5874 **Client Name/Project:** Geofon / JPL

1. Sample Arrival

Date/Time Received 10/29/03 1428 Date/Time Opened 10/29/03 1428 By (name): [illegible]

Custody Transfer: ☐ Client ☐ Golden State ☐ UPS ☐ US Mail ☐ FedEx ☑ APCL Empl: Richard Simon

2. Chain-of-Custody (CoC)

☑ With Samples? ☐ Faxed? ☑ Client has Copy? ☐ Signed, dated? By:
☑ Project ID? ☑ Analyses Clear? ☐ Hold Samples? #on Hold ____ # Received 12
☐ CoC/Docs Zip-Locked under lid? ☐ Compos.#: ☑ #Samples OK?
☐ Discrepancies? ☐ Client notified? ☐ Response (attach docs):

3. Shipping Container/Cooler

☑ Cooler Used? # of 1 Cooled by: ☑ Ice ☐ Blue Ice ☐ Dry Ice ☐ None

Temp °C 4.1

(Cooler temperature measured from temp blank if present, otherwise measured from the cooler).

Cooler Custody Seal? ☐ Absent ☐ Intact ☐ Tampered?

4. Sample Preservation

☐ **pH <2** ☐ pH >12

If Not, pH = ____ Preserved by: ☐ Client ☐ APCL ☐ Third Party

5. Holding-time Requirements

☐ pH 24hr ☐ BACT 6/24hr ☑ Cr^{VI} 24hr ☐ NO_3^- 48hr ☐ BOD 48hr
☐ Cl_2 ASAP ☐ Turbidity 48hr ☐ DO ASAP ☐ Fe(II) ASAP
☐ HT Expired? ☐ Client notified?

6. Sample Container Condition

☑ Intact? ☐ Broken? ☐ Documented? Number:
Type: ☑ plastic ☑ glass ☐ Tube: brass/SS ☐ Tedlar Bag
☑ Quantity OK? ☐ Leaking? ☐ Anomaly?
☑ Caps tight? ☐ Air Bubbles? ☐ Anomaly?
Labels: ☑ Unique ID? ☑ Date/Time ☐ Preserved?

7. Turn Around Time

☐ RUSH TAT: ____ ☑ Std (7-10 days) ☐ Not Marked

8. Sample Matrix

☐ Drinking H_2O ☐ Other Liq ☐ Soil ☐ Wipe ☐ Polymer ☐ Air ☐ Other:
☑ Ground H_2O ☐ Sludge ☐ Filter ☐ Oil/Petro ☐ Paint ☐ W. Water ☐ Extract ☐ Unknown

9. Pre-Login Check List Completed & OK?

☑ ALL OK? (if not, attach docs) ☐ Client Contact? (Name:____)Date/Time:____

Received/Checked by: [illegible] Printed: 29 Oct 2003 7:22 a.m.

*HT: Samples must be analyzed for results to reflect total concentrations. Results generated outside required of holding times are considered minimal values and may be used to define waste as hazardous but not as non-hazardous.

DocumentFile: [neal.texfiles]smpicl.tex.

Applied P & Ch Laboratory
13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Sample Login: Check List

03-05874 (0470_ 181) (2202777_ 181)

10/29/03

Part 1: General Information

☐ Company Information	Name:	*GEOFON, Inc.*
	Address:	*22632 Golden Spring Dr Ste 270 ,Diamond Bar ,CA 91765*
☐ Project Information	Project Description:	*JPL*
	Project #:	*04.4428.10*
☐ Billing Information	P.O. #:	
	Bill Address:	*22632 Golden Spring Dr Ste 270 ,Diamond Bar ,CA 91765*
	Lab Project ID:	
	Client Database #:	*3*
☐ Receiving Information	Who Received Sample?	*Kenny Chan*
	Receiving Date/Time:	*10/29/03 1428*
	COC No.	*0064 0065*
☐ Shipping Information	Shipping Company	*APCL pick up*
	Packing Information:	*Cooler/Ice Chester*
	Cooler Temperature:	*4.7 °C*
☐ Container Information	Container Provider:	*Client*
☐ Sampling Information	Sampling Person:	*Jr*
	Sampling Company:	*Client*
☐ Turn-Around-Time Option:		*Normal*
☐ QC Option:		*NEESA C*
☐ Disposal Option:		*Not specify*

Part 2: Sample Information

Seq. #	Sample ID (on COC)	Sample Sub-ID	APCL Sample ID	Matrix	Cont-tainer	Preser-vative	Vol, ml Am. g	# of Replica	Condition G, L, B	Collected mmddyy	Hold ?	Composite Group	TAT Days	
1	MW-18-5	524.2	03-05874-6-α	W	V	C	40	3	G	102903	N	0	9	☐
	MW-18-5	CrVI	03-05874-6-β	W	P		500	1	G	102903	N	0	9	☐
2	MW-18-4	524.2	03-05874-5-α	W	V	C	40	3	G	102903	N	0	9	☐
	MW-18-4	CrVI	03-05874-5-β	W	P		500	1	G	102903	N	0	9	☐
3	MW-18-3	524.2	03-05874-4-α	W	V	C	40	3	G	102903	N	0	9	☐
	MW-18-3	CrVI	03-05874-4-β	W	P		500	1	G	102903	N	0	9	☐
4	MW-18-2	524.2	03-05874-3-α	W	V	C	40	6	G	102903	N	0	9	☐
	MW-18-2	CrVI	03-05874-3-β	W	P		500	2	G	102903	N	0	9	☐
5	MW-18-1	524.2	03-05874-2-α	W	V	C	40	3	G	102903	N	0	9	☐
	MW-18-1	CrVI	03-05874-2-β	W	P		500	1	G	102903	N	0	9	☐
6	EB-6-10-29-03	524.2	03-05874-1-α	W	V	C	40	3	G	102903	N	0	9	☐
	EB-6-10-29-03	CrVI	03-05874-1-β	W	P		500	1	G	102903	N	0	9	☐
7	TB-6-10-29-03	524.2	03-05874-12	W	V	C	40	2	G	102903	N	0	9	☐
8	MW-19-5	524.2	03-05874-11-α	W	V	C	40	3	G	102903	N	0	9	☐
	MW-19-5	CrVI	03-05874-11-β	W	P		500	1	G	102903	N	0	9	☐
9	MW-19-4	524.2	03-05874-10-α	W	V	C	40	3	G	102903	N	0	9	☐
	MW-19-4	CrVI	03-05874-10-β	W	P		500	1	G	102903	N	0	9	☐
10	MW-19-3	524.2	03-05874-9-α	W	V	C	40	6	G	102903	N	0	9	☐
	MW-19-3	CrVI	03-05874-9-β	W	P		500	2	G	102903	N	0	9	☐
11	MW-19-2	524.2	03-05874-8-α	W	V	C	40	3	G	102903	N	0	9	☐
	MW-19-2	CrVI	03-05874-8-β	W	P		500	1	G	102903	N	0	9	☐
12	MW-19-1	524.2	03-05874-7-α	W	V	C	40	3	G	102903	N	0	9	☐
	MW-19-1	CrVI	03-05874-7-β	W	P		500	1	G	102903	N	0	9	☐

Part 3: Analysis Information

Test Items:

- ☐ *524.2* Volatile Organic Compounds
- ☐ *7196A* Chromium (VI)
- ☐ *314.0/300.0* Perchlorate, low level
- ☐ *300.0* Chloride Cl^- by IC
- ☐ *300.0* Sulfate (SO_4^{--}), by IC
- ☐ *300.0/SM4500NO3C* Nitrate (NO_3^-) as N by IC
- ☐ *SM2320B* Carbonate
- ☐ *SM2320B* Bicarbonate
- ☐ *9040B/150.1* pH
- ☐ *160.1* Solids, Total Dissolved (TDS)
- ☐ *200.7/6010B* Sodium, Na, by ICP
- ☐ *200.7/6010B* Calcium, Ca, by ICP

Level C Data Package Deliverables

Volatile Organics

Applied P & Ch Laboratory

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/31/2003
Project ID:	JPL	Service ID:	35874	Collected by:	
		Lab Sample ID:	03G4658-MB-01	Received Date:	10/31/2003
Sample ID:	**03G4658-MB-01**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Method Blank	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	15:43
Data File Name:	G4658K01	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	2	J
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
64	1,2-DICHLOROETHENE (TOTAL)	540-59-0	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	4-BROMO-FLUOROBENZENE (BFB)	460-00-4	70-129	97
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	93
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	95
4	TOLUENE-D8	2037-26-5	73-129	103
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	123
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	129
3	FLUOROBENZENE	462-06-6	50-200	136
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-1	Received Date:	10/29/2003
Sample ID:	EB-6-10-29-03	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	17:28
Data File Name:	5874-01	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	μg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	98
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	98
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	98
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	123
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	130
3	FLUOROBENZENE	462-06-6	50-200	133
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-2	Received Date:	10/29/2003
Sample ID:	**MW-18-1**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	17:54
Data File Name:	5874-02	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	98
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	95
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	96
4	TOLUENE-D8	2037-26-5	73-129	104
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	121
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	128
3	FLUOROBENZENE	462-06-6	50-200	132
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-3	Received Date:	10/29/2003
Sample ID:	MW-18-2	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	16:36
Data File Name:	5874-03	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	97
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	93
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	95
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	122
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	128
3	FLUOROBENZENE	462-06-6	50-200	133
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-4	Received Date:	10/29/2003
Sample ID:	**MW-18-3**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	18:20
Data File Name:	5874-04	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	1.1	
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	0.4	J
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	0.4	J
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	99
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	96
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	97
4	TOLUENE-D8	2037-26-5	73-129	104
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	120
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	128
3	FLUOROBENZENE	462-06-6	50-200	132
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-5	Received Date:	10/29/2003
Sample ID:	MW-18-4	Sample Matrix	Water	Moisture %:	-
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	-	Anal. Time:	18:46
Data File Name:	5874-05	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	-				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	3.4	
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	0.8	
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	1.5	
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	1.0	
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	98
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	94
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	97
4	TOLUENE-D8	2037-26-5	73-129	104
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	118
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	125
3	FLUOROBENZENE	462-06-6	50-200	129
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-6	Received Date:	10/29/2003
Sample ID:	**MW-18-5**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	19:12
Data File Name:	5874-06	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	μg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	98
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	97
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	97
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	119
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	128
3	FLUOROBENZENE	462-06-6	50-200	131
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-7	Received Date:	10/29/2003
Sample ID:	**MW-19-1**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	19:38
Data File Name:	5874-07	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	98
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	97
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	97
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	119
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	126
3	FLUOROBENZENE	462-06-6	50-200	129
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-8	Received Date:	10/29/2003
Sample ID:	MW-19-2	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	20:05
Data File Name:	5874-08	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	0.5	
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	0.4	J
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	0.8	
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	1.5	
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	0.3	J
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	100
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	97
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	97
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	119
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	124
3	FLUOROBENZENE	462-06-6	50-200	130
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-9	Received Date:	10/29/2003
Sample ID:	**MW-19-3**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	17:02
Data File Name:	5874-09	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	0.4	J
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	1.4	
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	0.3	J
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
64	1,2-DICHLOROETHENE (TOTAL)	540-59-0	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	98
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	92
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	96
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	120
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	125
3	FLUOROBENZENE	462-06-6	50-200	131
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-10	Received Date:	10/29/2003
Sample ID:	MW-19-4	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	20:31
Data File Name:	5874-10	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	1.3	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	0.7	
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
64	1,2-DICHLOROETHENE (TOTAL)	540-59-0	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	99
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	97
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	98
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	119
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	126
3	FLUOROBENZENE	462-06-6	50-200	132
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-11	Received Date:	10/29/2003
Sample ID:	**MW-19-5**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	20:57
Data File Name:	5874-11	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	0.3	J
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	μg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	3.9	
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	0.3	J
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U
64	1,2-DICHLOROETHENE (TOTAL)	540-59-0	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	97
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	97
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	97
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	118
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	129
3	FLUOROBENZENE	462-06-6	50-200	131
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Collection Date:	10/29/2003
Project ID:	JPL	Service ID:	35874	Collected by:	Jr
		Lab Sample ID:	03-5874-12	Received Date:	10/29/2003
Sample ID:	**TB-6-10-29-03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: A
Anal. Method:	524.2	Prep. Date:	10/31/03	Anal. Date:	10/31/03
Batch No:	03G4658	Prep. No:	–	Anal. Time:	16:10
Data File Name:	5874-12	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	0.4	J
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
64	1,2-DICHLOROETHENE (TOTAL)	540-59-0	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	98
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	96
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	98
4	TOLUENE-D8	2037-26-5	73-129	104
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	124
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	132
3	FLUOROBENZENE	462-06-6	50-200	136
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

FORM-2A

Applied P & Ch Laboratory

Surrogate Recovery Summary for Method 524.2

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		SDG Number:	035874
Project ID:	JPL	Project No:	04.4428.10	Sample Matrix:	Water
		Batch No:	03G4658		

#	Client Sample No	Lab Sample ID	S1 % #	S2 % #	S3 % #	S4 % #	TOT OUT
1	03G4658-LCS-01	03G4658-LCS-01	96	95	95	97	0
2	MW-18-2MS	03-5874-3MS	99	85	91	99	0
3	MW-18-2MSD	03-5874-3MSD	95	91	92	94	0
4	MW-19-3MS	03-5874-9MS	93	92	93	93	0
5	MW-19-3MSD	03-5874-9MSD	95	93	93	95	0
6	03G4658-MB-01	03G4658-MB-01	97	93	95	103	0
7	TB-6-10-29-03	03-5874-12	98	96	98	104	0
8	MW-18-2	03-5874-3	97	93	95	105	0
9	MW-19-3	03-5874-9	98	92	96	105	0
10	EB-6-10-29-03	03-5874-1	98	98	98	105	0
11	MW-18-1	03-5874-2	98	95	96	104	0
12	MW-18-3	03-5874-4	99	96	97	104	0
13	MW-18-4	03-5874-5	98	94	97	104	0
14	MW-18-5	03-5874-6	98	97	97	105	0
15	MW-19-1	03-5874-7	98	97	97	105	0
16	MW-19-2	03-5874-8	100	97	97	105	0
17	MW-19-4	03-5874-10	99	97	98	105	0
18	MW-19-5	03-5874-11	97	97	97	105	0
19							
20							
21							
22							
23							
24							
25							

	QC Control Limit
S1 = 1-BROMO-4-FLUOROBENZENE (4-BROMOFL	70-129
S2 = 1,2-DICHLOROETHANE-D4	70-129
S3 = DIBROMOFLUOROMETHANE	70-122
S4 = TOLUENE-D8	73-129

Column to be used to flag recovery values:

* – Values outside of contract required QC Limits D – Surrogate diluted out I – Matrix Interference

FORM-3A

Applied P & Ch Laboratory

Lab Control Spike/Lab Control Spike Duplicate Recovery for Method 524.2

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		Service ID:	35874
Project ID:	JPL	Project No:	04.4428.10	Sample Matrix:	Water
		Batch No:	03G4658		
LCS Filename:	G4658L01	Date Analyzed:	103103	Time Analyzed:	11:34
LCSD Filename:	–	Date Analyzed:	–	Time Analyzed:	–

Spiked Components	Unit	Spike Added	Concentration Unspiked	LCS	LCS Rec% #	QC Limit, % REC
BENZENE	μg/L	20	0	18.2	91	65-120
CHLOROBENZENE	μg/L	20	0	18.9	95	65-134
1,1-DICHLOROETHENE	μg/L	20	0	20.3	102	65-127
TOLUENE	μg/L	20	0	18.5	93	65-134
TRICHLOROETHENE	μg/L	20	0	19.5	98	67-122
# of Out-of-control					0	

Column to be used to flag recovery and RPD values:

* – Values outside of contract required QC Limits D – Spiked components diluted out

Comments: _____

FORM-3A

Applied P & Ch Laboratory

Matrix Spike/Matrix Spike Duplicate Recovery for Method 524.2

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		Service ID:	35874
Project ID:	JPL	Project No:	04.4428.10	Sample Matrix:	Water
		Batch No:	03G4658		
MS Filename:	G4658M01	Date Analyzed:	103103	Time Analyzed:	12:16
MSD Filename:	G4658N01	Date Analyzed:	103103	Time Analyzed:	12:42
MS Sample No:	**MW-18-2**	Sample Lab ID:	03-5874-3		

Spiked Components	Unit	Spike Added	Concentration Unspiked	Concentration MS	MS Rec% #	QC Limit, % REC
BENZENE	μg/L	20	0	18.4	92	65-121
CHLOROBENZENE	μg/L	20	0	19.0	95	65-134
1,1-DICHLOROETHENE	μg/L	20	0	20.9	105	65-127
TOLUENE	μg/L	20	0	18.8	94	65-134
TRICHLOROETHENE	μg/L	20	0	19.6	98	65-125
# of Out-of-control					0	

Spiked Components	Unit	Spike Added	MSD Concentration	MSD Rec% #	RPD% #	QC Limit, % RPD	QC Limit, % REC
BENZENE	μg/L	20	17.9	90	2	28	65-121
CHLOROBENZENE	μg/L	20	18.4	92	3	35	65-134
1,1-DICHLOROETHENE	μg/L	20	20.2	101	4	31	65-127
TOLUENE	μg/L	20	17.6	88	7	35	65-134
TRICHLOROETHENE	μg/L	20	19.2	96	2	30	65-125
# of Out-of-control				0	0		

Column to be used to flag recovery and RPD values:

* – Values outside of contract required QC Limits D – Spiked components diluted out

Comments: _____

FORM-3A

Applied P & Ch Laboratory

Matrix Spike/Matrix Spike Duplicate Recovery for Method 524.2

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		Service ID:	35874
Project ID:	JPL	Project No:	04.4428.10	Sample Matrix:	Water
		Batch No:	03G4658		
MS Filename:	G4658M02	Date Analyzed:	103103	Time Analyzed:	13:07
MSD Filename:	G4658N02	Date Analyzed:	103103	Time Analyzed:	13:33
MS Sample No:	MW-19-3	Sample Lab ID:	03-5874-9		

Spiked Components	Unit	Spike Added	Concentration Unspiked	Concentration MS	MS Rec% #	QC Limit, % REC
BENZENE	μg/L	20	0	17.8	89	65-121
CHLOROBENZENE	μg/L	20	0	18.3	92	65-134
1,1-DICHLOROETHENE	μg/L	20	0	20.1	101	65-127
TOLUENE	μg/L	20	0	17.5	88	65-134
TRICHLOROETHENE	μg/L	20	0.3	19.8	98	65-125
# of Out-of-control					0	

Spiked Components	Unit	Spike Added	MSD Concentration	MSD Rec% #	RPD% #	QC Limit, % RPD	QC Limit, % REC
BENZENE	μg/L	20	18.1	91	2	28	65-121
CHLOROBENZENE	μg/L	20	18.6	93	1	35	65-134
1,1-DICHLOROETHENE	μg/L	20	20.1	101	0	31	65-127
TOLUENE	μg/L	20	17.7	89	1	35	65-134
TRICHLOROETHENE	μg/L	20	19.9	98	0	30	65-125
# of Out-of-control				0	0		

Column to be used to flag recovery and RPD values:

* – Values outside of contract required QC Limits D – Spiked components diluted out

Comments: _____

FORM-4A

Applied P & Ch Laboratory

Method Blank Summary for Method 524.2

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		Service ID:	35874
Project ID:	JPL	Project No:	04.4428.10	Analysis Date:	10/31/03
		Sample Matrix:	Water	Analysis Time:	15:43
Sample ID:	**03G4658-MB-01**	Batch No:	03G4658	Instrument ID:	GC/MS: A
Lab Sample ID:	03G4658-MB-01	Data File Name:	G4658K01	GC Column:	HP-VOC
		Heated Purge: (Y/N)	N	Column ID:	0.20 mm

This Method Blank applies to the following samples and QC samples:

#	Client Sample No	Lab Sample ID	Sample Type	Data Filename	Analysis Date	Analysis Time
1	03G4658-LCS-01	03G4658-LCS-01	Lab Control Spike	G4658L01	10/31/03	11:34
2	MW-18-2MS	03-5874-3MS	Matrix Spike	G4658M01	10/31/03	12:16
3	MW-18-2MSD	03-5874-3MSD	Matrix Spike Duplicate	G4658N01	10/31/03	12:42
4	MW-19-3MS	03-5874-9MS	Matrix Spike	G4658M02	10/31/03	13:07
5	MW-19-3MSD	03-5874-9MSD	Matrix Spike Duplicate	G4658N02	10/31/03	13:33
6	TB-6-10-29-03	03-5874-12	Field Sample	5874-12	10/31/03	16:10
7	MW-18-2	03-5874-3	Field Sample	5874-03	10/31/03	16:36
8	MW-19-3	03-5874-9	Field Sample	5874-09	10/31/03	17:02
9	EB-6-10-29-03	03-5874-1	Field Sample	5874-01	10/31/03	17:28
10	MW-18-1	03-5874-2	Field Sample	5874-02	10/31/03	17:54
11	MW-18-3	03-5874-4	Field Sample	5874-04	10/31/03	18:20
12	MW-18-4	03-5874-5	Field Sample	5874-05	10/31/03	18:46
13	MW-18-5	03-5874-6	Field Sample	5874-06	10/31/03	19:12
14	MW-19-1	03-5874-7	Field Sample	5874-07	10/31/03	19:38
15	MW-19-2	03-5874-8	Field Sample	5874-08	10/31/03	20:05
16	MW-19-4	03-5874-10	Field Sample	5874-10	10/31/03	20:31
17	MW-19-5	03-5874-11	Field Sample	5874-11	10/31/03	20:57
18						
19						
20						
21						
22						
23						
24						
25						

Data File : C:\MSDCHEM\1\DATA\03G4612\G4612P01.D
Acq On : 21 Oct 2003 9:14 am
Sample : ##03g4565,w 50ng
Misc :
MS Integration Params: Lscint.p
Method : C:\MSDCHEM\1\METHODS\E524A003.M (RTE Integrator)
Title : **Applied P &Ch Lab** EPA 524.2

Vial: 1
Operator: zou
Inst : GCMS-A
Multiplr: 1.00



Spectrum Information: Average of 12.736 to 12.748 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.0	11745	PASS
75	95	30	60	45.0	31162	PASS
95	95	100	100	100.0	69181	PASS
96	95	5	9	6.6	4558	PASS
173	174	0.00	2	0.5	275	PASS
174	95	50	100	87.4	60456	PASS
175	174	5	9	7.7	4634	PASS
176	174	95	101	98.1	59312	PASS
177	176	5	9	7.0	4135	PASS

5A
VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name : APPLIED P & CH LAB Contract: _____

Lab Code: _____ Case No.: _____ SAS No.: _____ SDG No.: 035874

Lab File ID: G 4612 P01 BFB Injection Date: 10/21/2003

Instrument ID: GCMS-A BFB Injection Time: 0914

GC Column: HP-VOC ID: 0.20 (mm) Heated Purge: (Y/N) N

m/e	ION ABUNDANCE CRITERIA	%RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	17.0
75	30.0 - 60.0% of mass 95	45.0
95	Base peak, 100% relative abundance	100.0
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.4 (0.5)1
174	50.0 - 100.0% of mass 95	87.4
175	5.0 - 9.0% of mass 174	6.7 (7.7)1
176	95.0 - 101.0% of mass 174	85.7 (98.1)1
177	5.0 - 9.0% of mass 176	6.0 (7.0)2

1-Value is % mass 174 2-Value is % mass 176

This check applies to the following SAMPLES, MS, MSD, BLANKS and STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
01	VSTD0003	3-A0003	3-A0003.D	10/21/2003	1005
02	VSTD002	3-0002	3-0002.D	10/21/2003	1031
03	VSTD010	3-0010	3-0010.D	10/21/2003	1056
04	VSTD020	3-0020	3-0020.D	10/21/2003	1122
05	VSTD040	3-0040	3-0040.D	10/21/2003	1147
06	VSTD060	3-0060	3-0060.D	10/21/2003	1214
07					
08					
09					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					

INITIAL CALIBRATION SUMMARY

Method File	e524a003		
Last Calibration Update	Mon Oct 27 13:56:31 2003		
Level 1 File Name	3-A0003.D	**Level 1 ID**	.3
Level 2 File Name	3-002.D	**Level 2 ID**	2
Level 3 File Name	3-0010.D	**Level 3 ID**	10
Level 4 File Name	3-0020.D	**Level 4 ID**	20
Level 5 File Name	3-0040.D	**Level 5 ID**	40
Level 6 File Name	3-0060.D	**Level 6 ID**	60
Level 7 File Name	3-0020.D	**Level 7 ID**	cc

Compound Name	Level 1 Response	Level 2 Response	Level 3 Response	Level 4 Response	Level 5 Response	Level 6 Response	Level 7 Response	Coeff X^0	Coeff X^1 / ave RF	Coeff X^2	R^2 / RSD
1 Fluorobenzene	920067	916201	919392	955116	907151	1006211	-1	------	------	------	------
3 di-Cl-di-F-methane	6437	39097	199979	303885	731076	1099511	-1	0.0000	0.2011	0.0000	0.1330
4 Chloromethane	6136	37417	181676	348991	739859	1132557	-1	0.0000	0.1997	0.0000	0.0704
2 F114	1527	19306	101058	155932	367886	551592	-1	0.0028	0.0932	0.0000	0.9919
5 vinyl chloride	4876	33446	161213	285451	640709	982033	-1	0.0000	0.1705	0.0000	0.0717
6 bromomethane	2539	18408	87613	168214	383891	601141	-1	0.0000	0.0969	0.0000	0.0659
7 chloroethane	2197	17752	90280	159905	376464	536590	-1	0.0000	0.0918	0.0000	0.1013
8 tri-Cl-F-methane	8394	56082	277768	461283	1030751	1532166	-1	0.0000	0.2819	0.0000	0.0992
91 Acetonitrile X10	10689	82041	382813	794170	1713507	2428244	-1	0.0000	0.0431	0.0000	0.0662
9 acrolein X10	5645	40130	201945	381523	807796	1232252	-1	0.0000	0.0212	0.0000	0.0467
11 acetone X10	-1	64324	288996	538495	1267205	1816586	-1	0.0000	0.0319	0.0000	0.0947
12 ethyl ether X5	13693	93915	436367	874255	1704647	2654028	-1	0.0000	0.0950	0.0000	0.0552
13 11-dichloroethene	6637	46215	229524	410040	877445	1367991	-1	0.0000	0.2376	0.0000	0.0605
14 Iodomethane	3461	29384	183879	356970	751141	1123042	-1	0.0038	0.1913	0.0000	0.9951
15 F-113	4492	28088	146108	224035	516639	784890	-1	0.0000	0.1441	0.0000	0.1231
16 acrylonitrile X10	10574	78358	419836	794346	1709905	2532493	-1	0.0000	0.0429	0.0000	0.0731
17 carbon disulfide	13839	85463	420250	772812	1671671	2561651	-1	0.0000	0.4524	0.0000	0.0750
94 Isopropyl Alcoholx10	331	1284	1238	28714	240070	128634	-1	0.0000	0.0020	0.0000	1.1427
18 methylene chloride	10278	48191	203033	407143	804880	1192483	-1	0.0203	0.2012	0.0000	0.9947
19 t-12-di-Cl-ethene	6615	45733	217854	407874	774111	1126298	-1	0.0000	0.2233	0.0000	0.1039
20 t-Bu-Me-ether	9529	70095	362056	751228	1592599	2437498	-1	0.0000	0.3929	0.0000	0.0773
95 Tert butyl alcoholx10	1332	4624	88338	76985	452680	251872	-1	0.0000	0.0066	0.0000	0.6496

94 allyl chloride	10689	82041	426268	794170	1724483	2428244	-1	0.0000	0.4409	0.0000	0.0706
21 11-dichloroethane	10035	71189	342138	689918	1398310	2101878	-1	0.0000	0.3698	0.0000	0.0415
97 propionitrile	103	3119	18745	29270	62391	109975	-1	0.0000	0.0176	0.0000	0.1055
22 c-12-di-Cl-ethene	6459	46844	225560	436935	845644	1210430	-1	0.0000	0.2329	0.0000	0.0801
23 22-Dichloropropane	5693	38859	218612	456844	979998	1513030	-1	0.0000	0.2360	0.0000	0.1011
24 Br-Cl-methane	3294	23046	109590	215148	438926	655419	-1	0.0000	0.1177	0.0000	0.0523
25 chloroform	13164	80631	375460	741683	1492648	2229387	-1	0.0000	0.4157	0.0000	0.0920
26 tetrahydrofuranX5	4044	28808	169729	317711	726346	1114014	-1	0.0000	0.0346	0.0000	0.1155
98 Diisopropyl ether	16931	136442	688556	1387177	2723439	4088155	-1	0.0000	0.7101	0.0000	0.0772
27 Di-Br-F-Me (surr)	6137	45047	218874	439096	880145	1314263	-1	0.0000	0.2348	0.0000	0.0481
99 ETBE	9591	74005	411030	879689	1934215	2984281	-1	0.0000	0.4477	0.0000	0.1468
29 1,2-Di-Cl-Et-d4 (S1)	5167	39287	192209	385356	801084	1206050	-1	0.0000	0.2091	0.0000	0.0418
30 12-dichloroethane	2318	14864	77221	146461	303955	463659	-1	0.0000	0.0811	0.0000	0.0434
32 vinyl acetate X5	36279	291623	1683396	3463521	7301957	11062661	-1	0.0000	0.3465	0.0000	0.1412
92 Nitro Methane(x10)	1574	10571	358371	95786	134103	298076	-1	0.0000	0.0117	0.0000	1.3083
33 2-butanoneMEK X10	14739	98541	459782	956754	2014463	3034414	-1	0.0000	0.0522	0.0000	0.0453
93 Ethyl Acetate x2	6463	37342	237607	465630	1199719	1687653	-1	0.0000	0.1292	0.0000	0.1682
34 111-trichloroethane	9393	63933	333945	629233	1352461	2070811	-1	0.0000	0.3496	0.0000	0.0454
35 11-Di-Cl-propene	6749	49362	260487	484228	1020849	1516179	-1	0.0000	0.2639	0.0000	0.0625
36 benzene	23536	167998	811668	1581767	3067808	4519766	-1	0.0000	0.8457	0.0000	0.0673
37 CCl4	9052	59763	312695	555550	1228562	1842053	-1	0.0000	0.3215	0.0000	0.0608

Compound Name	Level 1 Response	Level 2 Response	Level 3 Response	Level 4 Response	Level 5 Response	Level 6 Response	Level 7 Response	Coeff X^0	Coeff X^1 / ave RF	Coeff X^2	R^2 / RSD
100 Isobutyl alcoholx10	2932	21047	112031	247109	552842	880424	-1	------	------	------	------
38 thiophene	10222	83170	428533	868274	1718044	2554816	-1	0.0000	0.4403	0.0000	0.0870
39 12-di-Cl-propane	4815	37555	178169	370946	752605	1124998	-1	0.0000	0.1935	0.0000	0.0628
40 trichloroethene	7696	50098	255616	501517	1020048	1520360	-1	0.0000	0.2710	0.0000	0.0423
41 dibromomethane	3770	26033	124373	247759	508912	747601	-1	0.0000	0.1346	0.0000	0.0506
101 TAME	7879	59032	335735	716981	1622049	2525816	-1	-0.0345	0.4304	0.0000	0.9963
42 Br-di-Cl-methane	9392	56733	274489	546104	1129675	1703118	-1	0.0000	0.3046	0.0000	0.0694
43 Me-methacrylate	1457	14747	87065	188318	435160	662734	-1	-0.0095	0.1138	0.0000	0.9946
44 2-ClEt-Vi-ether10	4887	45249	275793	591932	1294105	1950851	-1	-0.0367	0.0357	0.0000	0.9955
45 c-13-di-Cl-propene	6769	54923	293425	608486	1252896	1891415	-1	0.0000	0.3069	0.0000	0.1096
46 t-1,3-dichloropropene	5153	40425	231011	505285	1067475	1638452	-1	-0.0097	0.2792	0.0000	0.9968
47 Chlorobezene-d5	745928	732691	724575	731811	659336	705610	-1	0.0000	1.0000	0.0000	0.0000

48 112-tri-Cl-Et	4896	31570	158039	308164	644393	936372	-1	0.0000	0.2214	0.0000	0.0533
49 13-di-Cl-propane	7569	48826	251096	479955	958048	1392110	-1	0.0000	0.3397	0.0000	0.0396
50 Et methacrylate	2785	29334	189197	404574	897662	1374874	-1	-0.0430	0.3345	0.0000	0.9960

Compound Name	Level 1 Response	Level 2 Response	Level 3 Response	Level 4 Response	Level 5 Response	Level 6 Response	Level 7 Response	Coeff X^0	Coeff X^1 / ave RF	Coeff X^2	R^2 / RSD
51 di-Br-Cl-methane	7359	44728	219307	430191	895491	1332246	-1	0.0000	0.3141	0.0000	0.0548
52 bromoform	4029	26611	127022	257402	537161	813031	-1	0.0000	0.1848	0.0000	0.0599
53 1,4-dichlorobutane-2	7248	52525	270048	530472	1118631	1708578	-1	0.0000	0.3742	0.0000	0.0947
54 MIBK	2727	20267	114402	244409	567049	883730	-1	-0.0334	0.2144	0.0000	0.9947
55 toluene-d8	22176	162055	811859	1597728	3175057	4699033	-1	0.0000	1.1038	0.0000	0.0617
56 toluene	32453	204349	957684	1867212	3679239	5453434	-1	0.0000	1.3542	0.0000	0.0513
57 2-hexanone X5	8296	70148	400071	806303	1789801	2714840	-1	-0.0660	0.1318	0.0000	0.9958
58 12-dibromoethane	4393	32436	159989	322735	683020	1029927	-1	0.0000	0.2269	0.0000	0.0954
59 tetra-Cl-ethene	8603	59173	281865	498130	1001817	1436032	-1	0.0000	0.3728	0.0000	0.0719
60 chlorobenzene	21200	144953	676974	1307160	2492370	3536493	-1	0.0000	0.9241	0.0000	0.0576
61 1112-tetra-Cl-Et	7365	51367	249086	487380	969491	1437439	-1	0.0000	0.3439	0.0000	0.0403
62 1,4-Dichlorobenzene-d4	436784	441685	431820	433290	398842	428493	-1	0.0000	1.0000	0.0000	0.0000
63 1-chlorohexane	3125	20185	105292	181129	379300	546400	-1	0.0000	0.2284	0.0000	0.0636
64 Et-Bz	31613	225200	1128701	2162672	4307205	6378701	-1	0.0000	2.5420	0.0000	0.0404
65 m/p-Xylenes X2	51588	364290	1753788	3322945	6407191	9314343	-1	0.0000	1.9663	0.0000	0.0463
66 styrene	19899	158986	755749	1448349	2712603	3842415	-1	0.0000	1.6558	0.0000	0.0748
67 o-xylene	25030	181225	888564	1715323	3289680	4761198	-1	0.0000	1.9855	0.0000	0.0445
68 1122-Tetra-Cl-Et	4786	35408	176142	333638	679431	1005365	-1	0.0000	0.3960	0.0000	0.0524
69 123-tri-Cl-Pr	1463	11857	58506	111098	231129	345542	-1	0.0000	0.1315	0.0000	0.0844
70 4-Br-1-F-Bz (S3)	9236	60769	292727	571173	1114624	1638683	-1	0.0000	0.6776	0.0000	0.0376
71 isopropylbenzene	31809	228075	1204677	2257352	4576796	6821750	-1	0.0000	2.6544	0.0000	0.0592
72 bromobenzene	8279	62596	299551	577972	1123002	1610222	-1	0.0000	0.6719	0.0000	0.0539
92 t-1,4-dichloro-2-butene	370	3850	26890	58909	138278	214154	-1	-0.0139	0.0860	0.0000	0.9950
73 n-propylbenzene	9091	70513	355352	656540	1290074	1877926	-1	0.0000	0.7686	0.0000	0.0653
74 2-Cl-Toluene	7570	61543	297725	558317	1084812	1601091	-1	0.0000	0.6518	0.0000	0.0707
75 4-Cl-Toluene	9182	64367	296683	566381	1062470	1497703	-1	0.0000	0.6698	0.0000	0.0750
76 135-tri-Me-Benzene	28727	213694	1063069	2008540	3906361	5670674	-1	0.0000	2.3409	0.0000	0.0517
77 4-iso-Pr-toluene	31720	226621	1149787	2111126	4215398	6174914	-1	0.0000	2.5215	0.0000	0.0463
78 124-tri-Me-Benzene	29971	222021	1078747	2103275	4147713	6124395	-1	0.0000	2.4513	0.0000	0.0448
79 tert-butylbenzene	24279	177466	929162	1736357	3598467	5289165	-1	0.0000	2.0550	0.0000	0.0671

80 13-DCB	19447	131664	611576	1176380	2227519	3166301	-1	0.0000	1.3960	0.0000	0.0684
81 sec-butylbenzene	38627	268975	1387531	2530666	5231475	7732088	-1	0.0000	3.0688	0.0000	0.0475
82 14-DCB	21754	133604	619366	1184125	2334471	3417867	-1	0.0000	1.4610	0.0000	0.0806
83 Cl-benzyl	871	5857	44122	106539	268318	431461	-1	-0.0412	0.1731	0.0000	0.9920
84 12-DCB	17721	121367	559240	1055514	2046326	2953834	-1	0.0000	1.2785	0.0000	0.0657
85 n-butylbenzene	8164	59910	307483	567085	1127859	1596385	-1	0.0000	0.6659	0.0000	0.0599
86 12-diBr-2-Cl-Pra	1032	7477	40003	83208	177685	276212	-1	0.0000	0.0951	0.0000	0.1330
87 124-tri-Cl-Bz	9169	66792	354819	702123	1402301	2085639	-1	0.0000	0.7963	0.0000	0.0771
88 naphthalene	11079	87277	545491	1159824	2506296	3927146	-1	-0.1762	1.5623	0.0000	0.9976
89 hx-Cl-butadiene	6445	40287	205532	360400	746186	1096494	-1	0.0000	0.4557	0.0000	0.0644
90 123-Tri-Cl-Bz	7754	57941	307963	600915	1229181	1798500	-1	0.0000	0.6874	0.0000	0.0870

Response F or Report GCMS-A

Method : C:\MSDCHEM\1\METHODS\E524A003.M (RTE Integrator)
Title : **Applied P &Ch Lab** EPA 524.2
Last Update : Mon Oct 27 13:56:31 2003
Response via : Initial Calibration

Calibration Files
.3 =3-A0003.D 2 =3-002.D 10 =3-0010.D
20 =3-0020.D 40 =3-0040.D 60 =3-0060.D

		Compound	.3	2	10	20	40	60	Avg	%RSD	r2
1)	I 1	Fluorobenzene	----------ISTD----------								
2)	3	di-Cl-di-F-m	0.233	0.213	0.218	0.159	0.201	0.182	0.201	13.30	
3)	p 4	Chloromethan	0.222	0.204	0.198	0.183	0.204	0.188	0.200	7.04	
4)	2	F114	0.055	0.105	0.110	0.082	0.101	0.091	0.091	22.20	0.990
5)	C 5	vinyl chlori	0.177	0.183	0.175	0.149	0.177	0.163	0.171	7.17	
6)	6	bromomethane	0.092	0.100	0.095	0.088	0.106	0.100	0.097	6.59	
7)	7	chloroethane	0.080	0.097	0.098	0.084	0.104	0.089	0.092	10.13	
8)	8	tri-Cl-F-met	0.304	0.306	0.302	0.241	0.284	0.254	0.282	9.92	
9)	91	Acetonitrile		0.045	0.042	0.042	0.047	0.040	0.043	6.62	
10)	9	acrolein	0.020	0.022	0.022	0.020	0.022	0.020	0.021	4.67	
11)	11	acetone X		0.035	0.031	0.028	0.035	0.030	0.032	9.47	
12)	12	ethyl ether	0.099	0.103	0.095	0.092	0.094	0.088	0.095	5.52	
13)	M, C13	11-dichloroe	0.240	0.252	0.250	0.215	0.242	0.227	0.238	6.05#	
14)	14	Iodomethane	0.125	0.160	0.200	0.187	0.207	0.186	0.178	16.98	0.995
15)	15	F-113	0.163	0.153	0.159	0.117	0.142	0.130	0.144	12.31	
16)	16	acrylonitril	0.038	0.043	0.046	0.042	0.047	0.042	0.043	7.31	
17)	17	carbon disul	0.501	0.466	0.457	0.405	0.461	0.424	0.452	7.50	
18)	94	Isopropyl Al	0.001	0.001	0.000	0.002	0.007	0.002	0.002	114.27	
19)	18	methylene ch	0.372	0.263	0.221	0.213	0.222	0.198	0.248	26.04	0.995
20)	19	t-12-di-Cl-e	0.240	0.250	0.237	0.214	0.213	0.187	0.223	10.39	
21)	20	t-Bu-Me-ethe	0.345	0.383	0.394	0.393	0.439	0.404	0.393	7.73	
22)	95	Tert butyl a		0.003	0.010	0.004	0.012	0.004	0.007	64.96	
23)	94	allyl chlori		0.448	0.464	0.416	0.475	0.402	0.441	7.06	
24)	p 21	11-dichloroe	0.364	0.389	0.372	0.361	0.385	0.348	0.370	4.15	
25)	97	propionitril		0.017	0.020	0.015	0.017	0.018	0.018	10.55	
26)	22	c-12-di-Cl-e	0.234	0.256	0.245	0.229	0.233	0.200	0.233	8.01	
27)	23	22-Dichlorop	0.206	0.212	0.238	0.239	0.270	0.251	0.236	10.11	
28)	24	Br-Cl-methan	0.119	0.126	0.119	0.113	0.121	0.109	0.118	5.23	

(#) = Out of Range

Response F .or Report GCMS-A

Method : C:\MSDCHEM\1\METHODS\E524A003.M (RTE Integrator)
Title : **Applied P &Ch Lab** EPA 524.2
Last Update : Mon Oct 27 13:56:31 2003
Response via : Initial Calibration

Calibration Files
.3 =3-A0003.D 2 =3-002.D 10 =3-0010.D
20 =3-0020.D 40 =3-0040.D 60 =3-0060.D

		Compound	.3	2	10	20	40	60	Avg	%RSD	
29)	C	25 chloroform	0.477	0.440	0.408	0.388	0.411	0.369	0.416	9.20	
30)		26 tetrahydrofu	0.029	0.031	0.037	0.033	0.040	0.037	0.035	11.55	
31)		98 Diisopropyl	0.613	0.745	0.749	0.726	0.751	0.677	0.710	7.72	
32)	S	27 Di-Br-F-Me (		0.246	0.238	0.230	0.243	0.218	0.235	4.81	
33)		99 ETBE	0.347	0.404	0.447	0.461	0.533	0.494	0.448	14.68	
34)	S	29 1,2-Di-Cl-Et		0.214	0.209	0.202	0.221	0.200	0.209	4.18	
35)		30 12-dichloroe	0.084	0.081	0.084	0.077	0.084	0.077	0.081	4.34	
36)		32 vinyl acetat	0.263	0.318	0.366	0.363	0.402	0.366	0.346	14.12	
37)		92 Nitro Methan		0.006	0.039	0.005	0.004	0.005	0.012	130.83	
38)		33 2-butanoneME	0.053	0.054	0.050	0.050	0.056	0.050	0.052	4.53	
39)		93 Ethyl Acetat	0.117	0.102	0.129	0.122	0.165	0.140	0.129	16.82	
40)		34 111-trichlor	0.340	0.349	0.363	0.329	0.373	0.343	0.350	4.54	
41)		35 11-Di-Cl-pro	0.245	0.269	0.283	0.253	0.281	0.251	0.264	6.25	
42)	M	36 benzene	0.853	0.917	0.883	0.828	0.845	0.749	0.846	6.73	
43)		37 CCl4	0.328	0.326	0.340	0.291	0.339	0.305	0.321	6.08	
44)		100 Isobutyl al	0.011	0.011	0.012	0.013	0.015	0.015	0.013	13.92	
45)		38 thiophene	0.370	0.454	0.466	0.455	0.473	0.423	0.440	8.70	
46)	C	39 12-di-Cl-pro	0.174	0.205	0.194	0.194	0.207	0.186	0.194	6.28#	
47)	M	40 trichloroeth	0.279	0.273	0.278	0.263	0.281	0.252	0.271	4.23	
48)		41 dibromometha	0.137	0.142	0.135	0.130	0.140	0.124	0.135	5.06	
49)		101 TAME	0.285	0.322	0.365	0.375	0.447	0.418	0.369	16.15	0.995
50)		42 Br-di-Cl-met	0.340	0.310	0.299	0.286	0.311	0.282	0.305	6.94	
51)		43 Me-methacryl	0.053	0.080	0.095	0.099	0.120	0.110	0.093	25.59	0.994
52)		44 2-ClEt-Vi-et	0.018	0.025	0.030	0.031	0.036		0.028	24.68	0.993
53)		45 c-13-di-Cl-p	0.245	0.300	0.319	0.319	0.345	0.313	0.307	10.96	
54)		46 t-1,3-dichlo	0.187	0.221	0.251	0.265	0.294	0.271	0.248	15.58	0.997
55)	I	47 Chlorobezene-d5	----------ISTD----------								
56)		48 112-tri-Cl-E	0.219	0.215	0.218	0.211	0.244	0.221	0.221	5.33	

(#) = Out of Range

Response F :or Report GCMS-A

Method : C:\MSDCHEM\1\METHODS\E524A003.M (RTE Integrator)
Title : **Applied P &Ch Lab** EPA 524.2
Last Update : Mon Oct 27 13:56:31 2003
Response via : Initial Calibration

Calibration Files
.3 =3-A0003.D 2 =3-002.D 10 =3-0010.D
20 =3-0020.D 40 =3-0040.D 60 =3-0060.D

			Compound	.3	2	10	20	40	60	Avg	%RSD	
57)		49	13-di-Cl-pro	0.338	0.333	0.347	0.328	0.363	0.329	0.340	3.96	
58)		50	Et methacryl	0.124	0.200	0.261	0.276	0.340	0.325	0.255	31.77	0.995
59)		51	di-Br-Cl-met	0.329	0.305	0.303	0.294	0.340	0.315	0.314	5.48	
60)	P	52	bromoform	0.180	0.182	0.175	0.176	0.204	0.192	0.185	5.99	
61)		53	1,4-dichloro	0.324	0.358	0.373	0.362	0.424	0.404	0.374	9.47	
62)		54	MIBK	0.122	0.138	0.158	0.167	0.215	0.209	0.168	22.23	0.992
63)	s	55	toluene-d8	0.991	1.106	1.120	1.092	1.204	1.110	1.104	6.17	
64)	M,C	56	toluene	1.450	1.395	1.322	1.276	1.395	1.288	1.354	5.13	
65)		57	2-hexanone X	0.074	0.096	0.110	0.110	0.136	0.128	0.109	20.41	0.994
66)		58	12-dibromoet	0.196	0.221	0.221	0.221	0.259	0.243	0.227	9.54	
67)		59	tetra-Cl-eth	0.384	0.404	0.389	0.340	0.380	0.339	0.373	7.19	
68)	M,P	60	chlorobenzen	0.947	0.989	0.934	0.893	0.945	0.835	0.924	5.76	
69)		61	1112-tetra-C	0.329	0.351	0.344	0.333	0.368	0.340	0.344	4.03	
70)	I	62	1,4-Dichlorobenzen ----------ISTD----------									
71)		63	1-chlorohexa	0.238	0.228	0.244	0.209	0.238	0.213	0.228	6.36	
72)	C	64	Et-Bz	2.413	2.549	2.614	2.496	2.700	2.481	2.542	4.04#	
73)		65	m/p-Xylenes	1.968	2.062	2.031	1.917	2.008	1.811	1.966	4.63	
74)		66	styrene	1.519	1.800	1.750	1.671	1.700	1.495	1.656	7.48	
75)		67	o-xylene	1.910	2.052	2.058	1.979	2.062	1.852	1.985	4.45	
76)	p	68	1122-Tetra-C	0.365	0.401	0.408	0.385	0.426	0.391	0.396	5.24	
77)		69	123-tri-Cl-P	0.112	0.134	0.135	0.128	0.145	0.134	0.131	8.44	
78)	S	70	4-Br-1-F-Bz	0.705	0.688	0.678	0.659	0.699	0.637	0.678	3.76	
79)		71	isopropylben	2.428	2.582	2.790	2.605	2.869	2.653	2.654	5.92	
80)		72	bromobenzene	0.632	0.709	0.694	0.667	0.704	0.626	0.672	5.39	
81)		92	t-1,4-dichlo	0.028	0.044	0.062	0.068	0.087	0.083	0.062	36.62	0.993
82)		73	n-propylbenz	0.694	0.798	0.823	0.758	0.809	0.730	0.769	6.53	
83)		74	2-Cl-Toluene	0.578	0.697	0.689	0.644	0.680	0.623	0.652	7.07	
84)		75	4-Cl-Toluene	0.701	0.729	0.687	0.654	0.666	0.583	0.670	7.50	

(#) = Out of Range

Response F or Report GCMS-A

Method : C:\MSDCHEM\1\METHODS\E524A003.M (RTE Integrator)
Title : **Applied P &Ch Lab** EPA 524.2
Last Update : Mon Oct 27 13:56:31 2003
Response via : Initial Calibration

Calibration Files
.3 =3-A0003.D 2 =3-002.D 10 =3-0010.D
20 =3-0020.D 40 =3-0040.D 60 =3-0060.D

		Compound	.3	2	10	20	40	60	Avg	%RSD	
85)	76	135-tri-Me-B	2.192	2.419	2.462	2.318	2.449	2.206	2.341	5.17	
86)	77	4-iso-Pr-tol	2.421	2.565	2.663	2.436	2.642	2.402	2.522	4.63	
87)	78	124-tri-Me-B	2.287	2.513	2.498	2.427	2.600	2.382	2.451	4.48	
88)	79	tert-butylbe	1.853	2.009	2.152	2.004	2.256	2.057	2.055	6.71	
89)	80	13-DCB	1.484	1.490	1.416	1.357	1.396	1.232	1.396	6.84	
90)	81	sec-butylben	2.948	3.045	3.213	2.920	3.279	3.007	3.069	4.75	
91)	82	14-DCB	1.660	1.512	1.434	1.366	1.463	1.329	1.461	8.06	
92)	83	Cl-benzyl	0.066	0.066	0.102	0.123	0.168	0.168	0.116	39.76	0.989
93)	84	12-DCB	1.352	1.374	1.295	1.218	1.283	1.149	1.278	6.57	
94)	85	n-butylbenze	0.623	0.678	0.712	0.654	0.707	0.621	0.666	5.99	
95)	86	12-diBr-2-Cl	0.079	0.085	0.093	0.096	0.111	0.107	0.095	13.30	
96)	87	124-tri-Cl-B	0.700	0.756	0.822	0.810	0.879	0.811	0.796	7.71	
97)	88	naphthalene	0.845	0.988	1.263	1.338	1.571	1.528	1.256	23.09	0.997
98)	89	hx-Cl-butadi	0.492	0.456	0.476	0.416	0.468	0.426	0.456	6.44	
99)	90	123-Tri-Cl-B	0.592	0.656	0.713	0.693	0.770	0.700	0.687	8.70	

(#) = Out of Range

BFB

Data File : C:\MSDCHEM\1\DATA\03G4658\G4658P01.D Vial: 1
Acq On : 31 Oct 2003 10:43 am Operator: zou
Sample : ##03g4658,w 50ng Inst : GCMS-A
Misc : Multiplr: 1.00
MS Integration Params: Lscint.p
Method : C:\MSDCHEM\1\METHODS\E524A003.M (RTE Integrator)
Title : **Applied P &Ch Lab** EPA 524.2



Spectrum Information: Average of 12.717 to 12.741 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	10959	PASS
75	95	30	60	49.1	29050	PASS
95	95	100	100	100.0	59148	PASS
96	95	5	9	6.7	3982	PASS
173	174	0.00	2	0.3	131	PASS
174	95	50	100	83.6	49424	PASS
175	174	5	9	7.7	3792	PASS
176	174	95	101	95.2	47030	PASS
177	176	5	9	7.0	3282	PASS

G4658P01.D E524A003.M Mon Nov 03 11:05:44 2003

FORM-5A

Applied P & Ch Laboratory

Volatile Organic Instrument Performance Check for Method 524.2

Bromofluorobenzene (BFB), Part II

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		Service ID:	035874
Project ID:	JPL	BFB Inj. Date:	10/31/03	Batch No:	03G4658
		BFB Inj. Time:	10:43	Sequence No:	03G4658
Project No:	04.4428.10	Instrument ID:	A	GC Column:	HP-VOC
Data File Name:	G4658P01	Heated Purge: (Y/N)	N	Column ID:	0.20 mm

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS AND STANDARDS:

#	Client Sample No	Lab Sample ID	Data File Name	Date Analyzed	Time Analyzed
1	03G4658-CCV-01	03G4658-CCV-01	G4658Q01	10/31/03	11:09
2	03G4658-LCS-01	03G4658-LCS-01	G4658L01	10/31/03	11:34
3	MW-18-2MS	03-5874-3MS	G4658M01	10/31/03	12:16
4	MW-18-2MSD	03-5874-3MSD	G4658N01	10/31/03	12:42
5	MW-19-3MS	03-5874-9MS	G4658M02	10/31/03	13:07
6	MW-19-3MSD	03-5874-9MSD	G4658N02	10/31/03	13:33
7	03G4658-MB-01	03G4658-MB-01	G4658K01	10/31/03	15:43
8	TB-6-10-29-03	03-5874-12	5874-12	10/31/03	16:10
9	MW-18-2	03-5874-3	5874-03	10/31/03	16:36
10	MW-19-3	03-5874-9	5874-09	10/31/03	17:02
11	EB-6-10-29-03	03-5874-1	5874-01	10/31/03	17:28
12	MW-18-1	03-5874-2	5874-02	10/31/03	17:54
13	MW-18-3	03-5874-4	5874-04	10/31/03	18:20
14	MW-18-4	03-5874-5	5874-05	10/31/03	18:46
15	MW-18-5	03-5874-6	5874-06	10/31/03	19:12
16	MW-19-1	03-5874-7	5874-07	10/31/03	19:38
17	MW-19-2	03-5874-8	5874-08	10/31/03	20:05
18	MW-19-4	03-5874-10	5874-10	10/31/03	20:31
19	MW-19-5	03-5874-11	5874-11	10/31/03	20:57
20					
21					
22					
23					
24					
25					

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G4658\G4658Q01.D Vial: 2
Acq On : 31 Oct 2003 11:09 am Operator: zou
Sample : f=1 Inst : GCMS-A
Misc : Multiplr: 1.00
MS Integration Params: Lscint.p

Method : C:\MSDCHEM\1\METHODS\E524A003.M (RTE Integrator)
Title : **Applied P &Ch Lab** EPA 524.2
Last Update : Mon Oct 27 14:05:06 2003
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 200%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1	I	1 Fluorobenzene	1.000	1.000	0.0	58	0.00
2		3 di-Cl-di-F-methane	0.201	0.194	3.5	71	0.00
3	p	4 Chloromethane	0.200	0.148	26.0#	47#	0.00
4		2 F114	0.091	0.092	-1.1	66	0.00
5	C	5 vinyl chloride	0.171	0.156	8.8	61	0.00
6		6 bromomethane	0.097	0.065	33.0#	43#	0.00
7		7 chloroethane	0.092	0.079	14.1	55	0.00
8		8 tri-Cl-F-methane	0.282	0.323	-14.5	78	0.00
9		91 Acetonitrile X10	0.043	0.038	11.6	53	0.00
10		9 acrolein X10	0.021	0.020	4.8	59	0.00
11		11 acetone X10	0.032	0.057	-78.1#	117	-0.02
12		12 ethyl ether X5	0.095	0.089	6.3	56	0.00
13	M, C	13 11-dichloroethene	0.238	0.244	-2.5	66	0.00
14		14 Iodomethane	0.178	0.155	12.9	48#	0.00
15		15 F-113	0.144	0.166	-15.3	82	0.00
16		16 acrylonitrile X10	0.043	0.039	9.3	54	0.00
17		17 carbon disulfide	0.452	0.403	10.8	58	0.00
18		94 Isopropyl Alcoholx10	0.002	0.003	-50.0#	135	-0.11
19		18 methylene chloride	0.248	0.192	22.6#	52	0.00
20		19 t-12-di-Cl-ethene	0.223	0.211	5.4	58	0.00
21		20 t-Bu-Me-ether	0.393	0.406	-3.3	60	0.00
22		95 Tert butyl alcoholx10	0.007	0.010	-42.9#	145	-0.06
23		94 allyl chloride	0.441	0.376	14.7	53	0.00
24	p	21 11-dichloroethane	0.370	0.344	7.0	55	0.00
25		97 propionitrile	0.018	0.017	5.6	65	-0.01
26		22 c-12-di-Cl-ethene	0.233	0.217	6.9	55	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G4658\G4658Q01.D Vial: 2
Acq On : 31 Oct 2003 11:09 am Operator: zou
Sample : f=1 Inst : GCMS-A
Misc : Multiplr: 1.00
MS Integration Params: Lscint.p

Method : C:\MSDCHEM\1\METHODS\E524A003.M (RTE Integrator)
Title : **Applied P &Ch Lab** EPA 524.2
Last Update : Mon Oct 27 14:05:06 2003
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 200%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
27		23 22-Dichloropropane	0.236	0.306	-29.7#	74	0.00
28		24 Br-Cl-methane	0.118	0.108	8.5	56	0.00
29	C	25 chloroform	0.416	0.393	5.5	59	0.00
30		26 tetrahydrofuranX5	0.035	0.031	11.4	55	-0.01
31		98 Diisopropyl ether	0.710	0.662	6.8	53	0.00
32	S	27 Di-Br-F-Me (surr)	0.235	0.231	1.7	59	0.00
33		99 ETBE	0.448	0.488	-8.9	62	0.00
34	S	29 1,2-Di-Cl-Et-d4 (S1)	0.209	0.208	0.5	60	0.00
35		30 12-dichloroethane	0.081	0.083	-2.5	63	0.00
36		32 vinyl acetate X5	0.346	0.352	-1.7	57	0.00
37		92 Nitro Methane(x10)	0.012	0.005	58.3#	56	-0.04
38		33 2-butanoneMEK X10	0.052	0.066	-26.9#	77	-0.01
39		93 Ethyl Acetate x2	0.129	0.120	7.0	57	0.00
40		34 111-trichloroethane	0.350	0.381	-8.9	67	0.00
41		35 11-Di-Cl-propene	0.264	0.281	-6.4	64	0.00
42	M	36 benzene	0.846	0.787	7.0	55	0.00
43		37 CCl4	0.321	0.375	-16.8	75	0.00
44		100 Isobutyl alcoholx10	0.013	0.014	-7.7	61	0.00
45		38 thiophene	0.440	0.423	3.9	54	0.00
46	C	39 12-di-Cl-propane	0.194	0.179	7.7	54	0.00
47	M	40 trichloroethene	0.271	0.269	0.7	60	0.00
48		41 dibromomethane	0.135	0.129	4.4	58	0.00
49		101 TAME	0.369	0.395	-7.0	61	0.00
50		42 Br-di-Cl-methane	0.305	0.290	4.9	59	0.00
51		43 Me-methacrylate	0.093	0.097	-4.3	57	0.00
52		44 2-ClEt-Vi-ether10	0.028	0.026	7.1	49#	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G4658\G4658Q01.D Vial: 2
Acq On : 31 Oct 2003 11:09 am Operator: zou
Sample : f=1 Inst : GCMS-A
Misc : Multiplr: 1.00
MS Integration Params: Lscint.p

Method : C:\MSDCHEM\1\METHODS\E524A003.M (RTE Integrator)
Title : **Applied P &Ch Lab** EPA 524.2
Last Update : Mon Oct 27 14:05:06 2003
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 200%

			Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
53		45	c-13-di-Cl-propene	0.307	0.312	-1.6	57	0.00
54		46	t-1,3-dichloropropene	0.248	0.269	-8.5	59	0.00
55	I	47	Chlorobezene-d5	1.000	1.000	0.0	59	0.00
56		48	112-tri-Cl-Et	0.221	0.202	8.6	57	0.00
57		49	13-di-Cl-propane	0.340	0.316	7.1	57	0.00
58		50	Et methacrylate	0.255	0.267	-4.7	57	0.00
59		51	di-Br-Cl-methane	0.314	0.302	3.8	61	0.00
60	P	52	bromoform	0.185	0.181	2.2	61	0.00
61		53	1,4-dichlorobutane-2	0.374	0.346	7.5	57	0.00
62		54	MIBK	0.168	0.159	5.4	56	0.00
63	s	55	toluene-d8	1.104	1.041	5.7	57	0.00
64	M,C	56	toluene	1.354	1.200	11.4	56	0.00
65		57	2-hexanone X5	0.109	0.129	-18.3	70	0.00
66		58	12-dibromoethane	0.227	0.210	7.5	57	0.00
67		59	tetra-Cl-ethene	0.373	0.381	-2.1	67	0.00
68	M,P	60	chlorobenzene	0.924	0.870	5.8	58	0.00
69		61	1112-tetra-Cl-Et	0.344	0.337	2.0	60	0.00
70	I	62	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	61	0.00
71		63	1-chlorohexane	0.228	0.234	-2.6	68	0.00
72	C	64	Et-Bz	2.542	2.450	3.6	60	0.00
73		65	m/p-Xylenes X2	1.966	1.901	3.3	60	0.00
74		66	styrene	1.656	1.583	4.4	57	0.00
75		67	o-xylene	1.985	1.922	3.2	59	0.00
76	p	68	1122-Tetra-Cl-Et	0.396	0.368	7.1	58	0.00

(#) = Out of Range

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G4658\G4658Q01.D Vial: 2
Acq On : 31 Oct 2003 11:09 am Operator: zou
Sample : f=1 Inst : GCMS-A
Misc : Multiplr: 1.00
MS Integration Params: Lscint.p

Method : C:\MSDCHEM\1\METHODS\E524A003.M (RTE Integrator)
Title : **Applied P &Ch Lab** EPA 524.2
Last Update : Mon Oct 27 14:05:06 2003
Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 200%

		Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
77		69 123-tri-Cl-Pr	0.131	0.130	0.8	62	0.00
78	S	70 4-Br-1-F-Bz (S3)	0.678	0.653	3.7	60	0.00
79		71 isopropylbenzene	2.654	2.704	-1.9	63	0.00
80		72 bromobenzene	0.672	0.661	1.6	60	0.00
81		92 t-1,4-dichloro-2-butene	0.062	0.074	-19.4	66	0.00
82		73 n-propylbenzene	0.769	0.791	-2.9	63	0.00
83		74 2-Cl-Toluene	0.652	0.653	-0.2	62	0.00
84		75 4-Cl-Toluene	0.670	0.647	3.4	60	0.00
85		76 135-tri-Me-Benzene	2.341	2.368	-1.2	62	0.00
86		77 4-iso-Pr-toluene	2.522	2.612	-3.6	65	0.00
87		78 124-tri-Me-Benzene	2.451	2.418	1.3	60	0.00
88		79 tert-butylbenzene	2.055	2.133	-3.8	65	0.00
89		80 13-DCB	1.396	1.356	2.9	61	0.00
90		81 sec-butylbenzene	3.069	3.152	-2.7	66	0.00
91		82 14-DCB	1.461	1.378	5.7	61	0.00
92		83 Cl-benzyl	0.116	0.160	-37.9#	79	0.00
93		84 12-DCB	1.278	1.213	5.1	60	0.00
94		85 n-butylbenzene	0.666	0.718	-7.8	67	0.00
95		86 12-diBr-2-Cl-Pra	0.095	0.096	-1.1	61	0.00
96		87 124-tri-Cl-Bz	0.796	0.843	-5.9	63	0.00
97		88 naphthalene	1.256	1.352	-7.6	61	0.00
98		89 hx-Cl-butadiene	0.456	0.501	-9.9	73	0.00
99		90 123-Tri-Cl-Bz	0.687	0.728	-6.0	64	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 0

Continuing Calibration Concentration Summary

Data File G4658Q01
Method File E524A003

Compound Name	Amount	Actual	Units	%Dev	Target Response
1 Fluorobenzene	10	10.00	ppb	0.00	556313
3 di-Cl-di-F-methane	20	19.29	ppb	3.53	215887
4 Chloromethane	20	14.81	ppb	25.96	164527
2 F114	20	19.53	ppb	2.37	102782
5 vinyl chloride	20	18.26	ppb	8.68	173278
6 bromomethane	20	13.47	ppb	32.67	72561
7 chloroethane	20	17.21	ppb	13.93	87940
8 tri-Cl-F-methane	20	22.93	ppb	14.64	359623
91 Acetonitrile X10	200	174.71	ppb	12.65	418760
9 acrolein X10	200	190.08	ppb	4.96	223754
11 acetone X10	200	354.01	ppb	77.01	629191
12 ethyl ether X5	100	93.38	ppb	6.62	493588
13 11-dichloroethene	20	20.53	ppb	2.66	271340
14 Iodomethane	20	15.98	ppb	20.11	172152
15 F-113	20	22.97	ppb	14.86	184156
16 acrylonitrile X10	200	180.36	ppb	9.82	430423
17 carbon disulfide	20	17.81	ppb	10.96	448183
94 Isopropyl Alcoholx10	200	340.19	ppb	70.10	38748
18 methylene chloride	20	18.03	ppb	9.83	213113
19 t-12-di-Cl-ethene	20	18.93	ppb	5.36	235095
20 t-Bu-Me-ether	20	20.69	ppb	3.43	452155
95 Tert butyl alcoholx10	200	306.48	ppb	53.24	111880
94 allyl chloride	20	17.07	ppb	14.64	418760
21 11-dichloroethane	20	18.60	ppb	7.02	382598
97 propionitrile	20	19.46	ppb	2.70	19085
22 c-12-di-Cl-ethene	20	18.62	ppb	6.91	241204
23 22-Dichloropropane	20	25.90	ppb	29.51	340051
24 Br-Cl-methane	20	18.29	ppb	8.56	119785
25 chloroform	20	18.88	ppb	5.58	436718
26 tetrahydrofuranX5	100	90.59	ppb	9.41	174601
98 Diisopropyl ether	20	18.64	ppb	6.79	736454
27 Di-Br-F-Me (surr)	20	19.70	ppb	1.52	257265
99 ETBE	20	21.80	ppb	8.98	542848
29 1,2-Di-Cl-Et-d4 (S1)	20	19.87	ppb	0.64	231224
30 12-dichloroethane	20	20.55	ppb	2.75	92663
32 vinyl acetate X5	100	101.57	ppb	1.57	1957788
92 Nitro Methane(x10)	200	81.91	ppb	59.04	53219
33 2-butanoneMEK X10	200	252.25	ppb	26.13	732174
93 Ethyl Acetate x2	40	37.07	ppb	7.33	266405
34 111-trichloroethane	20	21.79	ppb	8.94	423733
35 11-Di-Cl-propene	20	21.27	ppb	6.35	312237
36 benzene	20	18.61	ppb	6.97	875449
37 CCl4	20	23.35	ppb	16.74	417547

Compound Name	Amount	Actual	Units	%Dev	Target Response
100 Isobutyl alcoholx10	200	211.36	ppb	5.68	150991
38 thiophene	20	19.23	ppb	3.86	470919
39 12-di-Cl-propane	20	18.49	ppb	7.53	199093
40 trichloroethene	20	19.82	ppb	0.90	298765
41 dibromomethane	20	19.23	ppb	3.84	144035
101 TAME	20	19.16	ppb	4.20	439558
42 Br-di-Cl-methane	20	19.07	ppb	4.65	323160
43 Me-methacrylate	20	17.82	ppb	10.88	107502
44 2-ClEt-Vi-ether10	200	156.51	ppb	21.74	290826
45 c-13-di-Cl-propene	20	20.32	ppb	1.60	346897
46 t-1,3-dichloropropene	20	19.64	ppb	1.80	299623
47 Chlorobezene-d5	10	10.00	ppb	0.00	434659
48 112-tri-Cl-Et	20	18.25	ppb	8.77	175596
49 13-di-Cl-propane	20	18.59	ppb	7.05	274465
50 Et methacrylate	20	17.27	ppb	13.65	232432

Compound Name	Amount	Actual	Units	%Dev	Target Response
51 di-Br-Cl-methane	20	19.21	ppb	3.94	262349
52 bromoform	20	19.54	ppb	2.29	156926
53 1,4-dichlorobutane-2	20	18.51	ppb	7.45	301058
54 MIBK	20	16.38	ppb	18.12	138071
55 toluene-d8	20	18.87	ppb	5.66	905206
56 toluene	20	17.72	ppb	11.42	1042797
57 2-hexanone X5	100	102.89	ppb	2.89	560749
58 12-dibromoethane	20	18.51	ppb	7.47	182491
59 tetra-Cl-ethene	20	20.46	ppb	2.30	331505
60 chlorobenzene	20	18.82	ppb	5.88	756046
61 1112-tetra-Cl-Et	20	19.60	ppb	1.98	293057
62 1,4-Dichlorobenzene-d4	10	10.00	ppb	0.00	262960
63 1-chlorohexane	20	20.48	ppb	2.40	122983
64 Et-Bz	20	19.28	ppb	3.62	1288504
65 m/p-Xylenes X2	40	38.67	ppb	3.33	1999374
66 styrene	20	19.12	PPB	4.41	832416
67 o-xylene	20	19.36	ppb	3.21	1010627
68 1122-Tetra-Cl-Et	20	18.60	ppb	6.99	193700
69 123-tri-Cl-Pr	20	19.77	ppb	1.13	68365
70 4-Br-1-F-Bz (S3)	20	19.27	ppb	3.64	343401
71 isopropylbenzene	20	20.37	ppb	1.87	1422140
72 bromobenzene	20	19.68	ppb	1.59	347755
92 t-1,4-dichloro-2-butene	20	18.94	ppb	5.32	39164
73 n-propylbenzene	20	20.58	ppb	2.90	415964
74 2-Cl-Toluene	20	20.04	ppb	0.18	343434
75 4-Cl-Toluene	20	19.31	ppb	3.46	340062
76 135-tri-Me-Benzene	20	20.23	ppb	1.17	1245553
77 4-iso-Pr-toluene	20	20.72	ppb	3.59	1373725
78 124-tri-Me-Benzene	20	19.73	ppb	1.36	1271660
79 tert-butylbenzene	20	20.76	ppb	3.79	1121745
80 13-DCB	20	19.43	ppb	2.86	713174
81 sec-butylbenzene	20	20.54	ppb	2.72	1657813

82 14-DCB	20	18.86	ppb	5.70	724582
83 Cl-benzyl	20	20.85	ppb	4.23	84054
84 12-DCB	20	18.98	ppb	5.10	638119
85 n-butylbenzene	20	21.55	ppb	7.77	377441
86 12-diBr-2-Cl-Pra	20	20.20	ppb	0.98	50531
87 124-tri-Cl-Bz	20	21.18	ppb	5.90	443512
88 naphthalene	20	18.44	ppb	7.79	711292
89 hx-Cl-butadiene	20	21.99	ppb	9.96	263512
90 123-Tri-Cl-Bz	20	21.19	ppb	5.96	383050

Average D % 9.4588396

FORM-8A

Applied P & Ch Laboratory

Internal Standard Area and RT Summary for Method 524.2

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		Service ID:	035874
Project ID:	JPL	Project No:	04.4428.10	Sample Matrix:	Water
		CCV Data File:	G4658Q01	Instrument ID:	A

Batch No: 03G4658

#	Client Sample No	Lab Sample ID	Analysis Date & Time	IS-1 Area #	IS-1 RT #	IS-2 Area #	IS-2 RT #	IS-3 Area #	IS-3 RT #
12 Hour CCV STD			10/31/03 11:09	556313	7.64	434659	11.54	262960	13.84
CCV Upper Limit				1112626	8.14	869318	12.04	525920	14.34
CCV Lower Limit				278156	7.14	217329	11.04	131480	13.34
1	03G4658-LCS-01	03G4658-LCS-01	10/31/03 11:34	684645	7.64	517030	11.54	311730	13.84
2	MW-18-2MS	03-5874-3MS	10/31/03 12:16	698416	7.64	510133	11.54	297085	13.84
3	MW-18-2MSD	03-5874-3MSD	10/31/03 12:42	717084	7.64	543647	11.54	321329	13.84
4	MW-19-3MS	03-5874-9MS	10/31/03 13:07	721584	7.64	553095	11.54	335106	13.84
5	MW-19-3MSD	03-5874-9MSD	10/31/03 13:33	713505	7.65	540269	11.54	324941	13.85
6	03G4658-MB-01	03G4658-MB-01	10/31/03 15:43	756269	7.65	535865	11.54	339476	13.84
7	TB-6-10-29-03	03-5874-12	10/31/03 16:10	757784	7.64	538838	11.54	347007	13.85
8	MW-18-2	03-5874-3	10/31/03 16:36	742319	7.64	528882	11.54	335503	13.84
9	MW-19-3	03-5874-9	10/31/03 17:02	728809	7.64	521701	11.54	329428	13.85
10	EB-6-10-29-03	03-5874-1	10/31/03 17:28	741879	7.65	533330	11.54	342153	13.84
11	MW-18-1	03-5874-2	10/31/03 17:54	736729	7.64	523844	11.54	336363	13.84
12	MW-18-3	03-5874-4	10/31/03 18:20	735830	7.64	521224	11.54	335756	13.84
13	MW-18-4	03-5874-5	10/31/03 18:46	719162	7.65	514781	11.55	328453	13.84
14	MW-18-5	03-5874-6	10/31/03 19:12	727213	7.64	516982	11.54	335701	13.85
15	MW-19-1	03-5874-7	10/31/03 19:38	720223	7.64	516849	11.54	330613	13.84
16	MW-19-2	03-5874-8	10/31/03 20:05	722159	7.65	517888	11.54	327072	13.84
17	MW-19-4	03-5874-10	10/31/03 20:31	731788	7.65	516435	11.54	330756	13.84
18	MW-19-5	03-5874-11	10/31/03 20:57	726259	7.65	514768	11.55	338898	13.84
19									
20									
21									
22									

IS-1 = FLUOROBENZENE
IS-2 = CHLOROBENZENE-D5
IS-3 = 1,4-DICHLOROBENZENE-D4

Area Upper Limit = +100% of CCV internal standard area
Area Lower Limit = - 50% of CCV internal standard area
RT Upper Limit = +0.50 minutes of CCV internal standard RT
RT Lower Limit = - 0.50 minutes of CCV internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits

A

VOC-112

pplied P & Ch Laboratory
0 Magnolia Ave. Chino CA 91710
(909) 590-1828 Fax: (909) 590-1498

VOC Analysis General Logbook

Sequence # 0394658 Batch # 0394658 Matrix: W Date: 10/31/03 Analyst: Zou

IS/Surrogate: GC-1576/1576 Methanol(mark-*M*): _____ PEG (mark-*PEG*): _____ Defoaming(mark-*DF*): _____

Calib. + Ini. Batch ☒ Initial Batch; ☐ Middle Batch; ☐ Final Batch. ☐ Internal Study Datafile Path: _____

#	Type	Sample ID	Method	$V/X=f_1$	$V_f/V_i=f_2$	$V_{ppg}/V_{inj}=f_3$	F	A-#	Datafile	Note	pH
553	SP	G4658P01	E544A 003	25/25 = 1	/ =	/ =	1		G4658P01	10/31/03	10:43am
554	CCV	Q01		/ =	/ =	/ =			Q01	GC1576	
555	LCS	L01		/ =	/ =	/ =			L01		
556	MS	M01		/ =	/ =	/ =			M01	S5874-03	<2
557	MSD	N01		/ =	/ =	/ =			N01	↓	↓
558	MS	M02		/ =	/ =	/ =			M02	S5874-09	
559	MSD	N02		/ =	/ =	/ =			N02	↓	
560	MB	K01		/ =	/ =	/ =			K01		
561	sample	5874-12		/ =	/ =	/ =			5874-12	tb	<2
562		03		/ =	/ =	/ =			03	ms	
563		09		/ =	/ =	/ =			09	ms	
564		01		/ =	/ =	/ =			01	eb	
565		02		/ =	/ =	/ =			02		
566		04		/ =	/ =	/ =			04		
567		05		/ =	/ =	/ =			05		
568		06		/ =	/ =	/ =			06		
569		07		/ =	/ =	/ =			07		
570		08		/ =	/ =	/ =			08		
571		10		/ =	/ =	/ =			10		
572		11		/ =	/ =	/ =			11		
573		5896-01		/ =	/ =	/ =			5896-01		
574		5892-02		/ =	/ =	/ =			5892-02		
575		03		/ =	/ =	/ =			03		
576				/ =	/ =	/ =					
577				/ =	/ =	/ =					
578				/ =	/ =	/ =					
579				/ =	/ =	/ =					
580				/ =	/ =	/ =					
581				/ =	/ =	/ =					
582				/ =	/ =	/ =					
583				/ =	/ =	/ =					
584				/ =	/ =	/ =					

11/03/03

Type	Op #	STD Lot #	$C_{std}(ng/\mu L) \times V_{std}(\mu L)/X(g \text{ or } mL) = T$	Op #	STD Lot #	$C_{std}(ng/\mu L) \times V_{std}(\mu L)/X(g \text{ or } mL) = T$
LCS/LCSD	4555	GC-	~~200~~ × ~~2.5~~ /X = ~~20~~ ppb		GC-	× /X = ppb
MS/MSD	4556/4557	GC-1577	× /X = ppb	4558/4559	GC-1577	200 × 2.5 /X = 20 ppb

Footnote/Anomaly:

pH for field water sample only.
Supervisor Initial

A

Applied P & Ch Laboratory

13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

VOC Analysis General Logbook

Sequence # 0364612 Batch # 0364612 Matrix: W. Date: 10/21/03 Analyst: Eddie

Lot #: IS/Surrogate: GC15714/15715 Methanol(mark-*M*): _____ PEG (mark-*PEG*): _____ Defoaming(mark-*DF*): _____

☑ Calib. + Ini. Batch ☐ Initial Batch; ☐ Middle Batch; ☐ Final Batch. ☐ Internal Study Datafile Path: _____

Op. #	Type	Sample ID	Method	$V/X=f_1$	$V_f/V_i=f_2$	$V_{app}/V_{inj}=f_3$	F	A-#	Datafile	Note	pH
4201	SP	G4612P01	E524A	25/25 = 1	/ =	/ =	1		G4612P01	10/21/03 9:14am	
4202	Calib	3-A003	003	/ =	/ =	/ =			3-A003	gc15731	
4203		3-002		/ =	/ =	/ =			3-002		
4204		3-0010		/ =	/ =	/ =			3-0010		
4205		3-0020		/ =	/ =	/ =			3-0020		
4206		3-0040		/ =	/ =	/ =			3-0040		
4207		3-0060		/ =	/ =	/ =			3-0060		
4208	MV	ICV		/ =	/ =	/ =			ICV	gc15732	
4209				/ =	/ =	/ =					
4210				/ =	/ =	/ =					
4211				/ =	/ =	/ =					
4212				/ =	/ =	/ =					
4213				/ =	/ =	/ =					
4214				/ =	/ =	/ =					
4215				/ =	/ =	/ =					
4216				/ =	/ =	/ =					
4217				/ =	/ =	/ =					
4218				/ =	/ =	/ =					
4219				/ =	/ =	/ =					
4220				/ =	/ =	/ =					
4221				/ =	/ =	/ =					
4222				/ =	/ =	/ =					
4223				/ =	/ =	/ =					
4224				/ =	10/21/03	/ =					
4225				/ =	/ =	/ =					
4226				/ =	/ =	/ =					
4227				/ =	/ =	/ =					
4228				/ =	/ =	/ =					
4229				/ =	/ =	/ =					
4230				/ =	/ =	/ =					
4231				/ =	/ =	/ =					
4232				/ =	/ =	/ =					

Type	Op #	STD Lot #	$C_{std}(ng/\mu L)\times V_{std}(\mu L)/X(g \text{ or } mL)=T$	Op #	STD Lot #	$C_{std}(ng/\mu L)\times V_{std}(\mu L)/X(g \text{ or } mL)=T$
LCS/LCSD		GC-	× /X= ppb		GC-	× /X= ppb
MS/MSD		GC-	× /X= ppb		GC-	× /X= ppb

Footnote/Anomaly:

858

FILE: [CUST.DOC.GC]VOC_LOG .TEX Root-File: VOC_LOG_ROOT.TEX 1-Page-File:[CUST.DOC.GC]VOC_LOG1.TEX; pH for field water sample only.
APCL form 7-201, April 03, 1996; Ver 4.1 No pencil. Use blue pen for record. Use red pen for correction. Supervisor Initial _____

Level C Data Package Deliverables

Wet Chemistry

Applied P & Ch Laboratory

Applied P & Ch Laboratory

Wet Analysis Results for Method 7196

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Anal. Method	7196
Project ID:	JPL	Service ID:	35874	Collected by:	Jr

Component Name: Chromium (VI)
CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-5874-1	EB-6-10-29-03	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-2	MW-18-1	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-3	MW-18-2	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-4	MW-18-3	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-5	MW-18-4	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-6	MW-18-5	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-7	MW-19-1	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-8	MW-19-2	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-9	MW-19-3	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-10	MW-19-4	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-11	MW-19-5	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03W4993-MB-01	03W4993-MB-01	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.
Qualifier: U - Not Detected or less than MDL
B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Applied P & Ch Laboratory

Wet Analysis Results for Method 314.0

Client Name:	GEOFON, Inc.	Project No:	04.4428.10	Anal. Method	314.0
Project ID:	JPL	Service ID:	35874	Collected by:	Jr

Component Name: Perchlorate
CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-5874-1	EB-6-10-29-03	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	<4	U
03-5874-2	MW-18-1	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	<4	U
03-5874-3	MW-18-2	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	<4	U
03-5874-4	MW-18-3	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	<4	U
03-5874-5	MW-18-4	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	17.2	
03-5874-6	MW-18-5	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	<4	U
03-5874-7	MW-19-1	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	<4	U
03-5874-8	MW-19-2	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	4.4	
03-5874-9	MW-19-3	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	5.1	
03-5874-10	MW-19-4	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	<4	U
03-5874-11	MW-19-5	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	<4	U
03W4987-MB-01	03W4987-MB-01	Water	10/29/03	10/29/03	10/29/03	03W4987	μg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

FORM-3

Applied P & Ch Laboratory

Lab Control Spike/Lab Control Spike Duplicate Recovery for Method 314.0

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		Service ID:	35874
Project ID:	JPL	Project No:	04.4428.10	Sample Matrix:	Water
		Batch No:	03W4987		
LCS Filename:	-	Date Analyzed:	102903	Time Analyzed:	
LCSD Filename:	-	Date Analyzed:	102903	Time Analyzed:	

Spiked Components	Unit	Spike Added	Concentration Unspiked	Concentration LCS	LCS Rec% #	QC Limit, % REC
PERCHLORATE	μg/L	25	0	28.6	114	80-120
# of Out-of-control					0	

Spiked Components	Unit	Spike Added	LCSD Concentration	LCSD Rec% #	RPD% #	QC Limit, % RPD	QC Limit, % REC
PERCHLORATE	μg/L	25	27.8	111	3	20	80-120
# of Out-of-control				0	0		

Column to be used to flag recovery and RPD values:

* – Values outside of contract required QC Limits D – Spiked components diluted out

Comments: _____

FORM-3

Applied P & Ch Laboratory

Matrix Spike/Matrix Spike Duplicate Recovery for Method 314.0

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		Service ID:	35874
Project ID:	JPL	Project No:	04.4428.10	Sample Matrix:	Water
		Batch No:	03W4987		
MS Filename:	-	Date Analyzed:	102903	Time Analyzed:	
MSD Filename:	-	Date Analyzed:	102903	Time Analyzed:	
MS Sample No:	**MW-19-3**	Sample Lab ID:	03-5874-9		

Spiked Components	Unit	Spike Added	Concentration Unspiked	Concentration MS	MS Rec% #	QC Limit, % REC
PERCHLORATE	μg/L	25	5.1	33.4	113	75-125
# of Out-of-control					0	

Spiked Components	Unit	Spike Added	MSD Concentration	MSD Rec% #	RPD% #	QC Limit, % RPD	QC Limit, % REC
PERCHLORATE	μg/L	25	36.3	125	10	20	75-125
# of Out-of-control				0	0		

Column to be used to flag recovery and RPD values:

* – Values outside of contract required QC Limits D – Spiked components diluted out

Comments: _____

FORM-3

Applied P & Ch Laboratory

Lab Control Spike/Lab Control Spike Duplicate Recovery for Method 7196

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		Service ID:	35874
Project ID:	JPL	Project No:	04.4428.10	Sample Matrix:	Water
		Batch No:	03W4993		
LCS Filename:	-	Date Analyzed:	102903	Time Analyzed:	15:39
LCSD Filename:	-	Date Analyzed:	102903	Time Analyzed:	15:39

Spiked Components	Unit	Spike Added	Concentration Unspiked	Concentration LCS	LCS Rec% #	QC Limit, % REC
CHROMIUM (VI)	mg/L	0.25	0	0.232	93	80-115
# of Out-of-control					0	

Spiked Components	Unit	Spike Added	LCSD Concentration	LCSD Rec% #	RPD% #	QC Limit, % RPD	QC Limit, % REC
CHROMIUM (VI)	mg/L	0.25	0.227	91	2	19	80-115
# of Out-of-control				0	0		

Column to be used to flag recovery and RPD values:

* – Values outside of contract required QC Limits D – Spiked components diluted out

Comments: _____

FORM-3

Applied P & Ch Laboratory

Matrix Spike/Matrix Spike Duplicate Recovery for Method 7196

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		Service ID:	35874
Project ID:	JPL	Project No:	04.4428.10	Sample Matrix:	Water
		Batch No:	03W4993		
MS Filename:	-	Date Analyzed:	102903	Time Analyzed:	15:39
MSD Filename:	-	Date Analyzed:	102903	Time Analyzed:	15:39
MS Sample No:	MW-19-3	Sample Lab ID:	03-5874-9		

Spiked Components	Unit	Spike Added	Concentration Unspiked	Concentration MS	MS Rec% #	QC Limit, % REC
CHROMIUM (VI)	mg/L	0.25	0	0.273	109	78-115
# of Out-of-control					0	

Spiked Components	Unit	Spike Added	MSD Concentration	MSD Rec% #	RPD% #	QC Limit, % RPD	QC Limit, % REC
CHROMIUM (VI)	mg/L	0.25	0.267	107	2	19	78-115
# of Out-of-control				0	0		

Column to be used to flag recovery and RPD values:

* – Values outside of contract required QC Limits D – Spiked components diluted out

Comments: _____

6A

INITIAL CALIBRATION DATA

Lab Name: Applied P & Ch Lab Contract: 35874

Analysis: Chromium (VI) Calibration Date: 07/28/2003

Concentration (mg/L)	0.000	0.0125	0.050	0.125	0.250	0.50
Absorbance	0.000	0.007	0.017	0.107	0.212	0.420

A=-0.001+ 0.846C

A=Absorbance

C=Concentration (mg/L)

r= 0.9999

FORM-7

Applied P & Ch Laboratory

CCV Recovery for Wet Analysis

Client Name:	GEOFON, Inc.	Contract No.:		Lab Code:	APCL
Case No:		SAS No.:		Service ID:	35874
Project ID:	JPL	Project No.:	04.4428.10		

#	Component Name	Method	Batch No.	Unit	Expe-cted	Test Result	Rec. %	Dev. %	Flag	Control Limit, %	Test Date
1	Perchlorate	314.0	03W4987	μg/L	50	57.3	115	15	√	85-115	10/29/2003
	Perchlorate	314.0	03W4987	μg/L	50	56.2	112	12	√	85-115	10/29/2003
	Perchlorate	314.0	03W4987	μg/L	50	55.9	112	12	√	85-115	10/29/2003
	Perchlorate	314.0	03W4987	μg/L	50	57.5	115	15	√	85-115	10/29/2003
2	Chromium (VI)	7196	03W4993	mg/L	0.25	0.245	98	-2	√	90-110	10/29/2003
	Chromium (VI)	7196	03W4993	mg/L	0.25	0.248	99	-1	√	90-110	10/29/2003

Applied P & Ch Laboratory
4760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Chromium (VI) (7196) Worksheet

Batch # PZ Matrix: W&GS [Holding Time: 24 hours!!] Test Date: 7/28/03 Analyst: Br.

Lot #: Reagent Water _____ Diphenylcazide solution _____ Test Time: _____ SOP: C

Calibration	STD Lot #	$C_{std} \times V_{std}/V_f = C_i$	A_i	$RF_i = A_i/C_i$	Calibration results	Note
STD-1	W-7757	× / =0.00 mg/L	0.000		Least Square [RF]=	Cal. Code:
STD-2	W-	× / =0.025 mg/L	0.007		Average $\bar{RF}$=	
STD-3	W-	× / 0.025 ~~0.050~~ mg/L	0.017		C.C.= 0.9999 (≥ 0.995)	
STD-4	W-	× / =0.125 mg/L	0.107		RSD= % (≤ 15%)	
STD-5	W-	× / =0.250 mg/L	0.212		Ref. page	
STD-6	W-	× / =0.50 mg/L	0.420		A=-0.001+0.846C	

Analysis Type	Sample ID or Lot #	Samp. Amnt X_0 (g or mL)	Dilu./Ext $X/X_0=f_1$	Treat. Ratio $V/X=f_2$	540 nm A	Concentration $C'=A/\bar{RF}$	C (Sample) $C=f_1 f_2 C'$	Anom Not
CCV	Lot: W-7853	Expected Conc.:	×	/ =0.25 mg/L	0.218	0.259 mg/L	REC. %	90-110
Method Blank	Bl. Lot:		/X_0 =	95.0/ =	0.000	0.000 mg/L	ppm	
LCS1	Bl. Lot:		/X_0 =	95.0/ =	0.210	0.250 mg/L	ppm	
Sample-1	4177-37	1ml → 100ml	/X_0 = 1	95.0/ = 2	0.290	0.689 mg/L	ppm	
MS on S-1	37	0.5ml → 100ml	/X_0 =	95.0/ = 2	0.287	0.682 mg/L	report	
MSD on S-1	4175-15	10.0g	50.0mL /X_0 = 5	95.0/ = 10	0.050	3.04 mg/L	ppm	
Sample 2	15	4	/X_0 =	95.0/ = 2	0.247	2.94 mg/L	report	
Sample 3			/X_0 =	95.0/ =		mg/L	ppm	
Sample 4			/X_0 =	95.0/ =		mg/L	ppm	
Sample 5			/X_0 =	95.0/ =		mg/L	ppm	
Sample 6			/X_0 =	95.0/ =		mg/L	ppm	
Sample 7			/X_0 =	95.0/ =		mg/L	ppm	
Sample 8			/X_0 =	95.0/ =		mg/L	ppm	
Sample 9			/X_0 =	95.0/ =		mg/L	ppm	
Sample 10			/X_0 =	95.0/ =		mg/L	ppm	
Blank	Lot:		/X_0 =	95.0/ =		mg/L	ppm	
LCS2	Bl. Lot:		/X_0 =	95.0/ =		mg/L	ppm	
Sample 11			/X_0 =	95.0/ =		mg/L	ppm	
Sample 12			/X_0 =	95.0/ =		mg/L	ppm	
Sample 13			/X_0 =	95.0/ =		mg/L	ppm	
Sample 14			/X_0 =	95.0/ =		mg/L	ppm	
Sample 15			/X_0 =	95.0/ =		mg/L	ppm	
Sample 16			/X_0 =	95.0/ =		mg/L	ppm	
Sample 17			/X_0 =	95.0/ =		mg/L	ppm	
Sample 18			/X_0 =	95.0/ =		mg/L	ppm	
Sample 19			/X_0 =	95.0/ =		mg/L	ppm	
Sample 20			/X_0 =	95.0/ =		mg/L	ppm	
MTX Dup.	closing on [illegible]		/X_0 =	95.0/ =	0.218	0.259 mg/L	ppm	

Type	STD Lot #	$C_{STD}(\mu g/mL) \times V_{STD}(mL)/X(g \text{ or } mL) = T$	Spike Rec.	Ctl Limit (W/S)
MS	W-	× / = ppm	%	80-120 %/80-120 %
MSD	W-	× / = ppm	%	.. ..
LCS	W-	× / = ppm	%	80-120 %/80-120 %
LCSD	W-	× / = ppm	%	.. ..

PQL/MDL (in ppm)	
PQL(w)	0.01
PQL(s)	0.05
MDL(w)	0.005
MDL(s)	0.025

868

APCL form 5-155 May 08, 1996. Ver. 3.1 No pencil. Use blue pen for record. Use red pen for correction.
File:[CUST.DOC.WET]CR6.TEX Root-file: CR6_ROOT.TEX 1-Page-File: CR61.TEX
Control limits are subjected to change. The updated values are given in the latest version of APCL Technical Handbook Vol. 2

plied P & Ch Laboratory
60 Magnolia Ave. Chino CA 91710
el: (909) 590-1828 Fax: (909) 590-1498

Chromium (VI) (7196) Worksheet

1

Batch #03W4993 Matrix: W

[Holding Time: 24 hours!!]

Test Date: 10/29/03 Analyst: Pr

Lot #: Reagent Water _____ Diphenylcazide solution _____

Test Time: 15:39 SOP: G-2

Calibration	STD Lot #	$C_{std} \times V_{std}/V_f = C_i$	A_i	$RF_i = A_i/C_i$	Calibration results	Note
STD-1	W-	× / = mg/L			Least Square [RF]=	Cal. Code:
STD-2	W-	× / = mg/L			Average RF=	
STD-3	W-	× / = mg/L			C.C.= 0.9999 (≥ 0.995)	
STD-4	W-	× / = mg/L			RSD= % (≤ 15%)	
STD-5	W-	× / = mg/L			Ref. page	
STD-6	W-	× / = mg/L			A=-0.001+0.8460	

Analysis Type	Sample ID or Lot #	Samp. Amnt X_0 (g or mL)	Dilu./Ext $X/X_0=f_1$	Treat. Ratio $V/X=f_2$	540 nm A	Concentration $C'=A/\bar{RF}$	C (Sample) $C=f_1 f_2 C'$	Anomaly Note
CCV	Lot: W- 7757	Expected Conc.:	×	/ = 0.25 mg/L	0.206	0.245 mg/L	REC. %	90-110 %
Method Blank	Bl. Lot: T1118		$/X_0$ = 1	95.0/ =	0.000	mg/L	0.00 ppm	
LCS1	Bl. Lot:		$/X_0$ =	95.0/ =	0.195	mg/L	0.232 ppm	
Sample-1	5874-1		$/X_0$ =	95.0/ =	0.002	mg/L	0.004 ppm	
MS on S-1	3		$/X_0$ =	95.0/ =	0.230	mg/L	0.273 ppm	
MSD on S-1	3		$/X_0$ =	95.0/ =	0.225	mg/L	0.267 ppm	
Sample 2	2		$/X_0$ =	95.0/ =	0.001	mg/L	0.002 ppm	
Sample 3	3		$/X_0$ =	95.0/ =	0.001	mg/L	0.002 ppm	
Sample 4	4		$/X_0$ =	95.0/ =	0.002	mg/L	0.004 ppm	
Sample 5	5		$/X_0$ =	95.0/ =	0.000	mg/L	0.001 ppm	
Sample 6	6		$/X_0$ =	95.0/ =	0.000	mg/L	0.001 ppm	
Sample 7	7		$/X_0$ =	95.0/ =	0.002	mg/L	0.004 ppm	
Sample 8	8		$/X_0$ =	95.0/ =	0.002	mg/L	0.004 ppm	
Sample 9	9		$/X_0$ =	95.0/ =	0.001	mg/L	0.002 ppm	
Sample 10	10		$/X_0$ =	95.0/ =	0.001	mg/L	0.002 ppm	
Blank	Lot: T1118		$/X_0$ =	95.0/ =	0.000	mg/L	0.001 ppm	
LCS2	Bl. Lot:		$/X_0$ =	95.0/ =	0.191	mg/L	0.227 ppm	
Sample 11	11		$/X_0$ =	95.0/ =	0.002	mg/L	0.004 ppm	
Sample 12 MS	9		$/X_0$ =	95.0/ =	0.229	mg/L	0.272 ppm	
Sample 13 MSD	9		$/X_0$ =	95.0/ =	0.224	mg/L	0.266 ppm	
Sample 14			$/X_0$ =	95.0/ =		mg/L	ppm	
Sample 15			$/X_0$ =	95.0/ =		mg/L	ppm	
Sample 16			$/X_0$ =	95.0/ =		mg/L	ppm	
Sample 17			$/X_0$ =	95.0/ =		mg/L	ppm	
Sample 18			$/X_0$ =	95.0/ =		mg/L	ppm	
Sample 19			$/X_0$ =	95.0/ =		mg/L	ppm	
Sample 20			$/X_0$ =	95.0/ =		mg/L	ppm	
MTX Dup.	closing 0.25 mg/L		$/X_0$ =	95.0/ =	0.209	0.248 mg/L	ppm	

10/29/03

Type	STD Lot #	$C_{STD}(\mu g/mL) \times V_{STD}(mL)/X(g \text{ or } mL) = T$	Spike Rec.	Ctl Limit (W/S)
MS	W- 7757	× / = 0.25 ppm	%	80-120 %/80-120 %
MSD	W-	× / = ppm	%	.. ..
LCS	W- 7853	× / = ppm	%	80-120 %/80-120 %
LCSD	W-	× / = ppm	%	.. ..

PQL/MDL (in ppm)	
PQL(w)	0.01
PQL(s)	0.05
MDL(w)	0.005
MDL(s)	0.025

869

APCL form 5-155 May 08, 1996. Ver. 3.1 No pencil. Use blue pen for record. Use red pen for correction.
File:[CUST.DOC.WET]CR6.TEX Root-file: CR6-ROOT.TEX 1-Page-File: CR61.TEX
Control limits are subjected to change. The updated values are given in the latest version of APCL Technical Handbook

chedule File: C:\PeakNet\schedule\03w4987k.sch

ine	Sample	Sample Type	Level	Method	Data File	Volume	Dilution	Weight
	##03w4987k ipc 25ppb w8032	Sample		e314-011.met	c:\data\03w4987k\w4987k ipc25ppb	1	1	1
:	ccv 50ppb w8082	Sample		e314-011.met	c:\data\03w4987k\w4987k q01	1	1	1
;	lcs 25ppb w8087	Sample		e314-011.met	c:\data\03w4987k\w4987k l01	1	1	1
.	Lcsd 25PPB W8033a	Sample		e314-011.met	c:\data\03w4987k\w4987k j01	1	1	1
.	ICCS 4ppb w8088	Sample		e314-011.met	c:\data\03w4987k\w4987k iccs 4ppb	1	1	1
:	mb	Sample		e314-011.met	c:\data\03w4987k\w4987k k01	1	1	1
	5826-02 f=1	Sample		e314-011.met	c:\data\03w4987k\5826-02	1	1	1
.	5852-01 f=1	Sample		e314-011.met	c:\data\03w4987k\5852-01	1	1	1
	5852-02 f=1	Sample		e314-011.met	c:\data\03w4987k\5852-02	1	1	1
)	5852-03 f=1	Sample		e314-011.met	c:\data\03w4987k\5852-03	1	1	1
l	5852-04 f=1	Sample		e314-011.met	c:\data\03w4987k\5852-04	1	1	1
2	5852-05 f=1	Sample		e314-011.met	c:\data\03w4987k\5852-05	1	1	1
3	ccv 50ppb w8082	Sample		e314-011.met	c:\data\03w4987k\w4987k q02	1	1	1
4	ccb	Sample		e314-011.met	c:\data\03w4987k\w4987k ccb	1	1	1
5	5852-06 F=1	Sample		e314-011.met	c:\data\03w4987k\5852-06	1	1	1
6	5852-07 F=1	Sample		e314-011.met	c:\data\03w4987k\5852-07	1	1	1
7	5852-03 f=50	Sample		e314-011.met	c:\data\03w4987k\5852-03a	1	50	1
8	5852-04 f=1	Sample		e314-011.met	c:\data\03w4987k\5852-04	1	1	1
9	5852-04 f=5	Sample		e314-011.met	c:\data\03w4987k\5852-04a	1	5	1
)	5874-03 f=1	Sample		e314-011.met	c:\data\03w4987k\5874-03	1	1	1
l	5874-09 f=1	Sample		e314-011.met	c:\data\03w4987k\5874-09	1	1	1
2	5874-03 ms 25ppb f=1w8033b	Sample		e314-011.met	c:\data\03w4987k\w4987k m01	1	1	1
3	5874-03 msd 25ppb f=1w8033b	Sample		e314-011.met	c:\data\03w4987k\w4987k n01	1	1	1
4	ccv 50ppb w8082	Sample		e314-011.met	c:\data\03w4987k\w4987k q03	1	1	1
5	ccb	Sample		e314-011.met	c:\data\03w4987k\w4987k ccb	1	1	1
6	5874-01 f=1	Sample		e314-011.met	c:\data\03w4987k\5874-01	1	1	1
7	5874-02 f=1	Sample		e314-011.met	c:\data\03w4987k\5874-02	1	1	1
8	5874-04 f=1	Sample		e314-011.met	c:\data\03w4987k\5874-04	1	1	1
9	5874-05 f=1	Sample		e314-011.met	c:\data\03w4987k\5874-05	1	1	1
)	5874-06 f=1	Sample		e314-011.met	c:\data\03w4987k\5874-06	1	1	1
1	5874-07 f=1	Sample		e314-011.met	c:\data\03w4987k\5874-07	1	1	1
2	5874-08 f=1	Sample		e314-011.met	c:\data\03w4987k\5874-08	1	1	1
3	5874-10 f=1	Sample		e314-011.met	c:\data\03w4987k\5874-10	1	1	1
4	5874-11 f=1	Sample		e314-011.met	c:\data\03w4987k\5874-11	1	1	1
5	ccv 50ppb w8082	Sample		e314-011.met	c:\data\03w4987k\w4987k q04	1	1	1
6	ccb	Sample		e314-011.met	c:\data\03w4987k\w4987k ccb	1	1	1
7	5874-09 ms 25ppb f=1w8033b	Sample		e314-011.met	c:\data\03w4987k\w4987k m02	1	1	1
8	5874-09 msd 25ppb f=1 w8033b	Sample		e314-011.met	c:\data\03w4987k\w4987k n02	1	1	1
9	ccv 50ppb w8082	Sample		e314-011.met	c:\data\03w4987k\w4987k q05	1	1	1
)		Sample		aastopcl.met		1	1	1

Analyst W. W

Date 10/29/03

Instrument IC-K

870

ne	Int. Std.	Comment
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	
	1	

›fault Method Path: C:\PEAKNET\METHOD
›fault Data Path: C:\DATA\03W4560K
›mment:

871

Method:C:\PEAKNET\METHOD\E314-011.MET Updated:03/13/2003 10:39:48 AM Total:1

1 Component:perchlorate
Standard:External Fit Type:Linear
Origin:Force Calibration:Area
r^2=0.999492
Amt=0.0005893*Resp+0



Calibration : 7 points , 0, 2, 10, 25, 50, 75, 100 ppb

Analyst C. W
Date 03/12/03
Instrument IC-1C