NASA JPL
Data Validation Reports
LDC# 10025

Volatiles

LDC

LDC Report# 10025A1

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL

Collection Date: February 14, 2003

LDC Report Date: April 1, 2003

Matrix: Water

Parameters: Volatiles

Validation Level: EPA Level III

Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-1651

Sample Identification

DUPE-4-1Q03
EB-8-2/14/03
MW-12-1
MW-12-2
MW-12-3
MW-12-4
MW-12-5
TB-8-2/14/03
DUPE-4-1Q03MS
DUPE-4-1Q03MSD

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

J Indicates an estimated value.

R Quality control indicates the data is not usable.

N Presumptive evidence of presence of the constituent.

UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples DUPE-4-1Q03 and MW-12-2 were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/L)		RPD
	DUPE-4-1Q03	MW-12-2	
Carbon tetrachloride	0.6	0.5	18
1,3-Dichloropropane	0.6	0.5	18

XVII. Field Blanks

Sample TB-8-2/14/03 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-8-2/14/03 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Compound	Concentration (ug/L)
EB-8-2/14/03	Chloroform 1,3-Dichloropropane	0.4 0.7

NASA JPL
Volatiles - Data Qualification Summary - SDG 03-1651

No Sample Data Qualified in this SDG

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-1651

No Sample Data Qualified in this SDG

10025A

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/14/2003
Project ID:	JPL	Service ID:	31651	Collected by:	
		Lab Sample ID:	03-1651-1	Received Date:	02/14/2003
Sample ID:	**DUPE-4-1Q03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/17/03	Anal. Date:	02/17/03
Batch No:	03G1445	Prep. No:	–	Anal. Time:	18:10
Data File Name:	1651-01	Sample Amount:	25 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	0.6	
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	0.6	
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	104
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	95
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	84
4	TOLUENE-D8	2037-26-5	73-129	98
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	98
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	98
3	FLUOROBENZENE	462-06-6	50-200	102
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

4/1/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/14/2003
Project ID:	JPL	Service ID:	31651	Collected by:	
		Lab Sample ID:	03-1651-2	Received Date:	02/14/2003
Sample ID:	**EB-8-2/14/03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/17/03	Anal. Date:	02/17/03
Batch No:	03G1445	Prep. No:	–	Anal. Time:	18:39
Data File Name:	1651-02	Sample Amount:	25 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	0.4	J
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	0.7	
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	109
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	108
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	97
4	TOLUENE-D8	2037-26-5	73-129	109
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	91
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	95
3	FLUOROBENZENE	462-06-6	50-200	95
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/14/2003
Project ID:	JPL	Service ID:	31651	Collected by:	
		Lab Sample ID:	03-1651-3	Received Date:	02/14/2003
Sample ID:	**MW-12-1**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/17/03	Anal. Date:	02/17/03
Batch No:	03G1445	Prep. No:	–	Anal. Time:	19:08
Data File Name:	1651-03	Sample Amount:	25 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	0.6	
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	μg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	112TRICHLOROTRIFLUOROETHANE	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	94
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	94
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	86
4	TOLUENE-D8	2037-26-5	73-129	95
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	101
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	106
3	FLUOROBENZENE	462-06-6	50-200	101
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/14/2003
Project ID:	JPL	Service ID:	31651	Collected by:	
		Lab Sample ID:	03-1651-4	Received Date:	02/14/2003
Sample ID:	**MW-12-2**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/17/03	Anal. Date:	02/17/03
Batch No:	03G1445	Prep. No:	–	Anal. Time:	19:37
Data File Name:	1651-04	Sample Amount:	25 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	0.5	J
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	0.5	
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

4/1/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	μg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	112TRICHLOROTRIFLUOROETHANE	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	101
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	92
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	80
4	TOLUENE-D8	2037-26-5	73-129	99
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	97
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	94
3	FLUOROBENZENE	462-06-6	50-200	95
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Ga/1/on

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/14/2003
Project ID:	JPL	Service ID:	31651	Collected by:	
		Lab Sample ID:	03-1651-5	Received Date:	02/14/2003
Sample ID:	**MW-12-3**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/17/03	Anal. Date:	02/17/03
Batch No:	03G1445	Prep. No:	–	Anal. Time:	20:06
Data File Name:	1651-05	Sample Amount:	25 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	4.9	
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	2.2	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	108
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	98
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	87
4	TOLUENE-D8	2037-26-5	73-129	99
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	99
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	97
3	FLUOROBENZENE	462-06-6	50-200	103
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

4/1/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/14/2003
Project ID:	JPL	Service ID:	31651	Collected by:	
		Lab Sample ID:	03-1651-6	Received Date:	02/14/2003
Sample ID:	**MW-12-4**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/17/03	Anal. Date:	02/17/03
Batch No:	03G1445	Prep. No:	–	Anal. Time:	20:35
Data File Name:	1651-06	Sample Amount:	25 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	2.3	
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	0.8	
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

4/1/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	0.4	J
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	97
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	95
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	88
4	TOLUENE-D8	2037-26-5	73-129	95
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	100
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	103
3	FLUOROBENZENE	462-06-6	50-200	105
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

Kallum

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/14/2003
Project ID:	JPL	Service ID:	31651	Collected by:	
		Lab Sample ID:	03-1651-7	Received Date:	02/14/2003
Sample ID:	**MW-12-5**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/17/03	Anal. Date:	02/17/03
Batch No:	03G1445	Prep. No:	–	Anal. Time:	21:03
Data File Name:	1651-07	Sample Amount:	25 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	< 0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	< 0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	< 0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	< 0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	< 0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	< 0.5	U
7	N-BUTYLBENZENE	104-51-8	μg/L	0.5	< 0.5	U
8	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	< 0.5	U
9	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	< 0.5	U
10	2-BUTANONE	78-93-3	μg/L	10	< 10	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	< 0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	< 0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	< 0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	< 0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	< 0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	< 0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	< 0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	< 0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	0.5	< 0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	< 0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	< 0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	< 0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	< 0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	< 0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	< 0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	< 0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	< 0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	< 0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	< 0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	< 0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	< 0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	< 0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	< 0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	< 0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	< 0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	< 0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	< 0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	< 0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	< 0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	μg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	112TRICHLOROTRIFLUOROETHANE	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	99
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	91
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	77
4	TOLUENE-D8	2037-26-5	73-129	97
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	112
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	115
3	FLUOROBENZENE	462-06-6	50-200	118
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

4/1/07

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/14/2003
Project ID:	JPL	Service ID:	31651	Collected by:	
		Lab Sample ID:	03-1651-8	Received Date:	02/14/2003
Sample ID:	**TB-8-2/14/03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/17/03	Anal. Date:	02/17/03
Batch No:	03G1445	Prep. No:	–	Anal. Time:	13:23
Data File Name:	1651-08	Sample Amount:	25 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N) N	

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	109
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	100
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	85
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	86
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	91
3	FLUOROBENZENE	462-06-6	50-200	90
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

LDC #: 10025A1

SDG #: 03-1651

Laboratory: Applied P & Ch Laboratory

VALIDATION COMPLETENESS WORKSHEET

Level III

Date: 3-26-03

Page: 1 of 1

Reviewer:

2nd Reviewer:

METHOD: GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 2-14-03
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	%RSD, r2
IV.	Continuing calibration	A	
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	SW	D = 1+4
XVII.	Field blanks	SW	TB = 8* EB=2

Note: A = Acceptable; N = Not provided/applicable; SW = See worksheet; *ND = No compounds detected; R = Rinsate; FB = Field blank; D = Duplicate; TB = Trip blank; EB = Equipment blank

Validated Samples: All H_2Os

1	DUPE-4-1Q03	11	03G1445MB01	21		31	
2	EB-8-2/14/03	12		22		32	
3	MW-12-1	13		23		33	
4	MW-12-2	14		24		34	
5	MW-12-3	15		25		35	
6	MW-12-4	16		26		36	
7	MW-12-5	17		27		37	
8	TB-8-2/14/03	18		28		38	
9	DUPE-4-1Q03MS	19		29		39	
10	DUPE-4-1Q03MSD	20		30		40	

TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method 524.2)

A. Chloromethane	Q. 1,2-Dichloropropane	GG. Xylenes, total	WW. Bromobenzene	MMM. Naphthalene
B. Bromomethane	R. cis-1,3-Dichloropropene	HH. Vinyl acetate	XX. 1,2,3-Trichloropropane	NNN. 1,2,3-Trichlorobenzene
C. Vinyl choride	S. Trichloroethene	II. 2-Chloroethylvinyl ether	YY. n-Propylbenzene	OOO. 1,3,5-Trichlorobenzene
D. Chloroethane	T. Dibromochloromethane	JJ. Dichlorodifluoromethane	ZZ. 2-Chlorotoluene	PPP. trans-1,2-Dichloroethene
E. Methylene chloride	U. 1,1,2-Trichloroethane	KK. Trichlorofluoromethane	AAA. 1,3,5-Trimethylbenzene	QQQ. cis-1,2-Dichloroethene
F. Acetone	V. Benzene	LL. Methyl-tert-butyl ether	BBB. 4-Chlorotoluene	RRR. m,p-Xylenes
G. Carbon disulfide	W. trans-1,3-Dichloropropene	MM. 1,2-Dibromo-3-chloropropane	CCC. tert-Butylbenzene	SSS. o-Xylene
H. 1,1-Dichloroethene	X. Bromoform	NN. Diethyl ether	DDD. 1,2,4-Trimethylbenzene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane
I. 1,1-Dichloroethane	Y. 4-Methyl-2-pentanone	OO. 2,2-Dichloropropane	EEE. sec-Butylbenzene	UUU. Benzyl chloride
J. 1,2-Dichloroethene, total	Z. 2-Hexanone	PP. Bromochloromethane	FFF. 1,3-Dichlorobenzene	VVV. 4-Ethyltoluene
K. Chloroform	AA. Tetrachloroethene	QQ. 1,1-Dichloropropene	GGG. p-Isopropyltoluene	WWW. Ethanol
L. 1,2-Dichloroethane	BB. 1,1,2,2-Tetrachloroethane	RR. Dibromomethane	HHH. 1,4-Dichlorobenzene	XXX. Ethyl ether
M. 2-Butanone	CC. Toluene	SS. 1,3-Dichloropropane	III. n-Butylbenzene	
N. 1,1,1-Trichloroethane	DD. Chlorobenzene	TT. 1,2-Dibromoethane	JJJ. 1,2-Dichlorobenzene	
O. Carbon tetrachloride	EE. Ethylbenzene	UU. 1,1,1,2-Tetrachloroethane	KKK. 1,2,4-Trichlorobenzene	
P. Bromodichloromethane	FF. Styrene	VV. Isopropylbenzene	LLL. Hexachlorobutadiene	

Notes:

LDC #: 10025A1
SDG #: 03-1651

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Reviewer: DM
2nd reviewer:

METHOD: GC/MS VOA (EPA Method 524.2)

Y N N/A Were field duplicate pairs identified in this SDG?
Y N N/A Were target compounds detected in the field duplicate pairs?

Compound	Concentration (ug/L)		RPD
	1	4	
O	0.6	0.5	18
SS	0.6	0.5	18

Compound	Concentration ( )		RPD

Compound	Concentration ( )		RPD

LDC #: 1002SA1
SDG #: 03-1651

VALIDATION FINDINGS WORKSHEET
Field Blanks

Page: 1 of 1
Reviewer:
2nd reviewer:

METHOD: GC/MS VOA (EPA Method 524.2)

Y N N/A Were field blanks identified in this SDG?
Y N N/A Were target compounds detected in the field blanks?

Sample: 2 Field Blank / Trip Blank / Rinsate (circle one)

Compound	Concentration Units (µg/L)
K	0.4
SS	0.7

Sample: _____ Field Blank / Trip Blank / Rinsate (circle one)

Compound	Concentration Units ( )

Sample: _____ Field Blank / Trip Blank / Rinsate (circle one)

Compound	Concentration Units ( )

Laboratory Data Consultants, Inc.
Data Validation Report

Project/Site Name: NASA JPL

Collection Date: February 17, 2003

LDC Report Date: April 1, 2003

Matrix: Water

Parameters: Volatiles

Validation Level: EPA Level III

Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-1684

Sample Identification

EB-9-2/17/03
MW-23-1
MW-23-2
MW-23-3
MW-24-1
MW-24-2
MW-24-3
TB-9-2/17/03
MW-23-3MS
MW-23-3MSD

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

J Indicates an estimated value.

R Quality control indicates the data is not usable.

N Presumptive evidence of presence of the constituent.

UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-9-2/17/03 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-9-2/17/03 was identified as an equipment blank. No volatile contaminants were found in this blank.

NASA JPL
Volatiles - Data Qualification Summary - SDG 03-1684

No Sample Data Qualified in this SDG

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-1684

No Sample Data Qualified in this SDG

10025B

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/17/2003
Project ID:	JPL	Service ID:	31684	Collected by:	
		Lab Sample ID:	03-1684-1	Received Date:	02/17/2003
Sample ID:	**EB-9-2/17/03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/18/03	Anal. Date:	02/18/03
Batch No:	03G1460	Prep. No:	–	Anal. Time:	22:09
Data File Name:	1684-01	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	μg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	1.1 [a]	<1.1	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

4/1/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
42	METHYLENE CHLORIDE	75-09-2	μg/L	1.8 [a]	<1.8	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	1,1,2-TRICHLOROTRIFLUORO	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	111
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	99
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	82
4	TOLUENE-D8	2037-26-5	73-129	107
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	79
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	85
3	FLUOROBENZENE	462-06-6	50-200	84
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

[a] MDL reported.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

4/1/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/17/2003
Project ID:	JPL	Service ID:	31684	Collected by:	
		Lab Sample ID:	03-1684-2	Received Date:	02/17/2003
Sample ID:	**MW-23-1**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/18/03	Anal. Date:	02/18/03
Batch No:	03G1460	Prep. No:	–	Anal. Time:	22:37
Data File Name:	1684-02	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge:	(Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	μg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	0.5	J
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	1.1 [a]	<1.1	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
42	METHYLENE CHLORIDE	75-09-2	µg/L	1.8 [a]	<1.8	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	1.0	
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	1.5	
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2-TRICHLOROTRIFLUORO	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	115
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	98
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	85
4	TOLUENE-D8	2037-26-5	73-129	106
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	75
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	80
3	FLUOROBENZENE	462-06-6	50-200	81
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

[a] MDL reported.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

4/1/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/17/2003
Project ID:	JPL	Service ID:	31684	Collected by:	
		Lab Sample ID:	03-1684-3	Received Date:	02/17/2003
Sample ID:	**MW-23-2**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/18/03	Anal. Date:	02/18/03
Batch No:	03G1460	Prep. No:	–	Anal. Time:	23:05
Data File Name:	1684-03	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	μg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	0.5	
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	1.1 [a]	<1.1	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

4/1/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
42	METHYLENE CHLORIDE	75-09-2	μg/L	1.8 [a]	<1.8	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	0.6	
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	0.7	
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	1,1,2-TRICHLOROTRIFLUORO	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	111
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	96
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	81
4	TOLUENE-D8	2037-26-5	73-129	103
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	80
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	85
3	FLUOROBENZENE	462-06-6	50-200	86
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

[a] MDL reported.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

4/1/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/17/2003
Project ID:	JPL	Service ID:	31684	Collected by:	
		Lab Sample ID:	03-1684-4	Received Date:	02/17/2003
Sample ID:	**MW-23-3**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/18/03	Anal. Date:	02/18/03
Batch No:	03G1460	Prep. No:	–	Anal. Time:	23:33
Data File Name:	1684-04	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	μg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	1.1 [a]	<1.1	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

9/1/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
42	METHYLENE CHLORIDE	75-09-2	μg/L	1.8 [a]	<1.8	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	1,1,2-TRICHLOROTRIFLUORO	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	115
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	102
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	88
4	TOLUENE-D8	2037-26-5	73-129	113
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	76
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	80
3	FLUOROBENZENE	462-06-6	50-200	80
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

[a] MDL reported.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

9/1/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/17/2003
Project ID:	JPL	Service ID:	31684	Collected by:	
		Lab Sample ID:	03-1684-6	Received Date:	02/17/2003
Sample ID:	**MW-24-1**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/19/03	Anal. Date:	02/19/03
Batch No:	03G1460	Prep. No:	–	Anal. Time:	00:02
Data File Name:	1684-06	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	μg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	4.7	
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	2.4	
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	1.1 [a]	<1.1	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

9/1/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
42	METHYLENE CHLORIDE	75-09-2	μg/L	1.8 [a]	<1.8	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	0.5	J
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	1.7	
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	1,1,2-TRICHLOROTRIFLUORO	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	107
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	95
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	83
4	TOLUENE-D8	2037-26-5	73-129	103
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	83
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	90
3	FLUOROBENZENE	462-06-6	50-200	87
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

[a] MDL reported.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

4/1/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/17/2003
Project ID:	JPL	Service ID:	31684	Collected by:	
		Lab Sample ID:	03-1684-7	Received Date:	02/17/2003
Sample ID:	**MW-24-2**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/19/03	Anal. Date:	02/19/03
Batch No:	03G1460	Prep. No:	–	Anal. Time:	00:31
Data File Name:	1684-07	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	µg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	8.9	
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	2.8	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	1.1 [a]	<1.1	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	0.5	J
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

4/1/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
42	METHYLENE CHLORIDE	75-09-2	μg/L	1.8 [a]	<1.8	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	1.3	
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	1,1,2-TRICHLOROTRIFLUORO	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	108
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	95
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	82
4	TOLUENE-D8	2037-26-5	73-129	101
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	86
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	88
3	FLUOROBENZENE	462-06-6	50-200	91
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

[a] MDL reported.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

4/1/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/17/2003
Project ID:	JPL	Service ID:	31684	Collected by:	
		Lab Sample ID:	03-1684-8	Received Date:	02/17/2003
Sample ID:	**MW-24-3**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/19/03	Anal. Date:	02/19/03
Batch No:	03G1460	Prep. No:	–	Anal. Time:	01:00
Data File Name:	1684-08	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	μg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	1.1 [a]	<1.1	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

4/1/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
42	METHYLENE CHLORIDE	75-09-2	μg/L	1.8 [a]	<1.8	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	1,1,2-TRICHLOROTRIFLUORO	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	108
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	97
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	80
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	82
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	86
3	FLUOROBENZENE	462-06-6	50-200	86
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

[a] MDL reported.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

4/1/03

Applied P & Ch Laboratory

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	02/17/2003
Project ID:	JPL	Service ID:	31684	Collected by:	
		Lab Sample ID:	03-1684-10	Received Date:	02/17/2003
Sample ID:	**TB-9-2/17/03**	Sample Matrix	Water	Moisture %:	–
Sample Type:	Field Sample	Prep. Method:	5030	Instrument ID:	GC/MS: G
Anal. Method:	524.2	Prep. Date:	02/19/03	Anal. Date:	02/19/03
Batch No:	03G1460	Prep. No:	–	Anal. Time:	01:29
Data File Name:	1684-10	Sample Amount:	25.0 mL	Dilution Factor:	1
Methanol Vol.	–				
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	μg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	μg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	μg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	μg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	μg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	μg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	μg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	μg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	μg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	μg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	μg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	μg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	μg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	μg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	μg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	μg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	μg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	μg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	μg/L	1.1 [a]	<1.1	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	μg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	μg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	μg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	μg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	μg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	μg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	μg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	μg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	μg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	μg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	μg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	μg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	μg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	μg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	μg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	μg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	μg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	μg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	μg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	μg/L	0.5	<0.5	U

4/1/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	μg/L	0.5	<0.5	U
41	4-METHYL-2-PENTANONE (MIBK)	108-10-1	μg/L	10	<10	U
42	METHYLENE CHLORIDE	75-09-2	μg/L	1.8 [a]	<1.8	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	μg/L	1	<1	U
44	NAPHTHALENE	91-20-3	μg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	μg/L	0.5	<0.5	U
46	STYRENE	100-42-5	μg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	μg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	μg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	μg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	μg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	μg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	μg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	μg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	μg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	μg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	μg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	μg/L	0.5	<0.5	U
58	1,1,2-TRICHLOROTRIFLUORO	76-13-1	μg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	μg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	μg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	μg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	μg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	μg/L	0.5	<0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4	70-129	110
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	102
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	86
4	TOLUENE-D8	2037-26-5	73-129	108
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	81
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	89
3	FLUOROBENZENE	462-06-6	50-200	85
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

[a] MDL reported.

Qualifier: U - Not Detected or less than MDL
J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
E - Exceed calibration range
B - A positive value was found in the method blank
D - Diluted

4/1/03

LDC #: 10025B1

SDG #: 03-1684

Laboratory: Applied P & Ch Laboratory

VALIDATION COMPLETENESS WORKSHEET

Level III

Date: 3-26-03

Page: 1 of 1

Reviewer:

2nd Reviewer:

METHOD: GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 2-17-03
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	%RSD, r2
IV.	Continuing calibration	A	
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	ND	TB = 8 EB = 1

Note: A = Acceptable; N = Not provided/applicable; SW = See worksheet; ND = No compounds detected; R = Rinsate; FB = Field blank; D = Duplicate; TB = Trip blank; EB = Equipment blank

Validated Samples: All H_2Os

1	EB-9-2/17/03	11	03G1460MB01	21		31	
2	MW-23-1	12		22		32	
3	MW-23-2	13		23		33	
4	MW-23-3	14		24		34	
5	MW-24-1	15		25		35	
6	MW-24-2	16		26		36	
7	MW-24-3	17		27		37	
8	TB-9-2/17/03	18		28		38	
9	MW-23-3MS	19		29		39	
10	MW-23-3MSD	20		30		40	