

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 19, 2001
LDC Report Date: August 14, 2001
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-4794

Sample Identification

ER-22
MW-8
MW-22-1
MW-22-2
MW-22-3
Trip Blank
MW-8MS
MW-8MSD

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals. All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0%.

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample ER-22 was identified as an equipment rinsate. No volatile contaminants were found in this blank with the following exceptions:

Equipment Rinsate	Sampling Date	Compound	Concentration	Associated Samples
ER-22	7/19/01	Toluene	0.5 ug/L	MW-8 MW-22-1 MW-22-2 MW-22-3

Sample "TRIP BLANK" was identified as a trip blank. No volatile contaminants were found in this blank.

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019
Volatiles - Data Qualification Summary - SDG 01-4794

No Sample Data Qualified in this SDG

JPL, 00HW019
Volatiles - Laboratory Blank Data Qualification Summary - SDG 01-4794

No Sample Data Qualified in this SDG

JPL, 00HW019
Volatiles - Field Blank Data Qualification Summary - SDG 01-4794

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 5, 2001
LDC Report Date: August 16, 2001
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 01-4536

Sample Identification

MW-18-5
MW-18-4
MW-18-3
MW-18-2
ER-18
Trip Blank
MW-18-2MS
MW-18-2MSD
ER-18RE

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met with the following exceptions:

Sample	Compound	Total Days From Sample Collection Until Analysis	Required Holding Time (in Days) From Sample Collection Until Analysis	Flag	A or P
ER-18RE	All TCL compounds	34	14	J (all detects) R (all non-detects)	A

Non-detected sample concentrations were qualified as unusable (R) due to a gross exceedance (>2X) of holding time.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals. All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
7/9/01	Bromomethane	41.15	MW-18-5 MW-18-4 MW-18-3 MW-18-2 ER-18 Trip Blank MW-18-2MS MW-18-2MSD	J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
01G3479MB01	7/9/01	Methylene chloride	1 ug/L	MW-18-5 MW-18-4 MW-18-3 MW-18-2 ER-18 Trip Blank

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
Trip Blank	Methylene chloride	2 ug/L	2U ug/L

Samples ER-18 and ER-18RE were identified as equipment rinsates. No volatile contaminants were found in these blanks with the following exceptions:

Equipment Rinsate	Sampling Date	Compound	Concentration	Associated Samples
ER-18	7/5/01	cis-1,2-Dichloroethene 1,2-Dichloroethene, total Trichloroethene	2.9 ug/L 2.9 ug/L 20.6 ug/L	MW-18-5 MW-18-4 MW-18-3 MW-18-2

Sample "TRIP BLANK" was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank	Sampling Date	Compound	Concentration	Associated Samples
Trip Blank	7/5/01	Methylene chloride	2 ug/L	MW-18-5 MW-18-4 MW-18-3 MW-18-2 ER-18

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks with the following exceptions:

Sample	Compound	Reported Concentration	Modified Final Concentration
MW-18-2	Trichloroethene	0.8 ug/L	0.8U ug/L
MW-18-3	Trichloroethene	0.4 ug/L	0.5U ug/L
MW-18-4	Trichloroethene	1.1 ug/L	1.1U ug/L

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

The original analysis of equipment-rinsate sample ER-18 resulted in high detections of 1,2-dichloroethane and trichloroethene. The sample was reanalyzed resulting in no detections however it exceeded the holding time criteria.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019
Volatiles - Data Qualification Summary - SDG 01-4536

SDG	Sample	Compound	Flag	A or P	Reason
01-4536	MW-18-5 MW-18-4 MW-18-3 MW-18-2 ER-18 Trip Blank	Bromomethane	J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

JPL, 00HW019
Volatiles - Laboratory Blank Data Qualification Summary - SDG 01-4536

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P
01-4536	Trip Blank	Methylene chloride	2U ug/L	A

JPL, 00HW019
Volatiles - Field Blank Data Qualification Summary - SDG 01-4536

SDG	Sample	Compound	Modified Final Concentration	A or P
01-4536	MW-18-2	Trichloroethene	0.8U ug/L	A
01-4536	MW-18-3	Trichloroethene	0.5U ug/L	A
01-4536	MW-18-4	Trichloroethene	1.1U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 31, 2001
LDC Report Date: August 23, 2001
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-5020

Sample Identification

ER-23
MW-23-3
MW-23-2
MW-23-1
TRIP BLANK

Introduction

This data review covers 5 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 with the following exceptions:

Date	Compound	r^2	Associated Samples	Flag	A or P
7/19/01	Chloromethane	0.9897	All samples in SDG 01-5020	J (all detects) UJ (all non-detects)	P

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample ER-23 was identified as an equipment rinsate. No volatile contaminants were found in this blank with the following exceptions:

Equipment Rinsate ID	Sampling Date	Compound	Concentration	Associated Samples
ER-23	7/31/01	Chloroform Bromodichloromethane	0.3 ug/L 0.4 ug/L	MW-23-3 MW-23-2 MW-23-1

Sample "TRIP BLANK" was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Sampling Date	Compound	Concentration	Associated Samples
TRIP BLANK	7/31/01	Methylene chloride	1.6 ug/L	ER-23 MW-23-3 MW-23-2 MW-23-1

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks with the following exceptions:

Sample	Compound	Reported Concentration	Modified Final Concentration
MW-23-2	Chloroform	0.5 ug/L	0.5U ug/L
MW-23-1	Chloroform	0.5 ug/L	0.5U ug/L

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

- All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

All tentatively identified compounds were within validation criteria.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 01-5020

SDG	Sample	Compound	Flag	A or P	Reason
01-5020	ER-23 MW-23-3 MW-23-2 MW-23-1 TRIP BLANK	Chloromethane	J (all detects) UJ (all non-detects)	P	Initial calibration (r ²)

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 01-5020

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 01-5020

SDG	Sample	Compound	Modified Final Concentration	A or P
01-5020	MW-23-2	Chloroform	0.5U ug/L	A
01-5020	MW-23-1	Chloroform	0.5U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 20, 2001
LDC Report Date: August 23, 2001
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-4824

Sample Identification

ER-11
MW-11-1
MW-11-4
MW-11-3
MW-11-2
TRIP BLANK
ER-11MS
ER-11MSD

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 with the following exceptions:

Date	Compound	r^2	Associated Samples	Flag	A or P
7/19/01	Chloromethane	0.9897	All samples in SDG 01-4824	J (all detects) UJ (all non-detects)	P

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
01G3700MB01	7/26/01	Methylene chloride	2.4 ug/L	ER-11

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
01G3679MB01	7/25/01	Methylene chloride	0.8 ug/L	MW-11-1 MW-11-4 MW-11-3 MW-11-2 TRIP BLANK

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
TRIP BLANK	Methylene chloride	2.1 ug/L	2.1U ug/L

Sample ER-11 was identified as an equipment rinsate. No volatile contaminants were found in this blank.

Sample "TRIP BLANK" was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Sampling Date	Compound	Concentration	Associated Samples
TRIP BLANK	7/20/01	Methylene chloride	2.1 ug/L	ER-11 MW-11-1 MW-11-4 MW-11-3 MW-11-2

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

All tentatively identified compounds were within validation criteria.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 01-4824

SDG	Sample	Compound	Flag	A or P	Reason
01-4824	ER-11 MW-11-1 MW-11-4 MW-11-3 MW-11-2 TRIP BLANK	Chloromethane	J (all detects) UJ (all non-detects)	P	Initial calibration (r^2)

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 01-4824

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P
01-4824	TRIP BLANK	Methylene chloride	2.1U ug/L	A

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 01-4824

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 25, 2001
LDC Report Date: August 23, 2001
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-4919

Sample Identification

ER-12
MW-12-1
MW-12-5
MW-12-4
MW-12-3
MW-12-2
TRIP BLANK

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 with the following exceptions:

Date	Compound	r^2	Associated Samples	Flag	A or P
7/19/01	Chloromethane	0.9897	All samples in SDG 01-4919	J (all detects) UJ (all non-detects)	P

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
7/30/01	Chloromethane	32.62	MW-12-1 MW-12-2 01G3754MB01	J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants

were found in the method blanks.

Sample ER-12 was identified as an equipment rinsate. No volatile contaminants were found in this blank.

Sample "TRIP BLANK" was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Sampling Date	Compound	Concentration	Associated Samples
TRIP BLANK	7/25/01	Methylene chloride	1.9 ug/L	ER-12 MW-12-1 MW-12-5 MW-12-4 MW-12-3 MW-12-2

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits with the following exceptions:

Sample	Surrogate	%R (Limits)	Compound	Flag	A or P
ER-12	Bromofluorobenzene	123 (75-121)	All TCL compounds	J (all detects)	P

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

Spike ID (Associated Samples)	Compound	MS (%R) (Limits)	MSD (%R) (Limits)	RPD (Limits)	Flag	A or P
MW-4-1MS/MSD (All samples in SDG 01-4919)	Trichloroethene	66 (70-129)	68 (70-129)	-	J (all detects) UJ (all non-detects)	A

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

All tentatively identified compounds were within validation criteria.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 01-4919

SDG	Sample	Compound	Flag	A or P	Reason
01-4919	ER-12 MW-12-1 MW-12-5 MW-12-4 MW-12-3 MW-12-2 TRIP BLANK	Chloromethane	J (all detects) UJ (all non-detects)	P	Initial calibration (r ²)
01-4919	MW-12-1 MW-12-2	Chloromethane	J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)
01-4919	ER-12	All TCL compounds	J (all detects)	P	Surrogate spikes (%R)
01-4919	ER-12 MW-12-1 MW-12-5 MW-12-4 MW-12-3 MW-12-2 TRIP BLANK	Trichloroethene	J (all detects) UJ (all non-detects)	A	Matrix spike/Matrix spike duplicates (%R)

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 01-4919

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 01-4919

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 26, 2001
LDC Report Date: August 23, 2001
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-4946

Sample Identification

ER-24
FIELD BLANK
MW-24-3
MW-24-3D
MW-24-2
MW-24-1
TRIP BLANK
ER-24MS
ER-24MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

- A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample ER-24 was identified as an equipment rinsate. No volatile contaminants were found in this blank.

Sample "FIELD BLANK" was identified as a field blank. No volatile contaminants were found in this blank.

Sample "TRIP BLANK" was identified as a trip blank. No volatile contaminants were found in this blank.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits with the following exceptions:

Sample	Surrogate	%R (Limits)	Compound	Flag	A or P
TRIP BLANK	Bromofluorobenzene	122 (75-121)	All TCL compounds	J (all detects)	P

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

All tentatively identified compounds were within validation criteria.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples MW-24-3 and MW-24-3D were identified as field duplicates. No volatiles were detected in any of the samples.

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 01-4946

SDG	Sample	Compound	Flag	A or P	Reason
01-4946	TRIP BLANK	All TCL compounds	J (all detects)	P	Surrogate spikes (%R)

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 01-4946

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 01-4946

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 27, 2001
LDC Report Date: August 24, 2001
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 01-4964

Sample Identification

ER-4
MW-4-1
MW-4-3
MW-4-2
TRIP BLANK
MW-4-1MS
MW-4-1MSD

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 with the following exceptions:

Date	Compound	r^2	Associated Samples	Flag	A or P
7/19/01	Chloromethane	0.9897	All samples in SDG 01-4964	J (all detects) UJ (all non-detects)	P

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
7/30/01	Chloromethane	32.62	All samples in SDG 01-4964	J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

Sample ER-4 was identified as an equipment rinsate. No volatile contaminants were found in this blank with the following exceptions:

Equipment Rinsate ID	Sampling Date	Compound	Concentration	Associated Samples
ER-4	7/27/01	Bromodichloromethane	0.5 ug/L	MW-4-1 MW-4-3 MW-4-2

Sample "TRIP BLANK" was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Sampling Date	Compound	Concentration	Associated Samples
TRIP BLANK	7/27/01	Methylene chloride	1.9 ug/L	ER-4 MW-4-1 MW-4-3 MW-4-2

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limit with the following exceptions:

Spike ID (Associated Samples)	Compound	MS (%R) (Limits)	MSD (%R) (Limits)	RPD (Limits)	Flag	A or P
MW-4-1MS/MSD (All samples in SDG 01-4964)	Trichloroethene	66 (70-129)	63 (70-129)	-	J (all detects) UJ (all non-detects)	A

VIII. Laboratory Control Samples (LCS)

- Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 01-4964

SDG	Sample	Compound	Flag	A or P	Reason
01-4964	ER-4 MW-4-1 MW-4-3 MW-4-2 TRIP BLANK	Chloromethane	J (all detects) UJ (all non-detects)	P	Initial calibration (r^2)
01-4964	ER-4 MW-4-1 MW-4-3 MW-4-2 TRIP BLANK	Chloromethane	J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)
01-4964	ER-4 MW-4-1 MW-4-3 MW-4-2 TRIP BLANK	Trichloroethene	J (all detects) UJ (all non-detects)	A	Matrix spike/Matrix spike duplicates (%R)

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 01-4964

No Sample Data Qualified in this SDG

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 01-4964

No Sample Data Qualified in this SDG