

**NASA JPL  
Data Validation Reports  
LDC# 11272**

Wet Chemistry



**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** NASA JPL  
**Collection Date:** October 30, 2003  
**LDC Report Date:** December 15, 2003  
**Matrix:** Water  
**Parameters:** Wet Chemistry  
**Validation Level:** EPA Level III  
**Laboratory:** Applied P & Ch Laboratory

**Sample Delivery Group (SDG):** 03-5892

**Sample Identification**

EB-7-10-30-03  
MW-21-1  
MW-21-2  
MW-21-3  
MW-21-4  
MW-21-5  
MW-21-4MS  
MW-21-4MSD

## Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196A for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## II. Calibration

### a. Initial Calibration

All criteria for the initial calibration of each method were met.

### b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable with the following exceptions:

Date	Lab. Reference/ID	Analyte	%R (Limits)	Associated Samples	Flag	A or P
11/4/03	CCV	Perchlorate	115 (90-110)	MW-21-4MS MW-21-4MSD	J (all detects)	P

## III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

## IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

## VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## VII. Sample Result Verification

Raw data were not reviewed for this SDG.

### **VIII. Overall Assessment of Data**

Data flags are summarized at the end of this report.

### **IX. Field Duplicates**

No field duplicates were identified in this SDG.

### **X. Field Blanks**

Sample EB-7-10-30-03 was identified as an equipment blank. No contaminant concentrations were found in this blank.

**NASA JPL**  
**Wet Chemistry - Data Qualification Summary - SDG 03-5892**

No Sample Data Qualified in this SDG

**NASA JPL**  
**Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-5892**

No Sample Data Qualified in this SDG

**NASA JPL**  
**Wet Chemistry - Field Blank Data Qualification Summary - SDG 03-5892**

No Sample Data Qualified in this SDG

A

Applied P & Ch Laboratory

### Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.

Project No: 04-4428.10

Anal. Method 7196

Project ID: JPL

Service ID: 35892

Collected by: JR

Component Name: Chromium (VI)

CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-5892-1	EB-7-10-30-03	Water	10/30/03	10/30/03	10/30/03	03W5000	mg/L	0.01	<0.01	U
03-5892-2	MW-21-1	Water	10/30/03	10/30/03	10/30/03	03W5000	mg/L	0.01	<0.01	U
03-5892-3	MW-21-2	Water	10/30/03	10/30/03	10/30/03	03W5000	mg/L	0.01	<0.01	U
03-5892-4	MW-21-3	Water	10/30/03	10/30/03	10/30/03	03W5000	mg/L	0.01	<0.01	U
03-5892-5	MW-21-4	Water	10/30/03	10/30/03	10/30/03	03W5000	mg/L	0.01	<0.01	U
03-5892-6	MW-21-5	Water	10/30/03	10/30/03	10/30/03	03W5000	mg/L	0.01	<0.01	U
03W5000-MB-01	03W5000-MB-01	Water	10/30/03	10/30/03	10/30/03	03W5000	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

*Handwritten signature*

Applied P & Ch Laboratory  
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc.  
Project ID: JPL

Project No: 04-4428.10  
Service ID: 35892

Anal. Method: 314.0  
Collected by: JR

Component Name: Perchlorate  
CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-5892-1	EB-7-10-30-03	Water	10/30/03	10/30/03	11/04/03	03W5052	µg/L	4	<4	U
03-5892-2	MW-21-1	Water	10/30/03	10/30/03	11/04/03	03W5052	µg/L	4	6.5	
03-5892-3	MW-21-2	Water	10/30/03	10/30/03	11/04/03	03W5052	µg/L	4	2.7	B
03-5892-4	MW-21-3	Water	10/30/03	10/30/03	11/04/03	03W5052	µg/L	4	3.6	B
03-5892-5	MW-21-4	Water	10/30/03	10/30/03	11/04/03	03W5052	µg/L	4	3.4	B
03-5892-6	MW-21-5	Water	10/30/03	10/30/03	11/04/03	03W5052	µg/L	4	2.6	B
03W5052-MB-01	03W5052-MB-01	Water	11/04/03	11/04/03	11/04/03	03W5052	µg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

*1/2/04*



LDC #: 11272A6

### VALIDATION COMPLETENESS WORKSHEET

Date: 12/17/03

SDG #: 03-5892

Level III

Page: (of)

Laboratory: Applied P & Ch Laboratory

Reviewer: MW

2nd Reviewer: A

**METHOD: (Analyte) Hexavalent Chromium (EPA SW846 Method 7196A), Perchlorate (EPA Method 314.0)**

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 10/30/03
IIa.	Initial calibration	SW	
IIb.	Calibration verification	A	
III.	Blanks	A	
IV	Matrix Spike/Matrix Spike Duplicates	A	3 MS/MSD
V	Duplicates	A	
VI.	Laboratory control samples	A	LS/LSD
VII.	Sample result verification	N	
VIII.	Overall assessment of data	A	
IX.	Field duplicates	N	
X	Field blanks	ND	EB = 1

Note: A = Acceptable  
 N = Not provided/applicable  
 SW = See worksheet

ND = No compounds detected  
 R = Rinsate  
 FB = Field blank  
 D = Duplicate  
 TB = Trip blank  
 EB = Equipment blank

Validated Samples:

1	EB-7-10-30-03	11		21		31	
2	MW-21-1	12		22		32	
3	MW-21-2	13		23		33	
4	MW-21-3	14		24		34	
5	MW-21-4	15		25		35	
6	MW-21-5	16		26		36	
7	MW-21-4MS	17		27		37	
8	MW-21-4MSD	18		28		38	
9	MS	19		29		39	
10		20		30		40	

Notes: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

LDC #: 11272A6  
 SDG #: 03-5892

**VALIDATION FINDINGS WORKSHEET**  
**Sample Specific Analysis Reference**

Page: 1 of 1  
 Reviewer: MI  
 2nd reviewer: [Signature]

All circled methods are applicable to each sample.

Sample ID	Parameter
1-6	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC <u>CR<sup>0+</sup></u> <u>(u04)</u>
27.8	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC <u>CR<sup>0+</sup></u> <u>(u04)</u>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>0+</sup>

Comments: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

**VALIDATION FINDINGS WORKSHEET**  
Calibration

LDC #: 1127286  
SDG #: 03-5872

METHOD: Inorganics, EPA Method see com

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".  
Were all instruments calibrated daily, each set-up time, and were the proper number of standards used?  
Were all initial and continuing calibration verification percent recoveries (%R) within the control limits of 90-110%?  
Are all correlation coefficients  $\geq 0.995$ ?

**LEVEL I/II ONLY:**  
Were recalculated results acceptable? See Level IV Initial and Continuing Calibration Recalculation Worksheet for recalculations.  
Was a balance check conducted prior to the TDS analysis?  
Was the titrant normality checked?

#	Date	Calibration ID	Analyte	%R	Associated Samples	Qualifications
	11/4/03	CCV	ce-04	115	7.8	Justo / P

Comments:

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** NASA JPL  
**Collection Date:** November 3, 2003  
**LDC Report Date:** December 15, 2003  
**Matrix:** Water  
**Parameters:** Wet Chemistry  
**Validation Level:** EPA Level III  
**Laboratory:** Applied P & Ch Laboratory

**Sample Delivery Group (SDG):** 03-5923

**Sample Identification**

EB-8-11-3-03  
MW-3-3  
MW-3-4  
MW-3-5  
MW-3-3MS  
MW-3-3MSD

## Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196A for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## II. Calibration

### a. Initial Calibration

All criteria for the initial calibration of each method were met.

### b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable with the following exceptions:

Date	Lab. Reference/ID	Analyte	%R (Limits)	Associated Samples	Flag	A or P
11/4/03	CCV	Perchlorate	115 (90-110)	All samples in SDG 03-5923	J (all detects)	P

## III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

## IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

## VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## VII. Sample Result Verification

Raw data were not reviewed for this SDG.

### **VIII. Overall Assessment of Data**

Data flags are summarized at the end of this report.

### **IX. Field Duplicates**

No field duplicates were identified in this SDG.

### **X. Field Blanks**

Sample EB-8-11-3-03 was identified as an equipment blank. No contaminant concentrations were found in this blank.

**NASA JPL**  
**Wet Chemistry - Data Qualification Summary - SDG 03-5923**

<b>SDG</b>	<b>Sample</b>	<b>Analyte</b>	<b>Flag</b>	<b>A or P</b>	<b>Reason</b>
03-5923	EB-8-11-3-03 MW-3-3 MW-3-4 MW-3-5	Perchlorate	J (all detects)	P	Calibration verification (%R)

**NASA JPL**  
**Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-5923**

No Sample Data Qualified in this SDG

**NASA JPL**  
**Wet Chemistry - Field Blank Data Qualification Summary - SDG 03-5923**

No Sample Data Qualified in this SDG



B

Applied P & Ch Laboratory  
**Wet Analysis Results for Method 7196**

Client Name: GEOFON, Inc.  
Project ID: JPL

Project No: 04-4428.10  
Service ID: 35923

Anal. Method 7196  
Collected by: JR

Component Name: Chromium (VI)  
CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-5923-1	EB-8-11-3-03	Water	11/03/03	11/03/03	11/03/03	03W5037	mg/L	0.01	<0.01	U
03-5923-2	MW-3-3	Water	11/03/03	11/03/03	11/03/03	03W5037	mg/L	0.01	<0.01	U
03-5923-3	MW-3-4	Water	11/03/03	11/03/03	11/03/03	03W5037	mg/L	0.01	<0.01	U
03-5923-4	MW-3-5	Water	11/03/03	11/03/03	11/03/03	03W5037	mg/L	0.01	<0.01	U
03W5037-MB-01	03W5037-MB-01	Water	11/03/03	11/03/03	11/03/03	03W5037	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

*Handwritten signature*

Applied P & Ch Laboratory  
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc.  
Project ID: JPL

Project No: 04-4428.10  
Service ID: 35923

Anal. Method: 314.0  
Collected by: JR

Component Name: Perchlorate  
CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-5923-1	EB-8-11-3-03	Water	11/03/03	11/03/03	11/04/03	03W5052	µg/L	4	<4	U
03-5923-2	MW-3-3	Water	11/03/03	11/03/03	11/04/03	03W5052	µg/L	4	<4	U
03-5923-3	MW-3-4	Water	11/03/03	11/03/03	11/04/03	03W5052	µg/L	4	<4	U
03-5923-4	MW-3-5	Water	11/03/03	11/03/03	11/04/03	03W5052	µg/L	4	<4	U
03W5052-MB-01	03W5052-MB-01	Water	11/04/03	11/04/03	11/04/03	03W5052	µg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

*12/18*

LDC #: 11272B6

**VALIDATION COMPLETENESS WORKSHEET**

Date: 12/1/03

SDG #: 03-5923

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: MS

2nd Reviewer: /

**METHOD: (Analyte) Perchlorate (EPA Method 314.0), Hexavalent Chromium (EPA SW846 Method 7196A)**

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 11/3/03
IIa.	Initial calibration	A	
IIb.	Calibration verification	SW	
III.	Blanks	A	
IV	Matrix Spike/Matrix Spike Duplicates	A	3 MS / MSD
V	Duplicates	A	
VI.	Laboratory control samples	A	1 MS / MSD
VII.	Sample result verification	N	
VIII.	Overall assessment of data	A	
IX.	Field duplicates	A	
X	Field blanks	ND	EB = 2

Note: A = Acceptable  
 N = Not provided/applicable  
 SW = See worksheet

ND = No compounds detected  
 R = Rinsate  
 FB = Field blank  
 D = Duplicate  
 TB = Trip blank  
 EB = Equipment blank

Validated Samples:

*MS*

1	EB-8-11-3-03	11		21		31	
2	MW-3-3	12		22		32	
3	MW-3-4	13		23		33	
4	MW-3-5	14		24		34	
5	MW-3-MS	15		25		35	
6	MW-3-MSD	16		26		36	
7	MB	17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

Notes: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

LDC #: 11292B6  
 SDG #: 03-8923

**VALIDATION FINDINGS WORKSHEET**  
**Sample Specific Analysis Reference**

Page: 1 of 1  
 Reviewer: MH  
 2nd reviewer: [Signature]

All circled methods are applicable to each sample.

Sample ID	Parameter
<u>14</u>	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup> <u>(CLOP)</u>
<u>2516</u>	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>

Comments: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

VALIDATION FINDINGS WORKSHEET  
Calibration

Page: ( 1 ) of ( 1 )  
Reviewer: 1/17  
2nd Reviewer: J

LDC #: 11272-86  
SDG #: 03-593

METHOD: Inorganics, EPA Method See below

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".  
 N N/A  
 Y N/A

Were all instruments calibrated daily, each set-up time, and were the proper number of standards used?  
 N N/A  
Were all initial and continuing calibration verification percent recoveries (%R) within the control limits of 90-110%?  
 N N/A  
Are all correlation coefficients  $\geq 0.995$ ?  
 N N/A

LEVEL IV/D ONLY:  
 Y N N/A  
 Y N N/A  
 Y N N/A  
Were recalculated results acceptable? See Level IV Initial and Continuing Calibration Recalculation Worksheet for recalculations.  
Was a balance check conducted prior to the TDS analysis?  
Was the titrant normality checked?

#	Date	Calibration ID	Analyte	%R	Associated Samples	Qualifications
1	11/4/03	cev	ce04	115	A1	J data / p

Comments: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** NASA JPL  
**Collection Date:** October 29, 2003  
**LDC Report Date:** December 15, 2003  
**Matrix:** Water  
**Parameters:** Wet Chemistry  
**Validation Level:** EPA Level III  
**Laboratory:** Applied P & Ch Laboratory

**Sample Delivery Group (SDG):** 03-5874

**Sample Identification**

EB-6-10-29-03  
MW-18-1  
MW-18-2  
MW-18-3  
MW-18-4  
MW-18-5  
MW-19-1  
MW-19-2  
MW-19-3  
MW-19-4  
MW-19-5  
MW-19-3MS  
MW-19-3MSD

## Introduction

This data review covers 13 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196A for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## II. Calibration

### a. Initial Calibration

All criteria for the initial calibration of each method were met.

### b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable with the following exceptions:

Date	Lab. Reference/ID	Analyte	%R (Limits)	Associated Samples	Flag	A or P
12/9/03	CCV2	Perchlorate	112 (90-110)	MW-18-2 MW-19-3	J (all detects)	P
12/9/03	CCV3	Perchlorate	112 (90-110)	EB-6-10-29-03 MW-18-1 MW-18-3 MW-18-4 MW-18-5 MW-19-1 MW-19-2 MW-19-4 MW-19-5	J (all detects)	P
12/9/03	CCV4	Perchlorate	115 (90-110)	MW-19-3MS MW-19-3MSD	J (all detects)	P
12/9/03	CCV1	Perchlorate	115 (90-110)	MB	J (all detects)	P

## III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

## IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.



## **V. Duplicates**

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

## **VI. Laboratory Control Samples**

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## **VII. Sample Result Verification**

Raw data were not reviewed for this SDG.

## **VIII. Overall Assessment of Data**

Data flags are summarized at the end of this report.

## **IX. Field Duplicates**

No field duplicates were identified in this SDG.

## **X. Field Blanks**

Sample EB-6-10-29-03 was identified as an equipment blank. No contaminant concentrations were found in this blank.

**NASA JPL  
Wet Chemistry - Data Qualification Summary - SDG 03-5874**

<b>SDG</b>	<b>Sample</b>	<b>Analyte</b>	<b>Flag</b>	<b>A or P</b>	<b>Reason</b>
03-5874	EB-6-10-29-03 MW-18-1 MW-18-2 MW-18-3 MW-18-4 MW-18-5 MW-19-1 MW-19-2 MW-19-3 MW-19-4 MW-19-5	Perchlorate	J (all detects)	P	Calibration verification (%R)

**NASA JPL  
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-5874**

No Sample Data Qualified in this SDG

**NASA JPL  
Wet Chemistry - Field Blank Data Qualification Summary - SDG 03-5874**

No Sample Data Qualified in this SDG

BC

# Applied P & Ch Laboratory Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.  
Project ID: JPL

Project No: 04.4428.10  
Service ID: 35874

Anal. Method 7196  
Collected by: Jr

Component Name: Chromium (VI)  
CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-5874-1	EB-6-10-29-03	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-2	MW-18-1	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-3	MW-18-2	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-4	MW-18-3	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-5	MW-18-4	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-6	MW-18-5	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-7	MW-19-1	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-8	MW-19-2	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-9	MW-19-3	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-10	MW-19-4	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03-5874-11	MW-19-5	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U
03W4993-MB-01	03W4993-MB-01	Water	10/29/03	10/29/03	10/29/03	03W4993	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

*Handwritten signature*

Applied P & Ch Laboratory  
**Wet Analysis Results for Method 314.0**

Client Name: GEOFON, Inc.  
 Project ID: JPL

Project No: 04.4428.10  
 Service ID: 35874

Anal. Method: 314.0  
 Collected by: Jr

Component Name: Perchlorate  
 CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-5874-1	EB-6-10-29-03	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	<4	U
03-5874-2	MW-18-1	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	<4	U
03-5874-3	MW-18-2	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	<4	U
03-5874-4	MW-18-3	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	<4	U
03-5874-5	MW-18-4	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	17.2	J
03-5874-6	MW-18-5	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	<4	U
03-5874-7	MW-19-1	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	<4	U
03-5874-8	MW-19-2	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	4.4	J
03-5874-9	MW-19-3	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	5.1	J
03-5874-10	MW-19-4	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	<4	U
03-5874-11	MW-19-5	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	<4	U
03W4987-MB-01	03W4987-MB-01	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

*Handwritten signature/initials*

LDC #: 11272C6

### VALIDATION COMPLETENESS WORKSHEET

Date: 12/12/03

SDG #: 03-5874

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: [Signature]

2nd Reviewer: [Signature]

**METHOD: (Analyte) Perchlorate (EPA Method 314.0), Hexavalent Chromium (EPA SW846 Method 7196A)**

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 10/29/03
IIa.	Initial calibration	A	
IIb.	Calibration verification	SW	
III.	Blanks	A	
IV.	Matrix Spike/Matrix Spike Duplicates	A	3 MS/MSD
V.	Duplicates	A	
VI.	Laboratory control samples	A	LCS/LCSD
VII.	Sample result verification	N	
VIII.	Overall assessment of data	N	
IX.	Field duplicates	N	
X.	Field blanks	ND	EB=1

Note: A = Acceptable  
 N = Not provided/applicable  
 SW = See worksheet

ND = No compounds detected  
 R = Rinsate  
 FB = Field blank  
 D = Duplicate  
 TB = Trip blank  
 EB = Equipment blank

Validated Samples:

1	EB-6-10-29-03	11	MW-19-5	21		31	
2	MW-18-1	12	MW-19-3MS	22		32	
3	MW-18-2	13	MW-19-3MSD	23		33	
4	MW-18-3	14	MB	24		34	
5	MW-18-4	15		25		35	
6	MW-18-5	16		26		36	
7	MW-19-1	17		27		37	
8	MW-19-2	18		28		38	
9	MW-19-3	19		29		39	
10	MW-19-4	20		30		40	

Notes: \_\_\_\_\_

LDC #: 1127206  
SDG #: 035874

VALIDATION FINDINGS WORKSHEET  
Sample Specific Analysis Reference

Page: 1 of 1  
Reviewer: MIT  
2nd reviewer:   

All circled methods are applicable to each sample.

Sample ID	Parameter
<u>1-11</u>	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> <u>CLC</u>
<u>2-12, 13</u>	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> <u>CLC</u>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN <sup>-</sup> NH <sub>3</sub> TKN TOC CR <sup>6+</sup> _____

Comments: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Page: ( of )  
 Reviewer: MY  
 2nd Reviewer: [Signature]

### VALIDATION FINDINGS WORKSHEET

Calibration

LDC #: 112/1206  
 SDG #: 03-8874

METHOD: Inorganics, EPA Method See below

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".  
 N  Y  
 N  Y  
 N  Y  
 N  Y  
 N  Y

**LEVEL I/II ONLY:**  
 Were instruments calibrated daily, each set-up time, and were the proper number of standards used?  
 Were all initial and continuing calibration verification percent recoveries (%R) within the control limits of 90-110%?  
 Are all correlation coefficients  $\geq 0.995$ ?

#	Date	Calibration ID	Analyte	%R	Associated Samples	Qualifications
1	<u>12/09/03</u>	<u>ccv2</u>	<u>ce04</u>	<u>112</u>	<u>3, 9</u>	<u>Jtk/p</u>
2	<u>↓</u>	<u>ccv3</u>	<u>↓</u>	<u>112</u>	<u>1, 2, 4-8, 10, 11</u>	<u>↓</u>
3	<u>↓</u>	<u>ccv4</u>	<u>↓</u>	<u>115</u>	<u>12, 13</u>	<u>↓</u>
4	<u>↓</u>	<u>ccv1</u>	<u>↓</u>	<u>115</u>	<u>MB</u>	<u>↓</u>

Comments: \_\_\_\_\_

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** NASA JPL  
**Collection Date:** October 28, 2003  
**LDC Report Date:** December 15, 2003  
**Matrix:** Water  
**Parameters:** Wet Chemistry  
**Validation Level:** EPA Level III & IV  
**Laboratory:** Applied P & Ch Laboratory  
**Sample Delivery Group (SDG):** 03-5852

**Sample Identification**

Dupe-1-4Q03\*\*  
EB-5-10-28-03  
MW-24-1  
MW-24-2  
MW-24-3  
MW-24-4  
MW-24-5  
Dupe-1-4Q03MS  
Dupe-1-4Q03MSD

\*\*Indicates sample underwent EPA Level IV review



## Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196A for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## II. Calibration

### a. Initial Calibration

All criteria for the initial calibration of each method were met.

### b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable with the following exceptions:

Date	Lab. Reference/ID	Analyte	%R (Limits)	Associated Samples	Flag	A or P
10/29/03	CCV1	Perchlorate	115 (90-110)	Dupe-1-4Q03** EB-5-10-28-03 MW-24-3 MB	J (all detects)	P
10/29/03	CCV2	Perchlorate	112 (90-110)	MW-24-1 MW-24-2 MW-24-4 MW-24-5	J (all detects)	P

## III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

## IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

## **VI. Laboratory Control Samples**

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## **VII. Sample Result Verification**

All sample result verifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

## **VIII. Overall Assessment of Data**

Data flags are summarized at the end of this report.

## **IX. Field Duplicates**

Samples Dupe-1-4Q03\*\* and MW-24-4 were identified as field duplicates. No contaminant concentrations were detected in any of the samples.

## **X. Field Blanks**

Sample EB-5-10-28-03 was identified as an equipment blank. No contaminant concentrations were found in this blank.

**NASA JPL  
Wet Chemistry - Data Qualification Summary - SDG 03-5852**

SDG	Sample	Analyte	Flag	A or P	Reason
03-5852	Dupe-1-4Q03** EB-5-10-28-03 MW-24-1 MW-24-2 MW-24-3 MW-24-4 MW-24-5	Perchlorate	J (all detects)	P	Calibration verification (%R)

**NASA JPL  
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-5852**

No Sample Data Qualified in this SDG

**NASA JPL  
Wet Chemistry - Field Blank Data Qualification Summary - SDG 03-5852**

No Sample Data Qualified in this SDG

b

Applied P & Ch Laboratory  
**Wet Analysis Results for Method 7196**

Client Name: GEOFON, Inc.  
Project ID: JPL

Project No: 04.4428.10  
Service ID: 35852

Anal. Method 7196  
Collected by: Jr

Component Name: Chromium (VI)  
CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-5852-1	DUPE-1-4Q03	Water	10/28/03	10/28/03	10/28/03	03W4979	mg/L	0.01	<0.01	U
03-5852-2	EB-5-10-28-03	Water	10/28/03	10/28/03	10/28/03	03W4979	mg/L	0.01	<0.01	U
03-5852-3	MW-24-1	Water	10/28/03	10/28/03	10/28/03	03W4979	mg/L	0.01	<0.01	U
03-5852-4	MW-24-2	Water	10/28/03	10/28/03	10/28/03	03W4979	mg/L	0.01	<0.01	U
03-5852-5	MW-24-3	Water	10/28/03	10/28/03	10/28/03	03W4979	mg/L	0.01	<0.01	U
03-5852-6	MW-24-4	Water	10/28/03	10/28/03	10/28/03	03W4979	mg/L	0.01	<0.01	U
03-5852-7	MW-24-5	Water	10/28/03	10/28/03	10/28/03	03W4979	mg/L	0.01	<0.01	U
03W4979-MB-01	03W4979-MB-01	Water	10/28/03	10/28/03	10/28/03	03W4979	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

*Handwritten signature*

Applied P & Ch Laboratory  
**Wet Analysis Results for Method 314.0**

Client Name: GEOFON, Inc.                      Project No: 04.4428.10                      Anal. Method 314.0  
 Project ID: JPL                                      Service ID: 35852                              Collected by: Jr

Component Name: Perchlorate  
 CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-5852-1	DUPE-1-4Q03	Water	10/28/03	10/28/03	10/29/03	03W4987	µg/L	4	<4	U
03-5852-2	EB-5-10-28-03	Water	10/28/03	10/28/03	10/29/03	03W4987	µg/L	4	<4	U
03-5852-3	MW-24-1	Water	10/28/03	10/28/03	10/29/03	03W4987	µg/L	200	2760	U
03-5852-4	MW-24-2	Water	10/28/03	10/28/03	10/29/03	03W4987	µg/L	20	155	U
03-5852-5	MW-24-3	Water	10/28/03	10/28/03	10/29/03	03W4987	µg/L	4	<4	U
03-5852-6	MW-24-4	Water	10/28/03	10/28/03	10/29/03	03W4987	µg/L	4	<4	U
03-5852-7	MW-24-5	Water	10/28/03	10/28/03	10/29/03	03W4987	µg/L	4	<4	U
03W4987-MB-01	03W4987-MB-01	Water	10/29/03	10/29/03	10/29/03	03W4987	µg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

LDC #: 11272D6

## VALIDATION COMPLETENESS WORKSHEET

Date: 12/21/03

SDG #: 03-5852

Level III/IV

Page: 1 of 1

Laboratory: Applied P &amp; Ch Laboratory

Reviewer: [Signature]

2nd Reviewer: [Signature]

**METHOD: (Analyte) Perchlorate (EPA Method 314.0), Hexavalent Chromium (EPA SW846 Method 7196A)**

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 10/28/03
IIa.	Initial calibration	A	
IIb.	Calibration verification	SW	
III.	Blanks	A	
IV	Matrix Spike/Matrix Spike Duplicates	A	MS/MSD
V	Duplicates	A	
VI.	Laboratory control samples	A	LES/LESD
VII.	Sample result verification	A	Not reviewed for Level III validation.
VIII.	Overall assessment of data	A	
IX.	Field duplicates	ND	(1, 6)
X	Field blanks	ND	EB = 2

Note: A = Acceptable  
N = Not provided/applicable  
SW = See worksheet

ND = No compounds detected  
R = Rinsate  
FB = Field blank

D = Duplicate  
TB = Trip blank  
EB = Equipment blank

Validated Samples: \*\* Indicates sample underwent Level IV validation.

1	✓ Dupe-1-4Q03**	11		21		31	
2	EB-5-10-28-03	12		22		32	
3	MW-24-1	13		23		33	
4	MW-24-2	14		24		34	
5	MW-24-3	15		25		35	
6	✓ MW-24-4	16		26		36	
7	MW-24-5	17		27		37	
8	Dupe-1-4Q03MS	18		28		38	
9	Dupe-1-4Q03MSD	19		29		39	
10	MB	20		30		40	

Notes: \_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

LDC #: 1127206  
 SDG #: 03-5852

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2  
 Reviewer: MJ  
 2nd Reviewer: [Signature]

Method: Inorganics (EPA Method *See copy*)

Validation Area	Yes	No	NA	Findings/Comments
<b>i. Technical holding times</b>				
All technical holding times were met.	/			
Cooler temperature criteria was met.	/			
<b>ii. Calibration</b>				
Were all instruments calibrated daily, each set-up time?	/			
Were the proper number of standards used?	/			
Were all initial calibration correlation coefficients $\geq 0.995$ ?	/			
Were all initial and continuing calibration verification %Rs within the 90-110% QC limits?		/		
Were titrant checks performed as required?			/	
Were balance checks performed as required?			/	
Were all initial and continuing calibration verification %Rs within the 90-110% QC limits?	/			
<b>iii. Blanks</b>				
Was a method blank associated with every sample in this SDG?	/			
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.		/		
<b>iv. Matrix spike/Matrix spike duplicates and Duplicates</b>				
Were a matrix spike (MS) and duplicate (DUP) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD or MS/DUP. Soil / Water.	/			
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the 75-125 QC limits? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken.	/			
Were the MS/MSD or duplicate relative percent differences (RPD) $\leq 20\%$ for waters and $\leq 35\%$ for soil samples? A control limit of $\leq \text{CRDL} (\leq 2X \text{ CRDL for soil})$ was used for samples that were $\leq 5X$ the CRDL, including when only one of the duplicate sample values were $\leq 5X$ the CRDL.	/			
<b>v. Laboratory control samples</b>				
Was an LCS analyzed for this SDG?	/			
Was an LCS analyzed per extraction batch?	/			
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the 80-120% (85-115% for Method 300.0) QC limits?	/			
<b>vi. Regional Quality Assurance and Quality Control</b>				
Were performance evaluation (PE) samples performed?			/	
Were the performance evaluation (PE) samples within the acceptance limits?			/	



LDC #: 1127-066  
 SDG #: 03-6852

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2  
 Reviewer: [Signature]  
 2nd Reviewer: [Signature]

Validation Area	Yes	No	NA	Findings/Comments
<b>VII. Sample Result Verification</b>				
Were RLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were detection limits < RL?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>VIII. Overall assessment of data</b>				
Overall assessment of data was found to be acceptable.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>IX. Field duplicates</b>				
Field duplicate pairs were identified in this SDG.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Target analytes were detected in the field duplicates.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>X. Field blanks</b>				
Field blanks were identified in this SDG.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Target analytes were detected in the field blanks.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	

LDC #: 1129206  
 SDG #: 03-5852

**VALIDATION FINDINGS WORKSHEET**  
**Sample Specific Analysis Reference**

Page: 1 of 1  
 Reviewer: MT  
 2nd reviewer: [Signature]

All circled methods are applicable to each sample.

Sample ID	Parameter
1-1	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup> (CR <sup>6+</sup> )
8,9	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup> (CR <sup>6+</sup> )
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>
	pH TDS Cl F NO <sub>3</sub> NO <sub>2</sub> SO <sub>4</sub> PO <sub>4</sub> ALK CN' NH <sub>3</sub> TKN TOC CR <sup>6+</sup>

Comments: \_\_\_\_\_  
 \_\_\_\_\_  
 \_\_\_\_\_

LDC #: 112206  
 SDG #: 03-882

**VALIDATION FINDINGS WORKSHEET**  
**Calibration**

Page: 1 of 1  
 Reviewer: MR  
 2nd Reviewer: A

METHOD: Inorganics, EPA Method See Com

- Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
- N N/A Were all instruments calibrated daily, each set-up time, and were the proper number of standards used?
  - N N/A Were all initial and continuing calibration verification percent recoveries (%R) within the control limits of 90-110%?
  - N N/A Are all correlation coefficients  $\geq 0.995$ ?
- LEVEL IV/D ONLY:
- N N/A Were recalculated results acceptable? See Level IV Initial and Continuing Calibration Recalculation Worksheet for recalculations.
  - N N/A Was a balance check conducted prior to the TDS analysis.?
  - N N/A Was the titrant normality checked?

#	Date	Calibration ID	Analyte	%R	Associated Samples	Qualifications
1	10/29/03	ccv1	clay	115	1, 2, 5, MB, <del>7, 8</del>	J <del>W</del> / P
2	↓	ccv2	↓	112	3, 4, 6, 7	↓

Comments:

LDC #: 12206  
 SDG #: 03882

**VALIDATION FINDINGS WORKSHEET**  
**Initial and Continuing Calibration Calculation Verification**

Page: 1 of 1  
 Reviewer: MLT  
 2nd Reviewer: A

METHOD: Inorganics, Method see cond

The correlation coefficient (r) for the calibration of Cr6+ was recalculated. Calibration date: 9/28/03

An initial or continuing calibration verification percent recovery (%R) was recalculated for each type of analysis using the following formula:

$\%R = \frac{\text{Found} \times 100}{\text{True}}$  Where, Found = concentration of each analyte measured in the analysis of the ICV or CCV solution  
 True = concentration of each analyte in the ICV or CCV source

Type of Analysis	Analyte	Conc. (mg/L)	Area (units)	Recalculated		Reported		Acceptable (Y/N)
				r	%R	r	%R	
Initial calibration		0	0					
Calibration verification	Standard 1	0.0125	0.007					
	Standard 2	0.025	0.017					
	Standard 3	0.125	0.107					
	Standard 4	0.250	0.212					
	Standard 5	0.50	0.424					
	Standard 6							
	Standard 7							
Calibration verification	Cr6+	59.3		115	115	115	115	Y
Calibration verification	Cr6+	0.235		94	94	94	94	Y
Calibration verification								

Comments: Refer to Calibration Verification findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 1121266  
 SDG #: 03-5852

**VALIDATION FINDINGS WORKSHEET**  
**Level IV Recalculation Worksheet**

Page: 1 of 1  
 Reviewer: MLH  
 2nd Reviewer: [Signature]

METHOD: Inorganics, Method See cover

Percent recoveries (%R) for a laboratory control sample and a matrix spike sample were recalculated using the following formula:

$\%R = \frac{\text{Found}}{\text{True}} \times 100$  Where, Found = concentration of each analyte measured in the analysis of the sample. For the matrix spike calculation, Found = SSR (spiked sample result) - SR (sample result).  
 True = concentration of each analyte in the source.

A sample and duplicate relative percent difference (RPD) was recalculated using the following formula:

$RPD = \frac{|S-D|}{(S+D)/2} \times 100$  Where, S = Original sample concentration  
 D = Duplicate sample concentration

Sample ID	Type of Analysis	Element	Found / S (units)	True / D (units)	Recalculated		Reported		Acceptable (Y/N)
					%R / RPD	%R / RPD			
107	Laboratory control sample	CO4	25.6	25	114	114	114	114	Y
8	Matrix spike sample	W6+	(SSR-SR) 0.222	0.25	89	89	89	89	Y
819	Duplicate sample	W6+	0.222	0.222	1	1	1	1	Y

Comments: Refer to appropriate worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 1/2/2006  
 SDG #: 03-8882

# VALIDATION FINDINGS WORKSHEET

## Sample Calculation Verification

Page: 1 of 1  
 Reviewer: MIT  
 2nd reviewer: [Signature]

METHOD: Inorganics, Method see code

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".


- Y  N  N/A Have results been reported and calculated correctly?
- Y  N  N/A Are results within the calibrated range of the instruments?
- Y  N  N/A Are all detection limits below the CRQL?

Compound (analyte) results for 1 reported with a positive detect were recalculated and verified using the following equation: ND

Concentration = \_\_\_\_\_ Recalculation: \_\_\_\_\_

#	Sample ID	Analyte	Reported Concentration ( )	Calculated Concentration ( )	Acceptable (Y/N)

Note: \_\_\_\_\_  
 \_\_\_\_\_



**LABORATORY DATA CONSULTANTS, INC.**

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

Geofon, Inc.  
22632 Golden Springs Drive, Suite 270  
Diamond Bar, CA 91765  
ATTN: Mr. Tony Ford

December 24, 2003

SUBJECT: NASA JPL, DO #01, Data Validation

Dear Mr. Ford,

Enclosed are the final validation reports for the fractions listed below. These SDGs were received on December 15, 2003. Attachment 1 is a summary of the samples that were reviewed for each analysis.

**LDC Project # 11280:**

<b><u>SDG #</u></b>	<b><u>Fraction</u></b>
03-5951, 03-5973, 03-6002, 03-6034, 03-6047	Volatiles, Wet Chemistry

The data validation was performed under EPA Level III and Level IV guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, February 1994
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996

Please feel free to contact us if you have any questions.

Sincerely,

Erlinda T. Rauto  
Operations Manager/Senior Chemist

**LDC #11280 (Geofon-Diamond Bar / NASA JPL, DO#0001)**

LDC	SDG#	DATE RECD	DATE DUE	VOA (524.2)		Cr(VI) (7196A)			CLO <sub>4</sub> (314.0)			S		W		S		W		S		W		S		W		S		W		S		W			
				W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S		
Matrix: Water/Soil																																					
A	03-5951	12-15-03	1-6-04	12	0	13	0	11	0																												
A	03-5951	12-15-03	1-6-04	1	0	1	0	1	0																												
B	03-5973	12-15-03	1-6-04	6	0	5	0	5	0																												
C	03-6002	12-15-03	1-6-04	6	0	5	0	5	0																												
C	03-6002	12-15-03	1-6-04	1	0	1	0	1	0																												
D	03-6034	12-15-03	1-6-04	6	0	5	0	5	0																												
D	03-6034	12-15-03	1-6-04	1	0	1	0	1	0																												
E	03-6047	12-15-03	1-6-04	4	0	5	0	3	0																												
				37	0	36	0	32	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	0	
Total				TH																																	



**NASA JPL  
Data Validation Reports  
LDC# 11280**

Volatiles

DDC

**Laboratory Data Consultants, Inc.  
Data Validation Report**

**Project/Site Name:** NASA JPL  
**Collection Date:** November 4, 2003  
**LDC Report Date:** December 22, 2003  
**Matrix:** Water  
**Parameters:** Volatiles  
**Validation Level:** EPA Level III & IV  
**Laboratory:** Applied P & Ch Laboratory  
**Sample Delivery Group (SDG):** 03-5951

**Sample Identification**

DUPE-5-4-Q03\*\*  
EB-9-11-4-03  
MW-14-1  
MW-14-2  
MW-14-3  
MW-14-4  
MW-14-5  
MW-17-1  
MW-17-2  
MW-17-3  
MW-17-4  
MW-17-5  
TB-9-11-4-03

\*\*Indicates sample underwent EPA Level IV review

## Introduction

This data review covers 13 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

## I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

## II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

## III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination ( $r^2$ ) was greater than or equal to 0.990 .

## IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
11/6/03 (10:27)	Chloromethane Bromomethane 2,2-Dichloropropane	30.98 44.52 32.14	All samples in SDG 03-5951	J (all detects) UJ (all non-detects)	P

## V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

## VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

## **VII. Matrix Spike/Matrix Spike Duplicates**

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

## **VIII. Laboratory Control Samples (LCS)**

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

## **IX. Regional Quality Assurance and Quality Control**

Not applicable.

## **X. Internal Standards**

All internal standard areas and retention times were within QC limits.

## **XI. Target Compound Identifications**

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

## **XII. Compound Quantitation and CRQLs**

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

## **XIII. Tentatively Identified Compounds (TICs)**

Tentatively identified compounds were not reported by the laboratory.

## **XIV. System Performance**

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

## **XV. Overall Assessment of Data**

Data flags have been summarized at the end of the report.

## XVI. Field Duplicates

Samples DUPE-5-4-Q03\*\* and MW-17-3 were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/L)		RPD
	DUPE-5-4-Q03**	MW-17-3	
Carbon tetrachloride	13.7	11.0	22
Chloroform	3.1	2.6	18
Tetrachloroethene	0.6	0.4	40
Trichloroethene	3.8	3.1	20

## XVII. Field Blanks

Sample TB-9-11-4-03 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Compound	Concentration (ug/L)
TB-9-11-4-03	Methylene chloride	0.4

Sample EB-9-11-4-03 was identified as an equipment blank. No volatile contaminants were found in this blank.

**NASA JPL**  
**Volatiles - Data Qualification Summary - SDG 03-5951**

SDG	Sample	Compound	Flag	A or P	Reason
03-5951	DUPE-5-4-Q03** EB-9-11-4-03 MW-14-1 MW-14-2 MW-14-3 MW-14-4 MW-14-5 MW-17-1 MW-17-2 MW-17-3 MW-17-4 MW-17-5 TB-9-11-4-03	Chloromethane Bromomethane 2,2-Dichloropropane	J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

**NASA JPL**  
**Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-5951**

No Sample Data Qualified in this SDG

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: <b>DUPE-5-4-Q03</b>	Lab Sample ID: 03-5951-1	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-01	Prep. No: -	Anal. Time: 16:30
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>UJ</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	13.7	
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	3.1	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>UJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U <i>UJ</i>
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U



#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	0.6	
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	3.8	
56	TRICHLOROFUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U

**Surrogates**

		Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	108
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	108
3	DIBROMOFUOROMETHANE	1868-53-7	70-122	107
4	TOLUENE-D8	2037-26-5	73-129	108
# of out-of-control			0	

**Internal Standard**

		Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4	50-200	96
2	1,4-DICHLOROETHANE-D4	3855-82-1	50-200	92
3	FLUOROBENZENE	462-06-6	50-200	99
# of out-of-control			0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL  
 J - Less than RL (PQL, EQ, or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)  
 E - Exceed calibration range  
 B - A positive value was found in the method blank  
 D - Diluted

*1272*

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: <b>EB-9-11-4-03</b>	Lab Sample ID: 03-5951-2	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-02	Prep. No: -	Anal. Time: 16:57
Methanol Vol. -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>W</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>W</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROETHANE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROETHANE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROETHANE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U <i>W</i>
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	< 0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	< 0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U
<b>Surrogates</b>				<b>Control Limit, %</b>	<b>Surro. Rec.%</b>	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	108	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	108	
3	DIBROMOFLUOROMETHANE	1868-53-7		70-122	107	
4	TOLUENE-D8	2037-26-5		73-129	109	
# of out-of-control					0	
<b>Internal Standard</b>				<b>Control Limit, %</b>	<b>IS Rec.%</b>	
1	CHLOROBENZENE-D5	3114-55-4		50-200	96	
2	1,4-DICHLOROETHANE-D4	3855-82-1		50-200	94	
3	FLUOROBENZENE	462-06-6		50-200	100	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

*1022*

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: MW-14-1	Lab Sample ID: 03-5951-3	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-03	Prep. No: -	Anal. Time: 17:23
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>UJ</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>UJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	0.5	J
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U <i>UJ</i>
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	0.4	J
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2,2-TETRACHLOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
<b>Surrogates</b>				<b>Control Limit, %</b>	<b>Surro. Rec.%</b>	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	108	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	108	
3	DIBROMOFLUOROMETHANE	1868-53-7		70-122	106	
4	TOLUENE-D8	2037-26-5		73-129	108	
# of out-of-control					0	
<b>Internal Standard</b>				<b>Control Limit, %</b>	<b>IS Rec.%</b>	
1	CHLOROBENZENE-D5	3114-55-4		50-200	94	
2	1,4-DICHLOROETHANE-D4	3855-82-1		50-200	90	
3	FLUOROBENZENE	462-06-6		50-200	97	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

1122

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: MW-14-2	Lab Sample ID: 03-5951-4	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-04	Prep. No: -	Anal. Time: 17:49
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	UWJ
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	0.5	J
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	UWJ
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	UW
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	0.7	
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	4.6	
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2,2-TRICHLORO-1,2,2-TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
<b>Surrogates</b>				<b>Control Limit, %</b>	<b>Surro. Rec.%</b>	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	110	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	106	
3	DIBROMOFLUOROMETHANE	1868-53-7		70-122	108	
4	TOLUENE-D8	2037-26-5		73-129	107	
# of out-of-control					0	
<b>Internal Standard</b>				<b>Control Limit, %</b>	<b>IS Rec.%</b>	
1	CHLOROBENZENE-D5	3114-55-4		50-200	96	
2	1,4-DICHLOROBENZENE-D4	3855-82-1		50-200	91	
3	FLUOROBENZENE	462-06-6		50-200	99	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: MW-14-3	Lab Sample ID: 03-5951-5	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-05	Prep. No: -	Anal. Time: 18:15
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U UJ
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	0.4	J
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U UJ
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U UJ
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

*Handwritten signature*



#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	0.6	
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	0.8	
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	1,1,2,2-TRICHLORO-1,1,2,2-TRIFLUOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U

## Surrogates

		Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	109
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	109
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	107
4	TOLUENE-D8	2037-26-5	73-129	109
# of out-of-control			0	

## Internal Standard

		Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4	50-200	94
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	92
3	FLUOROBENZENE	462-06-6	50-200	98
# of out-of-control			0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

*Handwritten signature*

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: MW-14-4	Lab Sample ID: 03-5951-6	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-06	Prep. No: -	Anal. Time: 18:41
Methanol Vol. -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	< 0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	< 0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	< 0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	< 0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	< 0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	< 0.5	U U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	< 0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	< 0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	< 0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	< 10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	< 0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	< 0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	< 0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	< 0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	< 0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	< 0.5	U U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	< 0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	< 0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	< 0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	< 0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	< 0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	< 0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	< 0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	< 0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	< 0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	< 0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	< 0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	< 0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	< 0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	< 0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	< 0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	< 0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	< 0.5	U U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	< 0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	< 0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	< 0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	< 0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	< 0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	< 0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	< 0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	< 0.5	U
56	TRICHLOROFUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	1,1,2,2-TETRACHLOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U
<b>Surrogates</b>				<b>Control Limit, %</b>	<b>Surro. Rec.%</b>	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	109	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	109	
3	DIBROMOFUOROMETHANE	1868-53-7		70-122	108	
4	TOLUENE-D8	2037-26-5		73-129	109	
# of out-of-control					0	
<b>Internal Standard</b>				<b>Control Limit, %</b>	<b>IS Rec.%</b>	
1	CHLOROBENZENE-D5	3114-55-4		50-200	93	
2	1,4-DICHLOROETHANE-D4	3855-82-1		50-200	90	
3	FLUOROBENZENE	462-06-6		50-200	96	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

*1202*

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: MW-14-5	Lab Sample ID: 03-5951-7	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-07	Prep. No: -	Anal. Time: 19:07
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>WT</i>
7	N-BUTYL BENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYL BENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYL BENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>WT</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U <i>WT</i>
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYL BENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYL BENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

*M202*

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

## Surrogates

		Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	110
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	108
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	108
4	TOLUENE-D8	2037-26-5	73-129	107
# of out-of-control			0	

## Internal Standard

		Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4	50-200	94
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	91
3	FLUOROBENZENE	462-06-6	50-200	98
# of out-of-control			0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: MW-17-1	Lab Sample ID: 03-5951-8	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-08	Prep. No: -	Anal. Time: 19:33
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>UJ</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>UJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U <i>UJ</i>
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

*N200*

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	< 0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	< 0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U

Surrogates			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	108
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	109
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	109
4	TOLUENE-D8	2037-26-5	73-129	108
# of out-of-control				0

Internal Standard			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	94
2	1,4-DICHLOROETHANE-D4	3855-82-1	50-200	91
3	FLUOROBENZENE	462-06-6	50-200	97
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

*Handwritten signature*

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: MW-17-2	Lab Sample ID: 03-5951-9	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-09	Prep. No: -	Anal. Time: 19:59
Methanol Vol. -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>WJ</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	1.0	
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	1.1	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>WJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROETHENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROETHENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROETHENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U <i>WJ</i>
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U



#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	0.4	J
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	6.2	
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2-TRICHLORO-1,2,2-TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

## Surrogates

			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	109
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	111
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	107
4	TOLUENE-D8	2037-26-5	73-129	108
# of out-of-control				0

## Internal Standard

			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	95
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	92
3	FLUOROBENZENE	462-06-6	50-200	98
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

*M/202*

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: MW-17-3	Lab Sample ID: 03-5951-10	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-10	Prep. No: -	Anal. Time: 20:25
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>UJ</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	11.0	
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	2.6	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>UJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U <i>UJ</i>
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

*1000*

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	0.4	J
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	3.1	
56	TRICHLOROFUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U
<b>Surrogates</b>				<b>Control Limit, %</b>	<b>Surro. Rec.%</b>	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	107	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	111	
3	DIBROMOFUOROMETHANE	1868-53-7		70-122	108	
4	TOLUENE-D8	2037-26-5		73-129	105	
# of out-of-control					0	
<b>Internal Standard</b>				<b>Control Limit, %</b>	<b>IS Rec.%</b>	
1	CHLOROBENZENE-D5	3114-55-4		50-200	96	
2	1,4-DICHLOROETHANE-D4	3855-82-1		50-200	94	
3	FLUOROBENZENE	462-06-6		50-200	98	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

*MBO*

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: MW-17-4	Lab Sample ID: 03-5951-11	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-11	Prep. No: -	Anal. Time: 20:51
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	< 0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	< 0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	< 0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	< 0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	< 0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	< 0.5	UWJ
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	< 0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	< 0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	< 0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	< 10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	< 0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	< 0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	< 0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	< 0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	< 0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	< 0.5	UWJ
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	< 0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	< 0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	< 0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	< 0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	< 0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	< 0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	< 0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	< 0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	< 0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	< 0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	< 0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	< 0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	< 0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	< 0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	< 0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	< 0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	< 0.5	UWJ
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	< 0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	< 0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	< 0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	< 0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	< 0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	< 0.5	U

*MJD*  
1639

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	0.8	
56	TRICHLOROFUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2,2-TETRACHLOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
<b>Surrogates</b>				<b>Control Limit, %</b>	<b>Surro. Rec.%</b>	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	109	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	110	
3	DIBROMOFUOROMETHANE	1868-53-7		70-122	109	
4	TOLUENE-D8	2037-26-5		73-129	107	
# of out-of-control					0	
<b>Internal Standard</b>				<b>Control Limit, %</b>	<b>IS Rec.%</b>	
1	CHLOROBENZENE-D5	3114-55-4		50-200	95	
2	1,4-DICHLOROETHANE-D4	3855-82-1		50-200	91	
3	FLUOROBENZENE	462-06-6		50-200	96	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank  
D - Diluted

*M/202*

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: MW-17-5	Lab Sample ID: 03-5951-12	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-12	Prep. No: -	Anal. Time: 21:17
Methanol Vol. -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	UWJ
7	N-BUTYL BENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYL BENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYL BENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	UWJ
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	UWJ
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	< 0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	< 0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	1,1,2,2-TETRACHLOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U

**Surrogates**

		Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	106
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	110
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	107
4	TOLUENE-D8	2037-26-5	73-129	107
# of out-of-control			0	

**Internal Standard**

		Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4	50-200	97
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	94
3	FLUOROBENZENE	462-06-6	50-200	99
# of out-of-control			0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory  
**Organic Analysis Results for Method 524.2**

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/04/2003
Project ID: JPL	Service ID: 35951	Collected by: JR
Sample ID: <b>TB-9-11-4-03</b>	Lab Sample ID: 03-5951-13	Received Date: 11/04/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5951-13	Prep. No: -	Anal. Time: 14:20
Methanol Vol. -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>UJ</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>UJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U <i>UJ</i>
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

*1.000*  
1643



#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	0.4	J
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	< 0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	< 0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U

## Surrogates

		Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	109
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	107
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	105
4	TOLUENE-D8	2037-26-5	73-129	107
# of out-of-control			0	

## Internal Standard

		Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4	50-200	100
2	1,4-DICHLOROETHANE-D4	3855-82-1	50-200	97
3	FLUOROBENZENE	462-06-6	50-200	104
# of out-of-control			0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

LDC #: 11280A1

## VALIDATION COMPLETENESS WORKSHEET

Date: 12-16-03

SDG #: 03-5951

Level III/IV

Page: 1 of 1

Laboratory: Applied P &amp; Ch Laboratory

Reviewer: Z. Pan

2nd Reviewer: [Signature]

METHOD: GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 11-4-03
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	%RSD OR $r^2 > 0.990$
IV.	Continuing calibration	SW	%D
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	MW-3-1
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	A	Not reviewed for Level III validation.
XII.	Compound quantitation/CRQLs	A	Not reviewed for Level III validation.
XIII.	Tentatively identified compounds (TICs)	N	Not reviewed for Level III validation. <del>Not required</del>
XIV.	System performance	A	Not reviewed for Level III validation.
XV.	Overall assessment of data	A	
XVI.	Field duplicates	SW	D = 1, 10
XVII.	Field blanks	SW	EB = 2*; 13 = TB

Note: A = Acceptable  
N = Not provided/applicable  
SW = See worksheet

ND = No compounds detected  
R = Rinsate  
FB = Field blank

D = Duplicate  
TB = Trip blank  
EB = Equipment blank

\* = ND

Validated Samples: \*\* Indicates sample underwent Level IV validation.

All Water

1	D	DUPE-5-4-Q03**	11	MW-17-4	21	03G 4717-MB	31
2		EB-9-11-4-03	12	MW-17-5	22		32
3		MW-14-1	13	TB-9-11-4-03	23		33
4		MW-14-2	14		24		34
5		MW-14-3	15		25		35
6		MW-14-4	16		26		36
7		MW-14-5	17		27		37
8		MW-17-1	18		28		38
9		MW-17-2	19		29		39
10	D	MW-17-3	20		30		40

LDC #: 11280A1  
 SDG #: 03-5951

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2  
 Reviewer: Z. Pan  
 2nd Reviewer: JL

Method: Volatiles (EPA Method 524.2)

Validation Area	Yes	No	NA	Findings/Comments
<b>I. Technical holding times</b>				
All technical holding times were met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Cooler temperature criteria was met.	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>II. GC/MS instrument performance check</b>				
Were the BFB performance results reviewed and found to be within the specified criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all samples analyzed within the 12 hour clock criteria?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>III. Initial calibration</b>				
Did the laboratory perform a 5 point calibration prior to sample analysis?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent relative standard deviations (%RSD) $\leq$ 20%?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>IV. Continuing calibration</b>				
Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were all percent differences (%D) $\leq$ 30%?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>V. Blanks</b>				
Was a method blank associated with every sample in this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was a method blank analyzed at least once every 12 hours for each matrix and concentration?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.	<input type="checkbox"/>	<input checked="" type="checkbox"/>	<input type="checkbox"/>	
<b>VI. Surrogate spikes</b>				
Were all surrogate %R within QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
If the percent recovery (%R) for one or more surrogates was out of QC limits, was a reanalysis performed to confirm samples with %R outside of criteria?	<input type="checkbox"/>	<input type="checkbox"/>	<input checked="" type="checkbox"/>	
<b>VII. Matrix spike/Matrix spike duplicates</b>				
Was a matrix spike (MS) and matrix spike duplicate (MSD) analyzed for this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
<b>VIII. Laboratory control samples</b>				
Was an LCS analyzed for this SDG?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Was an LCS analyzed per analytical batch?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the QC limits?	<input checked="" type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>	

LDC #: 11280A1  
 SDG #: 03-595

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2  
 Reviewer: Z. Pan  
 2nd Reviewer: WM

Validation Area	Yes	No	NA	Findings/Comments
<b>IX. Regional Quality Assurance and Quality Control</b>				
Were performance evaluation (PE) samples performed?			✓	
Were the performance evaluation (PE) samples within the acceptance limits?			✓	
<b>X. Internal standards</b>				
Were internal standard area counts within +/-40% from the associated calibration standard?	✓			
Were retention times within - 30% of the last continuing calibration or +/- 50% of the initial calibration?	✓			
<b>XI. Target compound identification</b>				
Were relative retention times (RRT's) within $\pm 0.06$ RRT units of the standard?	✓			
Did compound spectra meet specified EPA "Functional Guidelines" criteria?	✓			
Were chromatogram peaks verified and accounted for?	✓			
<b>XII. Compound quantitation/CRQLs</b>				
Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?	✓			
Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	✓			
<b>XIII. Tentatively identified compounds (TICs)</b>				
Were the major ions (> 25 percent relative intensity) in the reference spectrum evaluated in sample spectrum?			✓	
Were relative intensities of the major ions within $\pm 20\%$ between the sample and the reference spectra?			✓	
Did the raw data indicate that the laboratory performed a library search for all required peaks in the chromatograms (samples and blanks)?			✓	
<b>XIV. System performance</b>				
System performance was found to be acceptable.	✓			
<b>XV. Overall assessment of data</b>				
Overall assessment of data was found to be acceptable.	✓			
<b>XVI. Field duplicates</b>				
Field duplicate pairs were identified in this SDG.	✓			
Target compounds were detected in the field duplicates.	✓			
<b>XVII. Field blanks</b>				
Field blanks were identified in this SDG.	✓			
Target compounds were detected in the field blanks.	✓			

# TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method 524.2)

A. Chloromethane	G. 1,2-Dichloropropane	GG. Xylenes, total	WW. Bromobenzene	MMM. Naphthalene
B. Bromomethane	R. cis-1,3-Dichloropropene	HH. Vinyl acetate	XX. 1,2,3-Trichloropropane	NNN. 1,2,3-Trichlorobenzene
C. Vinyl chloride	S. Trichloroethene	II. 2-Chloroethylvinyl ether	YY. n-Propylbenzene	OOO. 1,3,5-Trichlorobenzene
D. Chloroethane	T. Dibromochloromethane	JJ. Dichlorodifluoromethane	ZZ. 2-Chlorotoluene	PPP. trans-1,2-Dichloroethene
E. Methylene chloride	U. 1,1,2-Trichloroethane	KK. Trichlorofluoromethane	AAA. 1,3,5-Trimethylbenzene	QQQ. cis-1,2-Dichloroethene
F. Acetone	V. Benzene	LL. Methyl-tert-butyl ether	BBB. 4-Chlorotoluene	RRR. m,p-Xylenes
G. Carbon disulfide	W. trans-1,3-Dichloropropene	MM. 1,2-Dibromo-3-chloropropane	CCC. tert-Butylbenzene	SSS. o-Xylene
H. 1,1-Dichloroethene	X. Bromoform	NN. Diethyl ether	DDD. 1,2,4-Trimethylbenzene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane
I. 1,1-Dichloroethane	Y. 4-Methyl-2-pentanone	OO. 2,2-Dichloropropane	EEE. sec-Butylbenzene	UUU. Benzyl chloride
J. 1,2-Dichloroethene, total	Z. 2-Hexanone	PP. Bromochloromethane	FFF. 1,3-Dichlorobenzene	VVV. 4-Ethyltoluene
K. Chloroform	AA. Tetrachloroethene	QQ. 1,1-Dichloropropene	GGG. p-isopropyltoluene	WWW. Ethanol
L. 1,2-Dichloroethane	BB. 1,1,2,2-Tetrachloroethane	RR. Dibromomethane	HHH. 1,4-Dichlorobenzene	XXX. Ethyl ether
M. 2-Butanone	CC. Toluene	SS. 1,3-Dichloropropane	III. n-Butylbenzene	
N. 1,1,1-Trichloroethane	DD. Chlorobenzene	TT. 1,2-Dibromoethane	JJJ. 1,2-Dichlorobenzene	
O. Carbon tetrachloride	EE. Ethylbenzene	UU. 1,1,1,2-Tetrachloroethane	KKK. 1,2,4-Trichlorobenzene	
P. Bromodichloromethane	FF. Styrene	VV. Isopropylbenzene	LLL. Hexachlorobutadiene	

Notes:

LDC #: 11280A1  
 SDG #: 03-595

**VALIDATION FINDINGS WORKSHEET I**  
**Continuing Calibration**

Page: 1 of 1  
 Reviewer: Z. Pan  
 2nd Reviewer: MH

METHOD: GC/MS VOA (EPA Method 524.2)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".  
 Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?

(Y) N N/A  
 (Y) (N) (N/A)

Were all percent differences (%D) ≤ 30%?

#	Date	Standard ID	Compound	Finding %D (Limit: ≤30.0%)	Associated Samples	Qualifications
1	11-6-03 (10:27 AM)	G4717801	A	30.98	AM + BLK	5/05/P
			B	44.52		
			00	32.14		

LDC #: 11280A1  
 SDG #: 03-5951

**VALIDATION FINDINGS WORKSHEET**  
**Field Duplicates**

Page: 1 of 1  
 Reviewer: Z. Pan  
 2nd reviewer: by

METHOD: GC/MS VOA (EPA Method 524.2)

Y  N  N/A  
 Y  N  N/A

Were field duplicate pairs identified in this SDG?

Were target compounds detected in the field duplicate pairs?

Compound	Concentration ( <u>ug/L</u> )		RPD
	# 1	10	
O	13.7	11.0	22
K	3.1	2.6	18
AA	0.6	0.4	40
S	3.8	3.1	20

Compound	Concentration ( )		RPD

Compound	Concentration ( )		RPD

LDC #: 11280A1  
SDG #: 03-5951

**VALIDATION FINDINGS WORKSHEET**  
**Field Blanks**

Page: 1 of 1  
Reviewer: Z. Pan  
2nd reviewer: UH

**METHOD:** GC/MS VOA (EPA Method 524.2)

N N/A Were field blanks identified in this SDG?  
 N N/A Were target compounds detected in the field blanks?

Sample: # 13 Field Blank /  Trip Blank / Rinsate (circle one)

Compound	Concentration Units ( $\mu\text{g/L}$ )
E	0.4

Sample: \_\_\_\_\_ Field Blank / Trip Blank / Rinsate (circle one)

Compound	Concentration Units ( )

Sample: \_\_\_\_\_ Field Blank / Trip Blank / Rinsate (circle one)

Compound	Concentration Units ( )



**VALIDATION FINDINGS WORKSHEET**  
**Initial Calibration Calculation Verification**

METHOD: GC/MS VOA (EPA Method 524.2)

The Relative Response Factor (RRF), average RRF, and percent relative standard deviation (%RSD) were recalculated for the compounds identified below using the following calculations:

$RRF = (A_x)(C_s)/(A_s)(C_x)$   
 average RRF = sum of the RRFs/number of standards  
 $\%RSD = 100 * (S/X)$   
 $A_x$  = Area of compound,  $A_s$  = Area of associated internal standard  
 $C_x$  = Concentration of compound,  $C_s$  = Concentration of internal standard  
 $S$  = Standard deviation of the RRFs  
 $X$  = Mean of the RRFs

#	Standard ID	Calibration Date	Compound (Reference Internal Standard)	Reported		Recalculated		Reported		Recalculated	
				RRF (10 std)	RRF (10 std)	Average RRF (initial)	Average RRF (initial)	%RSD	%RSD		
1	E524A003	10-21-03	Methylene chloride (1st Internal Standard)	0.221	0.221	See	Next page				
			toluene Trichloroethene (2nd internal standard)	1.322	1.322	1.354	1.354	5.13	5.12		
			Bromoforn Bromoforn (3rd internal standard)	2.614	2.614	2.542	2.542	4.04	4.04		
2			Methylene chloride (1st Internal Standard)								
			Trichloroethene (2nd internal standard)								
			Bromoforn (3rd internal standard)								
3			Methylene chloride (1st Internal Standard)								
			Trichloroethene (2nd internal standard)								
			Bromoforn (3rd internal standard)								
4			Methylene chloride (1st Internal Standard)								
			Trichloroethene (2nd internal standard)								
			Bromoforn (3rd internal standard)								

Comments: Refer to Initial Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC # 11280A1  
 SDG# 03-5951

VALIDATION FINDINGS WORKSHEET  
Initial Calibration Calculation Verification

Page: 1 of 1  
 Reviewer: Z. Pan  
 2nd Reviewer: MY

METHOD: GC / MS

Parameter: \_\_\_\_\_

Order of regression: 1.00

Date	IS Fluorobenzene	Methylene chloride	Y Area / (Area IS)	Methylene chloride Conc. (ppb)	X C / (C IS)
10/21/2003	920067	10278	0.0112	0.3	0.03
	916201	48191	0.0526	2	0.2
	919392	203033	0.2208	10	1
	955116	407143	0.4263	20	2
	907151	804880	0.8873	40	4
	1006211	1192483	1.1851	60	6

Regression Output:	Regression Output:	Reported
Constant	0.02025	c = 0.02025
Std Err of Y Est	0.04	
R Squared	0.9947	r <sup>2</sup> = 0.9947
No. of Observations	6.00	
Degrees of Freedom	4.00	
X Coefficient(s)	0.20119	b(X) a(X <sup>2</sup> )
Std Err of Coef.	0.01	0.20119 0.00000

LDC #: 11280A  
 SDG #: 03-5951

**VALIDATION FINDINGS WORKSHEET**  
**Continuing Calibration Results Verification**

Page: 1 of 1  
 Reviewer: Z. Pan  
 2nd Reviewer: MY

METHOD: GC/MS VOA (EPA Method 524.2)

The percent difference (%D) of the initial calibration average Relative Response Factors (RRFs) and the continuing calibration RRFs were recalculated for the compounds identified below using the following calculation:

% Difference =  $100 * (\text{ave. RRF} - \text{RRF}) / \text{ave. RRF}$   
 RRF =  $(A_x)(C_s) / (A_s)(C_x)$

Where: ave. RRF = initial calibration average RRF

RRF = continuing calibration RRF

A<sub>x</sub> = Area of compound,

C<sub>x</sub> = Concentration of compound,

A<sub>s</sub> = Area of associated internal standard

C<sub>s</sub> = Concentration of internal standard

#	Standard ID	Calibration Date	Compound (Reference Internal Standard)	Average RRF (Initial)	Reported		Recalculated	
					RRF (CC)	%D	RRF (CC)	%D
1	G9717801	11-6-03 (10:27AM)	Methylene chloride (1st Internal Standard)	20 *	17.33	13.37	17.33	13.35
			<del>Trichloroethane</del> (2nd internal standard)	1.354	11.1	1.204	11.1	
			<del>Bromoform</del> (3rd internal standard)	2.418	4.9	2.418	4.9	
2		Echylbenzol	Methylene chloride (1st Internal Standard)					
			Trichloroethene (2nd internal standard)					
			Bromoform (3rd internal standard)					
3			Methylene chloride (1st Internal Standard)					
			Trichloroethene (2nd internal standard)					
			Bromoform (3rd internal standard)					
4			Methylene chloride (1st Internal Standard)					
			Trichloroethene (2nd internal standard)					
			Bromoform (3rd internal standard)					

Comments: Refer to Continuing Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

\* Conc.

LDC #: 11280A1  
 SDG #: 03-5951

**VALIDATION FINDINGS WORKSHEET**  
**Surrogate Results Verification**

Page: 1 of 1  
 Reviewer: Z. Pan  
 2nd reviewer: HM

**METHOD: GC/MS VOA (EPA Method 524.2)**

The percent recoveries (%R) of surrogates were recalculated for the compounds identified below using the following calculation:

% Recovery:  $SF/SS * 100$

Where: SF = Surrogate Found  
 SS = Surrogate Spiked

Sample ID: #1

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8	20.0	21.50	108	108	0
Bromofluorobenzene		21.50	108	108	
<sup>ethane</sup> 1,2-Dichlorobenzene-d4		21.59	108	108	
Dibromofluoromethane		21.30	107	107	

Sample ID: \_\_\_\_\_

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID: \_\_\_\_\_

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID: \_\_\_\_\_

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID: \_\_\_\_\_

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

LDC #: 11280A1  
 SDG #: 03-5951

**VALIDATION FINDINGS WORKSHEET**  
**Matrix Spike/Matrix Spike Duplicates Results Verification**

Page: 1 of 1  
 Reviewer: Z. Pan  
 2nd Reviewer: HY

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the matrix spike and matrix spike duplicate were recalculated for the compounds identified below using the following calculation:

% Recovery =  $100 * (SSC - SC) / SA$  Where: SSC = Spiked sample concentration SC = Sample concentration  
 SA = Spike added

RPD =  $1 MSC - MSDC | * 2 / (MSC + MSDC)$  MSC = Matrix spike percent recovery MSDC = Matrix spike duplicate percent recovery

MS/MSD sample: MW-3-1

Compound	Spike Added (ug/L)		Sample Concentration (ug/L)	Spiked Sample Concentration (ug/L)		Matrix Spike		Matrix Spike Duplicate		MS/MSD	
	MS	MSD		MS	MSD	Percent Recovery		Percent Recovery		Reported	Recalculated
						Reported	Recalc.	Reported	Recalc.	Reported	Recalculated
1,1-Dichloroethene	20	20	0	21.1	20.9	106	106	105	105	1	1
Trichloroethene				19.5	<del>20.7</del>	98	98	<del>98</del>	98	0	0
Benzene				17.9	17.6	90	90	88	88	2	2
Toluene				17.6	17.3	88	88	87	87	1	1
Chlorobenzene	↓	↓	↓	18.6	18.1	93	93	91	91	2	2

Comments: Refer to Matrix Spike/Matrix Spike Duplicates findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

**VALIDATION FINDINGS WORKSHEET**  
**Laboratory Control Sample Results Verification**

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the laboratory control sample and laboratory control sample duplicate (if applicable) were recalculated for the compounds identified below using the following calculation:

% Recovery =  $100 * SSC/SA$  Where: SSC = Spiked sample concentration  
 SA = Spike added

RPD =  $100 * (LCS - LCS2) / (LCS + LCS2)$

LCS = Laboratory control sample percent recovery  
 LCS2 = Laboratory control sample duplicate percent recovery

LCS ID: G4717L01

Compound	Spike Added (ug/L)		Spiked Sample Concentration (ug/L)		LCS Percent Recovery		LCS2 Percent Recovery		RPD	
	LCS	LCS2	LCS	LCS2	Reported	Recalc.	Reported	Recalc.	Reported	Recalculated
	1,1-Dichloroethene	20		20.8		104	104			
Trichloroethene	↓		19.6		98	98				
Benzene	20		17.9		90	90				
Toluene	↓		17.8		89	89				
Chlorobenzene	↓		18.9		95	95				

Comments: Refer to Laboratory Control Sample findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 11280A1  
SDG #: 03-5951

### VALIDATION FINDINGS WORKSHEET

#### Sample Calculation Verification

Page: 1 of 1  
Reviewer: Z. Pan  
2nd reviewer: M

**METHOD:** GC/MS VOA (EPA Method 524.2)  
 Y  N  N/A Were all reported results recalculated and verified for all level IV samples?  
 Y  N  N/A Were all recalculated results for detected target compounds agree within 10.0% of the reported results?

Concentration =  $\frac{(A_x)(I_s)(DF)}{(A_s)(RRF)(V_s)(\%S)}$

$A_x$  = Area of the characteristic ion (EICP) for the compound to be measured  
 $A_s$  = Area of the characteristic ion (EICP) for the specific internal standard  
 $I_s$  = Amount of internal standard added in nanograms (ng)  
RRF = Relative response factor of the calibration standard.  
 $V_s$  = Volume or weight of sample pruged in milliliters (ml) or grams (g).  
Df = Dilution factor.  
%S = Percent solids, applicable to soils and solid matrices only.

Example:  
Sample I.D. #1, CCl<sub>4</sub>  
Conc. =  $(284.222) (10) ( ) ( )$   
 $(644.992) (0.321) ( ) ( )$   
= 13.7 ug/L

#	Sample ID	Compound	Reported Concentration ( )	Calculated Concentration ( )	Qualification