

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: November 5, 2003
LDC Report Date: December 22, 2003
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 03-5973

Sample Identification

MW-9
MW-3-1
MW-3-2
TB-10-11-5-03
MW-3-1MS
MW-3-1MSD

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
11/6/03	Chloromethane Bromomethane 2,2-Dichloropropane	30.98 44.52 32.14	All samples in SDG 03-5973	J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
03G4717MB	11/6/03	2-Butanone	0.5 ug/L	All samples in SDG 03-5973

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-10-11-5-03 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Compound	Concentration (ug/L)
TB-10-11-5-03	Methylene chloride	0.5

NASA JPL
Volatiles - Data Qualification Summary - SDG 03-5973

SDG	Sample	Compound	Flag	A or P	Reason
03-5973	MW-9 MW-3-1 MW-3-2 TB-10-11-5-03	Chloromethane Bromomethane 2,2-Dichloropropane	J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-5973

No Sample Data Qualified in this SDG

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/05/2003
Project ID: JPL	Service ID: 35973	Collected by: JR
Sample ID: MW-9	Lab Sample ID: 03-5973-1	Received Date: 11/05/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5973-01	Prep. No: -	Anal. Time: 15:38
Methanol Vol. -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>WJ</i>
7	N-BUTYL BENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYL BENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYL BENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>WJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLORO BENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLORO BENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLORO BENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U <i>WJ</i>
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYL BENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYL BENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2-TRICHLORO-1,2,2-TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	107
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	106
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	106
4	TOLUENE-D8	2037-26-5	73-129	108
# of out-of-control				0

Internal Standard

			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	97
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	95
3	FLUOROBENZENE	462-06-6	50-200	101
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

1072

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/05/2003
Project ID: JPL	Service ID: 35973	Collected by: JR
Sample ID: MW-3-1	Lab Sample ID: 03-5973-2	Received Date: 11/05/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5973-02	Prep. No: -	Anal. Time: 15:12
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>WJ</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>WJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U <i>WJ</i>
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2,2-TRICHLORO-1,1,2,2-TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
Surrogates				Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	108	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	107	
3	DIBROMOFLUOROMETHANE	1868-53-7		70-122	105	
4	TOLUENE-D8	2037-26-5		73-129	107	
# of out-of-control					0	
Internal Standard				Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4		50-200	97	
2	1,4-DICHLOROETHANE-D4	3855-82-1		50-200	94	
3	FLUOROBENZENE	462-06-6		50-200	100	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/05/2003
Project ID: JPL	Service ID: 35973	Collected by: JR
Sample ID: MW-3-2	Lab Sample ID: 03-5973-3	Received Date: 11/05/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5973-03	Prep. No: -	Anal. Time: 16:04
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	< 0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	< 0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	< 0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	< 0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	< 0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	< 0.5	UUT
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	< 0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	< 0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	< 0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	< 10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	0.8	
12	CHLOROBENZENE	108-90-7	µg/L	0.5	< 0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	< 0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	< 0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	< 0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	< 0.5	UUT
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	< 0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	< 0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	< 0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	< 0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	< 0.5	U
22	1,2-DICHLOROETHANE	95-50-1	µg/L	0.5	< 0.5	U
23	1,3-DICHLOROETHANE	541-73-1	µg/L	0.5	< 0.5	U
24	1,4-DICHLOROETHANE	106-46-7	µg/L	0.5	< 0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	< 0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	< 0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	< 0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	< 0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	< 0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	< 0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	< 0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	< 0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	< 0.5	UUT
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	< 0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	< 0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	< 0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	< 0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	< 0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	< 0.5	U

u 1222

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	0.3	J
56	TRICHLOROFUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	108
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	109
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	108
4	TOLUENE-D8	2037-26-5	73-129	109
# of out-of-control				0

Internal Standard

			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	96
2	1,4-DICHLOROENZENE-D4	3855-82-1	50-200	95
3	FLUOROBENZENE	462-06-6	50-200	100
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/05/2003
Project ID: JPL	Service ID: 35973	Collected by: JR
Sample ID: TB-10-11-5-03	Lab Sample ID: 03-5973-4	Received Date: 11/05/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4717	Prep. Date: 11/06/03	Anal. Date: 11/06/03
Data File Name: 5973-04	Prep. No: -	Anal. Time: 14:46
Methanol Vol. -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>UJ</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>UJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROETHANE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROETHANE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROETHANE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U <i>UJ</i>
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

1/22

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	0.5	J
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2,2-TETRACHLOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
Surrogates				Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	106	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	107	
3	DIBROMOFLUOROMETHANE	1868-53-7		70-122	105	
4	TOLUENE-D8	2037-26-5		73-129	107	
# of out-of-control					0	
Internal Standard				Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4		50-200	99	
2	1,4-DICHLOROBENZENE-D4	3855-82-1		50-200	96	
3	FLUOROBENZENE	462-06-6		50-200	102	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

LDC #: 11280B1**VALIDATION COMPLETENESS WORKSHEET**Date: 12-16-03SDG #: 03-5973

Level III

Page: 1 of 1Laboratory: Applied P & Ch LaboratoryReviewer: Z. Pan2nd Reviewer: [Signature]**METHOD:** GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>11-5-03</u>
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	$\%RSD$ OR $r^2 > 0.990$
IV.	Continuing calibration	SW	$\%D$
V.	Blanks	SW	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	TB = 4

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

All Water

1	MW-9	11	<u>03G4717-MB</u>	21		31	
2	MW-3-1	12		22		32	
3	MW-3-2	13		23		33	
4	TB-10-11-5-03	14		24		34	
5	MW-3-1MS	15		25		35	
6	MW-3-1MSD	16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: November 6, 2003
LDC Report Date: December 22, 2003
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 03-6002

Sample Identification

Dupe-7-4-Q03**
MW-5
MW-8
MW-10
TB-11-6-03
MW-5MS
MW-5MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
11/11/03 (20:21)	Chloromethane	39.69	All samples in SDG 03-6002	J (all detects)	P
	Bromomethane	37.19		UJ (all non-detects) J (all detects) UJ (all non-detects)	

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
03G4771-MB	11/12/03	Methylene chloride	0.4 ug/L	All samples in SDG 03-6002

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples Dupe-7-4-Q03** and MW-8 were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-11-6-03 was identified as a trip blank. No volatile contaminants were found in this blank.

**NASA JPL
Volatiles - Data Qualification Summary - SDG 03-6002**

SDG	Sample	Compound	Flag	A or P	Reason
03-6002	Dupe-7-4-Q03** MW-5 MW-8 MW-10 TB-11-6-03	Chloromethane Bromomethane	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

**NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-6002**

No Sample Data Qualified in this SDG

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/06/2003
Project ID: JPL GW Mon-4Q03	Service ID: 36002	Collected by:
Sample ID: DUPE-7-4-Q03	Lab Sample ID: 03-6002-1	Received Date: 11/06/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4771	Prep. Date: 11/12/03	Anal. Date: 11/12/03
Data File Name: 6002-01	Prep. No: -	Anal. Time: 03:59
Methanol Vol. -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U UJ
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U UJ
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2,2-TETRACHLOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
Surrogates				Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	107	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	116	
3	DIBROMOFLUOROMETHANE	1868-53-7		70-122	112	
4	TOLUENE-D8	2037-26-5		73-129	104	
# of out-of-control					0	
Internal Standard				Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4		50-200	103	
2	1,4-DICHLOROBENZENE-D4	3855-82-1		50-200	103	
3	FLUOROBENZENE	462-06-6		50-200	100	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/06/2003
Project ID: JPL GW Mon-4Q03	Service ID: 36002	Collected by:
Sample ID: MW-5	Lab Sample ID: 03-6002-2	Received Date: 11/06/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4771	Prep. Date: 11/12/03	Anal. Date: 11/12/03
Data File Name: 6002-02	Prep. No: -	Anal. Time: 03:33
Methanol Vol. -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>UT</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>UT</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2,2-TETRACHLOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
Surrogates				Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	105	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	114	
3	DIBROMOFUOROMETHANE	1868-53-7		70-122	112	
4	TOLUENE-D8	2037-26-5		73-129	102	
# of out-of-control					0	
Internal Standard				Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4		50-200	104	
2	1,4-DICHLOROETHANE-D4	3855-82-1		50-200	105	
3	FLUOROBENZENE	462-06-6		50-200	101	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/06/2003
Project ID: JPL GW Mon-4Q03	Service ID: 36002	Collected by:
Sample ID: MW-8	Lab Sample ID: 03-6002-3	Received Date: 11/06/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4771	Prep. Date: 11/12/03	Anal. Date: 11/12/03
Data File Name: 6002-03	Prep. No: -	Anal. Time: 04:26
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>UJ</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>UJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROETHANE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROETHANE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROETHANE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

1222

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates		Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	107
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	115
3	DIBROMOFUOROMETHANE	1868-53-7	70-122	113
4	TOLUENE-D8	2037-26-5	73-129	104
# of out-of-control			0	

Internal Standard		Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4	50-200	104
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	104
3	FLUOROBENZENE	462-06-6	50-200	100
# of out-of-control			0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

LDC #: 11280C1

VALIDATION COMPLETENESS WORKSHEET

Date: 12-19-03

SDG #: 03-6002

Level III/IV

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: Z. Pan

2nd Reviewer: [Signature]

METHOD: GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 11-6-03
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	%RSD OR $r^2 > 0.990$
IV.	Continuing calibration	SW	%D
V.	Blanks	SW	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	A	Not reviewed for Level III validation.
XII.	Compound quantitation/CRQLs	A	Not reviewed for Level III validation.
XIII.	Tentatively identified compounds (TICs)	N	Not reviewed for Level III validation. Not required
XIV.	System performance	A	Not reviewed for Level III validation.
XV.	Overall assessment of data	A	
XVI.	Field duplicates	ND	D = 1, 3
XVII.	Field blanks	ND	TB = 5

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: ** Indicates sample underwent Level IV validation.

AU Waters

1 D	Dupe-7-4-Q03**	11	03G4771-MB	21		31	
2	MW-5	12		22		32	
3 D	MW-8	13		23		33	
4	MW-10	14		24		34	
5	TB-11-6-03	15		25		35	
6	MW-5MS	16		26		36	
7	MW-5MSD	17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

LDC #: 11280 C1
 SDG #: 03-6002

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2
 Reviewer: Z. Pan
 2nd Reviewer: lmj

Method: Volatiles (EPA Method 524.2)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical holding times				
All technical holding times were met.	✓			
Cooler temperature criteria was met.	✓			
II. GC/MS Instrument performance check				
Were the BFB performance results reviewed and found to be within the specified criteria?	✓			
Were all samples analyzed within the 12 hour clock criteria?	✓			
III. Initial calibration				
Did the laboratory perform a 5 point calibration prior to sample analysis?	✓			
Were all percent relative standard deviations (%RSD) ≤ 20%?		✓		
IV. Continuing calibration				
Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?	✓			
Were all percent differences (%D) ≤ 30%?		✓		
V. Blanks				
Was a method blank associated with every sample in this SDG?	✓			
Was a method blank analyzed at least once every 12 hours for each matrix and concentration?	✓			
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.	✓			
VI. Surrogate spikes				
Were all surrogate %R within QC limits?	✓			
If the percent recovery (%R) for one or more surrogates was out of QC limits, was a reanalysis performed to confirm samples with %R outside of criteria?			✓	
VII. Matrix spike/Matrix spike duplicates				
Was a matrix spike (MS) and matrix spike duplicate (MSD) analyzed for this SDG?	✓			
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the QC limits?	✓			
VIII. Laboratory control samples				
Was an LCS analyzed for this SDG?	✓			
Was an LCS analyzed per analytical batch?	✓			
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the QC limits?	✓			

LDC #: 11280 C1
 SDG #: 03-6002

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: Z. Pan
 2nd Reviewer: lh

Validation Area	Yes	No	NA	Findings/Comments
IX. Regional Quality Assurance and Quality Control				
Were performance evaluation (PE) samples performed?			✓	
Were the performance evaluation (PE) samples within the acceptance limits?			✓	
X. Internal standards				
Were internal standard area counts within +/-40% from the associated calibration standard?	✓			
Were retention times within - 30% of the last continuing calibration or +/- 50% of the initial calibration?	✓			
XI. Target compound identification				
Were relative retention times (RRT's) within ± 0.06 RRT units of the standard?			✓	
Did compound spectra meet specified EPA "Functional Guidelines" criteria?			✓	
Were chromatogram peaks verified and accounted for?			✓	
XII. Compound quantitation/CRQLs				
Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?	✓			
Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	✓			
XIII. Tentatively identified compounds (TICs)				
Were the major ions (> 25 percent relative intensity) in the reference spectrum evaluated in sample spectrum?			✓	
Were relative intensities of the major ions within $\pm 20\%$ between the sample and the reference spectra?			✓	
Did the raw data indicate that the laboratory performed a library search for all required peaks in the chromatograms (samples and blanks)?			✓	
XIV. System performance				
System performance was found to be acceptable.	✓			
XV. Overall assessment of data				
Overall assessment of data was found to be acceptable.	✓			
XVI. Field duplicates				
Field duplicate pairs were identified in this SDG.	✓			
Target compounds were detected in the field duplicates.		✓		
XVII. Field blanks				
Field blanks were identified in this SDG.	✓			
Target compounds were detected in the field blanks.		✓		

TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method 524.2)

A. Chloromethane	Q. 1,2-Dichloropropane	GG. Xylenes, total	WW. Bromobenzene	MMM. Naphthalene
B. Bromomethane	R. cis-1,3-Dichloropropene	HH. Vinyl acetate	XX. 1,2,3-Trichloropropane	NNN. 1,2,3-Trichlorobenzene
C. Vinyl chloride	S. Trichloroethene	II. 2-Chloroethylvinyl ether	YY. n-Propylbenzene	OOO. 1,3,5-Trichlorobenzene
D. Chloroethane	T. Dibromochloromethane	JJ. Dichlorodifluoromethane	ZZ. 2-Chlorotoluene	PPP. trans-1,2-Dichloroethene
E. Methylene chloride	U. 1,1,2-Trichloroethane	KK. Trichlorofluoromethane	AAA. 1,3,5-Trimethylbenzene	QQQ. cis-1,2-Dichloroethene
F. Acetone	V. Benzene	LL. Methyl-tert-butyl ether	BBB. 4-Chlorotoluene	RRR. m,p-Xylenes
G. Carbon disulfide	W. trans-1,3-Dichloropropene	MM. 1,2-Dibromo-3-chloropropane	CCC. tert-Butylbenzene	SSS. o-Xylene
H. 1,1-Dichloroethene	X. Bromoform	NN. Diethyl ether	DDD. 1,2,4-Trimethylbenzene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane
I. 1,1-Dichloroethane	Y. 4-Methyl-2-pentanone	OO. 2,2-Dichloropropane	EEE. sec-Butylbenzene	UUU. Benzyl chloride
J. 1,2-Dichloroethene, total	Z. 2-Hexanone	PP. Bromochloromethane	FFF. 1,3-Dichlorobenzene	VVV. 4-Ethyltoluene
K. Chloroform	AA. Tetrachloroethene	QQ. 1,1-Dichloropropene	GGG. p-Isopropyltoluene	WWW. Ethanol
L. 1,2-Dichloroethane	BB. 1,1,2,2-Tetrachloroethane	RR. Dibromomethane	HHH. 1,4-Dichlorobenzene	XXX. Ethyl ether
M. 2-Butanone	CC. Toluene	SS. 1,3-Dichloropropane	III. n-Butylbenzene	
N. 1,1,1-Trichloroethane	DD. Chlorobenzene	TT. 1,2-Dibromoethane	JJJ. 1,2-Dichlorobenzene	
O. Carbon tetrachloride	EE. Ethylbenzene	UU. 1,1,1,2-Tetrachloroethane	KKK. 1,2,4-Trichlorobenzene	
P. Bromodichloromethane	FF. Styrene	VV. Isopropylbenzene	LLL. Hexachlorobutadiene	

Notes:

VALIDATION FINDINGS WORKSHEET I
Continuing Calibration

METHOD: GC/MS VOA (EPA Method 524.2)
 Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
 Y N N/A Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?
 Y N N/A Were all percent differences (%D) \leq 30%?

#	Date	Standard ID	Compound	Finding %D (Limit: $\leq 30.0\%$)	Associated Samples	Qualifications
1	11-11-2003 (8:21 PM)	G4771801	A	39.69	MU + BLK ↓	J/WJ/AP ↓
			B	37.19		

LDC #: 11280 C1
 SDG #: 03-600Z

VALIDATION FINDINGS WORKSHEET

Blanks

Page: 1 of 1
 Reviewer: Z. Pan
 2nd Reviewer: MH

METHOD: GC/MS VOA (EPA Method 524.2)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- Y N N/A Was a method blank associated with every sample in this SDG?
- Y N N/A Was a method blank analyzed at least once every 12 hours for each matrix and concentration?
- Y N N/A Was there contamination in the method blanks? If yes, please see the qualifications below.

Blank analysis date: 11-12-03 Associated Samples: MB
 Conc. units: ug/L

Compound	Blank ID	Sample Identification				
	03G4771-MB	#1	2	3	4	5
Methylene chloride	0.4	-	-	-	-	-
Acetone						
CRQL						

Blank analysis date: _____ Associated Samples: _____
 Conc. units: _____

Compound	Blank ID	Sample Identification				
Methylene chloride						
Acetone						
CRQL						

CIRCLED RESULTS WERE NOT QUALIFIED. ALL RESULTS NOT CIRCLED WERE QUALIFIED BY THE FOLLOWING STATEMENT:

Note: Common contaminants such as Methylene chloride, Acetone, 2-Butanone, Carbon disulfide and TICs that were detected in samples within ten times the associated method blank concentration were qualified as not detected, "U". Other contaminants within five times the method blank concentration were also qualified as not detected, "U".

LDC # 11200 C1
 SDG# 03-600 Z

VALIDATION FINDINGS WORKSHEET
Initial Calibration Calculation Verification

Page: 1 of 1
 Reviewer: Z. Pan
 2nd Reviewer: MH

METHOD: GC/MS

Parameter: _____

Order of regression: 1.00

Date	IS Fluorobenzene	Methylene chloride	Y Area / (Area IS)	Methylene chloride Conc. (ppb)	X C / (C IS)
10/21/2003	920067	10278	0.0112	0.3	0.03
	916201	48191	0.0526	2	0.2
	919392	203033	0.2208	10	1
	955116	407143	0.4263	20	2
	907151	804880	0.8873	40	4
	1006211	1192483	1.1851	60	6

Regression Output: Regression Output:		Reported
Constant	0.02025	c = 0.02025
Std Err of Y Est	0.04	
R Squared	0.9947	r ² = 0.9947
No. of Observations	6.00	
Degrees of Freedom	4.00	
X Coefficient(s)	0.20119	b(X) 0.20119
Std Err of Coef.	0.01	a(X ²) 0.00000

VALIDATION FINDINGS WORKSHEET
Initial Calibration Calculation Verification

METHOD: GC/MS VOA (EPA Method 524.2)

The Relative Response Factor (RRF), average RRF, and percent relative standard deviation (%RSD) were recalculated for the compounds identified below using the following calculations:

$RRF = (A_x)(C_s)/(A_s)(C_x)$
 average RRF = sum of the RRFs/number of standards
 $\%RSD = 100 * (S/X)$
 A_x = Area of compound, A_s = Area of associated internal standard
 C_x = Concentration of compound, C_s = Concentration of internal standard
 S = Standard deviation of the RRFs
 X = Mean of the RRFs

#	Standard ID	Calibration Date	Compound (Reference Internal Standard)	Reported		Recalculated		Reported		Recalculated	
				RRF (/O std)	RRF (/O std)	Average RRF (Initial)	Average RRF (Initial)	%RSD	%RSD		
1	E524A003	10-21-03 Ethylbenzene	Methylene chloride (1st Internal Standard)	0.221	0.221	See	Next page				
			toluene Trichloroethene (2nd internal standard)	1.322	1.322	1.354	1.354	5.13	5.12		
			Bromoform (3rd internal standard)	2.614	2.614	2.542	2.542	4.04	4.04		
2			Methylene chloride (1st Internal Standard)								
			Trichloroethene (2nd internal standard)								
			Bromoform (3rd internal standard)								
3			Methylene chloride (1st Internal Standard)								
			Trichloroethene (2nd internal standard)								
			Bromoform (3rd internal standard)								
4			Methylene chloride (1st Internal Standard)								
			Trichloroethene (2nd internal standard)								
			Bromoform (3rd internal standard)								

Comments: Refer to Initial Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 11280C1
 SDG #: 03-600Z

VALIDATION FINDINGS WORKSHEET
Continuing Calibration Results Verification

Page: 1 of 1
 Reviewer: Z. Pan
 2nd Reviewer: 47

METHOD: GC/MS VOA (EPA Method 524.2)

The percent difference (%D) of the initial calibration average Relative Response Factors (RRFs) and the continuing calibration RRFs were recalculated for the compounds identified below using the following calculation:

% Difference = $100 * (\text{ave. RRF} - \text{RRF}) / \text{ave. RRF}$ Where: ave. RRF = initial calibration average RRF
 RRF = $(A_x)(C_s) / (A_s)(C_x)$ RRF = continuing calibration RRF
 A_x = Area of compound, A_s = Area of associated internal standard
 C_x = Concentration of compound, C_s = Concentration of internal standard

#	Standard ID	Calibration Date	Compound (Reference Internal Standard)	Average RRF (Initial)	Reported		Recalculated	
					RRF (CC)	%D	RRF (CC)	%D
1	G477/Q01	11-11-03 (8:21 PM)	Methylene chloride (1st Internal Standard)	20 *	17.58	12.09	17.58	12.1
			Toluene Trichloroethene (2nd internal standard)	1.354	1.209	1.209	10.7	
			Ethyl benzene Bromoform (3rd internal standard)	2.542	2.456	2.456	3.4	
2			Methylene chloride (1st Internal Standard)					
			Trichloroethene (2nd internal standard)					
			Bromoform (3rd internal standard)					
3			Methylene chloride (1st Internal Standard)					
			Trichloroethene (2nd internal standard)					
			Bromoform (3rd internal standard)					
4			Methylene chloride (1st Internal Standard)					
			Trichloroethene (2nd internal standard)					
			Bromoform (3rd internal standard)					

Comments: Refer to Continuing Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

* Conc.

LDC #: 11280 C1
 SDG #: 03-6002

VALIDATION FINDINGS WORKSHEET
Surrogate Results Verification

Page: 1 of 1
 Reviewer: Z. Pan
 2nd reviewer: UH

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) of surrogates were recalculated for the compounds identified below using the following calculation:

% Recovery: SF/SS * 100

Where: SF = Surrogate Found
 SS = Surrogate Spiked

Sample ID: # 1

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8	20.0	20.78	104	104	0
Bromofluorobenzene		21.34	107	107	
1,2-Dichlorobenzene-d4		23.18	116	116	
<i>Dibromofluoromethane</i>		22.36	112	112	

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

LDC #: 11280C1
 SDG #: 03-600Z

VALIDATION FINDINGS WORKSHEET
Matrix Spike/Matrix Spike Duplicates Results Verification

Page: 1 of 1
 Reviewer: Z. Pan
 2nd Reviewer: HN

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the matrix spike and matrix spike duplicate were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * (SSC - SC) / SA$ Where: SSC = Spiked sample concentration SC = Sample concentration
 SA = Spike added

RPD = $1 MSC - MSDC | * 2 / (MSC + MSDC)$ MSC = Matrix spike percent recovery MSDC = Matrix spike duplicate percent recovery

MS/MSD sample: # 6/7

Compound	Spike Added (ug/L)		Sample Concentration (ug/L)	Spiked Sample Concentration (ug/L)		Matrix Spike Percent Recovery		Matrix Spike Duplicate Percent Recovery		MS/MSD RPD	
	MS	MSD		MS	MSD	Reported	Recalc.	Reported	Recalc.	Reported	Recalculated
1,1-Dichloroethene	20	20	0	20.4	20.1	102	102	101	101	1	1
Trichloroethene				18.8	18.9	94	94	95	95	1	1
Benzene				17.3	17.3	87	87	87	87	0	0
Toluene				17.3	17.1	87	87	86	86	1	1
Chlorobenzene				18.6	18.2	93	93	91	91	2	2

Comments: Refer to Matrix Spike/Matrix Spike Duplicates findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

VALIDATION FINDINGS WORKSHEET
Laboratory Control Sample Results Verification

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the laboratory control sample and laboratory control sample duplicate (if applicable) were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * SSC/SA$ Where: SSC = Spiked sample concentration
 SA = Spike added

RPD = $100 * (LCS - LCSD) / (LCS + LCSD)$

LCS = Laboratory control sample percent recovery

LCSD = Laboratory control sample duplicate percent recovery

LCS ID: G4771L01

Compound	Spike Added (ug/L)		Spiked Sample Concentration (ug/L)		LCS		LCSD		LCS/LCSD	
	LCS	LCSD	LCS	LCSD	Percent Recovery		Percent Recovery		RPD	
	Reported	Recalc.	Reported	Recalc.	Reported	Recalc.	Reported	Recalc.	Reported	Recalculated
1,1-Dichloroethene	20		20.4		102	102				
Trichloroethene			18.5		93	93				
Benzene			17.0		85	85				
Toluene			17.2		86	86				
Chlorobenzene			18.5		93	93				

Comments: Refer to Laboratory Control Sample findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 11280C1
 SDG #: 03-6002

VALIDATION FINDINGS WORKSHEET Sample Calculation Verification

Page: 1 of 1
 Reviewer: Z. Pan
 2nd reviewer: MM

METHOD: GC/MS VOA (EPA Method 524.2)

Y N N/A Were all reported results recalculated and verified for all level IV samples?
Y N N/A Were all recalculated results for detected target compounds agree within 10.0% of the reported results?

$$\text{Concentration} = \frac{(A_x)(I_s)(DF)}{(A_s)(RRF)(V_o)(\%S)}$$

- A_x = Area of the characteristic ion (EICP) for the compound to be measured
- A_s = Area of the characteristic ion (EICP) for the specific internal standard
- I_s = Amount of internal standard added in nanograms (ng)
- RRF = Relative response factor of the calibration standard.
- V_o = Volume or weight of sample pruged in milliliters (ml) or grams (g).
- Df = Dilution factor.
- %S = Percent solids, applicable to soils and solid matrices only.

Example:

Sample I.D. #1 ND :

Conc. = () () ()
 () () () ()
 =

#	Sample ID	Compound	Reported Concentration ()	Calculated Concentration ()	Qualification

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: November 7, 2003
LDC Report Date: December 22, 2003
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 03-6034

Sample Identification

DUPE-2-4-Q03**
MW-6
MW-13
MW-15
TB-12-11-7-03
MW-13MS
MW-13MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
11/11/03 (20:21)	Chloromethane	39.69	All samples in SDG 03-6034	J (all detects)	P
	Bromomethane	37.19		UJ (all non-detects)	
				J (all detects)	
				UJ (all non-detects)	

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
03G4771-MB	11/12/03	Methylene chloride	0.4 ug/L	All samples in SDG 03-6034

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples DUPE-2-4-Q03** and MW-15 were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-12-11-7-03 was identified as a trip blank. No volatile contaminants were found in this blank.

NASA JPL
Volatiles - Data Qualification Summary - SDG 03-6034

SDG	Sample	Compound	Flag	A or P	Reason
03-6034	DUPE-2-4-Q03** MW-6 MW-13 MW-15 TB-12-11-7-03	Chloromethane Bromomethane	J (all detects) UJ (all non-detects) J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-6034

No Sample Data Qualified in this SDG

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/07/2003
Project ID: JPL	Service ID: 36034	Collected by: JR
Sample ID: DUPE-2-4-Q03	Lab Sample ID: 03-6034-1	Received Date: 11/07/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4771	Prep. Date: 11/12/03	Anal. Date: 11/12/03
Data File Name: 6034-01	Prep. No: -	Anal. Time: 05:18
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2,2-TETRACHLOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

			Control Limit, %	Surro. Rec.%
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	107
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	115
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	113
4	TOLUENE-D8	2037-26-5	73-129	103
# of out-of-control				0

Internal Standard

			Control Limit, %	IS Rec.%
1	CHLOROBENZENE-D5	3114-55-4	50-200	102
2	1,4-DICHLOROETHANE-D4	3855-82-1	50-200	102
3	FLUOROBENZENE	462-06-6	50-200	97
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	11/07/2003
Project ID:	JPL	Service ID:	36034	Collected by:	JR
Sample ID:	MW-6	Lab Sample ID:	03-6034-2	Received Date:	11/07/2003
Sample Type:	Field Sample	Sample Matrix:	Water	Moisture %:	-
Anal. Method:	524.2	Prep. Method:	5030	Instrument ID:	GC/MS: A
Batch No:	03G4771	Prep. Date:	11/12/03	Anal. Date:	11/12/03
Data File Name:	6034-02	Prep. No:	-	Anal. Time:	05:44
Methanol Vol.	-	Sample Amount:	25.0 mL	Dilution Factor:	1
Test Level:	Low	Spurge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U ^{UJ}
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	0.3	J
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U ^{UJ}
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	0.9	
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	0.8	
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	3.0	
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	< 0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	1,1,2,2-TETRACHLOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U
Surrogates				Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	107	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	117	
3	DIBROMOFLUOROMETHANE	1868-53-7		70-122	113	
4	TOLUENE-D8	2037-26-5		73-129	104	
# of out-of-control					0	
Internal Standard				Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4		50-200	102	
2	1,4-DICHLOROETHANE-D4	3855-82-1		50-200	103	
3	FLUOROBENZENE	462-06-6		50-200	98	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/07/2003
Project ID: JPL	Service ID: 36034	Collected by: JR
Sample ID: MW-13	Lab Sample ID: 03-6034-3	Received Date: 11/07/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4771	Prep. Date: 11/12/03	Anal. Date: 11/12/03
Data File Name: 6034-03	Prep. No: -	Anal. Time: 03:07
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Spurge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>UJ</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	1.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	1.7	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>UJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	0.4	J
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	≥0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	0.9	
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	9.0	
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2,2-TETRACHLOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
Surrogates				Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	109	
2	1,2-DICHLOROETHANE-D4	17060-07-0		70-129	113	
3	DIBROMOFLUOROMETHANE	1868-53-7		70-122	111	
4	TOLUENE-D8	2037-26-5		73-129	103	
# of out-of-control					0	
Internal Standard				Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4		50-200	105	
2	1,4-DICHLOROETHANE-D4	3855-82-1		50-200	105	
3	FLUOROBENZENE	462-06-6		50-200	100	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	11/07/2003
Project ID:	JPL	Service ID:	36034	Collected by:	JR
Sample ID:	MW-15	Lab Sample ID:	03-6034-4	Received Date:	11/07/2003
Sample Type:	Field Sample	Sample Matrix:	Water	Moisture %:	-
Anal. Method:	524.2	Prep. Method:	5030	Instrument ID:	GC/MS: A
Batch No:	03G4771	Prep. Date:	11/12/03	Anal. Date:	11/12/03
Data File Name:	6034-04	Prep. No:	-	Anal. Time:	06:10
Methanol Vol.	-	Sample Amount:	25.0 mL	Dilution Factor:	1
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U <i>UJ</i>
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U <i>UJ</i>
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	< 0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	< 10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	< 0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	< 0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	1,1,2,2-TETRACHLOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U

Surrogates

			Control Limit, %	Surro. Rec. %
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	106
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	114
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	113
4	TOLUENE-D8	2037-26-5	73-129	103
	# of out-of-control			0

Internal Standard

			Control Limit, %	IS Rec. %
1	CHLOROBENZENE-D5	3114-55-4	50-200	102
2	1,4-DICHLOROETHANE-D4	3855-82-1	50-200	102
3	FLUOROBENZENE	462-06-6	50-200	98
	# of out-of-control			0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 11/07/2003
Project ID: JPL	Service ID: 36034	Collected by: JR
Sample ID: TB-12-11-7-03	Lab Sample ID: 03-6034-5	Received Date: 11/07/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: A
Batch No: 03G4771	Prep. Date: 11/12/03	Anal. Date: 11/12/03
Data File Name: 6034-05	Prep. No: -	Anal. Time: 02:41
Methanol Vol: -	Sample Amount: 25.0 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	UW
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	UW
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2,2-TRICHLORO-1,1,2,2-TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

			Control Limit, %	Surro. Rec. %
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	106
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	114
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	112
4	TOLUENE-D8	2037-26-5	73-129	102
# of out-of-control				0

Internal Standard

			Control Limit, %	IS Rec. %
1	CHLOROBENZENE-D5	3114-55-4	50-200	103
2	1,4-DICHLOROETHANE-D4	3855-82-1	50-200	104
3	FLUOROBENZENE	462-06-6	50-200	99
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

1827

LDC #: 11280D1

VALIDATION COMPLETENESS WORKSHEET

Date: 12-22-03

SDG #: 03-6034

Level III/IV

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: Z. Pan

2nd Reviewer: K

METHOD: GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 11-7-03
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	%RSD, OR r ²
IV.	Continuing calibration	SW	%D
V.	Blanks	SW	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	A	Not reviewed for Level III validation.
XII.	Compound quantitation/CRQLs	A	Not reviewed for Level III validation.
XIII.	Tentatively identified compounds (TICs)	N	Not reviewed for Level III validation.
XIV.	System performance	A	Not reviewed for Level III validation.
XV.	Overall assessment of data	A	
XVI.	Field duplicates	ND	D = 1, 4
XVII.	Field blanks	ND	TB = 5

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: ** Indicates sample underwent Level IV validation.

All Water

1	D	DUPE-2-4-Q03**	11	03G4771-MB	21		31	
2		MW-6	12		22		32	
3		MW-13	13		23		33	
4	D	MW-15	14		24		34	
5		TB-12-11-7-03	15		25		35	
6		MW-13MS	16		26		36	
7		MW-13MSD	17		27		37	
8			18		28		38	
9			19		29		39	
10			20		30		40	

LDC #: 11280D1
 SDG #: 03-6034

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2
 Reviewer: Z. Pan
 2nd Reviewer: [Signature]

Method: Volatiles (EPA Method 524.2)

Validation Area	Yes	No	NA	Findings/Comments
I. Technical holding times				
All technical holding times were met.	✓			
Cooler temperature criteria was met.	✓			
II. GC/MS Instrument performance check				
Were the BFB performance results reviewed and found to be within the specified criteria?	✓			
Were all samples analyzed within the 12 hour clock criteria?	✓			
III. Initial calibration				
Did the laboratory perform a 5 point calibration prior to sample analysis?	✓			
Were all percent relative standard deviations (%RSD) ≤ 20%?		✓		
IV. Continuing calibration				
Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?	✓			
Were all percent differences (%D) ≤ 30%?		✓		
V. Blanks				
Was a method blank associated with every sample in this SDG?	✓			
Was a method blank analyzed at least once every 12 hours for each matrix and concentration?	✓			
Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet.	✓			
VI. Surrogate spikes				
Were all surrogate %R within QC limits?	✓			
If the percent recovery (%R) for one or more surrogates was out of QC limits, was a reanalysis performed to confirm samples with %R outside of criteria?			✓	
VII. Matrix spike/Matrix spike duplicates				
Was a matrix spike (MS) and matrix spike duplicate (MSD) analyzed for this SDG?	✓			
Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the QC limits?	✓			
VIII. Laboratory control samples				
Was an LCS analyzed for this SDG?	✓			
Was an LCS analyzed per analytical batch?	✓			
Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the QC limits?	✓			

LDC #: 11280 D1
 SDG #: 03-6034

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: Z. Pan
 2nd Reviewer: [Signature]

Validation Area	Yes	No	NA	Findings/Comments
IX. Regional Quality Assurance and Quality Control				
Were performance evaluation (PE) samples performed?			✓	
Were the performance evaluation (PE) samples within the acceptance limits?			✓	
X. Internal standards				
Were internal standard area counts within +/-40% from the associated calibration standard?	✓			
Were retention times within - 30% of the last continuing calibration or +/- 50% of the initial calibration?	✓			
XI. Target compound identification				
Were relative retention times (RRT's) within ± 0.06 RRT units of the standard?			✓	
Did compound spectra meet specified EPA "Functional Guidelines" criteria?			✓	
Were chromatogram peaks verified and accounted for?			✓	
XII. Compound quantitation/CRQLs				
Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound?	✓			
Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation?	✓			
XIII. Tentatively identified compounds (TICs)				
Were the major ions (> 25 percent relative intensity) in the reference spectrum evaluated in sample spectrum?			✓	
Were relative intensities of the major ions within ± 20% between the sample and the reference spectra?			✓	
Did the raw data indicate that the laboratory performed a library search for all required peaks in the chromatograms (samples and blanks)?			✓	
XIV. System performance				
System performance was found to be acceptable.	✓			
XV. Overall assessment of data				
Overall assessment of data was found to be acceptable.	✓			
XVI. Field duplicates				
Field duplicate pairs were identified in this SDG.	✓			
Target compounds were detected in the field duplicates.		✓		
XVII. Field blanks				
Field blanks were identified in this SDG.	✓			
Target compounds were detected in the field blanks.		✓		

TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method 524.2)

A. Chloromethane	Q. 1,2-Dichloropropane	GG. Xylenes, total	WW. Bromobenzene	MMM. Naphthalene
B. Bromomethane	R. cis-1,3-Dichloropropene	HH. Vinyl acetate	XX. 1,2,3-Trichloropropane	NNN. 1,2,3-Trichlorobenzene
C. Vinyl chloride	S. Trichloroethene	II. 2-Chloroethylvinyl ether	YY. n-Propylbenzene	OOO. 1,3,5-Trichlorobenzene
D. Chloroethane	T. Dibromochloromethane	JJ. Dichlorodifluoromethane	ZZ. 2-Chlorotoluene	PPP. trans-1,2-Dichloroethene
E. Methylene chloride	U. 1,1,2-Trichloroethane	KK. Trichlorofluoromethane	AAA. 1,3,5-Trimethylbenzene	QQQ. cis-1,2-Dichloroethene
F. Acetone	V. Benzene	LL. Methyl-tert-butyl ether	BBB. 4-Chlorotoluene	RRR. m,p-Xylenes
G. Carbon disulfide	W. trans-1,3-Dichloropropene	MM. 1,2-Dibromo-3-chloropropane	CCC. tert-Butylbenzene	SSS. o-Xylene
H. 1,1-Dichloroethene	X. Bromoform	NN. Diethyl ether	DDD. 1,2,4-Trimethylbenzene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane
I. 1,1-Dichloroethane	Y. 4-Methyl-2-pentanone	OO. 2,2-Dichloropropane	EEE. sec-Butylbenzene	UUU. Benzyl chloride
J. 1,2-Dichloroethene, total	Z. 2-Hexanone	PP. Bromochloromethane	FFF. 1,3-Dichlorobenzene	VVV. 4-Ethyltoluene
K. Chloroform	AA. Tetrachloroethene	QQ. 1,1-Dichloropropene	GGG. p-Isopropyltoluene	WWW. Ethanol
L. 1,2-Dichloroethane	BB. 1,1,2,2-Tetrachloroethane	RR. Dibromomethane	HHH. 1,4-Dichlorobenzene	XXX. Ethyl ether
M. 2-Butanone	CC. Toluene	SS. 1,3-Dichloropropane	III. n-Butylbenzene	
N. 1,1,1-Trichloroethane	DD. Chlorobenzene	TT. 1,2-Dibromoethane	JJJ. 1,2-Dichlorobenzene	
O. Carbon tetrachloride	EE. Ethylbenzene	UU. 1,1,1,2-Tetrachloroethane	KKK. 1,2,4-Trichlorobenzene	
P. Bromodichloromethane	FF. Styrene	VV. Isopropylbenzene	LLL. Hexachlorobutadiene	

Notes:

LDC #: 11280D1
 SDG #: 03-6034

VALIDATION FINDINGS WORKSHEET
Continuing Calibration

Page: 1 of 1
 Reviewer: Z. Pan
 2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
 Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?

Y
 N
 N/A

Were all percent differences (%D) ≤ 30% ?

#	Date	Standard ID	Compound	Finding %D (Limit: ≤30.0%)	Associated Samples	Qualifications
1	11-11-2003 (8:21 PM)	G4771801	A	39.69	MU + BLK	3/03/AP
			B	37.19	↓	↓

LDC #: 11280 D1
 SDG #: 03-6034

VALIDATION FINDINGS WORKSHEET
Blanks

Page: 1 of 1
 Reviewer: Z. Pan
 2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
 N N/A Was a method blank associated with every sample in this SDG?
 N N/A Was a method blank analyzed at least once every 12 hours for each matrix and concentration?
 N N/A Was there contamination in the method blanks? If yes, please see the qualifications below.

Blank analysis date: 11-12-03
 Conc. units: ug/L Associated Samples: MC (ND)

Compound	Blank ID	Sample Identification
	<u>03G477H MB</u>	
Methylene chloride	<u>0.4</u>	
Acetone		
CRQL		

Blank analysis date: _____
 Conc. units: _____ Associated Samples: _____

Compound	Blank ID	Sample Identification
Methylene chloride		
Acetone		
CRQL		

CIRCLED RESULTS WERE NOT QUALIFIED. ALL RESULTS NOT CIRCLED WERE QUALIFIED BY THE FOLLOWING STATEMENT:

Note: Common contaminants such as Methylene chloride, Acetone, 2-Butanone, Carbon disulfide and TICs that were detected in samples within ten times the associated method blank concentration were qualified as not detected, "U". Other contaminants within five times the method blank concentration were also qualified as not detected, "U".

LDC # 11280 D1
 SDG# 03-6034

VALIDATION FINDINGS WORKSHEET
Initial Calibration Calculation Verification

Page: 1 of 1
 Reviewer: Z. Pan
 2nd Reviewer: [Signature]

METHOD: GC / MS

Parameter: _____

Order of regression: 1.00

Date	IS	Fluorobenzene	Methylene chloride	Y	Methylene chloride Conc. (ppb)	X
				Area / (Area IS)		C / (C IS)
10/21/2003	920067		10278	0.0112	0.3	0.03
	916201		48191	0.0526	2	0.2
	919392		203033	0.2208	10	1
	955116		407143	0.4263	20	2
	907151		804880	0.8873	40	4
	1006211		1192483	1.1851	60	6

Regression Output: Regression Output:		Reported
Constant	0.02025	c = 0.02025
Std Err of Y Est	0.04	
R Squared	0.9947	r ² = 0.9947
No. of Observations	6.00	
Degrees of Freedom	4.00	
X Coefficient(s)	0.20119	b(X) 0.20119
Std Err of Coef.	0.01	a(X ²) 0.00000

VALIDATION FINDINGS WORKSHEET
Initial Calibration Calculation Verification

METHOD: GC/MS VOA (EPA Method 524.2)

The Relative Response Factor (RRF), average RRF, and percent relative standard deviation (%RSD) were recalculated for the compounds identified below using the following calculations:

$RRF = (A_x)(C_{is}) / (A_{is})(C_x)$
 average RRF = sum of the RRFs/number of standards
 $\%RSD = 100 * (S/X)$
 A_x = Area of compound, A_{is} = Area of associated internal standard
 C_x = Concentration of compound, C_{is} = Concentration of internal standard
 S = Standard deviation of the RRFs
 X = Mean of the RRFs

#	Standard ID	Calibration Date	Compound (Reference Internal Standard)	Reported		Recalculated		Reported		Recalculated	
				RRF (10 std)	RRF (10 std)	Average RRF (Initial)	Average RRF (Initial)	%RSD	%RSD		
1	E524A003	10-21-03	Methylene chloride (1st Internal Standard)	0.221	0.221	See	Next page				
			Trichloroethene (2nd internal standard)	1.322	1.322	1.354	1.354	5.13	5.12		
			Bromoform (3rd internal standard)	2.614	2.614	2.542	2.542	4.04	4.04		
2			Methylene chloride (1st Internal Standard)								
			Trichloroethene (2nd internal standard)								
			Bromoform (3rd internal standard)								
3			Methylene chloride (1st Internal Standard)								
			Trichloroethene (2nd internal standard)								
			Bromoform (3rd internal standard)								
4			Methylene chloride (1st Internal Standard)								
			Trichloroethene (2nd internal standard)								
			Bromoform (3rd internal standard)								

Comments: Refer to Initial Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 11280 D1
 SDG #: 03-6034

VALIDATION FINDINGS WORKSHEET
Continuing Calibration Results Verification

Page: 1 of 1
 Reviewer: Z. Pan
 2nd Reviewer: A

METHOD: GC/MS VOA (EPA Method 524.2)

The percent difference (%D) of the initial calibration average Relative Response Factors (RRFs) and the continuing calibration RRFs were recalculated for the compounds identified below using the following calculation:

% Difference = $100 * (\text{ave. RRF} - \text{RRF}) / \text{ave. RRF}$
 RRF = $(A_x / C_x) / (A_s / C_s)$
 Where: ave. RRF = initial calibration average RRF
 RRF = continuing calibration RRF
 A_x = Area of compound,
 C_x = Concentration of compound,
 A_s = Area of associated internal standard
 C_s = Concentration of internal standard

#	Standard ID	Calibration Date	Compound (Reference Internal Standard)	Average RRF (Initial)	Reported		Recalculated	
					RRF (CC)	%D	RRF (CC)	%D
1	G477/Q01	11-11-03 (8:21 PM)	Methylene chloride (1st Internal Standard)	20 *	17.58	12.09	17.58	12.1
			Toluene Trichloroethene (2nd internal standard)	1.354	1.209	10.7	1.209	10.7
			Ethyl benzene Bromoform (3rd internal standard)	2.542	2.456	3.4	2.456	3.4
2			Methylene chloride (1st Internal Standard)					
			Trichloroethene (2nd internal standard)					
			Bromoform (3rd internal standard)					
3			Methylene chloride (1st Internal Standard)					
			Trichloroethene (2nd internal standard)					
			Bromoform (3rd internal standard)					
4			Methylene chloride (1st Internal Standard)					
			Trichloroethene (2nd internal standard)					
			Bromoform (3rd internal standard)					

Comments: Refer to Continuing Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

* Conc.

LDC #: 11280 D1
 SDG #: 03-6034

VALIDATION FINDINGS WORKSHEET
Surrogate Results Verification

Page: 1 of 1
 Reviewer: Z. Pan
 2nd reviewer: PA

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) of surrogates were recalculated for the compounds identified below using the following calculation:

% Recovery: SF/SS * 100

Where: SF = Surrogate Found
 SS = Surrogate Spiked

Sample ID: #1

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8	20	20.55	103	103	0
Bromofluorobenzene	↓	21.36	107	107	↓
^{ethene} 1,2-Dichlorobenzene-d4		22.93	115	115	
^{Dibromofluoromethane}		25.57 22.57	113	113	

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

Sample ID: _____

	Surrogate Spiked	Surrogate Found	Percent Recovery	Percent Recovery	Percent Difference
			Reported	Recalculated	
Toluene-d8					
Bromofluorobenzene					
1,2-Dichlorobenzene-d4					

LDC #: 11280 D1
 SDG #: 03-6034

VALIDATION FINDINGS WORKSHEET
Matrix Spike/Matrix Spike Duplicates Results Verification

Page: 1 of 1
 Reviewer: Z. Pan
 2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the matrix spike and matrix spike duplicate were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * (SSC - SC) / SA$ Where: SSC = Spiked sample concentration SC = Sample concentration
 SA = Spike added

RPD = $1 MSC - MSDC | * 2 / (MSC + MSDC)$ MSC = Matrix spike percent recovery MSDC = Matrix spike duplicate percent recovery

MS/MSD sample: # 6/7

Compound	Spike Added (ug/L)		Sample Concentration (ug/L)	Spiked Sample Concentration (ug/L)		Matrix Spike		Matrix Spike Duplicate		MS/MSD	
	MS	MSD		MS	MSD	Percent Recovery		Percent Recovery		Reported	Recalculated
						Reported	Recalc.	Reported	Recalc.	Reported	Recalculated
1,1-Dichloroethene	20	20	0	20.5	20.9	103	103	105	105	2	2
Trichloroethene			9.0	27.1	28.0	91	91	95	95	4	4
Benzene			0	17.0	17.8	85	85	87	87	5	5
Toluene				16.9	17.6	85	85	88	88	3	3
Chlorobenzene			↓	18.2	18.7	91	91	94	94	3	3

Comments: Refer to Matrix Spike/Matrix Spike Duplicates findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 11280 DI
SDG #: 03-6034

VALIDATION FINDINGS WORKSHEET
Laboratory Control Sample Results Verification

Page: 1 of 1
Reviewer: Z. Pan
2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the laboratory control sample and laboratory control sample duplicate (if applicable) were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 \times \text{SSC}/\text{SA}$

Where: SSC = Spiked sample concentration
SA = Spike added

RPD = $100 \times \frac{\text{LCS} - \text{LCSD}}{\text{LCS} + \text{LCSD}}$

LCS = Laboratory control sample percent recovery

LCSD = Laboratory control sample duplicate percent recovery

LCS ID: 9477101

Compound	Spike Added (ug/L)		Spiked Sample Concentration (ug/L)		LCS		LCSD		LCS/LCS D	
	LCS	LCSD	LCS	LCSD	Percent Recovery		Percent Recovery		RPD	
	Reported	Recalc.	Reported	Recalc.	Reported	Recalc.	Reported	Recalc.	Reported	Recalculated
1,1-Dichloroethene	20	/	20.4	/	102	102				
Trichloroethene		/	18.5	/	93	93				
Benzene		/	17.0	/	85	85				
Toluene		/	17.2	/	86	86				
Chlorobenzene		/	18.5	/	93	93				

Comments: Refer to Laboratory Control Sample findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 11290D1
 SDG #: 03-6034

VALIDATION FINDINGS WORKSHEET

Sample Calculation Verification

Page: 1 of 1
 Reviewer: Z. Pan
 2nd reviewer: F

METHOD: GC/MS VOA (EPA Method 524.2)

Y N /N/A
 Y N /N/A

Were all reported results recalculated and verified for all level IV samples?

Were all recalculated results for detected target compounds agree within 10.0% of the reported results?

$$\text{Concentration} = \frac{(A_x)(I_s)(DF)}{(A_s)(RRF)(V_o)(\%S)}$$

- A_x = Area of the characteristic ion (EICP) for the compound to be measured
- A_s = Area of the characteristic ion (EICP) for the specific internal standard
- I_s = Amount of internal standard added in nanograms (ng)
- RRF = Relative response factor of the calibration standard.
- V_o = Volume or weight of sample pruged in milliliters (ml) or grams (g).
- Df = Dilution factor.
- %S = Percent solids, applicable to soils and solid matrices only.

Example:

Sample I.D. #1 → ND

Conc. = () () ()
 () () () ()
 =

#	Sample ID	Compound	Reported Concentration ()	Calculated Concentration ()	Qualification