

**NASA JPL
Data Validation Reports
LDC# 9985**

Wet Chemistry

LDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: January 29, 2003
LDC Report Date: March 27, 2003
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-1369

Sample Identification

EB-1-1/29/03
MW-21-1
MW-21-2
MW-21-3
MW-21-4
MW-21-5
MW-21-5MS
MW-21-5MSD

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-1-1/29/03 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 03-1369

No Sample Data Qualified in this SDG

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-1369

No Sample Data Qualified in this SDG

B

Applied P & Ch Laboratory
Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-442810
 Service ID: 31369

Anal. Method 7196
 Collected by:

Component Name: Chromium (VI)

CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-1369-1	EB-1-1/29/03	Water	01/29/03	01/29/03	01/29/03	03W1295	mg/L	0.01	<0.01	U
03-1369-2	MW-21-1	Water	01/29/03	01/29/03	01/29/03	03W1295	mg/L	0.01	<0.01	U
03-1369-3	MW-21-2	Water	01/29/03	01/29/03	01/29/03	03W1295	mg/L	0.01	<0.01	U
03-1369-4	MW-21-3	Water	01/29/03	01/29/03	01/29/03	03W1295	mg/L	0.01	<0.01	U
03-1369-5	MW-21-4	Water	01/29/03	01/29/03	01/29/03	03W1295	mg/L	0.01	<0.01	U
03-1369-6	MW-21-5	Water	01/29/03	01/29/03	01/29/03	03W1295	mg/L	0.01	<0.01	U
03W1295-MB-01	03W1295-MB-01	Water	01/29/03	01/29/03	01/29/03	03W1295	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Applied P & Ch Laboratory
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc.
Project ID: JPL

Project No: 04-442810
Service ID: 31369

Anal. Method 314.0
Collected by:

Component Name: Perchlorate
CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-1369-1	EB-1-1/29/03	Water	01/29/03	01/29/03	01/29/03	03W1286	µg/L	4	<4	U
03-1369-2	MW-21-1	Water	01/29/03	01/29/03	01/29/03	03W1286	µg/L	4	3.1	B
03-1369-3	MW-21-2	Water	01/29/03	01/29/03	01/29/03	03W1286	µg/L	4	<4	U
03-1369-4	MW-21-3	Water	01/29/03	01/29/03	01/29/03	03W1286	µg/L	4	1.9	B
03-1369-5	MW-21-4	Water	01/29/03	01/29/03	01/29/03	03W1286	µg/L	4	1.8	B
03-1369-6	MW-21-5	Water	01/29/03	01/29/03	01/29/03	03W1286	µg/L	4	<4	U
03W1286-MB-01	03W1286-MB-01	Water	01/29/03	01/29/03	01/29/03	03W1286	µg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten signature/initials
3/20/03

LDC #: 9985B6

VALIDATION COMPLETENESS WORKSHEET

Date: 3-27-03

SDG #: 03-1369

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: MG

2nd Reviewer: MP

METHOD: Hexavalent chromium (EPA SW 846 Method 7196) Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 1-29-03
IIa.	Initial calibration	A	
IIb.	Calibration verification	A	
III.	Blanks	A	
IVa.	Matrix Spike/(Matrix Spike) Duplicates	A	
IVb.	Laboratory control samples	A	LCS / LCSD
V.	Sample result verification	N	
VI.	Overall assessment of data	A	
VII.	Field duplicates	N	
VIII.	Field blanks	ND	EB = 1

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

1	EB-1-1/29/03	W	11		21		31	
2	MW-21-1		12		22		32	
3	MW-21-2		13		23		33	
4	MW-21-3		14		24		34	
5	MW-21-4		15		25		35	
6	MW-21-5		16		26		36	
7	MW-21-5MS		17		27		37	
8	MW-21-5MSD	✓	18		28		38	
9	PBW		19		29		39	
10			20		30		40	

Notes: _____

LDC #: 9985B6

SDG #: 03-1369

VALIDATION FINDINGS WORKSHEET

Sample Specific Analysis Reference

Page: 1 of 1

Reviewer: MG

2nd reviewer: 144

All circled methods are applicable to each sample.

[illegible]

Comments: _____

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: February 3, 2003
LDC Report Date: March 27, 2003
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-1424

Sample Identification

DUPE-2-1Q03
EB-3-2/3/03
MW-19-1
MW-19-2
MW-19-3
MW-19-4
MW-19-5
DUPE-2-1Q03MS
DUPE-2-1Q03MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met.

Instrument detection limits, interelement corrections and linear range analysis were performed at the required frequency with the following exceptions:

Analyte	Calibration	Date of Last Report	Report Frequency Requirement	Date of Analysis	Associated Samples	Flag	A or P
Perchlorate	ICAL	7/31/02	Every 6 months	2/3/03	All samples in SDG 03-1424	None	P

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

Samples DUPE-2-1Q03 and MW-19-4 were identified as field duplicates. No contaminant concentrations were detected in any of the samples.

X. Field Blanks

Sample EB-3-2/3/03 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL**Wet Chemistry - Data Qualification Summary - SDG 03-1424**

SDG	Sample	Analyte	Flag	A or P	Reason
03-1424	DUPE-2-1Q03 EB-3-2/3/03 MW-19-1 MW-19-2 MW-19-3 MW-19-4 MW-19-5	Perchlorate	None	P	Calibration

NASA JPL**Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-1424**

No Sample Data Qualified in this SDG

C

Applied P & Ch Laboratory
Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.
Project ID: JPL

Project No: 04-4428.10
Service ID: 31424

Anal. Method 7196
Collected by:

Component Name: Chromium (VI)
CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-1424-1	DUPE-2-1Q03	Water	02/03/03	02/03/03	02/03/03	03W1348	mg/L	0.01	<0.01	U
03-1424-2	EB-3-2/3/03	Water	02/03/03	02/03/03	02/03/03	03W1348	mg/L	0.01	<0.01	U
03-1424-3	MW-19-1	Water	02/03/03	02/03/03	02/03/03	03W1348	mg/L	0.01	<0.01	U
03-1424-4	MW-19-2	Water	02/03/03	02/03/03	02/03/03	03W1348	mg/L	0.01	<0.01	U
03-1424-5	MW-19-3	Water	02/03/03	02/03/03	02/03/03	03W1348	mg/L	0.01	<0.01	U
03-1424-6	MW-19-4	Water	02/03/03	02/03/03	02/03/03	03W1348	mg/L	0.01	<0.01	U
03-1424-7	MW-19-5	Water	02/03/03	02/03/03	02/03/03	03W1348	mg/L	0.01	<0.01	U
03W1348-MB-01	03W1348-MB-01	Water	02/03/03	02/03/03	02/03/03	03W1348	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten: 3/28/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc.
Project ID: JPL

Project No: 04-4428.10
Service ID: 31424

Anal. Method 314.0
Collected by:

Component Name: Perchlorate
CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-1424-1	DUPE-2-1Q03	Water	02/03/03	02/03/03	02/03/03	03W1345	µg/L	4	<4	U
03-1424-2	EB-3-2/3/03	Water	02/03/03	02/03/03	02/03/03	03W1345	µg/L	4	<4	U
03-1424-3	MW-19-1	Water	02/03/03	02/03/03	02/03/03	03W1345	µg/L	4	<4	U
03-1424-4	MW-19-2	Water	02/03/03	02/03/03	02/03/03	03W1345	µg/L	4	<4	U
03-1424-5	MW-19-3	Water	02/03/03	02/03/03	02/03/03	03W1345	µg/L	4	<4	U
03-1424-6	MW-19-4	Water	02/03/03	02/03/03	02/03/03	03W1345	µg/L	4	<4	U
03-1424-7	MW-19-5	Water	02/03/03	02/03/03	02/03/03	03W1345	µg/L	4	<4	U
03W1345-MB-01	03W1345-MB-01	Water	02/03/03	02/03/03	02/03/03	03W1345	µg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

3/28/03

LDC #: 9985C6

VALIDATION COMPLETENESS WORKSHEET

Date: 3-27-03

SDG #: 03-1424

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: MG

2nd Reviewer: MJA

METHOD: Hexavalent chromium (EPA SW 846 Method 7196) Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 2-3-03
IIa.	Initial calibration	SW	
IIb.	Calibration verification	A	
III.	Blanks	A	
IVa.	Matrix Spike/(Matrix Spike) Duplicates	A	MW-20-3 MS/MSD (SDG: 03-1402)
IVb.	Laboratory control samples	A	LCS/LCSD
V.	Sample result verification	N	
VI.	Overall assessment of data	A	
VII.	Field duplicates	ND	D = 1+6
VIII.	Field blanks	ND	EB = 2

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

1	DUPE-2-1Q03	W	11		21		31	
2	EB-3-2/3/03		12		22		32	
3	MW-19-1		13		23		33	
4	MW-19-2		14		24		34	
5	MW-19-3		15		25		35	
6	MW-19-4		16		26		36	
7	MW-19-5		17		27		37	
8	DUPE-2-1Q03MS		18		28		38	
9	DUPE-2-1Q03MSD		19		29		39	
10	PBW		20		30		40	

Notes: _____

LDC #: 9985C6

SDG #: 03-1424

VALIDATION FINDINGS WORKSHEET

Sample Specific Analysis Reference

Page: 1 of 1

Reviewer: MG

2nd reviewer: Y/W

All circled methods are applicable to each sample.

[illegible]

Comments: _____

Page: 1 of 1
Reviewer: MG
2nd Reviewer: MZ

VALIDATION FINDINGS WORKSHEET

Calibration

METHOD: Inorganics, EPA Method See cover

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Were all instruments calibrated daily, each set-up time, and were the proper number of standards used?	Y	N	N/A
Were all initial and continuing calibration verification percent recoveries (%R) within the control limits of 90-110%?	Y	N	N/A
Are all correlation coefficients >0.995 ?	Y	N	N/A

LEVEL IV/D ONLY:

Y	N	N/A	Were recalculated results acceptable? See Level IV Initial and Continuing Calibration Recalculation Worksheet for recalculations.
Y	N	N/A	Was a balance check conducted prior to the TDS analysis?
Y	N	N/A	Was the titrant normality checked?

[illegible]

Comments:

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: February 4, 2003
LDC Report Date: March 27, 2003
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-1444

Sample Identification

EB-4-2/4/03
MW-17-2
MW-17-3
MW-17-4
MW-17-2MS
MW-17-2MSD

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met.

Instrument detection limits, interelement corrections and linear range analysis were performed at the required frequency with the following exceptions:

Analyte	Calibration	Date of Last Report	Report Frequency Requirement	Date of Analysis	Associated Samples	Flag	A or P
Perchlorate	ICAL	7/31/02	Every 6 months	2/10-2/11/03	All samples in SDG 03-1444	None	P

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

Spike ID (Associated Samples)	Analyte	MS (%R) (Limits)	MSD (%R) (Limits)	RPD (Limits)	Flag	A or P
MW-17-2MS/MSD (All samples in SDG 03-1444)	Hexavalent chromium	76 (78-115)	-	-	J (all detects) UJ (all non-detects)	A

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-4-2/4/03 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL**Wet Chemistry - Data Qualification Summary - SDG 03-1444**

SDG	Sample	Analyte	Flag	A or P	Reason
03-1444	EB-4-2/4/03 MW-17-2 MW-17-3 MW-17-4	Perchlorate	None	P	Calibration
03-1444	EB-4-2/4/03 MW-17-2 MW-17-3 MW-17-4	Hexavalent chromium	J (all detects) UJ (all non-detects)	A	Matrix spike/Matrix spike duplicates (%R)

NASA JPL**Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-1444**

No Sample Data Qualified in this SDG

D

Applied P & Ch Laboratory

Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.
Project ID: JPL

Project No: 04-4428.10
Service ID: 31444

Anal. Method 7196
Collected by:

Component Name: Chromium (VI)

CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-1444-1	EB-4-2/4/03	Water	02/04/03	02/04/03	02/04/03	03W1370	mg/L	0.01	<0.01	U <i>US</i>
03-1444-2	MW-17-2	Water	02/04/03	02/04/03	02/04/03	03W1370	mg/L	0.01	<0.01	U
03-1444-3	MW-17-3	Water	02/04/03	02/04/03	02/04/03	03W1370	mg/L	0.01	<0.01	U
03-1444-4	MW-17-4	Water	02/04/03	02/04/03	02/04/03	03W1370	mg/L	0.01	<0.01	U
03W1370-MB-01	03W1370-MB-01	Water	02/04/03	02/04/03	02/04/03	03W1370	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

2/26/03

Applied P & Ch Laboratory
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc.
Project ID: JPL

Project No: 04-4428.10
Service ID: 31444

Anal. Method 314.0
Collected by:

Component Name: Perchlorate
CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-1444-1	EB-4-2/4/03	Water	02/04/03	02/04/03	02/11/03	03W1424	µg/L	4	<4	U
03-1444-2	MW-17-2	Water	02/04/03	02/04/03	02/10/03	03W1424	µg/L	4	3.4	B
03-1444-3	MW-17-3	Water	02/04/03	02/04/03	02/11/03	03W1424	µg/L	8	145	
03-1444-4	MW-17-4	Water	02/04/03	02/04/03	02/10/03	03W1424	µg/L	4	<4	U
03W1424-MB-01	03W1424-MB-01	Water	02/10/03	02/10/03	02/10/03	03W1424	µg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

3/28/03

LDC #: 9985D6

VALIDATION COMPLETENESS WORKSHEET

Date: 3-27-03

SDG #: 03-1444

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: MG

2nd Reviewer: JMS

METHOD: Hexavalent chromium (EPA SW 846 Method 7196) Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 2-4-03
IIa.	Initial calibration	SW	
IIb.	Calibration verification	A	
III.	Blanks	A	
IVa.	Matrix Spike/(Matrix Spike) Duplicates	SW	
IVb.	Laboratory control samples	A	LCS/LCSD
V.	Sample result verification	N	
VI.	Overall assessment of data	A	
VII.	Field duplicates	N	
VIII.	Field blanks	ND	EB = 1

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

1	EB-4-2/4/03	W	11		21		31	
2	MW-17-2		12		22		32	
3	MW-17-3		13		23		33	
4	MW-17-4		14		24		34	
5	MW-17-2MS		15		25		35	
6	MW-17-2MSD	↓	16		26		36	
7	PBW		17		27		37	
8			18		28		38	
9			19		29		39	
10			20		30		40	

Notes: _____

LDC #: 9985D6

SDG #: 03-1444

VALIDATION FINDINGS WORKSHEET

Sample Specific Analysis Reference

Page: 1 of 1

Reviewer: MG

2nd reviewer: my

All circled methods are applicable to each sample.

[illegible]

Comments: _____

VALIDATION FINDINGS WORKSHEET

Matrix Spike/Matrix Spike Duplicates

Page: 1 of 1
Reviewer: MG
2nd Reviewer: hws

Y N N/A

$$\frac{Y(N)N/\Delta}{Y(N)N/\Delta}$$

DAY 10

of 4 or more, no action was taken.

12/13/51

Y	N	N/A	Were recalculated results acceptable? See Level IV Recalculation Worksheet for recalculations.

[illegible]

Comments:

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: February 5, 2003
LDC Report Date: March 27, 2003
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-1457

Sample Identification

EB-5-2/5/03
MW-18-2
MW-18-3
MW-18-4
MW-18-5
EB-5-2/5/03MS
EB-5-2/5/03MSD

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met.

Instrument detection limits, interelement corrections and linear range analysis were performed at the required frequency with the following exceptions:

Analyte	Calibration	Date of Last Report	Report Frequency Requirement	Date of Analysis	Associated Samples	Flag	A or P
Perchlorate	ICAL	7/31/02	Every 6 months	2/10/03	All samples in SDG 03-1457	None	P

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-5-2/5/03 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL

Wet Chemistry - Data Qualification Summary - SDG 03-1457

SDG	Sample	Analyte	Flag	A or P	Reason
03-1457	EB-5-2/5/03 MW-18-2 MW-18-3 MW-18-4 MW-18-5	Perchlorate	None	P	Calibration

NASA JPL

Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-1457

No Sample Data Qualified in this SDG

E

Applied P & Ch Laboratory
Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.
Project ID: JPL

Project No: 04-4428.10
Service ID: 31457

Anal. Method 7196
Collected by:

Component Name: Chromium (VI)

CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-1457-1	EB-5-2/5/03	Water	02/05/03	02/05/03	02/05/03	03W1379	mg/L	0.01	<0.01	U
03-1457-2	MW-18-2	Water	02/05/03	02/05/03	02/05/03	03W1379	mg/L	0.01	<0.01	U
03-1457-3	MW-18-3	Water	02/05/03	02/05/03	02/05/03	03W1379	mg/L	0.01	<0.01	U
03-1457-4	MW-18-4	Water	02/05/03	02/05/03	02/05/03	03W1379	mg/L	0.01	<0.01	U
03-1457-5	MW-18-5	Water	02/05/03	02/05/03	02/05/03	03W1379	mg/L	0.01	<0.01	U
03W1379-MB-01	03W1379-MB-01	Water	02/05/03	02/05/03	02/05/03	03W1379	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten signature/initials
3/26/03

Applied P & Ch Laboratory
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc.
Project ID: JPL

Project No: 04-4428.10
Service ID: 31457

Anal. Method 314.0
Collected by:

Component Name: Perchlorate
CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-1457-1	EB-5-2/5/03	Water	02/05/03	02/05/03	02/10/03	03W1424	µg/L	4	<4	U
03-1457-2	MW-18-2	Water	02/05/03	02/05/03	02/10/03	03W1424	µg/L	4	<4	U
03-1457-3	MW-18-3	Water	02/05/03	02/05/03	02/10/03	03W1424	µg/L	4	<4	U
03-1457-4	MW-18-4	Water	02/05/03	02/05/03	02/10/03	03W1424	µg/L	4	24.6	
03-1457-5	MW-18-5	Water	02/05/03	02/05/03	02/10/03	03W1424	µg/L	4	<4	U
03W1424-MB-01	03W1424-MB-01	Water	02/10/03	02/10/03	02/10/03	03W1424	µg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

2/27/03

LDC #: 9985E6

VALIDATION COMPLETENESS WORKSHEET

Date: 3-27-03

SDG #: 03-1457

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: MG

2nd Reviewer: Lm

METHOD: Hexavalent chromium (EPA SW 846 Method 7196) Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 2-5-03
IIa.	Initial calibration	SW	
IIb.	Calibration verification	A	
III.	Blanks	A	
IVa.	Matrix Spike/(Matrix Spike) Duplicates	A	MW-4-1 MS/MSD (SDG: 03-1501)
IVb.	Laboratory control samples	A	LCS/LCSD
V.	Sample result verification	N	
VI.	Overall assessment of data	A	
VII.	Field duplicates	N	
VIII.	Field blanks	ND	EB = 1

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

1	EB-5-2/5/03	W	11		21		31	
2	MW-18-2		12		22		32	
3	MW-18-3		13		23		33	
4	MW-18-4		14		24		34	
5	MW-18-5		15		25		35	
6	EB-5-2/5/03MS		16		26		36	
7	EB-5-2/5/03MSD		17		27		37	
8	PBW		18		28		38	
9			19		29		39	
10			20		30		40	

Notes: _____

LDC #: 9985E6
SDG #: 03-1457

VALIDATION FINDINGS WORKSHEET

Sample Specific Analysis Reference

Page: 1 of 1
Reviewer: MG
2nd reviewer: MB

All circled methods are applicable to each sample.

[illegible]

Comments: _____

VALIDATION FINDINGS WORKSHEET

Calibration

Page: 1 of 1
Reviewer: MG
2nd Reviewer: YK

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".	
Were all instruments calibrated daily, each set-up time, and were the proper number of standards used?	Y <u>N</u> N/A
Were all initial and continuing calibration verification percent recoveries (%R) within the control limits of 90-110%?	<u>Y</u> N N/A
Are all correlation coefficients ≥ 0.995 ?	<u>Y</u> N N/A
LEVEL IV/D ONLY:	
Were recalculated results acceptable? See Level IV Initial and Continuing Calibration Recalculation Worksheet for recalculations.	Y N <u>N/A</u>
Was a balance check conducted prior to the TDS analysis?	<u>Y</u> N N/A
Was the titrant normality checked?	Y N <u>N/A</u>

[illegible]

Comments:

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: February 6, 2003
LDC Report Date: March 27, 2003
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-1501

Sample Identification

EB-6-2/6/03
MW-4-1
MW-4-2
MW-4-3
MW-3-2
MW-3-3
MW-3-4
MW-4-1MS
MW-4-1MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate, and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met.

Instrument detection limits, interelement corrections and linear range analysis were performed at the required frequency with the following exceptions:

Analyte	Calibration	Date of Last Report	Report Frequency Requirement	Date of Analysis	Associated Samples	Flag	A or P
Perchlorate	ICAL	7/31/02	Every 6 months	2/10/03	All samples in SDG 03-1501	None	P

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-6-2/6/03 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL**Wet Chemistry - Data Qualification Summary - SDG 03-1501**

SDG	Sample	Analyte	Flag	A or P	Reason
03-1501	EB-6-2/6/03 MW-4-1 MW-4-2 MW-4-3 MW-3-2 MW-3-3 MW-3-4	Perchlorate	None	P	Calibration

NASA JPL**Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-1501**

No Sample Data Qualified in this SDG

F

Applied P & Ch Laboratory
Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 31501

Anal. Method 7196
 Collected by:

Component Name: Chromium (VI)
 CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-1501-1	EB-6-2/6/03	Water	02/06/03	02/06/03	02/06/03	03W1392	mg/L	0.01	<0.01	U
03-1501-2	MW-4-1	Water	02/06/03	02/06/03	02/06/03	03W1392	mg/L	0.01	<0.01	U
03-1501-3	MW-4-2	Water	02/06/03	02/06/03	02/06/03	03W1392	mg/L	0.01	<0.01	U
03-1501-4	MW-4-3	Water	02/06/03	02/06/03	02/06/03	03W1392	mg/L	0.01	<0.01	U
03-1501-6	MW-3-2	Water	02/06/03	02/06/03	02/06/03	03W1392	mg/L	0.01	<0.01	U
03-1501-7	MW-3-3	Water	02/06/03	02/06/03	02/06/03	03W1392	mg/L	0.01	<0.01	U
03-1501-8	MW-3-4	Water	02/06/03	02/06/03	02/06/03	03W1392	mg/L	0.01	<0.01	U
03W1392-MB-01	03W1392-MB-01	Water	02/06/03	02/06/03	02/06/03	03W1392	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

9/3/26/03

Applied P & Ch Laboratory
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc.

Project No: 04-4428.10

Anal. Method 314.0

Project ID: JPL

Service ID: 31501

Collected by:

Component Name: Perchlorate

CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-1501-1	EB-6-2/6/03	Water	02/06/03	02/06/03	02/10/03	03W1424	µg/L	4	<4	U
03-1501-2	MW-4-1	Water	02/06/03	02/06/03	02/10/03	03W1424	µg/L	4	2.9	B
03-1501-3	MW-4-2	Water	02/06/03	02/06/03	02/10/03	03W1424	µg/L	4	<4	U
03-1501-4	MW-4-3	Water	02/06/03	02/06/03	02/10/03	03W1424	µg/L	4	<4	U
03-1501-6	MW-3-2	Water	02/06/03	02/06/03	02/10/03	03W1424	µg/L	4	<4	U
03-1501-7	MW-3-3	Water	02/06/03	02/06/03	02/10/03	03W1424	µg/L	4	<4	U
03-1501-8	MW-3-4	Water	02/06/03	02/06/03	02/10/03	03W1424	µg/L	4	<4	U
03W1424-MB-01	03W1424-MB-01	Water	02/10/03	02/10/03	02/10/03	03W1424	µg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

3/26/03

LDC #: 9985F6

VALIDATION COMPLETENESS WORKSHEET

Date: 3-27-03

SDG #: 03-1501

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: MG

2nd Reviewer: MW

METHOD: Hexavalent chromium (EPA SW 846 Method 7196) Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 2-6-03
IIa.	Initial calibration	SW	
IIb.	Calibration verification	A	
III.	Blanks	A	
IVa.	Matrix Spike/(Matrix Spike) Duplicates	A	
IVb.	Laboratory control samples	A	LCS/LCSD
V.	Sample result verification	N	
VI.	Overall assessment of data	A	
VII.	Field duplicates	N	
VIII.	Field blanks	ND	EB = 1

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

1	EB-6-2/6/03	W	11		21		31	
2	MW-4-1		12		22		32	
3	MW-4-2		13		23		33	
4	MW-4-3		14		24		34	
5	MW-3-2		15		25		35	
6	MW-3-3		16		26		36	
7	MW-3-4		17		27		37	
8	MW-4-1MS		18		28		38	
9	MW-4-1MSD		19		29		39	
10	PBW		20		30		40	

Notes: _____

LDC #: 9985 FL

SDG #: 03-1501

VALIDATION FINDINGS WORKSHEET

Sample Specific Analysis Reference

Page: 1 of 1

Reviewer: MG

2nd reviewer: 

All circled methods are applicable to each sample.

[illegible]

Comments: _____

VALIDATION FINDINGS WORKSHEET

Calibration

LDC #: 9985F6
SDG #: 03-1501

METHOD: Inorganics, EPA Method See cover

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y	☐	N/A
☒	N	N/A
☐	N	N/A

LEVEL IV/D ONLY:

LEVEL 1750 ONLY	Were recalculated results acceptable? See Level IV Initial and Continuing Calibration Recalculation Worksheet for recalculations.		
Y	N	N/A	Was a balance check conducted prior to the TDS analysis?
Y	N	N/A	Was the titrant normality checked?

[illegible]

Comments:

TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method TO-14/TO-14A)

A. Chloromethane*	S. Trichloroethene	KK. Trichlorofluoromethane	CCC. tert-Butylbenzene	UUU. Benzyl chloride
B. Bromomethane	T. Dibromochloromethane	LL. Methyl-tert-butyl ether	DDD. 1,2,4-Trimethylbenzene	VVV. 4-Ethyltoluene
C. Vinyl chloride**	U. 1,1,2-Trichloroethane	MM. 1,2-Dibromo-3-chloropropane	EEE. sec-Butylbenzene	WWW. Ethanol
D. Chloroethane	V. Benzene	NN. Diethyl ether	FFF. 1,3-Dichlorobenzene	XXX. Ethyl ether
E. Methylene chloride	W. trans-1,3-Dichloropropene	OO. 2,2-Dichloropropane	GGG. p-Isopropyltoluene	YYY. tert-Butanol
F. Acetone	X. Bromoform*	PP. Bromochloromethane	HHH. 1,4-Dichlorobenzene	ZZZ. tert-Butyl alcohol
G. Carbon disulfide	Y. 4-Methyl-2-pentanone	QQ. 1,1-Dichloropropene	III. n-Butylbenzene	AAA. Ethyl tert-butyl ether
H. 1,1-Dichloroethene**	Z. 2-Hexanone	RR. Dibromomethane	JJJ. 1,2-Dichlorobenzene	BBB. tert-Amyl methyl ether
I. 1,1-Dichloroethane*	AA. Tetrachloroethene	SS. 1,3-Dichloropropane	KKK. 1,2,4-Trichlorobenzene	CCC. 1-Chlorohexane
J. 1,2-Dichloroethene, total	BB. 1,1,2,2-Tetrachloroethane*	TT. 1,2-Dibromoethane	LLL. Hexachlorobutadiene	DDD. Isopropyl alcohol
K. Chloroform**	CC. Toluene**	UU. 1,1,1,2-Tetrachloroethane	MMM. Naphthalene	EEE. Acetonitrile
L. 1,2-Dichloroethane	DD. Chlorobenzene*	VV. Isopropylbenzene	NNN. 1,2,3-Trichlorobenzene	FFF. Acrolein
M. 2-Butanone	EE. Ethylbenzene**	WW. Bromobenzene	OOO. 1,3,5-Trichlorobenzene	GGG. Acrylonitrile
N. 1,1,1-Trichloroethane	FF. Styrene	XX. 1,2,3-Trichloropropane	PPP. trans-1,2-Dichloroethene	HHH. 1,4-Dioxane
O. Carbon tetrachloride	GG. Xylenes, total	YY. n-Propylbenzene	QQQ. cis-1,2-Dichloroethene	III. Isobutyl alcohol
P. Bromodichloromethane	HH. Vinyl acetate	ZZ. 2-Chlorotoluene	RRR. m,p-Xylenes	JJJ. Methacrylonitrile
Q. 1,2-Dichloropropane**	II. 2-Chloroethylvinyl ether	AAA. 1,3,5-Trimethylbenzene	SSS. o-Xylene	KKK. Propionitrile
R. cis-1,3-Dichloropropene	JJ. Dichlorodifluoromethane	BBB. 4-Chlorotoluene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane	LLL

= System performance check compounds (SPCC) for RRF ; ** = Calibration check compounds (CCC) for %RSD.

VALIDATION FINDINGS WORKSHEET

Initial Calibration

Page: 1 of 1
Reviewer: DM
2nd Reviewer: _____

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

Y/N N/A Did the laboratory perform a 5 point calibration prior to sample analysis?

Were all percent relative standard deviations (%RSD) $\leq 30\%$ and relative response factors (RRF) ≥ 0.05 ?

[illegible]

VALIDATION FINDINGS WORKSHEET

Laboratory Control Samples (LCS)

Page: 1 of 1
Reviewer: DM
2nd Reviewer: K

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

(Y) N N/A

~~YCN)N/A~~[illegible]



LABORATORY DATA CONSULTANTS, INC.

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

Geofon, Inc.
22632 Golden Springs Drive, Suite 270
Diamond Bar, CA 91765
ATTN: Mr. Leo Williamson

March 28, 2003

SUBJECT: NASA JPL, DO #01, Data Validation

Dear Mr. Williamson,

Enclosed are the final validation reports for the fractions listed below. These SDGs were received on March 19, 2003. Attachment 1 is a summary of the samples that were reviewed for each analysis.

LDC Project # 10006:

<u>SDG #</u>	<u>Fraction</u>
03-1402, 03-1425, 03-1534	Volatiles (524.2), Chromium, Wet Chemistry

The data validation was performed under EPA Level III guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, February 1994
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996

Please feel free to contact us if you have any questions.

Sincerely,

Erlinda T. Rauto
Operations Manager/Senior Chemist

LDC #10006 (Geofon, Inc.-Diamond Bar / NASA Jet Propulsion Laboratory, DO#001)

[illegible]

Shaded cells indicate Level IV validation (all other cells are Level III validation).

**NASA JPL
Data Validation Reports
LDC# 10006**

Volatiles

LDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: January 30, 2003
LDC Report Date: March 27, 2003
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-1402

Sample Identification

DUPE-1-1Q03
EB-2-1/30/03
MW-20-1
MW-20-2
MW-20-3
MW-20-4
MW-20-5
TB-2-1/30/03
MW-20-3MS
MW-20-3MSD

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples DUPE-1-1Q03 and MW-20-1 were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-2-1/30/03 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-2-1/30/03 was identified as an equipment blank. No volatile contaminants were found in this blank.

NASA JPL

Volatiles - Data Qualification Summary - SDG 03-1402

No Sample Data Qualified in this SDG

NASA JPL

Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-1402

No Sample Data Qualified in this SDG

10006A

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 01/30/2003
Project ID: JPL	Service ID: 31402	Collected by:
Sample ID: DUPE-1-1Q03	Lab Sample ID: 03-1402-1	Received Date: 01/30/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: G
Batch No: 03G1286	Prep. Date: 02/03/03	Anal. Date: 02/03/03
Data File Name: 1402-01	Prep. No: -	Anal. Time: 17:50
Methanol Vol: -	Sample Amount: 25 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	µg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	0.3	J
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE	98-82-8	µg/L	0.5	<0.5	U

Handwritten signature/initials

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE (PCE)	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE (TCE)	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

Control Limit, %

Surro. Rec.%

1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4
2	DIBROMOFLUOROMETHANE	1868-53-7
3	1,2-DICHLOROETHANE-D4	17060-07-0
4	TOLUENE-D8	2037-26-5

70-129

107

70-129

99

70-122

90

73-129

102

of out-of-control

0

Internal Standard

Control Limit, %

IS Rec.%

1	CHLOROBENZENE-D5	3114-55-4
2	1,4-DICHLOROBENZENE-D4	3855-82-1
3	FLUOROBENZENE	462-06-6

50-200

136

50-200

136

50-200

140

of out-of-control

0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

3/28/03

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 01/30/2003
Project ID: JPL	Service ID: 31402	Collected by:
Sample ID: EB-2-1/30/03	Lab Sample ID: 03-1402-2	Received Date: 01/30/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: G
Batch No: 03G1286	Prep. Date: 02/03/03	Anal. Date: 02/03/03
Data File Name: 1402-02	Prep. No: -	Anal. Time: 18:19
Methanol Vol: -	Sample Amount: 25 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	< 0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	< 0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	< 0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	< 0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	< 0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	< 0.5	U
7	2-BUTANONE	78-93-3	µg/L	10	< 10	U
8	N-BUTYLBENZENE	104-51-8	µg/L	0.5	< 0.5	U
9	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	< 0.5	U
10	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	< 0.5	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	< 0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	< 0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	< 0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	< 0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	< 0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	< 0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	< 0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	< 0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	< 0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	< 0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	< 0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	< 0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	< 0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	< 0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	< 0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	< 0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	< 0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	< 0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	< 0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	< 0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	< 0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	< 0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	< 0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	< 0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	< 0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	< 0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	< 0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	< 0.5	U
39	ISOPROPYLBENZENE	98-82-8	µg/L	0.5	< 0.5	U

3/28/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE (PCE)	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE (TCE)	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

Control Limit, %

Surro. Rec.%

1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4
2	DIBROMOFLUOROMETHANE	1868-53-7
3	1,2-DICHLOROETHANE-D4	17060-07-0
4	TOLUENE-D8	2037-26-5

70-129

112

70-129

105

70-122

92

73-129

110

of out-of-control

0

Internal Standard

Control Limit, %

IS Rec.%

1	CHLOROBENZENE-D5	3114-55-4
2	1,4-DICHLOROBENZENE-D4	3855-82-1
3	FLUOROBENZENE	462-06-6

50-200

127

50-200

131

50-200

128

of out-of-control

0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

3/28/03

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 01/30/2003
Project ID: JPL	Service ID: 31402	Collected by:
Sample ID: MW-20-1	Lab Sample ID: 03-1402-3	Received Date: 01/30/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: G
Batch No: 03G1286	Prep. Date: 02/03/03	Anal. Date: 02/03/03
Data File Name: 1402-03	Prep. No: -	Anal. Time: 18:48
Methanol Vol: -	Sample Amount: 25 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	µg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE	98-82-8	µg/L	0.5	<0.5	U

3/28/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE (PCE)	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE (TCE)	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

Control Limit, % Surro. Rec.%

1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4
2	DIBROMOFLUOROMETHANE	1868-53-7
3	1,2-DICHLOROETHANE-D4	17060-07-0
4	TOLUENE-D8	2037-26-5

70-129	105
70-129	98
70-122	87
73-129	101

of out-of-control

0

Internal Standard

Control Limit, % IS Rec.%

1	CHLOROBENZENE-D5	3114-55-4
2	1,4-DICHLOROBENZENE-D4	3855-82-1
3	FLUOROBENZENE	462-06-6

50-200	133
50-200	135
50-200	132

of out-of-control

0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 01/30/2003
Project ID: JPL	Service ID: 31402	Collected by:
Sample ID: MW-20-2	Lab Sample ID: 03-1402-4	Received Date: 01/30/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: G
Batch No: 03G1286	Prep. Date: 02/03/03	Anal. Date: 02/03/03
Data File Name: 1402-04	Prep. No: -	Anal. Time: 19:16
Methanol Vol. -	Sample Amount: 25 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	µg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	1.3	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE	98-82-8	µg/L	0.5	<0.5	U

3/28/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE (PCE)	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE (TCE)	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

Control Limit, %

Surro. Rec.%

1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4
2	DIBROMOFLUOROMETHANE	1868-53-7
3	1,2-DICHLOROETHANE-D4	17060-07-0
4	TOLUENE-D8	2037-26-5

70-129

100

70-129

93

70-122

84

73-129

96

of out-of-control

0

Internal Standard

Control Limit, %

IS Rec.%

1	CHLOROBENZENE-D5	3114-55-4
2	1,4-DICHLOROBENZENE-D4	3855-82-1
3	FLUOROBENZENE	462-06-6

50-200

144

50-200

144

50-200

142

of out-of-control

0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 01/30/2003
Project ID: JPL	Service ID: 31402	Collected by:
Sample ID: MW-20-3	Lab Sample ID: 03-1402-5	Received Date: 01/30/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: G
Batch No: 03G1286	Prep. Date: 02/03/03	Anal. Date: 02/03/03
Data File Name: 1402-05	Prep. No: -	Anal. Time: 19:45
Methanol Vol. -	Sample Amount: 25 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	µg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE	98-82-8	µg/L	0.5	<0.5	U

3/28/03

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE (PCE)	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE (TCE)	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
Surrogates				Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL	460-00-4		70-129	100	
2	DIBROMOFLUOROMETHANE	1868-53-7		70-129	100	
3	1,2-DICHLOROETHANE-D4	17060-07-0		70-122	90	
4	TOLUENE-D8	2037-26-5		73-129	104	
# of out-of-control					0	
Internal Standard				Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4		50-200	137	
2	1,4-DICHLOROBENZENE-D4	3855-82-1		50-200	138	
3	FLUOROBENZENE	462-06-6		50-200	130	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

3/28/03

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 01/30/2003
Project ID: JPL	Service ID: 31402	Collected by:
Sample ID: MW-20-4	Lab Sample ID: 03-1402-6	Received Date: 01/30/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: G
Batch No: 03G1286	Prep. Date: 02/03/03	Anal. Date: 02/03/03
Data File Name: 1402-06	Prep. No: -	Anal. Time: 20:15
Methanol Vol. -	Sample Amount: 25 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	µg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE (PCE)	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE (TCE)	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

Control Limit, % Surro. Rec. %

1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4
2	DIBROMOFLUOROMETHANE	1868-53-7
3	1,2-DICHLOROETHANE-D4	17060-07-0
4	TOLUENE-D8	2037-26-5

70-129	110
70-129	99
70-122	89
73-129	102

of out-of-control

0

Internal Standard

Control Limit, % IS Rec. %

1	CHLOROBENZENE-D5	3114-55-4
2	1,4-DICHLOROBENZENE-D4	3855-82-1
3	FLUOROBENZENE	462-06-6

50-200	131
50-200	131
50-200	135

of out-of-control

0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 01/30/2003
Project ID: JPL	Service ID: 31402	Collected by:
Sample ID: MW-20-5	Lab Sample ID: 03-1402-7	Received Date: 01/30/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: G
Batch No: 03G1286	Prep. Date: 02/03/03	Anal. Date: 02/03/03
Data File Name: 1402-07	Prep. No: -	Anal. Time: 20:43
Methanol Vol. -	Sample Amount: 25 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	µg/L	10	3	J
8	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	0.6	
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE (PCE)	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE (TCE)	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

Control Limit, % Surro. Rec. %

1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	106
2	DIBROMOFLUOROMETHANE	1868-53-7	70-129	104
3	1,2-DICHLOROETHANE-D4	17060-07-0	70-122	92
4	TOLUENE-D8	2037-26-5	73-129	107

of out-of-control

0

Internal Standard

Control Limit, % IS Rec. %

1	CHLOROBENZENE-D5	3114-55-4	50-200	125
2	1,4-DICHLOROBENZENE-D4	3855-82-1	50-200	131
3	FLUOROBENZENE	462-06-6	50-200	125

of out-of-control

0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 01/30/2003
Project ID: JPL	Service ID: 31402	Collected by:
Sample ID: TB-2-1/30/03	Lab Sample ID: 03-1402-8	Received Date: 01/30/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: G
Batch No: 03G1286	Prep. Date: 02/03/03	Anal. Date: 02/03/03
Data File Name: 1402-08	Prep. No: -	Anal. Time: 21:11
Methanol Vol: -	Sample Amount: 25 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	2-BUTANONE	78-93-3	µg/L	10	<10	U
8	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
9	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
10	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	0.5	<0.5	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	0.5	<0.5	U
42	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	<10	U
43	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE (PCE)	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE (TCE)	79-01-6	µg/L	0.5	<0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	TRICHLOROTRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
58	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U
Surrogates				Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4		70-129	101	
2	DIBROMOFLUOROMETHANE	1868-53-7		70-129	93	
3	1,2-DICHLOROETHANE-D4	17060-07-0		70-122	85	
4	TOLUENE-D8	2037-26-5		73-129	97	
# of out-of-control					0	
Internal Standard				Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4		50-200	139	
2	1,4-DICHLOROBENZENE-D4	3855-82-1		50-200	140	
3	FLUOROBENZENE	462-06-6		50-200	140	
# of out-of-control					0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank
D - Diluted

Handwritten signature
3/28/03

LDC #: 10006A1 **VALIDATION COMPLETENESS WORKSHEET**
SDG #: 03-1402 Level III
Laboratory: Applied P & Ch Laboratory

Date: 3-21-03
Page: 1 of 1
Reviewer: DM
2nd Reviewer:

METHOD: GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 1-30-03
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	%RSD, r ²
IV.	Continuing calibration	A	
V.	Blanks	A	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	SW	D = 1 + 3
XVII.	Field blanks	ND	TB = 8 EB = 2

Note: A = Acceptable ND = No compounds detected D = Duplicate
N = Not provided/applicable R = Rinsate TB = Trip blank
SW = See worksheet FB = Field blank EB = Equipment blank

Validated Samples: All H₂O's

1	DUPE-1-1Q03	11	03G1286 Mbol	21		31	
2	EB-2-1/30/03	12		22		32	
3	MW-20-1	13		23		33	
4	MW-20-2	14		24		34	
5	MW-20-3	15		25		35	
6	MW-20-4	16		26		36	
7	MW-20-5	17		27		37	
8	TB-2-1/30/03	18		28		38	
9	MW-20-3MS	19		29		39	
10	MW-20-3MSD	20		30		40	

TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method 524.2)

A. Chloromethane	Q. 1,2-Dichloropropane	GG. Xylenes, total	WW. Bromobenzene	MMM. Naphthalene
B. Bromomethane	R. cis-1,3-Dichloropropene	HH. Vinyl acetate	XX. 1,2,3-Trichloropropane	NNN. 1,2,3-Trichlorobenzene
C. Vinyl chloride	S. Trichloroethene	II. 2-Chloroethylvinyl ether	YY. n-Propylbenzene	OOO. 1,3,5-Trichlorobenzene
D. Chloroethane	T. Dibromochloromethane	JJ. Dichlorodifluoromethane	ZZ. 2-Chlorotoluene	PPP. trans-1,2-Dichloroethene
E. Methylene chloride	U. 1,1,2-Trichloroethane	KK. Trichlorofluoromethane	AAA. 1,3,5-Trimethylbenzene	QQQ. cis-1,2-Dichloroethene
F. Acetone	V. Benzene	LL. Methyl-tert-butyl ether	BBB. 4-Chlorotoluene	RRR. m,p-Xylenes
G. Carbon disulfide	W. trans-1,3-Dichloropropene	MM. 1,2-Dibromo-3-chloropropane	CCC. tert-Butylbenzene	SSS. o-Xylene
H. 1,1-Dichloroethene	X. Bromoform	NN. Diethyl ether	DDD. 1,2,4-Trimethylbenzene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane
I. 1,1-Dichloroethane	Y. 4-Methyl-2-pentanone	OO. 2,2-Dichloropropane	EEE. sec-Butylbenzene	UUU. Benzyl chloride
J. 1,2-Dichloroethene, total	Z. 2-Hexanone	PP. Bromochloromethane	FFF. 1,3-Dichlorobenzene	VVV. 4-Ethyltoluene
K. Chloroform	AA. Tetrachloroethene	QQ. 1,1-Dichloropropene	GGG. p-Isopropyltoluene	WWW. Ethanol
L. 1,2-Dichloroethane	BB. 1,1,2,2-Tetrachloroethane	RR. Dibromomethane	HHH. 1,4-Dichlorobenzene	XXX. Ethyl ether
M. 2-Butanone	CC. Toluene	SS. 1,3-Dichloropropane	III. n-Butylbenzene	
N. 1,1,1-Trichloroethane	DD. Chlorobenzene	TT. 1,2-Dibromoethane	JJJ. 1,2-Dichlorobenzene	
O. Carbon tetrachloride	EE. Ethylbenzene	UU. 1,1,1,2-Tetrachloroethane	KKK. 1,2,4-Trichlorobenzene	
P. Bromodichloromethane	FF. Styrene	VV. Isopropylbenzene	LLL. Hexachlorobutadiene	

Notes:

LDC #: 10006A1
SDG #: 03-1402

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
Reviewer: DM
2nd reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

☒ Y ☐ N ☐ N/A
☒ Y ☐ N ☐ N/A

Were field duplicate pairs identified in this SDG?

Were target compounds detected in the field duplicate pairs?

Compound	Concentration (ug/L)		RPD
	1	3	
K	0.3	0.5u	—

Compound	Concentration ()		RPD

Compound	Concentration ()		RPD