

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 8, 2003
LDC Report Date: June 17, 2003
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 03-3130

Sample Identification

EB-12-5/8/03
MW-22-1
MW-22-2
MW-22-3
MW-22-4
MW-22-5
MW-22-1MS
MW-22-1MSD
MW-22-1DUP

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7 and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Iron, Magnesium, Potassium, and Sodium.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met with the following exceptions:

Date	Lab. Reference/ID	Analyte	%R (Limits)	Associated Samples	Flag	A or P
5/13/03	CCV (12:27)	Arsenic	113.2 (90-110)	EB-12-5/8/03 MW-22-1	J (all detects)	P
5/13/03	CCV (13:42)	Arsenic	111.2 (90-110)	MW-22-2 MW-22-3 MW-22-4 MW-22-5	J (all detects)	P

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Calcium Potassium Sodium	117 ug/L 64.7 ug/L 1040 ug/L	All samples in SDG 03-3130
ICB/CCB	Calcium Iron Magnesium Potassium Sodium	227.84 ug/L 19.77 ug/L 25.97 ug/L 115.69 ug/L 1314.67 ug/L	All samples in SDG 03-3130

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
EB-12-5/8/03	Iron Magnesium Potassium Sodium	31.4 ug/L 20.7 ug/L 93.5 ug/L 1240 ug/L	31.4U ug/L 20.7U ug/L 93.5U ug/L 1240U ug/L
MW-22-2	Iron	81.3 ug/L	81.3U ug/L
MW-22-4	Iron	53.8 ug/L	53.8U ug/L
MW-22-5	Iron	57.5 ug/L	57.5U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

Spike ID (Associated Samples)	Analyte	MS (%R) (Limits)	MSD (%R) (Limits)	RPD (Limits)	Flag	A or P
MW-22-1MS/MSD (All samples in SDG 03-3130)	Magnesium	147 (75-125)	-	-	J (all detects)	A

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards (ICP-MS)

ICP-MS was not utilized in this SDG.

IX. Furnace Atomic Absorption QC

Raw data were not reviewed for this SDG.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-22-1L	Magnesium Potassium	10.8 (≤ 10) 11.5 (≤ 10)	All samples in SDG 03-3130	J (all detects) J (all detects)	A

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-12-5/8/03 was identified as an equipment blank. No metal contaminants were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration (ug/L)
EB-12-5/8/03	Iron Magnesium Potassium Sodium	31.4 20.7 93.5 1240

**NASA JPL
Metals - Data Qualification Summary - SDG 03-3130**

SDG	Sample	Analyte	Flag	A or P	Reason
03-3130	EB-12-5/8/03 MW-22-1 MW-22-2 MW-22-3 MW-22-4 MW-22-5	Arsenic	J (all detects)	P	Calibration verification (%R)
03-3130	EB-12-5/8/03 MW-22-1 MW-22-2 MW-22-3 MW-22-4 MW-22-5	Magnesium	J (all detects)	A	Matrix spike/Matrix spike duplicates (%R)
03-3130	EB-12-5/8/03 MW-22-1 MW-22-2 MW-22-3 MW-22-4 MW-22-5	Magnesium Potassium	J (all detects) J (all detects)	A	ICP serial dilution (%D)

**NASA JPL
Metals - Laboratory Blank Data Qualification Summary - SDG 03-3130**

SDG	Sample	Analyte	Modified Final Concentration	A or P
03-3130	EB-12-5/8/03	Iron Magnesium Potassium Sodium	31.4U ug/L 20.7U ug/L 93.5U ug/L 1240U ug/L	A
03-3130	MW-22-2	Iron	81.3U ug/L	A
03-3130	MW-22-4	Iron	53.8U ug/L	A
03-3130	MW-22-5	Iron	57.5U ug/L	A

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
Project ID: JPL GW Mon-2Q03

Project No: 04-4428.10
Service ID: 33130
Lab Sample ID: 03-3130-1
Sample Matrix: Water

Collection Date: 05/08/2003
Collected by:
Received Date: 05/08/2003
Moisture %: -

Sample ID: **EB-12-5/8/03**
Sample Type: Field Sample

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	µg/L	5	< 5	U	F		03M1458E	05/13/03	05/13/03	1	200.9
CALCIUM	7440-70-2	µg/L	200	< 200	U	P		03M1452L	05/12/03	05/12/03	1	200.7
IRON	7439-89-6	µg/L	50	31.4		B		03M1452L	05/12/03	05/12/03	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	20.7		B		03M1452L	05/12/03	05/12/03	1	200.7
POTASSIUM	7440-09-7	µg/L	400	93.5		B		03M1452L	05/12/03	05/12/03	1	200.7
SODIUM	7440-23-5	µg/L	2000	1240		B		03M1452L	05/12/03	05/12/03	1	200.7

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor

C Qualifier: U - Not Detected or less than IDL

B - Less than RL (PQL, EQL or CRDL), but greater than IDL.

Q Qualifier: N - Spike recovery out of control

* - Duplicate analysis out of control

W - Post digestion spike for GFAA out of control

E - Serial dilution difference out of control

M Qualifier: P - ICP

A - FLAA

F - GFAA

CV - Cold Vapor

Handwritten signature: J. G. Geron
6/18/03

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL GW Mon-2Q03

Project No: 04-4428.10
 Service ID: 33130
 Lab Sample ID: 03-3130-2
 Sample Matrix: Water

Collection Date: 05/08/2003
 Collected by:
 Received Date: 05/08/2003
 Moisture %: -

Sample ID: MW-22-1
 Sample Type: Field Sample

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	µg/L	5	<5	U	F		03M1458E	05/13/03	05/13/03	1	200.9
CALCIUM	7440-70-2	µg/L	200	108000		P		03M1452L	05/12/03	05/12/03	1	200.7
IRON	7439-89-6	µg/L	50	115		P		03M1452L	05/12/03	05/12/03	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	39100	J	P		03M1452L	05/12/03	05/12/03	1	200.7
POTASSIUM	7440-09-7	µg/L	400	3110	J	P		03M1452L	05/12/03	05/12/03	1	200.7
SODIUM	7440-23-5	µg/L	2000	30700		P		03M1452L	05/12/03	05/12/03	1	200.7

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor

C Qualifier: U - Not Detected or less than IDL

B - Less than RL (PQL, EQL or CRDL), but greater than IDL.

Q Qualifier: N - Spike recovery out of control

* - Duplicate analysis out of control

W - Post digestion spike for GFAA out of control

E - Serial dilution difference out of control

M Qualifier: P - ICP

A - FLAA

F - GFAA

CV - Cold Vapor

6/15/03

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL GW Mon-2Q03

Project No: 04-4428.10
 Service ID: 33130
 Lab Sample ID: 03-3130-3
 Sample Matrix: Water

Collection Date: 05/08/2003
 Collected by:
 Received Date: 05/08/2003
 Moisture %: -

Sample ID: MW-22-2
 Sample Type: Field Sample

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	µg/L	5	< 5	U	F		03M1458E	05/13/03	05/13/03	1	200.9
CALCIUM	7440-70-2	µg/L	200	35000		P		03M1452L	05/12/03	05/12/03	1	200.7
IRON	7439-89-6	µg/L	50	81.3 U		P		03M1452L	05/12/03	05/12/03	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	16800 J		P		03M1452L	05/12/03	05/12/03	1	200.7
POTASSIUM	7440-09-7	µg/L	400	2360 J		P		03M1452L	05/12/03	05/12/03	1	200.7
SODIUM	7440-23-5	µg/L	2000	30400		P		03M1452L	05/12/03	05/12/03	1	200.7

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor

C Qualifier: U - Not Detected or less than IDL

B - Less than RL (PQL, EQL or CRDL), but greater than IDL.

Q Qualifier: N - Spike recovery out of control

* - Duplicate analysis out of control

W - Post digestion spike for GFAA out of control

E - Serial dilution difference out of control

M Qualifier: P - ICP

A - FLAA

F - GFAA

CV - Cold Vapor

6/15/03

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL GW Mon-2Q03
 Sample ID: MW-22-3
 Sample Type: Field Sample

Project No: 04-4428.10
 Service ID: 33130
 Lab Sample ID: 03-3130-4
 Sample Matrix: Water

Collection Date: 05/08/2003
 Collected by:
 Received Date: 05/08/2003
 Moisture %: -

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	µg/L	5	< 5	U	F		03M1458E	05/13/03	05/13/03	1	200.9
CALCIUM	7440-70-2	µg/L	200	29500		P		03M1452L	05/12/03	05/12/03	1	200.7
IRON	7439-89-6	µg/L	50	130		P		03M1452L	05/12/03	05/12/03	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	14500	J	P		03M1452L	05/12/03	05/12/03	1	200.7
POTASSIUM	7440-09-7	µg/L	400	2360	J	P		03M1452L	05/12/03	05/12/03	1	200.7
SODIUM	7440-23-5	µg/L	2000	36800		P		03M1452L	05/12/03	05/12/03	1	200.7

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: P - ICP A - FLAA F - GFAA CV - Cold Vapor

6/8/03

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL GW Mon-2Q03

Project No: 04-4428.10
 Service ID: 33130
 Lab Sample ID: 03-3130-5
 Sample Matrix: Water

Collection Date: 05/08/2003
 Collected by:
 Received Date: 05/08/2003
 Moisture %: -

Sample ID: MW-22-4
 Sample Type: Field Sample

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	µg/L	5	< 5	U	F		03M1458E	05/13/03	05/13/03	1	200.9
CALCIUM	7440-70-2	µg/L	200	29400		P		03M1452L	05/12/03	05/12/03	1	200.7
IRON	7439-89-6	µg/L	50	53.8		P		03M1452L	05/12/03	05/12/03	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	8450		P		03M1452L	05/12/03	05/12/03	1	200.7
POTASSIUM	7440-09-7	µg/L	400	1700		P		03M1452L	05/12/03	05/12/03	1	200.7
SODIUM	7440-23-5	µg/L	2000	26900		P		03M1452L	05/12/03	05/12/03	1	200.7

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor

C Qualifier: U - Not Detected or less than IDL

B - Less than RL (PQL, EQL or CRDL), but greater than IDL.

Q Qualifier: N - Spike recovery out of control

* - Duplicate analysis out of control

W - Post digestion spike for GFAA out of control

E - Serial dilution difference out of control

M Qualifier: P - ICP

A - FLAA

F - GFAA

CV - Cold Vapor

M/6/10/03

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL GW Mon-2Q03
 Sample ID: MW-22-5
 Sample Type: Field Sample

Project No: 04-4428.10
 Service ID: 33130
 Lab Sample ID: 03-3130-6
 Sample Matrix: Water

Collection Date: 05/08/2003
 Collected by:
 Received Date: 05/08/2003
 Moisture %: -

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	µg/L	5	<5	U	F		03M1458E	05/13/03	05/13/03	1	200.9
CALCIUM	7440-70-2	µg/L	200	4580		P		03M1452L	05/12/03	05/12/03	1	200.7
IRON	7439-89-6	µg/L	50	57.5		P		03M1452L	05/12/03	05/12/03	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	625		P		03M1452L	05/12/03	05/12/03	1	200.7
POTASSIUM	7440-09-7	µg/L	400	1130		P		03M1452L	05/12/03	05/12/03	1	200.7
SODIUM	7440-23-5	µg/L	2000	70200		P		03M1452L	05/12/03	05/12/03	1	200.7

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: P - ICP A - FLAA F - GFAA CV - Cold Vapor

6/18/03

LDC #: 1041404

VALIDATION COMPLETENESS WORKSHEET

Date: 6/17/03

SDG #: 03-3130

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: [Signature]

2nd Reviewer: [Signature]

METHOD: Metals (EPA Method 200.7 & 200.9)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 5/8/03
II.	Calibration	SW	
III.	Blanks	SW	
IV.	ICP Interference Check Sample (ICS) Analysis	A	
V.	Matrix Spike Analysis	SW	
VI.	Duplicate Sample Analysis	A	
VII.	Laboratory Control Samples (LCS)	A	LCS/LCSD
VIII.	Internal Standard (ICP-MS)	N	Not Utilized
IX.	Furnace Atomic Absorption QC	A	MSA was not performed
X.	ICP Serial Dilution	SW	
XI.	Sample Result Verification	N	
XII.	Overall Assessment of Data	A	
XIII.	Field Duplicates	N	
XIV.	Field Blanks	SW	EB=1

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples:

AA

1	EB-12-5/8/03	11		21		31	
2	MW-22-1	12		22		32	
3	MW-22-2	13		23		33	
4	MW-22-3	14		24		34	
5	MW-22-4	15		25		35	
6	MW-22-5	16		26		36	
7	MW-22-1MS	17		27		37	
8	MW-22-1MSD	18		28		38	
9	↓ Dup	19		29		39	
10	PB	20		30		40	

Notes: _____

LDC #: 10414D4
SDG #: 03-3120

VALIDATION FINDINGS WORKSHEET
Calibration

Page: 1 of 1
Reviewer: ML
2nd Reviewer: [signature]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
Were all instruments calibrated daily, each set-up time, and were the proper number of standards used? Y
Were all initial and continuing calibration verification percent recoveries (%R) within the control limits of 90-110% for all analytes except mercury (90-120%) and cyanide (85-115%)? Y

LEVEL IV ONLY:
Was a midrange cyanide standard distilled? Y
Are all correlation coefficients ≥ 0.995 ? N/A
Were recalculated results acceptable? See Level IV Initial and Continuing Calibration Recalculation Worksheet for recalculations. N/A

#	Date	Calibration ID	Analyte	%R	Associated Samples	Qualification of Data
1	5/13/03	COV (1227)	AS	103.2 (90-110)	1, 2	J <input checked="" type="checkbox"/> A <input checked="" type="checkbox"/> P <input checked="" type="checkbox"/>
2	5/13/03	COV (1342)	AS	111.2 (90-110)	3-6	d <input checked="" type="checkbox"/>

Comments:

VALIDATION FINDINGS WORKSHEET
PB/ICB/CCB QUALIFIED SAMPLES

LDC #: 1041404
 SDG #: 03-3130
 METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: U-4/A Associated Samples: A1
 Sample Concentration units, unless otherwise noted: _____

Analyte	Maximum PB* (mg/Kg)	Maximum PB* (ug/L)	Maximum ICB/CCB* (ug/L)	Blank Action Limit	Sample Identification						
					1	3	5	6			
Al											Al
Sb											Sb
As											As
Ba											Ba
Be											Be
Cd											Cd
Ce		117	227.84	119.2							Ce
Cr											Cr
Cc											Co
Cu											Cu
Fe			19.77	98.85	31.4	81.3	53.8	57.5			Fe
Pb											Pb
Mg			25.97	129.85	20.7						Mg
Mn											Mn
Hg											Hg
Ni											Ni
K		64.7	115.69	578.45	93.5						K
Se											Se
Ag											Ag
Na		1040	1314.67	6573.35	1240						Na
Tl											Tl
V											V
Zn											Zn
B											B
Mo											Mo
Sr											Sr

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".
 Note: a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

LDC #: 1041404
SDG #: 03-330

VALIDATION FINDINGS WORKSHEET ICP Serial Dilution

Page: 1 of 1
Reviewer: [Signature]
2nd Reviewer: [Signature]

METHOD: Trace Metals (EPA SW 846 Method 60107000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

N N/A If analyte concentrations were > 50X the IDL, was an ICP serial dilution analyzed?

Y N/A Were ICP serial dilution percent differences (%D) \leq 10%?

Y N/A Is there evidence of negative interference? If yes, professional judgement will be used to qualify the data.

LEVEL IV ONLY:

Y N N/A Were recalculated results acceptable? See Level IV Recalculation Worksheet for recalculations.

#	Diluted Sample ID	Matrix	Analyte	%D	Associated Samples	Qualifications
1	2	A2	Mg K	10.8 11.5	A1 ↓	J Leto / A ↓

Comments:

LDC #: 1041404
SDG #: 03-3/20

VALIDATION FINDINGS WORKSHEET
Field Blanks

Page: 1 of 1
Reviewer: MB
2nd reviewer: A

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Y N N/A Were field blanks identified in this SDG?
 Y N N/A Were target analytes detected in the field blanks?

Sample: 1 Field Blank / Trip Blank / Rinsate / Other EB (circle one)

Analyte	Concentration Units (µg/l)
Fe	31.4
Mg	20.7
K	93.5
Na	1240

Sample: _____ Field Blank / Trip Blank / Rinsate / Other _____ (circle one)

Analyte	Concentration Units ()

**NASA JPL
Data Validation Reports
LDC# 10414**

Wet Chemistry

LDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: April 17, 2003
LDC Report Date: June 17, 2003
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 03-2767

Sample Identification

EB-1-4/17/03
MW-21-1
MW-21-2
MW-21-3
MW-21-4
MW-21-5
MW-21-1MS
MW-21-1MSD
MW-21-4MS
MW-21-4MSD
MW-21-5MS
MW-21-5MSD
MW-21-5DUP

Introduction

This data review covers 13 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 314.0 for Perchlorate, EPA SW 846 Method 7196 for Hexavalent Chromium, EPA SW 846 Method 9040B for pH, and Standard Method 2320B for Carbonate and Bicarbonate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-1-4/17/03 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-1-4/17/03	pH Chloride Nitrate as N Sulfate	6.04 units 0.29 mg/L 0.079 mg/L 0.70 mg/L

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 03-2767

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-2767

No Sample Data Qualified in this SDG

A

Applied P & Ch Laboratory Wet Analysis Results for Method SM2320B

Client Name: GEOFON, Inc.
Project ID: JPL

Project No: 04-4428.10
Service ID: 32767

Anal. Method SM2320B
Collected by:

Component Name: Bicarbonate
CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2767-1	EB-1-4/17/03	Water	04/17/03	04/17/03	04/18/03	03W2446	mg/L	2	< 2	U
03-2767-2	MW-21-1	Water	04/17/03	04/17/03	04/18/03	03W2446	mg/L	2	181	
03-2767-3	MW-21-2	Water	04/17/03	04/17/03	04/18/03	03W2446	mg/L	2	319	
03-2767-4	MW-21-3	Water	04/17/03	04/17/03	04/18/03	03W2446	mg/L	2	286	
03-2767-5	MW-21-4	Water	04/17/03	04/17/03	04/18/03	03W2446	mg/L	2	220	
03-2767-6	MW-21-5	Water	04/17/03	04/17/03	04/18/03	03W2446	mg/L	2	98.4	
03W2446-MB-01	03W2446-MB-01	Water	04/18/03	04/18/03	04/18/03	03W2446	mg/L	2	< 2	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten signature
6/18/07

Applied P & Ch Laboratory
Wet Analysis Results for Method SM2320B

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32767

Anal. Method SM2320B
 Collected by:

Component Name: Carbonate
 CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2767-1	EB-1-4/17/03	Water	04/17/03	04/17/03	04/18/03	03W2446	mg-CaCO ₃ /L	2	<2	U
03-2767-2	MW-21-1	Water	04/17/03	04/17/03	04/18/03	03W2446	mg-CaCO ₃ /L	2	<2	U
03-2767-3	MW-21-2	Water	04/17/03	04/17/03	04/18/03	03W2446	mg-CaCO ₃ /L	2	<2	U
03-2767-4	MW-21-3	Water	04/17/03	04/17/03	04/18/03	03W2446	mg-CaCO ₃ /L	2	<2	U
03-2767-5	MW-21-4	Water	04/17/03	04/17/03	04/18/03	03W2446	mg-CaCO ₃ /L	2	<2	U
03-2767-6	MW-21-5	Water	04/17/03	04/17/03	04/18/03	03W2446	mg-CaCO ₃ /L	2	<2	U
03W2446-MB-01	03W2446-MB-01	Water	04/18/03	04/18/03	04/18/03	03W2446	mg-CaCO ₃ /L	2	<2	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

6/18/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 9040

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32767

Anal. Method 9040
 Collected by:

Component Name: pH
 CAS No: 10-29-7

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2767-1	EB-1-4/17/03	Water	04/17/03	04/17/03	04/17/03	03W2443	pH unit	0.01	6.04	
03-2767-2	MW-21-1	Water	04/17/03	04/17/03	04/17/03	03W2443	pH unit	0.01	6.88	
03-2767-3	MW-21-2	Water	04/17/03	04/17/03	04/17/03	03W2443	pH unit	0.01	7.28	
03-2767-4	MW-21-3	Water	04/17/03	04/17/03	04/17/03	03W2443	pH unit	0.01	7.44	
03-2767-5	MW-21-4	Water	04/17/03	04/17/03	04/17/03	03W2443	pH unit	0.01	7.23	
03-2767-6	MW-21-5	Water	04/17/03	04/17/03	04/17/03	03W2443	pH unit	0.01	7.71	
03W2443-MB-01	03W2443-MB-01	Water	04/17/03	04/17/03	04/17/03	03W2443	pH unit	0.01	6.85	

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten: 6/18/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 160.1

Client Name: GEOFON, Inc.
Project ID: JPL

Project No: 04-4428.10
Service ID: 32767

Anal. Method 160.1
Collected by:

Component Name: Solids, Total Dissolved (TDS)
CAS No: 10-33-3

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2767-1	EB-1-4/17/03	Water	04/17/03	04/17/03	04/21/03	03W2463	mg/L	10	<10	U
03-2767-2	MW-21-1	Water	04/17/03	04/17/03	04/21/03	03W2463	mg/L	10	637	
03-2767-3	MW-21-2	Water	04/17/03	04/17/03	04/21/03	03W2463	mg/L	10	755	
03-2767-4	MW-21-3	Water	04/17/03	04/17/03	04/21/03	03W2463	mg/L	10	651	
03-2767-5	MW-21-4	Water	04/17/03	04/17/03	04/21/03	03W2463	mg/L	10	413	
03-2767-6	MW-21-5	Water	04/17/03	04/17/03	04/21/03	03W2463	mg/L	10	503	
03W2463-MB-01	03W2463-MB-01	Water	04/21/03	04/21/03	04/21/03	03W2463	mg/L	10	<10	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten signature/initials
6/18/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32767

Anal. Method 7196
 Collected by:

Component Name: Chromium (VI)
 CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2767-1	EB-1-4/17/03	Water	04/17/03	04/17/03	04/17/03	03W2441	mg/L	0.01	<0.01	U
03-2767-2	MW-21-1	Water	04/17/03	04/17/03	04/17/03	03W2441	mg/L	0.01	<0.01	U
03-2767-3	MW-21-2	Water	04/17/03	04/17/03	04/17/03	03W2441	mg/L	0.01	<0.01	U
03-2767-4	MW-21-3	Water	04/17/03	04/17/03	04/17/03	03W2441	mg/L	0.01	<0.01	U
03-2767-5	MW-21-4	Water	04/17/03	04/17/03	04/17/03	03W2441	mg/L	0.01	<0.01	U
03-2767-6	MW-21-5	Water	04/17/03	04/17/03	04/17/03	03W2441	mg/L	0.01	<0.01	U
03W2441-MB-01	03W2441-MB-01	Water	04/17/03	04/17/03	04/17/03	03W2441	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

6/12/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32767

Anal. Method 314.0
 Collected by:

Component Name: Perchlorate
 CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2767-1	EB-1-4/17/03	Water	04/17/03	04/17/03	04/17/03	03W2440	µg/L	4	<4	U
03-2767-2	MW-21-1	Water	04/17/03	04/17/03	04/17/03	03W2440	µg/L	4	3.6	B
03-2767-3	MW-21-2	Water	04/17/03	04/17/03	04/17/03	03W2440	µg/L	4	2.9	B
03-2767-4	MW-21-3	Water	04/17/03	04/17/03	04/17/03	03W2440	µg/L	4	2.9	B
03-2767-5	MW-21-4	Water	04/17/03	04/17/03	04/17/03	03W2440	µg/L	4	2.1	B
03-2767-6	MW-21-5	Water	04/17/03	04/17/03	04/17/03	03W2440	µg/L	4	2.7	B
03W2440-MB-01	03W2440-MB-01	Water	04/17/03	04/17/03	04/17/03	03W2440	µg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten: 6/18/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 300.0

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32767

Anal. Method 300.0
 Collected by:

Component Name: Chloride
 CAS No: 16887-00-6

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2767-1	EB-1-4/17/03	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	0.25	0.29	
03-2767-2	MW-21-1	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	4	110	
03-2767-3	MW-21-2	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	4	127	
03-2767-4	MW-21-3	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	4	102	
03-2767-5	MW-21-4	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	1.6	52.7	
03-2767-6	MW-21-5	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	1.6	68.5	
03W2435-MB-01	03W2435-MB-01	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	0.2	<0.2	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQ, or CRDL), but greater than MDL.

6/18/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 300.0

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32767

Anal. Method 300.0
 Collected by:

Component Name: Nitrate as N
 CAS No: 14797-55-8

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2767-1	EB-1-4/17/03	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	0.05	0.079	
03-2767-2	MW-21-1	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	0.8	13.2	
03-2767-3	MW-21-2	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	0.8	5.3	
03-2767-4	MW-21-3	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	0.8	8.8	
03-2767-5	MW-21-4	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	0.32	7.1	
03-2767-6	MW-21-5	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	0.32	6.8	
03W2435-MB-01	03W2435-MB-01	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	0.04	<0.04	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

6/18/03

Applied P & Ch Laboratory
Wet Analysis Results for Method 300.0

Client Name: GEOFON, Inc.
Project ID: JPL

Project No: 04-4428.10
Service ID: 32767

Anal. Method 300.0
Collected by:

Component Name: Sulfate
CAS No: 14808-79-8

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2767-1	EB-1-4/17/03	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	0.63	0.70	
03-2767-2	MW-21-1	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	10	181	
03-2767-3	MW-21-2	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	10	144	
03-2767-4	MW-21-3	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	10	122	
03-2767-5	MW-21-4	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	4	65.2	
03-2767-6	MW-21-5	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	4	103	
03W2435-MB-01	03W2435-MB-01	Water	04/17/03	04/17/03	04/17/03	03W2435	mg/L	0.5	<0.5	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

*M
6/18/07*

LDC #: 10414A6

VALIDATION COMPLETENESS WORKSHEET

Date: 6/16/03

SDG #: 03-2767

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: MB

2nd Reviewer: [Signature]

Carbonate, Bicarbonate

METHOD: (Analyte) Alkalinity (SM2320B), Chloride, Nitrate-N, Sulfate (EPA Method 300.0), Hexavalent Chromium (EPA SW846 Method 7196), Perchlorate (EPA Method 314.0), pH (EPA SW846 Method 9040), TDS (EPA Method 160.1)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/17/03
Ila.	Initial calibration	A	
Ilb.	Calibration verification	A	
III.	Blanks	A	
IV	Matrix Spike/Matrix Spike Duplicates	A	
V	Duplicates	A	
VI.	Laboratory control samples	A	LOS/LOSD
VII.	Sample result verification	N	
VIII.	Overall assessment of data	A	
IX.	Field duplicates	N	
X	Field blanks	SW	EB = 1

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: [Signature]

1	EB-1-4/17/03	11	MW-21-5MS	21		31	
2	MW-21-1	12	MW-21-5MSD	22		32	
3	MW-21-2	13	MW-21-5DUP	23		33	
4	MW-21-3	14	MB	24		34	
5	MW-21-4	15		25		35	
6	MW-21-5	16		26		36	
7	MW-21-1MS	17		27		37	
8	MW-21-1MSD	18		28		38	
9	MW-21-4MS	19		29		39	
10	MW-21-4MSD	20		30		40	

Notes: _____

LDC #: 10414A6
 SDG #: 03-2767

VALIDATION FINDINGS WORKSHEET
Sample Specific Analysis Reference

Page: 1 of 1
 Reviewer: MM
 2nd reviewer: [Signature]

All circled methods are applicable to each sample.

Sample ID	Parameter
1-6	<p>(PH) (TDS) (Cl) F (NO₃) NO₂ (SO₄) PO₄ (ALK) (CN) NH₃ TKN TOC (CR⁶⁺) (CO₂) (Carbonate) (bicarbonate)</p>
7-8	<p>(PH) (TDS) (Cl) F (NO₃) NO₂ (SO₄) PO₄ (ALK) (CN) NH₃ TKN TOC (CR⁶⁺) _____</p>
9-10	<p>(PH) (TDS) (Cl) F NO₃ NO₂ SO₄ PO₄ ALK CN NH₃ TKN TOC CR⁶⁺ (CO₂) _____</p>
11-12	<p>(PH) (TDS) (Cl) F NO₃ NO₂ SO₄ PO₄ ALK CN NH₃ TKN TOC CR⁶⁺ _____</p>
13	<p>(PH) (TDS) (Cl) F NO₃ NO₂ SO₄ PO₄ ALK CN NH₃ TKN TOC CR⁶⁺ _____</p>
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____
	(PH) TDS (Cl) F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ _____

Comments: _____

LDC #: 10414A6
SDG #: 03-2967

VALIDATION FINDINGS WORKSHEET
Field Blanks

Page: 1 of 1
Reviewer: MW
2nd reviewer: [Signature]

METHOD: Inorganics, EPA Method See below

N N/A Were field blanks identified in this SDG?
 N N/A Were target analytes detected in the field blanks?

Sample: 1 Field Blank / Trip Blank / Rinsate (circle one) EB

Analyte	Concentration Units (<u>mg/L</u>)
<u>pH (unit)</u>	<u>6.04</u>
<u>Cl</u>	<u>0.29</u>
<u>NO₃-N</u>	<u>0.079</u>
<u>SO₄</u>	<u>0.90</u>

Sample: _____ Field Blank / Trip Blank / Rinsate (circle one)

Analyte	Concentration Units (_____)

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: April 21, 2003
LDC Report Date: June 17, 2003
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-2809

Sample Identification

DUPE-1-2Q03
EB-2-4/21/03
MW-4-1
MW-4-2
MW-4-3
MW-4-4
MW-4-5
SOURCE-2Q03
MW-4-1DUP
MW-4-2MS
MW-4-2MSD
DUPE-1-2Q03DUP

Introduction

This data review covers 12 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 160.1 for Total Dissolved Solids, EPA Method 300.0 for Chloride, Nitrate as Nitrogen, and Sulfate, EPA Method 314.0 for Perchlorate, EPA SW 846 Method 7196 for Hexavalent Chromium, EPA SW 846 Method 9040B for pH, and Standard Method 2320B for Carbonate and Bicarbonate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

Samples DUPE-1-2Q03 and MW-4-4 were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration		RPD
	DUPE-1-2Q03	MW-4-4	
Bicarbonate	161 mg/L	147 mg/L	9
pH	7.88 units	8.01 units	2
Total dissolved solids	221 mg/L	191 mg/L	15
Chloride	24.2 mg/L	22.7 mg/L	6
Nitrate as N	1.7 mg/L	1.7 mg/L	0
Sulfate	9.5 mg/L	10.2 mg/L	7

X. Field Blanks

Sample EB-2-4/21/03 was identified as an equipment blank. No contaminant concentrations were found in this blank with the following exceptions:

Equipment Blank ID	Analyte	Concentration
EB-2-4/21/03	pH Chloride Nitrate as N Sulfate	7.85 units 0.25 mg/L 0.078 mg/L 0.73 mg/L

Sample SOURCE-2Q03 was identified as a source blank. No contaminant concentrations were found in this blank with the following exceptions:

Source Blank ID	Analyte	Concentration
SOURCE-2Q03	pH Chloride Nitrate as N	6.60 units 0.21 mg/L 0.075 mg/L

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 03-2809

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 03-2809

No Sample Data Qualified in this SDG

Applied P & Ch Laboratory
Wet Analysis Results for Method SM2320B

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32809

Anal. Method SM2320B
 Collected by:

Component Name: Bicarbonate
 CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2809-1	DUPE-1-2Q03	Water	04/21/03	04/21/03	04/24/03	03W2539	mg/L	2	161	
03-2809-2	EB-2-4/21/03	Water	04/21/03	04/21/03	04/24/03	03W2539	mg/L	2	< 2	U
03-2809-3	MW-4-1	Water	04/21/03	04/21/03	04/24/03	03W2539	mg/L	2	179	
03-2809-4	MW-4-2	Water	04/21/03	04/21/03	04/24/03	03W2539	mg/L	2	224	
03-2809-5	MW-4-3	Water	04/21/03	04/21/03	04/24/03	03W2539	mg/L	2	197	
03-2809-6	MW-4-4	Water	04/21/03	04/21/03	04/24/03	03W2539	mg/L	2	147	
03-2809-7	MW-4-5	Water	04/21/03	04/21/03	04/24/03	03W2539	mg/L	2	120	
03-2809-8	SOURCE-2Q03	Water	04/21/03	04/21/03	04/24/03	03W2539	mg/L	2	< 2	U
03W2539-MB-01	03W2539-MB-01	Water	04/24/03	04/24/03	04/24/03	03W2539	mg/L	2	< 2	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

J
 6/18/03

Applied P & Ch Laboratory
Wet Analysis Results for Method SM2320B

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32809

Anal. Method SM2320B
 Collected by:

Component Name: Carbonate
 CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2809-1	DUPE-1-2Q03	Water	04/21/03	04/21/03	04/24/03	03W2539	mg-CaCO ₃ /L	2	<2	U
03-2809-2	EB-2-4/21/03	Water	04/21/03	04/21/03	04/24/03	03W2539	mg-CaCO ₃ /L	2	<2	U
03-2809-3	MW-4-1	Water	04/21/03	04/21/03	04/24/03	03W2539	mg-CaCO ₃ /L	2	<2	U
03-2809-4	MW-4-2	Water	04/21/03	04/21/03	04/24/03	03W2539	mg-CaCO ₃ /L	2	<2	U
03-2809-5	MW-4-3	Water	04/21/03	04/21/03	04/24/03	03W2539	mg-CaCO ₃ /L	2	<2	U
03-2809-6	MW-4-4	Water	04/21/03	04/21/03	04/24/03	03W2539	mg-CaCO ₃ /L	2	<2	U
03-2809-7	MW-4-5	Water	04/21/03	04/21/03	04/24/03	03W2539	mg-CaCO ₃ /L	2	<2	U
03-2809-8	SOURCE-2Q03	Water	04/21/03	04/21/03	04/24/03	03W2539	mg-CaCO ₃ /L	2	10.2	
03W2539-MB-01	03W2539-MB-01	Water	04/24/03	04/24/03	04/24/03	03W2539	mg-CaCO ₃ /L	2	<2	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

*M
6/18/07*

Applied P & Ch Laboratory
Wet Analysis Results for Method 9040B

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32809

Anal. Method 9040B
 Collected by:

Component Name: pH
 CAS No: 10-29-7

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2809-1	DUPE-1-2Q03	Water	04/21/03	04/21/03	04/21/03	03W2477	pH unit	0.01	7.88	
03-2809-2	EB-2-4/21/03	Water	04/21/03	04/21/03	04/21/03	03W2477	pH unit	0.01	7.85	
03-2809-3	MW-4-1	Water	04/21/03	04/21/03	04/21/03	03W2477	pH unit	0.01	7.28	
03-2809-4	MW-4-2	Water	04/21/03	04/21/03	04/21/03	03W2477	pH unit	0.01	7.05	
03-2809-5	MW-4-3	Water	04/21/03	04/21/03	04/21/03	03W2477	pH unit	0.01	7.58	
03-2809-6	MW-4-4	Water	04/21/03	04/21/03	04/21/03	03W2477	pH unit	0.01	8.01	
03-2809-7	MW-4-5	Water	04/21/03	04/21/03	04/21/03	03W2477	pH unit	0.01	8.47	
03-2809-8	SOURCE-2Q03	Water	04/21/03	04/21/03	04/21/03	03W2477	pH unit	0.01	6.60	
03W2477-MB-01	03W2477-MB-01	Water	04/21/03	04/21/03	04/21/03	03W2477	pH unit	0.01	6.86	

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten: 6/18/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 160.1

Client Name: GEOFON, Inc. Project No: 04-4428.10 Anal. Method 160.1
 Project ID: JPL Service ID: 32809 Collected by:

Component Name: Solids, Total Dissolved (TDS)

CAS No: 10-33-3

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2809-1	DUPE-1-2Q03	Water	04/21/03	04/21/03	04/22/03	03W2500	mg/L	10	221	
03-2809-2	EB-2-4/21/03	Water	04/21/03	04/21/03	04/22/03	03W2500	mg/L	10	<10	U
03-2809-3	MW-4-1	Water	04/21/03	04/21/03	04/22/03	03W2500	mg/L	10	244	
03-2809-4	MW-4-2	Water	04/21/03	04/21/03	04/22/03	03W2500	mg/L	10	587	
03-2809-5	MW-4-3	Water	04/21/03	04/21/03	04/22/03	03W2500	mg/L	10	226	
03-2809-6	MW-4-4	Water	04/21/03	04/21/03	04/22/03	03W2500	mg/L	10	191	
03-2809-7	MW-4-5	Water	04/21/03	04/21/03	04/22/03	03W2500	mg/L	10	133	
03-2809-8	SOURCE-2Q03	Water	04/21/03	04/21/03	04/22/03	03W2500	mg/L	10	<10	U
03W2500-MB-01	03W2500-MB-01	Water	04/22/03	04/22/03	04/22/03	03W2500	mg/L	10	<10	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

6/18/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32809

Anal. Method 7196
 Collected by:

Component Name: Chromium (VI)

CAS No: 1333-82-0

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2809-1	DUPE-1-2Q03	Water	04/21/03	04/21/03	04/21/03	03W2476	mg/L	0.01	<0.01	U
03-2809-2	EB-2-4/21/03	Water	04/21/03	04/21/03	04/21/03	03W2476	mg/L	0.01	<0.01	U
03-2809-3	MW-4-1	Water	04/21/03	04/21/03	04/21/03	03W2476	mg/L	0.01	<0.01	U
03-2809-4	MW-4-2	Water	04/21/03	04/21/03	04/21/03	03W2476	mg/L	0.01	<0.01	U
03-2809-5	MW-4-3	Water	04/21/03	04/21/03	04/21/03	03W2476	mg/L	0.01	<0.01	U
03-2809-6	MW-4-4	Water	04/21/03	04/21/03	04/21/03	03W2476	mg/L	0.01	<0.01	U
03-2809-7	MW-4-5	Water	04/21/03	04/21/03	04/21/03	03W2476	mg/L	0.01	<0.01	U
03-2809-8	SOURCE-2Q03	Water	04/21/03	04/21/03	04/21/03	03W2476	mg/L	0.01	<0.01	U
03W2476-MB-01	03W2476-MB-01	Water	04/21/03	04/21/03	04/21/03	03W2476	mg/L	0.01	<0.01	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

gr
6/18/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32809

Anal. Method 314.0
 Collected by:

Component Name: Perchlorate
 CAS No:

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2809-1	DUPE-1-2Q03	Water	04/21/03	04/21/03	04/22/03	03W2496	µg/L	4	<4	U
03-2809-2	EB-2-4/21/03	Water	04/21/03	04/21/03	04/22/03	03W2496	µg/L	4	<4	U
03-2809-3	MW-4-1	Water	04/21/03	04/21/03	04/22/03	03W2496	µg/L	4	<4	U
03-2809-4	MW-4-2	Water	04/21/03	04/21/03	04/22/03	03W2496	µg/L	4	6.6	
03-2809-5	MW-4-3	Water	04/21/03	04/21/03	04/22/03	03W2496	µg/L	4	<4	U
03-2809-6	MW-4-4	Water	04/21/03	04/21/03	04/22/03	03W2496	µg/L	4	<4	U
03-2809-7	MW-4-5	Water	04/21/03	04/21/03	04/22/03	03W2496	µg/L	4	<4	U
03-2809-8	SOURCE-2Q03	Water	04/21/03	04/21/03	04/22/03	03W2496	µg/L	4	<4	U
03W2496-MB-01	03W2496-MB-01	Water	04/22/03	04/22/03	04/22/03	03W2496	µg/L	4	<4	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten signature and date: 6/18/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 300.0

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32809

Anal. Method 300.0
 Collected by:

Component Name: Chloride Cl⁻
 CAS No: 16887-00-6

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2809-1	DUPE-1-2Q03	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.5	24.2	
03-2809-2	EB-2-4/21/03	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.25	0.25	B
03-2809-3	MW-4-1	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.4	18.8	
03-2809-4	MW-4-2	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	4	95.4	
03-2809-5	MW-4-3	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.8	27.4	
03-2809-6	MW-4-4	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.8	22.7	
03-2809-7	MW-4-5	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.8	22.5	
03-2809-8	SOURCE-2Q03	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.25	0.21	B
03W2479-MB-01	03W2479-MB-01	Water	04/22/03	04/22/03	04/22/03	03W2479	mg/L	0.2	<0.2	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten signature/initials
 6/18/07

Applied P & Ch Laboratory
Wet Analysis Results for Method 300.0

Client Name: GEOFON, Inc. Project No: 04-4428.10 Anal. Method 300.0
 Project ID: JPL Service ID: 32809 Collected by:

Component Name: Nitrate as N
 CAS No: 14797-55-8

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2809-1	DUPE-1-2Q03	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.1	1.7	
03-2809-2	EB-2-4/21/03	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.05	0.078	
03-2809-3	MW-4-1	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.08	1.5	
03-2809-4	MW-4-2	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.8	10.0	
03-2809-5	MW-4-3	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.16	0.24	
03-2809-6	MW-4-4	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.16	1.7	
03-2809-7	MW-4-5	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.16	0.20	
03-2809-8	SOURCE-2Q03	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.05	0.075	
03W2479-MB-01	03W2479-MB-01	Water	04/22/03	04/22/03	04/22/03	03W2479	mg/L	0.04	<0.04	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten signature/initials
 6/18/03

Applied P & Ch Laboratory
Wet Analysis Results for Method 300.0

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 32809

Anal. Method 300.0
 Collected by:

Component Name: Sulfate SO_4^{--}
 CAS No: 14808-79-8

Lab ID	Sample ID	Matrix	Coll. Date	Rcv Date	Anal. Date	Batch	Unit	RL	Result	Q
03-2809-1	DUPE-1-2Q03	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	1.3	9.5	
03-2809-2	EB-2-4/21/03	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.63	0.73	
03-2809-3	MW-4-1	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	1	40.7	
03-2809-4	MW-4-2	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	10	124	
03-2809-5	MW-4-3	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	2	2.3	
03-2809-6	MW-4-4	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	2	10.2	
03-2809-7	MW-4-5	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	2	8.2	
03-2809-8	SOURCE-2Q03	Water	04/21/03	04/21/03	04/22/03	03W2479	mg/L	0.63	<0.63	U
03W2479-MB-01	03W2479-MB-01	Water	04/22/03	04/22/03	04/22/03	03W2479	mg/L	0.5	<0.5	U

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

6/18/07

LDC #: 10414B6

VALIDATION COMPLETENESS WORKSHEET

Date: 6/16/03

SDG #: 03-2809

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: *WV*

2nd Reviewer: *A*

Bicarbonate, carbonate

METHOD: (Analyte) Alkalinity (SM2320B), Chloride, Nitrate-N, Sulfate (EPA Method 300.0), Hexavalent Chromium (EPA SW846 Method 7196), Perchlorate (EPA Method 314.0), pH (EPA SW846 Method 9040), TDS (EPA Method 160.1)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 4/21/03
IIa.	Initial calibration	A	
IIb.	Calibration verification	A	
III.	Blanks	A	
IV.	Matrix Spike/Matrix Spike Duplicates	A	} MS/MSD for TDS from SNG 03-2819
V.	Duplicates	A	
VI.	Laboratory control samples	A	LCs/LCSD
VII.	Sample result verification	N	
VIII.	Overall assessment of data	A	
IX.	Field duplicates	SW	(1,6)
X.	Field blanks	SW	EB=2, SB=8

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: *AD*

1	✓ DUPE-1-2Q03	11	MW-4-2MSD	21		31	
2	EB-2-4/21/03	12	DUPE-1-2Q03 Dup.	22		32	
3	MW-4-1	13	MB	23		33	
4	MW-4-2	14		24		34	
5	MW-4-3	15		25		35	
6	✓ MW-4-4	16		26		36	
7	MW-4-5	17		27		37	
8	SOURCE-2Q03	18		28		38	
9	MW-4-1DUP	19		29		39	
10	MW-4-2MS	20		30		40	

Notes: _____

LDC #: 10414B6
SDG #: 03-2809

VALIDATION FINDINGS WORKSHEET Sample Specific Analysis Reference

Page: 1 of 1
Reviewer: MH
2nd reviewer: ✓

All circled methods are applicable to each sample.

Sample ID	Parameter
1-8	<u>pH</u> <u>TDS</u> <u>Cl</u> F <u>NO₃</u> NO ₂ <u>SO₄</u> PO ₄ ALK CN' NH ₃ TKN TOC <u>CR⁶⁺</u> <u>Clog</u> <u>Carbonate</u> <u>bicarbonate</u>
9	<u>pH</u> TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
✓ 10-11	pH TDS <u>Cl</u> F <u>NO₃</u> NO ₂ <u>SO₄</u> PO ₄ ALK CN' NH ₃ TKN TOC <u>CR⁶⁺</u> <u>Clog</u>
✓ 12	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ <u>ALK</u> CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺
	pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN' NH ₃ TKN TOC CR ⁶⁺

Comments: _____

LDC #: 10414B6
 SDG #: 63-2809

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
 Reviewer: MW
 2nd reviewer: [Signature]

METHOD: Inorganics, Method see cover

Y N N/A Were field duplicate pairs identified in this SDG?
 Y N N/A Were target analytes detected in the field duplicate pairs?

Analyte	Concentration ($\mu\text{g/l}$)		RPD (Limit)	Difference (Limit)	Qualifier
	1	6			
Bicarbonate	161	147	9		
pH (unit)	7.88	8.01	2		
TDS	221	191	15		
Cl	24.2	22.7	6		
NO ₃ -N	1.7	1.7	0		

Analyte	Concentration ($\mu\text{g/l}$)		RPD (Limit)	Difference (Limit)	Qualifier
	1	6			
SO ₄	9.5	10.2	7		

Analyte	Concentration ()		RPD (Limit)	Difference (Limit)	Qualifier

Analyte	Concentration ()		RPD (Limit)	Difference (Limit)	Qualifier

LDC #: 10414B6
SDG #: 03-2809

VALIDATION FINDINGS WORKSHEET
Field Blanks

Page: 1 of 1
Reviewer: 147
2nd reviewer: [Signature]

METHOD: Inorganics, EPA Method _____


N N/A Were field blanks identified in this SDG?
 N N/A Were target analytes detected in the field blanks?

Sample: 2 Field Blank / Trip Blank / Rinsate (circle one) EB

Analyte	Concentration Units (mg/l)
PH (unit)	7.85
ce	0.25
NO ₃ -N	0.078
SO ₄	0.93

Sample: 8 Field Blank / Trip Blank / Rinsate (circle one) SB

Analyte	Concentration Units (mg/l)
PH (unit)	6.60
ce	0.21
NO ₃ -N	0.075



LABORATORY DATA CONSULTANTS, INC.

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

Geofon, Inc.
22632 Golden Springs Drive, Suite 270
Diamond Bar, CA 91765
ATTN: Mr. Leo Williamson

August 4, 2003

SUBJECT: NASA JPL, DO #01, Data Validation

Dear Mr. Williamson,

Enclosed are the final validation reports for the fractions listed below. These SDGs were received on July 23, 2003. Attachment 1 is a summary of the samples that were reviewed for each analysis.

LDC Project # 10609:

<u>SDG #</u>	<u>Fraction</u>
03-3391, 03-3414, 03-3444, 03-3465, 03-3484	Volatiles, 1,4-Dioxane, Metals, Wet Chemistry

The data validation was performed under EPA Level III guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, February 1994
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996

Please feel free to contact us if you have any questions.

Sincerely,

Erlinda T. Rauto
Operations Manager/Senior Chemist

Attachment 1

LDC #10609 (Geofon, inC.-Diamond Bar / NASA, JPL, DO#001)

LDC	SDG#	DATE REC'D	DATE DUE	VOA (524.2)		1,4-Dioxane (8270C-SIM)		Metals (200.7/200.9)		Pb,Cr (200.8)	ClO ₂ (314.0)		Alk. (2320B)		pH (9040B)		TDS (160.1)		Cr(VI) (7196A)		Cl,NO ₃ -N, SO ₄ (300.0)													
				W	S	W	S	W	S		W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S	W	S		
Matrix: Water/Soil																																		
A	03-3391	7-23-03	8-13-03	3	0	2	0	2	0	-	2	0	2	0	3	0	2	0	2	0	4	0	2	0										
B	03-3414	7-23-03	8-13-03	5	0	1	0	2	0	-	2	0	2	0	3	0	2	0	2	0	4	0	2	0										
C	03-3444	7-23-03	8-13-03	8	0	-	-	9	0	-	7	0	3	0	3	0	7	0	7	0	7	0	7	0										
D	03-3465	7-23-03	8-13-03	4	0	1	0	3	0	-	3	0	3	0	4	0	6	0	5	0	3	0	3	0										
E	03-3484	7-23-03	8-13-03	-	-	-	-	-	-	32	0	-	-	-	-	-	-	-	-	-	-	-	-											
				20	0	4	0	16	0	32	0	14	0	10	0	13	0	17	0	20	0	14	0	14	0	0	0	0	0	0	0	0	0	
Total				TH																														

**NASA JPL
Data Validation Reports
LDC# 10609**

Volatiles

LDC

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 27, 2003
LDC Report Date: July 29, 2003
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-3391

Sample Identification

MW-13
MW-16
TB-14-5/27/03

Introduction

This data review covers 3 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.

J Indicates an estimated value.

R Quality control indicates the data is not usable.

N Presumptive evidence of presence of the constituent.

UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.

A Indicates the finding is based upon technical validation criteria.

P Indicates the finding is related to a protocol/contractual deviation.

None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
5/29/03	Methylene chloride	30.07	All samples in SDG 03-3391	J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
03G2684-MB-01	5/29/03	Methylene chloride	5.9 ug/L	All samples in SDG 03-3391

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
MW-13	Methylene chloride	5.3 ug/L	5.3U ug/L
MW-16	Methylene chloride	2.8 ug/L	2.8U ug/L
TB-14-5/27/03	Methylene chloride	6.1 ug/L	6.1U ug/L

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-14-5/27/03 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Compound	Concentration (ug/L)
TB-14-5/27/03	Methylene chloride 4-Methyl-2-pentanone	6.1 5

**NASA JPL
Volatiles - Data Qualification Summary - SDG 03-3391**

SDG	Sample	Compound	Flag	A or P	Reason
03-3391	MW-13 MW-16 TB-14-5/27/03	Methylene chloride	J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

**NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 03-3391**

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P
03-3391	MW-13	Methylene chloride	5.3U ug/L	A
03-3391	MW-16	Methylene chloride	2.8U ug/L	A
03-3391	TB-14-5/27/03	Methylene chloride	6.1U ug/L	A

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name:	GEOFON, Inc.	Project No:	04-4428.10	Collection Date:	05/27/2003
Project ID:	JPL	Service ID:	33391	Collected by:	
Sample ID:	MW-13	Lab Sample ID:	03-3391-1	Received Date:	05/27/2003
Sample Type:	Field Sample	Sample Matrix:	Water	Moisture %:	-
Anal. Method:	524.2	Prep. Method:	5030	Instrument ID:	GC/MS: G
Batch No:	03G2684	Prep. Date:	05/29/03	Anal. Date:	05/29/03
Data File Name:	3391-01	Prep. No:	-	Anal. Time:	14:39
Methanol Vol.	-	Sample Amount:	25 mL	Dilution Factor:	1
Test Level:	Low	Sparge Size:	25 mL	Heated Purge: (Y/N)	N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	1.3	
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	1.5	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	1.1 (a)	<1.1	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	0.4	J
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	1.8 ^(a)	5.3 <i>UI</i>	B
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	5	J
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	1.0	
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	9.2	
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	1,1,2,2-TRICHLORO-1,1,2,2-TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

		Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	112
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	93
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	99
4	TOLUENE-D8	2037-26-5	73-129	105
# of out-of-control			0	

Internal Standard

		Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4	50-200	82
2	1,4-DICHLOROETHANE-D4	3855-82-1	50-200	80
3	FLUOROBENZENE	462-06-6	50-200	90
# of out-of-control			0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

^(a)MDL Reported.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

8/4/03

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 05/27/2003
Project ID: JPL	Service ID: 33391	Collected by:
Sample ID: MW-16	Lab Sample ID: 03-3391-2	Received Date: 05/27/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: G
Batch No: 03G2684	Prep. Date: 05/29/03	Anal. Date: 05/29/03
Data File Name: 3391-02	Prep. No: -	Anal. Time: 15:10
Methanol Vol. -	Sample Amount: 25 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	2.9	
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	3.8	
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	1.1 (a)	<1.1	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	0.9	
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

GEOFON

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	<0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	1.8 ^(a)	2.8 <i>µJ</i>	B
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	<1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	4	J
44	NAPHTHALENE	91-20-3	µg/L	0.5	<0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	<0.5	U
46	STYRENE	100-42-5	µg/L	0.5	<0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	<0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	<0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	<0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	<0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	<0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	<0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	<0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	<0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	1.6	
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	<0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	<0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	<0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	<0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	<0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	<0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	<0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	<0.5	U

Surrogates

		Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	99
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	89
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	94
4	TOLUENE-D8	2037-26-5	73-129	102
# of out-of-control				0

Internal Standard

		Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4	50-200	92
2	1,4-DICHLOROETHANE-D4	3855-82-1	50-200	94
3	FLUOROBENZENE	462-06-6	50-200	99
# of out-of-control				0

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

(a)MDL Reported.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Handwritten signature

Applied P & Ch Laboratory
Organic Analysis Results for Method 524.2

Client Name: GEOFON, Inc.	Project No: 04-4428.10	Collection Date: 05/27/2003
Project ID: JPL	Service ID: 33391	Collected by:
Sample ID: TB-14-5/27/03	Lab Sample ID: 03-3391-3	Received Date: 05/27/2003
Sample Type: Field Sample	Sample Matrix: Water	Moisture %: -
Anal. Method: 524.2	Prep. Method: 5030	Instrument ID: GC/MS: G
Batch No: 03G2684	Prep. Date: 05/29/03	Anal. Date: 05/29/03
Data File Name: 3391-03	Prep. No: -	Anal. Time: 15:40
Methanol Vol: -	Sample Amount: 25 mL	Dilution Factor: 1
Test Level: Low	Sparge Size: 25 mL	Heated Purge: (Y/N) N

#	Component Name	CAS No	Unit	RL	Result	Qualifier
1	BENZENE	71-43-2	µg/L	0.5	<0.5	U
2	BROMOBENZENE	108-86-1	µg/L	0.5	<0.5	U
3	BROMOCHLOROMETHANE	74-97-5	µg/L	0.5	<0.5	U
4	BROMODICHLOROMETHANE	75-27-4	µg/L	0.5	<0.5	U
5	BROMOFORM	75-25-2	µg/L	0.5	<0.5	U
6	BROMOMETHANE	74-83-9	µg/L	0.5	<0.5	U
7	N-BUTYLBENZENE	104-51-8	µg/L	0.5	<0.5	U
8	SEC-BUTYLBENZENE	135-98-8	µg/L	0.5	<0.5	U
9	TERT-BUTYLBENZENE	98-06-6	µg/L	0.5	<0.5	U
10	2-BUTANONE	78-93-3	µg/L	10	<10	U
11	CARBON TETRACHLORIDE	56-23-5	µg/L	0.5	<0.5	U
12	CHLOROBENZENE	108-90-7	µg/L	0.5	<0.5	U
13	CHLORODIBROMOMETHANE	124-48-1	µg/L	0.5	<0.5	U
14	CHLOROETHANE	75-00-3	µg/L	0.5	<0.5	U
15	CHLOROFORM	67-66-3	µg/L	0.5	<0.5	U
16	CHLOROMETHANE	74-87-3	µg/L	0.5	<0.5	U
17	2-CHLOROTOLUENE	95-49-8	µg/L	0.5	<0.5	U
18	4-CHLOROTOLUENE	106-43-4	µg/L	0.5	<0.5	U
19	1,2-DIBROMO-3-CHLOROPROPANE	96-12-8	µg/L	1.1 (a)	<1.1	U
20	1,2-DIBROMOETHANE (EDB)	106-93-4	µg/L	0.5	<0.5	U
21	DIBROMOMETHANE	74-95-3	µg/L	0.5	<0.5	U
22	1,2-DICHLOROBENZENE	95-50-1	µg/L	0.5	<0.5	U
23	1,3-DICHLOROBENZENE	541-73-1	µg/L	0.5	<0.5	U
24	1,4-DICHLOROBENZENE	106-46-7	µg/L	0.5	<0.5	U
25	DICHLORODIFLUOROMETHANE	75-71-8	µg/L	0.5	<0.5	U
26	1,1-DICHLOROETHANE	75-34-3	µg/L	0.5	<0.5	U
27	1,2-DICHLOROETHANE	107-06-2	µg/L	0.5	<0.5	U
28	1,1-DICHLOROETHENE	75-35-4	µg/L	0.5	<0.5	U
29	CIS-1,2-DICHLOROETHENE	156-59-2	µg/L	0.5	<0.5	U
30	TRANS-1,2-DICHLOROETHENE	156-60-5	µg/L	0.5	<0.5	U
31	1,2-DICHLOROPROPANE	78-87-5	µg/L	0.5	<0.5	U
32	1,3-DICHLOROPROPANE	142-28-9	µg/L	0.5	<0.5	U
33	2,2-DICHLOROPROPANE	594-20-7	µg/L	0.5	<0.5	U
34	1,1-DICHLOROPROPENE	563-58-6	µg/L	0.5	<0.5	U
35	CIS-1,3-DICHLOROPROPENE	10061-01-5	µg/L	0.5	<0.5	U
36	TRANS-1,3-DICHLOROPROPENE	10061-02-6	µg/L	0.5	<0.5	U
37	ETHYLBENZENE	100-41-4	µg/L	0.5	<0.5	U
38	HEXACHLOROBUTADIENE	87-68-3	µg/L	0.5	<0.5	U
39	ISOPROPYLBENZENE (CUMENE)	98-82-8	µg/L	0.5	<0.5	U

#	Component Name	CAS No	Unit	RL	Result	Qualifier
40	P-ISOPROPYLTOLUENE	99-87-6	µg/L	0.5	< 0.5	U
41	METHYLENE CHLORIDE	75-09-2	µg/L	1.8 ^(a)	6.1 <i>MS</i>	B
42	METHYL-T-BUTYL ETHER (MTBE)	1634-04-4	µg/L	1	< 1	U
43	4-METHYL-2-PENTANONE (MIBK)	108-10-1	µg/L	10	5	J
44	NAPHTHALENE	91-20-3	µg/L	0.5	< 0.5	U
45	N-PROPYLBENZENE	103-65-1	µg/L	0.5	< 0.5	U
46	STYRENE	100-42-5	µg/L	0.5	< 0.5	U
47	1,1,1,2-TETRACHLOROETHANE	630-20-6	µg/L	0.5	< 0.5	U
48	1,1,2,2-TETRACHLOROETHANE	79-34-5	µg/L	0.5	< 0.5	U
49	TETRACHLOROETHENE	127-18-4	µg/L	0.5	< 0.5	U
50	TOLUENE	108-88-3	µg/L	0.5	< 0.5	U
51	1,2,3-TRICHLOROBENZENE	87-61-6	µg/L	0.5	< 0.5	U
52	1,2,4-TRICHLOROBENZENE	120-82-1	µg/L	0.5	< 0.5	U
53	1,1,1-TRICHLOROETHANE	71-55-6	µg/L	0.5	< 0.5	U
54	1,1,2-TRICHLOROETHANE	79-00-5	µg/L	0.5	< 0.5	U
55	TRICHLOROETHENE	79-01-6	µg/L	0.5	< 0.5	U
56	TRICHLOROFLUOROMETHANE	75-69-4	µg/L	0.5	< 0.5	U
57	1,2,3-TRICHLOROPROPANE	96-18-4	µg/L	0.5	< 0.5	U
58	112TRICHLORO-122TRIFLUOROETHANE	76-13-1	µg/L	0.5	< 0.5	U
59	1,2,4-TRIMETHYLBENZENE	95-63-6	µg/L	0.5	< 0.5	U
60	1,3,5-TRIMETHYLBENZENE	108-67-8	µg/L	0.5	< 0.5	U
61	VINYL CHLORIDE	75-01-4	µg/L	0.5	< 0.5	U
62	O-XYLENE	95-47-6	µg/L	0.5	< 0.5	U
63	M/P-XYLENE	108-38-3	µg/L	0.5	< 0.5	U

Surrogates

		Control Limit, %	Surro. Rec.%	
1	1-BROMO-4-FLUOROBENZENE (4-BROMOFL)	460-00-4	70-129	105
2	1,2-DICHLOROETHANE-D4	17060-07-0	70-129	92
3	DIBROMOFLUOROMETHANE	1868-53-7	70-122	99
4	TOLUENE-D8	2037-26-5	73-129	106
# of out-of-control			0	

Internal Standard

		Control Limit, %	IS Rec.%	
1	CHLOROBENZENE-D5	3114-55-4	50-200	93
2	1,4-DICHLOROETHANE-D4	3855-82-1	50-200	96
3	FLUOROBENZENE	462-06-6	50-200	101
# of out-of-control			0	

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

^(a)MDL Reported.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

8/4/02

LDC #: 10609A1

VALIDATION COMPLETENESS WORKSHEET

SDG #: 03-3391

Level III

Laboratory: Applied P & Ch Laboratory

Date: 7/28/03

Page: 1 of 1

Reviewer: [Signature]

2nd Reviewer: [Signature]

METHOD: GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 5/27/03
II.	GC/MS Instrument performance check	A	
III.	Initial calibration	A	% RSD ≤ 20%, r ² 20.990
IV.	Continuing calibration	SW	% RSD ≤ 30%
V.	Blanks	SW	
VI.	Surrogate spikes	A	
VII.	Matrix spike/Matrix spike duplicates	A	03-3414-1
VIII.	Laboratory control samples	A	LCS
IX.	Regional Quality Assurance and Quality Control	N	
X.	Internal standards	A	
XI.	Target compound identification	N	
XII.	Compound quantitation/CRQLs	N	
XIII.	Tentatively identified compounds (TICs)	N	
XIV.	System performance	N	
XV.	Overall assessment of data	A	
XVI.	Field duplicates	N	
XVII.	Field blanks	SW	TB = 3

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

1 †	MW-13	11		21		31	
2 †	MW-16	12		22		32	
3 †	TB-14-5/27/03	13		23		33	
4	03G2684-MB-01	14		24		34	
5		15		25		35	
6		16		26		36	
7		17		27		37	
8		18		28		38	
9		19		29		39	
10		20		30		40	

TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method 524.2)

A. Chloromethane	Q. 1,2-Dichloropropane	GG. Xylenes, total	WW. Bromobenzene	MMM. Naphthalene
B. Bromomethane	R. cis-1,3-Dichloropropene	HH. Vinyl acetate	XX. 1,2,3-Trichloropropane	NNN. 1,2,3-Trichlorobenzene
C. Vinyl chloride	S. Trichloroethene	II. 2-Chloroethylvinyl ether	YY. n-Propylbenzene	OOO. 1,3,5-Trichlorobenzene
D. Chloroethane	T. Dibromochloromethane	JJ. Dichlorodifluoromethane	ZZ. 2-Chlorotoluene	PPP. trans-1,2-Dichloroethene
E. Methylene chloride	U. 1,1,2-Trichloroethane	KK. Trichlorofluoromethane	AAA. 1,3,5-Trimethylbenzene	QQQ. cis-1,2-Dichloroethene
F. Acetone	V. Benzene	LL. Methyl-tert-butyl ether	BBB. 4-Chlorotoluene	RRR. m,p-Xylenes
G. Carbon disulfide	W. trans-1,3-Dichloropropene	MM. 1,2-Dibromo-3-chloropropane	CCC. tert-Butylbenzene	SSS. o-Xylene
H. 1,1-Dichloroethene	X. Bromoform	NN. Diethyl ether	DDD. 1,2,4-Trimethylbenzene	TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane
I. 1,1-Dichloroethane	Y. 4-Methyl-2-pentanone	OO. 2,2-Dichloropropane	EEE. sec-Butylbenzene	UUU. Benzyl chloride
J. 1,2-Dichloroethene, total	Z. 2-Hexanone	PP. Bromochloromethane	FFF. 1,3-Dichlorobenzene	VVV. 4-Ethyltoluene
K. Chloroform	AA. Tetrachloroethene	QQ. 1,1-Dichloropropane	GGG. p-Isopropyltoluene	WWW. Ethanol
L. 1,2-Dichloroethane	BB. 1,1,2,2-Tetrachloroethane	RR. Dibromomethane	HHH. 1,4-Dichlorobenzene	XXX. Ethyl ether
M. 2-Butanone	CC. Toluene	SS. 1,3-Dichloropropane	III. n-Butylbenzene	
N. 1,1,1-Trichloroethane	DD. Chlorobenzene	TT. 1,2-Dibromoethane	JJJ. 1,2-Dichlorobenzene	
O. Carbon tetrachloride	EE. Ethylbenzene	UU. 1,1,1,2-Tetrachloroethane	KKK. 1,2,4-Trichlorobenzene	
P. Bromodichloromethane	FF. Styrene	VV. Isopropylbenzene	LLL. Hexachlorobutadiene	

Notes:

LDC #: 10609A1
SDG #: 03-3391

VALIDATION FINDINGS WORKSHEET
Continuing Calibration

Page: 1 of 1
Reviewer: F2
2nd Reviewer: R

METHOD: GC/MS VOA (EPA Method 524.2)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
Y N N/A Was a continuing calibration standard analyzed at least once every 12 hours for each instrument?
Y N N/A Were all percent differences (%D) ≤ 30%?

#	Date	Standard ID	Compound	Finding %D (Limit: ≤30.0%)	Associated Samples	Qualifications
	5/29/03	G2684801	E	30.07	A11 + B1K	J/JJ/RP

VALIDATION FINDINGS WORKSHEET
Blanks

LDC #: 10609A1
SDG #: 03-3391

Page: 1 of 1
Reviewer: [Signature]
2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
 N N/A Was a method blank associated with every sample in this SDG?
 N N/A Was a method blank analyzed at least once every 12 hours for each matrix and concentration?
 N N/A Was there contamination in the method blanks? If yes, please see the qualifications below.

Blank analysis date: 5/29/03
Conc. units: ug/l Associated Samples: ALL

Compound	Blank ID	Sample Identification		
	<u>03 G2684</u>	<u>1</u>	<u>2</u>	<u>3</u>
Methylene chloride	<u>5.9</u>	<u>5.3/u</u>	<u>2.8/u</u>	<u>6.1/u</u>
Acetone				
CRQL				

Blank analysis date: _____
Conc. units: _____ Associated Samples: _____

Compound	Blank ID	Sample Identification		
Methylene chloride				
Acetone				
CRQL				

CIRCLED RESULTS WERE NOT QUALIFIED. ALL RESULTS NOT CIRCLED WERE QUALIFIED BY THE FOLLOWING STATEMENT:

Note: Common contaminants such as Methylene chloride, Acetone, 2-Butanone, Carbon disulfide and TICs that were detected in samples within ten times the associated method blank concentration were qualified as not detected, "U". Other contaminants within five times the method blank concentration were also qualified as not detected, "U".

LDC #: 10609A1
SDG #: 03-3391

VALIDATION FINDINGS WORKSHEET

Field Blanks

Page: 1 of 1
Reviewer: #7
2nd reviewer: [Signature]

METHOD: GC/MS VOA (EPA SW-846 Method 8260B)^{524.2}

Y N N/A Were field blanks identified in this SDG?
 Y N N/A Were target compounds detected in the field blanks?

Sample: 3 Field Blank / Trip Blank / Rinsate / Other (circle one)

Compound	Concentration Units (ug/L)
E	6.1
Y	5

Sample: Field Blank / Trip Blank / Rinsate / Other (circle one)

Compound	Concentration Units ()

Sample: Field Blank / Trip Blank / Rinsate / Other (circle one)

Compound	Concentration Units ()

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 28, 2003
LDC Report Date: July 29, 2003
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-3414

Sample Identification

MW-5
MW-8
TB-15-5/28/03
MW-5MS
MW-5MSD

Introduction

This data review covers 5 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 28, 2003
LDC Report Date: July 29, 2003
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 03-3414

Sample Identification

MW-5
MW-8

Introduction

This data review covers 2 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7 and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Iron, Magnesium, Potassium, and Sodium.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Iron Magnesium Potassium	5.1 ug/L 49.4 ug/L 156 ug/L	All samples in SDG 03-3414
ICB/CCB	Calcium Iron Magnesium Potassium	98.11 ug/L 12.65 ug/L 61.02 ug/L 156.87 ug/L	All samples in SDG 03-3414

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
MW-8	Iron	48.7 ug/L	48.7U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards (ICP-MS)

ICP-MS was not utilized in this SDG.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria.

X. ICP Serial Dilution

ICP serial dilution was not required by the method.

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

No field blanks were identified in this SDG.

**NASA JPL
Metals - Data Qualification Summary - SDG 03-3414**

No Sample Data Qualified in this SDG

**NASA JPL
Metals - Laboratory Blank Data Qualification Summary - SDG 03-3414**

SDG	Sample	Analyte	Modified Final Concentration	A or P
03-3414	MW-8	Iron	48.7U ug/L	A

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 33414
 Lab Sample ID: 03-3414-1
 Sample Matrix: Water

Collection Date: 05/28/2003
 Collected by:
 Received Date: 05/28/2003
 Moisture %: -

Sample ID: MW-5
 Sample Type: Field Sample

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	$\mu\text{g/L}$	5	<5	U	F		03M1535E	06/02/03	06/02/03	1	200.9
CALCIUM	7440-70-2	$\mu\text{g/L}$	200	45300		P		03M1525L	05/29/03	05/29/03	1	200.7
IRON	7439-89-6	$\mu\text{g/L}$	50	357		P		03M1525L	05/29/03	05/29/03	1	200.7
MAGNESIUM	7439-95-4	$\mu\text{g/L}$	100	13800		P		03M1525L	05/29/03	05/29/03	1	200.7
POTASSIUM	7440-09-7	$\mu\text{g/L}$	400	2740		P		03M1525L	05/29/03	05/29/03	1	200.7
SODIUM	7440-23-5	$\mu\text{g/L}$	2000	16500		P		03M1525L	05/29/03	05/29/03	1	200.7

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: P - ICP A - FLAA F - GFAA CV - Cold Vapor

MJG/m

LDC #: 10609B4

VALIDATION COMPLETENESS WORKSHEET

Date: 7-25-03

SDG #: 03-3414

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: MG

2nd Reviewer: [Signature]

METHOD: Metals (EPA Method 200.7 / 200.9)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 5-28-03
II.	Calibration	A	
III.	Blanks	SW	
IV.	ICP Interference Check Sample (ICS) Analysis	A	
V.	Matrix Spike Analysis	A	MW-6 MS/MSD (SDG: 03-3444)
VI.	Duplicate Sample Analysis	A	MW-6 DUP (↓)
VII.	Laboratory Control Samples (LCS)	A	LCS/LCSD
VIII.	Internal Standard (ICP-MS)	N	Not utilized
IX.	Furnace Atomic Absorption QC	A	MSA not performed
X.	ICP Serial Dilution	N	Not required
XI.	Sample Result Verification	N	
XII.	Overall Assessment of Data	A	
XIII.	Field Duplicates	N	
XIV.	Field Blanks	N	

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

1	MW-5	W	11		21		31	
2	MW-8	↓	12		22		32	
3			13		23		33	
4			14		24		34	
5			15		25		35	
6			16		26		36	
7			17		27		37	
8			18		28		38	
9			19		29		39	
10			20		30		40	

Notes:

LDC #: 10609B4
 SDG #: 03-3414

VALIDATION FINDINGS WORKSHEET
Sample Specific Element Reference

Page: 1 of 1
 Reviewer: MG
 2nd reviewer: MH

All circled elements are applicable to each sample.

Sample ID	Matrix	Target Analyte List (TAL):
1, 2	W	Al, Sb, (As), Ba, Be, Cd, (Ca), Cr, Co, Cu, (Fe), Pb, (Mg), Mn, Hg, Ni, (K), Se, Ag, (Na), Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
Analysis Method		
ICP	W	Al, Sb, As, Ba, Be, Cd, (Ca), Cr, Co, Cu, (Fe), Pb, (Mg), Mn, Hg, Ni, (K), Se, Ag, (Na), Ti, V, Zn, Mo, B, Si, CN, _____
ICP Trace		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
ICP-MS		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
GFAA	↓	Al, Sb, (As), Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____

Comments: Mercury by CVAA if performed

LDC #: 1060984

SDG #: 03-3414

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: N/A

Sample Concentration units, unless otherwise noted: µg/L Associated Samples: all

Page: 1 of 1
 Reviewer: MG
 2nd Reviewer: MH

VALIDATION FINDINGS WORKSHEET

PB/ICB/CCB QUALIFIED SAMPLES

Analyte	Maximum PB* (mg/kg)	Maximum PB* (µg/L)	Maximum ICB/CCB* (µg/L)	Blank Action Limit	2	Sample Identification
Al						
Sb						
As						
Ba						
Be						
Cd						
Ca			98.11	490.55		
Cr						
Cu						
Fe	5.1		12.65	63.25	48.7	
Pb						
Mg	49.4		61.02	305.10		
Mn						
Hg						
Ni						
K	156		156.87	784.35		
Se						
Ag						
Na						
Tl						
V						
Zn						
B						
Mo						
Sr						

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".
 Note: a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 29, 2003
LDC Report Date: July 29, 2003
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-3444

Sample Identification

MW-6
MW-7
MW-15
MW-6MS
MW-6MSD
MW-6DUP
MW-6DUP2
MW-7MS
MW-7MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7 and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Iron, Magnesium, Potassium, and Sodium.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	73.2 ug/L	All samples in SDG 03-3444
ICB/CCB	Calcium Iron Potassium	222.94 ug/L 13.83 ug/L 164.74 ug/L	All samples in SDG 03-3444

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks.

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards (ICP-MS)

ICP-MS was not utilized in this SDG.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met with the following exceptions:

Diluted Sample	Analyte	%D (Limits)	Associated Samples	Flag	A or P
MW-6L	Potassium	11.1 (≤ 10)	All samples in SDG 03-3444	J (all detects)	A

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

No field blanks were identified in this SDG.

NASA JPL
Metals - Data Qualification Summary - SDG 03-3444

SDG	Sample	Analyte	Flag	A or P	Reason
03-3444	MW-6 MW-7 MW-15	Potassium	J (all detects)	A	ICP serial dilution (%D)

NASA JPL
Metals - Laboratory Blank Data Qualification Summary - SDG 03-3444

No Sample Data Qualified in this SDG

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.

Project No: 04-4428.10

Collection Date: 05/29/2003

Project ID: JPL

Service ID: 33444

Collected by:

Lab Sample ID: 03-3444-1

Received Date: 05/29/2003

Sample ID: MW-6

Sample Matrix: Water

Moisture %: -

Sample Type: Field Sample

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	µg/L	5	< 5	U	F		03M1535E	06/02/03	06/02/03	1	200.9
CALCIUM	7440-70-2	µg/L	200	162000		P		03M1537M	06/02/03	06/02/03	1	200.7
IRON	7439-89-6	µg/L	50	785		P		03M1537M	06/02/03	06/02/03	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	51200		P		03M1537M	06/02/03	06/02/03	1	200.7
POTASSIUM	7440-09-7	µg/L	400	2690	J	P		03M1537M	06/02/03	06/02/03	1	200.7
SODIUM	7440-23-5	µg/L	2000	35100		P		03M1537M	06/02/03	06/02/03	1	200.7

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor

C Qualifier: U - Not Detected or less than IDL

B - Less than RL (PQL, EQL or CRDL), but greater than IDL.

Q Qualifier: N - Spike recovery out of control

* - Duplicate analysis out of control

W - Post digestion spike for GFAA out of control

E - Serial dilution difference out of control

M Qualifier: P - ICP

A - FLAA

F - GFAA

CV - Cold Vapor

8/4/03

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL
 Sample ID: MW-7
 Sample Type: Field Sample

Project No: 04-4428.10
 Service ID: 33444
 Lab Sample ID: 03-3444-2
 Sample Matrix: Water

Collection Date: 05/29/2003
 Collected by:
 Received Date: 05/29/2003
 Moisture %: -

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	µg/L	5	< 5	U	F		03M1535E	06/02/03	06/02/03	1	200.9
CALCIUM	7440-70-2	µg/L	200	54200		P		03M1537M	06/02/03	06/02/03	1	200.7
IRON	7439-89-6	µg/L	50	138		P		03M1537M	06/02/03	06/02/03	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	17200		P		03M1537M	06/02/03	06/02/03	1	200.7
POTASSIUM	7440-09-7	µg/L	400	2250	J	P		03M1537M	06/02/03	06/02/03	1	200.7
SODIUM	7440-23-5	µg/L	2000	17800		P		03M1537M	06/02/03	06/02/03	1	200.7

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor

C Qualifier: U - Not Detected or less than IDL

B - Less than RL (PQL, EQL or CRDL), but greater than IDL.

Q Qualifier: N - Spike recovery out of control

* - Duplicate analysis out of control

W - Post digestion spike for GFAA out of control

E - Serial dilution difference out of control

M Qualifier: P - ICP

A - FLAA

F - GFAA

CV - Cold Vapor

8/4/03

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 33444
 Lab Sample ID: 03-3444-3
 Sample Matrix: Water

Collection Date: 05/29/2003
 Collected by:
 Received Date: 05/29/2003
 Moisture %: -

Sample ID: MW-15
 Sample Type: Field Sample

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	µg/L	5	2.1	B	F		03M1535E	06/02/03	06/02/03	1	200.9
CALCIUM	7440-70-2	µg/L	200	57400		P		03M1537M	06/02/03	06/02/03	1	200.7
IRON	7439-89-6	µg/L	50	389		P		03M1537M	06/02/03	06/02/03	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	18400		P		03M1537M	06/02/03	06/02/03	1	200.7
POTASSIUM	7440-09-7	µg/L	400	2800	J	P		03M1537M	06/02/03	06/02/03	1	200.7
SODIUM	7440-23-5	µg/L	2000	24500		P		03M1537M	06/02/03	06/02/03	1	200.7

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: P - ICP A - FLAA F - GFAA CV - Cold Vapor

M 8/4/03

LDC #: 10609C4

VALIDATION COMPLETENESS WORKSHEET

Date: 7-25-03

SDG #: 03-3444

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: MG

2nd Reviewer: MH

METHOD: Metals (EPA Method 200.7 / 200.9)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: 5-29-03
II.	Calibration	A	
III.	Blanks	SW	
IV.	ICP Interference Check Sample (ICS) Analysis	A	
V.	Matrix Spike Analysis	A	
VI.	Duplicate Sample Analysis	A	
VII.	Laboratory Control Samples (LCS)	A	LCS/LCSD
VIII.	Internal Standard (ICP-MS)	N	Not utilized
IX.	Furnace Atomic Absorption QC	A	MSA not performed
X.	ICP Serial Dilution	SW	
XI.	Sample Result Verification	N	
XII.	Overall Assessment of Data	A	
XIII.	Field Duplicates	N	
XIV.	Field Blanks	N	

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

1	MW-6	w	11	PBW	21		31	
2	MW-7		12		22		32	
3	MW-15		13		23		33	
4	TB-16-5/29/03		14		24		34	
5	MW-6MS		15		25		35	
6	MW-6MSD		16		26		36	
7	MW-6DUP		17		27		37	
8	MW-6DUP2		18		28		38	
9	MW-7MS		19		29		39	
10	MW-7MSD	v	20		30		40	

Notes: _____

LDC #: 10609C4
SDG #: 03-3444

VALIDATION FINDINGS WORKSHEET
Sample Specific Element Reference

Page: 1 of 1
Reviewer: MG
2nd reviewer: MH

All circled elements are applicable to each sample.

Sample ID	Matrix	Target Analyte List (TAL)
1 → 3	W	Al, Sb, (As) Ba, Be, Cd, (Ca) Cr, Co, Cu, (Fe) Pb, (Mg) Mn, Hg, Ni, (K) Se, Ag, (Na) Ti, V, Zn, Mo, B, Si, CN', ____
QC 5 → 7 9, 10	↓	Al, Sb, (As) Ba, Be, Cd, (Ca) Cr, Co, Cu, (Fe) Pb, (Mg) Mn, Hg, Ni, (K) Se, Ag, (Na) Ti, V, Zn, Mo, B, Si, CN', ____
↓ 8	↓	Al, Sb, As, Ba, Be, Cd, (Ca) Cr, Co, Cu, (Fe) Pb, (Mg) Mn, Hg, Ni, (K) Se, Ag, (Na) Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
Analyte Method		
ICP	W	Al, Sb, As, Ba, Be, Cd, (Ca) Cr, Co, Cu, (Fe) Pb, (Mg) Mn, Hg, Ni, (K) Se, Ag, (Na) Ti, V, Zn, Mo, B, Si, CN', ____
ICP Trace	↓	Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
ICP-MS	↓	Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____
GFAA	↓	Al, Sb, (As) Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN', ____

Comments: Mercury by CVAA if performed

LDC #: 10609C4

SDG #: 03-3444

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: N/A

Sample Concentration units, unless otherwise noted: $\mu\text{g}/\text{L}$ Associated Samples: all (No findings)

VALIDATION FINDINGS WORKSHEET

PB/ICB/CCB QUALIFIED SAMPLES

Page: 1 of 1

Reviewer: MG

2nd Reviewer: MH

Analyte	Maximum PB* (mg/kg)	Maximum PB* (ug/L)	Maximum ICB/CCB* (ug/L)	Blank Action Limit	Sample Identification
Al					
Sb					
As					
Ba					
Be					
Cd					
Ca			222.94	1114.70	
Cr					
Co					
Cu					
Fe			13.83	69.15	
Pb					
Mg					
Mn					
Hg					
Ni					
K		73.2	164.74	823.70	
Se					
Ag					
Na					
Tl					
V					
Zn					
B					
Mo					
Sr					

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".

Note: a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

LDC #: 1060924
 SDG #: 03-3444

VALIDATION FINDINGS WORKSHEET
ICP Serial Dilution

Page: 1 of 1
 Reviewer: MG
 2nd Reviewer: MH

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
 N N/A if analyte concentrations were > 50X the IDL, was an ICP serial dilution analyzed?
 N N/A Were ICP serial dilution percent differences (%D) ≤ 10%?
 N N/A Is there evidence of negative interference? if yes, professional judgement will be used to qualify the data.
LEVEL IV ONLY:
 N N/A Were recalculated results acceptable? See Level IV Recalculation Worksheet for recalculations.

#	Diluted Sample ID	Matrix	Analyte	%D	Associated Samples	Qualifications
1		water	K	11.1	all	5 dots / A

Comments:

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: May 30, 2003
LDC Report Date: July 29, 2003
Matrix: Water
Parameters: Metals
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 03-3465

Sample Identification

MW-1
MW-9
MW-10

Introduction

This data review covers 3 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Methods 200.7 and 200.9 for Metals. The metals analyzed were Arsenic, Calcium, Iron, Magnesium, Potassium, and Sodium.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

Method Blank ID	Analyte	Maximum Concentration	Associated Samples
PB (prep blank)	Potassium	73.2 ug/L	All samples in SDG 03-3465
ICB/CCB	Calcium Iron Potassium	222.94 ug/L 13.83 ug/L 164.74 ug/L	All samples in SDG 03-3465

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Analyte	Reported Concentration	Modified Final Concentration
MW-10	Iron	43.6 ug/L	43.6U ug/L

IV. ICP Interference Check Sample (ICS) Analysis

The frequency of analysis was met.

The criteria for analysis were met.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards (ICP-MS)

ICP-MS was not utilized in this SDG.

IX. Furnace Atomic Absorption QC

All graphite furnace atomic absorption QC were within validation criteria.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met.

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

No field blanks were identified in this SDG.

NASA JPL
Metals - Data Qualification Summary - SDG 03-3465

No Sample Data Qualified in this SDG

NASA JPL
Metals - Laboratory Blank Data Qualification Summary - SDG 03-3465

SDG	Sample	Analyte	Modified Final Concentration	A or P
03-3465	MW-10	Iron	43.6U ug/L	A

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL

Project No: 04-4428.10
 Service ID: 33465
 Lab Sample ID: 03-3465-1
 Sample Matrix: Water

Collection Date: 05/30/2003
 Collected by:
 Received Date: 05/30/2003
 Moisture %: -

Sample ID: MW-1
 Sample Type: Field Sample

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	µg/L	5	< 5	U	F		03M1535E	06/02/03	06/02/03	1	200.9
CALCIUM	7440-70-2	µg/L	200	65100		P		03M1537M	06/02/03	06/02/03	1	200.7
IRON	7439-89-6	µg/L	50	72.0		P		03M1537M	06/02/03	06/02/03	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	20000		P		03M1537M	06/02/03	06/02/03	1	200.7
POTASSIUM	7440-09-7	µg/L	400	2730		P		03M1537M	06/02/03	06/02/03	1	200.7
SODIUM	7440-23-5	µg/L	2000	24600		P		03M1537M	06/02/03	06/02/03	1	200.7

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: P - ICP A - FLAA F - GFAA CV - Cold Vapor

Handwritten signature/initials: 5/31/03

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
Project ID: JPL

Project No: 04-4428.10
Service ID: 33465
Lab Sample ID: 03-3465-2
Sample Matrix: Water

Collection Date: 05/30/2003
Collected by:
Received Date: 05/30/2003
Moisture %: ~

Sample ID: MW-9
Sample Type: Field Sample

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	µg/L	5	2.1	B	F		03M1535E	06/02/03	06/02/03	1	200.9
CALCIUM	7440-70-2	µg/L	200	51600		P		03M1537M	06/02/03	06/02/03	1	200.7
IRON	7439-89-6	µg/L	50	832		P		03M1537M	06/02/03	06/02/03	1	200.7
MAGNESIUM	7439-95-4	µg/L	100	16300		P		03M1537M	06/02/03	06/02/03	1	200.7
POTASSIUM	7440-09-7	µg/L	400	2750		P		03M1537M	06/02/03	06/02/03	1	200.7
SODIUM	7440-23-5	µg/L	2000	19400		P		03M1537M	06/02/03	06/02/03	1	200.7

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
M Qualifier: P - ICP A - FLAA F - GFAA CV - Cold Vapor

Handwritten: 8/4/03

Applied P & Ch Laboratory
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL
 Sample ID: MW-10
 Sample Type: Field Sample

Project No: 04-4428.10
 Service ID: 33465
 Lab Sample ID: 03-3465-3
 Sample Matrix: Water

Collection Date: 05/30/2003
 Collected by:
 Received Date: 05/30/2003
 Moisture %: -

Element Name	CAS No	Unit	RL	Result	C	M	Q	Batch	D-Date	A-Date	DF	Method
ARSENIC	7440-38-2	μg/L	5	< 5	U	F		03M1535E	06/02/03	06/02/03	1	200.9
CALCIUM	7440-70-2	μg/L	200	121000			P	03M1537M	06/02/03	06/02/03	1	200.7
IRON	7439-89-6	μg/L	50	43.6 <i>u</i>	B	P		03M1537M	06/02/03	06/02/03	1	200.7
MAGNESIUM	7439-95-4	μg/L	100	38800			P	03M1537M	06/02/03	06/02/03	1	200.7
POTASSIUM	7440-09-7	μg/L	400	3130			P	03M1537M	06/02/03	06/02/03	1	200.7
SODIUM	7440-23-5	μg/L	2000	26200			P	03M1537M	06/02/03	06/02/03	1	200.7

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: P - ICP A - FLAA F - GFAA CV - Cold Vapor

18/4/03

LDC #: 10609D4

VALIDATION COMPLETENESS WORKSHEET

Date: 7-28-03

SDG #: 03-3465

Level III

Page: 1 of 1

Laboratory: Applied P & Ch Laboratory

Reviewer: MG

2nd Reviewer: ML

METHOD: Metals (EPA Method 200.7 / 200.9)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

	Validation Area		Comments
I.	Technical holding times	A	Sampling dates: <u>5-30-03</u>
II.	Calibration	A	
III.	Blanks	SW	
IV.	ICP Interference Check Sample (ICS) Analysis	A	
V.	Matrix Spike Analysis	A	MS/MSD MW-6 (SDG: 03-3444)
VI.	Duplicate Sample Analysis	A	MW-6 DUP (↓)
VII.	Laboratory Control Samples (LCS)	A	LCS/LCSD
VIII.	Internal Standard (ICP-MS)	N	Not utilized
IX.	Furnace Atomic Absorption QC	A	MSA not performed
X.	ICP Serial Dilution	A	
XI.	Sample Result Verification	N	
XII.	Overall Assessment of Data	A	
XIII.	Field Duplicates	N	
XIV.	Field Blanks	N	

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples:

1	MW-1	W	11	21	31
2	MW-9	↓	12	22	32
3	MW-10	↓	13	23	33
4	PBW		14	24	34
5			15	25	35
6			16	26	36
7			17	27	37
8			18	28	38
9			19	29	39
10			20	30	40

Notes: _____

LDC #: 10609D4
SDG #: 03-3465

VALIDATION FINDINGS WORKSHEET
Sample Specific Element Reference

Page: 1 of 1
Reviewer: MG
2nd reviewer: fuy

All circled elements are applicable to each sample.

Sample ID	Matrix	Target Analyte List (TAL)
1 → 3	W	Al, Sb, (As)Ba, Be, Cd, (Ca)Cr, Co, Cu, (Fe)Pb, (Mg)Mn, Hg, Ni, (K)Se, Ag, (Na)Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
		Analysis Method
ICP	W	Al, Sb, As, Ba, Be, Cd, (Ca)Cr, Co, Cu, (Fe)Pb, (Mg)Mn, Hg, Ni, (K)Se, Ag, (Na)Ti, V, Zn, Mo, B, Si, CN, _____
ICP Trace	↓	Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
ICP-MS	↓	Al, Sb, As, Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____
GFAA	↓	Al, Sb, (As)Ba, Be, Cd, Ca, Cr, Co, Cu, Fe, Pb, Mg, Mn, Hg, Ni, K, Se, Ag, Na, Ti, V, Zn, Mo, B, Si, CN, _____

Comments: Mercury by CVAA if performed

LDC #: 10609D4
 SDG #: 03-3465

VALIDATION FINDINGS WORKSHEET
 PB/ICB/CCB QUALIFIED SAMPLES

Page: 1 of 1
 Reviewer: MG
 2nd Reviewer: MJ

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: NA
 Sample Concentration units, unless otherwise noted: µg/L Associated Samples: all

Analyte	Maximum PB* (mg/Kg)	Maximum PB* (µg/L)	Maximum ICB/CCB* (µg/L)	Blank Action Limit	3	Sample Identification
Al						Al
Sb						Sb
As						As
Ba						Ba
Be						Be
Cd						Cd
Ca			222.94	114.70		Ca
Cr						Cr
Cc						Co
Cu						Cu
Fe			13.83	69.15	43.6	Fe
Pb						Pb
Mg						Mg
Mn						Mn
Hg						Hg
Ni						Ni
K		73.2	164.74	823.70		K
Se						Se
Ag						Ag
Na						Na
Tl						Tl
V						V
Zn						Zn
B						B
Mo						Mo
Sr						Sr

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".
 Note: a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.