

5A
VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name : APPLIED P & CH LAB Contract: _____
 Lab Code: _____ Case No.: _____ SAS No.: _____ SDG No.: 034406
 Lab File ID: G 3415 P01 BFB Injection Date: 7/22/03
 Instrument ID: GCMS-A BFB Injection Time: 1708
 GC Column: HP-VOC ID: 0.20 (mm) Heated Purge: (Y/N) N

m/e	ION ABUNDANCE CRITERIA	%RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.3
75	30.0 - 60.0% of mass 95	44.1
95	Base peak, 100% relative abundance	100.0
96	5.0 - 9.0% of mass 95	7.0
173	Less than 2.0% of mass 174	0.6 (0.7)1
174	50.0 - 100.0% of mass 95	81.3
175	5.0 - 9.0% of mass 174	5.8 (7.1)1
176	95.0 - 101.0% of mass 174	80.2 (98.6)1
177	5.0 - 9.0% of mass 176	6.1 (7.6)2

1-Value is % mass 174

2-Value is % mass 176

This check applies to the following SAMPLES, MS, MSD, BLANKS and STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
01	VSTD0003	20-0003	20-00003.D	7/22/03	1925
02	VSTD002	20-0002	20-0002.D	7/22/03	2111
03	VSTD010	20-0010	20-0010.D	7/22/03	2137
04	VSTD020	20-0020	20-0020.D	7/22/03	2256
05	VSTD040	20-0040	20-0040.D	7/23/03	0043
06	VSTD060	20-0060	20-0060.D	7/23/03	0245
07					
08					
09					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Initial Calibration

Calibration Files
 .3 =2-00003.D 2 =2-0002A.D 10 =2-00010.D
 20 =2-00020.D 40 =2-00040.D 60 =2-00060.D

Compound	ISTD						Avg	%RSD
	3	2	10	20	40	60		
1) I Fluorobenzene	0.279	0.253	0.228	0.251	0.250	0.246	0.251	6.61
2) di-Cl-di-F-m	0.300	0.265	0.208	0.221	0.218	0.200	0.235	16.51
3) p Chloromethan	0.116	0.138	0.126	0.136	0.137	0.133	0.131	6.45
4) F114	0.268	0.267	0.219	0.239	0.240	0.223	0.242	8.61
5) C vinyl chlori	0.143	0.128	0.105	0.113	0.104	0.106	0.116	13.63
6) bromomethane	0.162	0.171	0.136	0.131	0.138	0.129	0.144	12.08
7) chloroethane	0.426	0.367	0.310	0.357	0.356	0.338	0.359	10.76
8) tri-Cl-F-met	0.055	0.041	0.042	0.042	0.041	0.040	0.044	14.16
9) Acetonitrile	0.034	0.034	0.026	0.025	0.024	0.022	0.028	18.73
10) acrolein	0.106	0.053	0.028	0.024	0.024	0.021	0.043	77.29
11) acetone X	0.162	0.157	0.117	0.117	0.110	0.098	0.127	20.67
12) ethyl ether	0.345	0.312	0.258	0.277	0.275	0.253	0.287	12.31#
13) M, Cl3 11-dichloro	0.266	0.286	0.244	0.262	0.258	0.205	0.253	10.75
14) Iodomethane	0.313	0.246	0.193	0.211	0.205	0.185	0.225	21.21
15) F-113	0.053	0.052	0.042	0.041	0.040	0.040	0.045	13.61
16) acrylonitril	0.838	0.780	0.654	0.729	0.722	0.649	0.729	10.00
17) carbon disul	0.015	0.008	0.005	0.006	0.007	0.006	0.008	46.15
18) Isopropyl Al	1.323	0.482	0.257	0.251	0.242	0.244	0.467	92.15
19) methylene ch	0.297	0.305	0.248	0.262	0.256	0.253	0.270	8.99
20) t-12-di-Cl-e	0.474	0.509	0.419	0.427	0.427	0.442	0.450	7.81
21) t-Bu-Me-ethe	0.014	0.010	0.010	0.010	0.012	0.013	0.012	16.51
22) Tert butyl a	0.549	0.412	0.417	0.307	0.282	0.282	0.393	26.97
23) allyl chlori	0.518	0.484	0.380	0.407	0.403	0.409	0.433	12.52
24) p 11-dichloro	0.020	0.014	0.014	0.016	0.016	0.017	0.017	14.53
25) propionitril	0.296	0.311	0.252	0.263	0.257	0.255	0.272	9.12
26) c-12-di-Cl-e	0.392	0.360	0.280	0.290	0.277	0.262	0.310	16.98
27) 22-Dichlorop	0.158	0.150	0.117	0.123	0.122	0.122	0.132	13.12
28) Br-Cl-methan								

(#) = Out of Range
 E524A002.M

Thu Jul 24 12:41:07 2003

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Initial Calibration

Calibration Files
 .3 =2-00003.D 2 =2-0002A.D 10 =2-00010.D
 20 =2-00020.D 40 =2-00040.D 60 =2-00060.D

Compound	.3	2	10	20	40	60	Avg	%RSD
29) C 25 chloroform	0.724	0.528	0.396	0.413	0.403	0.406	0.478	27.21#
30) 26 tetrahydrofu	0.036	0.035	0.031	0.031	0.032	0.032	0.033	6.66
31) 98 Disopropyl	0.648	0.792	0.634	0.653	0.636	0.641	0.667	9.23
32) S 27 Di-Br-F-Me	0.294	0.230	0.240	0.239	0.243	0.249	0.249	10.27
33) 99 ETBE	0.437	0.534	0.485	0.524	0.532	0.549	0.510	8.20
34) S 29 1,2-Di-Cl-Et	0.248	0.186	0.191	0.191	0.192	0.192	0.202	12.96
35) 30 12-dichloroe	0.094	0.093	0.070	0.070	0.070	0.070	0.078	15.50
36) 32 vinyl acetat	0.319	0.377	0.321	0.291	0.285	0.280	0.312	11.53
37) 92 Nitro Methan	0.006	0.005	0.005	0.005	0.006	0.006	0.006	9.09
38) 33 2-butanoneME	0.081	0.061	0.044	0.043	0.042	0.041	0.052	30.99
39) 93 Ethyl Acetat	0.175	0.135	0.110	0.107	0.107	0.110	0.124	22.07
40) 34 111-trichlor	0.479	0.436	0.353	0.387	0.385	0.378	0.403	11.36
41) 35 11-Di-Cl-pro	0.271	0.306	0.284	0.325	0.324	0.316	0.304	7.40
42) M 36 benzene	1.117	1.160	0.923	0.983	0.960	0.948	1.015	9.69
43) 37 CCl4	0.434	0.397	0.329	0.361	0.358	0.350	0.371	10.20
44) 100 Isobutyl al	0.002	0.005	0.004	0.004	0.005	0.005	0.004	30.78
45) 38 thiophene	0.471	0.573	0.471	0.505	0.504	0.503	0.505	7.36
46) C 39 12-di-Cl-pro	0.247	0.251	0.204	0.220	0.221	0.220	0.227	7.93#
47) M 40 trichloroeth	0.354	0.335	0.274	0.312	0.315	0.309	0.316	8.59
48) 41 dibromometha	0.177	0.159	0.126	0.133	0.134	0.135	0.144	13.68
49) 101 TAME	0.352	0.451	0.420	0.462	0.472	0.498	0.442	11.61
50) 42 Br-di-Cl-met	0.433	0.359	0.280	0.298	0.298	0.297	0.327	17.81
51) 43 Me-methacryl	0.065	0.098	0.100	0.112	0.117	0.121	0.102	20.03
52) 44 2-ClEt-Vl-et	0.010	0.016	0.019	0.025	0.029		0.020	37.07
53) 45 c-13-di-Cl-p	0.312	0.358	0.307	0.333	0.335	0.336	0.330	5.59
54) 46 t-1,3-dichlo	0.227	0.272	0.247	0.270	0.278	0.281	0.262	8.00
55) I 47 Chlorobezene-d5	0.288	0.275	0.213	0.222	0.224	0.226	0.241	13.17
56) 48 112-tri-Cl-E								

(#) = Out of Range
 E524A002.M Thu Jul 24 12:41:08 2003

Response Factor Report GCMS-A

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Initial Calibration

Calibration Files
 .3 =2-00003.D 2 =2-0002A.D 10 =2-00010.D
 20 =2-00020.D 40 =2-00040.D 60 =2-00060.D

Compound	.3	2	10	20	40	60	Avg	%RSD
57) 49 1,3-di-Cl-pro	0.436	0.432	0.347	0.357	0.359	0.353	0.381	10.91
58) 50 Et methacryl	0.189	0.269	0.272	0.287	0.301	0.321	0.273	16.74
59) 51 di-Br-Cl-met	0.422	0.341	0.275	0.296	0.304	0.309	0.325	16.16
60) 52 bromoform	0.222	0.192	0.159	0.171	0.181	0.187	0.185	11.70
61) 53 1,4-dichloro	0.405	0.396	0.312	0.330	0.340	0.344	0.354	10.53
62) 54 MIBK	0.132	0.144	0.134	0.144	0.155	0.166	0.146	8.89
63) 55 toluene-d8	1.345	1.364	1.130	1.244	1.257	1.223	1.260	6.78
64) 56 toluene	1.448	1.627	1.315	1.441	1.448	1.412	1.448	6.98
65) 57 2-hexanone X	0.100	0.109	0.092	0.094	0.098	0.100	0.099	5.95
66) 58 1,2-dibromoet	0.274	0.254	0.210	0.222	0.233	0.235	0.238	9.68
67) 59 tetra-Cl-eth	0.496	0.464	0.378	0.427	0.424	0.408	0.433	9.68
68) 60 chlorobenzen	1.297	1.156	0.893	0.967	0.959	0.919	1.032	15.48
69) 61 1,1,1,2-tetra-C	0.447	0.412	0.317	0.340	0.348	0.348	0.369	13.47
70) 62 1,4-Dichlorobenzen	0.269	0.309	0.291	0.356	0.374	0.358	0.326	12.94
71) 63 1-chlorohexa	3.072	3.263	2.780	3.211	3.292	3.156	3.129	6.01#
72) 64 Et-Bz	2.292	2.598	2.175	2.438	2.422	2.268	2.365	6.37
73) 65 m/p-Xylenes	1.472	1.978	1.707	1.902	1.915	1.818	1.798	10.32
74) 66 styrene	1.955	2.508	2.249	2.527	2.563	2.438	2.373	9.83
75) 67 o-xylene	0.659	0.575	0.456	0.475	0.501	0.503	0.528	14.34
76) 68 1,1,2,2-Tetra-C	0.186	0.176	0.138	0.146	0.151	0.152	0.158	11.76
77) 69 1,2,3-tri-Cl-P	0.953	0.876	0.706	0.798	0.823	0.806	0.827	10.01
78) 70 4-Br-1-F-Bz	2.262	3.010	2.857	3.361	3.462	3.290	3.040	14.58
79) 71 isopropylben	0.840	0.881	0.719	0.807	0.842	0.810	0.817	6.73
80) 72 bromobenzene	0.038	0.078	0.075	0.082	0.094	0.095	0.077	27.12
81) 92 t-1,4-dichlo	0.759	0.999	0.867	1.019	1.061	1.009	0.952	12.10
82) 73 n-propylbenz	0.756	0.887	0.758	0.870	0.905	0.873	0.841	7.90
83) 74 2-Cl-Toluene	0.853	0.955	0.757	0.860	0.880	0.830	0.856	7.52
84) 75 4-Cl-Toluene								

✓

0.997

0.999

0.997

(#) = Out of Range
 E524A002.M Thu Jul 24 12:41:09 2003

Response Factor Report GCMS-A

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTF Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Initial Calibration

Calibration Files
 .3 =2-00003.D 2 =2-0002A.D 10 =2-00010.D
 20 =2-00020.D 40 =2-00040.D 60 =2-00060.D

Compound	.3	2	10	20	40	60	Avg	%RSD
85) 76 135-tri-Me-B	2.031	2.864	2.465	2.815	2.851	2.678	2.617	12.38
86) 77 4-iso-Pr-tol	2.280	3.155	2.726	3.159	3.220	2.998	2.923	12.39
87) 78 124-tri-Me-B	2.267	3.012	2.517	2.854	2.932	2.773	2.726	10.34
88) 79 tert-butylbe	2.041	2.391	2.526	2.636	2.748	2.655	2.499	10.24
89) 80 13-DCB	1.989	1.888	1.484	1.667	1.680	1.592	1.717	10.95
90) 81 sec-butylben	2.715	3.632	3.274	3.857	3.936	3.690	3.517	12.95
91) 82 14-DCB	2.240	1.859	1.457	1.644	1.684	1.613	1.749	15.60
92) 83 Cl-benzyl	0.164	0.159	0.155	0.169	0.181	0.165	0.165	5.41
93) 84 12-DCB	1.762	1.734	1.326	1.464	1.479	1.408	1.529	11.67
94) 85 n-butylbenze	0.528	0.789	0.731	0.875	0.910	0.843	0.779	17.73
95) 86 12-diBr-2-Cl	0.110	0.104	0.091	0.102	0.118	0.120	0.108	10.09
96) 87 124-tri-Cl-B	0.843	0.926	0.884	1.065	1.130	1.091	0.990	12.16
97) 88 naphthalene	2.218	1.820	1.426	1.651	1.832	1.831	1.796	14.50
98) 89 hx-Cl-butadi	0.644	0.621	0.514	0.599	0.629	0.588	0.599	7.74
99) 90 123-Tri-Cl-B	0.849	0.875	0.777	0.914	0.969	0.944	0.888	7.86

Y

0.999

0.999

(#) = Out of Range
 E524A002.M Thu Jul 24 12:41:09 2003

INITIAL CALIBRATION SUMMARY

1677

Method File: E524A002
 Last Calibration Update: Thu Jul 24 12:40:35 2003

Level	File Name	ID
Level 1	File Name	Level 1 ID
Level 2	File Name	Level 2 ID
Level 3	File Name	Level 3 ID
Level 4	File Name	Level 4 ID
Level 5	File Name	Level 5 ID
Level 6	File Name	Level 6 ID
Level 7	File Name	Level 7 ID

Compound Name	Level 1		Level 2		Level 3		Level 4		Level 5		Level 6		Level 7		Coeff X ⁿ 0	Coeff X ⁿ 1 / ave RF	Coeff X ⁿ 2	R ² / RSD
	Response	Level 1 ID	Response	Level 2 ID	Response	Level 3 ID	Response	Level 4 ID	Response	Level 5 ID	Response	Level 6 ID	Response	Level 7 ID				
1 Fluorobenzene	919393	1006828	1003977	1015499	990930	1444214	980047	-1	0.0000	0.2510	0.0000	0.0661	0.9966					
3 di-Cl-di-F-methane	7698	51044	228452	508778	990705	1174857	1174857	-1	0.0169	0.2027	0.0000	0.9966	0.9966					
4 Chloromethane	8264	53373	208536	448729	864850	783345	783345	-1	0.0000	0.1313	0.0000	0.0645	0.0645					
2 F114	3209	27835	126949	275940	544934	1312837	1312837	-1	0.0000	0.2424	0.0000	0.0861	0.0861					
5 vinyl chloride	7381	53695	219399	484855	949966	626111	626111	-1	0.0000	0.1165	0.0000	0.1363	0.1363					
6 bromomethane	3952	25743	105105	228520	411844	758703	758703	-1	0.0000	0.1443	0.0000	0.1208	0.1208					
7 chloroethane	4460	34347	136103	266250	546077	1986417	1986417	-1	0.0000	0.3591	0.0000	0.1076	0.1076					
8 tri-Cl-F-methane	11754	73985	311036	724847	1412342	2365932	2365932	-1	0.0000	0.0439	0.0000	0.1416	0.1416					
91 Acetonitrile X10	-1	110646	413272	845923	1639456	1313901	1313901	-1	0.0306	0.0223	0.0000	0.9978	0.9978					
9 acrolein X10	9511	67561	262941	513017	937791	1241915	1241915	-1	0.0622	0.0208	0.0000	0.9944	0.9944					
11 acetone X10	29124	106138	281400	497276	943071	2883396	2883396	-1	0.0885	0.0991	0.0000	0.9929	0.9929					
12 ethyl ether X5	22348	158277	587938	1189867	2188257	1490545	1490545	-1	0.0000	0.2867	0.0000	0.1231	0.1231					
13 11-dichloroethene	9522	62800	259067	562011	1088836	1206486	1206486	-1	0.0000	0.2534	0.0000	0.1075	0.1075					
14 iodomethane	7324	57536	244664	532902	1022337	1086062	1086062	-1	0.0189	0.1881	0.0000	0.9947	0.9947					
15 F-113	8636	49466	193654	427782	814332	2352958	2352958	-1	0.0000	0.0447	0.0000	0.1361	0.1361					
16 acrylonitrile X10	14700	104060	420553	831510	1601521	3817157	3817157	-1	-0.0027	0.0064	0.0000	0.9954	0.9954					
17 carbon disulfide	23125	156993	656825	1480814	2862257	1435510	1435510	-1	0.0322	0.2370	0.0000	0.9996	0.9996					
94 Isopropyl Alcoholx10	4021	15273	46223	124874	265806	1488348	1488348	-1	0.0000	0.2701	0.0000	0.0899	0.0899					
18 methylene chloride	36501	97000	257935	510536	960233	2598506	2598506	-1	0.0000	0.4496	0.0000	0.0781	0.0781					
19 t-12-di-Cl-ethene	8180	61414	249312	531623	1015155	785022	785022	-1	0.0000	0.0135	0.0000	0.9918	0.9918					
20 t-Bu-Me-ether	13066	102540	420692	866584	1693079			-1										
95 Tert butyl alcoholx10	-1	28855	98345	205914	482970			-1	-0.0326		0.0000							

94 allyl chloride	-1	110646	413272	845923	1216384	1660968	-1	0.0000	0.3934	0.0000	0.2697
21 11-dichloroethane	14292	97373	381929	825889	1598207	2404444	-1	0.0000	0.4335	0.0000	0.1252
97 propionitrile	-1	4103	13637	33328	64498	101359	-1	0.0000	0.0168	0.0000	0.1453
22 c-12-di-Cl-ethene	8154	62676	253306	533706	1019103	1499328	-1	0.0000	0.2723	0.0000	0.0912
23 22-Dichloropropane	10809	72464	281116	589735	1098975	1539718	-1	0.0244	0.2629	0.0000	0.9982
24 Br-Cl-methane	4359	30212	117713	250220	483734	716534	-1	0.0000	0.1321	0.0000	0.1312
25 chloroform	19972	106303	398005	838697	1597118	2387736	-1	0.0108	0.4031	0.0000	0.9998
26 tetrahydrofuranX5	4965	35404	154519	318868	624422	946730	-1	0.0000	0.0328	0.0000	0.0666
98 Diisopropyl ether	17882	159533	636576	1326012	2521921	3767835	-1	0.0000	0.6674	0.0000	0.0923
27 Di-Br-F-Me (surr)	-1	59230	230740	486919	949109	1428311	-1	0.0000	0.2492	0.0000	0.1027
99 ETBE	12061	107588	486462	1064815	2106754	3231163	-1	0.0000	0.5102	0.0000	0.0820
29 1,2-Di-Cl-Et-d4 (S1)	-1	50036	187219	388784	757790	411282	-1	0.0020	0.2018	0.0000	0.1296
30 12-dichloroethane	2582	18795	70404	142937	277276	1126883	-1	0.0000	0.0695	0.0000	0.9999
32 vinyl acetate X5	43949	379164	1608933	2957025	5653569	8226056	-1	0.0000	0.3120	0.0000	0.1153
92 Nitro Methane(x10)	-1	12060	47784	111334	237578	325914	-1	0.0000	0.0056	0.0000	0.0909
33 2-butanoneMEK X10	22369	123309	444332	875377	1648493	2411259	-1	0.0340	0.0406	0.0000	0.9998
93 Ethyl Acetate x2	9671	54370	220002	434881	845931	1295021	-1	0.0014	0.1088	0.0000	0.9995
34 111-trichloroethane	13204	87716	354801	785718	1524561	2222673	-1	0.0000	0.4029	0.0000	0.1136
35 11-Di-Cl-propene	7463	61583	285223	660985	1285689	1859527	-1	0.0000	0.3044	0.0000	0.0740
36 benzene	30808	233485	926403	1995763	3805709	5575015	-1	0.0000	1.0150	0.0000	0.0969
37 CCl4	11983	79844	329846	732441	1420059	2060040	-1	0.0000	0.3715	0.0000	0.1020

Compound Name	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6	Level 7	Coeff Xv0	Coeff Xv1 / ave RF	Coeff Xv2	Rv2 / RSD
	Response	Response	Response	Response	Response	Response	Response				
100 Isobutyl alcoholx10	457	10371	40745	87337	182637	287972	-1	-----	-----	-----	-----
38 thiophene	12985	115327	473132	1025220	1999457	2960047	-1	0.0000	0.5046	0.0000	0.0736
39 12-di-Cl-propane	6808	50516	204989	446632	875952	1293157	-1	0.0000	0.2271	0.0000	0.0793
40 trichloroethene	9774	67466	274815	633192	1246934	1815606	-1	0.0000	0.3164	0.0000	0.0859
41 dibromomethane	4889	32002	126997	270225	530196	793122	-1	0.0000	0.1441	0.0000	0.1368
101 TAME	9699	90788	421379	938326	1872529	2927534	-1	0.0000	0.4424	0.0000	0.1161
42 Br-di-Cl-methane	11934	72312	281223	604352	1181950	1745648	-1	0.0014	0.2967	0.0000	0.9998
43 Me-methacrylate	1794	19729	100519	228095	464768	714088	-1	-0.0114	0.1219	0.0000	0.9990
44 2-ClEt-Vi-ether10	2834	31796	189485	514977	1130652	-1	-1	-0.0446	0.0290	0.0000	0.9927
45 c-13-di-Cl-propene	8596	71992	308098	676158	1328342	1977341	-1	0.0000	0.3301	0.0000	0.0559
46 t-1,3-dichloropropene	6273	54825	247619	547714	1100371	1653172	-1	0.0000	0.2625	0.0000	0.0800
47 Chlorobezene-d5	683445	775455	774184	772935	739886	736062	-1	0.0000	1.0000	0.0000	0.0000

Compound	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6	Level 7	Coeff X ⁰	Coeff X ¹ / ave RF	Coeff X ²	R ² / RSD
	Response	Response	Response	Response	Response	Response	Response				
48 112-tri-Cl-Et	5902	42713	164770	343401	662895	997315	-1	0.0000	0.2413	0.0000	0.1317
49 13-di-Cl-propane	8941	66977	268452	552557	1062190	1558760	-1	0.0000	0.3807	0.0000	0.1091
50 Et methacrylate	3865	41714	210769	443099	890456	1418605	-1	-0.0322	0.3197	0.0000	0.9978
Compound	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6	Level 7	Coeff	Coeff	Coeff	R² / RSD
Name	Response	Response	Response	Response	Response	Response	Response	X⁰	X¹ / ave RF	X²	RSD
51 di-Br-Cl-methane	8661	52843	213283	456818	899440	1365305	-1	-0.0111	0.3087	0.0000	0.9994
52 bromoform	4560	29809	123106	263988	534402	823804	-1	0.0000	0.1852	0.0000	0.1170
53 1,4-dichlorobutane-2	8296	61397	241554	510059	1005306	1519089	-1	0.0000	0.3544	0.0000	0.1053
54 MIBK	2701	22334	103475	222364	457366	733371	-1	0.0000	0.1456	0.0000	0.0889
55 toluene-d8	27570	211476	874470	1923324	3719816	5400714	-1	0.0000	1.2603	0.0000	0.0678
56 toluene	29681	252340	1018191	2227432	4284454	6235319	-1	0.0000	1.4484	0.0000	0.0698
57 2-hexanone X5	10224	84522	356146	728526	1453530	2205707	-1	0.0000	0.0989	0.0000	0.0595
58 12-dibromoethane	5621	39459	162435	343609	688415	1038913	-1	0.0000	0.2381	0.0000	0.0968
59 tetra-Cl-ethene	10167	72027	292262	659769	1256084	1800306	-1	0.0000	0.4328	0.0000	0.0968
60 chlorobenzene	26600	179300	691236	1494861	2837497	4059222	-1	0.0338	0.9255	0.0000	0.9991
61 1112-tetra-Cl-Et	9159	63898	245258	525072	1031299	1537036	-1	0.0000	0.3686	0.0000	0.1347
62 1,4-Dichlorobenzene-d4	358277	419640	407774	393961	367099	365124	-1	0.0000	1.0000	0.0000	0.0000
63 1-chlorohexane	2893	25946	118729	280119	548810	784184	-1	0.0000	0.3261	0.0000	0.1294
64 Et-Bz	33015	273886	1133719	2529977	4833949	6913463	-1	0.0000	3.1290	0.0000	0.0601
65 m/p-Xylenes X2	49281	436018	1773476	3842318	7112016	9936549	-1	0.0000	2.3654	0.0000	0.0637
66 styrene	15818	166023	695954	1498467	2811895	3981902	-1	0.0000	1.7985	0.0000	0.1032
67 o-xylene	21010	210456	917264	1990843	3763083	5341093	-1	0.0000	2.3732	0.0000	0.0983
68 1122-Tetra-Cl-Et	7086	48249	186045	374484	736112	1101133	-1	0.0000	0.5283	0.0000	0.1434
69 123-tri-Cl-Pr	1994	14801	56224	114784	222436	332760	-1	0.0000	0.1581	0.0000	0.1176
70 4-Br-1-F-Bz (S3)	10243	73535	287803	629095	1208574	1766234	-1	0.0000	0.8271	0.0000	0.1001
71 isopropylbenzene	24316	252585	1164869	2648278	5083237	7208570	-1	0.0000	3.0403	0.0000	0.1458
72 bromobenzene	9030	73980	293172	635461	1236729	1774887	-1	0.0000	0.8166	0.0000	0.0673
92 t-1,4-dichloro-2-butene	406	6524	30509	64587	137632	207953	-1	-0.0118	0.0961	0.0000	0.9978
73 n-propylbenzene	8157	83853	353478	802663	1558540	2210026	-1	0.0000	0.9523	0.0000	0.1210
74 2-Cl-Toluene	8123	74427	309252	685694	1328259	1912837	-1	0.0000	0.8415	0.0000	0.0790
75 4-Cl-Toluene	9172	80121	308786	677984	1291985	1819259	-1	0.0000	0.8560	0.0000	0.0752
76 135-tri-Me-Benzene	21830	240336	1005234	2217776	4186569	5867539	-1	0.0000	2.6173	0.0000	0.1238
77 4-iso-Pr-toluene	24508	264794	1111631	2489320	4727712	6567533	-1	0.0000	2.9230	0.0000	0.1239
78 124-tri-Me-Benzene	24365	252796	1026195	2248874	4305294	6074118	-1	0.0000	2.7257	0.0000	0.1034
79 tert-butylbenzene	21932	200692	1030129	2077031	4034706	5816456	-1	0.0000	2.4995	0.0000	0.1024

C:\RP\T\cal-ivoc.xls

80 13-DCB	21374	158442	605258	1313556	2467081	3488098	-1	0.0000	1.7167	0.0000	0.1095
81 sec-butylbenzene	29181	304851	1334960	3039109	5779243	8084394	-1	0.0000	3.5174	0.0000	0.1295
82 14-DCB	24079	155998	594167	1295125	2472135	3533637	-1	-0.0003	1.6325	0.0000	0.9988
83 Cl-benzyl	1764	13306	63403	132793	265795	361085	-1	0.0000	0.1654	0.0000	0.0541
84 12-DCB	18942	145496	540526	1153351	2171646	3083911	-1	0.0000	1.5286	0.0000	0.1167
85 n-butylbenzene	5679	66251	298215	689097	1335543	1846727	-1	-0.0178	0.8660	0.0000	0.9968
86 12-diBr-2-Cl-Pra	1180	8757	37194	80335	173925	263112	-1	0.0000	0.1076	0.0000	0.1009
87 124-tri-Cl-Bz	9061	77700	360446	838928	1658990	2390948	-1	0.0000	0.9898	0.0000	0.1216
88 naphthalene	23839	152715	581293	1300980	2689379	4011635	-1	0.0000	1.7961	0.0000	0.1450
89 hx-Cl-butadiene	6922	52135	209577	472099	923440	1288580	-1	0.0000	0.5992	0.0000	0.0774
90 123-Tri-Cl-Bz	9124	73440	317007	719921	1423460	2068944	-1	0.0000	0.8881	0.0000	0.0786

Quantitation Report: **Applied P &Ch Lab** EPA 524.2

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00003.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 19:25 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:52 2003
 Print Time : Thu Jul 24 11:52 2003
 Miscellaneous :

Sample : F=1 0.3 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	919.393	10.00	0.5	2.38%	189
47	Chlorobenzene-d5	11.55	11.54	0.000	117	82	683.445	10.00	0.4	2.23%	99
62	1,4-Dichlorobenze	13.84	13.84	0.000	152	150	358.277	10.00	0.3	1.58%	1

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.61	6.58	0.002	111	113	10.081	0.48	0.5	2.38%	
29	1,2-Di-Cl-Et-d4 (7.11	7.08	0.002	65	102	8.754	0.45	0.4	2.23%	
55	toluene-d8	9.91	9.89	0.000	98	100	27.570	0.32	0.3	1.58%	
70	4-Br-1-F-Bz (S3)	12.75	12.74	0.000	174	95	10.243	0.38	0.4	1.88%	

Target Compounds												
<<< 11	: ISTD ID = 1	>>>										Qvalue
3	di-Cl-di-F-methan	1.89	1.85	0.005	85	87	7.698	0.34	0.3	0.3	97	<i>De</i>
4	Chloromethane	2.11	2.07	0.006	50	52	8.264	0.34	0.3	0.3	97	<i>De</i>
2	F114	2.04	2.00	0.006	85	135	3.209	0.27	0.3	0.3	96	<i>De</i>
5	vinyl chloride	2.22	2.19	0.004	62	64	7.381	0.32	0.3	0.3	89	<i>De</i>
6	bromomethane	2.61	2.58	0.005	94	96	3.952	0.30	0.3	0.3	95	<i>De</i>
7	chloroethane	2.73	2.70	0.004	64	66	4.460	0.30	0.3	0.3	0	<i>De</i>
8	tri-Cl-F-methane	3.03	3.00	0.005	101	103	11.754	0.34	0.3	0.3	99	<i>De</i>
91	Acetonitrile X10	4.07	4.04	0.004	41	40	24.302	8.33	8.3	8.3	73	<i>De</i>
9	acrolein X10	3.51	3.48	0.005	56	55	9.511	4.28	4.3	4.3	0	<i>De</i>
11	acetone X10	3.72	3.69	0.004	43	58	29.124	12.37	12.4	12.4	94	<i>De</i>
12	ethyl ether X5	3.37	3.34	0.005	59	74	22.348	1.91	1.9	1.9	92	<i>De</i>
13	11-dichloroethene	3.64	3.60	0.005	61	96	9.522	0.34	0.3	0.3	0	<i>De</i>
14	Iodomethane	3.82	3.78	0.006	142	127	7.324	0.47	0.5	0.5	84	<i>De</i>
15	F-113	3.65	3.62	0.005	101	151	8.636	0.42	0.4	0.4	0	<i>De</i>
16	acrylonitrile X10	4.53	4.49	0.005	53	52	14.700	4.72	4.7	4.7	89	<i>De</i>
17	carbon disulfide	3.90	3.87	0.004	76	78	23.125	0.28	0.3	0.3	98	<i>De</i>
94	Isopropyl Alcohol	4.08	4.01	0.010	45	43	4.021	14.57	14.6	14.6	100	<i>De</i>
18	methylene chlorid	4.22	4.19	0.004	84	49	36.501	1.52	1.5	1.5	94	<i>De</i>
19	t-12-di-Cl-ethene	4.58	4.55	0.004	96	61	8.180	0.35	0.3	0.3	83	<i>De</i>

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Quantitation Report: **Applied P

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00003.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 19:25 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:52 2003
 Print Time : Thu Jul 24 11:52 2003
 Miscellaneous :

Sample : F=1 0.3 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.60	4.56	0.005	73	57	13.066	0.31	0.3	93	?
95	Tert butyl alcoho	4.47	4.47	0.000	59	57	6.515	8.83	8.8	100	m?
94	allyl chloride	4.07	4.04	0.004	41	76	24.302	0.71	0.7	62	#?
21	11-dichloroethane	5.12	5.09	0.005	63	83	14.292	0.37	0.4	95	m?
22	c-12-di-Cl-ethene	5.92	5.89	0.004	96	61	8.154	0.35	0.3	96	m?
23	22-Dichloropropan	5.92	5.89	0.003	77	97	10.809	0.38	0.4	95	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	4.359	0.47	0.5	87	?
25	chloroform	6.37	6.35	0.002	83	85	19.972	0.51	0.5	100	?
26	tetrahydrofuranX5	6.36	6.32	0.005	42	72	4.965	1.89	1.9	89	?
98	Disopropyl ether	5.24	5.22	0.003	45	87	17.882	0.28	0.3	69	#
99	ETBE	5.75	5.72	0.004	59	87	12.061	0.25	0.2	97	#
30	12-dichloroethane	7.23	7.20	0.004	64	62	2.582	0.36	0.4	83	m?
32	vinyl acetate X5	5.20	5.17	0.005	43	86	43.949	1.55	1.5	99	m?
92	Nitro Methane(X10	5.81	5.80	0.002	61	46	1.737	0.82	0.8	47	#
33	2-butanoneMEK X10	5.95	5.92	0.004	43	72	22.369	7.70	7.7	96	#
93	Ethyl Acetate x2	6.04	6.02	0.003	43	61	9.671	0.81	0.8	78	#
34	111-trichloroetha	6.65	6.63	0.002	97	99	13.204	0.38	0.4	99	?
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	7.463	0.25	0.2	91	?
36	benzene	7.21	7.19	0.003	78	52	30.808	0.33	0.3	96	?
37	CCl4	6.91	6.89	0.002	117	119	11.983	0.40	0.4	92	?
100	Isobutyl alcohol	7.19	7.39	-0.027	43	42	0.457	2.50	2.5	89	m?
38	thiophene	7.53	7.51	0.003	84	58	12.985	0.28	0.3	94	m?
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	6.808	0.34	0.3	94	m?
40	trichloroethene	8.24	8.23	0.002	130	132	9.774	0.39	0.4	96	?
41	dibromomethane	8.73	8.71	0.002	174	172	4.889	0.51	0.5	87	?
101	TAME	7.42	7.39	0.003	73	43	9.699	0.22	0.2	83	#
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	11.934	0.44	0.4	100	#
43	Me-methacrylate	8.76	8.75	0.002	69	100	1.794	0.19	0.2	99	#
44	2-ClEt-Vi-ether10	9.38	9.37	0.002	63	43	2.834	1.36	1.4	84	#
45	c-13-di-Cl-propen	9.57	9.55	0.002	75	110	8.596	0.29	0.3	97	#
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	6.273	0.28	0.3	90	#

Quantitation Report: **Applied P &Ch Lab** EPA 524.2

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00003.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 19:25 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:52 2003
 Print Time : Thu Jul 24 11:52 2003
 Miscellaneous :

Sample : f=1 0.3 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	8.941	0.36	0.4	90	?
50	Et methacrylate	10.38	10.37	0.001	69	99	3.865	0.19	0.2	92	
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	8.661	0.50	0.5	99	
52	bromoform	12.42	12.41	0.000	173	174	4.560	0.49	0.5	95	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	8.296	0.37	0.4	91	
54	MIBK	9.77	9.76	0.000	43	58	2.701	0.28	0.3	85	
56	toluene	9.99	9.98	0.000	91	92	29.681	0.29	0.3	96	
57	2-hexanone X5	10.76	10.75	0.001	43	58	10.224	1.59	1.6	94	
58	12-dibromoethane	11.03	11.03	0.000	107	109	5.621	0.42	0.4	90	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	10.167	0.39	0.4	95	?
60	chlorobenzene	11.58	11.57	0.000	112	77	26.600	0.40	0.4	90	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	9.159	0.43	0.4	95	
<<< I3 : ISTD ID = 62 >>>											
63	1-chlorohexane	11.56	11.56	0.000	93	55	2.893	0.20	0.2	54	#?
64	Et-Bz	11.70	11.69	0.000	91	106	33.015	0.26	0.3	99	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	49.281	0.51	0.5	99	
66	styrene	12.24	12.23	0.000	104	78	15.818	0.21	0.2	80	?
67	O-xylene	12.22	12.22	0.000	91	106	21.010	0.21	0.2	97	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	7.086	0.41	0.4	98	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	1.994	0.39	0.4	100	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	24.316	0.20	0.2	97	?
72	bromobenzene	12.89	12.89	0.000	156	158	9.030	0.33	0.3	97	?
92	t-1,4-dichloro-2-	12.93	12.92	0.000	89	53	0.406	0.16	0.2	19	#?
73	n-propylbenzene	13.00	12.99	0.000	120	78	8.157	0.23	0.2	82	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	8.123	0.27	0.3	94	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	9.172	0.29	0.3	88	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	21.830	0.21	0.2	96	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	24.508	0.21	0.2	97	?
78	124-tri-Me-Benzen	13.52	13.51	0.000	105	120	24.365	0.23	0.2	98	
79	tert-butylbenzene	13.48	13.47	0.000	119	91	21.932	0.23	0.2	94	
80	13-DCB	13.79	13.78	0.000	146	148	21.374	0.37	0.4	97	
81	sec-butylbenzene	13.68	13.68	0.000	105	134	29.181	0.21	0.2	98	

ms
2003/7/24/63

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00003.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 19:25 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:52 2003
 Print Time : Thu Jul 24 11:52 2003
 Miscellaneous :


Sample : f=1 0.3 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	CO,ppb	C,ppb	Quality	Note
82	14-DCB	13.87	13.87	0.000	146	148	24.079	0.41	0.4	92	
83	Cl-benzyl	13.98	13.98	0.000	126	91	1.764	0.41	0.4	82	#
84	12-DCB	14.21	14.21	0.000	146	148	18.942	0.38	0.4	95	
85	n-butylbenzene	14.18	14.18	0.000	134	91	5.679	0.19	0.2	87	#
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	1.180	0.41	0.4	81	#
87	124-tri-Cl-Bz	15.59	15.58	0.000	180	182	9.061	0.25	0.2	96	
88	naphthalene	15.79	15.78	0.000	128	129	23.839	0.45	0.4	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	6.922	0.32	0.3	88	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	9.124	0.30	0.3	91	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00003.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 19:25 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:52 2003
 Print Time : Thu Jul 24 11:52 2003
 Miscellaneous :

Sample : f=1 0.3 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000



Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-0002A.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:11 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:56 2003
 Print Time : Thu Jul 24 11:56 2003
 Miscellaneous :

Sample : F=1 2 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	1006.828	10.00		0.02	
47	Chlorobenzene-d5	11.54	11.54	0.000	117	82	775.455	10.00		0.00	
62	1,4-Dichlorobenze	13.85	13.84	0.000	152	150	419.640	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	59.230	2.56		2.6	12.78%
29	1,2-Di-Cl-Et-d4	7.11	7.08	0.002	65	102	50.036	2.33		2.3	11.64%
55	toluene-d8	9.91	9.89	0.000	98	100	211.476	2.14		2.1	10.71%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	73.535	2.30		2.3	11.51%

Target Compounds <<< I1 : ISTD ID = 1 >>> Qvalue

3	di-Cl-di-F-methan	1.89	1.85	0.005	85	87	51.044	2.04		2.0	96
4	Chloromethane	2.11	2.07	0.006	50	52	53.373	2.02		2.0	98
2	F114	2.03	2.00	0.005	85	135	27.835	2.13		2.1	54
5	vinyl chloride	2.23	2.19	0.005	62	64	53.695	2.14		2.1	98
6	bromomethane	2.61	2.58	0.005	94	96	25.743	1.77		1.8	100
7	chloroethane	2.73	2.70	0.004	64	66	34.347	2.11		2.1	96
8	tri-Cl-F-methane	3.03	3.00	0.005	101	103	73.985	1.97		2.0	100
91	Acetonitrile X10	4.07	4.04	0.004	41	40	110.646	34.62		34.6	84
9	acrolein X10	3.52	3.48	0.006	56	55	67.561	27.75		27.8	99
11	acetone X10	3.73	3.69	0.005	43	58	106.138	41.18		41.2	0
12	ethyl ether X5	3.37	3.34	0.005	59	74	158.277	12.35		12.4	89
13	11-dichloroethene	3.64	3.60	0.005	61	96	62.800	2.04		2.0	94
14	Iodomethane	3.82	3.78	0.005	142	127	57.536	3.40		3.4	93
15	F-113	3.65	3.62	0.005	101	151	49.466	2.20		2.2	90
16	acrylonitrile X10	4.53	4.49	0.005	53	52	104.060	30.48		30.5	95
17	carbon disulfide	3.90	3.87	0.004	76	78	156.993	1.77		1.8	99
94	Isopropyl Alcohol	4.10	4.01	0.011	45	43	15.273	50.54		50.5	100
18	methylene chlorid	4.22	4.19	0.005	84	49	97.000	3.70		3.7	94
19	t-12-di-Cl-ethene	4.58	4.55	0.005	96	61	61.414	2.39		2.4	90

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

De
07/24/03
m
De
07/24/03
m
De
07/24/03
m
De
07/24/03
m

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-0002A.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:11 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:56 2003
 Print Time : Thu Jul 24 11:56 2003
 Miscellaneous :

Sample : F=1 2 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.60	4.56	0.005	73	57	102.540	2.19	2.2	97	?
95	Tert butyl alcoho	4.50	4.47	0.005	59	57	28.855	35.70	35.7	100	me
94	allyl chloride	4.07	4.04	0.004	41	76	110.646	2.96	3.0	77	#?
21	11-dichloroethane	5.12	5.09	0.004	63	83	97.373	2.28	2.3	95	me
97	propionitrile	6.03	5.99	0.004	54	51	4.103	2.96	3.0	100	me
22	c-12-di-Cl-ethene	5.92	5.89	0.004	96	61	62.676	2.45	2.5	90	?
23	22-Dichloropropan	5.92	5.89	0.004	77	97	72.464	2.34	2.3	99	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	30.212	2.99	3.0	97	?
25	chloroform	6.38	6.35	0.003	83	85	106.303	2.47	2.5	99	?
26	tetrahydrofuranX5	6.36	6.32	0.005	42	72	35.404	12.31	12.3	89	?
98	Disopropyl ether	5.25	5.22	0.004	45	87	159.533	2.30	2.3	97	me
99	ETBE	5.75	5.72	0.004	59	87	107.588	2.02	2.0	95	#?
30	12-dichloroethane	7.22	7.20	0.002	64	62	18.795	2.38	2.4	89	#?
32	vinyl acetate X5	5.20	5.17	0.004	43	86	379.164	12.18	12.2	98	me
92	Nitro Methane(X10	5.84	5.80	0.005	61	46	12.060	5.18	5.2	2	me
33	2-butanoneMEK X10	5.95	5.92	0.003	43	72	123.309	38.78	38.8	95	#?
93	Ethyl Acetate x2	6.05	6.02	0.004	43	61	54.370	4.14	4.1	88	#?
34	111-trichloroetha	6.65	6.63	0.002	97	99	87.716	2.28	2.3	98	?
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	61.583	1.86	1.9	88	?
36	benzene	7.22	7.19	0.004	78	52	233.485	2.31	2.3	100	?
37	CCl4	6.91	6.89	0.003	117	119	79.844	2.46	2.5	100	?
100	Isobutyl alcohol	7.19	7.39	-0.027	43	42	10.371	51.83	51.8	91	me
38	thiophene	7.53	7.51	0.002	84	58	115.327	2.25	2.3	95	me
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	50.516	2.33	2.3	97	me
40	trichloroethene	8.24	8.23	0.002	130	132	67.466	2.44	2.4	95	me
41	di bromomethane	8.73	8.71	0.002	174	172	32.002	3.05	3.0	100	me
101	TAME	7.42	7.39	0.003	73	43	90.788	1.90	1.9	95	me
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	72.312	2.42	2.4	97	me
43	Me-methacrylate	8.76	8.75	0.002	69	100	19.729	1.93	1.9	90	me
44	2-ClEt-VI-ether10	9.38	9.37	0.002	63	43	31.796	13.91	13.9	97	me
45	c-13-di-Cl-propen	9.57	9.55	0.002	75	110	71.992	2.24	2.2	95	me
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	54.825	2.21	2.2	90	me

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Quantitation Report: **Applied P & Ch Lab** EPA 524.2

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-0002A.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:11 2003
 Method Update: Mon Jul 21 15:01 2003
 Quart. Time : Jul 24 11:56 2003
 Print Time : Thu Jul 24 11:56 2003
 Miscellaneous :

Sample : f=1 2 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<< I2	: ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	42.713	2.65	2.6	97	
49	13-di-Cl-propane	10.65	10.64	0.001	76	78	66.977	2.35	2.3	96	
50	Et methacrylate	10.38	10.37	0.001	69	99	41.714	1.85	1.8	95	?
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	52.843	2.67	2.7	100	
52	bromofom	12.42	12.41	0.000	173	174	29.809	2.82	2.8	100	
53	1,4-dichlorobutan	12.67	12.67	0.000	55	41	61.397	2.38	2.4	96	
54	MIBK	9.77	9.76	0.001	43	58	22.334	2.04	2.0	94	
56	toluene	9.99	9.98	0.001	91	92	252.340	2.17	2.2	99	
57	2-hexanone X5	10.76	10.75	0.001	43	58	84.522	11.62	11.6	94	
58	12-dibromoethane	11.03	11.03	0.000	107	109	39.459	2.61	2.6	94	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	72.027	2.41	2.4	99	
60	chlorobenzene	11.58	11.57	0.000	112	77	179.300	2.40	2.4	93	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	63.898	2.67	2.7	98	
<<< I3	: ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	25.946	1.54	1.5	91	?
64	Et-Bz	11.70	11.69	0.000	91	106	273.886	1.85	1.9	96	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	436.018	3.85	3.8	96	
66	styrene	12.24	12.23	0.000	104	78	166.023	1.92	1.9	92	?
67	o-xylene	12.22	12.22	0.000	91	106	210.456	1.81	1.8	99	
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	48.249	2.37	2.4	93	
69	123-tri-Cl-Pr	12.92	12.91	0.000	110	97	14.801	2.50	2.5	96	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	252.585	1.76	1.8	99	
72	bromobenzene	12.89	12.89	0.000	156	158	73.980	2.31	2.3	99	
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	6.524	2.14	2.1	92	?
73	n-propylbenzene	13.00	12.99	0.000	120	78	83.853	1.98	2.0	94	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	74.427	2.08	2.1	100	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	80.121	2.19	2.2	97	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	240.336	1.95	1.9	98	?
77	4-Iso-Pr-toluene	13.81	13.81	0.000	119	134	264.794	1.97	2.0	99	?
78	124-tri-Me-Benzen	13.52	13.51	0.000	105	120	252.796	1.99	2.0	96	
79	tert-butylbenzene	13.48	13.47	0.000	119	91	200.692	1.81	1.8	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Quantitation Report: **Applied P &Ch Lab** EPA 524.2

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-0002A.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:11 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:56 2003
 Print Time : Thu Jul 24 11:56 2003
 Miscellaneous :


Sample : f=1 2 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	158.442	2.31	2.3	99	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	304.851	1.87	1.9	100	
82	14-DCB	13.87	13.87	0.000	146	148	155.998	2.27	2.3	97	
83	Cl-benzyl	13.98	13.98	0.000	126	91	13.306	2.65	2.6	81	#
84	12-DCB	14.21	14.21	0.000	146	148	145.496	2.46	2.5	100	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	66.251	1.89	1.9	94	?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	8.757	2.61	2.6	89	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	77.700	1.82	1.8	97	
88	naphthalene	15.79	15.78	0.000	128	129	152.715	2.44	2.4	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	52.135	2.05	2.0	100	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	73.440	2.07	2.1	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-0002A.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:11 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:56 2003
 Print Time : Thu Jul 24 11:56 2003
 Miscellaneous :

Sample : F=1 2 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000



Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00010.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:37 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:58 2003
 Print Time : Thu Jul 24 11:58 2003
 Miscellaneous :

Sample : f=1 10 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

169

ID	Component Name	R.T.	RT0	DRRT	Qion	Ql	RF/1000	C0,ppb	C,ppb	Quality	Note
----	----------------	------	-----	------	------	----	---------	--------	-------	---------	------

Internal Standards

1	Fluorobenzene	7.65	7.63	0.003	96	70	1003.977	10.00		0.02	
47	Chlorobenzene-d5	11.55	11.54	0.000	117	82	774.184	10.00		0.00	
62	1,4-Dichlorobenze	13.85	13.84	0.000	152	150	407.774	10.00		0.00	

System Monitoring Compounds (Surrogate)

27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	230.740	9.99	10.0	49.93%	
29	1,2-Di-Cl-Et-d4	7.11	7.08	0.002	65	102	187.219	8.73	8.7	43.66%	
55	toluene-d8	9.91	9.89	0.000	98	100	874.470	8.87	8.9	44.35%	
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	287.803	9.27	9.3	46.37%	

Target Compounds

<<< ID	: ISTD ID = 1	>>>	1.89	1.85	0.005	85	87	228.452	9.14	9.1	98	
3	di-Cl-di-F-methan		2.11	2.07	0.006	50	52	208.536	7.92	7.9	98	
4	Chloromethane		2.03	2.00	0.005	85	135	126.949	9.74	9.7	59	
2	F114		2.22	2.19	0.004	62	64	219.399	8.78	8.8	100	
5	vinyl chloride		2.61	2.58	0.005	94	96	105.105	7.25	7.2	97	
6	bromomethane		2.73	2.70	0.004	64	66	136.103	8.39	8.4	100	
7	chloroethane		3.03	3.00	0.005	101	103	311.036	8.29	8.3	100	
8	tri-Cl-F-methane		4.07	4.04	0.004	41	40	413.272	129.67	129.7	89	
91	Acetonitrile X10		3.51	3.48	0.005	56	55	262.941	108.33	108.3	0	
9	acrolein X10		3.73	3.69	0.005	43	58	281.400	109.48	109.5	0	
11	acetone X10		3.37	3.34	0.005	59	74	587.938	46.01	46.0	88	
12	ethyl ether X5		3.64	3.60	0.005	61	96	259.067	8.43	8.4	98	
13	11-dichloroethene		3.82	3.78	0.005	142	127	244.664	14.51	14.5	93	
14	Iodomethane		3.65	3.62	0.005	101	151	193.654	8.65	8.6	87	
15	F-113		4.53	4.49	0.005	53	52	420.553	123.54	123.5	97	
16	acrylonitrile X10		3.91	3.87	0.005	76	78	656.825	7.41	7.4	100	
17	carbon disulfide		4.11	4.01	0.014	45	43	46.223	153.39	153.4	100	
94	Isopropyl Alcohol		4.22	4.19	0.005	84	49	257.935	9.86	9.9	100	
18	methylene chlorid		4.58	4.55	0.005	96	61	249.312	9.75	9.7	93	
19	t-12-di-Cl-ethene											

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

m *[Signature]*
 m *[Signature]*
 m *[Signature]*
 m *[Signature]*

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00010.D
 Method : C:\MSDCHEM\1\METHODS\E5224A002.M
 Acq. Time : Jul 22 21:37 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:58 2003
 Print Time : Thu Jul 24 11:58 2003
 Miscellaneous :

Sample : f=1 10 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	QIon	QI	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.60	4.56	0.005	73	57	420.692	9.03	9.0	97	?
95	Tert butyl alcoho	4.58	4.47	0.014	59	57	98.345	122.02	122.0	100	m?
94	allyl chloride	4.07	4.04	0.004	41	76	413.272	11.07	11.1	81	#?
21	11-dichloroethane	5.12	5.09	0.004	63	83	381.929	8.98	9.0	99	
97	propionitrile	6.03	5.99	0.004	54	51	13.637	9.86	9.9	100	#?
22	c-12-di-Cl-ethene	5.92	5.89	0.003	96	61	253.306	9.94	9.9	88	?
23	22-Dichloropropan	5.92	5.89	0.004	77	97	281.116	9.09	9.1	95	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	117.713	11.68	11.7	96	?
25	chloroform	6.37	6.35	0.002	83	85	398.005	9.29	9.3	98	?
26	tetrahydrofuranX5	6.35	6.32	0.003	42	72	154.519	53.88	53.9	88	?
98	Disopropyl ether	5.24	5.22	0.003	45	87	636.576	9.21	9.2	96	?
99	ETBE	5.74	5.72	0.003	59	87	486.462	9.16	9.2	95	?
30	12-dichloroethane	7.22	7.20	0.003	64	62	70.404	8.93	8.9	96	?
32	vinyl acetate X5	5.20	5.17	0.004	43	86	1608.933	51.83	51.8	99	?
92	Nitro Methane(X10	5.83	5.80	0.004	61	46	47.784	20.59	20.6	88	?
33	2-butanoneMEK X10	5.95	5.92	0.003	43	72	444.332	140.12	140.1	95	?
93	Ethyl Acetate x2	6.04	6.02	0.003	43	61	220.002	16.80	16.8	92	#?
34	111-trichloroetha	6.65	6.63	0.002	97	99	354.801	9.25	9.2	100	?
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	285.223	8.66	8.7	88	?
36	benzene	7.21	7.19	0.003	78	52	926.403	9.17	9.2	99	?
37	CCl4	6.91	6.89	0.003	117	119	329.846	10.19	10.2	99	?
100	Isobutyl alcohol	7.19	7.39	-0.027	43	42	40.745	204.19	204.2	95	m?
38	thiophene	7.53	7.51	0.003	84	58	473.132	9.26	9.3	98	
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	204.989	9.47	9.5	98	
40	trichloroethene	8.24	8.23	0.002	130	132	274.815	9.97	10.0	99	
41	dibromomethane	8.73	8.71	0.002	174	172	126.997	12.13	12.1	99	
101	TAME	7.41	7.39	0.002	73	43	421.379	8.86	8.9	98	
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	281.223	9.42	9.4	99	
43	Me-methacrylate	8.76	8.75	0.002	69	100	100.519	9.84	9.8	97	
44	2-ClEt-Vi-ether10	9.38	9.37	0.002	63	43	189.485	83.14	83.1	98	
45	c-13-di-Cl-propen	9.56	9.55	0.002	75	110	308.098	9.63	9.6	94	
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	247.619	10.02	10.0	93	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Handwritten signature

Handwritten signature

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00010.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:37 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:58 2003
 Print Time : Thu Jul 24 11:58 2003
 Miscellaneous :

Sample : f=1 10 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Ql	RF/1000	C0,ppb	C,ppb	Quality	Note
<<< 12	: ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	164.770	10.23	10.2	98	
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	268.452	9.42	9.4	100	?
50	Et methacrylate	10.37	10.37	0.000	69	99	210.769	9.34	9.3	94	
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	213.283	10.78	10.8	100	
52	bromofom	12.42	12.41	0.000	173	174	123.106	11.66	11.7	100	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	241.554	9.38	9.4	95	
54	MIBK	9.77	9.76	0.000	43	58	103.475	9.46	9.5	99	
56	toluene	9.99	9.98	0.000	91	92	1018.191	8.79	8.8	100	
57	2-hexanone X5	10.76	10.75	0.000	43	58	356.146	49.05	49.0	95	
58	12-dibromoethane	11.03	11.03	0.000	107	109	162.435	10.76	10.8	97	
59	tetra-Cl-ethene	10.65	10.64	0.001	166	168	292.262	9.80	9.8	97	?
60	chlorobenzene	11.58	11.57	0.000	112	77	691.236	9.26	9.3	91	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	245.258	10.28	10.3	100	
<<< 13	: ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	118.729	7.23	7.2	96	?
64	Et-Bz	11.70	11.69	0.000	91	106	1133.719	7.90	7.9	95	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	1773.476	16.10	16.1	96	
66	styrene	12.24	12.23	0.000	104	78	695.954	8.27	8.3	91	?
67	o-xylene	12.22	12.22	0.000	91	106	917.264	8.12	8.1	97	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	186.045	9.39	9.4	99	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	56.224	9.77	9.8	98	?
71	Isopropylbenzene	12.60	12.59	0.000	105	120	1164.869	8.34	8.3	98	
72	bromobenzene	12.89	12.89	0.000	156	158	293.172	9.42	9.4	97	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	30.509	10.30	10.3	90	?
73	n-propylbenzene	13.00	12.99	0.000	120	78	353.478	8.59	8.6	94	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	309.252	8.90	8.9	98	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	308.786	8.67	8.7	96	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	1005.234	8.38	8.4	97	?
77	4-Iso-Pr-toluene	13.81	13.81	0.000	119	134	1111.631	8.52	8.5	99	?
78	124-tri-Me-Benzen	13.51	13.51	0.000	105	120	1026.195	8.33	8.3	98	?
79	tert-butylbenzene	13.47	13.47	0.000	119	91	1030.129	9.57	9.6	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Quantitation Report: **Applied P & Ch Lab** EPA 524.2

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00010.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:37 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:58 2003
 Print Time : Thu Jul 24 11:58 2003
 Miscellaneous :


Sample : f=1 10 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	CO,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	605.258	9.08	9.1	99	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	1334.960	8.44	8.4	97	
82	14-DCB	13.87	13.87	0.000	146	148	594.167	8.88	8.9	98	
83	Cl-benzyl	13.98	13.98	0.000	126	91	63.403	12.97	13.0	76	#
84	12-DCB	14.21	14.21	0.000	146	148	540.526	9.41	9.4	100	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	298.215	8.75	8.8	90	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	37.194	11.39	11.4	99	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	360.446	8.69	8.7	98	
88	naphthalene	15.79	15.78	0.000	128	129	581.293	9.57	9.6	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	209.577	8.47	8.5	96	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	317.007	9.19	9.2	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00010.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:37 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:58 2003
 Print Time : Thu Jul 24 11:58 2003
 Miscellaneous :

Sample : F=1 10 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000



Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00020.D Sample : f=1 20 ppb
Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
Acq. Time : Jul 22 22:56 2003 RF via : Multiple Level Calibration
Method Update: Mon Jul 21 15:01 2003 Operator: zou
Quant. Time : Jul 24 12:01 2003 Multiplr: 1.000000
Print Time : Thu Jul 24 12:01 2003
Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	QION	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.64	7.63	0.002	96	70	1015.499	10.00		0.02	
47	Chlorobenzene-d5	11.54	11.54	0.000	117	82	772.935	10.00		0.00	
62	1,4-Dichlorobenze	13.84	13.84	0.000	152	150	393.961	10.00		0.00	

System Monitoring Compounds (Surrogate)											
ID	Component Name	R.T.	RT0	DRRT	QION	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	486.919	20.83	20.8	104.16%	
29	1,2-Di-Cl-Et-d4	7.10	7.08	0.001	65	102	388.784	17.93	17.9	89.64%	
55	toluene-d8	9.91	9.89	0.000	98	100	1923.324	19.54	19.5	97.70%	
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	629.095	20.98	21.0	104.91%	

Target Compounds <<< I1 : ISTD ID = 1 >>>

ID	Component Name	R.T.	RT0	DRRT	QION	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
3	di-Cl-di-F-methan	1.89	1.85	0.005	85	87	508.778	20.12	20.1	100	
4	Chloromethane	2.11	2.07	0.006	50	52	448.729	16.85	16.9	99	
2	F114	2.03	2.00	0.005	85	135	275.940	20.92	20.9	51	
5	vinyl chloride	2.23	2.19	0.005	62	64	484.855	19.18	19.2	98	
6	bromomethane	2.61	2.58	0.005	94	96	228.520	15.58	15.6	97	
7	chloroethane	2.73	2.70	0.004	64	66	266.250	16.22	16.2	0	
8	tri-Cl-F-methane	3.03	3.00	0.005	101	103	724.847	19.09	19.1	98	
91	Acetonitrile X10	4.07	4.04	0.004	41	40	845.923	262.40	262.4	90	
9	acrolein X10	3.51	3.48	0.004	56	55	513.017	208.95	209.0	0	
11	acetone X10	3.70	3.69	0.002	43	58	497.276	191.27	191.3	99	
12	ethyl ether X5	3.37	3.34	0.005	59	74	1189.867	92.06	92.1	90	
13	11-dichloroethene	3.64	3.60	0.005	61	96	562.011	18.08	18.1	0	
14	Iodomethane	3.82	3.78	0.005	142	127	532.902	31.25	31.3	95	
15	F-113	3.65	3.62	0.005	101	151	427.782	18.88	18.9	90	
16	acrylonitrile X10	4.52	4.49	0.003	53	52	831.510	241.48	241.5	98	
17	carbon disulfide	3.90	3.87	0.004	76	78	1480.814	16.51	16.5	99	
94	Isopropyl Alcohol	3.98	4.01	-0.004	45	43	124.874	409.68	409.7	100	
18	methylene chlorid	4.22	4.19	0.005	84	49	510.536	19.30	19.3	99	
19	t-12-di-Cl-ethene	4.58	4.55	0.004	96	61	531.623	20.55	20.5	95	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Quantitation Report: **Applied P

sch Lab**

EPA 524.2

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00020.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 22:56 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:01 2003
 Print Time : Thu Jul 24 12:01 2003
 Miscellaneous :

Sample : F=1 20 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Ql	RF/1000	C0, ppb	C, ppb	Quality	Note
20	t-Bu-Me-ether	4.59	4.56	0.004	73	57	866.584	18.38	18.4	94	?
95	Tert butyl alcoho	4.45	4.47	-0.002	59	57	205.914	252.58	252.6	100	De
94	allyl chloride	4.07	4.04	0.004	41	76	845.923	22.41	22.4	83	#? 07/24/03
21	11-dichloroethane	5.12	5.09	0.004	63	83	825.889	19.20	19.2	99	
97	propionitrile	6.01	5.99	0.002	54	51	33.328	23.83	23.8	100	#
22	c-12-di-Cl-ethene	5.92	5.89	0.003	96	61	533.706	20.71	20.7	91	?
23	22-Dichloropropan	5.92	5.89	0.004	77	97	589.735	18.85	18.9	95	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	250.220	24.55	24.5	100	
25	chloroform	6.38	6.35	0.003	83	85	838.697	19.34	19.3	99	
26	tetrahydrofuranX5	6.34	6.32	0.002	42	72	318.868	109.93	109.9	94	
98	Diisopropyl ether	5.24	5.22	0.003	45	87	1326.012	18.97	19.0	97	
99	ETBE	5.74	5.72	0.003	59	87	1064.815	19.83	19.8	96	
30	12-dichloroethane	7.22	7.20	0.003	64	62	142.937	17.93	17.9	99	?
32	vinyl acetate X5	5.19	5.17	0.003	43	86	2957.025	94.18	94.2	99	
92	Nitro Methane(X10	5.81	5.80	0.000	61	46	111.334	47.43	47.4	78	
33	2-butanoneMEK X10	5.93	5.92	0.002	43	72	875.377	272.92	272.9	94	?
93	Ethyl Acetate x2	6.04	6.02	0.002	43	61	434.881	32.82	32.8	91	#
34	111-trichloroetha	6.65	6.63	0.002	97	99	785.718	20.25	20.2	100	
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	660.985	19.83	19.8	92	?
36	benzene	7.21	7.19	0.003	78	52	1995.763	19.54	19.5	100	?
37	CCl4	6.91	6.89	0.002	117	119	732.441	22.36	22.4	99	?
100	Isobutyl alcohol	7.14	7.39	-0.034	43	42	87.337	432.72	432.7	93	De
38	thiophene	7.53	7.51	0.002	84	58	1025.220	19.84	19.8	99	m 07/24/03
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	446.632	20.39	20.4	93	
40	trichloroethene	8.24	8.23	0.002	130	132	633.192	22.70	22.7	100	
41	di bromomethane	8.73	8.71	0.002	174	172	270.225	25.52	25.5	96	
101	TAME	7.41	7.39	0.002	73	43	938.326	19.51	19.5	99	
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	604.352	20.02	20.0	100	
43	Me-methacrylate	8.76	8.75	0.002	69	100	228.095	22.09	22.1	93	
44	2-ClEt-VI-ether10	9.38	9.37	0.000	63	43	514.977	223.40	223.4	98	
45	c-13-di-Cl-propen	9.56	9.55	0.002	75	110	676.158	20.89	20.9	94	
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	547.714	21.91	21.9	93	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00020.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 22:56 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:01 2003
 Print Time : Thu Jul 24 12:01 2003
 Miscellaneous :

Sample : F=1 20 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<< I2 : ISTD ID = 47 >>>											
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	343.401	21.35	21.4	94	
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	552.557	19.42	19.4	100	
50	Et methacrylate	10.38	10.37	0.000	69	99	443.099	19.68	19.7	94	?
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	456.818	23.12	23.1	100	
52	bromoform	12.42	12.41	0.000	173	174	263.988	25.03	25.0	100	
53	1,4-dichlorobutan	12.67	12.67	0.000	55	41	510.059	19.84	19.8	94	
54	MIBK	9.77	9.76	0.000	43	58	222.364	20.36	20.4	92	
56	toluene	9.99	9.98	0.000	91	92	2227.432	19.25	19.3	99	
57	2-hexanone X5	10.76	10.75	0.000	43	58	728.526	100.50	100.5	94	
58	12-dibromoethane	11.03	11.03	0.000	107	109	343.609	22.79	22.8	96	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	659.769	22.17	22.2	99	?
60	chlorobenzene	11.58	11.57	0.000	112	77	1494.861	20.06	20.1	92	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	525.072	22.05	22.0	99	
<<< I3 : ISTD ID = 62 >>>											
63	1-chlorohexane	11.56	11.56	0.000	93	55	280.119	17.67	17.7	97	?
64	Et-Bz	11.70	11.69	0.000	91	106	2529.977	18.24	18.2	96	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	3842.318	36.10	36.1	97	
66	styrene	12.24	12.23	0.000	104	78	1498.467	18.43	18.4	92	?
67	o-xylene	12.22	12.22	0.000	91	106	1990.843	18.25	18.2	96	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	374.484	19.56	19.6	98	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	114.784	20.64	20.6	100	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	2648.278	19.63	19.6	98	
72	bromobenzene	12.89	12.89	0.000	156	158	635.461	21.14	21.1	99	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	64.587	22.57	22.6	85	?
73	n-propylbenzene	13.00	12.99	0.000	120	78	802.663	20.20	20.2	93	?
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	685.694	20.42	20.4	98	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	677.984	19.70	19.7	99	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	2217.776	19.13	19.1	97	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	2489.320	19.74	19.7	97	?
78	124-tri-Me-Benzen	13.51	13.51	0.000	105	120	2248.874	18.90	18.9	96	?
79	tert-butylbenzene	13.47	13.47	0.000	119	91	2077.031	19.97	20.0	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Quantitation Report: **Applied P & Chn Lab** EPA 524.2

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00020.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 22:56 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:01 2003
 Print Time : Thu Jul 24 12:01 2003
 Miscellaneous :


Sample : f=1 20 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	1313.556	20.40	20.4	100	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	3039.109	19.88	19.9	98	
82	14-DCB	13.87	13.87	0.000	146	148	1295.125	20.04	20.0	99	
83	Cl-benzyl	13.98	13.98	0.000	126	91	132.793	28.12	28.1	76	#
84	12-DCB	14.21	14.21	0.000	146	148	1153.351	20.77	20.8	99	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	689.097	20.94	20.9	84	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	80.335	25.47	25.5	99	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	838.928	20.94	20.9	99	
88	naphthalene	15.79	15.78	0.000	128	129	1300.980	22.16	22.2	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	472.099	19.75	19.8	98	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	719.921	21.61	21.6	98	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00020.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 22:56 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:01 2003
 Print Time : Thu Jul 24 12:02 2003
 Miscellaneous :

Sample : F=1 20 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000



Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00040.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 00:43 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:04 2003
 Print Time : Thu Jul 24 12:05 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	990.930	10.00		0.02	
47	Chlorobenzene-d5	11.54	11.54	0.000	117	82	739.886	10.00		0.00	
62	1,4-Dichlorobenzene	13.85	13.84	0.000	152	150	367.099	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	949.109	41.61		41.6	208.07%
29	1,2-Di-Cl-Et-d4 (7.11	7.08	0.002	65	102	757.790	35.81		35.8	179.05%
55	toluene-d8	9.91	9.89	0.000	98	100	3719.816	39.48		39.5	197.40%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	1208.574	43.26		43.3	216.30%

Target Compounds											
<<< I1 : ISTD ID = 1 >>>											
#	Target Compound	RT	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
3	di-Cl-di-F-methan	1.89	1.85	0.005	85	87	990.705	40.14		40.1	99
4	Chloromethane	2.11	2.07	0.006	50	52	864.850	33.28		33.3	98
2	F114	2.03	2.00	0.005	85	135	544.934	42.35		42.3	45
5	vinyl chloride	2.22	2.19	0.004	62	64	949.966	38.50		38.5	100
6	bromomethane	2.61	2.58	0.004	94	96	411.844	28.77		28.8	98
7	chloroethane	2.73	2.70	0.004	64	66	546.077	34.09		34.1	0
8	tri-Cl-F-methane	3.03	3.00	0.005	101	103	1412.342	38.12		38.1	98
91	Acetonitrile X10	4.07	4.04	0.004	41	40	1639.456	521.17		521.2	95
9	acrolein X10	3.51	3.48	0.004	56	55	937.791	391.44		391.4	0
11	acetone X10	3.70	3.69	0.000	43	58	943.071	371.74		371.7	100
12	ethyl ether X5	3.37	3.34	0.005	59	74	2188.257	173.50		173.5	91
13	1,1-dichloroethene	3.64	3.60	0.005	61	96	1088.836	35.91		35.9	0
14	Iodomethane	3.82	3.78	0.005	142	127	1022.337	61.44		61.4	94
15	F-113	3.65	3.62	0.005	101	151	814.332	36.83		36.8	0
16	acrylonitrile X10	4.52	4.49	0.003	53	52	1601.521	476.63		476.6	98
17	carbon disulfide	3.90	3.87	0.004	76	78	2862.257	32.71		32.7	100
94	Isopropyl Alcohol	3.92	4.01	-0.012	45	43	265.806	893.67		893.7	100
18	methylene chlorid	4.22	4.19	0.004	84	49	960.233	37.20		37.2	94
19	t-12-di-Cl-ethene	4.58	4.55	0.004	96	61	1015.155	40.21		40.2	93

m? m? m? m? m? m? m? m? m? m? m? m? m? m? m? m? m? m? m? m? m? m?

07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03 07/24/03

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00040.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 00:43 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:04 2003
 Print Time : Thu Jul 24 12:05 2003
 Miscellaneous :

Sample : F=1 40 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

1702

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.59	4.56	0.004	73	57	1693.079	36.81	36.8	97	?
95	Tert butyl alcoho	4.41	4.47	-0.008	59	57	482.970	607.12	607.1	100	?
94	allyl chloride	4.07	4.04	0.004	41	76	1216.384	33.02	33.0	99	?
21	11-dichloroethane	5.12	5.09	0.004	63	83	1598.207	38.08	38.1	100	?
97	propionitrile	6.00	5.99	0.000	54	51	64.498	47.25	47.3	100	#
22	c-12-di-Cl-ethene	5.92	5.89	0.003	96	61	1019.103	40.53	40.5	93	?
23	22-Dichloropropan	5.92	5.89	0.003	77	97	1098.975	36.00	36.0	96	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	483.734	48.64	48.6	98	?
25	chloroform	6.37	6.35	0.002	83	85	1597.118	37.75	37.8	97	?
26	tetrahydrofuranX5	6.34	6.32	0.002	42	72	624.422	220.61	220.6	95	?
98	Disopropyl ether	5.24	5.22	0.003	45	87	2521.921	36.97	37.0	97	?
99	ETBE	5.74	5.72	0.003	59	87	2106.754	40.20	40.2	97	?
30	12-dichloroethane	7.22	7.20	0.003	64	62	277.276	35.64	35.6	96	?
32	vinyl acetate X5	5.19	5.17	0.003	43	86	5653.569	184.52	184.5	99	?
92	Nitro Methane(X10	5.79	5.80	0.000	61	46	237.578	103.73	103.7	79	?
33	2-butanoneMEK X10	5.93	5.92	0.002	43	72	1648.493	526.70	526.7	98	?
93	Ethyl Acetate x2	6.04	6.02	0.002	43	61	845.931	65.43	65.4	88	#
34	111-trichloroetha	6.65	6.63	0.002	97	99	1524.561	40.26	40.3	100	?
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	1285.689	39.53	39.5	91	?
36	benzene	7.21	7.19	0.003	78	52	3805.709	38.18	38.2	98	?
37	CCl4	6.91	6.89	0.003	117	119	1420.059	44.44	44.4	100	?
100	Isobutyl alcohol	7.11	7.39	-0.037	43	42	182.637	927.34	927.3	95	?
38	thiophene	7.53	7.51	0.003	84	58	1999.457	39.65	39.7	98	?
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	875.952	40.98	41.0	93	?
40	trichloroethene	8.24	8.23	0.002	130	132	1246.934	45.81	45.8	99	?
41	dibromomethane	8.73	8.71	0.002	174	172	530.196	51.31	51.3	100	?
101	TAME	7.41	7.39	0.002	73	43	1872.529	39.90	39.9	99	?
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	1181.950	40.13	40.1	100	?
43	Me-methacrylate	8.76	8.75	0.002	69	100	464.768	46.12	46.1	92	?
44	2-ClEt-Vi-ether10	9.38	9.37	0.000	63	43	1130.652	502.64	502.6	98	?
45	c-13-di-Cl-propen	9.56	9.55	0.002	75	110	1328.342	42.05	42.1	93	?
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	1100.371	45.10	45.1	93	?

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

me
27/July/03

me
27/July/03

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00040.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 00:43 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:04 2003
 Print Time : Thu Jul 24 12:05 2003
 Miscellaneous :

Sample : f=1 40 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<< I2 :	ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	662.895	43.06	43.1	96	
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	1062.190	39.00	39.0	100	?
50	Et methacrylate	10.38	10.37	0.000	69	99	890.456	41.31	41.3	94	
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	899.440	47.55	47.6	100	
52	bromoform	12.42	12.41	0.000	173	174	534.402	52.94	52.9	100	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	1005.306	40.86	40.9	94	
54	MIBK	9.76	9.76	0.000	43	58	457.366	43.74	43.7	94	
56	toluene	9.99	9.98	0.001	91	92	4284.454	38.69	38.7	98	
57	2-hexanone X5	10.76	10.75	0.000	43	58	1453.530	209.46	209.5	94	
58	12-dibromoethane	11.03	11.03	0.000	107	109	688.415	47.70	47.7	97	
59	tetra-Cl-ethene	10.65	10.64	0.001	166	168	1256.084	44.09	44.1	98	?
60	chlorobenzene	11.58	11.57	0.000	112	77	2837.497	39.78	39.8	92	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	1031.299	45.23	45.2	99	
<<< I3 :	ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	548.810	37.15	37.1	97	?
64	Et-Bz	11.70	11.69	0.000	91	106	4833.949	37.41	37.4	94	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	7112.016	71.71	71.7	95	
66	styrene	12.24	12.23	0.000	104	78	2811.895	37.11	37.1	94	?
67	O-xylene	12.22	12.22	0.000	91	106	3763.083	37.01	37.0	96	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	736.112	41.27	41.3	99	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	222.436	42.93	42.9	99	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	5083.237	40.43	40.4	97	
72	bromobenzene	12.89	12.89	0.000	156	158	1236.729	44.14	44.1	98	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	137.632	51.61	51.6	83	#?
73	n-propylbenzene	13.00	12.99	0.000	120	78	1558.540	42.09	42.1	93	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	1328.259	42.46	42.5	99	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	1291.985	40.28	40.3	99	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	4186.569	38.76	38.8	97	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	4727.712	40.24	40.2	98	?
78	124-tri-Me-Benzen	13.52	13.51	0.000	105	120	4305.294	38.84	38.8	95	
79	tert-butylbenzene	13.47	13.47	0.000	119	91	4034.706	41.63	41.6	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00040.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 00:43 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:04 2003
 Print Time : Thu Jul 24 12:05 2003
 Miscellaneous :

Sample : f=1 40 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000


1704

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	2467.081	41.12	41.1	100	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	5779.243	40.56	40.6	97	
82	14-DCB	13.87	13.87	0.000	146	148	2472.135	41.06	41.1	100	
83	Cl-benzyl	13.98	13.98	0.000	126	91	265.795	60.40	60.4	76	#
84	12-DCB	14.21	14.21	0.000	146	148	2171.646	41.97	42.0	100	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	1335.543	43.55	43.6	81	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	173.925	59.18	59.2	99	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	1658.990	44.44	44.4	99	
88	naphthalene	15.79	15.78	0.000	128	129	2689.379	49.16	49.2	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	923.440	41.46	41.5	98	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	1423.460	45.86	45.9	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00040.D
Method : C:\MSDCHEM\1\METHODS\E524A002.M
Acq. Time : Jul 23 00:43 2003
Method Update: Mon Jul 21 15:01 2003
Quant. Time : Jul 24 12:04 2003
Print Time : Thu Jul 24 12:05 2003
Miscellaneous :

Sample : F=1 40 ppb
Inst. : GCMS-A
RF via : Multiple Level Calibration
Operator: zou
Multiplier: 1.000000



```

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00060.D
Method       : C:\MSDCHEM\1\METHODS\E524A002.M
Acq. Time   : Jul 23 02:45 2003
Method Update: Mon Jul 21 15:01 2003
Quant. Time : Jul 24 12:07 2003
Print Time  : Thu Jul 24 12:07 2003
Miscellaneous :
    
```

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0, ppb	C, ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.64	7.63	0.002	96	70	980.047	10.00		0.01	
47	Chlorobenzene-d5	11.54	11.54	0.000	117	82	736.062	10.00		0.00	
62	1,4-Dichlorobenze	13.85	13.84	0.000	152	150	365.124	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.59	6.58	0.000	111	113	1428.311	63.32		63.3	316.60%
29	1,2-Di-Cl-Et-d4 (7.10	7.08	0.000	65	102	1126.883	53.84		53.8	269.21%
55	toluene-d8	9.90	9.89	0.000	98	100	5400.714	57.62		57.6	288.09%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	1766.234	63.56		63.6	317.81%

Target Compounds											
<<< I1 : ISTD ID = 1 >>>											
3	di-Cl-di-F-methan	1.88	1.85	0.004	85	87	1444.214	59.17		59.2	99
4	Chloromethane	2.14	2.07	0.010	50	52	1174.857	45.71		45.7	99
2	F114	2.03	2.00	0.004	85	135	783.345	61.55		61.5	12
5	vinyl chloride	2.22	2.19	0.003	62	64	1312.837	53.80		53.8	99
6	bromomethane	2.59	2.58	0.002	94	96	626.111	44.22		44.2	97
7	chloroethane	2.71	2.70	0.002	64	66	758.703	47.89		47.9	0
8	tri-Cl-F-methane	3.01	3.00	0.002	101	103	1986.417	54.21		54.2	98
91	Acetonitrile X10	4.05	4.04	0.002	41	40	2365.932	760.46		760.5	95
9	acrolein X10	3.49	3.48	0.002	56	55	1313.901	554.52		554.5	0
11	acetone X10	3.67	3.69	-0.002	43	58	1241.915	494.97		495.0	99
12	ethyl ether X5	3.35	3.34	0.002	59	74	2883.396	231.16		231.2	92
13	11-dichloroethene	3.62	3.60	0.002	61	96	1490.545	49.70		49.7	0
14	Iodomethane	3.80	3.78	0.002	142	127	1206.486	73.32		73.3	95
15	F-113	3.63	3.62	0.002	101	151	1086.062	49.67		49.7	87
16	acrylonitrile X10	4.50	4.49	0.000	53	52	2352.958	708.05		708.0	99
17	carbon disulfide	3.88	3.87	0.002	76	78	3817.157	44.11		44.1	99
94	Isopropyl Alcohol	3.87	4.01	-0.019	45	43	369.178	1255.00		1255.0	100
18	methylene chlorid	4.20	4.19	0.002	84	49	1435.510	56.24		56.2	95
19	t-12-di-Cl-ethene	4.57	4.55	0.002	96	61	1488.348	59.61		59.6	92

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

m? 7/24/03
m? 7/24/03
m? 7/24/03
m? 7/24/03
m? 7/24/03
m? 7/24/03

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00060.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 02:45 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:07 2003
 Print Time : Thu Jul 24 12:07 2003
 Miscellaneous :

Sample : F=1 60 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

1707

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.58	4.56	0.002	73	57	2598.506	57.12	57.1	96	?
95	Tert butyl alcoho	4.37	4.47	-0.013	59	57	785.022	997.78	997.8	100	?
94	allyl chloride	4.05	4.04	0.002	41	76	1660.968	45.58	45.6	93	?
21	11-dichloroethane	5.11	5.09	0.002	63	83	2404.444	57.92	57.9	99	?
97	propionitrile	5.98	5.99	-0.002	54	51	101.359	75.08	75.1	100	#
22	c-12-di-Cl-ethene	5.90	5.89	0.002	96	61	1499.328	60.28	60.3	93	?
23	22-Dichloropropan	5.90	5.89	0.002	77	97	1539.718	51.00	51.0	96	?
24	Br-Cl-methane	6.24	6.23	0.002	128	130	716.534	72.84	72.8	99	?
25	chloroform	6.36	6.35	0.000	83	85	2387.736	57.07	57.1	97	?
26	tetrahydrofuranX5	6.32	6.32	0.000	42	72	946.730	338.19	338.2	95	?
98	Disopropyl ether	5.23	5.22	0.002	45	87	3767.835	55.85	55.8	97	?
99	ETBE	5.73	5.72	0.002	59	87	3231.163	62.34	62.3	97	?
30	12-dichloroethane	7.21	7.20	0.002	64	62	411.282	53.45	53.5	98	?
32	vinyl acetate X5	5.18	5.17	0.002	43	86	8226.056	271.47	271.5	100	?
92	Nitro Methane(X10	5.78	5.80	-0.003	61	46	325.914	143.88	143.9	78	?
33	2-butanoneMEK X10	5.92	5.92	0.000	43	72	2411.259	778.97	779.0	98	?
93	Ethyl Acetate x2	6.03	6.02	0.000	43	61	1295.021	101.28	101.3	87	#
34	111-trichloroetha	6.64	6.63	0.000	97	99	2222.673	59.34	59.3	98	?
35	11-Di-Cl-propene	6.89	6.88	0.000	75	110	1859.527	57.81	57.8	90	?
36	benzene	7.20	7.19	0.002	78	52	5575.015	56.55	56.6	98	?
37	CCl4	6.90	6.89	0.002	117	119	2060.040	65.18	65.2	99	?
100	Isobutyl alcohol	7.08	7.39	-0.041	43	42	287.972	1478.42	1478.4	93	?
38	thiophene	7.52	7.51	0.002	84	58	2960.047	59.36	59.4	99	?
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	1293.157	61.17	61.2	94	?
40	trichloroethene	8.23	8.23	0.000	130	132	1815.606	67.45	67.4	98	?
41	dibromomethane	8.72	8.71	0.000	174	172	793.122	77.61	77.6	99	?
101	TAME	7.41	7.39	0.002	73	43	2927.534	63.08	63.1	98	?
42	Br-di-Cl-methane	8.95	8.95	0.000	83	85	1745.648	59.93	59.9	99	?
43	Me-methacrylate	8.76	8.75	0.000	69	100	714.088	71.64	71.6	92	?
44	2-ClEt-Vi-ether10	9.38	9.37	0.000	63	43	198.558	89.25	89.3	97	?
45	c-13-di-Cl-propen	9.55	9.55	0.000	75	110	1977.341	63.30	63.3	94	?
46	t-1,3-dichloropro	10.24	10.24	0.000	75	110	1653.172	68.51	68.5	92	?

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

m
07/24/03

m
07/24/03

#
07/24/03

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00060.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 02:45 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:07 2003
 Print Time : Thu Jul 24 12:07 2003
 Miscellaneous :

Sample : f=1 60 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<< 12	: ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.45	10.45	0.000	97	83	997.315	65.11	65.1	96	
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	1558.760	57.54	57.5	99	
50	Et methacrylate	10.38	10.37	0.000	69	99	1418.605	66.15	66.2	94	?
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	1365.305	72.56	72.6	100	
52	bromoform	12.41	12.41	0.000	173	174	823.804	82.04	82.0	100	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	1519.089	62.06	62.1	94	
54	MIBK	9.76	9.76	0.000	43	58	733.371	70.50	70.5	92	
56	toluene	9.99	9.98	0.000	91	92	6235.319	56.59	56.6	98	
57	2-hexanone X5	10.75	10.75	0.000	43	58	2205.707	319.50	319.5	93	
58	12-dibromoethane	11.03	11.03	0.000	107	109	1038.913	72.36	72.4	97	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	1800.306	63.52	63.5	98	?
60	chlorobenzene	11.57	11.57	0.000	112	77	4059.222	57.21	57.2	95	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	1537.036	67.77	67.8	99	
<<< 13	: ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	784.184	53.36	53.4	99	?
64	Et-Bz	11.70	11.69	0.000	91	106	6913.463	53.79	53.8	94	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	9936.549	100.73	100.7	95	
66	styrene	12.24	12.23	0.000	104	78	3981.902	52.84	52.8	95	?
67	o-xylene	12.22	12.22	0.000	91	106	5341.093	52.81	52.8	97	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	1101.133	62.07	62.1	99	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	332.760	64.57	64.6	97	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	7208.570	57.65	57.6	96	
72	bromobenzene	12.89	12.89	0.000	156	158	1774.887	63.70	63.7	100	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	207.953	78.40	78.4	83	#?
73	n-propylbenzene	13.00	12.99	0.000	120	78	2210.026	60.01	60.0	93	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	1912.837	61.48	61.5	98	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	1819.259	57.03	57.0	99	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	5867.539	54.61	54.6	97	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	6567.533	56.20	56.2	97	?
78	124-tri-Me-Benzen	13.52	13.51	0.000	105	120	6074.118	55.09	55.1	94	
79	tert-butylbenzene	13.47	13.47	0.000	119	91	5816.456	60.34	60.3	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00060.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 02:45 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:07 2003
 Print Time : Thu Jul 24 12:07 2003
 Miscellaneous :

Sample : f=1 60 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000


1709

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	3488.098	58.45	58.5	99	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	8084.394	57.05	57.0	96	
82	14-DCB	13.87	13.87	0.000	146	148	3533.637	59.01	59.0	99	
83	Cl-benzyl	13.98	13.98	0.000	126	91	361.085	82.50	82.5	75	#
84	12-DCB	14.21	14.21	0.000	146	148	3083.911	59.93	59.9	99	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	1846.727	60.55	60.5	79	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	263.112	90.01	90.0	99	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	2390.948	64.40	64.4	100	
88	naphthalene	15.79	15.78	0.000	128	129	4011.635	73.73	73.7	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	1288.580	58.17	58.2	99	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	2068.944	67.02	67.0	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00060.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 02:45 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:07 2003
 Print Time : Thu Jul 24 12:07 2003
 Miscellaneous :

Sample : f=1 60 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000



Data File : C:\MSDCHEM\1\DATA\03G3570\G3570P01.D
Acq On : 6 Aug 2003 10:07 am

Sample : #03g3570,w 50ng
Misc :
MS Integration Params: Lscint.p
Operator: zou
Inst : GCMS-A
Multiplier: 1.00


Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
Title : **Applied P &Ch Lab** EPA 524.2

Vial: 1

Operator: zou

Inst : GCMS-A

Multiplier: 1.00



Spectrum Information: Average of 12.742 to 12.760 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result
50	95	15	40	15.1	2944	PASS
75	95	30	60	40.6	7912	PASS
95	95	100	100	100.0	19475	PASS
96	95	5	9	7.0	1357	PASS
173	174	0.00	2	0.4	68	PASS
174	95	50	100	93.6	18225	PASS
175	174	5	9	7.8	1416	PASS
176	174	95	101	98.7	17985	PASS
177	176	5	9	7.8	1394	PASS

FORM-5A

Applied P & Ch Laboratory

Volatile Organic Instrument Performance Check for Method 524.2

Bromofluorobenzene (BFB), Part II

Client Name:	GEOFON, Inc.	Contract No:		Lab Code:	APCL
Case No:		SAS No:		Service ID:	034406
Project ID:	JPL	BFB Inj. Date:	<u>08/06/03</u>	Batch No:	03G3570
		BFB Inj. Time:	<u>10:07</u>	Sequence No:	03G3570
Project No:	04-4428.10	Instrument ID:	A	GC Column:	HP-VOC
Data File Name:	G3570P01	Heated Purge:	(Y/N) N	Column ID:	0.20 mm

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS AND STANDARDS:

#	Client Sample No	Lab Sample ID	Data File Name	Date Analyzed	Time Analyzed
1	03G3570-CCV-01	03G3570-CCV-01	G3570Q01	08/06/03	10:32
2	03G3570-LCS-01	03G3570-LCS-01	G3570L01	08/06/03	10:58
3	MW-19-2MS	03-4406-10MS	G3570M01	08/06/03	11:24
4	MW-19-2MSD	03-4406-10MSD	G3570N01	08/06/03	11:50
5	03G3570-MB-01	03G3570-MB-01	G3570K01	08/06/03	13:08
6	MW-19-2	03-4406-10	4406-10	08/06/03	14:27
7	DUPE-1-3Q03	03-4406-1	4406-01	08/06/03	17:05
8	MW-3-2	03-4406-3	4406-03	08/06/03	17:31
9	MW-3-3	03-4406-4	4406-04	08/06/03	17:57
10	MW-3-4	03-4406-5	4406-05	08/06/03	18:23
11	MW-17-2	03-4406-6	4406-06	08/06/03	18:50
12	MW-17-3	03-4406-7	4406-07	08/06/03	19:16
13	MW-17-4	03-4406-8	4406-08	08/06/03	19:42
14	MW-19-1	03-4406-9	4406-09	08/06/03	20:08
15	MW-19-3	03-4406-11	4406-11	08/06/03	20:34
16	MW-19-4	03-4406-12	4406-12	08/06/03	21:00
17	MW-19-5	03-4406-13	4406-13	08/06/03	21:26
18					
19					
20					
21					
22					
23					
24					
25					

Continuing Calibration Concentration Summary

Data File G3570Q01

Method File E524A002

<u>Compound Name</u>	<u>Amount</u>	<u>Actual</u>	<u>Units</u>	<u>%Dev</u>	<u>Target Response</u>
1 Fluorobenzene	10	10.00	ppb	0.00	790353
3 di-Cl-di-F-methane	20	18.07	ppb	9.64	358542
4 Chloromethane	20	16.59	ppb	17.06	279095
2 F114	20	19.23	ppb	3.86	199480
5 vinyl chloride	20	17.96	ppb	10.18	344173
6 bromomethane	20	17.74	ppb	11.30	163266
7 chloroethane	20	18.77	ppb	6.14	214063
8 tri-Cl-F-methane	20	18.85	ppb	5.73	535063
91 Acetonitrile X10	200	176.79	ppb	11.60	613006
9 acrolein X10	200	204.92	ppb	2.46	385070
11 acetone X10	200	347.18	ppb	73.59	620100
12 ethyl ether X5	100	104.62	ppb	4.62	889704
13 11-dichloroethene	20	18.87	ppb	5.66	427487
14 Iodomethane	20	10.38	ppb	48.08	207968
15 F-113	20	21.23	ppb	6.15	330645
16 acrylonitrile X10	200	173.06	ppb	13.47	611466
17 carbon disulfide	20	18.48	ppb	7.60	1064381
94 Isopropyl Alcoholx10	200	176.36	ppb	11.82	87379
18 methylene chloride	20	18.04	ppb	9.81	363409
19 t-12-di-Cl-ethene	20	18.24	ppb	8.79	389462
20 t-Bu-Me-ether	20	18.46	ppb	7.71	655909
95 Tert butyl alcoholx10	200	175.44	ppb	12.28	161293
94 allyl chloride	20	19.72	ppb	1.42	613006
21 11-dichloroethane	20	17.50	ppb	12.49	599613
97 propionitrile	20	20.10	ppb	0.50	26651
22 c-12-di-Cl-ethene	20	18.43	ppb	7.87	396616
23 22-Dichloropropane	20	24.81	ppb	24.04	534764
24 Br-Cl-methane	20	17.99	ppb	10.06	187762
25 chloroform	20	19.68	ppb	1.60	635499
26 tetrahydrofuranX5	100	87.94	ppb	12.06	228264
98 Diisopropyl ether	20	17.94	ppb	10.28	946543
27 Di-Br-F-Me (surr)	20	18.80	ppb	5.98	370366
99 ETBE	20	19.53	ppb	2.35	787537
29 1,2-Di-Cl-Et-d4 (S1)	20	18.78	ppb	6.09	299618
30 12-dichloroethane	20	20.11	ppb	0.54	112076
32 vinyl acetate X5	100	97.31	ppb	2.69	2399525
92 Nitro Methane(x10)	200	169.16	ppb	15.42	74246
33 2-butanoneMEK X10	200	230.39	ppb	15.19	766132
93 Ethyl Acetate x2	40	36.69	ppb	8.27	316586
34 111-trichloroethane	20	18.76	ppb	6.19	597390
35 11-Di-Cl-propene	20	20.32	ppb	1.60	488889
36 benzene	20	18.15	ppb	9.26	1455820
37 CCl4	20	19.47	ppb	2.64	571650

<u>Compound Name</u>	<u>Amount</u>	<u>Actual</u>	<u>Units</u>	<u>%Dev</u>	<u>Target Response</u>
100 Isobutyl alcoholx10	200	170.18	ppb	14.91	61706
38 thiophene	20	19.06	ppb	4.69	760139
39 12-di-Cl-propane	20	17.94	ppb	10.30	322021
40 trichloroethene	20	18.93	ppb	5.36	473283
41 dibromomethane	20	17.86	ppb	10.68	203405
101 TAME	20	20.01	ppb	0.06	699756
42 Br-di-Cl-methane	20	19.39	ppb	3.03	455934
43 Me-methacrylate	20	17.50	ppb	12.48	159581
44 2-ClEt-Vi-ether10	200	116.80	ppb	41.60	232170
45 c-13-di-Cl-propene	20	19.78	ppb	1.09	516018
46 t-1,3-dichloropropene	20	20.36	ppb	1.81	422370
47 Chlorobezene-d5	10	10.00	ppb	0.00	610095
48 112-tri-Cl-Et	20	17.61	ppb	11.95	259291
49 13-di-Cl-propane	20	17.77	ppb	11.16	412652
50 Et methacrylate	20	18.39	ppb	8.05	338969

<u>Compound Name</u>	<u>Amount</u>	<u>Actual</u>	<u>Units</u>	<u>%Dev</u>	<u>Target Response</u>
51 di-Br-Cl-methane	20	18.92	ppb	5.41	349560
52 bromoform	20	17.82	ppb	10.92	201352
53 1,4-dichlorobutane-2	20	17.18	ppb	14.10	371406
54 MIBK	20	17.43	ppb	12.86	154861
55 toluene-d8	20	18.67	ppb	6.63	1435838
56 toluene	20	18.71	ppb	6.47	1652885
57 2-hexanone X5	100	94.13	ppb	5.87	567671
58 12-dibromoethane	20	17.75	ppb	11.27	257766
59 tetra-Cl-ethene	20	18.83	ppb	5.84	497251
60 chlorobenzene	20	19.73	ppb	1.37	1134367
61 1112-tetra-Cl-Et	20	17.91	ppb	10.46	402707
62 1,4-Dichlorobenzene-d4	10	10.00	ppb	0.00	311119
63 1-chlorohexane	20	21.04	ppb	5.20	213480
64 Et-Bz	20	19.58	ppb	2.10	1906033
65 m/p-Xylenes X2	40	39.84	ppb	0.40	2931988
66 styrene	20	20.34	PPB	1.72	1138373
67 o-xylene	20	20.49	ppb	2.47	1513242
68 1122-Tetra-Cl-Et	20	17.06	ppb	14.68	280470
69 123-tri-Cl-Pr	20	17.80	ppb	11.00	87570
70 4-Br-1-F-Bz (S3)	20	18.71	ppb	6.45	481440
71 isopropylbenzene	20	21.24	ppb	6.21	2009263
72 bromobenzene	20	19.03	ppb	4.85	483453
92 t-1,4-dichloro-2-butene	20	19.11	ppb	4.44	53439
73 n-propylbenzene	20	20.87	ppb	4.36	618383
74 2-Cl-Toluene	20	20.01	ppb	0.03	523766
75 4-Cl-Toluene	20	19.77	ppb	1.17	526408
76 135-tri-Me-Benzene	20	21.03	ppb	5.14	1712315
77 4-iso-Pr-toluene	20	21.30	ppb	6.52	1937365
78 124-tri-Me-Benzene	20	20.46	ppb	2.32	1735411
79 tert-butylbenzene	20	20.59	ppb	2.95	1601166
80 13-DCB	20	18.95	ppb	5.27	1011879
81 sec-butylbenzene	20	21.32	ppb	6.60	2333010

82 14-DCB	20	19.61	ppb	1.97	995672
83 Cl-benzyl	20	26.70	ppb	33.52	137437
84 12-DCB	20	18.62	ppb	6.92	885389
85 n-butylbenzene	20	20.33	ppb	1.66	542253
86 12-diBr-2-Cl-Pra	20	18.29	ppb	8.57	61240
87 124-tri-Cl-Bz	20	21.43	ppb	7.13	659771
88 naphthalene	20	17.81	ppb	10.94	995381
89 hx-Cl-butadiene	20	19.80	ppb	1.02	369056
90 123-Tri-Cl-Bz	20	20.10	ppb	0.51	555466

Average D % 8.5139949

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G3570\G3570Q01.D Vial: 2
 Acq On : 6 Aug 2003 10:32 am Operator: zou
 Sample : f=1 Inst : GCMS-A
 Misc : Multiplr: 1.00
 MS Integration Params: Lscint.p

1716

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AvgRRF	CCRF	%Dev	Area#	Dev(min)
1 I Fluorobenzene	1.000	1.000	0.0	78	0.02
2 di-Cl-dl-F-methane	0.251	0.227	9.6	70	0.04
3 P Chloromethane	0.235	0.177	24.7#	62	0.04
4 F114	0.131	0.126	3.8	72	0.04
5 C vinyl chloride	0.242	0.218	9.9	71	0.04
6 bromomethane	0.116	0.103	11.2	71	0.04
7 chloroethane	0.144	0.135	6.2	80	0.03
8 tri-Cl-F-methane	0.359	0.338	5.8	74	0.04
9 Acetonitrile X10	0.044	0.039	11.4	72	0.03
10 acrolein X10	0.028	0.024	14.3	75	0.04
11 acetone X10	0.043	0.039	9.3	125	0.04
12 ethyl ether X5	0.127	0.113	11.0	75	0.04
13 M, C 13 11-dichloroethene	0.287	0.270	5.9	76	0.04
14 Iodomethane	0.253	0.132	47.8#	39#	0.04
15 F-113	0.225	0.209	7.1	77	0.04
16 acrylonitrile X10	0.045	0.039	13.3	74	0.04
17 carbon disulfide	0.729	0.673	7.7	72	0.04
18 Isopropyl AlcoholX10	0.008	0.006	25.0#	70	0.10
19 methylene chloride	0.467	0.230	50.7#	71	0.04
20 t-12-di-Cl-ethene	0.270	0.246	8.9	73	0.04
21 t-Bu-Me-ether	0.450	0.415	7.8	76	0.04
22 Tert butyl alcoholX10	0.012	0.010	16.7	78	0.11
23 allyl chloride	0.393	0.388	1.3	72	0.03
24 P 11-dichloroethane	0.433	0.379	12.5	73	0.03
25 propionitrile	0.017	0.017	0.0	80	0.03
26 c-12-di-Cl-ethene	0.272	0.251	7.7	74	0.02

(#) = Out of Range
 G3570Q01.D E524A002.M Thu Aug 07 16:13:47 2003

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G3570\G3570Q01.D
 Acq On : 6 Aug 2003 10:32 am
 Sample : f=1
 Misc :
 MS Integration Params: Lscint.p

Vial: 2
 Operator: zou
 Inst : GCMS-A
 Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AvgRRF	CCRF	%Dev Area	Dev (min)
27 22-Dichloropropane	0.310	0.338	-9.0	91
28 Br-Cl-methane	0.132	0.119	9.8	75
29 C 25 chloroform	0.478	0.402	15.9	76
30 26 tetrahydrofuranX5	0.033	0.029	12.1	72
31 Disopropyl ether	0.667	0.599	10.2	71
32 S 27 Di-Br-F-Me (surr)	0.249	0.234	6.0	76
33 ETBE	0.510	0.498	2.4	74
34 S 29 1,2-Di-Cl-Et-d4 (S1)	0.202	0.190	5.9	77
35 30 12-dichloroethane	0.078	0.071	9.0	78
36 vinyl acetate X5	0.312	0.304	2.6	81
37 Nitro Methane(X10)	0.006	0.005	16.7	67
38 2-butanoneMEK X10	0.052	0.048	7.7	88
39 Ethyl Acetate x2	0.124	0.100	19.4	73
40 34 111-trichloroethane	0.403	0.378	6.2	76
41 35 11-Di-Cl-propene	0.304	0.309	-1.6	74
42 M 36 benzene	1.015	0.921	9.3	73
43 CCl4	0.371	0.362	2.4	78
44 100 Isobutyl alcoholX10	0.004	0.004	0.0	71
45 thiophene	0.505	0.481	4.8	74
46 C 39 12-di-Cl-propane	0.227	0.204	10.1	72
47 M 40 trichloroethene	0.316	0.299	5.4	75
48 41 dibromomethane	0.144	0.129	10.4	75
49 101 TAME	0.442	0.443	-0.2	75
50 42 Br-di-Cl-methane	0.327	0.288	11.9	75
51 43 Me-methacrylate	0.102	0.101	1.0	70
52 44 2-ClEt-Vl-ether10	0.020	0.015	25.0#	45#

(#) = Out of Range
 G3570Q01.D E524A002.M Thu Aug 07 16:13:48 2003

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G3570\G3570Q01.D Vial: 2
 Acq On : 6 Aug 2003 10:32 am Operator: zou
 Sample : F=1 Inst : GCMS-A
 Misc : Multiplr: 1.00
 MS Integration Params: Lscint.p

1718

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AvgRRF	CCRF	%Dev	Area%	Dev (min)
53 c-13-di-Cl-propene	0.330	0.326	1.2	76	0.01
54 t-1,3-dichloropropene	0.262	0.267	-1.9	77	0.00
55 I Chlorobenzene-d5	1.000	1.000	0.0	79	0.00
56 112-tri-Cl-Et	0.241	0.213	11.6	76	0.00
57 13-di-Cl-propane	0.381	0.338	11.3	75	0.01
58 Et methacrylate	0.273	0.278	-1.8	76	0.00
59 di-Br-Cl-methane	0.325	0.286	12.0	77	0.00
60 P bromoform	0.185	0.165	10.8	76	0.00
61 1,4-dichlorobutane-2	0.354	0.304	14.1	73	0.00
62 MIBK	0.146	0.127	13.0	70	0.00
63 s toluene-d8	1.260	1.177	6.6	75	0.01
64 M,C toluene	1.448	1.355	6.4	74	0.00
65 2-hexanone X5	0.099	0.093	6.1	78	0.00
66 12-dibromoethane	0.238	0.211	11.3	75	0.00
67 tetra-Cl-ethene	0.433	0.408	5.8	75	0.00
68 M,P chlorobenzene	1.032	0.930	9.9	76	0.00
69 1112-tetra-Cl-Et	0.369	0.330	10.6	77	0.00
70 I 1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
71 1-chlorohexane	0.326	0.343	-5.2	76	0.00
72 C Et-Bz	3.129	3.063	2.1	75	0.00
73 m/p-Xylenes X2	2.365	2.356	0.4	76	0.00
74 styrene	1.798	1.829	-1.7	76	0.00
75 o-xylene	2.373	2.432	-2.5	76	0.00
76 p 1122-Tetra-Cl-Et	0.528	0.451	14.6	75	0.00

(#) = Out of Range

G3570Q01.D E524A002.M Thu Aug 07 16:13:49 2003

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G3570\G3570Q01.D
 Acq On : 6 Aug 2003 10:32 am
 Sample : f=1
 Misc :
 MS Integration Params: lscint.p

Vial: 2
 Operator: zou
 Inst : GCMS-A
 Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
77 69 123-tri-Cl-Pr	0.158	0.141	10.8	76	0.00
78 S 70 4-Br-1-F-Bz (S3)	0.827	0.774	6.4	77	0.00
79 71 isopropylbenzene	3.040	3.229	-6.2	76	0.00
80 72 bromobenzene	0.817	0.777	4.9	76	0.00
81 92 t-1,4-dichloro-2-butene	0.077	0.086	-11.7	83	0.00
82 73 n-propylbenzene	0.952	0.994	-4.4	77	0.00
83 74 2-Cl-Toluene	0.841	0.842	-0.1	76	0.00
84 75 4-Cl-Toluene	0.856	0.846	1.2	78	0.00
85 76 135-tri-Me-Benzene	2.617	2.752	-5.2	77	0.00
86 77 4-iso-Pr-toluene	2.923	3.114	-6.5	78	0.00
87 78 124-tri-Me-Benzene	2.726	2.789	-2.3	77	0.00
88 79 tert-butylbenzene	2.499	2.573	-3.0	77	0.00
89 80 13-DCB	1.717	1.626	5.3	77	0.00
90 81 sec-butylbenzene	3.517	3.749	-6.6	77	0.00
91 82 14-DCB	1.749	1.600	8.5	77	0.00
92 83 Cl-benzyl	0.165	0.221	-33.9#	103	0.00
93 84 12-DCB	1.529	1.423	6.9	77	0.00
94 85 n-butylbenzene	0.779	0.871	-11.8	79	0.00
95 86 12-diBr-2-Cl-Pra	0.108	0.098	9.3	76	0.00
96 87 124-tri-Cl-Bz	0.990	1.060	-7.1	79	0.00
97 88 naphthalene	1.796	1.600	10.9	77	0.00
98 89 hx-Cl-butadiene	0.599	0.593	1.0	78	0.00
99 90 123-Tri-Cl-Bz	0.888	0.893	-0.6	77	0.00

(#) = Out of Range
 G3570Q01.D E524A002.M Thu Aug 07 16:13:49 2003
 SPCC's out = 0 CCC's out = 0

Data Filename: C:\MSDCHEM\1\DATA\03G3570\G3570Q01.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 6 10:32 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 07 16:09 2003
 Print Time : Thu Aug 07 16:10 2003
 Miscellaneous :

Sample : F=1
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

1720

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	790.353	10.00		0.02	
47	Chlorobenzene-d5	11.55	11.54	0.000	117	82	610.095	10.00		0.00	
62	1,4-Dichlorobenze	13.85	13.84	0.000	152	150	311.119	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	370.366	18.80	18.8	94.02%	
29	1,2-Di-Cl-Et-d4 (7.11	7.08	0.002	65	102	299.618	18.78	18.8	93.91%	
55	toluene-d8	9.91	9.89	0.000	98	100	1435.838	18.67	18.7	93.37%	
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	481.440	18.71	18.7	93.55%	

Target Compounds	<<< I1 : ISTD ID = 1 >>>	Qvalue									
3	di-Cl-di-F-methan	1.89	1.85	0.006	85	87	358.542	18.07	18.1	99	
4	Chloromethane	2.11	2.07	0.006	50	52	279.095	16.59	16.6	99	
2	F114	2.04	2.00	0.006	85	135	199.480	19.23	19.2	51	
5	vinyl chloride	2.23	2.19	0.005	62	64	344.173	17.96	18.0	99	
6	bromomethane	2.61	2.58	0.005	94	96	163.266	17.74	17.7	98	
7	chloroethane	2.73	2.70	0.004	64	66	214.063	18.77	18.8	98	
8	tri-Cl-F-methane	3.04	3.00	0.006	101	103	535.063	18.85	18.9	98	
91	Acetonitrile X10	4.07	4.04	0.004	41	40	613.006	176.79	176.8	92	
9	acrolein X10	3.52	3.48	0.006	56	55	385.070	204.92	204.9	0	
11	acetone X10	3.73	3.69	0.006	43	58	620.100	347.18	347.2	0	
12	ethyl ether X5	3.37	3.34	0.005	59	74	889.704	104.62	104.6	91	
13	1,1-dichloroethene	3.64	3.60	0.005	61	96	427.487	18.87	18.9	0	
14	Iodomethane	3.82	3.78	0.005	142	127	207.968	10.38	10.4	94	
15	F-113	3.65	3.62	0.005	101	151	330.645	21.23	21.2	88	
16	acrylonitrile X10	4.53	4.49	0.005	53	52	611.466	173.06	173.1	98	
17	carbon disulfide	3.91	3.87	0.005	76	78	1064.381	18.48	18.5	99	
94	Isopropyl Alcohol	4.12	4.01	0.014	45	43	87.379	176.36	176.4	100	
18	methylene chlorid	4.22	4.19	0.005	84	49	363.409	18.04	18.0	99	
19	t-12-di-Cl-ethene	4.58	4.55	0.005	96	61	389.462	18.24	18.2	93	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Handwritten signatures and dates:
 mDe 8/6/03
 mDe 8/6/03
 mDe 8/6/03
 ?
 ?
 ?
 ?

Data Filename: C:\MSDCHEM\1\DATA\03G3570\G3570Q01.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 6 10:32 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 07 16:09 2003
 Print Time : Thu Aug 07 16:10 2003
 Miscellaneous :

Sample : f=1
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000

1721

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.60	4.56	0.005	73	57	655.909	18.46	18.5	97	?
95	Tert butyl alcoho	4.58	4.47	0.014	59	57	161.293	175.44	175.4	100	m? <i>De</i>
94	allyl chloride	4.07	4.04	0.004	41	76	613.006	19.72	19.7	85	#? <i>De</i>
21	11-dichloroethane	5.12	5.09	0.004	63	83	599.613	17.50	17.5	98	
97	propionitrile	6.03	5.99	0.004	54	51	26.651	20.10	20.1	100	#?
22	c-12-di-Cl-ethene	5.92	5.89	0.003	96	61	396.616	18.43	18.4	87	?
23	22-Dichloropropan	5.92	5.89	0.004	77	97	534.764	24.81	24.8	99	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	187.762	17.99	18.0	98	?
25	chloroform	6.38	6.35	0.003	83	85	635.499	19.68	19.7	97	?
26	tetrahydrofuranX5	6.35	6.32	0.004	42	72	228.264	17.94	17.9	94	?
98	Diisopropyl ether	5.24	5.22	0.003	45	87	946.543	19.53	19.5	98	
99	ETBE	5.74	5.72	0.003	59	87	787.537	20.11	20.1	100	?
30	12-dichloroethane	7.22	7.20	0.004	64	62	112.076	97.31	97.3	99	
32	vinyl acetate X5	5.20	5.17	0.005	61	46	74.246	169.16	169.2	92	
92	Nitro Methane(X10	5.84	5.80	0.004	43	86	2399.525	20.11	20.1	100	
33	2-butanoneMEK X10	5.95	5.92	0.004	43	72	766.132	230.39	230.4	98	
93	Ethyl Acetate X2	6.04	6.02	0.003	43	61	316.586	36.69	36.7	91	#?
34	111-trichloroetha	6.66	6.63	0.003	97	99	597.390	18.76	18.8	99	
35	11-Di-Cl-propene	6.90	6.88	0.002	75	110	488.889	20.32	20.3	94	?
36	benzene	7.21	7.19	0.003	78	52	1455.820	18.15	18.1	98	?
37	CCl4	6.91	6.89	0.003	117	119	571.650	19.47	19.5	98	?
100	Isobutyl alcohol	7.19	7.39	-0.026	43	42	61.706	170.18	170.2	94	m? <i>De</i>
38	thiophene	7.53	7.51	0.003	84	58	760.139	19.06	19.1	96	
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	322.021	17.94	17.9	98	
40	trichloroethene	8.25	8.23	0.002	130	132	473.283	18.93	18.9	98	
41	dibromomethane	8.73	8.71	0.002	174	172	203.405	17.86	17.9	100	
101	TAME	7.41	7.39	0.002	73	43	699.756	20.01	20.0	97	
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	455.934	19.39	19.4	100	
43	Me-methacrylate	8.76	8.75	0.002	69	100	159.581	17.50	17.5	92	
44	2-ClEt-Vi-ether10	9.38	9.37	0.002	63	43	232.170	116.80	116.8	96	
45	c-13-di-Cl-propen	9.56	9.55	0.002	75	110	516.018	19.78	19.8	93	
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	422.370	20.36	20.4	90	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3570\G3570Q01.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 6 10:32 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 07 16:09 2003
 Print Time : Thu Aug 07 16:10 2003
 Misceneous :

Sample : F=1
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<< 12	: ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	259.291	17.61	17.6	96	
49	13-di-Cl-propane	10.65	10.64	0.001	76	78	412.652	17.77	17.8	99	?
50	Et methacrylate	10.38	10.37	0.000	69	99	338.969	18.39	18.4	96	
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	349.560	18.92	18.9	99	
52	bromoform	12.42	12.41	0.000	173	174	201.352	17.82	17.8	100	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	371.406	17.18	17.2	95	
54	MIBK	9.77	9.76	0.000	43	58	154.861	17.43	17.4	89	
56	toluene	9.99	9.98	0.000	91	92	1652.885	18.71	18.7	100	
57	2-hexanone X5	10.76	10.75	0.000	43	58	567.671	94.13	94.1	94	
58	12-dibromoethane	11.03	11.03	0.000	107	109	257.766	17.75	17.7	97	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	497.251	18.83	18.8	98	?
60	chlorobenzene	11.58	11.57	0.000	112	77	1134.367	19.73	19.7	91	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	402.707	17.91	17.9	98	
<<< 13	: ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	213.480	21.04	21.0	99	?
64	Et-Bz	11.70	11.69	0.000	91	106	1906.033	19.58	19.6	96	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	2931.988	39.84	39.8	97	
66	styrene	12.24	12.23	0.000	104	78	1138.373	20.34	20.3	93	?
67	o-xylene	12.22	12.22	0.000	91	106	1513.242	20.49	20.5	96	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	280.470	17.06	17.1	98	
69	123-tri-Cl-Pr	12.92	12.91	0.000	110	97	87.570	17.80	17.8	97	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	2009.263	21.24	21.2	97	
72	bromobenzene	12.89	12.89	0.000	156	158	483.453	19.03	19.0	100	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	53.439	19.11	19.1	85	?
73	n-propylbenzene	13.00	12.99	0.000	120	78	618.383	20.87	20.9	93	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	523.766	20.01	20.0	99	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	526.408	19.77	19.8	99	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	1712.315	21.03	21.0	98	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	1937.365	21.30	21.3	98	?
78	124-tri-Me-Benzen	13.52	13.51	0.000	105	120	1735.411	20.46	20.5	96	
79	tert-butylbenzene	13.47	13.47	0.000	119	91	1601.166	20.59	20.6	95	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3570\G3570Q01.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 6 10:32 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 07 16:09 2003
 Print Time : Thu Aug 07 16:10 2003
 Miscellaneous :


ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	1011.879	18.95	18.9	100	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	2333.010	21.32	21.3	97	
82	14-DCB	13.87	13.87	0.000	146	148	995.672	19.61	19.6	97	
83	Cl-benzyl	13.98	13.98	0.000	126	91	137.437	26.70	26.7	74	#
84	12-DCB	14.21	14.21	0.000	146	148	885.389	18.62	18.6	99	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	542.253	20.33	20.3	85	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	61.240	18.29	18.3	99	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	659.771	21.43	21.4	98	
88	naphthalene	15.79	15.78	0.000	128	129	995.381	17.81	17.8	99	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	369.056	19.80	19.8	99	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	555.466	20.10	20.1	99	

Sample : F=1
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06


Data Filename: C:\MSDCHEM\1\DATA\03G3570Q01.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 6 10:32 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 07 16:09 2003
 Print Time : Thu Aug 07 16:10 2003
 Miscellaneous :

Sample : F=1
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000



Data File : C:\MSDCHEM\1\DATA\03G3595\G3595P01.D
 Acq On : 7 Aug 2003 10:15 am
 Sample : #03g3595,w 50ng
 Misc :
 MS Integration Params: Lscint.p
 Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2

Vial: 1
 Operator: zou
 Inst : GCMS-A
 Multiplr: 1.00



Spectrum Information: Average of 12.723 to 12.766 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result
50	95	15	40	16.0	2280	PASS
75	95	30	60	42.8	6115	PASS
95	95	100	100	100.0	14285	PASS
96	95	5	9	7.1	1020	PASS
173	174	0.00	2	0.6	76	PASS
174	95	50	100	85.3	12189	PASS
175	174	5	9	8.7	1066	PASS
176	174	95	101	96.9	11811	PASS
177	176	5	9	7.1	839	PASS

FORM-5A

Applied P & Ch Laboratory

Volatile Organic Instrument Performance Check for Method 524.2

Bromofluorobenzene (BFB), Part II

Client Name: GEOFON, Inc.	Contract No:	Lab Code: APCL
Case No:	SAS No:	Service ID: 034406
Project ID: JPL	BFB Inj. Date: <u>08/07/03</u>	Batch No: 03G3595
Project No: 04-4428.10	BFB Inj. Time: <u>10:15</u>	Sequence No: 03G3595
Data File Name: G3595P01	Instrument ID: A	GC Column: HP-VOC
	Heated Purge: (Y/N) N	Column ID: 0.20 mm

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS AND STANDARDS:

#	Client Sample No	Lab Sample ID	Data File Name	Date Analyzed	Time Analyzed
1	03G3595-CCV-01	03G3595-CCV-01	G3595Q01	08/07/03	10:41
2	03G3595-LCS-01	03G3595-LCS-01	G3595L01	08/07/03	11:06
3	MW-20-4MS	03-4425-6MS	G3595M01	08/07/03	11:32
4	MW-20-4MSD	03-4425-6MSD	G3595N01	08/07/03	11:58
5	03G3595-MB-01	03G3595-MB-01	G3595K01	08/07/03	13:17
6	TB-2-7/30/03	03-4406-14	4406-14	08/07/03	13:43
7	EB-2-7/30/03	03-4406-2	4406-02	08/07/03	14:10
8					
9					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					
23					
24					
25					

82 14-DCB	20	20.20	ppb	0.99	1031759
83 Cl-benzyl	20	28.05	ppb	40.25	145214
84 12-DCB	20	19.18	ppb	4.09	917679
85 n-butylbenzene	20	20.76	ppb	3.82	557165
86 12-diBr-2-Cl-Pra	20	19.08	ppb	4.61	64268
87 124-tri-Cl-Bz	20	21.76	ppb	8.82	674132
88 naphthalene	20	18.30	ppb	8.52	1028466
89 hx-Cl-butadiene	20	20.47	ppb	2.35	383886
90 123-Tri-Cl-Bz	20	20.86	ppb	4.31	579865

Average D % 7.4449627

<u>Compound Name</u>	<u>Amount</u>	<u>Actual</u>	<u>Units</u>	<u>%Dev</u>	<u>Target Response</u>
100 Isobutyl alcoholx10	200	183.27	ppb	8.37	64748
38 thiophene	20	19.96	ppb	0.22	771980
39 12-di-Cl-propane	20	18.69	ppb	6.55	325445
40 trichloroethene	20	19.66	ppb	1.70	476880
41 dibromomethane	20	18.70	ppb	6.48	206592
101 TAME	20	20.73	ppb	3.64	703089
42 Br-di-Cl-methane	20	20.42	ppb	2.08	465535
43 Me-methacrylate	20	18.17	ppb	9.15	161032
44 2-ClEt-Vi-ether10	200	117.61	ppb	41.20	227015
45 c-13-di-Cl-propene	20	20.85	ppb	4.27	527707
46 t-1,3-dichloropropene	20	21.34	ppb	6.68	429352
47 Chlorobezene-d5	10	10.00	ppb	0.00	597320
48 112-tri-Cl-Et	20	18.48	ppb	7.60	266404
49 13-di-Cl-propane	20	18.68	ppb	6.60	424741
50 Et methacrylate	20	19.01	ppb	4.93	343789

<u>Compound Name</u>	<u>Amount</u>	<u>Actual</u>	<u>Units</u>	<u>%Dev</u>	<u>Target Response</u>
51 di-Br-Cl-methane	20	19.96	ppb	0.20	361466
52 bromoform	20	18.66	ppb	6.72	206444
53 1,4-dichlorobutane-2	20	18.19	ppb	9.07	384932
54 MIBK	20	18.40	ppb	7.98	160102
55 toluene-d8	20	19.50	ppb	2.50	1468019
56 toluene	20	19.39	ppb	3.06	1677394
57 2-hexanone X5	100	106.29	ppb	6.29	627565
58 12-dibromoethane	20	18.64	ppb	6.82	265041
59 tetra-Cl-ethene	20	19.66	ppb	1.72	508108
60 chlorobenzene	20	20.66	ppb	3.29	1162225
61 1112-tetra-Cl-Et	20	18.89	ppb	5.56	415886
62 1,4-Dichlorobenzene-d4	10	10.00	ppb	0.00	312953
63 1-chlorohexane	20	21.39	ppb	6.95	218294
64 Et-Bz	20	19.83	ppb	0.87	1941342
65 m/p-Xylenes X2	40	40.51	ppb	1.27	2998617
66 styrene	20	20.57	PPB	2.84	1157661
67 o-xylene	20	20.82	ppb	4.08	1545970
68 1122-Tetra-Cl-Et	20	17.60	ppb	12.02	290912
69 123-tri-Cl-Pr	20	18.31	ppb	8.47	90592
70 4-Br-1-F-Bz (S3)	20	19.02	ppb	4.89	492386
71 isopropylbenzene	20	21.60	ppb	8.02	2055482
72 bromobenzene	20	19.55	ppb	2.27	499474
92 t-1,4-dichloro-2-butene	20	19.62	ppb	1.90	55283
73 n-propylbenzene	20	21.18	ppb	5.91	631252
74 2-Cl-Toluene	20	20.48	ppb	2.38	539229
75 4-Cl-Toluene	20	20.09	ppb	0.46	538263
76 135-tri-Me-Benzene	20	21.40	ppb	7.02	1753181
77 4-iso-Pr-toluene	20	21.77	ppb	8.85	1991490
78 124-tri-Me-Benzene	20	20.86	ppb	4.32	1779697
79 tert-butylbenzene	20	21.03	ppb	5.14	1644833
80 13-DCB	20	19.42	ppb	2.89	1043489
81 sec-butylbenzene	20	21.75	ppb	8.76	2394290

Continuing Calibration Concentration Summary

Data File G3595Q01

Method File E524A002

<u>Compound Name</u>	<u>Amount</u>	<u>Actual</u>	<u>Units</u>	<u>%Dev</u>	<u>Target Response</u>
1 Fluorobenzene	10	10.00	ppb	0.00	766703
3 di-Cl-di-F-methane	20	18.80	ppb	5.98	361919
4 Chloromethane	20	17.47	ppb	12.66	284413
2 F114	20	19.00	ppb	5.02	191179
5 vinyl chloride	20	18.51	ppb	7.44	344051
6 bromomethane	20	20.94	ppb	4.68	186931
7 chloroethane	20	19.40	ppb	2.99	214645
8 tri-Cl-F-methane	20	20.23	ppb	1.13	556813
91 Acetonitrile X10	200	183.60	ppb	8.20	617560
9 acrolein X10	200	221.95	ppb	10.98	402646
11 acetone X10	200	404.53	ppb	102.26	693038
12 ethyl ether X5	100	109.69	ppb	9.69	901636
13 11-dichloroethene	20	19.60	ppb	1.98	430876
14 Iodomethane	20	11.00	ppb	44.99	213763
15 F-113	20	22.34	ppb	11.72	336809
16 acrylonitrile X10	200	181.75	ppb	9.12	622938
17 carbon disulfide	20	19.10	ppb	4.52	1067044
94 Isopropyl Alcoholx10	200	180.25	ppb	9.87	86681
18 methylene chloride	20	18.77	ppb	6.13	365911
19 t-12-di-Cl-ethene	20	19.05	ppb	4.77	394497
20 t-Bu-Me-ether	20	19.34	ppb	3.30	666689
95 Tert butyl alcoholx10	200	188.39	ppb	5.80	169870
94 allyl chloride	20	20.48	ppb	2.38	617560
21 11-dichloroethane	20	18.25	ppb	8.77	606390
97 propionitrile	20	20.72	ppb	3.59	26648
22 c-12-di-Cl-ethene	20	19.23	ppb	3.84	401595
23 22-Dichloropropane	20	26.02	ppb	30.11	543217
24 Br-Cl-methane	20	18.86	ppb	5.72	190934
25 chloroform	20	20.59	ppb	2.97	644722
26 tetrahydrofuranX5	100	91.61	ppb	8.39	230672
98 Diisopropyl ether	20	18.82	ppb	5.91	962962
27 Di-Br-F-Me (surr)	20	19.88	ppb	0.58	379920
99 ETBE	20	20.31	ppb	1.57	794685
29 1,2-Di-Cl-Et-d4 (S1)	20	19.62	ppb	1.88	303684
30 12-dichloroethane	20	21.09	ppb	5.44	113945
32 vinyl acetate X5	100	101.71	ppb	1.71	2432956
92 Nitro Methane(x10)	200	189.09	ppb	5.45	80510
33 2-butanoneMEK X10	200	260.01	ppb	30.00	835421
93 Ethyl Acetate x2	40	40.46	ppb	1.16	338571
34 111-trichloroethane	20	19.84	ppb	0.78	612912
35 11-Di-Cl-propene	20	20.99	ppb	4.93	489795
36 benzene	20	18.89	ppb	5.56	1469920
37 CCl4	20	20.60	ppb	2.98	586558

Evaluate Continuing Calibration Report

1730

Data File : C:\MSDCHEM\1\DATA\03G3595\G3595Q01.D
 Acq On : 7 Aug 2003 10:41 am
 Sample : f=1
 Misc :
 MS Integration Params: Lscint.p

Vial: 2
 Operator: zou
 Inst : GCMS-A
 Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P &Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AvgRRF	CCRF	%Dev	Area	% Dev	(min)
1 I	1.000	1.000	0.0	76	0.02	
2	0.251	0.236	6.0	71	0.04	
3 p	0.235	0.185	21.3#	63	0.04	
4	0.131	0.125	4.6	69	0.04	
5 C	0.242	0.224	7.4	71	0.04	
6	0.116	0.122	-5.2	82	0.04	
7	0.144	0.140	2.8	81	0.03	
8	0.359	0.363	-1.1	77	0.04	
9	0.044	0.040	9.1	73	0.03	
10	0.028	0.026	7.1	78	0.04	
11	0.043	0.045	-4.7	139	0.04	
12	0.127	0.118	7.1	76	0.04	
13 M, C	0.287	0.281	2.1	77	0.04	
14	0.253	0.139	45.1#	40#	0.04	
15	0.225	0.220	2.2	79	0.04	
16	0.045	0.041	8.9	75	0.03	
17	0.729	0.696	4.5	72	0.04	
18	0.008	0.006	25.0#	69	0.09	
19	0.467	0.239	48.8#	72	0.04	
20	0.270	0.257	4.8	74	0.04	
21	0.450	0.435	3.3	77	0.04	
22	0.012	0.011	8.3	82	0.10	
23	0.393	0.403	-2.5	73	0.03	
24 P	0.433	0.395	8.8	73	0.03	
25	0.017	0.017	0.0	80	0.04	
26	0.272	0.262	3.7	75	0.02	

(#) = Out of Range
 G3595Q01.D E524A002.M Fri Aug 08 12:25:06 2003

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G3595\G3595Q01.D
 Acq On : 7 Aug 2003 10:41 am
 Sample : f=1
 Misc :
 MS Integration Params: Lscint.p
 Vial: 2
 Operator: zou
 Inst : GCMS-A
 Multiplr: 1.00

1731

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P &h Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AVGRF	CCRF	%Dev Area	Dev(min)		
27	23 22-Dichloropropane	0.310	0.354	-14.2	92	0.03
28	24 Br-Cl-methane	0.132	0.125	5.3	76	0.02
29	25 chloroform	0.478	0.420	12.1	77	0.02
30	26 tetrahydrofuranX5	0.033	0.030	9.1	72	0.03
31	98 Disopropyl ether	0.667	0.628	5.8	73	0.02
32	27 Di-Br-F-Me (surr)	0.249	0.248	0.4	78	0.03
33	99 ETBE	0.510	0.518	-1.6	75	0.02
34	29 1,2-Di-Cl-Et-d4 (S1)	0.202	0.198	2.0	78	0.03
35	30 12-dichloroethane	0.078	0.074	5.1	80	0.02
36	32 vinyl acetate X5	0.312	0.317	-1.6	82	0.03
37	92 Nitro Methane (X10)	0.006	0.005	16.7	72	0.03
38	33 2-butanoneMEK X10	0.052	0.054	-3.8	95	0.03
39	93 Ethyl Acetate X2	0.124	0.110	11.3	78	0.02
40	34 111-trichloroethane	0.403	0.400	0.7	78	0.02
41	35 11-Di-Cl-propene	0.304	0.319	-4.9	74	0.02
42	36 benzene	1.015	0.959	5.5	74	0.02
43	37 CCl4	0.371	0.383	-3.2	80	0.02
44	100 Isobutyl alcoholX10	0.004	0.004	0.0	74	-0.21
45	38 thiophene	0.505	0.503	0.4	75	0.02
46	39 12-di-Cl-propane	0.227	0.212	6.6	73	0.02
47	40 trichloroethene	0.316	0.311	1.6	75	0.02
48	41 dibromomethane	0.144	0.135	6.2	76	0.01
49	101 TAME	0.442	0.459	-3.8	75	0.02
50	42 Br-di-Cl-methane	0.327	0.304	7.0	77	0.01
51	43 Me-methacrylate	0.102	0.105	-2.9	71	0.01
52	44 2-ClEt-Vi-ether10	0.020	0.015	25.0#	44#	0.01

(#) = Out of Range
 G3595Q01.D E524A002.M Fri Aug 08 12:25:07 2003

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G3595\G3595Q01.D
 Acq On : 7 Aug 2003 10:41 am
 Sample : F=1
 Misc :
 MS Integration Params: Lscint.p

1732

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AvgRF	CCRF	%Dev	Area	% Dev		
53 I	45	c-13-di-Cl-propene	0.330	0.344	-4.2	78	0.01
54	46	t-1,3-dichloropropene	0.262	0.280	-6.9	78	0.00
55 I	47	Chlorobezene-d5	1.000	1.000	0.0	77	0.00
56	48	112-tri-Cl-Et	0.241	0.223	7.5	78	0.00
57	49	13-di-Cl-propane	0.381	0.356	6.6	77	0.00
58	50	Et methacrylate	0.273	0.288	-5.5	78	0.00
59	51	di-Br-Cl-methane	0.325	0.303	6.8	79	0.00
60 P	52	bromoform	0.185	0.173	6.5	78	0.00
61	53	1,4-dichlorobutane-2	0.354	0.322	9.0	75	0.00
62	54	MTBK	0.146	0.134	8.2	72	0.00
63 s	55	toluene-d8	1.260	1.229	2.5	76	0.01
64 M,C	56	toluene	1.448	1.404	3.0	75	0.00
65	57	2-hexanone X5	0.099	0.105	-6.1	86	0.00
66	58	12-dibromoethane	0.238	0.222	6.7	77	0.00
67	59	tetra-Cl-ethene	0.433	0.425	1.8	77	0.00
68 M,P	60	chlorobenzene	1.032	0.973	5.7	78	0.00
69	61	1112-tetra-Cl-Et	0.369	0.348	5.7	79	0.00
70 I	62	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
71	63	1-chlorohexane	0.326	0.349	-7.1	78	0.00
72 C	64	Et-Bz	3.129	3.102	0.9	77	0.00
73	65	m/p-Xylenes X2	2.365	2.395	-1.3	78	0.00
74	66	styrene	1.798	1.850	-2.9	77	0.00
75	67	o-xylene	2.373	2.470	-4.1	78	0.00
76 p	68	1122-Tetra-Cl-Et	0.528	0.465	11.9	78	0.00

(#) = Out of Range
 G3595Q01.D E524A002.M Fri Aug 08 12:25:07 2003

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G3595\G3595Q01.D
 Acq On : 7 Aug 2003 10:41 am
 Sample : f=1
 Misc :
 MS Integration Params: Lscint.p

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & H Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AVGRF	CCRF	%Dev	Area%	Dev(min)
77 69 123-tri-Cl-Pr	0.158	0.145	8.2	79	0.00
78 S 70 4-Br-1-F-Bz (S3)	0.827	0.787	4.8	78	0.00
79 71 isopropylbenzene	3.040	3.284	-8.0	78	0.00
80 72 bromobenzene	0.817	0.798	2.3	79	0.00
81 92 t-1,4-dichloro-2-butene	0.077	0.088	-14.3	86	0.00
82 73 n-propylbenzene	0.952	1.009	-6.0	79	0.00
83 74 2-Cl-Toluene	0.841	0.862	-2.5	79	0.00
84 75 4-Cl-Toluene	0.856	0.860	-0.5	79	0.00
85 76 135-tri-Me-Benzene	2.617	2.801	-7.0	79	0.00
86 77 4-iso-Pr-toluene	2.923	3.182	-8.9	80	0.00
87 78 124-tri-Me-Benzene	2.726	2.843	-4.3	79	0.00
88 79 tert-butylbenzene	2.499	2.628	-5.2	79	0.00
89 80 13-DCB	1.717	1.667	2.9	79	0.00
90 81 sec-butylbenzene	3.517	3.825	-8.8	79	0.00
91 82 14-DCB	1.749	1.648	5.8	80	0.00
92 83 Cl-benzyl	0.165	0.232	-40.6#	109	0.00
93 84 12-DCB	1.529	1.466	4.1	80	0.00
94 85 n-butylbenzene	0.779	0.890	-14.2	81	0.00
95 86 12-diBr-2-Cl-Pra	0.108	0.103	4.6	80	0.00
96 87 124-tri-Cl-Bz	0.990	1.077	-8.8	80	0.00
97 88 naphthalene	1.796	1.643	8.5	79	0.00
98 89 hx-Cl-butadiene	0.599	0.613	-2.3	81	0.00
99 90 123-Tri-Cl-Bz	0.888	0.926	-4.3	81	0.00

(#) = Out of Range
 G3595Q01.D E524A002.M
 SPCC's out = 0
 CGC's out = 0
 Fri Aug 08 12:25:08 2003

FORM-8A

Applied P & Ch Laboratory

Internal Standard Area and RT Summary for Method 524.2

Client Name: GEOFON, Inc. Contract No: Lab Code: APCL
 Case No: SAS No: Service ID: 034406
 Project ID: JPL Project No: 04-4428.10 Sample Matrix: Water
 CCV Data File: G3570Q01 Instrument ID: A
 Batch No: 03G3570

#	Client Sample No	Lab Sample ID	Analysis Date & Time	IS-1		IS-2		IS-3	
				Area #	RT #	Area #	RT #	Area #	RT #
12 Hour CCV STD			08/06/03 10:32	790353	7.65	610095	11.55	311119	13.85
CCV Upper Limit				1580706	8.15	1220190	12.05	622238	14.35
CCV Lower Limit				395176	7.15	305047	11.05	155559	13.35
1	03G3570-LCS-01	03G3570-LCS-01	08/06/03 10:58	829058	7.65	648481	11.54	336853	13.84
2	MW-19-2MS	03-4406-10MS	08/06/03 11:24	818407	7.65	629647	11.54	320116	13.84
3	MW-19-2MSD	03-4406-10MSD	08/06/03 11:50	849231	7.65	658725	11.54	333596	13.84
4	03G3570-MB-01	03G3570-MB-01	08/06/03 13:08	728257	7.65	583860	11.55	305134	13.85
5	MW-19-2	03-4406-10	08/06/03 14:27	724079	7.65	585964	11.55	303150	13.84
6	DUPE-1-3Q03	03-4406-1	08/06/03 17:05	707817	7.65	564353	11.55	296802	13.85
7	MW-3-2	03-4406-3	08/06/03 17:31	715315	7.65	574418	11.55	301910	13.85
8	MW-3-3	03-4406-4	08/06/03 17:57	707644	7.65	566761	11.55	291695	13.84
9	MW-3-4	03-4406-5	08/06/03 18:23	707570	7.65	569834	11.55	295103	13.85
10	MW-17-2	03-4406-6	08/06/03 18:50	701194	7.65	566840	11.55	292706	13.84
11	MW-17-3	03-4406-7	08/06/03 19:16	705668	7.65	569201	11.55	299969	13.85
12	MW-17-4	03-4406-8	08/06/03 19:42	703564	7.65	572925	11.55	296607	13.85
13	MW-19-1	03-4406-9	08/06/03 20:08	697238	7.65	567408	11.55	295466	13.85
14	MW-19-3	03-4406-11	08/06/03 20:34	695949	7.65	561774	11.55	292982	13.85
15	MW-19-4	03-4406-12	08/06/03 21:00	695436	7.65	563333	11.55	297212	13.84
16	MW-19-5	03-4406-13	08/06/03 21:26	698226	7.65	562520	11.55	294748	13.85
17									
18									
19									
20									
21									
22									

IS-1 = FLUOROBENZENE
 IS-2 = CHLOROBENZENE-D5
 IS-3 = 1,4-DICHLOROBENZENE-D4

Area Upper Limit = +100% of CCV internal standard area
 Area Lower Limit = - 50% of CCV internal standard area
 RT Upper Limit = +0.50 minutes of CCV internal standard RT
 RT Lower Limit = - 0.50 minutes of CCV internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits

FORM-8A

Applied P & Ch Laboratory

Internal Standard Area and RT Summary for Method 524.2

Client Name: GEOFON, Inc. Contract No: Lab Code: APCL
 Case No: SAS No: Service ID: 034406
 Project ID: JPL Project No: 04-4428.10 Sample Matrix: Water
 CCV Data File: G3595Q01 Instrument ID: A
 Batch No: 03G3595

#	Client Sample No	Lab Sample ID	Analysis Date & Time	IS-1		IS-2		IS-3	
				Area #	RT #	Area #	RT #	Area #	RT #
12 Hour CCV STD			08/07/03 10:41	766703	7.65	597320	11.54	312953	13.84
CCV Upper Limit				1533406	8.15	1194640	12.04	625906	14.34
CCV Lower Limit				383351	7.15	298660	11.04	156476	13.34
1	03G3595-LCS-01	03G3595-LCS-01	08/07/03 11:06	793930	7.65	618724	11.55	326278	13.84
2	MW-20-4MS	03-4425-6MS	08/07/03 11:32	813408	7.65	623673	11.54	310405	13.84
3	MW-20-4MSD	03-4425-6MSD	08/07/03 11:58	820159	7.65	638486	11.54	325382	13.84
4	03G3595-MB-01	03G3595-MB-01	08/07/03 13:17	719160	7.65	584637	11.55	301712	13.84
5	TB-2-7/30/03	03-4406-14	08/07/03 13:43	708589	7.65	576217	11.54	297747	13.84
6	EB-2-7/30/03	03-4406-2	08/07/03 14:10	723006	7.65	577856	11.55	305816	13.85
7									
8									
9									
10									
11									
12									
13									
14									
15									
16									
17									
18									
19									
20									
21									
22									

IS-1 = FLUOROBENZENE
 IS-2 = CHLOROBENZENE-D5
 IS-3 = 1,4-DICHLOROBENZENE-D4

Area Upper Limit = +100% of CCV internal standard area
 Area Lower Limit = - 50% of CCV internal standard area
 RT Upper Limit = +0.50 minutes of CCV internal standard RT
 RT Lower Limit = - 0.50 minutes of CCV internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits

United P & C Laboratory

100 Magnolia Ave. Chino CA 91710
 (909) 590-1828 Fax: (909) 590-1498

VOC Analysis General Logbook

Sample # 0393595 Batch # 0393595 Matrix: W Date: 08/07/03 Analyst: TCU

IS/Surrogate: GC-1514/1515 Methanol(mark-M): _____ PEG (mark-PEG): _____ Defoaming(mark-DF): _____

Initial Batch: Initial Batch; Middle Batch; Final Batch. Internal Study Datafile Path: _____

Type	Sample ID	Method	V/X=f ₁	V _i /V _i =f ₂	V _{spg} /V _{inj} =f ₃	F	A-#	Datafile	Note	pH
SP	G3595 P01	ES24A	25/25 = 1	/ =	/ =	1		G3595 P01	08/07/03	10:15 am
CCV	Q01		/ =	/ =	/ =			Q01	GC1552	
LCS	L01		/ =	/ =	/ =			L01		
MS	M01		/ =	/ =	/ =			M01	\$4425-06	<2
MSD	N01		/ =	/ =	/ =			N01	↓	↓
MB	↓ K01		/ =	/ =	/ =			↓ K01		
Sample	4406-14		/ =	/ =	/ =			4406-14	tb	<2
	↓ 02		/ =	/ =	/ =			↓ 02	tb	
	4425-08		/ =	/ =	/ =			4425-08	tb	
	↓ 02		/ =	/ =	/ =			↓ 02	tb	
	4455-10		/ =	/ =	/ =			4455-10	tb	
	↓ 02		/ =	/ =	/ =			↓ 02	tb	
	4425-06		/ =	/ =	/ =			4425-06	ms	
	01		/ =	/ =	/ =			01		
	03		/ =	/ =	/ =			03		
	04		/ =	/ =	/ =			04		
	05		/ =	/ =	/ =			05		
	↓ 07		/ =	/ =	/ =			↓ 07		
	4455-01		/ =	/ =	/ =			4455-01		
	03		/ =	/ =	/ =			03		
	05		/ =	/ =	/ =			05		
	06		/ =	/ =	/ =			06		
	07		/ =	/ =	/ =			07		
	08		/ =	/ =	/ =			08		
	↓ 09		/ =	/ =	/ =			↓ 09		
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					

Type	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X(g or mL) = T	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X(g or mL) = T
IS/LCSD	2635	GC-	200 × 1/X = ppb		GC-	200 × 1/X = ppb
IS/MSD	2636/2637	GC-1552	200 × 2.5 / X = 20 ppb		GC-	200 × 2.5 / X = ppb

Footnote/Anomaly:

Applied P & Ch Laboratory

VOC Analysis General Logbook

Magnolia Ave. Chino CA 91710
 (909) 590-1828 Fax: (909) 590-1498

Sample # 0383570 Batch # 0383570 Matrix: W Date: 8/06/03 Analyst: Zou

IS/Surrogate: GC-15114/15115 Methanol(mark-M): _____ PEG (mark-PEG): _____ Defoaming(mark-DF): _____

☐ Ini. Batch Initial Batch; ☐ Middle Batch; ☐ Final Batch. ☐ Internal Study Datafile Path: _____

Type	Sample ID	Method	V/X=f ₁	V ₁ /V _i =f ₂	V _{esp} /V _{inj} =f ₃	F	A-#	Datafile	Note	pH
SP	63570101	ESL4A 202	25/25 = 1	/ =	/ =	1		63570101	8/06/03	10.07 am
CCV	Q01		/ =	/ =	/ =			Q01	GC15524	
LCS	L01		/ =	/ =	/ =			L01		
MS	M01		/ =	/ =	/ =			M01	\$4406-10	<2
MSD	N01		/ =	/ =	/ =			N01	↓	↓
MB	✓ K01		/ =	/ =	/ =			✓ K01		
Sample	4389-07		/ =	/ =	/ =			4389-07	#5	<2
	↓ 01		/ =	/ =	/ =			↓ 01	e _h	
	4406-10		/ =	/ =	/ =			4406-10	ms	
	4389-02		/ =	/ =	/ =			4389-02		
	03		/ =	/ =	/ =			03		
	04		/ =	/ =	/ =			04		
	05		/ =	/ =	/ =			05		
	↓ 06		/ =	/ =	/ =			↓ 06		
	4406-01		/ =	/ =	/ =			4406-01		
	03		/ =	/ =	/ =			03		
	04		/ =	/ =	/ =			04		
	05		/ =	/ =	/ =			05		
	06		/ =	/ =	/ =			06		
	07		/ =	/ =	/ =			07		
	08		/ =	/ =	/ =			08		
	09		/ =	/ =	/ =			09		
	11		/ =	/ =	/ =			11		
	12		/ =	/ =	/ =			12		
↓	↓ 13	↓	↓ / = ↓	/ =	/ =	↓		↓ 13		↓
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					

~~08/11/03~~

Type	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X (g or mL) = T	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X (g or mL) = T
S/LCSD	2603	GC-15522	200 × 2.5 / X = ppb		GC-	x / X = ppb
S/MSD	2604/2605	GC-	x × 2.5 / X = 20 ppb		GC-	x / X = ppb

note/Anomaly:

Applied P & Ch Laboratory

13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

VOC Analysis General Logbook

Sequence # 0393452 Batch # 0393452 Matrix: W Date: 07/24/03 Analyst: Zou
Lot #: IS/Surrogate: GC-1514/1515 Methanol(mark-M): _____ PEG (mark-PEG): _____ Defoaming(mark-DF): _____

Calib. + Ini. Batch Initial Batch; Middle Batch; Final Batch. Internal Study Datafile Path: _____

Op. #	Type	Sample ID	Method	V/X=f ₁	V ₁ /V _i =f ₂	V _{spg} /V _{inj} =f ₃	F'	A-#	Datafile	Note	pH
2313	SP	63452/01	E524A	25/25=1	/ =	/ =	1		63452/01	07/24/03	10:07 am
2314	CCV	01		/ =	/ =	/ =			01	GC15447	
2315	LCS	01		/ =	/ =	/ =			01		
2316	LCS0	01		/ =	/ =	/ =			01		
2317	MB	01		/ =	/ =	/ =			01		
2318	Sample	476-08		/ =	/ =	/ =			476-08	Internal PE	
2319		482-05		/ =	/ =	/ =			482-05	PE	
2320		18		1/5 = 5	/ =	/ =	5		18		
2321		03		1/25 = 1	/ =	/ =	1		03		
2322		03A		1/5 = 5	/ =	/ =	5		03A		
2323		04		1/5 = 5	/ =	/ =	5		04		
2324		04A		1/25 = 1	/ =	/ =	1		04A		
2325		04B		1/5 = 5	/ =	/ =	5		04B		
2326		475-038		1/25 = 1	/ =	/ =	1		475-038		
2327		03C		/ =	/ =	/ =			03C		
2328				/ =	/ =	/ =					
2329				/ =	/ =	/ =					
2330				/ =	/ =	/ =					
2331				/ =	/ =	/ =					
2332				/ =	/ =	/ =					
2333				/ =	/ =	/ =					
2334				/ =	/ =	/ =					
2335				/ =	/ =	/ =					
2336				/ =	/ =	/ =					
2337				/ =	/ =	/ =					
2338				/ =	/ =	/ =					
2339				/ =	/ =	/ =					
2340				/ =	/ =	/ =					
2341				/ =	/ =	/ =					
2342				/ =	/ =	/ =					
2343				/ =	/ =	/ =					
2344				/ =	/ =	/ =					

Se

107/25/01

Type	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X(g or mL) = T	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X(g or mL) = T
LCS/LCS0	2315/036	GC-15448	200 × 2.5 / X = 20 ppb		GC-	x / X = ppb
MS/MSD		GC-	x / X = ppb		GC-	x / X = ppb

Footnote/Anomaly:

Applied P & Ch Laboratory
 13760 Magnolia Ave. Chino CA 91710
 Tel: (909) 590-1828 Fax: (909) 590-1498

VOC Analysis General Logbook

Sequence # 0393415 Batch # 0393415 Matrix: W Date: 07/24/03 Analyst: Zou
 Lot #: IS/Surrogate: GC-1514/1515 Methanol(mark-M): _____ PEG (mark-PEG): _____ Defoaming(mark-DF): _____

Calib. + Ini. Batch Initial Batch; Middle Batch; Final Batch. Internal Study Datafile Path: _____

Op. #	Type	Sample ID	Method	$V/X=f_1$	$V_j/V_i=f_2$	$V_{spj}/V_{inj}=f_3$	F	A-#	Datafile	Note	pH
2217	SP	G3415 P01	E524A	25/25 = 1	/ =	/ =	1		G3415 P01	07/24/03	17:08
2218	Calib.	02-00003		/ =	/ =	/ =			02-00003	2.3 ppb	
2219		0002A		/ =	/ =	/ =			0002A	2 ppb	
2220		00010		/ =	/ =	/ =			00010	10 ppb	
2221		00020		/ =	/ =	/ =			00020	20 ppb	
2222		00040		/ =	/ =	/ =			00040	40 ppb	
2223		00060		/ =	/ =	/ =			00060	60 ppb	
2224				/ =	/ =	/ =					
2225				/ =	/ =	/ =					
2226				/ =	/ =	/ =					
2227				/ =	/ =	/ =					
2228				/ =	/ =	/ =					
2229				/ =	/ =	/ =					
2230				/ =	/ =	/ =					
2231				/ =	/ =	/ =					
2232				/ =	/ =	/ =					
2233				/ =	/ =	/ =					
2234				/ =	/ =	/ =					
2235				/ =	/ =	/ =					
2236				/ =	/ =	/ =					
2237				/ =	/ =	/ =					
2238				/ =	/ =	/ =					
2239				/ =	/ =	/ =					
2240				/ =	/ =	/ =					
2241				/ =	/ =	/ =					
2242				/ =	/ =	/ =					
2243				/ =	/ =	/ =					
2244				/ =	/ =	/ =					
2245				/ =	/ =	/ =					
2246				/ =	/ =	/ =					
2247				/ =	/ =	/ =					
2248				/ =	/ =	/ =					

De
 07/23/03

Type	Op #	STD Lot #	$C_{std}(ng/\mu L) \times V_{std}(\mu L) / X(g \text{ or } mL) = T$	Op #	STD Lot #	$C_{std}(ng/\mu L) \times V_{std}(\mu L) / X(g \text{ or } mL) = T$
LCS/LCSD		GC-	x /X= ppb		GC-	x /X= ppb
MS/MSD		GC-	x /X= ppb		GC-	x /X= ppb

Footnote/Anomaly: