

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659L01.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 11:06 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 11 16:04 2003
 Print Time : Mon Aug 11 16:04 2003
 Miscellaneous :
 Sample : f=1
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

2438

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	777.095	10.00		0.02	
47	Chlorobenzene-d5	11.55	11.54	0.000	117	82	607.472	10.00		0.00	
62	1,4-Dichlorobenze	13.84	13.84	0.000	152	150	315.375	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	379.297	19.59		19.6	97.93%
29	1,2-Di-Cl-Et-d4 (7.10	7.08	0.001	65	102	302.358	19.28		19.3	96.38%
55	toluene-d8	9.91	9.89	0.000	98	100	1462.759	19.11		19.1	95.53%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	497.074	19.06		19.1	95.28%

Target Compounds <<< 11 : ISTD ID = 1 >>> Qvalue

3	di-Cl-di-F-methan	1.89	1.85	0.005	85	87	356.148	18.26		18.3	100	
4	Chloromethane	2.11	2.07	0.005	50	52	270.305	16.33		16.3	99	
2	F114	2.03	2.00	0.005	85	135	207.624	20.35		20.4	52	
5	vinyl chloride	2.23	2.19	0.005	62	64	348.596	18.51		18.5	98	
6	bromomethane	2.61	2.58	0.005	94	96	151.977	16.79		16.8	97	
7	chloroethane	2.73	2.70	0.004	64	66	214.462	19.13		19.1	0	
8	tri-Cl-F-methane	3.03	3.00	0.005	101	103	551.729	19.77		19.8	98	
91	Acetonitrile X10	4.07	4.04	0.004	41	40	629.774	184.73		184.7	89	
9	acrolein X10	3.51	3.48	0.005	56	55	407.177	221.42		221.4	99	
11	acetone X10	3.72	3.69	0.004	43	58	444.300	244.89		244.9	0	
12	ethyl ether X5	3.37	3.34	0.005	59	74	910.910	109.31		109.3	89	
13	11-dichloroethene	3.64	3.60	0.005	61	96	428.751	19.25		19.2	0	
14	Iodomethane	3.82	3.78	0.005	142	127	228.825	11.62		11.6	96	
15	F-113	3.65	3.62	0.005	101	151	338.054	22.12		22.1	88	
16	acrylonitrile X10	4.52	4.49	0.004	53	52	639.255	184.02		184.0	99	
17	carbon disulfide	3.90	3.87	0.004	76	78	1030.530	18.20		18.2	100	
94	Isopropyl Alcohol	4.07	4.01	0.007	45	43	98.899	202.37		202.4	100	
18	methylene chlorid	4.22	4.19	0.004	84	49	387.410	19.67		19.7	99	
19	t-12-di-Cl-ethene	4.58	4.55	0.004	96	61	394.573	18.80		18.8	93	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Handwritten notes:
 m² 08/11/03
 m² 08/11/03
 m² 08/11/03
 m² 08/11/03
 ?
 ?
 ?

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659L01.D Sample : F=1
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Aug 11 11:06 2003 RF via : Multiple Level Calibration
 Method Update: Thu Jul 24 12:40 2003 Operator: zou
 Quant. Time : Aug 11 16:04 2003 Multiplr: 1.000000
 Print Time : Mon Aug 11 16:04 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.59	4.56	0.004	73	57	683.338	19.56	19.6	96	?
95	Tert butyl alcoho	4.50	4.47	0.005	59	57	178.057	194.01	194.0	100	m? 8/11/03
94	allyl chloride	4.07	4.04	0.004	41	76	629.774	20.60	20.6	85	#? 8/11/03
21	11-dichloroethane	5.12	5.09	0.004	63	83	610.991	18.14	18.1	100	
97	propionitrile	6.02	5.99	0.003	54	51	27.129	20.81	20.8	100	#?
22	c-12-di-Cl-ethene	5.92	5.89	0.003	96	61	403.767	19.08	19.1	89	?
23	22-Dichloropropan	5.92	5.89	0.004	77	97	541.706	25.59	25.6	99	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	189.458	18.46	18.5	96	?
25	chloroform	6.37	6.35	0.002	83	85	648.085	20.42	20.4	96	?
26	tetrahydrofuranX5	6.35	6.32	0.003	42	72	233.897	91.65	91.6	98	?
98	Diisopropyl ether	5.24	5.22	0.003	45	87	977.174	18.84	18.8	99	?
99	ETBE	5.74	5.72	0.003	59	87	805.877	20.32	20.3	98	
30	12-dichloroethane	7.22	7.20	0.003	64	62	113.585	20.74	20.7	93	?
32	vinyl acetate X5	5.20	5.17	0.004	43	86	2450.780	101.08	101.1	99	
92	Nitro Methane (X10	5.82	5.80	0.002	61	46	86.283	199.94	199.9	83	
33	2-butanoneMEK X10	5.95	5.92	0.003	43	72	682.535	207.96	208.0	98	?
93	Ethyl Acetate X2	6.04	6.02	0.002	43	61	334.602	39.45	39.5	94	?
34	111-trichloroetha	6.65	6.63	0.002	97	99	610.793	19.51	19.5	99	
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	490.291	20.73	20.7	90	?
36	benzene	7.21	7.19	0.003	78	52	1465.350	18.58	18.6	99	?
37	CCl4	6.91	6.89	0.002	117	119	584.813	20.26	20.3	99	?
100	Isobutyl alcohol	7.17	7.39	-0.030	43	42	65.908	184.01	184.0	94	m 8/11/03
38	thiophene	7.53	7.51	0.002	84	58	761.537	19.42	19.4	98	
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	325.553	18.45	18.4	96	
40	trichloroethene	8.24	8.23	0.002	130	132	474.437	19.30	19.3	100	
41	dibromomethane	8.73	8.71	0.002	174	172	209.617	18.72	18.7	99	
101	TAME	7.41	7.39	0.002	73	43	709.098	20.63	20.6	97	
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	464.302	20.09	20.1	99	
43	Me-methacrylate	8.76	8.75	0.002	69	100	166.971	18.57	18.6	90	
44	2-ClEt-Vl-ether10	9.38	9.37	0.002	63	43	216.998	111.79	111.8	95	
45	c-13-di-Cl-propen	9.56	9.55	0.002	75	110	524.917	20.47	20.5	93	
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	435.424	21.35	21.3	92	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659L01.D Sample : f=1
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Aug 11 11:06 2003 RF via : Multiple Level Calibration
 Method Update: Thu Jul 24 12:40 2003 Operator: zou
 Quant. Time : Aug 11 16:04 2003 Multiplr: 1.000000
 Print Time : Mon Aug 11 16:04 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<< I2	: ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	267.549	18.25	18.2	94	
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	427.293	18.48	18.5	100	?
50	Et methacrylate	10.37	10.37	0.000	69	99	352.024	19.14	19.1	93	
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	362.023	19.66	19.7	100	
52	bromoform	12.42	12.41	0.000	173	174	212.687	18.90	18.9	99	
53	1,4-dichlorobutan	12.67	12.67	0.000	55	41	388.280	18.04	18.0	95	
54	MIBK	9.77	9.76	0.000	43	58	166.327	18.80	18.8	92	
56	toluene	9.99	9.98	0.000	91	92	1678.622	19.08	19.1	99	
57	2-hexanone X5	10.76	10.75	0.000	43	58	567.957	94.58	94.6	96	
58	12-dibromoethane	11.03	11.03	0.000	107	109	266.465	18.42	18.4	98	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	504.967	19.21	19.2	99	?
60	chlorobenzene	11.58	11.57	0.000	112	77	1164.107	20.34	20.3	90	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	415.632	18.56	18.6	99	
<<< I3	: ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	219.220	21.31	21.3	100	?
64	Et-Bz	11.70	11.69	0.000	91	106	1942.228	19.68	19.7	95	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	3008.056	40.32	40.3	96	
66	styrene	12.24	12.23	0.000	104	78	1164.005	20.52	20.5	93	?
67	o-xylene	12.22	12.22	0.000	91	106	1551.616	20.73	20.7	97	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	291.303	17.48	17.5	99	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	89.710	17.99	18.0	99	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	2059.639	21.48	21.5	98	
72	bromobenzene	12.89	12.89	0.000	156	158	501.933	19.49	19.5	100	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	56.423	19.86	19.9	84	#?
73	n-propylbenzene	13.00	12.99	0.000	120	78	633.141	21.08	21.1	92	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	541.400	20.40	20.4	99	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	535.765	19.85	19.8	98	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	1762.816	21.36	21.4	97	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	1999.886	21.69	21.7	98	?
78	124-tri-Me-Benzen	13.52	13.51	0.000	105	120	1788.620	20.81	20.8	96	
79	tert-butylbenzene	13.47	13.47	0.000	119	91	1655.117	21.00	21.0	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659L01.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 11:06 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 11 16:04 2003
 Print Time : Mon Aug 11 16:04 2003
 Miscellaneous :


Sample : f=1
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	1040.756	19.22	19.2	99	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	2397.356	21.61	21.6	97	
82	14-DCB	13.87	13.87	0.000	146	148	1029.724	20.00	20.0	99	
83	Cl-benzyl	13.98	13.98	0.000	126	91	141.639	27.15	27.2	77	#
84	12-DCB	14.21	14.21	0.000	146	148	915.452	18.99	19.0	99	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	556.698	20.59	20.6	82	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	63.838	18.81	18.8	98	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	678.822	21.75	21.7	99	
88	naphthalene	15.79	15.78	0.000	128	129	1086.794	19.19	19.2	99	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	385.427	20.39	20.4	99	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	580.574	20.73	20.7	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659L01.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 11:06 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 11 16:04 2003
 Print Time : Mon Aug 11 16:04 2003
 Miscellaneous :

Sample : F=1
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000



FORM-3A

Applied P & Ch Laboratory

Matrix Spike/Matrix Spike Duplicate Recovery for Method 524.2

Client Name: GEOFON, Inc.	Contract No:	Lab Code: APCL
Case No:	SAS No:	Service ID: 34572
Project ID: JPL	Project No: 04-4428.10	Sample Matrix: Water
	Batch No: 03G3659	
MS Filename: G3659M02	Date Analyzed: 081103	Time Analyzed: 21:14
MSD Filename: G3659N02	Date Analyzed: 081103	Time Analyzed: 21:40
MS Sample No: MW-12-4	Sample Lab ID: 03-4572-6	

Spiked Components	Unit	Spike Added	Concentration		MS Rec% #	QC Limit, % REC
			Unspiked	MS		
BENZENE	µg/L	20	0	19.0	95	65-121
CHLOROBENZENE	µg/L	20	0	20.9	105	65-134
1,1-DICHLOROETHENE	µg/L	20	0	19.2	96	65-127
TOLUENE	µg/L	20	0	19.5	98	65-134
TRICHLOROETHENE	µg/L	20	0.4	20.3	100	65-125
# of Out-of-control					0	

Spiked Components	Unit	Spike Added	MSD Concentration	MSD Rec% #	RPD% #	QC Limit, %	
						RPD	REC
BENZENE	µg/L	20	18.8	94	1	28	65-121
CHLOROBENZENE	µg/L	20	20.6	103	2	35	65-134
1,1-DICHLOROETHENE	µg/L	20	19.2	96	0	31	65-127
TOLUENE	µg/L	20	19.3	97	1	35	65-134
TRICHLOROETHENE	µg/L	20	20.2	99	1	30	65-125
# of Out-of-control					0	0	

Column to be used to flag recovery and RPD values:

* - Values outside of contract required QC Limits D - Spiked components diluted out

Comments: _____

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659M02.D Sample : F=1 \$4572-06
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Aug 11 21:14 2003 RF via : Multiple Level Calibration
 Method Update: Thu Jul 24 12:40 2003 Operator: zou
 Quant. Time : Aug 12 10:04 2003 Multiplr: 1.000000
 Print Time : Tue Aug 12 10:05 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.64	7.63	0.002	96	70	757.966	10.00		0.02	
47	Chlorobenzene-d5	11.54	11.54	0.000	117	82	592.654	10.00		0.00	
62	1,4-Dichlorobenzene	13.84	13.84	0.000	152	150	310.822	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	383.289	20.29		20.3	101.46%
29	1,2-Di-Cl-Et-d4	7.11	7.08	0.002	65	102	315.157	20.60		20.6	103.00%
55	toluene-d8	9.91	9.89	0.000	98	100	1454.410	19.47		19.5	97.36%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	500.472	19.47		19.5	97.34%

Target Compounds	<<< I1 : ISTD ID = 1 >>>	Qvalue
3 di-Cl-di-F-methan	1.89 1.85 0.005 85 87 350.155 18.40 18.4 99	Dr
4 Chloromethane	2.11 2.07 0.006 50 52 244.797 15.10 15.1 98	Dr
2 F114	2.03 2.00 0.005 85 135 200.090 20.11 20.1 50	Dr 8/12/03
5 vinyl chloride	2.22 2.19 0.004 62 64 326.537 17.77 17.8 98	
6 bromomethane	2.61 2.58 0.004 94 96 114.494 12.97 13.0 99	
7 chloroethane	2.73 2.70 0.004 64 66 209.753 19.18 19.2 0	Dr
8 tri-Cl-F-methane	3.03 3.00 0.005 101 103 545.857 20.06 20.1 99	
91 Acetonitrile X10	4.07 4.04 0.004 41 40 601.155 180.78 180.8 95	?
9 acrolein X10	3.52 3.48 0.006 56 55 326.627 179.66 179.7 99	
11 acetone X10	3.73 3.69 0.006 43 58 426.832 240.75 240.7 0	Dr
12 ethyl ether X5	3.37 3.34 0.005 59 74 920.545 113.57 113.6 91	Dr 8/12/03
13 1,1-dichloroethene	3.64 3.60 0.005 61 96 417.599 19.22 19.2 0	
14 Iodomethane	3.82 3.78 0.005 142 127 230.863 12.02 12.0 96	Dr 8/12/03
15 F-113	3.65 3.62 0.005 101 151 327.277 21.94 21.9 87	?
16 acrylonitrile X10	4.53 4.49 0.005 53 52 669.263 197.52 197.5 98	
17 carbon disulfide	3.90 3.87 0.004 76 78 993.926 17.99 18.0 99	
94 Isopropyl Alcohol	4.13 4.01 0.015 45 43 85.742 180.35 180.4 100	Dr 8/12/03
18 methylene chlorid	4.22 4.19 0.005 84 49 371.344 19.31 19.3 99	
19 t-12-di-Cl-ethene	4.58 4.55 0.004 96 61 387.189 18.91 18.9 93	?

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659M02.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 21:14 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 12 10:04 2003
 Print Time : Tue Aug 12 10:05 2003
 Miscellaneous :

Sample : F=1 \$4572-06
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.59	4.56	0.004	73	57	697.058	20.45	20.5	96	?
95	Tert butyl alcoho	4.57	4.47	0.014	59	57	176.982	197.24	197.2	100	m? 08/12/03
94	allyl chloride	4.07	4.04	0.004	41	76	601.155	20.16	20.2	85	#? 08/12/03
21	11-dichloroethane	5.12	5.09	0.004	63	83	606.403	18.46	18.5	100	
97	propionitrile	6.03	5.99	0.004	54	51	25.156	19.78	19.8	100	#?
22	c-12-di-Cl-ethene	5.92	5.89	0.003	96	61	404.205	19.58	19.6	90	?
23	22-Dichloropropan	5.92	5.89	0.004	77	97	457.473	22.03	22.0	96	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	192.107	19.19	19.2	98	?
25	chloroform	6.37	6.35	0.003	83	85	671.170	21.70	21.7	97	?
26	tetrahydrofuranX5	6.35	6.32	0.004	42	72	251.028	100.84	100.8	96	?
98	Disopropyl ether	5.24	5.22	0.003	45	87	975.703	19.29	19.3	99	?
99	ETBE	5.74	5.72	0.003	59	87	798.548	20.65	20.6	97	?
30	12-dichloroethane	7.22	7.20	0.003	64	62	118.284	22.16	22.2	99	?
32	vinyl acetate X5	5.20	5.17	0.004	43	86	2194.926	92.81	92.8	98	?
92	Nitro Methane(X10	5.84	5.80	0.005	61	46	87.305	207.41	207.4	81	m? 08/12/03
33	2-butanoneMEK X10	5.95	5.92	0.004	43	72	687.231	214.95	214.9	97	?
93	Ethyl Acetate x2	6.04	6.02	0.003	43	61	344.135	41.61	41.6	95	?
34	111-trichloroetha	6.65	6.63	0.002	97	99	604.068	19.78	19.8	100	?
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	477.995	20.72	20.7	91	?
36	benzene	7.21	7.19	0.003	78	52	1463.223	19.02	19.0	99	?
37	CCl4	6.91	6.89	0.003	117	119	617.169	21.92	21.9	99	?
100	Isobutyl alcohol	7.19	7.39	-0.026	43	42	66.665	190.46	190.5	96	m? 08/12/03
38	thiophene	7.53	7.51	0.002	84	58	774.967	20.26	20.3	99	?
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	329.535	19.14	19.1	98	?
40	trichloroethene	8.24	8.23	0.002	130	132	477.309	19.90	19.9	100	?
41	dibromomethane	8.73	8.71	0.002	174	172	214.948	19.69	19.7	98	?
101	TAME	7.41	7.39	0.002	73	43	716.403	21.36	21.4	97	?
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	475.400	21.09	21.1	100	?
43	Me-methacrylate	8.76	8.75	0.002	69	100	177.232	20.12	20.1	96	?
44	2-ClEt-Vl-ether10	9.56	9.37	0.025	63	43	2.074	16.35	16.4	28	#?
45	c-13-di-Cl-propen	9.56	9.55	0.002	75	110	498.131	19.91	19.9	92	?
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	431.459	21.69	21.7	92	?

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659M02.D Sample : f=1 \$4572-06
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Aug 11 21:14 2003 RF via : Multiple Level Calibration
 Method Update: Thu Jul 24 12:40 2003 Operator: zou
 Quant. Time : Aug 12 10:04 2003 Multiplr: 1.000000
 Print Time : Tue Aug 12 10:05 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<< I2	: ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	279.757	19.56	19.6	96	
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	443.526	19.66	19.7	99	
50	Et methacrylate	10.38	10.37	0.000	69	99	370.308	20.55	20.6	93	?
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	372.391	20.71	20.7	99	
52	bromoform	12.42	12.41	0.000	173	174	219.390	19.98	20.0	100	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	407.980	19.43	19.4	94	
54	MIBK	9.77	9.76	0.000	43	58	176.000	20.39	20.4	96	
56	toluene	9.99	9.98	0.000	91	92	1671.766	19.48	19.5	99	
57	2-hexanone X5	10.76	10.75	0.000	43	58	608.718	103.91	103.9	93	
58	12-dibromoethane	11.03	11.03	0.000	107	109	277.347	19.66	19.7	98	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	499.307	19.47	19.5	98	
60	chlorobenzene	11.58	11.57	0.000	112	77	1168.115	20.93	20.9	90	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	427.365	19.56	19.6	98	
<<< I3	: ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	211.547	20.87	20.9	96	?
64	Et-Bz	11.70	11.69	0.000	91	106	1927.271	19.82	19.8	96	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	2985.338	40.60	40.6	97	
66	styrene	12.24	12.23	0.000	104	78	1161.803	20.78	20.8	93	?
67	o-xylene	12.22	12.22	0.000	91	106	1556.774	21.10	21.1	96	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	312.054	19.00	19.0	99	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	96.764	19.69	19.7	100	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	2047.226	21.66	21.7	99	
72	bromobenzene	12.89	12.89	0.000	156	158	507.646	20.00	20.0	99	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	56.559	20.17	20.2	86	?
73	n-propylbenzene	13.00	12.99	0.000	120	78	626.904	21.18	21.2	91	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	538.053	20.57	20.6	99	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	538.883	20.25	20.3	98	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	1754.332	21.56	21.6	97	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	1984.113	21.84	21.8	100	?
78	124-tri-Me-Benzen	13.51	13.51	0.000	105	120	1782.477	21.04	21.0	96	
79	tert-butylbenzene	13.47	13.47	0.000	119	91	1645.429	21.18	21.2	95	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06


Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659M02.D Sample : f=1 \$4572-06
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Aug 11 21:14 2003 RF via : Multiple Level Calibration
 Method Update: Thu Jul 24 12:40 2003 Operator: zou
 Quant. Time : Aug 12 10:04 2003 Multiplr: 1.000000
 Print Time : Tue Aug 12 10:05 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	QIon	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	1053.180	19.74	19.7	99	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	2387.653	21.84	21.8	97	
82	14-DCB	13.87	13.87	0.000	146	148	1047.537	20.65	20.6	98	
83	Cl-benzyl	13.98	13.98	0.000	126	91	115.597	22.48	22.5	79	#
84	12-DCB	14.21	14.21	0.000	146	148	938.736	19.76	19.8	99	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	550.969	20.68	20.7	84	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	69.402	20.74	20.7	97	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	683.604	22.22	22.2	100	
88	naphthalene	15.78	15.78	0.000	128	129	1117.935	20.02	20.0	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	384.951	20.67	20.7	97	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	600.425	21.75	21.8	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659M02.D
Method : C:\MSDCHEM\1\METHODS\E524A002.M
Acq. Time : Aug 11 21:14 2003
Method Update: Thu Jul 24 12:40 2003
Quant. Time : Aug 12 10:04 2003
Print Time : Tue Aug 12 10:05 2003
Miscellaneous :

Sample : F=1 \$4572-06
Inst. : GCMS-A
RF via : Multiple Level Calibration
Operator: zou
Multiplier: 1.000000



Quantitation Report: **Applied P & Ch Lab** EPA 524.2

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659N02.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 21:40 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 12 10:07 2003
 Print Time : Tue Aug 12 10:07 2003
 Miscellaneous :
 Sample : F=1 \$4572-06
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zcu
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	QIon	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	801.156	10.00		0.02	
47	Chlorobenzene-d5	11.54	11.54	0.000	117	82	622.097	10.00		0.00	
62	1,4-Dichlorobenze	13.84	13.84	0.000	152	150	325.477	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	394.696	19.77		19.8	98.84%
29	1,2-Di-Cl-Et-d4 (7.11	7.08	0.002	65	102	318.828	19.72		19.7	98.58%
55	toluene-d8	9.91	9.89	0.000	98	100	1519.123	19.38		19.4	96.88%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	515.683	19.16		19.2	95.78%

Target Compounds											
<<< I1 : ISFD ID = 1 >>>											
3	di-Cl-di-F-methan	1.89	1.85	0.005	85	87	365.151	18.16		18.2	99
4	Chloromethane	2.11	2.07	0.006	50	52	272.131	15.92		15.9	98
2	F114	2.03	2.00	0.005	85	135	208.514	19.83		19.8	54
5	vinyl chloride	2.22	2.19	0.004	62	64	355.993	18.33		18.3	99
6	bromomethane	2.61	2.58	0.004	94	96	148.081	15.87		15.9	99
7	chloroethane	2.73	2.70	0.004	64	66	224.060	19.38		19.4	0
8	tri-Cl-F-methane	3.03	3.00	0.005	101	103	564.113	19.61		19.6	98
91	Acetonitrile X10	4.07	4.04	0.004	41	40	626.902	178.36		178.4	93
9	acrolein X10	3.52	3.48	0.006	56	55	330.736	171.54		171.5	0
11	acetone X10	3.73	3.69	0.006	43	58	415.589	219.41		219.4	0
12	ethyl ether X5	3.37	3.34	0.005	59	74	947.834	110.41		110.4	90
13	11-dichloroethene	3.64	3.60	0.005	61	96	439.912	19.15		19.2	0
14	Iodomethane	3.82	3.78	0.005	142	127	326.944	16.10		16.1	95
15	F-113	3.65	3.62	0.005	101	151	346.011	21.95		22.0	87
16	acrylonitrile X10	4.53	4.49	0.005	53	52	666.376	186.06		186.1	97
17	carbon disulfide	3.90	3.87	0.004	76	78	1054.129	18.05		18.1	99
94	Isopropyl Alcohol	4.13	4.01	0.015	45	43	97.082	192.89		192.9	100
18	methylene chlorid	4.22	4.19	0.004	84	49	384.194	18.87		18.9	100
19	t-12-di-Cl-ethene	4.58	4.55	0.004	96	61	403.583	18.65		18.6	94

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

m
08/12/03
?
08/12/03
?
08/12/03
?
08/12/03
?
08/12/03
?
08/12/03
?
08/12/03

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659N02.D Sample : F=1 \$4572-06
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Aug 11 21:40 2003 RF via : Multiple Level Calibration
 Method Update: Thu Jul 24 12:40 2003 Operator: zou
 Quant. Time : Aug 12 10:07 2003 Multiplr: 1.000000
 Print Time : Tue Aug 12 10:07 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.59	4.56	0.004	73	57	712.925	19.79	19.8	95	?
95	Tert butyl alcoho	4.57	4.47	0.013	59	57	187.133	197.30	197.3	100	m? 08/12/03
94	allyl chloride	4.07	4.04	0.004	41	76	626.902	19.89	19.9	85	#?
21	11-dichloroethane	5.12	5.09	0.004	63	83	632.274	18.21	18.2	97	#?
97	propionitrile	6.03	5.99	0.005	54	51	27.116	20.18	20.2	100	#?
22	c-12-di-Cl-ethene	5.92	5.89	0.003	96	61	423.171	19.39	19.4	87	?
23	22-Dichloropropan	5.92	5.89	0.003	77	97	473.193	21.54	21.5	96	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	198.133	18.73	18.7	99	?
25	chloroform	6.37	6.35	0.003	83	85	696.146	21.29	21.3	98	?
26	tetrahydrofuranX5	6.35	6.32	0.003	42	72	249.115	94.68	94.7	94	?
98	Disopropyl ether	5.24	5.22	0.003	45	87	1014.049	18.96	19.0	99	?
99	ETBE	5.74	5.72	0.003	59	87	840.667	20.57	20.6	98	?
30	12-dichloroethane	7.22	7.20	0.003	64	62	118.629	21.01	21.0	100	?
32	vinyl acetate X5	5.20	5.17	0.004	43	86	2238.714	89.56	89.6	99	?
92	Nitro Methane (X10	5.84	5.80	0.005	61	46	80.415	180.75	180.7	87	?
33	2-butanoneMEK X10	5.95	5.92	0.003	43	72	677.907	200.04	200.0	97	#?
93	Ethyl Acetate X2	6.04	6.02	0.002	43	61	332.284	38.00	38.0	93	#?
34	111-trichloroetha	6.65	6.63	0.002	97	99	633.234	19.62	19.6	100	?
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	509.044	20.87	20.9	89	?
36	benzene	7.21	7.19	0.003	78	52	1524.883	18.75	18.8	99	?
37	CCl4	6.91	6.89	0.003	117	119	641.900	21.57	21.6	98	?
100	Isobutyl alcohol	7.19	7.39	-0.027	43	42	69.608	188.27	188.3	96	m? 08/12/03
38	thiophene	7.53	7.51	0.003	84	58	799.396	19.78	19.8	96	?
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	343.625	18.89	18.9	98	?
40	trichloroethene	8.25	8.23	0.002	130	132	502.444	19.82	19.8	100	?
41	dibromomethane	8.73	8.71	0.002	174	172	220.098	19.07	19.1	99	?
101	TAME	7.41	7.39	0.002	73	43	742.897	20.96	21.0	97	?
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	488.881	20.52	20.5	99	?
43	Me-methacrylate	8.76	8.75	0.002	69	100	177.786	19.14	19.1	96	?
44	2-ClEt-Vi-ether10	9.57	9.37	0.026	63	43	2.369	16.43	16.4	28	#?
45	c-13-di-Cl-propen	9.56	9.55	0.002	75	110	524.873	19.85	19.8	95	?
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	439.795	20.92	20.9	92	?

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659N02.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 21:40 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 12 10:07 2003
 Print Time : Tue Aug 12 10:07 2003
 Miscellaneous :

Sample : F=1 \$4572-06
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<< 12	: ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	281.920	18.78	18.8	94	
49	13-di-Cl-propane	10.65	10.64	0.001	76	78	442.946	18.70	18.7	100	
50	Et methacrylate	10.38	10.37	0.000	69	99	381.648	20.20	20.2	92	?
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	381.275	20.21	20.2	100	
52	bromoform	12.42	12.41	0.000	173	174	220.685	19.15	19.1	100	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	403.874	18.32	18.3	94	
54	MIBK	9.77	9.76	0.000	43	58	178.982	19.75	19.8	94	
56	toluene	9.99	9.98	0.000	91	92	1740.654	19.32	19.3	99	
57	2-hexanone X5	10.76	10.75	0.000	43	58	596.144	96.94	96.9	94	
58	12-dibromoethane	11.03	11.03	0.000	107	109	279.197	18.85	18.9	97	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	524.165	19.47	19.5	99	?
60	chlorobenzene	11.57	11.57	0.000	112	77	1209.200	20.64	20.6	92	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	433.345	18.90	18.9	100	
<<< 13	: ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	221.610	20.88	20.9	95	?
64	Et-Bz	11.70	11.69	0.000	91	106	2007.399	19.71	19.7	96	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	3088.341	40.11	40.1	96	
66	styrene	12.24	12.23	0.000	104	78	1189.187	20.32	20.3	92	?
67	o-xylene	12.22	12.22	0.000	91	106	1599.438	20.71	20.7	96	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	307.272	17.87	17.9	99	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	95.689	18.59	18.6	99	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	2135.113	21.58	21.6	98	
72	bromobenzene	12.89	12.89	0.000	156	158	522.288	19.65	19.7	100	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	54.487	18.66	18.7	86	?
73	n-propylbenzene	13.00	12.99	0.000	120	78	650.269	20.98	21.0	93	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	555.976	20.30	20.3	98	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	553.820	19.88	19.9	98	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	1805.191	21.19	21.2	97	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	2050.715	21.56	21.6	99	?
78	124-tri-Me-Benzen	13.51	13.51	0.000	105	120	1839.476	20.73	20.7	97	
79	tert-butylbenzene	13.47	13.47	0.000	119	91	1709.894	21.02	21.0	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659N02.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 21:40 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 12 10:07 2003
 Print Time : Tue Aug 12 10:07 2003
 Miscelaneous :


Sample : F=1 \$4572-06
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	QIon	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	1084.028	19.40	19.4	99	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	2471.474	21.59	21.6	97	
82	14-DCB	13.87	13.87	0.000	146	148	1071.198	20.16	20.2	99	
83	Cl-benzyl	13.98	13.98	0.000	126	91	110.776	20.58	20.6	77	#
84	12-DCB	14.21	14.21	0.000	146	148	956.332	19.22	19.2	99	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	565.560	20.27	20.3	86	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	67.355	19.23	19.2	99	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	703.241	21.83	21.8	100	
88	naphthalene	15.79	15.78	0.000	128	129	1179.640	20.18	20.2	99	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	395.971	20.30	20.3	97	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	602.920	20.86	20.9	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659N02.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 21:40 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 12 10:07 2003
 Print Time : Tue Aug 12 10:07 2003
 Miscellaneous :

Sample : F=1 \$4572-06
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000



Data Filename: C:\MSDCHEM\1\DATA\03G3659\4572-06.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 18:36 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 12 09:53 2003
 Print Time : Tue Aug 12 10:15 2003
 Miscellaneous :

Sample : f=1 ms
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000


ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	645.377	10.00		0.02	
47	Chlorobenzene-d5	11.55	11.54	0.000	117	82	554.567	10.00		0.00	
62	1,4-Dichlorobenze	13.85	13.84	0.000	152	150	294.159	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	356.891	22.19		22.2	110.95%
29	1,2-Di-Cl-Et-d4 (7.11	7.08	0.002	65	102	295.758	22.70		22.7	113.52%
55	toluene-d8	9.91	9.89	0.000	98	100	1351.905	19.34		19.3	96.72%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	468.279	19.25		19.2	96.23%

Target Compounds											
<<<	I1 : ISTD ID = 1	>>>									
94	Isopropyl Alcohol	3.85	4.01	-0.021	45	43	0.220	4.81		4.8	100
25	chloroform	6.38	6.35	0.003	83	85	21.764	0.57		0.6	94
92	Nitro Methane(x10	5.89	5.80	0.012	61	46	0.335	0.93		0.9	16
37	CCl4	6.91	6.89	0.003	117	119	38.423	1.60		1.6	99
40	trichloroethene	8.24	8.23	0.002	130	132	7.318	0.36		0.4	94
<<<	I2 : ISTD ID = 47	>>>									
54	MIBK	9.91	9.76	0.013	43	58	4.038	0.50		0.5	1

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\4572-06.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 18:36 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 12 09:53 2003
 Print Time : Tue Aug 12 10:15 2003
 Miscellaneous :
 Sample : f=1 ms
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000



FORM-4A

Applied P & Ch Laboratory

Method Blank Summary for Method 524.2

Client Name: GEOFON, Inc.	Contract No:	Lab Code: APCL
Case No:	SAS No:	Service ID: 34572
Project ID: JPL	Project No: 04-4428.10	Analysis Date: 08/11/03
Sample ID: 03G3659-MB-01	Sample Matrix: Water	Analysis Time: 13:18
Lab Sample ID: 03G3659-MB-01	Batch No: 03G3659	Instrument ID: GC/MS: A
	Data File Name: G3659K01	GC Column: HP-VOC
	Heated Purge: (Y/N) N	Column ID: 0.20 mm

This Method Blank applies to the following samples and QC samples:


#	Client Sample No	Lab Sample ID	Sample Type	Data Filename	Analysis Date	Analysis Time
1	03G3659-LCS-01	03G3659-LCS-01	Lab Control Spike	G3659L01	08/11/03	11:06
2	DUPE-5-3-Q03	03-4572-1	Field Sample	4572-01	08/11/03	16:24
3	EB-7-8-11-03	03-4572-2	Field Sample	4572-02	08/11/03	16:50
4	MW-12-1	03-4572-3	Field Sample	4572-03	08/11/03	17:17
5	MW-12-2	03-4572-4	Field Sample	4572-04	08/11/03	17:43
6	MW-12-3	03-4572-5	Field Sample	4572-05	08/11/03	18:10
7	MW-12-4	03-4572-6	Field Sample	4572-06	08/11/03	18:36
8	MW-12-5	03-4572-7	Field Sample	4572-07	08/11/03	19:03
9	MW-22-1	03-4572-8	Field Sample	4572-08	08/11/03	19:29
10	MW-22-2	03-4572-9	Field Sample	4572-09	08/11/03	19:55
11	MW-22-3	03-4572-10	Field Sample	4572-10	08/11/03	20:21
12	TB-7-8-11-03	03-4572-11	Field Sample	4572-11	08/11/03	20:47
13	MW-12-4MS	03-4572-6MS	Matrix Spike	G3659M02	08/11/03	21:14
14	MW-12-4MSD	03-4572-6MSD	Matrix Spike Duplicate	G3659N02	08/11/03	21:40
15						
16						
17						
18						
19						
20						
21						
22						
23						
24						
25						

Data File : C:\MSDCHEM\1\DATA\03G3415\G3415P01.D

Vial: 2
Operator: zou
Inst : GCMS-A
Multiplr: 1.00

Acq On : 22 Jul 2003 5:08 pm
Sample : #03G3415, w 50 ng
Misc :

MS Integration Params: Lscint.p
Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
Title : **Applied P & Ch Lab** EPA 524.2



Spectrum Information: Average of 12.729 to 12.741 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result
50	95	15	40	15.3	3208	PASS
75	95	30	60	44.1	9259	PASS
95	95	100	100	100.0	20984	PASS
96	95	5	9	7.0	1467	PASS
173	174	0.00	2	0.7	120	PASS
174	95	50	100	81.3	17069	PASS
175	174	5	9	7.1	1218	PASS
176	174	95	101	98.6	16830	PASS
177	176	5	9	7.6	1277	PASS

5A
VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name : APPLIED P & CH LAB Contract: _____
 Lab Code: _____ Case No.: _____ SAS No.: _____ SDG No.: 034572
 Lab File ID: G 3415 P01 BFB Injection Date: 7/22/03
 Instrument ID: GCMS-A BFB Injection Time: 1708
 GC Column: HP-VOC ID: 0.20 (mm) Heated Purge: (Y/N) N

m/e	ION ABUNDANCE CRITERIA	%RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.3
75	30.0 - 60.0% of mass 95	44.1
95	Base peak, 100% relative abundance	100.0
96	5.0 - 9.0% of mass 95	7.0
173	Less than 2.0% of mass 174	0.6 (0.7)1
174	50.0 - 100.0% of mass 95	81.3
175	5.0 - 9.0% of mass 174	5.8 (7.1)1
176	95.0 - 101.0% of mass 174	80.2 (98.6)1
177	5.0 - 9.0% of mass 176	6.1 (7.6)2

1-Value is % mass 174

2-Value is % mass 176

This check applies to the following SAMPLES, MS, MSD, BLANKS and STANDARDS:

	EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
01	VSTD0003	20-0003	20-00003.D	7/22/03	1925
02	VSTD0002	20-0002	20-00002.D	7/22/03	2111
03	VSTD010	20-0010	20-0010.D	7/22/03	2137
04	VSTD020	20-0020	20-0020.D	7/22/03	2256
05	VSTD040	20-0040	20-0040.D	7/23/03	0043
06	VSTD060	20-0060	20-0060.D	7/23/03	0245
07					
08					
09					
10					
11					
12					
13					
14					
15					
16					
17					
18					
19					
20					
21					
22					

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Initial Calibration

Calibration Files
 .3 =2-00003.D 2 =2-0002A.D 10 =2-00010.D
 20 =2-00020.D 40 =2-00040.D 60 =2-00060.D

Compound	.3	2	10	20	40	60	Avg	%RSD	
1) I	1 Fluorobenzene	0.279	0.253	0.228	0.251	0.250	0.246	0.251	6.61
2)	3 di-Cl-di-F-m	0.300	0.265	0.208	0.221	0.218	0.200	0.235	16.51
3) P	4 Chloromethan	0.116	0.138	0.126	0.136	0.137	0.133	0.131	6.45
4)	2 F114	0.268	0.267	0.219	0.239	0.240	0.223	0.242	8.61
5) C	5 vinyl chlori	0.143	0.128	0.105	0.113	0.104	0.106	0.116	13.63
6)	6 bromomethane	0.162	0.171	0.136	0.131	0.138	0.129	0.144	12.08
7)	7 chloroethane	0.426	0.367	0.310	0.357	0.356	0.338	0.359	10.76
8)	8 tri-Cl-F-met	0.034	0.034	0.026	0.025	0.024	0.022	0.028	18.73
9)	91 Acetonitrile	0.106	0.053	0.028	0.024	0.024	0.021	0.043	77.29
10)	9 acrolein	0.162	0.157	0.117	0.117	0.110	0.098	0.127	20.67
11)	11 acetone X	0.345	0.312	0.258	0.277	0.275	0.253	0.287	12.31#
12)	12 ethyl ether	0.266	0.286	0.244	0.262	0.258	0.205	0.253	10.75
13)	M, Cl3 11-dichloroe	0.313	0.246	0.193	0.211	0.205	0.185	0.225	21.21
14)	14 Iodomethane	0.053	0.052	0.042	0.041	0.040	0.040	0.045	13.61
15)	15 F-113	0.838	0.780	0.654	0.729	0.722	0.649	0.729	10.00
16)	16 acrylonitril	0.015	0.008	0.005	0.006	0.007	0.006	0.008	46.15
17)	17 carbon disul	1.323	0.482	0.257	0.251	0.242	0.244	0.467	92.15
18)	94 Isopropyl Al	0.297	0.305	0.248	0.262	0.256	0.253	0.270	8.99
19)	18 methylene ch	0.474	0.509	0.419	0.427	0.427	0.442	0.450	7.81
20)	19 t-12-di-Cl-e	0.014	0.010	0.010	0.010	0.012	0.013	0.012	16.51
21)	20 t-Bu-Me-ethe	0.549	0.412	0.417	0.307	0.282	0.393	0.393	26.97
22)	95 Tert butyl a	0.518	0.484	0.380	0.407	0.403	0.409	0.433	12.52
23)	94 allyl chlori	0.020	0.014	0.014	0.016	0.016	0.017	0.017	14.53
24)	P 21 11-dichloroe	0.296	0.311	0.252	0.263	0.257	0.255	0.272	9.12
25)	97 propionitril	0.392	0.360	0.280	0.290	0.277	0.262	0.310	16.98
26)	22 c-12-di-Cl-e	0.158	0.150	0.117	0.123	0.122	0.122	0.132	13.12
27)	23 22-Dichlorop								
28)	24 Br-Cl-methan								

(#) = Out of Range
 E524A002.M

Thu Jul 24 12:41:07 2003

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Initial Calibration

Calibration Files
 .3 =2-00003.D 2 =2-0002A.D 10 =2-00010.D
 20 =2-00020.D 40 =2-00040.D 60 =2-00060.D

Compound	.3	2	10	20	40	60	Avg	%RSD
29) C 25 Chloroform	0.724	0.528	0.396	0.413	0.403	0.406	0.478	27.21#
30) 26 tetrahydrofu	0.036	0.035	0.031	0.031	0.032	0.032	0.033	6.66
31) 98 Disopropyl	0.648	0.792	0.634	0.653	0.636	0.641	0.667	9.23
32) S 27 Di-Br-F-Me (0.294	0.230	0.240	0.239	0.243	0.243	0.249	10.27
33) 99 ETBE	0.437	0.534	0.485	0.524	0.532	0.549	0.510	8.20
34) S 29 1,2-Di-Cl-Et	0.248	0.186	0.191	0.191	0.191	0.192	0.202	12.96
35) 30 12-dichloroe	0.094	0.093	0.070	0.070	0.070	0.070	0.078	15.50
36) 32 vinyl acetat	0.319	0.377	0.321	0.291	0.285	0.280	0.312	11.53
37) 92 Nitro Methan	0.006	0.005	0.005	0.005	0.006	0.006	0.006	9.09
38) 33 2-butanoneME	0.081	0.061	0.044	0.043	0.042	0.041	0.052	30.99
39) 93 Ethyl Acetat	0.175	0.135	0.110	0.107	0.107	0.110	0.124	22.07
40) 34 111-trichlor	0.479	0.436	0.353	0.387	0.385	0.378	0.403	11.36
41) 35 11-Di-Cl-pro	0.271	0.306	0.284	0.325	0.324	0.316	0.304	7.40
42) M 36 benzene	1.117	1.160	0.923	0.983	0.960	0.948	1.015	9.69
43) 37 CCl4	0.434	0.397	0.329	0.361	0.358	0.350	0.371	10.20
44) 100 Isobutyl al	0.002	0.005	0.004	0.004	0.005	0.005	0.004	30.78
45) 38 thiophene	0.471	0.573	0.471	0.505	0.504	0.503	0.505	7.36
46) C 39 12-di-Cl-pro	0.247	0.251	0.204	0.220	0.221	0.220	0.227	7.93#
47) M 40 trichloroeth	0.354	0.335	0.274	0.312	0.315	0.309	0.316	8.59
48) 41 dibromometha	0.177	0.159	0.126	0.133	0.134	0.135	0.144	13.68
49) .101 TAME	0.352	0.451	0.420	0.462	0.472	0.498	0.442	11.61
50) 42 Br-di-Cl-met	0.433	0.359	0.280	0.298	0.298	0.297	0.327	17.81
51) 43 Me-methacryl	0.065	0.098	0.100	0.112	0.117	0.121	0.102	20.03
52) 44 2-ClEt-Vi-et	0.010	0.016	0.019	0.025	0.029		0.020	37.07
53) 45 C-13-di-Cl-p	0.312	0.358	0.307	0.333	0.335	0.336	0.330	5.59
54) 46 t-1,3-dichlo	0.227	0.272	0.247	0.270	0.278	0.281	0.262	8.00
55) I 47 Chlorobezene-d5	0.288	0.275	0.213	0.222	0.224	0.226	0.241	13.17
56) 48 112-tri-Cl-E								

Y2

1.00

1.00

0.999

0.997

1.00

0.999

0.992

(#) = Out of Range
 E524A002.M Thu Jul 24 12:41:08 2003

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Initial Calibration

Calibration Files
 .3 =2-00003.D 2 =2-0002A.D 10 =2-00010.D
 20 =2-00020.D 40 =2-00040.D 60 =2-00060.D

Compound	.3	2	10	20	40	60	Avg	%RSD
57) 49 1,3-di-Cl-pro	0.436	0.432	0.347	0.357	0.359	0.353	0.381	10.91
58) 50 Et methacryl	0.189	0.269	0.272	0.287	0.301	0.321	0.273	16.74
59) 51 di-Br-Cl-met	0.422	0.341	0.275	0.296	0.304	0.309	0.325	16.16
60) 52 bromoform	0.222	0.192	0.159	0.171	0.181	0.187	0.185	11.70
61) 53 1,4-dichloro	0.405	0.396	0.312	0.330	0.340	0.344	0.354	10.53
62) 54 MIBK	0.132	0.144	0.134	0.144	0.155	0.166	0.146	8.89
63) 55 toluene-d8	1.345	1.364	1.130	1.244	1.257	1.223	1.260	6.78
64) 56 toluene	1.448	1.627	1.315	1.441	1.448	1.412	1.448	6.98
65) 57 2-hexanone X	0.100	0.109	0.092	0.094	0.098	0.100	0.099	5.95
66) 58 12-dibromoet	0.274	0.254	0.210	0.222	0.233	0.235	0.238	9.68
67) 59 tetra-Cl-eth	0.496	0.464	0.378	0.427	0.424	0.408	0.433	9.68
68) 60 chlorobenzen	1.297	1.156	0.893	0.967	0.959	0.919	1.032	15.48
69) 61 1112-tetra-C	0.447	0.412	0.317	0.340	0.348	0.348	0.369	13.47
70) 62 1,4-Dichlorobenzen	0.269	0.309	0.291	0.356	0.374	0.358	0.326	12.94
71) 63 1-chlorohexa	3.072	3.263	2.780	3.211	3.292	3.156	3.129	6.01#
72) 64 Et-Bz	2.292	2.598	2.175	2.438	2.422	2.268	2.365	6.37
73) 65 m/p-Xylenes	1.472	1.978	1.707	1.902	1.915	1.818	1.798	10.32
74) 66 styrene	1.955	2.508	2.249	2.527	2.563	2.438	2.373	9.83
75) 67 o-xylene	0.659	0.575	0.456	0.475	0.501	0.503	0.528	14.34
76) 68 1122-Tetra-C	0.186	0.176	0.138	0.146	0.151	0.152	0.158	11.76
77) 69 123-tri-Cl-P	0.953	0.876	0.706	0.798	0.823	0.806	0.827	10.01
78) 70 4-Br-1-F-Bz	2.262	3.010	2.857	3.361	3.462	3.290	3.040	14.58
79) 71 isopropylben	0.840	0.881	0.719	0.807	0.842	0.810	0.817	6.73
80) 72 bromobenzene	0.038	0.078	0.075	0.082	0.094	0.095	0.077	27.12
81) 92 t-1,4-dichlo	0.759	0.999	0.867	1.019	1.061	1.009	0.952	12.10
82) 73 n-propylbenz	0.756	0.887	0.758	0.870	0.905	0.873	0.841	7.90
83) 74 2-Cl-Toluene	0.853	0.955	0.757	0.860	0.880	0.830	0.856	7.52
84) 75 4-Cl-Toluene								

✓

0.997

0.999

0.997

(#) = Out of Range
 E524A002.M

Thu Jul 24 12:41:09 2003

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Initial Calibration

Calibration Files
 .3 =2-00003.D 2 =2-0002A.D 10 =2-00010.D
 20 =2-00020.D 40 =2-00040.D 60 =2-00060.D

Compound	.3	2	10	20	40	60	Avg	%RSD
85) 76 135-tri-Me-B	2.031	2.864	2.465	2.815	2.851	2.678	2.617	12.38
86) 77 4-iso-Pr-tol	2.280	3.155	2.726	3.159	3.220	2.998	2.923	12.39
87) 78 124-tri-Me-B	2.267	3.012	2.517	2.854	2.932	2.773	2.726	10.34
88) 79 tert-butylbe	2.041	2.391	2.526	2.636	2.748	2.655	2.499	10.24
89) 80 13-DCB	1.989	1.888	1.484	1.667	1.680	1.592	1.717	10.95
90) 81 sec-butylben	2.715	3.632	3.274	3.857	3.936	3.690	3.517	12.95
91) 82 14-DCB	2.240	1.859	1.457	1.644	1.684	1.613	1.749	15.60
92) 83 Cl-benzyl	0.164	0.159	0.155	0.169	0.181	0.165	0.165	5.41
93) 84 12-DCB	1.762	1.734	1.326	1.464	1.479	1.408	1.529	11.67
94) 85 n-butylbenze	0.528	0.789	0.731	0.875	0.910	0.843	0.779	17.73
95) 86 12-diBr-2-Cl	0.110	0.104	0.091	0.102	0.118	0.120	0.108	10.09
96) 87 124-tri-Cl-B	0.843	0.926	0.884	1.065	1.130	1.091	0.990	12.16
97) 88 naphthalene	2.218	1.820	1.426	1.651	1.832	1.831	1.796	14.50
98) 89 hx-Cl-butadi	0.644	0.621	0.514	0.599	0.629	0.588	0.599	7.74
99) 90 123-Tri-Cl-B	0.849	0.875	0.777	0.914	0.969	0.944	0.888	7.86

Y

0.999

(#) = Out Of Range
 E524A002.M

Thu Jul 24 12:41:09 2003

INITIAL CALIBRATION SUMMARY

Method File		E524A002	
Last Calibration Update		Thu Jul 24 12:40:35 2003	
Level 1 File Name	2-00003.D	Level 1 ID	.3
Level 2 File Name	2-0002A.D	Level 2 ID	2
Level 3 File Name	2-00010.D	Level 3 ID	10
Level 4 File Name	2-00020.D	Level 4 ID	20
Level 5 File Name	2-00040.D	Level 5 ID	40
Level 6 File Name	2-00060.D	Level 6 ID	60
Level 7 File Name	2-00020.D	Level 7 ID	cc

Compound	Response							Level 7	Coeff X ^{v0}	Coeff X ^{v1} / ave RF	Coeff X ^{v2}	R ^{v2} / RSD
	Level 1	Level 2	Level 3	Level 4	Level 5	Level 6	Level 7					
1 Fluorobenzene	919393	1006828	1003977	1015499	990930	980047	-1	-----	-----	-----	-----	
3 di-Ci-di-F-methane	7698	51044	228452	508778	990705	1444214	-1	0.0000	0.2510	0.0000	0.0661	
4 Chloromethane	8264	53373	208536	448729	864850	11774857	-1	0.0169	0.2027	0.0000	0.9966	
2 F114	3209	27835	126949	275940	544934	783345	-1	0.0000	0.1313	0.0000	0.0645	
5 vinyl chloride	7381	53695	219399	484855	949966	1312837	-1	0.0000	0.2424	0.0000	0.0861	
6 bromomethane	3952	25743	105105	228520	411844	626111	-1	0.0000	0.1165	0.0000	0.1363	
7 chloroethane	4460	34347	136103	266250	546077	758703	-1	0.0000	0.1443	0.0000	0.1208	
8 tri-Ci-F-methane	11754	73985	311036	724847	1412342	1986417	-1	0.0000	0.3591	0.0000	0.1076	
91 Acetonitrile X10	-1	110646	413272	845923	1639456	2365932	-1	0.0000	0.0439	0.0000	0.1416	
9 acrolein X10	9511	67561	262941	513017	937791	1313901	-1	0.0306	0.0223	0.0000	0.9978	
11 acetone X10	29124	106138	281400	497276	943071	1241915	-1	0.0622	0.0208	0.0000	0.9944	
12 ethyl ether X5	22348	158277	587938	1189867	2188257	2883396	-1	0.0885	0.0991	0.0000	0.9929	
13 11-dichloroethene	9522	62800	259067	562011	1088836	1490545	-1	0.0000	0.2867	0.0000	0.1231	
14 Iodomethane	7324	57536	244664	532902	1022337	1206486	-1	0.0000	0.2534	0.0000	0.1075	
15 F-113	8636	49466	193654	427782	814332	1086062	-1	0.0189	0.1881	0.0000	0.9947	
16 acrylonitrile X10	14700	104060	420553	831510	1601521	2352958	-1	0.0000	0.0447	0.0000	0.1361	
17 carbon disulfide	23125	156993	656825	1480814	2862257	3817157	-1	0.0000	0.7288	0.0000	0.1000	
94 Isopropyl Alcoholx10	4021	15273	46223	124874	265806	369178	-1	-0.0027	0.0064	0.0000	0.9954	
18 methylene chloride	36501	97000	257935	510536	960233	1435510	-1	0.0322	0.2370	0.0000	0.9996	
19 t-12-di-Ci-ethene	8180	61414	249312	531623	1015155	1488348	-1	0.0000	0.2701	0.0000	0.0899	
20 t-Bu-Me-ether	13066	102540	420692	866584	1693079	2598506	-1	0.0000	0.4496	0.0000	0.0781	
95 Tert butyl alcoholx10	-1	28855	98345	205914	482970	785022	-1	-0.0326	0.0135	0.0000	0.9918	

Compound Name	Level 1 Response	Level 2 Response	Level 3 Response	Level 4 Response	Level 5 Response	Level 6 Response	Level 7 Response	Coeff X^0	Coeff X^1 / ave RF	Coeff X^2	R^2 / RSD
94 allyl chloride	-1	110646	413272	845923	1216384	1660968	-1	0.0000	0.3934	0.0000	0.2697
21 1,1-dichloroethane	14292	97373	381929	825889	1598207	2404444	-1	0.0000	0.4335	0.0000	0.1252
97 propionitrile	-1	4103	13637	33328	64498	101359	-1	0.0000	0.0168	0.0000	0.1453
22 c-12-di-Cl-ethene	8154	62676	253306	533706	1019103	1499328	-1	0.0000	0.2723	0.0000	0.0912
23 2,2-Dichloropropane	10809	72464	281116	589735	1098975	1539718	-1	0.0244	0.2629	0.0000	0.9982
24 Br-Cl-methane	4359	30212	117713	250220	483734	716534	-1	0.0000	0.1321	0.0000	0.1312
25 chloroform	19972	106303	398005	838697	1597118	2387736	-1	0.0108	0.4031	0.0000	0.9998
26 tetrahydrofuranX5	4965	35404	154519	318868	624422	946730	-1	0.0000	0.0328	0.0000	0.0666
98 Diisopropyl ether	17882	159533	636576	1326012	2521921	3767835	-1	0.0000	0.6674	0.0000	0.0923
27 Di-Br-F-Me (surr)	-1	59230	230740	486919	949109	1428311	-1	0.0000	0.2492	0.0000	0.1027
99 ETBE	12061	107588	486462	1064815	2106754	3231163	-1	0.0000	0.5102	0.0000	0.0820
29 1,2-Di-Cl-Et-d4 (S1)	-1	50036	187219	388784	757790	1126883	-1	0.0000	0.2018	0.0000	0.1296
30 1,2-dichloroethane	2582	18795	70404	142937	277276	411282	-1	0.0020	0.0695	0.0000	0.9999
32 vinyl acetate X5	43949	379164	1608933	2957025	5653569	8226056	-1	0.0000	0.3120	0.0000	0.1153
92 Nitro Methane(x10)	-1	12060	47784	111334	237578	325914	-1	0.0000	0.0056	0.0000	0.0909
33 2-butanoneMEK X10	22369	123309	444332	875377	1648493	2411259	-1	0.0340	0.0406	0.0000	0.9998
93 Ethyl Acetate x2	9671	54370	220002	434881	845931	1295021	-1	0.0014	0.1088	0.0000	0.9995
34 1,1-trichloroethane	13204	87716	354801	785718	1524561	2222673	-1	0.0000	0.4029	0.0000	0.1136
35 1,1-Di-Cl-propene	7463	61583	285223	660985	1285689	1859527	-1	0.0000	0.3044	0.0000	0.0740
36 benzene	30808	233485	926403	1995763	3805709	5575015	-1	0.0000	1.0150	0.0000	0.0969
37 CC14	11983	79844	329846	732441	1420059	2060040	-1	0.0000	0.3715	0.0000	0.1020
100 Isobutyl alcoholx10	457	10371	40745	87337	182637	287972	-1	-----	-----	-----	-----
38 thiophene	12985	115327	473132	1025220	1999457	2960047	-1	0.0000	0.5046	0.0000	0.0736
39 1,2-di-Cl-propane	6808	50516	204989	446632	875952	1293157	-1	0.0000	0.2271	0.0000	0.0793
40 trichloroethene	9774	67466	274815	633192	1246934	1815606	-1	0.0000	0.3164	0.0000	0.0859
41 dibromomethane	4889	32002	126997	270225	530196	793122	-1	0.0000	0.1441	0.0000	0.1368
101 TAME	9699	90788	421379	938326	1872529	2927534	-1	0.0000	0.4424	0.0000	0.1161
42 Br-di-Cl-methane	11934	72312	281223	604352	1181950	1745648	-1	0.0014	0.2967	0.0000	0.9998
43 Me-methacrylate	1794	19729	100519	228095	464768	714088	-1	-0.0114	0.1219	0.0000	0.9990
44 2-ClEt-VI-ether10	2834	31796	189485	514977	1130652	-1	-1	-0.0446	0.0290	0.0000	0.9927
45 c-1,3-di-Cl-propene	8596	71992	308098	676158	1328342	1977341	-1	0.0000	0.3301	0.0000	0.0559
46 t-1,3-dichloropropene	6273	54825	247619	547714	1100371	1653172	-1	0.0000	0.2625	0.0000	0.0800
47 Chlorobezene-d5	683445	775455	774184	772935	739886	736062	-1	0.0000	1.0000	0.0000	0.0000

C:\RPtical-voc.xls

Compound	Level 1 Response	Level 2 Response	Level 3 Response	Level 4 Response	Level 5 Response	Level 6 Response	Level 7 Response	Coeff Xv0	Coeff Xv1 / ave RF	Coeff Xv2	Rv2/ RSD
48 112-tri-Cl-Et	5902	42713	164770	343401	662895	997315	-1	0.0000	0.2413	0.0000	0.1317
49 13-di-Cl-propane	8941	66977	268452	552557	1062190	1558760	-1	0.0000	0.3807	0.0000	0.1091
50 Et methacrylate	3865	41714	210769	443099	890456	1418605	-1	-0.0322	0.3197	0.0000	0.9978
51 di-Br-Cl-methane	8661	52843	213283	456818	899440	1365305	-1	-0.0111	0.3087	0.0000	0.9994
52 bromoform	4560	29809	123106	263988	534402	823804	-1	0.0000	0.1852	0.0000	0.1170
53 1,4-dichlorobutane-2	8296	61397	241554	510059	1005306	1519089	-1	0.0000	0.3544	0.0000	0.1053
54 MIBK	2701	22334	103475	222364	457366	733371	-1	0.0000	0.1456	0.0000	0.0889
55 toluene-d8	27570	211476	874470	1923324	3719816	5400714	-1	0.0000	1.2603	0.0000	0.0678
56 toluene	29681	252340	1018191	2227432	4284454	6235319	-1	0.0000	1.4484	0.0000	0.0698
57 2-hexanone X5	10224	84522	356146	728526	1453530	2205707	-1	0.0000	0.0989	0.0000	0.0595
58 12-dibromoethane	5621	39459	162435	343609	688415	1038913	-1	0.0000	0.2381	0.0000	0.0968
59 tetra-Cl-ethene	10167	72027	292262	659769	1256084	1800306	-1	0.0000	0.4328	0.0000	0.0968
60 chlorobenzene	26600	179300	691236	1494861	2837497	4059222	-1	0.0338	0.9255	0.0000	0.9991
61 1112-tetra-Cl-Et	9159	63898	245258	525072	1031299	1537036	-1	0.0000	0.3686	0.0000	0.1347
62 1,4-Dichlorobenzene-d4	358277	419640	407774	393961	367099	365124	-1	0.0000	1.0000	0.0000	0.0000
63 1-chlorohexane	2893	25946	118729	280119	548810	784184	-1	0.0000	0.3261	0.0000	0.1294
64 Et-Bz	33015	273886	1133719	2529977	4833949	6913463	-1	0.0000	3.1290	0.0000	0.0601
65 m/p-Xylenes X2	49281	436018	1773476	3842318	7112016	9936549	-1	0.0000	2.3654	0.0000	0.0637
66 styrene	15818	166023	695954	1498467	2811895	3981902	-1	0.0000	1.7985	0.0000	0.1032
67 o-xylene	21010	210456	917264	1990843	3763083	5341093	-1	0.0000	2.3732	0.0000	0.0983
68 1122-Tetra-Cl-Et	7086	48249	186045	374484	736112	1101133	-1	0.0000	0.5283	0.0000	0.1434
69 123-tri-Cl-Pr	1994	14801	56224	114784	222436	332760	-1	0.0000	0.1581	0.0000	0.1176
70 4-Br-1-F-Bz (S3)	10243	73535	287803	629095	1208574	1766234	-1	0.0000	0.8271	0.0000	0.1001
71 isopropylbenzene	24316	252585	1164869	2648278	5083237	7208570	-1	0.0000	3.0403	0.0000	0.1458
72 bromobenzene	9030	73980	293172	635461	1236729	1774887	-1	0.0000	0.8166	0.0000	0.0673
92 t-1,4-dichloro-2-butene	406	6524	30509	64587	137632	207953	-1	-0.0118	0.0961	0.0000	0.9978
73 n-propylbenzene	8157	83853	353478	802663	1558540	2210026	-1	0.0000	0.9523	0.0000	0.1210
74 2-Cl-Toluene	8123	74427	309252	685694	1328259	1912837	-1	0.0000	0.8415	0.0000	0.0790
75 4-Cl-Toluene	9172	80121	308786	677984	1291985	1819259	-1	0.0000	0.8560	0.0000	0.0752
76 135-tri-Me-Benzene	21830	240336	1005234	2217776	4186569	5867539	-1	0.0000	2.6173	0.0000	0.1238
77 4-iso-Pr-toluene	24508	264794	1111631	2489320	4727712	6667533	-1	0.0000	2.9230	0.0000	0.1239
78 124-tri-Me-Benzene	24365	252796	1026195	2248874	4305294	6074118	-1	0.0000	2.7257	0.0000	0.1034
79 tert-butylbenzene	21932	200692	1030129	2077031	4034706	5816456	-1	0.0000	2.4995	0.0000	0.1024

80	13-DCB	21374	158442	605258	1313556	2467081	3488098	-1	0.0000	1.7167	0.0000	0.1095
81	sec-butylbenzene	29181	304851	1334960	3039109	5779243	8084394	-1	0.0000	3.5174	0.0000	0.1295
82	14-DCB	24079	155998	594167	1295125	2472135	3533637	-1	-0.0003	1.6325	0.0000	0.9988
83	Cl-benzyl	1764	13306	63403	132793	265795	361085	-1	0.0000	0.1654	0.0000	0.0541
84	12-DCB	18942	145496	540526	1153351	2171646	3083911	-1	0.0000	1.5286	0.0000	0.1167
85	n-butylbenzene	5679	66251	298215	689097	1335543	1846727	-1	-0.0178	0.8660	0.0000	0.9968
86	12-diBr-2-Cl-Pra	1180	8757	37194	80335	173925	263112	-1	0.0000	0.1076	0.0000	0.1009
87	124-tri-Cl-Bz	9061	77700	360446	838928	1658990	2390948	-1	0.0000	0.9898	0.0000	0.1216
88	naphthalene	23839	152715	581293	1300980	2689379	4011635	-1	0.0000	1.7961	0.0000	0.1450
89	hx-Cl-butadiene	6922	52135	209577	472099	923440	1288580	-1	0.0000	0.5992	0.0000	0.0774
90	123-Tri-Cl-Bz	9124	73440	317007	719921	1423460	2068944	-1	0.0000	0.8881	0.0000	0.0786

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00003.D Sample : f=1 0.3 ppb
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Jul 22 19:25 2003 RF via : Multiple Level Calibration
 Method Update: Mon Jul 21 15:01 2003 Operator: zou
 Quant. Time : Jul 24 11:52 2003 Multiplr: 1.000000
 Print Time : Thu Jul 24 11:52 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	919.393	10.00		0.02	
47	Chlorobenzene-d5	11.55	11.54	0.000	117	82	683.445	10.00		0.00	
62	1,4-Dichlorobenze	13.84	13.84	0.000	152	150	358.277	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.61	6.58	0.002	111	113	10.081	0.48		0.5	2.38%
29	1,2-Di-Cl-Et-d4 (7.11	7.08	0.002	65	102	8.754	0.45		0.4	2.23%
55	toluene-d8	9.91	9.89	0.000	98	100	27.570	0.32		0.3	1.58%
70	4-Br-1-F-Bz (S3)	12.75	12.74	0.000	174	95	10.243	0.38		0.4	1.88%

Target Compounds	<<< I1 : ISTD ID = 1 >>>	Qvalue
3 di-Cl-di-F-methan	1.89 1.85 0.005 85 87 7.698 0.34 0.3 0.3 97	<i>De</i>
4 Chloromethane	2.11 2.07 0.006 50 52 8.264 0.34 0.3 0.3 97	<i>De</i>
2 F114	2.04 2.00 0.006 85 135 3.209 0.27 0.3 0.3 96	<i>De</i>
5 vinyl chloride	2.22 2.19 0.004 62 64 7.381 0.32 0.3 0.3 89	<i>De</i>
6 bromomethane	2.61 2.58 0.005 94 96 3.952 0.30 0.3 0.3 95	<i>De</i>
7 chloroethane	2.73 2.70 0.004 64 66 4.460 0.30 0.3 0.3 0	<i>De</i>
8 tri-Cl-F-methane	3.03 3.00 0.005 101 103 11.754 0.34 0.3 0.3 99	<i>De</i>
91 Acetonitrile X10	4.07 4.04 0.004 41 40 24.302 8.33 8.3 73	<i>De</i>
9 acrolein X10	3.51 3.48 0.005 56 55 9.511 4.28 4.3 0	<i>De</i>
11 acetone X10	3.72 3.69 0.004 43 58 29.124 12.37 12.4 94	<i>De</i>
12 ethyl ether X5	3.37 3.34 0.005 59 74 22.348 1.91 1.9 92	<i>De</i>
13 11-dichloroethene	3.64 3.60 0.005 61 96 9.522 0.34 0.3 0.3 0	<i>De</i>
14 Iodomethane	3.82 3.78 0.006 142 127 7.324 0.47 0.5 0.5 84	<i>De</i>
15 F-113	3.65 3.62 0.005 101 151 8.636 0.42 0.4 0.4 0	<i>De</i>
16 acrylonitrile X10	4.53 4.49 0.005 53 52 14.700 4.72 4.7 89	<i>De</i>
17 carbon disulfide	3.90 3.87 0.004 76 78 23.125 0.28 0.3 0.3 98	<i>De</i>
94 Isopropyl Alcohol	4.08 4.01 0.010 45 43 4.021 14.57 14.6 100	<i>De</i>
18 methylene chlorid	4.22 4.19 0.004 84 49 36.501 1.52 1.5 94	<i>De</i>
19 t-12-di-Cl-ethene	4.58 4.55 0.004 96 61 8.180 0.35 0.3 83	<i>De</i>

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00003.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 19:25 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:52 2003
 Print Time : Thu Jul 24 11:52 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.60	4.56	0.005	73	57	13.066	0.31	0.3	93	?
95	Tert butyl alcoho	4.47	4.47	0.000	59	57	6.515	8.83	8.8	100	m
94	allyl chloride	4.07	4.04	0.004	41	76	24.302	0.71	0.7	62	#?
21	11-dichloroethane	5.12	5.09	0.005	63	83	14.292	0.37	0.4	95	m
22	c-12-di-Cl-ethene	5.92	5.89	0.004	96	61	8.154	0.35	0.3	96	?
23	22-Dichloropropan	5.92	5.89	0.003	77	97	10.809	0.38	0.4	95	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	4.359	0.47	0.5	87	?
25	chloroform	6.37	6.35	0.002	83	85	19.972	0.51	0.5	100	?
26	tetrahydrofuranX5	6.36	6.32	0.005	42	72	4.965	1.89	1.9	89	?
98	Disopropyl ether	5.24	5.22	0.003	45	87	17.882	0.28	0.3	69	#
99	ETBE	5.75	5.72	0.004	59	87	12.061	0.25	0.2	97	#
30	12-dichloroethane	7.23	7.20	0.004	64	62	2.582	0.36	0.4	83	m
32	vinyl acetate X5	5.20	5.17	0.005	43	86	43.949	1.55	1.5	99	m
92	Nitro Methane(X10	5.81	5.80	0.002	61	46	1.737	0.82	0.8	47	#
33	2-butanoneMEK X10	5.95	5.92	0.004	43	72	22.369	7.70	7.7	96	#
93	Ethyl Acetate x2	6.04	6.02	0.003	43	61	9.671	0.81	0.8	78	#
34	111-trichloroetha	6.65	6.63	0.002	97	99	13.204	0.38	0.4	99	?
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	7.463	0.25	0.2	91	?
36	benzene	7.21	7.19	0.003	78	52	30.808	0.33	0.3	96	?
37	CCl4	6.91	6.89	0.002	117	119	11.983	0.40	0.4	92	?
100	Isobutyl alcohol	7.19	7.39	-0.027	43	42	0.457	2.50	2.5	89	m
38	thiophene	7.53	7.51	0.003	84	58	12.985	0.28	0.3	94	m
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	6.808	0.34	0.3	94	m
40	trichloroethene	8.24	8.23	0.002	130	132	9.774	0.39	0.4	96	m
41	dibromomethane	8.73	8.71	0.002	174	172	4.889	0.51	0.5	87	m
101	TAME	7.42	7.39	0.003	73	43	9.699	0.22	0.2	83	#
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	11.934	0.44	0.4	100	#
43	Me-methacrylate	8.76	8.75	0.002	69	100	1.794	0.19	0.2	99	#
44	2-ClEt-VI-ether10	9.38	9.37	0.002	63	43	2.834	1.36	1.4	84	#
45	c-13-di-Cl-propen	9.57	9.55	0.002	75	110	8.596	0.29	0.3	97	#
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	6.273	0.28	0.3	90	#

<<< I2 : ISTD ID = 47 >>>
 # = qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00003.D Sample : F=1 0.3 ppb
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Jul 22 19:25 2003 RF via : Multiple Level Calibration 2469
 Method Update: Mon Jul 21 15:01 2003 Operator: zou
 Quant. Time : Jul 24 11:52 2003 Multiplr: 1.000000
 Print Time : Thu Jul 24 11:52 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	QIon	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	8.941	0.36	0.4	90	?
50	Et methacrylate	10.38	10.37	0.001	69	99	3.865	0.19	0.2	92	
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	8.661	0.50	0.5	99	
52	bromoform	12.42	12.41	0.000	173	174	4.560	0.49	0.5	95	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	8.296	0.37	0.4	91	
54	MIBK	9.77	9.76	0.000	43	58	2.701	0.28	0.3	85	
56	toluene	9.99	9.98	0.000	91	92	29.681	0.29	0.3	96	
57	2-hexanone X5	10.76	10.75	0.001	43	58	10.224	1.59	1.6	94	
58	12-dibromoethane	11.03	11.03	0.000	107	109	5.621	0.42	0.4	90	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	10.167	0.39	0.4	95	?
60	chlorobenzene	11.58	11.57	0.000	112	77	26.600	0.40	0.4	90	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	9.159	0.43	0.4	95	
<<< I3 : ISTD ID = 62 >>>											
63	1-chlorohexane	11.56	11.56	0.000	93	55	2.893	0.20	0.2	54	#?
64	Et-Bz	11.70	11.69	0.000	91	106	33.015	0.26	0.3	99	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	49.281	0.51	0.5	99	
66	styrene	12.24	12.23	0.000	104	78	15.818	0.21	0.2	80	?
67	o-xylene	12.22	12.22	0.000	91	106	21.010	0.21	0.2	97	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	7.086	0.41	0.4	98	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	1.994	0.39	0.4	100	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	24.316	0.20	0.2	97	
72	bromobenzene	12.89	12.89	0.000	156	158	9.030	0.33	0.3	97	?
92	t-1,4-dichloro-2-	12.93	12.92	0.000	89	53	0.406	0.16	0.2	19	#?
73	n-propylbenzene	13.00	12.99	0.000	120	78	8.157	0.23	0.2	82	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	8.123	0.27	0.3	94	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	9.172	0.29	0.3	88	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	21.830	0.21	0.2	96	?
77	4-Iso-Pr-toluene	13.81	13.81	0.000	119	134	24.508	0.21	0.2	97	?
78	124-tri-Me-Benzen	13.52	13.51	0.000	105	120	24.365	0.23	0.2	98	
79	tert-butylbenzene	13.48	13.47	0.000	119	91	21.932	0.23	0.2	94	
80	13-DCB	13.79	13.78	0.000	146	148	21.374	0.37	0.4	97	
81	sec-butylbenzene	13.68	13.68	0.000	105	134	29.181	0.21	0.2	98	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06


John
2/27/03

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00003.D Sample : f=1 0.3 ppb
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Jul 22 19:25 2003 RF via : Multiple Level Calibration
 Method Update: Mon Jul 21 15:01 2003 Operator: zou
 Quant. Time : Jul 24 11:52 2003 Multiplr: 1.000000
 Print Time : Thu Jul 24 11:52 2003
 Miscelaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
82	14-DCB	13.87	13.87	0.000	146	148	24.079	0.41	0.4	92	
83	Cl-benzy1	13.98	13.98	0.000	126	91	1.764	0.41	0.4	82	#
84	12-DCB	14.21	14.21	0.000	146	148	18.942	0.38	0.4	95	
85	n-butylbenzene	14.18	14.18	0.000	134	91	5.679	0.19	0.2	87	#
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	1.180	0.41	0.4	81	
87	124-tri-Cl-Bz	15.59	15.58	0.000	180	182	9.061	0.25	0.2	96	
88	naphthalene	15.79	15.78	0.000	128	129	23.839	0.45	0.4	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	6.922	0.32	0.3	88	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	9.124	0.30	0.3	91	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00003.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 19:25 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:52 2003
 Print Time : Thu Jul 24 11:52 2003
 Miscellaneous :
 Sample : f=1 0.3 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000



Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-0002A.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:11 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:56 2003
 Print Time : Thu Jul 24 11:56 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	1006.828	10.00		0.02	
47	Chlorobenzene-d5	11.54	11.54	0.000	117	82	775.455	10.00		0.00	
62	1,4-Dichlorobenzene	13.85	13.84	0.000	152	150	419.640	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	59.230	2.56		2.6	12.78%
29	1,2-Di-Cl-Et-d4 (7.11	7.08	0.002	65	102	50.036	2.33		2.3	11.64%
55	toluene-d8	9.91	9.89	0.000	98	100	211.476	2.14		2.1	10.71%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	73.535	2.30		2.3	11.51%

Target Compounds <<< 11 : ISTD ID = 1 >>>

3	di-Cl-di-F-methan	1.89	1.85	0.005	85	87	51.044	2.04		2.0	96	
4	Chloromethane	2.11	2.07	0.006	50	52	53.373	2.02		2.0	98	
2	F114	2.03	2.00	0.005	85	135	27.835	2.13		2.1	54	
5	vinyl chloride	2.23	2.19	0.005	62	64	53.695	2.14		2.1	98	
6	bromomethane	2.61	2.58	0.005	94	96	25.743	1.77		1.8	100	
7	chloroethane	2.73	2.70	0.004	64	66	34.347	2.11		2.1	96	
8	tri-Cl-F-methane	3.03	3.00	0.005	101	103	73.985	1.97		2.0	100	
91	Acetonitrile X10	4.07	4.04	0.004	41	40	110.646	34.62		34.6	84	
9	acrolein X10	3.52	3.48	0.006	56	55	67.561	27.75		27.8	99	
11	acetone X10	3.73	3.69	0.005	43	58	106.138	41.18		41.2	0	
12	ethyl ether X5	3.37	3.34	0.005	59	74	158.277	12.35		12.4	89	
13	11-dichloroethene	3.64	3.60	0.005	61	96	62.800	2.04		2.0	94	
14	Iodomethane	3.82	3.78	0.005	142	127	57.536	3.40		3.4	93	
15	F-113	3.65	3.62	0.005	101	151	49.466	2.20		2.2	90	
16	acrylonitrile X10	4.53	4.49	0.005	53	52	104.060	30.48		30.5	95	
17	carbon disulfide	3.90	3.87	0.004	76	78	156.993	1.77		1.8	99	
94	Isopropyl Alcohol	4.10	4.01	0.011	45	43	15.273	50.54		50.5	100	
18	methylene chlorid	4.22	4.19	0.005	84	49	97.000	3.70		3.7	94	
19	t-12-di-Cl-ethene	4.58	4.55	0.005	96	61	61.414	2.39		2.4	90	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

De
07/24/03
De
07/24/03
De
07/24/03
De
07/24/03

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-0002A.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:11 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:56 2003
 Print Time : Thu Jul 24 11:56 2003
 Miscellaneous :

Sample : f=1 2 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000

2473

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.60	4.56	0.005	73	57	102.540	2.19	2.2	97	?
95	Tert butyl alcoho	4.50	4.47	0.005	59	57	28.855	35.70	35.7	100	me
94	allyl chloride	4.07	4.04	0.004	41	76	110.646	2.96	3.0	77	#?
21	11-dichloroethane	5.12	5.09	0.004	63	83	97.373	2.28	2.3	95	me
97	propionitrile	6.03	5.99	0.004	54	51	4.103	2.96	3.0	100	me
22	c-12-di-Cl-ethene	5.92	5.89	0.004	96	61	62.676	2.45	2.5	90	?
23	22-Dichloropropan	5.92	5.89	0.004	77	97	72.464	2.34	2.3	99	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	30.212	2.99	3.0	97	?
25	chloroform	6.38	6.35	0.003	83	85	106.303	2.47	2.5	99	?
26	tetrahydrofuranX5	6.36	6.32	0.005	42	72	35.404	12.31	12.3	89	?
98	Disopropyl ether	5.25	5.22	0.004	45	87	159.533	2.30	2.3	97	?
99	ETBE	5.75	5.72	0.004	59	87	107.588	2.02	2.0	95	#?
30	12-dichloroethane	7.22	7.20	0.002	64	62	18.795	2.38	2.4	89	#?
32	vinyl acetate X5	5.20	5.17	0.004	43	86	379.164	12.18	12.2	98	me
92	Nitro Methane(X10	5.84	5.80	0.005	61	46	12.060	5.18	5.2	2	me
33	2-butanoneMEK X10	5.95	5.92	0.003	43	72	123.309	38.78	38.8	95	#?
93	Ethyl Acetate x2	6.05	6.02	0.004	43	61	54.370	4.14	4.1	88	#?
34	111-trichloroetha	6.65	6.63	0.002	97	99	87.716	2.28	2.3	98	?
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	61.583	1.86	1.9	88	?
36	benzene	7.22	7.19	0.004	78	52	233.485	2.31	2.3	100	?
37	CCl4	6.91	6.89	0.003	117	119	79.844	2.46	2.5	100	?
100	Isobutyl alcohol	7.19	7.39	-0.027	43	42	10.371	51.83	51.8	91	me
38	thiophene	7.53	7.51	0.002	84	58	115.327	2.25	2.3	95	me
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	50.516	2.33	2.3	97	me
40	trichloroethene	8.24	8.23	0.002	130	132	67.466	2.44	2.4	95	me
41	dibromomethane	8.73	8.71	0.002	174	172	32.002	3.05	3.0	100	me
101	TAME	7.42	7.39	0.003	73	43	90.788	1.90	1.9	95	me
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	72.312	2.42	2.4	97	me
43	Me-methacrylate	8.76	8.75	0.002	69	100	19.729	1.93	1.9	90	me
44	2-ClEt-VI-ether10	9.38	9.37	0.002	63	43	31.796	13.91	13.9	97	me
45	c-13-di-Cl-propen	9.57	9.55	0.002	75	110	71.992	2.24	2.2	95	me
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	54.825	2.21	2.2	90	me

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-0002A.D Sample : f=1 2 ppb
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Jul 22 21:11 2003 RF via : Multiple Level Calibration
 Method Update: Mon Jul 21 15:01 2003 Operator: zou
 Quant. Time : Jul 24 11:56 2003 Multiplr: 1.000000
 Print Time : Thu Jul 24 11:56 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<<	I2 : ISTD ID = 47	>>>									
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	42.713	2.65	2.6	97	
49	13-di-Cl-propane	10.65	10.64	0.001	76	78	66.977	2.35	2.3	96	?
50	Et methacrylate	10.38	10.37	0.001	69	99	41.714	1.85	1.8	95	
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	52.843	2.67	2.7	100	
52	bromoform	12.42	12.41	0.000	173	174	29.809	2.82	2.8	100	
53	1,4-dichlorobutan	12.67	12.67	0.000	55	41	61.397	2.38	2.4	96	
54	MIBK	9.77	9.76	0.001	43	58	22.334	2.04	2.0	94	
56	toluene	9.99	9.98	0.001	91	92	252.340	2.17	2.2	99	
57	2-hexanone X5	10.76	10.75	0.001	43	58	84.522	11.62	11.6	94	
58	12-dibromoethane	11.03	11.03	0.000	107	109	39.459	2.61	2.6	94	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	72.027	2.41	2.4	99	?
60	chlorobenzene	11.58	11.57	0.000	112	77	179.300	2.40	2.4	93	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	63.898	2.67	2.7	98	
<<<	I3 : ISTD ID = 62	>>>									
63	1-chlorohexane	11.56	11.56	0.000	93	55	25.946	1.54	1.5	91	?
64	Et-Bz	11.70	11.69	0.000	91	106	273.886	1.85	1.9	96	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	436.018	3.85	3.8	96	
66	styrene	12.24	12.23	0.000	104	78	166.023	1.92	1.9	92	?
67	o-xylene	12.22	12.22	0.000	91	106	210.456	1.81	1.8	99	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	48.249	2.37	2.4	93	
69	123-tri-Cl-Pr	12.92	12.91	0.000	110	97	14.801	2.50	2.5	96	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	252.585	1.76	1.8	99	
72	bromobenzene	12.89	12.89	0.000	156	158	73.980	2.31	2.3	99	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	6.524	2.14	2.1	92	?
73	n-propylbenzene	13.00	12.99	0.000	120	78	83.853	1.98	2.0	94	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	74.427	2.08	2.1	100	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	80.121	2.19	2.2	97	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	240.336	1.95	1.9	98	?
77	4-Iso-Pr-toluene	13.81	13.81	0.000	119	134	264.794	1.97	2.0	99	?
78	124-tri-Me-Benzen	13.52	13.51	0.000	105	120	252.796	1.99	2.0	96	
79	tert-butylbenzene	13.48	13.47	0.000	119	91	200.692	1.81	1.8	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06


Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-0002A.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:11 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:56 2003
 Print Time : Thu Jul 24 11:56 2003
 Miscellaneous :
 Sample : F=1 2 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	158.442	2.31	2.3	99	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	304.851	1.87	1.9	100	
82	14-DCB	13.87	13.87	0.000	146	148	155.998	2.27	2.3	97	
83	Cl-benzy1	13.98	13.98	0.000	126	91	13.306	2.65	2.6	81	#
84	12-DCB	14.21	14.21	0.000	146	148	145.496	2.46	2.5	100	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	66.251	1.89	1.9	94	?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	8.757	2.61	2.6	89	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	77.700	1.82	1.8	97	
88	naphthalene	15.79	15.78	0.000	128	129	152.715	2.44	2.4	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	52.135	2.05	2.0	100	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	73.440	2.07	2.1	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-0002A.D
Method : C:\MSDCHEM\1\METHODS\E524A002.M
Acq. Time : Jul 22 21:11 2003
Method Update: Mon Jul 21 15:01 2003
Quant. Time : Jul 24 11:56 2003
Print Time : Thu Jul 24 11:56 2003
Miscellaneous :

Sample : F=1 2 ppb
Inst. : GCMS-A
RF via : Multiple Level Calibration
Operator: zou
Multiplier: 1.000000



Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00010.D Sample : F=1 10 ppb
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Jul 22 21:37 2003 RF via : Multiple Level Calibration
 Method Update: Mon Jul 21 15:01 2003 Operator: zou
 Quant. Time : Jul 24 11:58 2003 Multiplr: 1.000000
 Print Time : Thu Jul 24 11:58 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	QIon	Q1	RF/1000	CO,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	1003.977	10.00		0.02	
47	Chlorobenzene-d5	11.55	11.54	0.000	117	82	774.184	10.00		0.00	
62	1,4-Dichlorobenzene	13.85	13.84	0.000	152	150	407.774	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	230.740	9.99	10.0	49.93%	
29	1,2-Di-Cl-Et-d4 (7.11	7.08	0.002	65	102	187.219	8.73	8.7	43.66%	
55	toluene-d8	9.91	9.89	0.000	98	100	874.470	8.87	8.9	44.35%	
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	287.803	9.27	9.3	46.37%	

Target Compounds <<< 11 : ISTD ID = 1 >>> Qvalue

3	di-Cl-di-F-methan	1.89	1.85	0.005	85	87	228.452	9.14	9.1	98	
4	Chloromethane	2.11	2.07	0.006	50	52	208.536	7.92	7.9	98	
2	F114	2.03	2.00	0.005	85	135	126.949	9.74	9.7	59	
5	vinyl chloride	2.22	2.19	0.004	62	64	219.399	8.78	8.8	100	
6	bromomethane	2.61	2.58	0.005	94	96	105.105	7.25	7.2	97	
7	chloroethane	2.73	2.70	0.004	64	66	136.103	8.39	8.4	100	
8	tri-Cl-F-methane	3.03	3.00	0.005	101	103	311.036	8.29	8.3	100	
91	Acetonitrile X10	4.07	4.04	0.004	41	40	413.272	129.67	129.7	89	
9	acrolein X10	3.51	3.48	0.005	56	55	262.941	108.33	108.3	0	
11	acetone X10	3.73	3.69	0.005	43	58	281.400	109.48	109.5	0	
12	ethyl ether X5	3.37	3.34	0.005	59	74	587.938	46.01	46.0	88	
13	1,1-dichloroethene	3.64	3.60	0.005	61	96	259.067	8.43	8.4	98	
14	Iodomethane	3.82	3.78	0.005	142	127	244.664	14.51	14.5	93	
15	F-113	3.65	3.62	0.005	101	151	193.654	8.65	8.6	87	
16	acrylonitrile X10	4.53	4.49	0.005	53	52	420.553	123.54	123.5	97	
17	carbon disulfide	3.91	3.87	0.005	76	78	656.825	7.41	7.4	100	
94	Isopropyl Alcohol	4.11	4.01	0.014	45	43	46.223	153.39	153.4	100	
18	methylene chlorid	4.22	4.19	0.005	84	49	257.935	9.86	9.9	100	
19	t-12-di-Cl-ethene	4.58	4.55	0.005	96	61	249.312	9.75	9.7	93	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Handwritten notes:
 m 07/24/03
 m 07/24/03
 m 07/24/03
 ?
 ?
 ?
 ?
 ?

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00010.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:37 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:58 2003
 Print Time : Thu Jul 24 11:58 2003
 Miscellaneous :

Sample : F=1 10 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.60	4.56	0.005	73	57	420.692	9.03	9.0	97	?
95	Tert butyl alcoho	4.58	4.47	0.014	59	57	98.345	122.02	122.0	100	m?
94	allyl chloride	4.07	4.04	0.004	41	76	413.272	11.07	11.1	81	#?
21	11-dichloroethane	5.12	5.09	0.004	63	83	381.929	8.98	9.0	99	
97	propionitrile	6.03	5.99	0.004	54	51	13.637	9.86	9.9	100	#?
22	c-12-di-Cl-ethene	5.92	5.89	0.003	96	61	253.306	9.94	9.9	88	?
23	22-Dichloropropan	5.92	5.89	0.004	77	97	281.116	9.09	9.1	95	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	117.713	11.68	11.7	96	?
25	chloroform	6.37	6.35	0.002	83	85	398.005	9.29	9.3	98	?
26	tetrahydrofuranX5	6.35	6.32	0.003	42	72	154.519	53.88	53.9	88	?
98	Diisopropyl ether	5.24	5.22	0.003	45	87	636.576	9.21	9.2	96	
99	ETBE	5.74	5.72	0.003	59	87	486.462	9.16	9.2	95	
30	12-dichloroethane	7.22	7.20	0.003	64	62	70.404	8.93	8.9	96	?
32	vinyl acetate X5	5.20	5.17	0.004	43	86	1608.933	51.83	51.8	99	
92	Nitro Methane(x10	5.83	5.80	0.004	61	46	47.784	20.59	20.6	88	
33	2-butanoneMEK X10	5.95	5.92	0.003	43	72	444.332	140.12	140.1	95	?
93	Ethyl Acetate x2	6.04	6.02	0.003	43	61	220.002	16.80	16.8	92	#?
34	111-trichloroetha	6.65	6.63	0.002	97	99	354.801	9.25	9.2	100	
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	285.223	8.66	8.7	88	?
36	benzene	7.21	7.19	0.003	78	52	926.403	9.17	9.2	99	?
37	CCl4	6.91	6.89	0.003	117	119	329.846	10.19	10.2	99	?
100	Isobutyl alcohol	7.19	7.39	-0.027	43	42	40.745	204.19	204.2	95	m?
38	thiophene	7.53	7.51	0.003	84	58	473.132	9.26	9.3	98	
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	204.989	9.47	9.5	98	
40	trichloroethene	8.24	8.23	0.002	130	132	274.815	9.97	10.0	99	
41	dibromomethane	8.73	8.71	0.002	174	172	126.997	12.13	12.1	99	
101	TAME	7.41	7.39	0.002	73	43	421.379	8.86	8.9	98	
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	281.223	9.42	9.4	99	
43	Me-methacrylate	8.76	8.75	0.002	69	100	100.519	9.84	9.8	97	
44	2-ClEt-Vi-ether10	9.38	9.37	0.002	63	43	189.485	83.14	83.1	98	
45	c-13-di-Cl-propen	9.56	9.55	0.002	75	110	308.098	9.63	9.6	94	
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	247.619	10.02	10.0	93	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00010.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:37 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:58 2003
 Print Time : Thu Jul 24 11:58 2003
 Miscellaneous :

Sample : F=1 10 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<< I2 : ISTD ID = 47 >>>											
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	164.770	10.23	10.2	98	
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	268.452	9.42	9.4	100	?
50	Et methacrylate	10.37	10.37	0.000	69	99	210.769	9.34	9.3	94	
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	213.283	10.78	10.8	100	
52	bromoform	12.42	12.41	0.000	173	174	123.106	11.66	11.7	100	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	241.554	9.38	9.4	95	
54	MIBK	9.77	9.76	0.000	43	58	103.475	9.46	9.5	99	
56	toluene	9.99	9.98	0.000	91	92	1018.191	8.79	8.8	100	
57	2-hexanone X5	10.76	10.75	0.000	43	58	356.146	49.05	49.0	95	
58	12-dibromoethane	11.03	11.03	0.000	107	109	162.435	10.76	10.8	97	
59	tetra-Cl-ethene	10.65	10.64	0.001	166	168	292.262	9.80	9.8	97	?
60	chlorobenzene	11.58	11.57	0.000	112	77	691.236	9.26	9.3	91	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	245.258	10.28	10.3	100	
<<< I3 : ISTD ID = 62 >>>											
63	1-chlorohexane	11.56	11.56	0.000	93	55	118.729	7.23	7.2	96	?
64	Et-Bz	11.70	11.69	0.000	91	106	1133.719	7.90	7.9	95	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	1773.476	16.10	16.1	96	
66	styrene	12.24	12.23	0.000	104	78	695.954	8.27	8.3	91	?
67	o-xylene	12.22	12.22	0.000	91	106	917.264	8.12	8.1	97	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	186.045	9.39	9.4	99	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	56.224	9.77	9.8	98	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	1164.869	8.34	8.3	98	
72	bromobenzene	12.89	12.89	0.000	156	158	293.172	9.42	9.4	97	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	30.509	10.30	10.3	90	?
73	n-propylbenzene	13.00	12.99	0.000	120	78	353.478	8.59	8.6	94	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	309.252	8.90	8.9	98	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	308.786	8.67	8.7	96	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	1005.234	8.38	8.4	97	?
77	4-Iso-Pr-toluene	13.81	13.81	0.000	119	134	1111.631	8.52	8.5	99	?
78	124-tri-Me-Benzen	13.51	13.51	0.000	105	120	1026.195	8.33	8.3	98	?
79	tert-butylbenzene	13.47	13.47	0.000	119	91	1030.129	9.57	9.6	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06


Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00010.D Sample : f=1 10 ppb
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Jul 22 21:37 2003 RF via : Multiple Level Calibration
 Method Update: Mon Jul 21 15:01 2003 Operator: zou
 Quant. Time : Jul 24 11:58 2003 Multiplr: 1.000000
 Print Time : Thu Jul 24 11:58 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	605.258	9.08	9.1	99	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	1334.960	8.44	8.4	97	
82	14-DCB	13.87	13.87	0.000	146	148	594.167	8.88	8.9	98	
83	Cl-benzy1	13.98	13.98	0.000	126	91	63.403	12.97	13.0	76	#
84	12-DCB	14.21	14.21	0.000	146	148	540.526	9.41	9.4	100	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	298.215	8.75	8.8	90	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	37.194	11.39	11.4	99	
87	124-tr1-Cl-Bz	15.58	15.58	0.000	180	182	360.446	8.69	8.7	98	
88	naphthalene	15.79	15.78	0.000	128	129	581.293	9.57	9.6	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	209.577	8.47	8.5	96	
90	123-Tr1-Cl-Bz	15.98	15.98	0.000	180	182	317.007	9.19	9.2	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00010.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 21:37 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 11:58 2003
 Print Time : Thu Jul 24 11:58 2003
 Miscellaneous :

Sample : F=1 10 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000



Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00020.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 22:56 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:01 2003
 Print Time : Thu Jul 24 12:01 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0, ppb	C, ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.64	7.63	0.002	96	70	1015.499	10.00		0.02	
47	Chlorobenzene-d5	11.54	11.54	0.000	117	82	772.935	10.00		0.00	
62	1,4-Dichlorobenze	13.84	13.84	0.000	152	150	393.961	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	486.919	20.83		20.8	104.16%
29	1,2-Di-Cl-Et-d4	7.10	7.08	0.001	65	102	388.784	17.93		17.9	89.64%
55	toluene-d8	9.91	9.89	0.000	98	100	1923.324	19.54		19.5	97.70%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	629.095	20.98		21.0	104.91%

Target Compounds	<<< I1 : ISTD ID = 1 >>>	1.89	1.85	0.005	85	87	508.778	20.12	20.1	100	Qvalue
3	di-Cl-di-F-methan	1.89	1.85	0.005	85	87	508.778	20.12	20.1	100	
4	Chloromethane	2.11	2.07	0.006	50	52	448.729	16.85	16.9	99	
2	F114	2.03	2.00	0.005	85	135	275.940	20.92	20.9	51	
5	vinyl chloride	2.23	2.19	0.005	62	64	484.855	19.18	19.2	98	
6	bromomethane	2.61	2.58	0.005	94	96	228.520	15.58	15.6	97	
7	chloroethane	2.73	2.70	0.004	64	66	266.250	16.22	16.2	0	
8	tri-Cl-F-methane	3.03	3.00	0.005	101	103	724.847	19.09	19.1	98	
91	Acetonitrile X10	4.07	4.04	0.004	41	40	845.923	262.40	262.4	90	
9	acrolein X10	3.51	3.48	0.004	56	55	513.017	208.95	209.0	0	
11	acetone X10	3.70	3.69	0.002	43	58	497.276	191.27	191.3	99	
12	ethyl ether X5	3.37	3.34	0.005	59	74	1189.867	92.06	92.1	90	
13	11-dichloroethene	3.64	3.60	0.005	61	96	562.011	18.08	18.1	0	
14	Iodomethane	3.82	3.78	0.005	142	127	532.902	31.25	31.3	95	
15	F-113	3.65	3.62	0.005	101	151	427.782	18.88	18.9	90	
16	acrylonitrile X10	4.52	4.49	0.003	53	52	831.510	241.48	241.5	98	
17	carbon disulfide	3.90	3.87	0.004	76	78	1480.814	16.51	16.5	99	
94	Isopropyl Alcohol	3.98	4.01	-0.004	45	43	124.874	409.68	409.7	100	
18	methylene chlorid	4.22	4.19	0.005	84	49	510.536	19.30	19.3	99	
19	t-12-di-Cl-ethene	4.58	4.55	0.004	96	61	531.623	20.55	20.5	95	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Quantitation Report: **Applied P & Ch Lab**

EPA 524.2

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00020.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 22:56 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:01 2003
 Print Time : Thu Jul 24 12:01 2003
 Miscellaneous :

Sample : F=1 20 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0, ppb	C, ppb	Quality	Note
20	t-Bu-Me-ether	4.59	4.56	0.004	73	57	866.584	18.38	18.4	94	?
95	Tert butyl alcoho	4.45	4.47	-0.002	59	57	205.914	252.58	252.6	100	me
94	allyl chloride	4.07	4.04	0.004	41	76	845.923	22.41	22.4	83	#? 07/14/03
21	11-dichloroethane	5.12	5.09	0.004	63	83	825.889	19.20	19.2	99	
97	propionitrile	6.01	5.99	0.002	54	51	33.328	23.83	23.8	100	#
22	c-12-di-Cl-ethene	5.92	5.89	0.003	96	61	533.706	20.71	20.7	91	?
23	22-Dichloropropan	5.92	5.89	0.004	77	97	589.735	18.85	18.9	95	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	250.220	24.55	24.5	100	
25	chloroform	6.38	6.35	0.003	83	85	838.697	19.34	19.3	99	
26	tetrahydrofuranX5	6.34	6.32	0.002	42	72	318.868	109.93	109.9	94	
98	Disopropyl ether	5.24	5.22	0.003	45	87	1326.012	18.97	19.0	97	
99	ETBE	5.74	5.72	0.003	59	87	1064.815	19.83	19.8	96	
30	12-dichloroethane	7.22	7.20	0.003	64	62	142.937	17.93	17.9	99	?
32	vinyl acetate X5	5.19	5.17	0.003	43	86	2957.025	94.18	94.2	99	
92	Nitro Methane(x10	5.81	5.80	0.000	61	46	111.334	47.43	47.4	78	
33	2-butanoneMEK X10	5.93	5.92	0.002	43	72	875.377	272.92	272.9	94	?
93	Ethyl Acetate x2	6.04	6.02	0.002	43	61	434.881	32.82	32.8	91	#
34	111-trichloroetha	6.65	6.63	0.002	97	99	785.718	20.25	20.2	100	
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	660.985	19.83	19.8	92	?
36	benzene	7.21	7.19	0.003	78	52	1995.763	19.54	19.5	100	?
37	CCl4	6.91	6.89	0.002	117	119	732.441	22.36	22.4	99	?
100	Isobutyl alcohol	7.14	7.39	-0.034	43	42	87.337	432.72	432.7	93	me
38	thiophene	7.53	7.51	0.002	84	58	1025.220	19.84	19.8	99	07/14/03
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	446.632	20.39	20.4	93	
40	trichloroethene	8.24	8.23	0.002	130	132	633.192	22.70	22.7	100	
41	dibromomethane	8.73	8.71	0.002	174	172	270.225	25.52	25.5	96	
101	TAME	7.41	7.39	0.002	73	43	938.326	19.51	19.5	99	
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	604.352	20.02	20.0	100	
43	Me-methacrylate	8.76	8.75	0.002	69	100	228.095	22.09	22.1	93	
44	2-ClEt-Vi-ether10	9.38	9.37	0.000	63	43	514.977	223.40	223.4	98	
45	c-13-di-Cl-propen	9.56	9.55	0.002	75	110	676.158	20.89	20.9	94	
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	547.714	21.91	21.9	93	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00020.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 22:56 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:01 2003
 Print Time : Thu Jul 24 12:01 2003
 Miscellaneous :

Sample : f=1 20 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<< 12	: ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	343.401	21.35	21.4	94	
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	552.557	19.42	19.4	100	
50	Et methacrylate	10.38	10.37	0.000	69	99	443.099	19.68	19.7	94	?
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	456.818	23.12	23.1	100	
52	bromoform	12.42	12.41	0.000	173	174	263.988	25.03	25.0	100	
53	1,4-dichlorobutan	12.67	12.67	0.000	55	41	510.059	19.84	19.8	94	
54	MIBK	9.77	9.76	0.000	43	58	222.364	20.36	20.4	92	
56	toluene	9.99	9.98	0.000	91	92	2227.432	19.25	19.3	99	
57	2-hexanone X5	10.76	10.75	0.000	43	58	728.526	100.50	100.5	94	
58	12-dibromoethane	11.03	11.03	0.000	107	109	343.609	22.79	22.8	96	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	659.769	22.17	22.2	99	
60	chlorobenzene	11.58	11.57	0.000	112	77	1494.861	20.06	20.1	92	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	525.072	22.05	22.0	99	
<<< 13	: ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	280.119	17.67	17.7	97	?
64	Et-Bz	11.70	11.69	0.000	91	106	2529.977	18.24	18.2	96	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	3842.318	36.10	36.1	97	
66	styrene	12.24	12.23	0.000	104	78	1498.467	18.43	18.4	92	?
67	o-xylene	12.22	12.22	0.000	91	106	1990.843	18.25	18.2	96	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	374.484	19.56	19.6	98	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	114.784	20.64	20.6	100	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	2648.278	19.63	19.6	98	
72	bromobenzene	12.89	12.89	0.000	156	158	635.461	21.14	21.1	99	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	64.587	22.57	22.6	85	?
73	n-propylbenzene	13.00	12.99	0.000	120	78	802.663	20.20	20.2	93	?
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	685.694	20.42	20.4	98	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	677.984	19.70	19.7	99	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	2217.776	19.13	19.1	97	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	2489.320	19.74	19.7	97	?
78	124-tri-Me-Benzen	13.51	13.51	0.000	105	120	2248.874	18.90	18.9	96	?
79	tert-butylbenzene	13.47	13.47	0.000	119	91	2077.031	19.97	20.0	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Quantitation Report: **Applied P & Ch Lab** EPA 524.2


Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00020.D Sample : f=1 20 ppb
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Jul 22 22:56 2003 RF via : Multiple Level Calibration
 Method Update: Mon Jul 21 15:01 2003 Operator: zou
 Quant. Time : Jul 24 12:01 2003 Multiplr: 1.000000
 Print Time : Thu Jul 24 12:01 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	1313.556	20.40	20.4	100	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	3039.109	19.88	19.9	98	
82	14-DCB	13.87	13.87	0.000	146	148	1295.125	20.04	20.0	99	
83	Cl-benzy1	13.98	13.98	0.000	126	91	132.793	28.12	28.1	76	#
84	12-DCB	14.21	14.21	0.000	146	148	1153.351	20.77	20.8	99	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	689.097	20.94	20.9	84	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	80.335	25.47	25.5	99	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	838.928	20.94	20.9	99	
88	naphthalene	15.79	15.78	0.000	128	129	1300.980	22.16	22.2	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	472.099	19.75	19.8	98	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	719.921	21.61	21.6	98	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00020.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 22 22:56 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:01 2003
 Print Time : Thu Jul 24 12:02 2003
 Miscellaneous :

Sample : F=1 20 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000



Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00040.D
Method : C:\MSDCHEM\1\METHODS\E524A002.M
Acq. Time : Jul 23 00:43 2003
Method Update: Mon Jul 21 15:01 2003
Quant. Time : Jul 24 12:04 2003
Print Time : Thu Jul 24 12:05 2003
Miscellaneous :
Sample : F=1 40 ppb
Inst. : GCMS-A
RF via : Multiple Level Calibration
Operator: zou
Multiplr: 1.000000

2487

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	990.930	10.00		0.02	
47	Chlorobenzene-d5	11.54	11.54	0.000	117	82	739.886	10.00		0.00	
62	1,4-Dichlorobenze	13.85	13.84	0.000	152	150	367.099	10.00		0.00	

System Monitoring Compounds (Surrogate)											
ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
27	Di-Br-F-Me (sur)	6.60	6.58	0.002	111	113	949.109	41.61		41.6	208.07%
29	1,2-Di-Cl-Et-d4 (7.11	7.08	0.002	65	102	757.790	35.81		35.8	179.05%
55	toluene-d8	9.91	9.89	0.000	98	100	3719.816	39.48		39.5	197.40%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	1208.574	43.26		43.3	216.30%

Target Compounds											
ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
3	di-Cl-di-F-methan	1.89	1.85	0.005	85	87	990.705	40.14		40.1	99
4	Chloromethane	2.11	2.07	0.006	50	52	864.850	33.28		33.3	98
2	F114	2.03	2.00	0.005	85	135	544.934	42.35		42.3	45
5	vinyl chloride	2.22	2.19	0.004	62	64	949.966	38.50		38.5	100
6	bromomethane	2.61	2.58	0.004	94	96	411.844	28.77		28.8	98
7	chloroethane	2.73	2.70	0.004	64	66	546.077	34.09		34.1	0
8	tri-Cl-F-methane	3.03	3.00	0.005	101	103	1412.342	38.12		38.1	98
91	Acetonitrile X10	4.07	4.04	0.004	41	40	1639.456	521.17		521.2	95
9	acrolein X10	3.51	3.48	0.004	56	55	937.791	391.44		391.4	0
11	acetone X10	3.70	3.69	0.000	43	58	943.071	371.74		371.7	100
12	ethyl ether X5	3.37	3.34	0.005	59	74	2188.257	173.50		173.5	91
13	11-dichloroethene	3.64	3.60	0.005	61	96	1088.836	35.91		35.9	0
14	Iodomethane	3.82	3.78	0.005	142	127	1022.337	61.44		61.4	94
15	F-113	3.65	3.62	0.005	101	151	814.332	36.83		36.8	0
16	acrylonitrile X10	4.52	4.49	0.003	53	52	1601.521	476.63		476.6	98
17	carbon disulfide	3.90	3.87	0.004	76	78	2862.257	32.71		32.7	100
94	Isopropyl Alcohol	3.92	4.01	-0.012	45	43	265.806	893.67		893.7	100
18	methylene chlorid	4.22	4.19	0.004	84	49	960.233	37.20		37.2	94
19	t-12-di-Cl-ethene	4.58	4.55	0.004	96	61	1015.155	40.21		40.2	93

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00040.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 00:43 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:04 2003
 Print Time : Thu Jul 24 12:05 2003
 Miscellaneous :

Sample : f=1 40 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.59	4.56	0.004	73	57	1693.079	36.81	36.8	97	?
95	Tert butyl alcoho	4.41	4.47	-0.008	59	57	482.970	607.12	607.1	100	?
94	allyl chloride	4.07	4.04	0.004	41	76	1216.384	33.02	33.0	99	?
21	11-dichloroethane	5.12	5.09	0.004	63	83	1598.207	38.08	38.1	100	?
97	propionitrile	6.00	5.99	0.000	54	51	64.498	47.25	47.3	100	#
22	c-12-di-Cl-ethene	5.92	5.89	0.003	96	61	1019.103	40.53	40.5	93	?
23	22-Dichloropropan	5.92	5.89	0.003	77	97	1098.975	36.00	36.0	96	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	483.734	48.64	48.6	98	?
25	chloroform	6.37	6.35	0.002	83	85	1597.118	37.75	37.8	97	?
26	tetrahydrofuranx5	6.34	6.32	0.002	42	72	624.422	220.61	220.6	95	?
98	Diisopropyl ether	5.24	5.22	0.003	45	87	2521.921	36.97	37.0	97	?
99	ETBE	5.74	5.72	0.003	59	87	2106.754	40.20	40.2	97	?
30	12-dichloroethane	7.22	7.20	0.003	64	62	277.276	35.64	35.6	96	?
32	vinyl acetate X5	5.19	5.17	0.003	43	86	5653.569	184.52	184.5	99	?
92	Nitro Methane(X10	5.79	5.80	0.000	61	46	237.578	103.73	103.7	79	?
33	2-butanoneMEK X10	5.93	5.92	0.002	43	72	1648.493	526.70	526.7	98	?
93	Ethyl Acetate x2	6.04	6.02	0.002	43	61	845.931	65.43	65.4	88	#
34	111-trichloroetha	6.65	6.63	0.002	97	99	1524.561	40.26	40.3	100	?
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	1285.689	39.53	39.5	91	?
36	benzene	7.21	7.19	0.003	78	52	3805.709	38.18	38.2	98	?
37	CCl4	6.91	6.89	0.003	117	119	1420.059	44.44	44.4	100	?
100	Isobutyl alcohol	7.11	7.39	-0.037	43	42	182.637	927.34	927.3	95	?
38	thiophene	7.53	7.51	0.003	84	58	1999.457	39.65	39.7	98	?
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	875.952	40.98	41.0	93	?
40	trichloroethene	8.24	8.23	0.002	130	132	1246.934	45.81	45.8	99	?
41	dibromomethane	8.73	8.71	0.002	174	172	530.196	51.31	51.3	100	?
101	TAME	7.41	7.39	0.002	73	43	1872.529	39.90	39.9	99	?
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	1181.950	40.13	40.1	100	?
43	Me-methacrylate	8.76	8.75	0.002	69	100	464.768	46.12	46.1	92	?
44	2-ClEt-Vl-ether10	9.38	9.37	0.000	63	43	1130.652	502.64	502.6	98	?
45	c-13-di-Cl-propen	9.56	9.55	0.002	75	110	1328.342	42.05	42.1	93	?
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	1100.371	45.10	45.1	93	?

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

me
07/24/03

me
07/24/03

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00040.D Sample : f=1 40 ppb
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Jul 23 00:43 2003 RF via : Multiple Level Calibration
 Method Update: Mon Jul 21 15:01 2003 Operator: zou
 Quant. Time : Jul 24 12:04 2003 Multiplr: 1.000000
 Print Time : Thu Jul 24 12:05 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<<	I2 : ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	662.895	43.06	43.1	96	
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	1062.190	39.00	39.0	100	?
50	Et methacrylate	10.38	10.37	0.000	69	99	890.456	41.31	41.3	94	
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	899.440	47.55	47.6	100	
52	bromoform	12.42	12.41	0.000	173	174	534.402	52.94	52.9	100	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	1005.306	40.86	40.9	94	
54	MIBK	9.76	9.76	0.000	43	58	457.366	43.74	43.7	94	
56	toluene	9.99	9.98	0.001	91	92	4284.454	38.69	38.7	98	
57	2-hexanone X5	10.76	10.75	0.000	43	58	1453.530	209.46	209.5	94	
58	12-dibromoethane	11.03	11.03	0.000	107	109	688.415	47.70	47.7	97	
59	tetra-Cl-ethene	10.65	10.64	0.001	166	168	1256.084	44.09	44.1	98	?
60	chlorobenzene	11.58	11.57	0.000	112	77	2837.497	39.78	39.8	92	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	1031.299	45.23	45.2	99	
<<<	I3 : ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	548.810	37.15	37.1	97	?
64	Et-Bz	11.70	11.69	0.000	91	106	4833.949	37.41	37.4	94	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	7112.016	71.71	71.7	95	
66	styrene	12.24	12.23	0.000	104	78	2811.895	37.11	37.1	94	?
67	o-xylene	12.22	12.22	0.000	91	106	3763.083	37.01	37.0	96	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	736.112	41.27	41.3	99	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	222.436	42.93	42.9	99	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	5083.237	40.43	40.4	97	
72	bromobenzene	12.89	12.89	0.000	156	158	1236.729	44.14	44.1	98	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	137.632	51.61	51.6	83	#?
73	n-propylbenzene	13.00	12.99	0.000	120	78	1558.540	42.09	42.1	93	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	1328.259	42.46	42.5	99	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	1291.985	40.28	40.3	99	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	4186.569	38.76	38.8	97	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	4727.712	40.24	40.2	98	?
78	124-tri-Me-Benzen	13.52	13.51	0.000	105	120	4305.294	38.84	38.8	95	
79	tert-butylbenzene	13.47	13.47	0.000	119	91	4034.706	41.63	41.6	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06


Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00040.D Sample : f=1 40 ppb
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Jul 23 00:43 2003 RF via : Multiple Level Calibration
 Method Update: Mon Jul 21 15:01 2003 Operator: zou
 Quant. Time : Jul 24 12:04 2003 Multiplr: 1.000000
 Print Time : Thu Jul 24 12:05 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	2467.081	41.12	41.1	100	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	5779.243	40.56	40.6	97	
82	14-DCB	13.87	13.87	0.000	146	148	2472.135	41.06	41.1	100	
83	Cl-benzy1	13.98	13.98	0.000	126	91	265.795	60.40	60.4	76	#
84	12-DCB	14.21	14.21	0.000	146	148	2171.646	41.97	42.0	100	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	1335.543	43.55	43.6	81	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	173.925	59.18	59.2	99	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	1658.990	44.44	44.4	99	
88	naphthalene	15.79	15.78	0.000	128	129	2689.379	49.16	49.2	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	923.440	41.46	41.5	98	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	1423.460	45.86	45.9	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00040.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 00:43 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:04 2003
 Print Time : Thu Jul 24 12:05 2003
 Miscellaneous :

Sample : F=1 40 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000



Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00060.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 02:45 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:07 2003
 Print Time : Thu Jul 24 12:07 2003
 Miscellaneous :

Sample : F=1 60 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0, ppb	C, ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.64	7.63	0.002	96	70	980.047	10.00		0.01	
47	Chlorobenzene-d5	11.54	11.54	0.000	117	82	736.062	10.00		0.00	
62	1,4-Dichlorobenze	13.85	13.84	0.000	152	150	365.124	10.00		0.00	

System Monitoring Compounds (Surrogate)											
ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0, ppb	C, ppb	Quality	Note
27	Di-Br-F-Me (surr)	6.59	6.58	0.000	111	113	1428.311	63.32		63.3	316.60%
29	1,2-Di-Cl-Et-d4	7.10	7.08	0.000	65	102	1126.883	53.84		53.8	269.21%
55	toluene-d8	9.90	9.89	0.000	98	100	5400.714	57.62		57.6	288.09%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	1766.234	63.56		63.6	317.81%

Target Compounds											
ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0, ppb	C, ppb	Quality	Note
3	di-Cl-di-F-methan	1.88	1.85	0.004	85	87	1444.214	59.17		59.2	99
4	Chloromethane	2.14	2.07	0.010	50	52	1174.857	45.71		45.7	99
2	F114	2.03	2.00	0.004	85	135	783.345	61.55		61.5	12
5	vinyl chloride	2.22	2.19	0.003	62	64	1312.837	53.80		53.8	99
6	bromomethane	2.59	2.58	0.002	94	96	626.111	44.22		44.2	97
7	chloroethane	2.71	2.70	0.002	64	66	758.703	47.89		47.9	0
8	tri-Cl-F-methane	3.01	3.00	0.002	101	103	1986.417	54.21		54.2	98
91	Acetonitrile X10	4.05	4.04	0.002	41	40	2365.932	760.46		760.5	95
9	acrolein X10	3.49	3.48	0.002	56	55	1313.901	554.52		554.5	0
11	acetone X10	3.67	3.69	-0.002	43	58	1241.915	494.97		495.0	99
12	ethyl ether X5	3.35	3.34	0.002	59	74	2883.396	231.16		231.2	92
13	1,1-dichloroethene	3.62	3.60	0.002	61	96	1490.545	49.70		49.7	0
14	Iodomethane	3.80	3.78	0.002	142	127	1206.486	73.32		73.3	95
15	F-113	3.63	3.62	0.002	101	151	1086.062	49.67		49.7	87
16	acrylonitrile X10	4.50	4.49	0.000	53	52	2352.958	708.05		708.0	99
17	carbon disulfide	3.88	3.87	0.002	76	78	3817.157	44.11		44.1	99
94	Isopropyl Alcohol	3.87	4.01	-0.019	45	43	369.178	1255.00		1255.0	100
18	methylene chlorid	4.20	4.19	0.002	84	49	1435.510	56.24		56.2	95
19	t-12-di-Cl-ethene	4.57	4.55	0.002	96	61	1488.348	59.61		59.6	92

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Handwritten signatures and initials:
 m? *[Signature]*
 m? *[Signature]*
 m? *[Signature]*
 m? *[Signature]*
 m? *[Signature]*
 m? *[Signature]*

Quantitation Report: **Applied P & Ch Lab** EPA 524.2

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00060.D Sample : f=1 60 ppb
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Jul 23 02:45 2003 RF via : Multiple Level Calibration
 Method Update: Mon Jul 21 15:01 2003 Operator: zou
 Quant. Time : Jul 24 12:07 2003 Multiplr: 1.000000
 Print Time : Thu Jul 24 12:07 2003
 Miscelaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.58	4.56	0.002	73	57	2598.506	57.12	57.1	96	?
95	Tert butyl alcoho	4.37	4.47	-0.013	59	57	785.022	997.78	997.8	100	?
94	allyl chloride	4.05	4.04	0.002	41	76	1660.968	45.58	45.6	93	?
21	11-dichloroethane	5.11	5.09	0.002	63	83	2404.444	57.92	57.9	99	?
97	propionitrile	5.98	5.99	-0.002	54	51	101.359	75.08	75.1	100	#
22	c-12-di-Cl-ethene	5.90	5.89	0.002	96	61	1499.328	60.28	60.3	93	?
23	22-Dichloropropan	5.90	5.89	0.002	77	97	1539.718	51.00	51.0	96	?
24	Br-Cl-methane	6.24	6.23	0.002	128	130	716.534	72.84	72.8	99	?
25	chloroform	6.36	6.35	0.000	83	85	2387.736	57.07	57.1	97	?
26	tetrahydrofuranX5	6.32	6.32	0.000	42	72	946.730	338.19	338.2	95	?
98	Disopropyl ether	5.23	5.22	0.002	45	87	3767.835	55.85	55.8	97	?
99	ETBE	5.73	5.72	0.002	59	87	3231.163	62.34	62.3	97	?
30	12-dichloroethane	7.21	7.20	0.002	64	62	411.282	53.45	53.5	98	?
32	vinyl acetate X5	5.18	5.17	0.002	43	86	8226.056	271.47	271.5	100	?
92	Nitro Methane(x10	5.78	5.80	-0.003	61	46	325.914	143.88	143.9	78	?
33	2-butanoneMEK X10	5.92	5.92	0.000	43	72	2411.259	778.97	779.0	98	?
93	Ethyl Acetate x2	6.03	6.02	0.000	43	61	1295.021	101.28	101.3	87	#
34	111-trichloroetha	6.64	6.63	0.000	97	99	2222.673	59.34	59.3	98	?
35	11-Di-Cl-propene	6.89	6.88	0.000	75	110	1859.527	57.81	57.8	90	?
36	benzene	7.20	7.19	0.002	78	52	5575.015	56.55	56.6	98	?
37	CCl4	6.90	6.89	0.002	117	119	2060.040	65.18	65.2	99	?
100	Isobutyl alcohol	7.08	7.39	-0.041	43	42	287.972	1478.42	1478.4	93	?
38	thiophene	7.52	7.51	0.002	84	58	2960.047	59.36	59.4	99	?
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	1293.157	61.17	61.2	94	?
40	trichloroethene	8.23	8.23	0.000	130	132	1815.606	67.45	67.4	98	?
41	dibromomethane	8.72	8.71	0.000	174	172	793.122	77.61	77.6	99	?
101	TAME	7.41	7.39	0.002	73	43	2927.534	63.08	63.1	98	?
42	Br-di-Cl-methane	8.95	8.95	0.000	83	85	1745.648	59.93	59.9	99	?
43	Me-methacrylate	8.76	8.75	0.000	69	100	714.088	71.64	71.6	92	?
44	2-ClEt-Vi-ether10	9.38	9.37	0.000	63	43	198.558	89.25	89.3	97	?
45	c-13-di-Cl-propen	9.55	9.55	0.000	75	110	1977.341	63.30	63.3	94	?
46	t-1,3-dichloropro	10.24	10.24	0.000	75	110	1653.172	68.51	68.5	92	?

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00060.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 02:45 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:07 2003
 Print Time : Thu Jul 24 12:07 2003
 Miscellaneous :

Sample : F=1 60 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
<<<	I2 : ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.45	10.45	0.000	97	83	997.315	65.11	65.1	96	
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	1558.760	57.54	57.5	99	
50	Et methacrylate	10.38	10.37	0.000	69	99	1418.605	66.15	66.2	94	?
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	1365.305	72.56	72.6	100	
52	bromoforn	12.41	12.41	0.000	173	174	823.804	82.04	82.0	100	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	1519.089	62.06	62.1	94	
54	MIBK	9.76	9.76	0.000	43	58	733.371	70.50	70.5	92	
56	toluene	9.99	9.98	0.000	91	92	6235.319	56.59	56.6	98	
57	2-hexanone X5	10.75	10.75	0.000	43	58	2205.707	319.50	319.5	93	
58	12-dibromoethane	11.03	11.03	0.000	107	109	1038.913	72.36	72.4	97	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	1800.306	63.52	63.5	98	?
60	chlorobenzene	11.57	11.57	0.000	112	77	4059.222	57.21	57.2	95	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	1537.036	67.77	67.8	99	
<<<	I3 : ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	784.184	53.36	53.4	99	?
64	Et-Bz	11.70	11.69	0.000	91	106	6913.463	53.79	53.8	94	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	9936.549	100.73	100.7	95	
66	styrene	12.24	12.23	0.000	104	78	3981.902	52.84	52.8	95	?
67	o-xylene	12.22	12.22	0.000	91	106	5341.093	52.81	52.8	97	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	1101.133	62.07	62.1	99	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	332.760	64.57	64.6	97	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	7208.570	57.65	57.6	96	
72	bromobenzene	12.89	12.89	0.000	156	158	1774.887	63.70	63.7	100	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	207.953	78.40	78.4	83	#?
73	n-propylbenzene	13.00	12.99	0.000	120	78	2210.026	60.01	60.0	93	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	1912.837	61.48	61.5	98	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	1819.259	57.03	57.0	99	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	5867.539	54.61	54.6	97	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	6567.533	56.20	56.2	97	?
78	124-tri-Me-Benzen	13.52	13.51	0.000	105	120	6074.118	55.09	55.1	94	
79	tert-butylbenzene	13.47	13.47	0.000	119	91	5816.456	60.34	60.3	94	


= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00060.D Sample : f=1 60 ppb
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Jul 23 02:45 2003 RF via : Multiple Level Calibration
 Method Update: Mon Jul 21 15:01 2003 Operator: zou
 Quant. Time : Jul 24 12:07 2003 Multiplr: 1.000000
 Print Time : Thu Jul 24 12:07 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	3488.098	58.45	58.5	99	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	8084.394	57.05	57.0	96	
82	14-DCB	13.87	13.87	0.000	146	148	3533.637	59.01	59.0	99	
83	Cl-benzyl	13.98	13.98	0.000	126	91	361.085	82.50	82.5	75	#
84	12-DCB	14.21	14.21	0.000	146	148	3083.911	59.93	59.9	99	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	1846.727	60.55	60.5	79	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	263.112	90.01	90.0	99	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	2390.948	64.40	64.4	100	
88	naphthalene	15.79	15.78	0.000	128	129	4011.635	73.73	73.7	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	1288.580	58.17	58.2	99	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	2068.944	67.02	67.0	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3415\2-00060.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Jul 23 02:45 2003
 Method Update: Mon Jul 21 15:01 2003
 Quant. Time : Jul 24 12:07 2003
 Print Time : Thu Jul 24 12:07 2003
 Miscellaneous :
 Sample : F=1 60 ppb
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000




Data File : C:\MSDCHEM\1\DATA\03G3659\G3659P01.D
 Acq On : 11 Aug 2003 10:14 am

Sample : ##03G3659,w 50ng
 Misc :
 MS Integration Params: Lscint.P

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P &ch Lab** EPA 524.2

Vial : 1
 Operator : zou
 Inst : GCMS-A
 Multiplr : 1.00



Spectrum Information: Average of 12.729 to 12.742 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result
50	95	15	40	15.5	4665	PASS
75	95	30	60	44.1	13288	PASS
95	95	100	100	100.0	30159	PASS
96	95	5	9	7.5	2271	PASS
173	174	0.00	2	0.5	114	PASS
174	95	50	100	82.3	24811	PASS
175	174	5	9	8.0	1991	PASS
176	174	95	101	97.9	24285	PASS
177	176	5	9	6.3	1535	PASS

FORM-5A

Applied P & Ch Laboratory

Volatile Organic Instrument Performance Check for Method 524.2

Bromofluorobenzene (BFB), Part II

Client Name:	GEOFON, Inc.	Contract No:	Lab Code:	APCL
Case No:		SAS No:	Service ID:	034572
Project ID:	JPL	BFB Inj. Date:	Batch No:	03G3659
		BFB Inj. Time:	Sequence No:	03G3659
Project No:	04-4428.10	Instrument ID:	GC Column:	HP-VOC
Data File Name:	G3659P01	Heated Purge: (Y/N)	Column ID:	0.20 mm

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS AND STANDARDS:

#	Client Sample No	Lab Sample ID	Data File Name	Date Analyzed	Time Analyzed
1	03G3659-CCV-01	03G3659-CCV-01	G3659Q01	08/11/03	10:40
2	03G3659-LCS-01	03G3659-LCS-01	G3659L01	08/11/03	11:06
3	03G3659-MB-01	03G3659-MB-01	G3659K01	08/11/03	13:18
4	DUPE-5-3-Q03	03-4572-1	4572-01	08/11/03	16:24
5	EB-7-8-11-03	03-4572-2	4572-02	08/11/03	16:50
6	MW-12-1	03-4572-3	4572-03	08/11/03	17:17
7	MW-12-2	03-4572-4	4572-04	08/11/03	17:43
8	MW-12-3	03-4572-5	4572-05	08/11/03	18:10
9	MW-12-4	03-4572-6	4572-06	08/11/03	18:36
10	MW-12-5	03-4572-7	4572-07	08/11/03	19:03
11	MW-22-1	03-4572-8	4572-08	08/11/03	19:29
12	MW-22-2	03-4572-9	4572-09	08/11/03	19:55
13	MW-22-3	03-4572-10	4572-10	08/11/03	20:21
14	TB-7-8-11-03	03-4572-11	4572-11	08/11/03	20:47
15	MW-12-4MS	03-4572-6MS	G3659M02	08/11/03	21:14
16	MW-12-4MSD	03-4572-6MSD	G3659N02	08/11/03	21:40
17					
18					
19					
20					
21					
22					
23					
24					
25					

Continuing Calibration Concentration Summary

Data File G3659Q01

Method File E524A002

<u>Compound Name</u>	<u>Amount</u>	<u>Actual</u>	<u>Units</u>	<u>%Dev</u>	<u>Target Response</u>
1 Fluorobenzene	10	10.00	ppb	0.00	764133
3 di-Cl-di-F-methane	20	17.70	ppb	11.52	339443
4 Chloromethane	20	15.64	ppb	21.82	255103
2 F114	20	19.72	ppb	1.42	197749
5 vinyl chloride	20	17.40	ppb	12.98	322358
6 bromomethane	20	12.76	ppb	36.22	113517
7 chloroethane	20	18.70	ppb	6.52	206129
8 tri-Cl-F-methane	20	19.13	ppb	4.34	524924
91 Acetonitrile X10	200	174.31	ppb	12.84	584362
9 acrolein X10	200	197.96	ppb	1.02	360446
11 acetone X10	200	292.02	ppb	46.01	511834
12 ethyl ether X5	100	102.57	ppb	2.57	844649
13 11-dichloroethene	20	18.56	ppb	7.20	406590
14 Iodomethane	20	8.59	ppb	57.07	166238
15 F-113	20	21.13	ppb	5.64	318205
16 acrylonitrile X10	200	166.07	ppb	16.97	567276
17 carbon disulfide	20	17.54	ppb	12.31	976666
94 Isopropyl Alcoholx10	200	166.25	ppb	16.87	79517
18 methylene chloride	20	18.96	ppb	5.18	368132
19 t-12-di-Cl-ethene	20	17.88	ppb	10.62	368993
20 t-Bu-Me-ether	20	17.60	ppb	12.00	604643
95 Tert butyl alcoholx10	200	169.66	ppb	15.17	149992
94 allyl chloride	20	19.44	ppb	2.80	584362
21 11-dichloroethane	20	17.33	ppb	13.35	574058
97 propionitrile	20	18.62	ppb	6.90	23869
22 c-12-di-Cl-ethene	20	18.24	ppb	8.82	379501
23 22-Dichloropropane	20	24.79	ppb	23.94	516622
24 Br-Cl-methane	20	17.45	ppb	12.74	176116
25 chloroform	20	19.68	ppb	1.62	614292
26 tetrahydrofuranX5	100	82.15	ppb	17.85	206156
98 Diisopropyl ether	20	17.85	ppb	10.77	910176
27 Di-Br-F-Me (surr)	20	18.81	ppb	5.97	358122
99 ETBE	20	18.67	ppb	6.66	727874
29 1,2-Di-Cl-Et-d4 (S1)	20	18.49	ppb	7.56	285144
30 12-dichloroethane	20	19.93	ppb	0.37	107389
32 vinyl acetate X5	100	93.14	ppb	6.86	2220559
92 Nitro Methane(x10)	200	186.18	ppb	6.91	79006
33 2-butanoneMEK X10	200	210.61	ppb	5.30	679357
93 Ethyl Acetate x2	40	32.96	ppb	17.60	275070
34 111-trichloroethane	20	18.79	ppb	6.07	578308
35 11-Di-Cl-propene	20	19.70	ppb	1.52	458172
36 benzene	20	17.89	ppb	10.56	1387345
37 CCl4	20	19.69	ppb	1.57	558782

<u>Compound Name</u>	<u>Amount</u>	<u>Actual</u>	<u>Units</u>	<u>%Dev</u>	<u>Target Response</u>
100 Isobutyl.alcoholx10	200	158.40	ppb	20.80	55273
38 thiophene	20	18.62	ppb	6.90	717865
39 1,2-di-Cl-propane	20	17.61	ppb	11.94	305636
40 trichloroethene	20	18.51	ppb	7.44	447547
41 dibromomethane	20	17.43	ppb	12.86	191858
101 TAME	20	18.74	ppb	6.29	633612
42 Br-di-Cl-methane	20	19.18	ppb	4.10	435941
43 Me-methacrylate	20	16.44	ppb	17.82	144340
44 2-ClEt-Vi-ether10	200	98.28	ppb	50.86	183472
45 c-1,3-di-Cl-propene	20	19.55	ppb	2.27	492981
46 t-1,3-dichloropropene	20	20.01	ppb	0.07	401404
47 Chlorobezene-d5	10	10.00	ppb	0.00	595141
48 1,1,2-tri-Cl-Et	20	17.14	ppb	14.32	246121
49 1,3-di-Cl-propane	20	17.42	ppb	12.88	394744
50 Et methacrylate	20	17.26	ppb	13.72	309091

<u>Compound Name</u>	<u>Amount</u>	<u>Actual</u>	<u>Units</u>	<u>%Dev</u>	<u>Target Response</u>
51 di-Br-Cl-methane	20	18.64	ppb	6.78	335964
52 bromoform	20	17.42	ppb	12.89	192070
53 1,4-dichlorobutane-2	20	17.02	ppb	14.88	359028
54 MIBK	20	16.01	ppb	19.95	138771
55 toluene-d8	20	18.41	ppb	7.94	1380968
56 toluene	20	18.32	ppb	8.39	1579323
57 2-hexanone X5	100	86.48	ppb	13.52	508775
58 1,2-dibromoethane	20	17.13	ppb	14.35	242728
59 tetra-Cl-ethene	20	18.62	ppb	6.88	479686
60 chlorobenzene	20	19.60	ppb	2.02	1099490
61 1,1,1,2-tetra-Cl-Et	20	17.79	ppb	11.03	390340
62 1,4-Dichlorobenzene-d4	10	10.00	ppb	0.00	318055
63 1-chlorohexane	20	19.99	ppb	0.07	207296
64 Et-Bz	20	18.55	ppb	7.27	1845608
65 m/p-Xylenes X2	40	37.95	ppb	5.12	2855210
66 styrene	20	19.20	PPB	4.02	1097995
67 o-xylene	20	19.42	ppb	2.88	1466176
68 1,1,2,2-Tetra-Cl-Et	20	15.78	ppb	21.12	265064
69 1,2,3-tri-Cl-Pr	20	16.52	ppb	17.38	83112
70 4-Br-1-F-Bz (S3)	20	17.74	ppb	11.29	466731
71 isopropylbenzene	20	20.19	ppb	0.96	1952530
72 bromobenzene	20	18.09	ppb	9.54	469862
92 t-1,4-dichloro-2-butene	20	17.90	ppb	10.49	50934
73 n-propylbenzene	20	19.96	ppb	0.22	604443
74 2-Cl-Toluene	20	19.12	ppb	4.40	511745
75 4-Cl-Toluene	20	18.85	ppb	5.76	513142
76 1,3,5-tri-Me-Benzene	20	20.09	ppb	0.46	1672635
77 4-iso-Pr-toluene	20	20.52	ppb	2.59	1907530
78 1,2,4-tri-Me-Benzene	20	19.69	ppb	1.53	1707247
79 tert-butylbenzene	20	19.74	ppb	1.28	1569659
80 1,3-DCB	20	18.08	ppb	9.59	987293
81 sec-butylbenzene	20	20.38	ppb	1.88	2279521

82 14-DCB	20	18.83	ppb	5.84	977724
83 Cl-benzyl	20	24.78	ppb	23.91	130389
84 12-DCB	20	17.69	ppb	11.57	859875
85 n-butylbenzene	20	19.56	ppb	2.19	533156
86 12-diBr-2-Cl-Pra	20	16.72	ppb	16.42	57226
87 124-tri-Cl-Bz	20	20.19	ppb	0.96	635630
88 naphthalene	20	16.78	ppb	16.09	958708
89 hx-Cl-butadiene	20	19.46	ppb	2.70	370879
90 123-Tri-Cl-Bz	20	19.27	ppb	3.64	544416

Average D % 9.9459377

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G3659\G3659Q01.D
 Acq On : 11 Aug 2003 10:40 am
 Sample : f=1
 Misc :
 MS Integration Params: Lscint.p
 Vial: 2
 Operator: zou
 Inst : GCMS-A
 Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P &ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AVGRF	CCRF	%Dev Area	% Dev (min)
1 I	1.000	1.000	0.0	75
2	0.251	0.222	11.6	67
3 p	0.235	0.167	28.9#	57
4	0.131	0.129	1.5	72
5 C	0.242	0.211	12.8	66
6	0.116	0.074	36.2#	50#
7	0.144	0.135	6.2	77
8	0.359	0.343	4.5	72
9	0.044	0.038	13.6	69
10	0.028	0.024	14.3	70
11	0.043	0.033	23.3#	103
12	0.127	0.111	12.6	71
13 M, C	0.287	0.266	7.3	72
14	0.253	0.109	56.9#	31#
15	0.225	0.208	7.6	74
16	0.045	0.037	17.8	68
17	0.729	0.639	12.3	66
18	0.008	0.005	37.5#	64
19	0.467	0.241	48.4#	72
20	0.270	0.241	10.7	69
21	0.450	0.396	12.0	70
22	0.012	0.010	16.7	73
23	0.393	0.382	2.8	69
24 p	0.433	0.376	13.2	70
25	0.017	0.016	5.9	72
26	0.272	0.248	8.8	71

(#) = Out of Range
 G3659Q01.D E524A002.M Mon Aug 11 16:02:15 2003

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G3659\G3659Q01.D
 Acq On : 11 Aug 2003 10:40 am
 Sample : F=1
 Misc :
 MS Integration Params: Lscint.p

Vial: 2
 Operator: zou
 Inst : GCMS-A
 Multiplr: 1.00

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AVGRF	CCRF	%Dev Area	Dev(min)
27 22-Dichloropropane	0.310	0.338	-9.0	88
28 Br-Cl-methane	0.132	0.115	12.9	70
29 C 25 chloroform	0.478	0.402	15.9	73
30 26 tetrahydrofuranX5	0.033	0.027	18.2	65
31 98 Disopropyl ether	0.667	0.596	10.6	69
32 S 27 Di-Br-F-Me (surr)	0.249	0.234	6.0	74
33 99 ETBE	0.510	0.476	6.7	68
34 S 29 1,2-Di-Cl-Et-d4 (S1)	0.202	0.187	7.4	73
35 30 12-dichloroethane	0.078	0.070	10.3	75
36 32 vinyl acetate X5	0.312	0.291	6.7	75
37 92 Nitro Methane (X10)	0.006	0.005	16.7	71
38 33 2-butanoneMEK X10	0.052	0.044	15.4	78
39 93 Ethyl Acetate X2	0.124	0.090	27.4#	63
40 34 111-trichloroethane	0.403	0.378	6.2	74
41 35 11-Di-Cl-propene	0.304	0.300	1.3	69
42 M 36 benzene	1.015	0.908	10.5	70
43 37 CC14	0.371	0.366	1.3	76
44 100 Isobutyl alcoholX10	0.004	0.004	0.0	63
45 38 thiophene	0.505	0.470	6.9	70
46 C 39 12-di-Cl-propane	0.227	0.200	11.9	68
47 M 40 trichloroethene	0.316	0.293	7.3	71
48 41 dibromomethane	0.144	0.126	12.5	71
49 101 TAME	0.442	0.415	6.1	68
50 42 Br-di-Cl-methane	0.327	0.285	12.8	72
51 43 Me-methacrylate	0.102	0.094	7.8	63
52 44 2-ClEt-VI-ether10	0.020	0.012	40.0#	36#

(#) = Out of Range

G3659Q01.D E524A002.M Mon Aug 11 16:02:16 2003

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G3659\G3659Q01.D
 Acq On : 11 Aug 2003 10:40 am
 Sample : f=1
 Misc :
 MS Integration Params: Lscint.p

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Max. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AVGRF	CCRF	%Dev Area	% Dev(min)			
53 I	45	c-13-di-Cl-propene	0.330	0.323	2.1	73	0.01
54	46	t-1,3-dichloropropene	0.262	0.263	-0.4	73	0.00
55 I	47	Chlorobenzene-d5	1.000	1.000	0.0	77	0.00
56	48	112-tri-Cl-Et	0.241	0.207	14.1	72	0.00
57	49	13-di-Cl-propane	0.381	0.332	12.9	71	0.00
58	50	Et methacrylate	0.273	0.260	4.8	70	0.00
59	51	di-Br-Cl-methane	0.325	0.282	13.2	74	0.00
60 P	52	bromoform	0.185	0.161	13.0	73	0.00
61	53	1,4-dichlorobutane-2	0.354	0.302	14.7	70	0.00
62	54	MIBK	0.146	0.117	19.9	62	0.00
63 s	55	toluene-d8	1.260	1.160	7.9	72	0.01
64 M,C	56	toluene	1.448	1.327	8.4	71	0.00
65	57	2-hexanone X5	0.099	0.085	14.1	70	0.00
66	58	12-dibromoethane	0.238	0.204	14.3	71	0.00
67	59	tetra-Cl-ethene	0.433	0.403	6.9	73	0.00
68 M,P	60	chlorobenzene	1.032	0.924	10.5	74	0.00
69	61	1112-tetra-Cl-Et	0.369	0.328	11.1	74	0.00
70 I	62	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	81	0.00
71	63	1-chlorohexane	0.326	0.326	0.0	74	0.00
72 C	64	Et-Bz	3.129	2.901	7.3	73	0.00
73	65	m/p-Xylenes X2	2.365	2.244	5.1	74	0.00
74	66	styrene	1.798	1.726	4.0	73	0.00
75	67	o-xylene	2.373	2.305	2.9	74	0.00
76 p	68	1122-Tetra-Cl-Et	0.528	0.417	21.0#	71	0.00

(#) = Out of Range

G3659Q01.D E524A002.M Mon Aug 11 16:02:17 2003

Evaluate Continuing Calibration Report

Data File : C:\MSDCHEM\1\DATA\03G3659\G3659Q01.D
 Acq On : 11 Aug 2003 10:40 am
 Sample : f=1
 Misc :
 MS Integration Params: Lscint.p

Method : C:\MSDCHEM\1\METHODS\E524A002.M (RTE Integrator)
 Title : **Applied P & Ch Lab** EPA 524.2
 Last Update : Thu Jul 24 12:40:35 2003
 Response via : Multiple Level Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 200%

Compound	AVGRF	CCRF	%Dev Area	Dev (min)
77 69 123-tri-Cl-Pr	0.158	0.131	17.1	72 0.00
78 S 70 4-Br-1-F-Bz (S3)	0.827	0.734	11.2	74 0.00
79 71 isopropylbenzene	3.040	3.069	-1.0	74 0.00
80 72 bromobenzene	0.817	0.739	9.5	74 0.00
81 92 t-1,4-dichloro-2-butene	0.077	0.080	-3.9	79 0.00
82 73 n-propylbenzene	0.952	0.950	0.2	75 0.00
83 74 2-Cl-Toluene	0.841	0.804	4.4	75 0.00
84 75 4-Cl-Toluene	0.856	0.807	5.7	76 0.00
85 76 135-tri-Me-Benzene	2.617	2.629	-0.5	75 0.00
86 77 4-iso-Pr-toluene	2.923	2.999	-2.6	77 0.00
87 78 124-tri-Me-Benzene	2.726	2.684	1.5	76 0.00
88 79 tert-butylbenzene	2.499	2.468	1.2	76 0.00
89 80 13-DCB	1.717	1.552	9.6	75 0.00
90 81 sec-butylbenzene	3.517	3.584	-1.9	75 0.00
91 82 14-DCB	1.749	1.537	12.1	75 0.00
92 83 Cl-benzyl	0.165	0.205	-24.2#	98 0.00
93 84 12-DCB	1.529	1.352	11.6	75 0.00
94 85 n-butylbenzene	0.779	0.838	-7.6	77 0.00
95 86 12-diBr-2-Cl-Pra	0.108	0.090	16.7	71 0.00
96 87 124-tri-Cl-Bz	0.990	0.999	-0.9	76 0.00
97 88 naphthalene	1.796	1.507	16.1	74 0.00
98 89 hx-Cl-butadiene	0.599	0.583	2.7	79 0.00
99 90 123-Tri-Cl-Bz	0.888	0.856	3.6	76 0.00

(#) = Out of Range
 G3659Q01.D E524A002.M
 SPPC's out = 0
 CCC's out = 0
 Mon Aug 11 16:02:17 2003

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659Q01.D Sample : F=1
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Aug 11 10:40 2003 RF via : Multiple Level Calibration
 Method Update: Thu Jul 24 12:40 2003 Operator: zou
 Quant. Time : Aug 11 16:00 2003 Multiplr: 1.000000
 Print Time : Mon Aug 11 16:01 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
Internal Standards											
1	Fluorobenzene	7.65	7.63	0.003	96	70	764.133	10.00		0.02	
47	Chlorobenzene-d5	11.55	11.54	0.000	117	82	595.141	10.00		0.00	
62	1,4-Dichlorobenze	13.84	13.84	0.000	152	150	318.055	10.00		0.00	

System Monitoring Compounds (Surrogate)											
27	Di-Br-F-Me (surr)	6.60	6.58	0.002	111	113	358.122	18.81		18.8	94.03%
29	1,2-Di-Cl-Et-d4 (7.11	7.08	0.002	65	102	285.144	18.49		18.5	92.44%
55	toluene-d8	9.91	9.89	0.000	98	100	1380.968	18.41		18.4	92.06%
70	4-Br-1-F-Bz (S3)	12.74	12.74	0.000	174	95	466.731	17.74		17.7	88.71%

Target Compounds													
<<<	I1	: ISTD	ID = 1	>>>								Qvalue	
3	3	di-Cl-di-F-methan	1.89	1.85	0.005	85	87	339.443	17.70		17.7	100	
4	4	Chloromethane	2.11	2.07	0.005	50	52	255.103	15.64		15.6	99	
2	2	F114	2.04	2.00	0.006	85	135	197.749	19.72		19.7	55	
5	5	vinyl chloride	2.23	2.19	0.005	62	64	322.358	17.40		17.4	97	
6	6	bromomethane	2.61	2.58	0.005	94	96	113.517	12.76		12.8	99	
7	7	chloroethane	2.73	2.70	0.004	64	66	206.129	18.70		18.7	0	
8	8	tri-Cl-F-methane	3.04	3.00	0.006	101	103	524.924	19.13		19.1	98	
91	91	Acetonitrile X10	4.07	4.04	0.004	41	40	584.362	174.31		174.3	91	
9	9	acrolein X10	3.52	3.48	0.006	56	55	360.446	197.96		198.0	0	
11	11	acetone X10	3.73	3.69	0.005	43	58	511.834	292.02		292.0	0	
12	12	ethyl ether X5	3.37	3.34	0.005	59	74	844.649	102.57		102.6	91	
13	13	1,1-dichloroethene	3.64	3.60	0.005	61	96	406.590	18.56		18.6	0	
14	14	Iodomethane	3.82	3.78	0.005	142	127	166.238	8.59		8.6	97	
15	15	F-113	3.65	3.62	0.005	101	151	318.205	21.13		21.1	89	
16	16	acrylonitrile X10	4.52	4.49	0.004	53	52	567.276	166.07		166.1	98	
17	17	carbon disulfide	3.91	3.87	0.005	76	78	976.666	17.54		17.5	99	
94	94	Isopropyl Alcohol	4.11	4.01	0.013	45	43	79.517	166.25		166.3	100	
18	18	methylene chlorid	4.22	4.19	0.005	84	49	368.132	18.96		19.0	98	
19	19	t-12-di-Cl-ethene	4.58	4.55	0.005	96	61	368.993	17.88		17.9	92	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659Q01.D Sample : F=1
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Aug 11 10:40 2003 RF via : Multiple Level Calibration
 Method Update: Thu Jul 24 12:40 2003 Operator: zou
 Quant. Time : Aug 11 16:00 2003 Multiplr: 1.000000
 Print Time : Mon Aug 11 16:01 2003
 Miscellaneous :

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
20	t-Bu-Me-ether	4.60	4.56	0.005	73	57	604.643	17.60	17.6	96	?
95	Tert butyl alcoho	4.52	4.47	0.007	59	57	149.992	169.66	169.7	100	m 08/11/03
94	allyl chloride	4.07	4.04	0.004	41	76	584.362	19.44	19.4	86	#? 08/11/03
21	11-dichloroethane	5.12	5.09	0.004	63	83	574.058	17.33	17.3	99	
97	propionitrile	6.03	5.99	0.004	54	51	23.869	18.62	18.6	100	#?
22	C-12-di-Cl-ethene	5.92	5.89	0.003	96	61	379.501	18.24	18.2	90	?
23	22-Dichloropropan	5.92	5.89	0.004	77	97	516.622	24.79	24.8	99	?
24	Br-Cl-methane	6.25	6.23	0.003	128	130	176.116	17.45	17.5	99	
25	chloroform	6.38	6.35	0.003	83	85	614.292	19.68	19.7	98	
26	tetrahydrofuranx5	6.35	6.32	0.003	42	72	206.156	82.15	82.1	98	
98	Diisopropyl ether	5.24	5.22	0.003	45	87	910.176	17.85	17.8	98	
99	ETBE	5.74	5.72	0.003	59	87	727.874	18.67	18.7	97	
30	12-dichloroethane	7.22	7.20	0.003	64	62	107.389	19.93	19.9	97	?
32	vinyl acetate X5	5.20	5.17	0.004	43	86	2220.559	93.14	93.1	100	
92	Nitro Methane (x10	5.83	5.80	0.004	61	46	79.006	186.18	186.2	84	
33	2-butanoneMEK X10	5.95	5.92	0.003	43	72	679.357	210.61	210.6	96	?
93	Ethyl Acetate x2	6.04	6.02	0.003	43	61	275.070	32.96	33.0	88	#?
34	111-trichloroetha	6.65	6.63	0.002	97	99	578.308	18.79	18.8	100	
35	11-Di-Cl-propene	6.91	6.88	0.003	75	110	458.172	19.70	19.7	89	?
36	benzene	7.21	7.19	0.003	78	52	1387.345	17.89	17.9	99	?
37	CCl4	6.91	6.89	0.003	117	119	558.782	19.69	19.7	98	?
100	Isobutyl alcohol	7.18	7.39	-0.029	43	42	55.273	158.40	158.4	88	m 08/11/03
38	thiophene	7.53	7.51	0.002	84	58	717.865	18.62	18.6	96	
39	12-di-Cl-propane	8.56	8.54	0.002	63	76	305.636	17.61	17.6	100	
40	trichloroethene	8.24	8.23	0.002	130	132	447.547	18.51	18.5	98	
41	dibromomethane	8.73	8.71	0.002	174	172	191.858	17.43	17.4	100	
101	TAME	7.41	7.39	0.002	73	43	633.612	18.74	18.7	98	
42	Br-di-Cl-methane	8.96	8.95	0.002	83	85	435.941	19.18	19.2	100	
43	Me-methacrylate	8.76	8.75	0.002	69	100	144.340	16.44	16.4	94	
44	2-ClEt-Vi-ether10	9.38	9.37	0.000	63	43	183.472	98.28	98.3	97	
45	C-13-di-Cl-propen	9.56	9.55	0.002	75	110	492.981	19.55	19.5	94	
46	t-1,3-dichloropro	10.25	10.24	0.000	75	110	401.404	20.01	20.0	93	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659Q01.D Sample : f=1
 Method : C:\MSDCHEM\1\METHODS\E524A002.M Inst. : GCMS-A
 Acq. Time : Aug 11 10:40 2003 RF via : Multiple Level Calibration
 Method Update: Thu Jul 24 12:40 2003 Operator: zou
 Quant. Time : Aug 11 16:00 2003 Multiplr: 1.000000
 Print Time : Mon Aug 11 16:01 2003
 Miscellaneous :

2508

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	CO,ppb	C,ppb	Quality	Note
<<< 12	: ISTD ID = 47 >>>										
48	112-tri-Cl-Et	10.46	10.45	0.000	97	83	246.121	17.14	17.1	96	
49	13-di-Cl-propane	10.65	10.64	0.000	76	78	394.744	17.42	17.4	100	
50	Et methacrylate	10.38	10.37	0.000	69	99	309.091	17.26	17.3	94	?
51	di-Br-Cl-methane	10.91	10.90	0.000	129	127	335.964	18.64	18.6	100	
52	bromoform	12.42	12.41	0.000	173	174	192.070	17.42	17.4	100	
53	1,4-dichlorobutan	12.68	12.67	0.000	55	41	359.028	17.02	17.0	95	
54	MIBK	9.77	9.76	0.000	43	58	138.771	16.01	16.0	93	
56	toluene	9.99	9.98	0.000	91	92	1579.323	18.32	18.3	98	
57	2-hexanone X5	10.76	10.75	0.000	43	58	508.775	86.48	86.5	94	
58	12-dibromoethane	11.03	11.03	0.000	107	109	242.728	17.13	17.1	97	
59	tetra-Cl-ethene	10.64	10.64	0.000	166	168	479.686	18.62	18.6	99	?
60	chlorobenzene	11.58	11.57	0.000	112	77	1099.490	19.60	19.6	91	?
61	1112-tetra-Cl-Et	11.66	11.66	0.000	131	133	390.340	17.79	17.8	100	
<<< 13	: ISTD ID = 62 >>>										
63	1-chlorohexane	11.56	11.56	0.000	93	55	207.296	19.99	20.0	98	?
64	Et-Bz	11.70	11.69	0.000	91	106	1845.608	18.55	18.5	97	
65	m/p-Xylenes X2	11.82	11.82	0.000	91	106	2855.210	37.95	38.0	97	
66	styrene	12.24	12.23	0.000	104	78	1097.995	19.20	19.2	93	?
67	o-xylene	12.22	12.22	0.000	91	106	1466.176	19.42	19.4	97	?
68	1122-Tetra-Cl-Et	12.87	12.87	0.000	83	85	265.064	15.78	15.8	98	
69	123-tri-Cl-Pr	12.91	12.91	0.000	110	97	83.112	16.52	16.5	98	?
71	isopropylbenzene	12.60	12.59	0.000	105	120	1952.530	20.19	20.2	97	
72	bromobenzene	12.89	12.89	0.000	156	158	469.862	18.09	18.1	99	?
92	t-1,4-dichloro-2-	12.92	12.92	0.000	89	53	50.934	17.90	17.9	87	?
73	n-propylbenzene	13.00	12.99	0.000	120	78	604.443	19.96	20.0	92	
74	2-Cl-Toluene	13.08	13.08	0.000	126	128	511.745	19.12	19.1	99	
75	4-Cl-Toluene	13.19	13.18	0.000	126	128	513.142	18.85	18.8	99	?
76	135-tri-Me-Benzen	13.16	13.16	0.000	105	120	1672.635	20.09	20.1	98	?
77	4-iso-Pr-toluene	13.81	13.81	0.000	119	134	1907.530	20.52	20.5	97	?
78	124-tri-Me-Benzen	13.52	13.51	0.000	105	120	1707.247	19.69	19.7	96	
79	tert-butylbenzene	13.47	13.47	0.000	119	91	1569.659	19.74	19.7	94	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06


Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659Q01.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 10:40 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 11 16:00 2003
 Print Time : Mon Aug 11 16:01 2003
 Miscellaneous :

Sample : f=1
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplr: 1.000000

ID	Component Name	R.T.	RT0	DRRT	Qion	Q1	RF/1000	C0,ppb	C,ppb	Quality	Note
80	13-DCB	13.79	13.78	0.000	146	148	987.293	18.08	18.1	100	?
81	sec-butylbenzene	13.68	13.68	0.000	105	134	2279.521	20.38	20.4	98	
82	14-DCB	13.87	13.87	0.000	146	148	977.724	18.83	18.8	97	
83	Cl-benzyl	13.98	13.98	0.000	126	91	130.389	24.78	24.8	77	#
84	12-DCB	14.21	14.21	0.000	146	148	859.875	17.69	17.7	100	?
85	n-butylbenzene	14.18	14.18	0.000	134	91	533.156	19.56	19.6	88	#?
86	12-diBr-2-Cl-Pra	14.88	14.88	0.000	157	155	57.226	16.72	16.7	97	
87	124-tri-Cl-Bz	15.58	15.58	0.000	180	182	635.630	20.19	20.2	99	
88	naphthalene	15.79	15.78	0.000	128	129	958.708	16.78	16.8	100	
89	hx-Cl-butadiene	15.72	15.72	0.000	225	258	370.879	19.46	19.5	99	
90	123-Tri-Cl-Bz	15.98	15.98	0.000	180	182	544.416	19.27	19.3	99	

= qualifier out of range, m = manual integration, ? = RT coelution, * = DRRT > 0.06

Data Filename: C:\MSDCHEM\1\DATA\03G3659\G3659Q01.D
 Method : C:\MSDCHEM\1\METHODS\E524A002.M
 Acq. Time : Aug 11 10:40 2003
 Method Update: Thu Jul 24 12:40 2003
 Quant. Time : Aug 11 16:00 2003
 Print Time : Mon Aug 11 16:01 2003
 Miscellaneous :
 Sample : f=1
 Inst. : GCMS-A
 RF via : Multiple Level Calibration
 Operator: zou
 Multiplier: 1.000000



FORM-8A

Applied P & Ch Laboratory

Internal Standard Area and RT Summary for Method 524.2

Client Name: GEOFON, Inc.

Contract No:

Lab Code: APCL

Case No:

SAS No:

Service ID: 034572

Project ID: JPL

Project No: 04-4428.10

Sample Matrix: Water

CCV Data File: G3659Q01

Instrument ID: A

Batch No: 03G3659

#	Client Sample No	Lab Sample ID	Analysis Date & Time	IS-1		IS-2		IS-3	
				Area #	RT #	Area #	RT #	Area #	RT #
12 Hour CCV STD			08/11/03 10:40	764133	7.65	595141	11.55	318055	13.84
CCV Upper Limit				1528266	8.15	1190282	12.05	636110	14.34
CCV Lower Limit				382066	7.15	297570	11.05	159027	13.34
1	03G3659-LCS-01	03G3659-LCS-01	08/11/03 11:06	777095	7.65	607472	11.55	315375	13.84
2	03G3659-MB-01	03G3659-MB-01	08/11/03 13:18	681858	7.65	580347	11.54	305392	13.84
3	DUPE-5-3-Q03	03-4572-1	08/11/03 16:24	657170	7.65	565234	11.54	299145	13.84
4	EB-7-8-11-03	03-4572-2	08/11/03 16:50	655472	7.65	554083	11.55	294054	13.85
5	MW-12-1	03-4572-3	08/11/03 17:17	650799	7.65	557909	11.55	295228	13.85
6	MW-12-2	03-4572-4	08/11/03 17:43	652972	7.65	558405	11.54	293286	13.84
7	MW-12-3	03-4572-5	08/11/03 18:10	647873	7.65	555552	11.55	294270	13.85
8	MW-12-4	03-4572-6	08/11/03 18:36	645377	7.65	554567	11.55	294159	13.85
9	MW-12-5	03-4572-7	08/11/03 19:03	639092	7.65	551981	11.55	291989	13.84
10	MW-22-1	03-4572-8	08/11/03 19:29	641314	7.65	551289	11.55	290597	13.85
11	MW-22-2	03-4572-9	08/11/03 19:55	642809	7.65	553028	11.55	294253	13.85
12	MW-22-3	03-4572-10	08/11/03 20:21	643023	7.65	552333	11.55	295914	13.85
13	TB-7-8-11-03	03-4572-11	08/11/03 20:47	637799	7.65	550519	11.55	288129	13.85
14	MW-12-4MS	03-4572-6MS	08/11/03 21:14	757966	7.64	592654	11.54	310822	13.84
15	MW-12-4MSD	03-4572-6MSD	08/11/03 21:40	801156	7.65	622097	11.54	325477	13.84
16									
17									
18									
19									
20									
21									
22									

IS-1 = FLUOROBENZENE

IS-2 = CHLOROBENZENE-D5

IS-3 = 1,4-DICHLOROBENZENE-D4

Area Upper Limit = +100% of CCV internal standard area

Area Lower Limit = - 50% of CCV internal standard area

RT Upper Limit = +0.50 minutes of CCV internal standard RT

RT Lower Limit = - 0.50 minutes of CCV internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits

Applied P & Ch Laboratory

VOC Analysis General Logbook

13760 Magnolia Ave. Chino CA 91710
Tel: (909) 590-1828 Fax: (909) 590-1498

Sequence # 0393415 Batch # 0393415 Matrix: W Date: 07/21/03 Analyst: Zou
Lot #: IS/Surrogate: GC-1514/1515 Methanol(mark-M): _____ PEG (mark-PEG): _____ Defoaming(mark-DF): _____

Calib. + Ini. Batch Initial Batch; Middle Batch; Final Batch. Internal Study Datafile Path: _____

Op. #	Type	Sample ID	Method	V/X=f ₁	V ₁ /V _i =f ₂	V _{sp0} /V _{inj} =f ₃	F	A-#	Datafile	Note	pH
2217	SP	G3415 P01	E54A	25/25 = 1	/ =	/ =	1		G3415 P01	07/21/03	7.08
2218	Calib.	02-0003		/ =	/ =	/ =			02-0003	0.3 ppb	
2219		0002A		/ =	/ =	/ =			0002A	2 ppb	
2220		00010		/ =	/ =	/ =			00010	10 ppb	
2221		00020		/ =	/ =	/ =			00020	20 ppb	
2222		00040		/ =	/ =	/ =			00040	40 ppb	
2223		00060		/ =	/ =	/ =			00060	60 ppb	
2224				/ =	/ =	/ =					
2225				/ =	/ =	/ =					
2226				/ =	/ =	/ =					
2227				/ =	/ =	/ =					
2228				/ =	/ =	/ =					
2229				/ =	/ =	/ =					
2230				/ =	/ =	/ =					
2231				/ =	/ =	/ =					
2232				/ =	/ =	/ =					
2233				/ =	/ =	/ =					
2234				/ =	/ =	/ =					
2235				/ =	/ =	/ =					
2236				/ =	/ =	/ =					
2237				/ =	/ =	/ =					
2238				/ =	/ =	/ =					
2239				/ =	/ =	/ =					
2240				/ =	/ =	/ =					
2241				/ =	/ =	/ =					
2242				/ =	/ =	/ =					
2243				/ =	/ =	/ =					
2244				/ =	/ =	/ =					
2245				/ =	/ =	/ =					
2246				/ =	/ =	/ =					
2247				/ =	/ =	/ =					
2248				/ =	/ =	/ =					

Type	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X(g or mL) = T	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X(g or mL) = T
LCS/LCSD		GC-	x / X = ppb		GC-	x / X = ppb
MS/MSD		GC-	x / X = ppb		GC-	x / X = ppb

Footnote/Anomaly:

Applied P & Ch Laboratory
 13760 Magnolia Ave. Chino CA 91710
 Tel: (909) 590-1828 Fax: (909) 590-1498

VOC Analysis General Logbook

Sequence # 0393452 Batch # 0393452 Matrix: W Date: 07/24/03 Analyst: Zou
 Lot #: IS/Surrogate: GC-1514/1515 Methanol(mark-M): _____ PEG (mark-PEG): _____ Defoaming(mark-DF): _____

Calib. + Ini. Batch Initial Batch; Middle Batch; Final Batch. Internal Study Datafile Path: _____

Op. #	Type	Sample ID	Method	V/X=f ₁	V ₁ /V _i =f ₂	V _{ppg} /V _{inj} =f ₃	F	A-#	Datafile	Note	pH
2313	SP	63452901	E524A	25/25 = 1	/ =	/ =	1		63452901	07/24/03	10:07 am
2314	CCV	001		/ =	/ =	/ =	1		001	GC15447	
2315	LCS	L01		/ =	/ =	/ =	1		L01		
2316	LCSD	J01		/ =	/ =	/ =	1		J01		
2317	MB	↓ K01		/ =	/ =	/ =	1		↓ K01		
2318	Sample	476-08		/ =	/ =	/ =	1		476-08	Internal PE	
2319		482-05		/ =	/ =	/ =	1		482-05	PE	
2320		18		↓ = ↓	/ =	/ =	↓		18		
2321		03		15 = 5	/ =	/ =	5		03		
2322		03A		125 = 1	/ =	/ =	1		03A		
2323		04		15 = 5	/ =	/ =	5		04		
2324		04A		15 = 5	/ =	/ =	5		04A		
2325		↓ 04B		125 = 1	/ =	/ =	1		↓ 04B		
2326		475-03B		15 = 5	/ =	/ =	5		475-03B		
2327	↓	03C	↓	↓ 125 = 1	/ =	/ =	1		↓ 03C		
2328				/ =	/ =	/ =					
2329				/ =	/ =	/ =					
2330				/ =	/ =	/ =					
2331				/ =	/ =	/ =					
2332				/ =	/ =	/ =					
2333				/ =	/ =	/ =					
2334				/ =	/ =	/ =					
2335				/ =	/ =	/ =					
2336				/ =	/ =	/ =					
2337				/ =	/ =	/ =					
2338				/ =	/ =	/ =					
2339				/ =	107/25/01	/ =					
2340				/ =	/ =	/ =					
2341				/ =	/ =	/ =					
2342				/ =	/ =	/ =					
2343				/ =	/ =	/ =					
2344				/ =	/ =	/ =					

Type	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X(ε or mL) = T	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X(ε or mL) = T
LCS/LCSD	2315/2316	GC-15448	200 × 2.5 / X = 20 ppb		GC-	x / X = ppb
MS/MSD		GC-	x / X = ppb		GC-	x / X = ppb

Footnote/Anomaly:

Applied P & Ch Laboratory

179 Magnolia Ave. Chino CA 91710
(909) 590-1828 Fax: (909) 590-1498

VOC Analysis General Logbook

Source # 0393659 Batch # 0393659 Matrix: W Date: 08/11/03 Analyst: Joe

IS/Surrogate: GC-1514/1515 Methanol(mark-M): _____ PEG (mark-PEG): _____ Defoaming(mark-DF): _____

Sub: Ini. Batch Initial Batch; Middle Batch; Final Batch. Internal Study Datafile Path: _____

Type	Sample ID	Method	V/X=f ₁	V _j /V _i =f ₂	V _{1pg} /V _{inj} =f ₃	F	A-#	Datafile	Note	pH
SP	G3659P01	E54A	25/25 = 1	/ =	/ =	1		G3659P01	08/11/03 10:40am	
CCV	Q01		/ =	/ =	/ =			Q01	GC15535	
LCS	L01		/ =	/ =	/ =			L01		
MS	M01		/ =	/ =	/ =			M01	\$4509-08	<2
MSD	N01		/ =	/ =	/ =			N01	↓	↓
MB	✓ K01		/ =	/ =	/ =			✓ K01		
Sample	4534-12		/ =	/ =	/ =			4534-12		<2
	↓ 01		/ =	/ =	/ =			↓ 01		
	4509-08		/ =	/ =	/ =			4509-08		
	4534-08		/ =	/ =	/ =			4534-08		
	↓ 09		/ =	/ =	/ =			↓ 09		
	↓ 10		/ =	/ =	/ =			↓ 10		
	4572-01		/ =	/ =	/ =			4572-01		
	02		/ =	/ =	/ =			02		
	03		/ =	/ =	/ =			03		
	04		/ =	/ =	/ =			04		
	05		/ =	/ =	/ =			05		
	06		/ =	/ =	/ =			06		
	07		/ =	/ =	/ =			07		
	08		/ =	/ =	/ =			08		
	09		/ =	/ =	/ =			09		
	10		/ =	/ =	/ =			10		
	↓ 11		/ =	/ =	/ =			↓ 11		
	G3659M02		/ =	/ =	/ =			G3659M02	\$4572-06	
	↓ N02		↓ / =	↓ / =	↓ / =	↓		↓ N02	↓	↓
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					
			/ =	/ =	/ =					

Type	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X(g or mL) = T	Op #	STD Lot #	C _{std} (ng/μL) × V _{std} (μL) / X(g or mL) = T
CS/LCSD	2699	GC-15536	200 × 2.5 / X = ppb		GC-	x / X = ppb
MS/MSD	2700/2701	GC-	x × 2.5 / X = ppb		GC-	x / X = ppb

Footnote/Anomaly: