


APPENDIX D

DATA VALIDATION REPORTS



LABORATORY DATA CONSULTANTS, INC.

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

Geofon, Inc.
22632 Golden Springs Drive, Suite 270
Diamond Bar, CA 91765
ATTN: Mr. Scott Brehmer

September 10, 2004

SUBJECT: NASA JPL, DO #12, Data Validation

Dear Mr. Brehmer,

Enclosed are the final validation reports for the fractions listed below. These SDGs were received on September 3, 2004. Attachment 1 is a summary of the samples that were reviewed for each analysis.

LDC Project # 12429:

| <u>SDG #</u> | <u>Fraction</u> |
|---------------------------------------|------------------------------------|
| 04-3915, 04-3925, 04-3938, 04-3962 | Volatiles, Chromium, Wet Chemistry |

The data validation was performed under EPA Level III and Level IV guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, February 1994
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996; update IIIA, April 1998

Please feel free to contact us if you have any questions.

Sincerely,

Erlinda T. Rauto
Operations Manager/Senior Chemist

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 2, 2004
LDC Report Date: September 8, 2004
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-3915

Sample Identification

DUPE-1-3Q04
EB-1-8/2/04
MW-14-1
MW-14-2
MW-14-3
MW-14-4
MW-14-5**
MW-21-1
MW-21-2
MW-21-3
MW-21-4
MW-21-5
TB-1-8/2/04
MW-21-4MS
MW-21-4MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 15 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

| Date | Compound | %D | Associated Samples | Flag | A or P |
|--------|------------|-------|----------------------------|---|--------|
| 8/6/04 | 2-Butanone | 30.83 | All samples in SDG 04-3195 | J (all detects) UJ (all non-detects) | P |

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

| Method Blank ID | Analysis Date | Compound TIC (RT in minutes) | Concentration | Associated Samples |
|-----------------|---------------|---------------------------------|---------------|----------------------------|
| 04G3115-MB-01 | 8/6/04 | Methylene chloride | 0.5 ug/L | All samples in SDG 04-3915 |

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples DUPE-1-3Q04 and MW-14-5** were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-1-8/2/04 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-1-8/2/04 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

| Equipment Blank ID | Compound | Concentration (ug/L) |
|---------------------------|-----------------|-----------------------------|
| EB-1-8/2/04 | 2-Butanone | 2 |

NASA JPL
Volatiles - Data Qualification Summary - SDG 04-3915

| SDG | Sample | Compound | Flag | A or P | Reason |
|---------|---|------------|---|--------|-----------------------------|
| 04-3915 | DUPE-1-3Q04 EB-1-8/2/04 MW-14-1 MW-14-2 MW-14-3 MW-14-4 MW-14-5** MW-21-1 MW-21-2 MW-21-3 MW-21-4 MW-21-5 TB-1-8/2/04 | 2-Butanone | J (all detects) UJ (all non-detects) | P | Continuing calibration (%D) |

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 04-3915

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 3, 2004
LDC Report Date: September 8, 2004
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-3925

Sample Identification

EB-2-8/3/04
MW-7
MW-18-2
MW-18-3
MW-18-4
MW-18-5
TB-2-8/3/04

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

| Date | Compound | %D | Associated Samples | Flag | A or P |
|--------|------------|-------|----------------------------|---|--------|
| 8/6/04 | 2-Butanone | 30.83 | All samples in SDG 04-3925 | J (all detects) UJ (all non-detects) | P |

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

| Method Blank ID | Analysis Date | Compound TIC (RT in minutes) | Concentration | Associated Samples |
|-----------------|---------------|---------------------------------|---------------|----------------------------|
| 04G3115-MB-01 | 8/6/04 | Methylene chloride | 0.5 ug/L | All samples in SDG 04-3925 |

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-2-8/3/04 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-2-8/3/04 was identified as an equipment blank. No volatile contaminants were found in this blank.

**NASA JPL
Volatiles - Data Qualification Summary - SDG 04-3925**

| SDG | Sample | Compound | Flag | A or P | Reason |
|---------|--|------------|---|--------|-----------------------------|
| 04-3925 | EB-2-8/3/04 MW-7 MW-18-2 MW-18-3 MW-18-4 MW-18-5 TB-2-8/3/04 | 2-Butanone | J (all detects) UJ (all non-detects) | P | Continuing calibration (%D) |

**NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 04-3925**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 4, 2004
LDC Report Date: September 8, 2004
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-3938

Sample Identification

EB-3-8/4/04
MW-20-1
MW-20-2
MW-20-3
MW-20-4
MW-20-5
TB-3-8/4/04
MW-20-5MS
MW-20-5MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

| Method Blank ID | Analysis Date | Compound TIC (RT in minutes) | Concentration | Associated Samples |
|-----------------|---------------|---------------------------------|---------------|----------------------------|
| 04G3136-MB-01 | 8/7/04 | Methylene chloride | 1.1 ug/L | All samples in SDG 04-3938 |

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks with the following exceptions:

| Sample | Compound TIC (RT in minutes) | Reported Concentration | Modified Final Concentration |
|-------------|---------------------------------|---------------------------|---------------------------------|
| EB-3-8/4/04 | Methylene chloride | 1.6 ug/L | 1.6U ug/L |
| MW-20-1 | Methylene chloride | 0.6 ug/L | 0.6U ug/L |
| MW-20-2 | Methylene chloride | 1.2 ug/L | 1.2U ug/L |
| MW-20-3 | Methylene chloride | 0.8 ug/L | 0.8U ug/L |
| MW-20-4 | Methylene chloride | 1.2 ug/L | 1.2U ug/L |
| MW-20-5 | Methylene chloride | 0.9 ug/L | 0.9U ug/L |
| TB-3-8/4/04 | Methylene chloride | 0.5 ug/L | 0.5U ug/L |

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-3-8/4/04 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

| Trip Blank ID | Compound | Concentration (ug/L) |
|----------------------|--------------------|-----------------------------|
| TB-3-8/4/04 | Methylene chloride | 1.6 |

Sample EB-3-8/4/04 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

| Equipment Blank ID | Compound | Concentration (ug/L) |
|---------------------------|--------------------|-----------------------------|
| EB-3-8/4/04 | Methylene chloride | 0.5 |

NASA JPL
Volatiles - Data Qualification Summary - SDG 04-3938

No Sample Data Qualified in this SDG

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 04-3938

| SDG | Sample | Compound TIC (RT in minutes) | Modified Final Concentration | A or P |
|---------|-------------|---------------------------------|---------------------------------|--------|
| 04-3938 | EB-3-8/4/04 | Methylene chloride | 1.6U ug/L | A |
| 04-3938 | MW-20-1 | Methylene chloride | 0.6U ug/L | A |
| 04-3938 | MW-20-2 | Methylene chloride | 1.2U ug/L | A |
| 04-3938 | MW-20-3 | Methylene chloride | 0.8U ug/L | A |
| 04-3938 | MW-20-4 | Methylene chloride | 1.2U ug/L | A |
| 04-3938 | MW-20-5 | Methylene chloride | 0.9U ug/L | A |
| 04-3938 | TB-3-8/4/04 | Methylene chloride | 0.5U ug/L | A |

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 5, 2004
LDC Report Date: September 8, 2004
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-3962

Sample Identification

DUPE-2-3Q04
EB-4-8/5/04
MW-17-2
MW-17-3
MW-17-4
MW-19-1
MW-19-2
MW-19-3
MW-19-4
MW-19-5**
TB-4-8/5/04
MW-17-4MS
MW-17-4MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 13 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

| Date | Compound | %D | Associated Samples | Flag | A or P |
|---------|---------------|-------|---|---|--------|
| 8/11/04 | Chloromethane | 31.58 | MW-17-2 MW-17-3 MW-17-4 MW-19-1 MW-19-2 MW-19-3 MW-19-4 MW-19-5** TB-4-8/5/04 MW-17-4MS MW-17-4MSD 04G3172-MB-01 | J (all detects) UJ (all non-detects) | P |

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

| Method Blank ID | Analysis Date | Compound TIC (RT in minutes) | Concentration | Associated Samples |
|-----------------|---------------|---------------------------------|---------------|----------------------------|
| 04G3136-MB-01 | 8/7/04 | Methylene chloride | 1.1 ug/L | DUPE-2-3Q04 EB-4-8/5/04 |

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples DUPE-2-3Q04 and MW-19-1 were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-4-8/5/04 was identified as a trip blank. No volatile contaminants were found in this blank.

Sample EB-4-8/5/04 was identified as an equipment blank. No volatile contaminants were found in this blank.

NASA JPL
Volatiles - Data Qualification Summary - SDG 04-3962

| SDG | Sample | Compound | Flag | A or P | Reason |
|---------|---|---------------|---|--------|-----------------------------|
| 04-3962 | MW-17-2 MW-17-3 MW-17-4 MW-19-1 MW-19-2 MW-19-3 MW-19-4 MW-19-5** TB-4-8/5/04 | Chloromethane | J (all detects) UJ (all non-detects) | P | Continuing calibration (%D) |

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 04-3962

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 2, 2004
LDC Report Date: September 8, 2004
Matrix: Water
Parameters: Chromium
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-3915

Sample Identification

EB-1-8/2/04
MW-14-1
MW-14-2
MW-14-3
MW-21-1
MW-21-2
MW-21-3
MW-21-4
MW-21-5
MW-21-4MS
MW-21-4MSD
MW-21-4DUP

Introduction

This data review covers 12 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

| Method Blank ID | Analyte | Maximum Concentration | Associated Samples |
|-----------------|----------|-----------------------|----------------------------|
| ICB/CCB | Chromium | 0.041 ug/L | All samples in SDG 04-3915 |

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks.

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample data were not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards (ICP-MS)

ICP-MS was not reviewed in this SDG.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met with the following exceptions:

| Diluted Sample | Analyte | %D (Limits) | Associated Samples | Flag | A or P |
|----------------|----------|--------------------|----------------------------|-----------------|--------|
| MW-21-4L | Chromium | 18.5 (≤ 10) | All samples in SDG 04-3915 | J (all detects) | A |

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-1-8/2/04 was identified as an equipment blank. No chromium contaminants were found in this blank with the following exceptions:

| Equipment Blank ID | Analyte | Concentration (ug/L) |
|--------------------|----------|----------------------|
| EB-1-8/2/04 | Chromium | 0.52 |

**NASA JPL
Chromium - Data Qualification Summary - SDG 04-3915**

| SDG | Sample | Analyte | Flag | A or P | Reason |
|---------|---|----------|-----------------|--------|--------------------------|
| 04-3915 | EB-1-8/2/04 MW-14-1 MW-14-2 MW-14-3 MW-21-1 MW-21-2 MW-21-3 MW-21-4 MW-21-5 | Chromium | J (all detects) | A | ICP serial dilution (%D) |

**NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 04-3915**

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 3, 2004
LDC Report Date: September 8, 2004
Matrix: Water
Parameters: Chromium
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-3925

Sample Identification

EB-2-8/3/04
MW-7
MW-18-2
MW-18-3
MW-18-4

Introduction

This data review covers 5 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

| Method Blank ID | Analyte | Maximum Concentration | Associated Samples |
|-----------------|----------|-----------------------|----------------------------|
| ICB/CCB | Chromium | 0.042 ug/L | All samples in SDG 04-3925 |

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks.

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample data were not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards (ICP-MS)

ICP-MS was not reviewed in this SDG.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met with the following exceptions:

| Diluted Sample | Analyte | %D (Limits) | Associated Samples | Flag | A or P |
|----------------|----------|--------------------|----------------------------|-----------------|--------|
| MW-21-4L | Chromium | 18.5 (≤ 10) | All samples in SDG 04-3925 | J (all detects) | A |

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-2-8/3/04 was identified as an equipment blank. No chromium contaminants were found in this blank with the following exceptions:

| Equipment Blank ID | Analyte | Concentration (ug/L) |
|--------------------|----------|----------------------|
| EB-2-8/3/04 | Chromium | 0.35 |

NASA JPL

Chromium - Data Qualification Summary - SDG 04-3925

| SDG | Sample | Analyte | Flag | A or P | Reason |
|---------|--|----------|-----------------|--------|--------------------------|
| 04-3925 | EB-2-8/3/04 MW-7 MW-18-2 MW-18-3 MW-18-4 | Chromium | J (all detects) | A | ICP serial dilution (%D) |

NASA JPL

Chromium - Laboratory Blank Data Qualification Summary - SDG 04-3925

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 4, 2004
LDC Report Date: September 8, 2004
Matrix: Water
Parameters: Chromium
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-3938

Sample Identification

EB-3-8/4/04
MW-20-1
MW-20-2
MW-20-3
MW-20-4
MW-20-5
MW-20-5MS
MW-20-5DUP
MW-20-5MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

| Method Blank ID | Analyte | Maximum Concentration | Associated Samples |
|-----------------|----------|-----------------------|----------------------------|
| ICB/CCB | Chromium | 0.145 ug/L | All samples in SDG 04-3938 |

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

| Sample | Analyte | Reported Concentration | Modified Final Concentration |
|-------------|----------|------------------------|------------------------------|
| EB-3-8/4/04 | Chromium | 0.41 ug/L | 0.41U ug/L |

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample data were not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards (ICP-MS)

ICP-MS was not reviewed in this SDG.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met.

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-3-8/4/04 was identified as an equipment blank. No chromium contaminants were found in this blank with the following exceptions:

| Equipment Blank ID | Analyte | Concentration (ug/L) |
|--------------------|----------|----------------------|
| EB-3-8/4/04 | Chromium | 0.41 |

NASA JPL
Chromium - Data Qualification Summary - SDG 04-3938

No Sample Data Qualified in this SDG

NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 04-3938

| SDG | Sample | Analyte | Modified Final Concentration | A or P |
|------------|---------------|----------------|-------------------------------------|---------------|
| 04-3938 | EB-3-8/4/04 | Chromium | 0.41U ug/L | A |

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 5, 2004
LDC Report Date: September 8, 2004
Matrix: Water
Parameters: Chromium
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-3962

Sample Identification

MW-17-2
MW-17-3
MW-17-4

Introduction

This data review covers 3 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

| Method Blank ID | Analyte | Maximum Concentration | Associated Samples |
|-----------------|----------|-----------------------|----------------------------|
| ICB/CCB | Chromium | 0.145 ug/L | All samples in SDG 04-3962 |

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks.

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample data were not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards (ICP-MS)

ICP-MS was not reviewed in this SDG.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met.

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

No field blanks were identified in this SDG.

NASA JPL
Chromium - Data Qualification Summary - SDG 04-3962

No Sample Data Qualified in this SDG

NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 04-3962

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 2, 2004
LDC Report Date: September 7, 2004
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-3915

Sample Identification

DUPE-1-3Q04
EB-1-8/2/04
MW-14-1
MW-14-2
MW-14-3
MW-14-4
MW-14-5**
MW-21-1
MW-21-2
MW-21-3
MW-21-4
MW-21-5
MW-21-4MS
MW-21-4MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 14 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

Samples DUPE-1-3Q04 and MW-14-5** were identified as field duplicates. No contaminant concentrations were detected in any of the samples.

X. Field Blanks

Sample EB-1-8/2/04 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 04-3915

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 04-3915

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 3, 2004
LDC Report Date: September 7, 2004
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-3925

Sample Identification

EB-2-8/3/04
MW-7
MW-18-2
MW-18-3
MW-18-4
MW-18-5

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-2-8/3/04 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 04-3925

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 04-3925

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 4, 2004
LDC Report Date: September 7, 2004
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-3938

Sample Identification

EB-3-8/4/04
MW-20-1
MW-20-2
MW-20-3
MW-20-4
MW-20-5
MW-20-5MS
MW-20-5MSD

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-3-8/4/04 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 04-3938

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 04-3938

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 5, 2004
LDC Report Date: September 7, 2004
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-3962

Sample Identification

DUPE-2-3Q04
EB-4-8/5/04
MW-17-2
MW-17-3
MW-17-4
MW-19-1
MW-19-2
MW-19-3
MW-19-4
MW-19-5**
MW-17-2MS
MW-17-2MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 12 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

Samples DUPE-2-3Q04 and MW-19-5** were identified as field duplicates. No contaminant concentrations were detected in any of the samples.

X. Field Blanks


Sample EB-4-8/5/04 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 04-3962

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 04-3962

No Sample Data Qualified in this SDG



LABORATORY DATA CONSULTANTS, INC.

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

Geofon, Inc.
22632 Golden Springs Drive, Suite 270
Diamond Bar, CA 91765
ATTN: Mr. Scott Brehmer

September 27, 2004

SUBJECT: NASA JPL, DO #12, Data Validation

Dear Mr. Brehmer,

Enclosed are the final validation reports for the fractions listed below. These SDGs were received on September 14, 2004. Attachment 1 is a summary of the samples that were reviewed for each analysis.

LDC Project # 12465:

| <u>SDG #</u> | <u>Fraction</u> |
|------------------------------|------------------------------------|
| 04-4000, 04-4017, 04-4031 | Volatiles, Chromium, Wet Chemistry |

The data validation was performed under EPA Level III and Level IV guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, February 1994
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996; update IIIA, April 1998

Please feel free to contact us if you have any questions.

Sincerely,

Erlinda T. Rauto
Operations Manager/Senior Chemist

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 9, 2004
LDC Report Date: September 24, 2004
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-4000

Sample Identification

DUPE-3-3Q04
EB-5-8/9/04
MW-4-1
MW-4-2
MW-4-3
MW-11-1
MW-11-2
MW-11-3
MW-11-4**
TB-5-8/9/04

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

| Date | Compound | %D | Associated Samples | Flag | A or P |
|---------|---------------|-------|----------------------------|---|--------|
| 8/11/04 | Chloromethane | 31.58 | All samples in SDG 04-4000 | J (all detects) UJ (all non-detects) | P |

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were not required by the method.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples DUPE-3-3Q04 and MW-11-4** were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-5-8/9/04 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

| Trip Blank ID | Compound | Concentration (ug/L) |
|---------------|------------|----------------------|
| TB-5-8/9/04 | 2-Butanone | 1 |

Sample EB-5-8/9/04 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

| Equipment Blank ID | Compound | Concentration (ug/L) |
|--------------------|--------------|----------------------|
| EB-5-8/9/04 | 2-Butanone | 1 |
| | Ethylbenzene | 1.7 |
| | Toluene | 2.8 |
| | o-Xylene | 1 |
| | m,p-Xylenes | 7.2 |

NASA JPL
Volatiles - Data Qualification Summary - SDG 04-4000

| SDG | Sample | Compound | Flag | A or P | Reason |
|---------|---|---------------|---|--------|-----------------------------|
| 04-4000 | DUPE-3-3Q04 EB-5-8/9/04 MW-4-1 MW-4-2 MW-4-3 MW-11-1 MW-11-2 MW-11-3 MW-11-4** TB-5-8/9/04 | Chloromethane | J (all detects) UJ (all non-detects) | P | Continuing calibration (%D) |

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 04-4000

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 10, 2004
LDC Report Date: September 24, 2004
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-4017

Sample Identification

EB-6-8/10/04
MW-23-1
MW-23-2
MW-23-3
MW-24-1
MW-24-2
MW-24-3
TB-6-8/10/04
MW-23-2MS
MW-23-2MSD
MW-24-1MS
MW-24-1MSD

Introduction

This data review covers 12 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-6-8/10/04 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

| Trip Blank ID | Compound | Concentration (ug/L) |
|----------------------|-----------------|-----------------------------|
| TB-6-8/10/04 | 2-Butanone | 0.9 |

Sample EB-6-8/10/04 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

| Equipment Blank ID | Compound | Concentration (ug/L) |
|--------------------|--|--------------------------|
| EB-6-8/10/04 | Ethylbenzene Toluene o-Xylene m,p-Xylenes | 0.6 1.0 0.4 2.6 |

NASA JPL
Volatiles - Data Qualification Summary - SDG 04-4017

No Sample Data Qualified in this SDG

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 04-4017

No Sample Data Qualified in this SDG

Laboratory Data Consultants, Inc. Data Validation Report

Project/Site Name: NASA JPL
Collection Date: August 11, 2004
LDC Report Date: September 24, 2004
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-4031

Sample Identification

DUPE-4-3Q04
EB-7-8/11/04
MW-3-2**
MW-3-3
MW-3-4
MW-12-1
MW-12-2
MW-12-3
MW-12-4
MW-12-5
TB-7-8/11/04

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 11 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples DUPE-4-3Q04 and MW-3-3 were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

| Compound | Concentration (ug/L) | | RPD |
|-------------------------|----------------------|--------|-----|
| | DUPE-4-3Q04 | MW-3-3 | |
| Ethylbenzene | 0.7 | 0.6 | 15 |
| Methyl-tert-butyl ether | 0.3 | 0.4 | 29 |
| Toluene | 0.4 | 0.3 | 29 |

XVII. Field Blanks

Sample TB-7-8/11/04 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

| Trip Blank ID | Compound | Concentration (ug/L) |
|---------------|------------|----------------------|
| TB-7-8/11/04 | 2-Butanone | 1 |

Sample EB-7-8/11/04 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

| Equipment Blank ID | Compound | Concentration (ug/L) |
|--------------------|-------------|----------------------|
| EB-7-8/11/04 | m,p-Xylenes | 0.8 |

NASA JPL
Volatiles - Data Qualification Summary - SDG 04-4031

No Sample Data Qualified in this SDG

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 04-4031

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 9, 2004
LDC Report Date: September 15, 2004
Matrix: Water
Parameters: Chromium
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-4000

Sample Identification

EB-5-8/9/04
MW-4-1
MW-4-2
MW-4-3
MW-11-1
MW-11-2
MW-11-3

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

| Method Blank ID | Analyte | Maximum Concentration | Associated Samples |
|-----------------|----------|-----------------------|----------------------------|
| ICB/CCB | Chromium | 0.145 ug/L | All samples in SDG 04-4000 |

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

| Sample | Analyte | Reported Concentration | Modified Final Concentration |
|----------------|----------|------------------------|------------------------------|
| EB-5-8/9/04 | Chromium | 0.67 ug/L | 0.67U ug/L |
| MW-4-1 (1.25x) | Chromium | 0.76 ug/L | 0.76U ug/L |

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample not required by the method.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

Raw data were not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met.

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-5-8/9/04 was identified as an equipment blank. No chromium was found in this blank with the following exceptions:

| Equipment Blank ID | Analyte | Concentration (ug/L) |
|--------------------|----------|----------------------|
| EB-5-8/9/04 | Chromium | 0.67 |

NASA JPL
Chromium - Data Qualification Summary - SDG 04-4000

No Sample Data Qualified in this SDG

NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 04-4000

| SDG | Sample | Analyte | Modified Final Concentration | A or P |
|------------|----------------|----------------|-------------------------------------|---------------|
| 04-4000 | EB-5-8/9/04 | Chromium | 0.67U ug/L | A |
| 04-4000 | MW-4-1 (1.25x) | Chromium | 0.76U ug/L | A |

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 10, 2004
LDC Report Date: September 15, 2004
Matrix: Water
Parameters: Chromium
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-4017

Sample Identification

EB-6-8/10/04
MW-23-1
MW-23-2
MW-23-3
MW-23-4
MW-24-1
MW-24-2
MW-24-3
MW-24-4
MW-23-2MS
MW-23-2MSD
MW-23-2DUP
MW-24-1MS
MW-24-1MSD
MW-24-1DUP

Introduction

This data review covers 15 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

| Method Blank ID | Analyte | Maximum Concentration | Associated Samples |
|-----------------|----------|-----------------------|----------------------------|
| PB (prep blank) | Chromium | 0.030 ug/L | All samples in SDG 04-4017 |
| ICB/CCB | Chromium | 0.235 ug/L | All samples in SDG 04-4017 |

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

| Sample | Analyte | Reported Concentration | Modified Final Concentration |
|--------------|----------|------------------------|------------------------------|
| EB-6-8/10/04 | Chromium | 0.87 ug/L | 0.87U ug/L |

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample not required by the method.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

Raw data were not reviewed for this SDG.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met.

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-6-8/10/04 was identified as an equipment blank. No chromium was found in this blank with the following exceptions:

| Equipment Blank ID | Analyte | Concentration (ug/L) |
|---------------------------|----------------|-----------------------------|
| EB-6-8/10/04 | Chromium | 0.87 |

**NASA JPL
Chromium - Data Qualification Summary - SDG 04-4017**

No Sample Data Qualified in this SDG

**NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 04-4017**

| SDG | Sample | Analyte | Modified Final Concentration | A or P |
|---------|--------------|----------|------------------------------|--------|
| 04-4017 | EB-6-8/10/04 | Chromium | 0.87U ug/L | A |

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 11, 2004
LDC Report Date: September 15, 2004
Matrix: Water
Parameters: Chromium
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-4031

Sample Identification

DUPE-4-3Q04
EB-7-8/11/04
MW-3-2**
MW-3-3
MW-3-4
MW-12-1
MW-12-2
MW-12-3

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 8 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

| Method Blank ID | Analyte | Maximum Concentration | Associated Samples |
|-----------------|----------|-----------------------|----------------------------|
| PB (prep blank) | Chromium | 0.030 ug/L | All samples in SDG 04-4031 |
| ICB/CCB | Chromium | 0.235 ug/L | All samples in SDG 04-4031 |

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

| Sample | Analyte | Reported Concentration | Modified Final Concentration |
|--------------|----------|------------------------|------------------------------|
| EB-7-8/11/04 | Chromium | 0.29 ug/L | 0.29U ug/L |

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample not required by the method.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met.

XI. Sample Result Verification

All sample result verifications met validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

Samples DUPE-4-3Q04 and MW-3-3 were identified as field duplicates. No chromium was detected in any of the samples with the following exceptions:

| Analyte | Concentration (ug/L) | | RPD |
|----------|----------------------|--------|-----|
| | DUPE-4-3Q04 | MW-3-3 | |
| Chromium | 7.4 | 7.2 | 3 |

XIV. Field Blanks

Sample EB-7-8/11/04 was identified as an equipment blank. No chromium was found in this blank with the following exceptions:

| Equipment Blank ID | Analyte | Concentration (ug/L) |
|--------------------|----------|----------------------|
| EB-7-8/11/04 | Chromium | 0.29 |

**NASA JPL
Chromium - Data Qualification Summary - SDG 04-4031**

No Sample Data Qualified in this SDG

**NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 04-4031**

| SDG | Sample | Analyte | Modified Final Concentration | A or P |
|------------|---------------|----------------|-------------------------------------|---------------|
| 04-4031 | EB-7-8/11/04 | Chromium | 0.29U ug/L | A |

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 9, 2004
LDC Report Date: September 15, 2004
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-4000

Sample Identification

DUPE-3-3Q04
EB-5-8/9/04
MW-4-1
MW-4-2
MW-4-3
MW-11-1
MW-11-2
MW-11-3
MW-11-4**
MW-4-1MS
MW-4-1MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 11 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

Samples DUPE-3-3Q04 and MW-11-4** were identified as field duplicates. No contaminant concentrations were detected in any of the samples.

X. Field Blanks

Sample EB-5-8/9/04 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 04-4000

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 04-4000

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 10, 2004
LDC Report Date: September 15, 2004
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-4017

Sample Identification

EB-6-8/10/04
MW-23-1
MW-23-2
MW-23-3
MW-23-4
MW-24-1
MW-24-2
MW-24-3
MW-24-4
MW-23-2MS
MW-23-2MSD
MW-24-1MS
MW-24-1MSD

Introduction

This data review covers 13 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits with the following exceptions:

| Spike ID (Associated Samples) | Analyte | MS (%R) (Limits) | MSD (%R) (Limits) | RPD (Limits) | Flag | A or P |
|---|-------------|---------------------|----------------------|-----------------|--|--------|
| MW-24-1MS/MSD (EB-6-8/10/04 MW-23-1 MW-23-2 MW-23-3 MW-24-1 MW-24-2 MW-24-3) | Perchlorate | 4 (75-125) | 6 (75-125) | - | J (all detects) R (all non-detects) | A |

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks


Sample EB-6-8/10/04 was identified as an equipment blank. No contaminant concentrations were found in this blank.

**NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 04-4017**

| SDG | Sample | Analyte | Flag | A or P | Reason |
|---------|--|-------------|--|--------|--|
| 04-4017 | EB-6-8/10/04 MW-23-1 MW-23-2 MW-23-3 MW-24-1 MW-24-2 MW-24-3 | Perchlorate | J (all detects) R (all non-detects) | A | Matrix spike/Matrix spike duplicates (%R) |

**NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 04-4017**

No Sample Data Qualified in this SDG



LABORATORY DATA CONSULTANTS, INC.

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

Geofon, Inc.
22632 Golden Springs Drive, Suite 270
Diamond Bar, CA 91765
ATTN: Mr. Scott Brehmer

October 13, 2004

SUBJECT: NASA JPL, DO #12, Data Validation

Dear Mr. Brehmer,

Enclosed is the revised data validation report for the fraction listed below. Please replace the previously submitted report with the enclosed revised report.

| <u>SDG#</u> | <u>LDC#</u> | <u>Fraction</u> |
|-------------|-------------|-----------------|
| 04-4017 | 12465B6 | Wet Chemistry |

Please feel free to contact us if you have any questions.

Sincerely,

Erlinda T. Rauto
Operations Manager/Senior Chemist

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 10, 2004
LDC Report Date: October 10, 2004
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-4017

Sample Identification

EB-6-8/10/04
MW-23-1
MW-23-2
MW-23-3
MW-23-4
MW-24-1
MW-24-2
MW-24-3
MW-24-4
MW-23-2MS
MW-23-2MSD
MW-24-1MS
MW-24-1MSD

Introduction

This data review covers 13 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

***IV. Matrix Spike/Matrix Spike Duplicates**

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

*Removed MS/MSD finding

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-6-8/10/04 was identified as an equipment blank. No contaminant concentrations were found in this blank.

***NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 04-4017**

No Sample Data Qualified in this SDG

**NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 04-4017**

No Sample Data Qualified in this SDG

*Removed MS/MSD finding from Data Qualification Summary

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 11, 2004
LDC Report Date: September 15, 2004
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-4031

Sample Identification

DUPE-4-3Q04
EB-7-8/11/04
MW-3-2**
MW-3-3
MW-3-4
MW-12-1
MW-12-2
MW-12-3
MW-12-4
MW-12-5
DUPE-4-3Q04MS
DUPE-4-3Q04MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 12 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

Samples DUPE-4-3Q04 and MW-3-3 were identified as field duplicates. No contaminant concentrations were detected in any of the samples.

X. Field Blanks


Sample EB-7-8/11/04 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 04-4031

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 04-4031

No Sample Data Qualified in this SDG



LABORATORY DATA CONSULTANTS, INC.

7750 El Camino Real, Suite 2L Carlsbad, CA 92009 Phone: 760/634-0437 Fax: 760/634-0439

Geofon, Inc.
22632 Golden Springs Drive, Suite 270
Diamond Bar, CA 91765
ATTN: Mr. Scott Brehmer

September 27, 2004

SUBJECT: NASA JPL, DO #12, Data Validation

Dear Mr. Brehmer,

Enclosed are the final validation reports for the fractions listed below. These SDGs were received on September 20, 2004. Attachment 1 is a summary of the samples that were reviewed for each analysis.

LDC Project # 12484:

| <u>SDG #</u> | <u>Fraction</u> |
|------------------------------|------------------------------------|
| 04-4044, 04-4085, 04-4096 | Volatiles, Chromium, Wet Chemistry |

The data validation was performed under EPA Level III and Level IV guidelines. The analyses were validated using the following documents, as applicable to each method:

- USEPA, Contract Laboratory Program National Functional Guidelines for Organic Data Review, October 1999
- USEPA, Contract Laboratory Program National Functional Guidelines for Inorganic Data Review, February 1994
- EPA SW 846, Third Edition, Test Methods for Evaluating Solid Waste, update 1, July 1992; update IIA, August 1993; update II, September 1994; update IIB, January 1995; update III, December 1996; update IIIA, April 1998

Please feel free to contact us if you have any questions.

Sincerely,

Erlinda T. Rauto
Operations Manager/Senior Chemist

LDC #12484 (Geofon, Inc.-Diamond Bar / NASA Jet Propulsion Laboratory, DO#012)

| LDC | SDG# | DATE RECD | DATE DUE | VOA (524.2) | | Cr (200.8) | | Cr(VI) (7196) | | CLO, (314.0) | | W | | S | | W | | S | | W | | S | | |
|--------------------|---------|-----------|----------|-------------|---|------------|---|---------------|---|--------------|---|----|---|---|---|---|---|---|---|---|---|---|---|----|
| | | | | W | S | W | S | W | S | W | S | W | S | W | S | W | S | W | S | W | S | W | S | W |
| Matrix: Water/Soil | | | | | | | | | | | | | | | | | | | | | | | | |
| A | 04-4044 | 09/20/04 | 10/11/04 | 5 | 0 | 4 | 0 | 4 | 0 | 4 | 0 | 4 | 0 | | | | | | | | | | | |
| B | 04-4085 | 09/20/04 | 10/11/04 | 4 | 0 | 4 | 0 | 4 | 0 | 4 | 0 | 3 | 0 | | | | | | | | | | | |
| B | 04-4085 | 09/20/04 | 10/11/04 | 1 | 0 | 1 | 0 | 1 | 0 | 1 | 0 | 1 | 0 | | | | | | | | | | | |
| C | 04-4096 | 09/20/04 | 10/11/04 | 4 | 0 | 3 | 0 | 3 | 0 | 3 | 0 | 3 | 0 | | | | | | | | | | | |
| C | 04-4096 | 09/20/04 | 10/11/04 | 1 | 0 | 1 | 0 | 1 | 0 | 1 | 0 | 1 | 0 | | | | | | | | | | | |
| | | | | 15 | 0 | 13 | 0 | 13 | 0 | 12 | 0 | 12 | 0 | | | | | | | | | | | |
| Total | B | | | | | | | | | | | | | | | | | | | | | | | 53 |

**NASA JPL
Data Validation Reports
LDC# 12484**

Volatiles

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 12, 2004
LDC Report Date: September 27, 2004
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-4044

Sample Identification

EB-8-8/12/04
MW-22-1
MW-22-2
MW-22-3
TB-8-8/12/04
MW-22-3MS
MW-22-3MSD

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

Raw data were not reviewed for this SDG.

XII. Compound Quantitation and CRQLs

Raw data were not reviewed for this SDG.

XIII. Tentatively Identified Compounds (TICs)

Raw data were not reviewed for this SDG.

XIV. System Performance

Raw data were not reviewed for this SDG.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

No field duplicates were identified in this SDG.

XVII. Field Blanks

Sample TB-8-8/12/04 was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

| TRIP Blank ID | Compound | Concentration (ug/L) |
|---------------|------------|----------------------|
| TB-8-8/12/04 | 2-Butanone | 1 |

Sample EB-8-8/12/04 was identified as an equipment blank. No volatile contaminants were found in this blank with the following exceptions:

| Equipment Blank ID | Compound | Concentration (ug/L) |
|--------------------|--------------------|----------------------|
| EB-8-8/12/04 | Methylene chloride | 0.5 |

NASA JPL
Volatiles - Data Qualification Summary - SDG 04-4044

No Sample Data Qualified in this SDG

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 04-4044

No Sample Data Qualified in this SDG

A

Applied P & CH Laboratories
Organic Analysis Results for Method 524.2

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/12/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44044 | Collected by: JJ/MM |
| Sample ID: EB-8-8/12/04 | Lab Sample ID: 04-4044-1 | Received Date: 08/12/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |
| Anal. Method: 524.2 | Prep. Method: 5030 | Instrument ID: GC/MS: A |
| Batch No: 04G3223 | Prep. Date: 08/18/04 | Anal. Date: 08/18/04 |
| Data File Name: 4044-01 | Prep. No: - | Anal. Time: 13:25 |
| Methanol Vol: - | Sample Amount: 25.0 mL | Dilution Factor: 1 |
| Test Level: Low | Sparge Size: 25 mL | Heated Purge: (Y/N) N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | <0.5 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | <0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

aw
8/27/04

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---------------------------------|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | 0.5 | |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | <0.5 | U |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | <0.5 | U |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 112-TRICHLORO122TRIFLUOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec.% | |
|---------------------|------------------------------------|------------------|--------------|-----|
| 1 | 1-BROMO-4-FLUOROENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 99 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 97 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 105 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 95 |
| # of out-of-control | | | 0 | |

Internal Standard

| | | Control Limit, % | IS Rec.% | |
|---------------------|-----------------------|------------------|----------|----|
| 1 | CHLOROENZENE-D5 | 3114-55-4 | 50-200 | 89 |
| 2 | 1,4-DICHLOROENZENE-D4 | 3855-82-1 | 50-200 | 91 |
| 3 | FLUROENZENE | 462-06-6 | 50-200 | 78 |
| # of out-of-control | | | 0 | |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Applied P & CH Laboratories
Organic Analysis Results for Method 524.2

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/12/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44044 | Collected by: JJ/MM |
| Sample ID: MW-22-1 | Lab Sample ID: 04-4044-2 | Received Date: 08/12/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |
| Anal. Method: 524.2 | Prep. Method: 5030 | Instrument ID: GC/MS: A |
| Batch No: 04G3223 | Prep. Date: 08/18/04 | Anal. Date: 08/18/04 |
| Data File Name: 4044-02 | Prep. No: - | Anal. Time: 13:51 |
| Methanol Vol: - | Sample Amount: 25.0 mL | Dilution Factor: 1 |
| Test Level: Low | Sparge Size: 25 mL | Heated Purge: (Y/N) N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | < 0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | < 0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | < 0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | < 0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | < 0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | < 0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | < 0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | < 0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | < 0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | < 10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | < 0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | < 0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | < 0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | < 0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | < 0.5 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | < 0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | < 0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | < 0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | < 0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | < 0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | < 0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | < 0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | < 0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | < 0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | < 0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | < 0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | < 0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | < 0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | < 0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | < 0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | < 0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | < 0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | < 0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | < 0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | < 0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | < 0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | < 0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | < 0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | < 0.5 | U |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---------------------------------|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | 0.7 | |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | 0.9 | |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROBENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROBENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | 0.3 | J |
| 56 | TRICHLOROFUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 112-TRICHLORO122TRIFLUOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec.% | |
|---------------------|-------------------------------------|------------------|--------------|-----|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 100 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 95 |
| 3 | DIBROMOFUOROMETHANE | 1868-53-7 | 70-122 | 105 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 96 |
| # of out-of-control | | | 0 | |

Internal Standard

| | | Control Limit, % | IS Rec.% | |
|---------------------|-----------------------|------------------|----------|----|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 | 84 |
| 2 | 1,4-DICHLOROETHANE-D4 | 3855-82-1 | 50-200 | 86 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 | 74 |
| # of out-of-control | | | 0 | |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank
D - Diluted

Handwritten signature

Applied P & CH Laboratories
Organic Analysis Results for Method 524.2

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/12/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44044 | Collected by: JJ/MM |
| Sample ID: MW-22-2 | Lab Sample ID: 04-4044-3 | Received Date: 08/12/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |
| Anal. Method: 524.2 | Prep. Method: 5030 | Instrument ID: GC/MS: A |
| Batch No: 04G3223 | Prep. Date: 08/18/04 | Anal. Date: 08/18/04 |
| Data File Name: 4044-03 | Prep. No: - | Anal. Time: 14:17 |
| Methanol Vol: - | Sample Amount: 25.0 mL | Dilution Factor: 1 |
| Test Level: Low | Sparge Size: 25 mL | Heated Purge: (Y/N) N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | <0.5 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | <0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---------------------------------------|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | 0.8 | |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | <0.5 | U |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | <0.5 | U |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,2,2-TRIFLUOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec.% | |
|---|------------------------------------|------------------|--------------|-----|
| 1 | 1-BROMO-4-FLUOROENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 101 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 95 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 105 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 96 |
| # | of out-of-control | | | 0 |

Internal Standard

| | | Control Limit, % | IS Rec.% | |
|---|-----------------------|------------------|----------|----|
| 1 | CHLOROENZENE-D5 | 3114-55-4 | 50-200 | 81 |
| 2 | 1,4-DICHLOROENZENE-D4 | 3855-82-1 | 50-200 | 81 |
| 3 | FLUROENZENE | 462-06-6 | 50-200 | 71 |
| # | of out-of-control | | | 0 |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Applied P & CH Laboratories
Organic Analysis Results for Method 524.2

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/12/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44044 | Collected by: JJ/MM |
| Sample ID: MW-22-3 | Lab Sample ID: 04-4044-4 | Received Date: 08/12/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |
| Anal. Method: 524.2 | Prep. Method: 5030 | Instrument ID: GC/MS: A |
| Batch No: 04G3223 | Prep. Date: 08/18/04 | Anal. Date: 08/18/04 |
| Data File Name: 4044-04 | Prep. No: - | Anal. Time: 14:44 |
| Methanol Vol: - | Sample Amount: 25.0 mL | Dilution Factor: 1 |
| Test Level: Low | Spurge Size: 25 mL | Heated Purge: (Y/N) N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | <0.5 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | <0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---------------------------------------|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | 0.7 | |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | <0.5 | U |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 63 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | <0.5 | U |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,2,2-TRIFLUOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | | Control Limit, % | Surro. Rec.% |
|---------------------|-------------------------------------|------------|------------------|--------------|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 102 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 96 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 106 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 96 |
| # of out-of-control | | | | 0 |

Internal Standard

| | | | Control Limit, % | IS Rec.% |
|---------------------|-----------------------|-----------|------------------|----------|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 | 79 |
| 2 | 1,4-DICHLOROENZENE-D4 | 3855-82-1 | 50-200 | 79 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 | 69 |
| # of out-of-control | | | | 0 |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Organic Analysis Results for Method 524.2

| | | | | | |
|-----------------|--------------|----------------|-----------|---------------------|------------|
| Client Name: | GEOFON, Inc. | Project No: | 4-12812 | Collection Date: | 08/12/2004 |
| Project ID: | JPL GW-3Q04 | Service ID: | 44044 | Collected by: | JJ/MM |
| Sample ID: | TB-8-8/12/04 | Lab Sample ID: | 04-4044-5 | Received Date: | 08/12/2004 |
| Sample Type: | Field Sample | Sample Matrix: | Water | Moisture %: | - |
| Anal. Method: | 524.2 | Prep. Method: | 5030 | Instrument ID: | GC/MS: A |
| Batch No: | 04G3223 | Prep. Date: | 08/18/04 | Anal. Date: | 08/18/04 |
| Data File Name: | 4044-05 | Prep. No: | - | Anal. Time: | 15:10 |
| Methanol Vol. | - | Sample Amount: | 25.0 mL | Dilution Factor: | 1 |
| Test Level: | Low | Sparge Size: | 25 mL | Heated Purge: (Y/N) | N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | 1 | J |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | <0.5 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROETHANE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROETHANE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROETHANE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | <0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | <0.5 | U |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | <0.5 | U |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | <0.5 | U |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,1,2,2,2-TRIFLUOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec.% | |
|---------------------|-------------------------------------|------------------|--------------|-----|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 100 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 98 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 108 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 97 |
| # of out-of-control | | | | 0 |

Internal Standard

| | | Control Limit, % | IS Rec.% | |
|---------------------|-----------------------|------------------|----------|----|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 | 77 |
| 2 | 1,4-DICHLOROENZENE-D4 | 3855-82-1 | 50-200 | 79 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 | 67 |
| # of out-of-control | | | | 0 |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

E - Exceed calibration range

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

B - A positive value was found in the method blank

D - Diluted

LDC #: 12484A1

VALIDATION COMPLETENESS WORKSHEET

SDG #: 04-4044

Level III

Laboratory: Applied Physics & Chemistry Laboratory

Date: 9/23/04

Page: 1 of 7

Reviewer: [Signature]

2nd Reviewer: [Signature]

METHOD: GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

| | Validation Area | | Comments |
|-------|--|----|------------------------------|
| I. | Technical holding times | A | Sampling dates: 8/12/04 |
| II. | GC/MS Instrument performance check | A | |
| III. | Initial calibration | A | % RSD, r ² 20.990 |
| IV. | Continuing calibration | A | |
| V. | Blanks | A | |
| VI. | Surrogate spikes | A | |
| VII. | Matrix spike/Matrix spike duplicates | A | |
| VIII. | Laboratory control samples | A | LC > |
| IX. | Regional Quality Assurance and Quality Control | N | |
| X. | Internal standards | A | |
| XI. | Target compound identification | N | |
| XII. | Compound quantitation/CRQLs | N | |
| XIII. | Tentatively identified compounds (TICs) | N | |
| XIV. | System performance | N | |
| XV. | Overall assessment of data | A | |
| XVI. | Field duplicates | N | |
| XVII. | Field blanks | SW | EB = 1 TB = 5 |

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

water

| | | | | |
|----|----------------|----|-----------------|----|
| 1 | EB-8-8/12/04 ✓ | 11 | 0403223-MB-210/ | 31 |
| 2 | MW-22-1 ✓ | 12 | | 32 |
| 3 | MW-22-2 ✓ | 13 | | 33 |
| 4 | MW-22-3 ✓ | 14 | | 34 |
| 5 | TB-8-8/12/04 ✓ | 15 | | 35 |
| 6 | MW-22-3MS | 16 | | 36 |
| 7 | MW-22-3MSD | 17 | | 37 |
| 8 | | 18 | | 38 |
| 9 | | 19 | | 39 |
| 10 | | 20 | | 40 |

TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method 524.2)

| A. Chloromethane | G. 1,2-Dichloropropane | GG. Xylenes, total | VW. Bromobenzene | MMM. Naphthalene |
|------------------------------|-------------------------------|---------------------------------|-----------------------------|--|
| B. Bromomethane | R. cis-1,3-Dichloropropene | HH. Vinyl acetate | XX. 1,2,3-Trichloropropane | NNN. 1,2,3-Trichlorobenzene |
| C. Vinyl chloride | S. Trichloroethene | II. 2-Chloroethylvinyl ether | YY. n-Propylbenzene | OOO. 1,3,5-Trichlorobenzene |
| D. Chloroethane | T. Dibromochloromethane | JJ. Dichlorodifluoromethane | ZZ. 2-Chlorotoluene | PPP. trans-1,2-Dichloroethene |
| E. Methylene chloride | U. 1,1,2-Trichloroethane | KK. Trichlorofluoromethane | AAA. 1,3,5-Trimethylbenzene | QQQ. cis-1,2-Dichloroethene |
| F. Acetone | V. Benzene | LL. Methyl-tert-butyl ether | BBB. 4-Chlorotoluene | RRR. m,p-Xylenes |
| G. Carbon disulfide | W. trans-1,3-Dichloropropene | MM. 1,2-Dibromo-3-chloropropane | CCC. tert-Butylbenzene | SSS. o-Xylene |
| H. 1,1-Dichloroethane | X. Bromoform | NN. Diethyl ether | DDD. 1,2,4-Trimethylbenzene | TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane |
| I. 1,1-Dichloroethane | Y. 4-Methyl-2-pentanone | OO. 2,2-Dichloropropane | EEE. sec-Butylbenzene | UUU. Benzyl chloride |
| J. 1,2-Dichloroethene, total | Z. 2-Hexanone | PP. Bromochloromethane | FFF. 1,3-Dichlorobenzene | VVV. 4-Ethyltoluene |
| K. Chloroform | AA. Tetrachloroethene | QQ. 1,1-Dichloropropene | GGG. p-Isopropyltoluene | WWW. Ethanol |
| L. 1,2-Dichloroethane | BB. 1,1,2,2-Tetrachloroethane | RR. Dibromomethane | HHH. 1,4-Dichlorobenzene | XXX. Ethyl ether |
| M. 2-Butanone | CC. Toluene | SS. 1,3-Dichloropropane | III. n-Butylbenzene | |
| N. 1,1,1-Trichloroethane | DD. Chlorobenzene | TT. 1,2-Dibromoethane | JJJ. 1,2-Dichlorobenzene | |
| O. Carbon tetrachloride | EE. Ethylbenzene | UU. 1,1,1,2-Tetrachloroethane | KKK. 1,2,4-Trichlorobenzene | |
| P. Bromodichloromethane | FF. Styrene | VV. Isopropylbenzene | LLL. Hexachlorobutadiene | |

Notes: _____

LDC #: 12484A/
SDG #: 04-4044

VALIDATION FINDINGS WORKSHEET
Field Blanks

Page: 1 of 1
Reviewer: [Signature]
2nd reviewer: [Signature]

METHOD: GC/MS VOA (EPA SW-846 Method-⁵²⁴⁻²8260B)

Y N N/A Were field blanks identified in this SDG?
 Y N N/A Were target compounds detected in the field blanks?

Sample: 1 Field Blank / Trip Blank / Rinsate / Other EB (circle one)

| Compound | Concentration Units (ug/l) |
|----------|----------------------------|
| <u>E</u> | <u>0.5</u> |
| | |
| | |
| | |
| | |
| | |

Sample: 5 Field Blank / Trip Blank / Rinsate / Other TB (circle one)

| Compound | Concentration Units (ug/l) |
|----------|----------------------------|
| <u>M</u> | <u>1</u> |
| | |
| | |
| | |
| | |
| | |

Sample: _____ Field Blank / Trip Blank / Rinsate / Other _____ (circle one)

| Compound | Concentration Units () |
|----------|-------------------------|
| | |
| | |
| | |
| | |
| | |

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 16, 2004
LDC Report Date: September 27, 2004
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-4085

Sample Identification

DUPE-5-3Q04
MW-5
MW-6**
MW-16
TB-9-8/16/04

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 5 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 .

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples DUPE-5-3Q04 and MW-5 were identified as field duplicates. No volatiles were detected in any of the samples.

XVII. Field Blanks

Sample TB-9-8/16/04 was identified as a trip blank. No volatile contaminants were found in this blank.

NASA JPL
Volatiles - Data Qualification Summary - SDG 04-4085

No Sample Data Qualified in this SDG

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 04-4085

No Sample Data Qualified in this SDG

B

Applied P & CH Laboratories

Organic Analysis Results for Method 524.2

| | | | | | |
|-----------------|--------------|----------------|-----------|------------------|------------|
| Client Name: | GEOFON, Inc. | Project No: | 4-12812 | Collection Date: | 08/16/2004 |
| Project ID: | JPL GW-3Q04 | Service ID: | 44085 | Collected by: | JJ/MM |
| Sample ID: | DUPE-5-3Q04 | Lab Sample ID: | 04-4085-1 | Received Date: | 08/16/2004 |
| Sample Type: | Field Sample | Sample Matrix: | Water | Moisture %: | - |
| Anal. Method: | 524.2 | Prep. Method: | 5030 | Instrument ID: | GC/MS: A |
| Batch No: | 04G3223 | Prep. Date: | 08/18/04 | Anal. Date: | 08/18/04 |
| Data File Name: | 4085-01 | Prep. No: | - | Anal. Time: | 15:36 |
| Methanol Vol. | - | Sample Amount: | 25.0 mL | Dilution Factor: | 1 |
| Test Level: | Low | Sparge Size: | 25 mL | Heated Purge: | (Y/N) N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | <0.5 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROETHANE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROETHANE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROETHANE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | <0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

Handwritten signature

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | <0.5 | U |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | <0.5 | U |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROBENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROBENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | <0.5 | U |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,1,2,2-TRIFLUOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec.% | |
|---|-------------------------------------|------------------|--------------|-----|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 102 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 102 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 110 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 96 |
| # | of out-of-control | | | 0 |

Internal Standard

| | | Control Limit, % | IS Rec.% | |
|---|-----------------------|------------------|----------|----|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 | 77 |
| 2 | 1,4-DICHLOROETHANE-D4 | 3855-82-1 | 50-200 | 77 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 | 67 |
| # | of out-of-control | | | 0 |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

9/2/04

Applied P & CH Laboratories
Organic Analysis Results for Method 524.2

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/16/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44085 | Collected by: JJ/MM |
| Sample ID: MW-5 | Lab Sample ID: 04-4085-2 | Received Date: 08/16/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |
| Anal. Method: 524.2 | Prep. Method: 5030 | Instrument ID: GC/MS: A |
| Batch No: 04G3223 | Prep. Date: 08/18/04 | Anal. Date: 08/18/04 |
| Data File Name: 4085-02 | Prep. No: - | Anal. Time: 16:02 |
| Methanol Vol: - | Sample Amount: 25.0 mL | Dilution Factor: 1 |
| Test Level: Low | Sparge Size: 25 mL | Heated Purge: (Y/N) N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | <0.5 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | <0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---------------------------------------|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | <0.5 | U |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | <0.5 | U |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | <0.5 | U |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,2,2-TRIFLUOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec. % | |
|---|-------------------------------------|------------------|---------------|-----|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 101 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 99 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 109 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 97 |
| | # of out-of-control | | | 0 |

Internal Standard

| | | Control Limit, % | IS Rec. % | |
|---|-----------------------|------------------|-----------|----|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 | 76 |
| 2 | 1,4-DICHLOROENZENE-D4 | 3855-82-1 | 50-200 | 75 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 | 66 |
| | # of out-of-control | | | 0 |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Handwritten signature
9/27/04

Applied P & CH Laboratories
Organic Analysis Results for Method 524.2

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/16/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44085 | Collected by: JJ/MM |
| Sample ID: MW-6 | Lab Sample ID: 04-4085-3 | Received Date: 08/16/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |
| Anal. Method: 524.2 | Prep. Method: 5030 | Instrument ID: GC/MS: A |
| Batch No: 04G3223 | Prep. Date: 08/18/04 | Anal. Date: 08/18/04 |
| Data File Name: 4085-03 | Prep. No: - | Anal. Time: 16:29 |
| Methanol Vol: - | Sample Amount: 25.0 mL | Dilution Factor: 1 |
| Test Level: Low | Sparge Size: 25 mL | Heated Purge: (Y/N) N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | <0.5 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROENZENE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROENZENE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROENZENE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | 0.6 | |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

9/27/04

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | <0.5 | U |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | 1.1 | |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROBENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROBENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | <0.5 | U |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | 0.4 | J |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2,2-TRICHLORO-1,1,2,2-TRIFLUOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec.% | |
|---------------------|-------------------------------------|------------------|--------------|-----|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 102 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 104 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 112 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 96 |
| # of out-of-control | | | 0 | |

Internal Standard

| | | Control Limit, % | IS Rec.% | |
|---------------------|-----------------------|------------------|----------|----|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 | 77 |
| 2 | 1,4-DICHLOROETHANE-D4 | 3855-82-1 | 50-200 | 77 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 | 66 |
| # of out-of-control | | | 0 | |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Applied P & CH Laboratories
Organic Analysis Results for Method 524.2

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/16/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44085 | Collected by: JJ/MM |
| Sample ID: MW-16 | Lab Sample ID: 04-4085-5 | Received Date: 08/16/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |
| Anal. Method: 524.2 | Prep. Method: 5030 | Instrument ID: GC/MS: A |
| Batch No: 04G3223 | Prep. Date: 08/18/04 | Anal. Date: 08/18/04 |
| Data File Name: 4085-05 | Prep. No: - | Anal. Time: 16:55 |
| Methanol Vol: - | Sample Amount: 25.0 mL | Dilution Factor: 1 |
| Test Level: Low | Sparge Size: 25 mL | Heated Purge: (Y/N) N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | < 0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | < 0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | < 0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | < 0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | < 0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | < 0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | < 0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | < 0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | < 0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | < 10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | 4.0 | |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | < 0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | < 0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | < 0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | 5.1 | |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | < 0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | < 0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | < 0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | < 0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | < 0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | < 0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | < 0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | < 0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | < 0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | < 0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | < 0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | < 0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | 1.3 | |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | < 0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | < 0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | < 0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | < 0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | < 0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | < 0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | < 0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | < 0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | < 0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | < 0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | < 0.5 | U |

9/21/04

| # | Component Name | CAS No | Unit, | RL | Result | Qualifier |
|----|---------------------------------------|-----------|-------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | <0.5 | U |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | 0.5 | |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | 1.0 | |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,2,2-TRIFLUOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec.% | |
|---------------------|-------------------------------------|------------------|--------------|-----|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 101 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 103 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 111 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 96 |
| # of out-of-control | | | 0 | |

Internal Standard

| | | Control Limit, % | IS Rec.% | |
|---------------------|-----------------------|------------------|----------|----|
| 1 | CHLOROENZENE-D5 | 3114-55-4 | 50-200 | 76 |
| 2 | 1,4-DICHLOROENZENE-D4 | 3855-82-1 | 50-200 | 77 |
| 3 | FLUOROENZENE | 462-06-6 | 50-200 | 66 |
| # of out-of-control | | | 0 | |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

9/27/04

Applied P & CH Laboratories
Organic Analysis Results for Method 524.2

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/16/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44085 | Collected by: JJ/MM |
| Sample ID: TB-9-8/16/04 | Lab Sample ID: 04-4085-6 | Received Date: 08/16/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |
| Anal. Method: 524.2 | Prep. Method: 5030 | Instrument ID: GC/MS: A |
| Batch No: 04G3223 | Prep. Date: 08/18/04 | Anal. Date: 08/18/04 |
| Data File Name: 4085-06 | Prep. No: - | Anal. Time: 17:21 |
| Methanol Vol: - | Sample Amount: 25.0 mL | Dilution Factor: 1 |
| Test Level: Low | Sparge Size: 25 mL | Heated Purge: (Y/N) N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | <0.5 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | <0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---------------------------------------|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | <0.5 | U |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | <0.5 | U |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROBENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROBENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | <0.5 | U |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,2,2-TRIFLUOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec.% |
|---|-------------------------------------|------------------|--------------|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 |
| | # of out-of-control | | 0 |

Internal Standard

| | | Control Limit, % | IS Rec.% |
|---|------------------------|------------------|----------|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 |
| 2 | 1,4-DICHLOROBENZENE-D4 | 3855-82-1 | 50-200 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 |
| | # of out-of-control | | 0 |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

LDC #: 12484B1

VALIDATION COMPLETENESS WORKSHEET

SDG #: 04-4085

Level III/IV

Laboratory: Applied Physics & Chemistry Laboratory

Date: 9/23/04

Page: 1 of 1

Reviewer: [Signature]

2nd Reviewer: [Signature]

METHOD: GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

| | Validation Area | | Comments |
|-------|--|----|--|
| I. | Technical holding times | Δ | Sampling dates: 8/16/04 |
| II. | GC/MS Instrument performance check | Δ | |
| III. | Initial calibration | Δ | % RSD, r ² 10.990 |
| IV. | Continuing calibration | Δ | |
| V. | Blanks | Δ | |
| VI. | Surrogate spikes | A | |
| VII. | Matrix spike/Matrix spike duplicates | A | MW - 22 - 3 |
| VIII. | Laboratory control samples | A | LCS |
| IX. | Regional Quality Assurance and Quality Control | N | |
| X. | Internal standards | A | |
| XI. | Target compound identification | Δ | Not reviewed for Level III validation. |
| XII. | Compound quantitation/CRQLs | Δ | Not reviewed for Level III validation. |
| XIII. | Tentatively identified compounds (TICs) | N | Not reviewed for Level III validation. |
| XIV. | System performance | Δ | Not reviewed for Level III validation. |
| XV. | Overall assessment of data | A | |
| XVI. | Field duplicates | ND | D = 1 + 2 |
| XVII. | Field blanks | ND | TB = 5 |

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: ** Indicates sample underwent Level IV validation

| | | | | | |
|----|-----------------|----|--------------|----|----|
| 1 | DUPE-5-3Q04 - D | 11 | 04G3223-MB01 | 31 | |
| 2 | MW-5 - D | 12 | | 22 | 32 |
| 3 | MW-6** | 13 | | 23 | 33 |
| 4 | MW-16 | 14 | | 24 | 34 |
| 5 | TB-9-8/16/04 | 15 | | 25 | 35 |
| 6 | | 16 | | 26 | 36 |
| 7 | | 17 | | 27 | 37 |
| 8 | | 18 | | 28 | 38 |
| 9 | | 19 | | 29 | 39 |
| 10 | | 20 | | 30 | 40 |

LDC #: 1248481
 SDG #: 04-4085

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2
 Reviewer: R
 2nd Reviewer: T

Method: Volatiles (EPA Method 524.2)

| Validation Area | Yes | No | NA | Findings/Comments |
|--|-----|----|----|-------------------|
| I. Technical holding times | | | | |
| All technical holding times were met. | ✓ | | | |
| Cooler temperature criteria was met. | ✓ | | | |
| II. GC/MS Instrument performance check | | | | |
| Were the BFB performance results reviewed and found to be within the specified criteria? | ✓ | | | |
| Were all samples analyzed within the 12 hour clock criteria? | ✓ | | | |
| III. Initial calibration | | | | |
| Did the laboratory perform a 5 point calibration prior to sample analysis? | ✓ | | | |
| Were all percent relative standard deviations (%RSD) ≤ 20%? | ✓ | | | |
| IV. Continuing calibration | | | | |
| Was a continuing calibration standard analyzed at least once every 12 hours for each instrument? | ✓ | | | |
| Were all percent differences (%D) ≤ 30%? | ✓ | | | |
| V. Blanks | | | | |
| Was a method blank associated with every sample in this SDG? | ✓ | | | |
| Was a method blank analyzed at least once every 12 hours for each matrix and concentration? | ✓ | | | |
| Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet. | | ✓ | | |
| VI. Surrogate spikes | | | | |
| Were all surrogate %R within QC limits? | ✓ | | | |
| If the percent recovery (%R) for one or more surrogates was out of QC limits, was a reanalysis performed to confirm samples with %R outside of criteria? | | | • | |
| VII. Matrix spike/Matrix spike duplicates | | | | |
| Was a matrix spike (MS) and matrix spike duplicate (MSD) analyzed for this SDG? | ✓ | | NA | |
| Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the QC limits? | ✓ | | NA | |
| VIII. Laboratory control samples | | | | |
| Was an LCS analyzed for this SDG? | ✓ | | | |
| Was an LCS analyzed per analytical batch? | ✓ | | | |
| Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the QC limits? | ✓ | | | |

LDC #: 12484B1
 SDG #: 04-4085

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

| Validation Area | Yes | No | NA | Findings/Comments |
|--|-----|----|----|-------------------|
| IX. Regional Quality Assurance and Quality Control | | | | |
| Were performance evaluation (PE) samples performed? | | | / | |
| Were the performance evaluation (PE) samples within the acceptance limits? | | | / | |
| X. Internal standards | | | | |
| Were internal standard area counts within +/-40% from the associated calibration standard? | / | | | |
| Were retention times within - 30% of the last continuing calibration or +/- 50% of the initial calibration? | / | | | |
| XI. Target compound identification | | | | |
| Were relative retention times (RRT's) within \pm 0.06 RRT units of the standard? | / | | | |
| Did compound spectra meet specified EPA "Functional Guidelines" criteria? | / | | | |
| Were chromatogram peaks verified and accounted for? | / | | | |
| XII. Compound quantitation/CRQLs | | | | |
| Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound? | / | | | |
| Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation? | / | | | |
| XIII. Tentatively identified compounds (TICs) | | | | |
| Were the major ions (> 25 percent relative intensity) in the reference spectrum evaluated in sample spectrum? | | | / | |
| Were relative intensities of the major ions within \pm 20% between the sample and the reference spectra? | | | / | |
| Did the raw data indicate that the laboratory performed a library search for all required peaks in the chromatograms (samples and blanks)? | | | / | |
| XIV. System performance | | | | |
| System performance was found to be acceptable. | / | | | |
| XV. Overall assessment of data | | | | |
| Overall assessment of data was found to be acceptable. | / | | | |
| XVI. Field duplicates | | | | |
| Field duplicate pairs were identified in this SDG. | / | | | |
| Target compounds were detected in the field duplicates. | | | / | |
| XVII. Field blanks | | | | |
| Field blanks were identified in this SDG. | / | | | |
| Target compounds were detected in the field blanks. | | | / | |

TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method 524.2)

| A. Chloromethane | Q. 1,2-Dichloropropane | GG. Xylenes, total | WW. Bromobenzene | MMM. Naphthalene |
|------------------------------|-------------------------------|---------------------------------|-----------------------------|--|
| B. Bromomethane | R. cis-1,3-Dichloropropane | HH. Vinyl acetate | XX. 1,2,3-Trichloropropane | NNN. 1,2,3-Trichlorobenzene |
| C. Vinyl chloride | S. Trichloroethene | II. 2-Chloroethylvinyl ether | YY. n-Propylbenzene | OOO. 1,3,5-Trichlorobenzene |
| D. Chloroethane | T. Dibromochloromethane | JJ. Dichlorodifluoromethane | ZZ. 2-Chlorotoluene | PPP. trans-1,2-Dichloroethene |
| E. Methylene chloride | U. 1,1,2-Trichloroethane | KK. Trichlorofluoromethane | AAA. 1,3,5-Trimethylbenzene | QQQ. cis-1,2-Dichloroethene |
| F. Acetone | V. Benzene | LL. Methyl-tert-butyl ether | BBB. 4-Chlorotoluene | RRR. m,p-Xylenes |
| G. Carbon disulfide | W. trans-1,3-Dichloropropane | MM. 1,2-Dibromo-3-chloropropane | CCC. tert-Butylbenzene | SSS. o-Xylene |
| H. 1,1-Dichloroethene | X. Bromoform | NN. Diethyl ether | DDD. 1,2,4-Trimethylbenzene | TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane |
| I. 1,1-Dichloroethane | Y. 4-Methyl-2-pentanone | OO. 2,2-Dichloropropane | EEE. sec-Butylbenzene | UUU. Benzyl chloride |
| J. 1,2-Dichloroethene, total | Z. 2-Hexanone | PP. Bromochloromethane | FFF. 1,3-Dichlorobenzene | VVV. 4-Ethyltoluene |
| K. Chloroform | AA. Tetrachloroethene | QQ. 1,1-Dichloropropane | GGG. p-Isopropyltoluene | WWW. Ethanol |
| L. 1,2-Dichloroethane | BB. 1,1,2,2-Tetrachloroethane | RR. Dibromomethane | HHH. 1,4-Dichlorobenzene | XXX. Ethyl ether |
| M. 2-Butanone | CC. Toluene | SS. 1,3-Dichloropropane | III. n-Butylbenzene | |
| N. 1,1,1-Trichloroethane | DD. Chlorobenzene | TT. 1,2-Dibromoethane | JJJ. 1,2-Dichlorobenzene | |
| O. Carbon tetrachloride | EE. Ethylbenzene | UU. 1,1,1,2-Tetrachloroethane | KKK. 1,2,4-Trichlorobenzene | |
| P. Bromodichloromethane | FF. Styrene | VV. Isopropylbenzene | LLL. Hexachlorobutadiene | |

Notes:

LDC #: 1248421
 SDG #: 04-4085

VALIDATION FINDINGS WORKSHEET
Initial Calibration Calculation Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

The Relative Response Factor (RRF), average RRF, and percent relative standard deviation (%RSD) were recalculated for the compounds identified below using the following calculations:

$RRF = (A_x)(C_{is}) / (A_{is})(C_x)$
 average RRF = sum of the RRFs/number of standards
 $\%RSD = 100 * (S/X)$

A_x = Area of compound, A_{is} = Area of associated internal standard
 C_x = Concentration of compound, C_{is} = Concentration of internal standard
 S = Standard deviation of the RRFs
 X = Mean of the RRFs

| # | Standard ID | Calibration Date | Compound (Reference Internal Standard) | Reported | | Recalculated | | Reported | | Recalculated | |
|---|---------------|------------------|--|--------------|--------------|--------------|-----------------------|-----------------------|------|--------------|------|
| | | | | RRF (10 std) | RRF (10 std) | RRF (10 std) | Average RRF (Initial) | Average RRF (Initial) | %RSD | %RSD | |
| 1 | 524.2 16A2 | 4/22/04 | Bromomethane | 0.099 | 0.099 | 0.104 | 0.104 | 4.77 | 4.77 | 4.77 | 4.77 |
| | | | Methylene chloride (1st internal standard) | 1.497 | 1.497 | 1.516 | 1.516 | 4.06 | 4.06 | 4.06 | 4.06 |
| | | | Toluene | 3.301 | 3.301 | 3.345 | 3.345 | 3.96 | 3.96 | 3.96 | 3.96 |
| | | | Trichlorethene (2nd internal standard) | | | | | | | | |
| 2 | | | Bromoform (3rd internal standard) | | | | | | | | |
| | | | Methylene chloride (1st internal standard) | | | | | | | | |
| | | | Trichlorethene (2nd internal standard) | | | | | | | | |
| | | | Bromoform (3rd internal standard) | | | | | | | | |
| 3 | | | Methylene chloride (1st internal standard) | | | | | | | | |
| | | | Trichlorethene (2nd internal standard) | | | | | | | | |
| | | | Bromoform (3rd internal standard) | | | | | | | | |
| | | | Methylene chloride (1st internal standard) | | | | | | | | |
| 4 | | | Trichlorethene (2nd internal standard) | | | | | | | | |
| | | | Bromoform (3rd internal standard) | | | | | | | | |
| | | | Methylene chloride (1st internal standard) | | | | | | | | |
| | | | Trichlorethene (2nd internal standard) | | | | | | | | |

Comments: Refer to Initial Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12484131
 SDG #: 04-4085

VALIDATION FINDINGS WORKSHEET
Continuing Calibration Results Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent difference (%D) of the initial calibration average Relative Response Factors (RRFs) and the continuing calibration RRFs were recalculated for the compounds identified below using the following calculation:

% Difference = $100 * (\text{ave. RRF} - \text{RRF}) / \text{ave. RRF}$ Where: ave. RRF = initial calibration average RRF
 RRF = $(A_x)(C_s) / (A_s)(C_x)$ RRF = continuing calibration RRF
 A_x = Area of compound, A_s = Area of associated internal standard
 C_x = Concentration of compound, C_s = Concentration of internal standard

| # | Standard ID | Calibration Date | Compound (Reference Internal Standard) | Average RRF (Initial) | Reported | | Recalculated | |
|---|-------------|------------------|--|-----------------------|----------|------|--------------|----|
| | | | | | RRF (CC) | %D | RRF (CC) | %D |
| 1 | 04G3223 | 8/18/04 | Methylene chloride (1st Internal Standard) Trichloroethene (2nd internal standard) Bromoform (3rd internal standard) Bromoform (3rd internal standard) | 0.104 | 0.120 | 15.4 | 15.4 | |
| | | | Methylene chloride (1st Internal Standard) | 1.514 | 1.440 | 5.0 | 5.0 | |
| | | | Trichloroethene (2nd internal standard) | 3.345 | 3.306 | 1.2 | 1.2 | |
| 2 | | | Methylene chloride (1st Internal Standard) | | | | | |
| | | | Trichloroethene (2nd internal standard) | | | | | |
| | | | Bromoform (3rd internal standard) | | | | | |
| 3 | | | Methylene chloride (1st Internal Standard) | | | | | |
| | | | Trichloroethene (2nd internal standard) | | | | | |
| | | | Bromoform (3rd internal standard) | | | | | |
| 4 | | | Methylene chloride (1st Internal Standard) | | | | | |
| | | | Trichloroethene (2nd internal standard) | | | | | |
| | | | Bromoform (3rd internal standard) | | | | | |

Comments: Refer to Continuing Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12484B1
 SDG #: 04-4085

VALIDATION FINDINGS WORKSHEET
Surrogate Results Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) of surrogates were recalculated for the compounds identified below using the following calculation:

% Recovery: $SF/SS * 100$

Where: SF = Surrogate Found
 SS = Surrogate Spiked

Sample ID: #3

| | Surrogate Spiked | Surrogate Found | Percent Recovery | Percent Recovery | Percent Difference |
|------------------------------|------------------|-----------------|------------------|------------------|--------------------|
| | | | Reported | Recalculated | |
| Toluene-d8 | 20 | 19.1 | 96 | 96 | 0 |
| Bromofluorobenzene | | 20.2 | 102 | 102 | |
| 1,2-Dichlorobenzene-d4 | | 20.8 | 104 | 104 | |
| <i>Di bromofluorobenzene</i> | | 22.4 | 112 | 112 | |

Sample ID: _____

| | Surrogate Spiked | Surrogate Found | Percent Recovery | Percent Recovery | Percent Difference |
|------------------------|------------------|-----------------|------------------|------------------|--------------------|
| | | | Reported | Recalculated | |
| Toluene-d8 | | | | | |
| Bromofluorobenzene | | | | | |
| 1,2-Dichlorobenzene-d4 | | | | | |

Sample ID: _____

| | Surrogate Spiked | Surrogate Found | Percent Recovery | Percent Recovery | Percent Difference |
|------------------------|------------------|-----------------|------------------|------------------|--------------------|
| | | | Reported | Recalculated | |
| Toluene-d8 | | | | | |
| Bromofluorobenzene | | | | | |
| 1,2-Dichlorobenzene-d4 | | | | | |

Sample ID: _____

| | Surrogate Spiked | Surrogate Found | Percent Recovery | Percent Recovery | Percent Difference |
|------------------------|------------------|-----------------|------------------|------------------|--------------------|
| | | | Reported | Recalculated | |
| Toluene-d8 | | | | | |
| Bromofluorobenzene | | | | | |
| 1,2-Dichlorobenzene-d4 | | | | | |

Sample ID: _____

| | Surrogate Spiked | Surrogate Found | Percent Recovery | Percent Recovery | Percent Difference |
|------------------------|------------------|-----------------|------------------|------------------|--------------------|
| | | | Reported | Recalculated | |
| Toluene-d8 | | | | | |
| Bromofluorobenzene | | | | | |
| 1,2-Dichlorobenzene-d4 | | | | | |

LDC #: 12484B
 SDG #: 04-4005

VALIDATION FINDINGS WORKSHEET
Matrix Spike/Matrix Spike Duplicates Results Verification

Page: 6 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the matrix spike and matrix spike duplicate were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * (SSC - SC) / SA$ Where: SSC = Spiked sample concentration SC = Sample concentration
 SA = Spike added

RPD = $1 MSC - MSDC | * 2 / (MSC + MSDC)$ MSC = Matrix spike percent recovery MSDC = Matrix spike duplicate percent recovery

MS/MSD sample: MW - 22 - 3

| Compound | Spike Added (ug/L) | | Sample Concentration (ug/L) | Spiked Sample Concentration (ug/L) | | Matrix Spike Percent Recovery | | Matrix Spike Duplicate Percent Recovery | | MS/MSD RPD | |
|--------------------|--------------------|-----|-----------------------------|------------------------------------|------|-------------------------------|---------|---|---------|------------|--------------|
| | MS | MSD | | MS | MSD | Reported | Recalc. | Reported | Recalc. | Reported | Recalculated |
| 1,1-Dichloroethene | 20 | 20 | 0 | 17.2 | 18.9 | 86 | 86 | 95 | 95 | 10 | 9.4 |
| Trichloroethene | | | | 18.7 | 20.5 | 94 | 94 | 103 | 103 | 9 | 9 |
| Benzene | | | | 18.5 | 20.4 | 93 | 93 | 102 | 102 | 9 | 10 |
| Toluene | | | | 19.4 | 21.3 | 97 | 97 | 107 | 107 | 10 | 9 |
| Chlorobenzene | | | | 19.7 | 21.4 | 99 | 99 | 107 | 107 | 8 | 8 |

Comments: Refer to Matrix Spike/Matrix Spike Duplicates findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

VALIDATION FINDINGS WORKSHEET
Laboratory Control Sample Results Verification

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the laboratory control sample and laboratory control sample duplicate (if applicable) were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * \frac{SSC}{SA}$ Where: SSC = Spiked sample concentration
 SA = Spike added

RPD = $100 * \frac{LCS - LCS2}{LCS + LCS2}$ LCS = Laboratory control sample percent recovery LCS2 = Laboratory control sample duplicate percent recovery

LCS ID: LCS

| Compound | Spike Added (ug/L) | | Spiked Sample Concentration (ug/L) | | LCS | | LCS2 | |
|--------------------|--------------------|------|------------------------------------|------|----------|---------|----------|---------|
| | LCS | LCS2 | LCS | LCS2 | Reported | Recalc. | Reported | Recalc. |
| 1,1-Dichloroethene | 20 | NA | 19.8 19.9 | NA | 95 | 95 | | |
| Trichloroethene | | | 20.2 | | 101 | 101 | | |
| Benzene | | | 19.8 | | 99 | 99 | | |
| Toluene | | | 20.6 | | 103 | 103 | | |
| Chlorobenzene | | | 20.7 | | 104 | 104 | NA | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |
| | | | | | | | | |

Comments: Refer to Laboratory Control Sample findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 17484B1
 SDG #: 04-4085

VALIDATION FINDINGS WORKSHEET

Sample Calculation Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

Y N N/A Were all reported results recalculated and verified for all level IV samples?
Y N N/A Were all recalculated results for detected target compounds agree within 10.0% of the reported results?

$$\text{Concentration} = \frac{(A_x)(I_s)(DF)}{(A_s)(RRF)(V_o)(\%S)}$$

- A_x = Area of the characteristic ion (EICP) for the compound to be measured
- A_s = Area of the characteristic ion (EICP) for the specific internal standard
- I_s = Amount of internal standard added in nanograms (ng)
- RRF = Relative response factor of the calibration standard.
- V_o = Volume or weight of sample pruged in milliliters (ml) or grams (g).
- Df = Dilution factor.
- %S = Percent solids, applicable to soils and solid matrices only.

Example:

Sample I.D. #3, 1,1: DEA

Conc. = $\frac{(346859)(10)}{(346859)(0.464)}$

= 0.6 ug/l2

| # | Sample ID | Compound | Reported Concentration () | Calculated Concentration () | Qualification |
|---|-----------|----------|-------------------------------|---------------------------------|---------------|
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 17, 2004
LDC Report Date: September 27, 2004
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-4096

Sample Identification

DUPE-6-3Q04
MW-8**
MW-10
MW-13
TB-10-8/17/04

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 5 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for all compounds.

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% .

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks.

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples DUPE-6-3Q04 and MW-10 were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

| Compound | Concentration (ug/L) | | RPD |
|--------------------|----------------------|-------|-----|
| | DUPE-6-3Q04 | MW-10 | |
| 1,1-Dichloroethane | 1 | 0.9 | 11 |
| Tetrachloroethane | 1.8 | 1.5 | 18 |

| Compound | Concentration (ug/L) | | RPD |
|-----------------|----------------------|-------|-----|
| | DUPE-6-3Q04 | MW-10 | |
| Trichloroethene | 16.6 | 14.6 | 13 |
| Chloroform | 1.4 | 1.3 | 7 |

XVII. Field Blanks

Sample TB-10-8/17/04 was identified as a trip blank. No volatile contaminants were found in this blank.

NASA JPL
Volatiles - Data Qualification Summary - SDG 04-4096

No Sample Data Qualified in this SDG

NASA JPL
Volatiles - Laboratory Blank Data Qualification Summary - SDG 04-4096

No Sample Data Qualified in this SDG

C

Applied P & CH Laboratories
Organic Analysis Results for Method 524.2

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/17/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44096 | Collected by: JJ/MM/TM |
| Sample ID: DUPE-6-3Q04 | Lab Sample ID: 04-4096-1 | Received Date: 08/17/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |
| Anal. Method: 524.2 | Prep. Method: 5030 | Instrument ID: GC/MS: A |
| Batch No: 04G3223 | Prep. Date: 08/18/04 | Anal. Date: 08/18/04 |
| Data File Name: 4096-01 | Prep. No: - | Anal. Time: 17:47 |
| Methanol Vol: - | Sample Amount: 25.0 mL | Dilution Factor: 1 |
| Test Level: Low | Sparge Size: 25 mL | Heated Purge: (Y/N) N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | 1.4 | |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | 1 | |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

Handwritten signature

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | <0.5 | U |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | 1.8 | |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROBENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROBENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | 16.6 | |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,1,2,2-TETRACHLOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec.% | |
|---------------------|-------------------------------------|------------------|--------------|-----|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 100 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 102 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 111 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 99 |
| # of out-of-control | | | 0 | |

Internal Standard

| | | Control Limit, % | IS Rec.% | |
|---------------------|------------------------|------------------|----------|----|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 | 73 |
| 2 | 1,4-DICHLOROBENZENE-D4 | 3855-82-1 | 50-200 | 75 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 | 64 |
| # of out-of-control | | | 0 | |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

gator

Organic Analysis Results for Method 524.2

| | | | | | |
|-----------------|--------------|----------------|-----------|---------------------|------------|
| Client Name: | GEOFON, Inc. | Project No: | 4-12812 | Collection Date: | 08/17/2004 |
| Project ID: | JPL GW-3Q04 | Service ID: | 44096 | Collected by: | JJ/MM/TM |
| Sample ID: | MW-8 | Lab Sample ID: | 04-4096-2 | Received Date: | 08/17/2004 |
| Sample Type: | Field Sample | Sample Matrix: | Water | Moisture %: | - |
| Anal. Method: | 524.2 | Prep. Method: | 5030 | Instrument ID: | GC/MS: A |
| Batch No: | 04G3223 | Prep. Date: | 08/18/04 | Anal. Date: | 08/18/04 |
| Data File Name: | 4096-02 | Prep. No: | - | Anal. Time: | 18:13 |
| Methanol Vol. | - | Sample Amount: | 25.0 mL | Dilution Factor: | 1 |
| Test Level: | Low | Sparge Size: | 25 mL | Heated Purge: (Y/N) | N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | <0.5 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | <0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | <0.5 | U |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | <0.5 | U |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | <0.5 | U |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,1,2,2-TETRACHLOROETHANE | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

| Surrogates | | | Control Limit, % | Surro. Rec.% |
|---------------------|-------------------------------------|------------|------------------|--------------|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 103 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 108 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 114 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 96 |
| # of out-of-control | | | | 0 |

| Internal Standard | | | Control Limit, % | IS Rec.% |
|---------------------|-----------------------|-----------|------------------|----------|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 | 75 |
| 2 | 1,4-DICHLOROENZENE-D4 | 3855-82-1 | 50-200 | 75 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 | 64 |
| # of out-of-control | | | | 0 |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL
 J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)
 E - Exceed calibration range
 B - A positive value was found in the method blank
 D - Diluted

9/27/04

Organic Analysis Results for Method 524.2

| | | | | | |
|-----------------|--------------|----------------|-----------|---------------------|------------|
| Client Name: | GEOFON, Inc. | Project No: | 4-12812 | Collection Date: | 08/17/2004 |
| Project ID: | JPL GW-3Q04 | Service ID: | 44096 | Collected by: | JJ/MM/TM |
| Sample ID: | MW-10 | Lab Sample ID: | 04-4096-3 | Received Date: | 08/17/2004 |
| Sample Type: | Field Sample | Sample Matrix: | Water | Moisture %: | - |
| Anal. Method: | 524.2 | Prep. Method: | 5030 | Instrument ID: | GC/MS: A |
| Batch No: | 04G3223 | Prep. Date: | 08/18/04 | Anal. Date: | 08/18/04 |
| Data File Name: | 4096-03 | Prep. No: | - | Anal. Time: | 18:39 |
| Methanol Vol. | - | Sample Amount: | 25.0 mL | Dilution Factor: | 1 |
| Test Level: | Low | Sparge Size: | 25 mL | Heated Purge: (Y/N) | N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | 1.3 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | 0.9 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

| # | Component Name | CAS No | Unit, | RL | Result | Qualifier |
|----|---------------------------------|-----------|-------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | <0.5 | U |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | 1.5 | |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROBENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROBENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | 14.6 | |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,2,2-TRIFLUORO | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec.% | |
|---------------------|-------------------------------------|------------------|--------------|-----|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 102 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 106 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 114 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 97 |
| # of out-of-control | | | | 0 |

Internal Standard

| | | Control Limit, % | IS Rec.% | |
|---------------------|-----------------------|------------------|----------|----|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 | 76 |
| 2 | 1,4-DICHLOROETHANE-D4 | 3855-82-1 | 50-200 | 75 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 | 64 |
| # of out-of-control | | | | 0 |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

Handwritten signature/initials

Applied P & CH Laboratories
Organic Analysis Results for Method 524.2

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/17/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44096 | Collected by: JJ/MM/TM |
| Sample ID: MW-13 | Lab Sample ID: 04-4096-4 | Received Date: 08/17/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |
| Anal. Method: 524.2 | Prep. Method: 5030 | Instrument ID: GC/MS: A |
| Batch No: 04G3223 | Prep. Date: 08/18/04 | Anal. Date: 08/18/04 |
| Data File Name: 4096-04 | Prep. No: - | Anal. Time: 19:05 |
| Methanol Vol: - | Sample Amount: 25.0 mL | Dilution Factor: 1 |
| Test Level: Low | Sparge Size: 25 mL | Heated Purge: (Y/N) N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | 2.0 | |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | 3.5 | |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | <0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---------------------------------|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | <0.5 | U |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | 0.9 | |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROBENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROBENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | 15.4 | |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,2,2-TRIFLUORO | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec. % | |
|---------------------|-------------------------------------|------------------|---------------|-----|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 104 |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 104 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 113 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 98 |
| # of out-of-control | | | 0 | |

Internal Standard

| | | Control Limit, % | IS Rec. % | |
|---------------------|-----------------------|------------------|-----------|----|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 | 73 |
| 2 | 1,4-DICHLOROETHANE-D4 | 3855-82-1 | 50-200 | 71 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 | 62 |
| # of out-of-control | | | 0 | |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

g
9/18/04

Organic Analysis Results for Method 524.2

| | | | | | |
|-----------------|---------------|----------------|-----------|---------------------|------------|
| Client Name: | GEOFON, Inc. | Project No: | 4-12812 | Collection Date: | 08/17/2004 |
| Project ID: | JPL GWJ3Q04 | Service ID: | 44096 | Collected by: | JJ/MM/TM |
| Sample ID: | TB-10-8/17/04 | Lab Sample ID: | 04-4096-5 | Received Date: | 08/17/2004 |
| Sample Type: | Field Sample | Sample Matrix: | Water | Moisture %: | - |
| Anal. Method: | 524.2 | Prep. Method: | 5030 | Instrument ID: | GC/MS: A |
| Batch No: | 04G3223 | Prep. Date: | 08/18/04 | Anal. Date: | 08/18/04 |
| Data File Name: | 4096-05 | Prep. No: | - | Anal. Time: | 19:31 |
| Methanol Vol. | - | Sample Amount: | 25.0 mL | Dilution Factor: | 1 |
| Test Level: | Low | Sparge Size: | 25 mL | Heated Purge: (Y/N) | N |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|-----------------------------|------------|------|-----|--------|-----------|
| 1 | BENZENE | 71-43-2 | µg/L | 0.5 | <0.5 | U |
| 2 | BROMOBENZENE | 108-86-1 | µg/L | 0.5 | <0.5 | U |
| 3 | BROMOCHLOROMETHANE | 74-97-5 | µg/L | 0.5 | <0.5 | U |
| 4 | BROMODICHLOROMETHANE | 75-27-4 | µg/L | 0.5 | <0.5 | U |
| 5 | BROMOFORM | 75-25-2 | µg/L | 0.5 | <0.5 | U |
| 6 | BROMOMETHANE | 74-83-9 | µg/L | 0.5 | <0.5 | U |
| 7 | N-BUTYLBENZENE | 104-51-8 | µg/L | 0.5 | <0.5 | U |
| 8 | SEC-BUTYLBENZENE | 135-98-8 | µg/L | 0.5 | <0.5 | U |
| 9 | TERT-BUTYLBENZENE | 98-06-6 | µg/L | 0.5 | <0.5 | U |
| 10 | 2-BUTANONE | 78-93-3 | µg/L | 10 | <10 | U |
| 11 | CARBON TETRACHLORIDE | 56-23-5 | µg/L | 0.5 | <0.5 | U |
| 12 | CHLOROBENZENE | 108-90-7 | µg/L | 0.5 | <0.5 | U |
| 13 | CHLORODIBROMOMETHANE | 124-48-1 | µg/L | 0.5 | <0.5 | U |
| 14 | CHLOROETHANE | 75-00-3 | µg/L | 0.5 | <0.5 | U |
| 15 | CHLOROFORM | 67-66-3 | µg/L | 0.5 | <0.5 | U |
| 16 | CHLOROMETHANE | 74-87-3 | µg/L | 0.5 | <0.5 | U |
| 17 | 2-CHLOROTOLUENE | 95-49-8 | µg/L | 0.5 | <0.5 | U |
| 18 | 4-CHLOROTOLUENE | 106-43-4 | µg/L | 0.5 | <0.5 | U |
| 19 | 1,2-DIBROMO-3-CHLOROPROPANE | 96-12-8 | µg/L | 0.5 | <0.5 | U |
| 20 | 1,2-DIBROMOETHANE (EDB) | 106-93-4 | µg/L | 0.5 | <0.5 | U |
| 21 | DIBROMOMETHANE | 74-95-3 | µg/L | 0.5 | <0.5 | U |
| 22 | 1,2-DICHLOROBENZENE | 95-50-1 | µg/L | 0.5 | <0.5 | U |
| 23 | 1,3-DICHLOROBENZENE | 541-73-1 | µg/L | 0.5 | <0.5 | U |
| 24 | 1,4-DICHLOROBENZENE | 106-46-7 | µg/L | 0.5 | <0.5 | U |
| 25 | DICHLORODIFLUOROMETHANE | 75-71-8 | µg/L | 0.5 | <0.5 | U |
| 26 | 1,1-DICHLOROETHANE | 75-34-3 | µg/L | 0.5 | <0.5 | U |
| 27 | 1,2-DICHLOROETHANE | 107-06-2 | µg/L | 0.5 | <0.5 | U |
| 28 | 1,1-DICHLOROETHENE | 75-35-4 | µg/L | 0.5 | <0.5 | U |
| 29 | CIS-1,2-DICHLOROETHENE | 156-59-2 | µg/L | 0.5 | <0.5 | U |
| 30 | TRANS-1,2-DICHLOROETHENE | 156-60-5 | µg/L | 0.5 | <0.5 | U |
| 31 | 1,2-DICHLOROPROPANE | 78-87-5 | µg/L | 0.5 | <0.5 | U |
| 32 | 1,3-DICHLOROPROPANE | 142-28-9 | µg/L | 0.5 | <0.5 | U |
| 33 | 2,2-DICHLOROPROPANE | 594-20-7 | µg/L | 0.5 | <0.5 | U |
| 34 | 1,1-DICHLOROPROPENE | 563-58-6 | µg/L | 0.5 | <0.5 | U |
| 35 | CIS-1,3-DICHLOROPROPENE | 10061-01-5 | µg/L | 0.5 | <0.5 | U |
| 36 | TRANS-1,3-DICHLOROPROPENE | 10061-02-6 | µg/L | 0.5 | <0.5 | U |
| 37 | ETHYLBENZENE | 100-41-4 | µg/L | 0.5 | <0.5 | U |
| 38 | HEXACHLOROBUTADIENE | 87-68-3 | µg/L | 0.5 | <0.5 | U |
| 39 | ISOPROPYLBENZENE (CUMENE) | 98-82-8 | µg/L | 0.5 | <0.5 | U |

| # | Component Name | CAS No | Unit | RL | Result | Qualifier |
|----|---------------------------------|-----------|------|-----|--------|-----------|
| 40 | P-ISOPROPYLTOLUENE | 99-87-6 | µg/L | 0.5 | <0.5 | U |
| 41 | METHYLENE CHLORIDE | 75-09-2 | µg/L | 0.5 | <0.5 | U |
| 42 | METHYL-T-BUTYL ETHER (MTBE) | 1634-04-4 | µg/L | 1 | <1 | U |
| 43 | 4-METHYL-2-PENTANONE (MIBK) | 108-10-1 | µg/L | 10 | <10 | U |
| 44 | NAPHTHALENE | 91-20-3 | µg/L | 0.5 | <0.5 | U |
| 45 | N-PROPYLBENZENE | 103-65-1 | µg/L | 0.5 | <0.5 | U |
| 46 | STYRENE | 100-42-5 | µg/L | 0.5 | <0.5 | U |
| 47 | 1,1,1,2-TETRACHLOROETHANE | 630-20-6 | µg/L | 0.5 | <0.5 | U |
| 48 | 1,1,2,2-TETRACHLOROETHANE | 79-34-5 | µg/L | 0.5 | <0.5 | U |
| 49 | TETRACHLOROETHENE | 127-18-4 | µg/L | 0.5 | <0.5 | U |
| 50 | TOLUENE | 108-88-3 | µg/L | 0.5 | <0.5 | U |
| 51 | 1,2,3-TRICHLOROBENZENE | 87-61-6 | µg/L | 0.5 | <0.5 | U |
| 52 | 1,2,4-TRICHLOROBENZENE | 120-82-1 | µg/L | 0.5 | <0.5 | U |
| 53 | 1,1,1-TRICHLOROETHANE | 71-55-6 | µg/L | 0.5 | <0.5 | U |
| 54 | 1,1,2-TRICHLOROETHANE | 79-00-5 | µg/L | 0.5 | <0.5 | U |
| 55 | TRICHLOROETHENE | 79-01-6 | µg/L | 0.5 | <0.5 | U |
| 56 | TRICHLOROFLUOROMETHANE | 75-69-4 | µg/L | 0.5 | <0.5 | U |
| 57 | 1,2,3-TRICHLOROPROPANE | 96-18-4 | µg/L | 0.5 | <0.5 | U |
| 58 | 1,1,2-TRICHLORO-1,2,2-TRIFLUORO | 76-13-1 | µg/L | 0.5 | <0.5 | U |
| 59 | 1,2,4-TRIMETHYLBENZENE | 95-63-6 | µg/L | 0.5 | <0.5 | U |
| 60 | 1,3,5-TRIMETHYLBENZENE | 108-67-8 | µg/L | 0.5 | <0.5 | U |
| 61 | VINYL CHLORIDE | 75-01-4 | µg/L | 0.5 | <0.5 | U |
| 62 | O-XYLENE | 95-47-6 | µg/L | 0.5 | <0.5 | U |
| 63 | M/P-XYLENE | 108-38-3 | µg/L | 0.5 | <0.5 | U |

Surrogates

| | | Control Limit, % | Surro. Rec.% | |
|---------------------|-------------------------------------|------------------|--------------|------|
| 1 | 1-BROMO-4-FLUOROBENZENE (4-BROMOFL) | 460-00-4 | 70-129 | 102. |
| 2 | 1,2-DICHLOROETHANE-D4 | 17060-07-0 | 70-129 | 101 |
| 3 | DIBROMOFLUOROMETHANE | 1868-53-7 | 70-122 | 112 |
| 4 | TOLUENE-D8 | 2037-26-5 | 73-129 | 98 |
| # of out-of-control | | | 0 | |

Internal Standard

| | | Control Limit, % | IS Rec.% | |
|---------------------|-----------------------|------------------|----------|----|
| 1 | CHLOROBENZENE-D5 | 3114-55-4 | 50-200 | 73 |
| 2 | 1,4-DICHLOROETHANE-D4 | 3855-82-1 | 50-200 | 73 |
| 3 | FLUOROBENZENE | 462-06-6 | 50-200 | 63 |
| # of out-of-control | | | 0 | |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Qualifier: U - Not Detected or less than MDL

J - Less than RL (PQL, EQL or CRDL), but greater than MDL, or an estimated result (e.g. for TIC)

E - Exceed calibration range

B - A positive value was found in the method blank

D - Diluted

g/m/04

LDC #: 12484C1

VALIDATION COMPLETENESS WORKSHEET

Date: 9/24/04

SDG #: 04-4096

Level III/IV

Page: 1 of 1

Laboratory: Applied Physics & Chemistry Laboratory

Reviewer: [Signature]

2nd Reviewer: [Signature]

METHOD: GC/MS Volatiles (EPA Method 524.2)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

| | Validation Area | | Comments |
|-------|--|----|---|
| I. | Technical holding times | A | Sampling dates: 8/17/04 |
| II. | GC/MS Instrument performance check | A | |
| III. | Initial calibration | Δ | |
| IV. | Continuing calibration | Δ | |
| V. | Blanks | Δ | |
| VI. | Surrogate spikes | A | |
| VII. | Matrix spike/Matrix spike duplicates | A | MW-22-3 |
| VIII. | Laboratory control samples | A | LC |
| IX. | Regional Quality Assurance and Quality Control | N | |
| X. | Internal standards | Δ | |
| XI. | Target compound identification | Δ | Not reviewed for Level III validation. |
| XII. | Compound quantitation/CRQLs | Δ | Not reviewed for Level III validation. |
| XIII. | Tentatively identified compounds (TICs) | N | Not reviewed for Level III validation. Not reported |
| XIV. | System performance | Δ | Not reviewed for Level III validation. |
| XV. | Overall assessment of data | A | |
| XVI. | Field duplicates | SW | D = 1 + 3 |
| XVII. | Field blanks | ND | TB = 5 |

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: ** Indicates sample underwent Level IV validation

| | | | | | |
|----|---------------|----|---------------|----|----|
| 1+ | DUPE-3Q04, D | 11 | 04G3223-MB-01 | 21 | 31 |
| 2 | MW-8** | 12 | | 22 | 32 |
| 3+ | MW-10, D | 13 | | 23 | 33 |
| 4+ | MW-13 | 14 | | 24 | 34 |
| 5 | TB-10-8/17/04 | 15 | | 25 | 35 |
| 6 | | 16 | | 26 | 36 |
| 7 | | 17 | | 27 | 37 |
| 8 | | 18 | | 28 | 38 |
| 9 | | 19 | | 29 | 39 |
| 10 | | 20 | | 30 | 40 |

LDC #: 12484C1
 SDG #: 04-4096

VALIDATION FINDINGS CHECKLIST

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

Method: Volatiles (EPA Method 524.2)

| Validation Area | Yes | No | NA | Findings/Comments |
|--|-----|----|----|-------------------|
| I. Technical holding times | | | | |
| All technical holding times were met. | ✓ | | | |
| Cooler temperature criteria was met. | ✓ | | | |
| II. GC/MS Instrument performance check | | | | |
| Were the BFB performance results reviewed and found to be within the specified criteria? | ✓ | | | |
| Were all samples analyzed within the 12 hour clock criteria? | ✓ | | | |
| III. Initial calibration | | | | |
| Did the laboratory perform a 5 point calibration prior to sample analysis? | ✓ | | | |
| Were all percent relative standard deviations (%RSD) ≤ 20%? | ✓ | | | |
| IV. Continuing calibration | | | | |
| Was a continuing calibration standard analyzed at least once every 12 hours for each instrument? | ✓ | | | |
| Were all percent differences (%D) ≤ 30%? | ✓ | | | |
| V. Blanks | | | | |
| Was a method blank associated with every sample in this SDG? | ✓ | | | |
| Was a method blank analyzed at least once every 12 hours for each matrix and concentration? | ✓ | | | |
| Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet. | | | ✓ | |
| VI. Surrogate spikes | | | | |
| Were all surrogate %R within QC limits? | ✓ | | | |
| If the percent recovery (%R) for one or more surrogates was out of QC limits, was a reanalysis performed to confirm samples with %R outside of criteria? | | | ✓ | |
| VII. Matrix spike/Matrix spike duplicates | | | | |
| Was a matrix spike (MS) and matrix spike duplicate (MSD) analyzed for this SDG? | ✓ | | | |
| Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the QC limits? | ✓ | | | |
| VIII. Laboratory control samples | | | | |
| Was an LCS analyzed for this SDG? | ✓ | | | |
| Was an LCS analyzed per analytical batch? | ✓ | | | |
| Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the QC limits? | ✓ | | | |

LDC #: 12484C1
 SDG #: 04-4096

VALIDATION FINDINGS CHECKLIST

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

| Validation Area | Yes | No | NA | Findings/Comments |
|--|-----|----|----|-------------------|
| IX. Regional Quality Assurance and Quality Control | | | | |
| Were performance evaluation (PE) samples performed? | | | / | |
| Were the performance evaluation (PE) samples within the acceptance limits? | | | / | |
| X. Internal standards | | | | |
| Were internal standard area counts within +/-40% from the associated calibration standard? | / | | | |
| Were retention times within - 30% of the last continuing calibration or +/- 50% of the initial calibration? | / | | | |
| XI. Target compound identification | | | | |
| Were relative retention times (RRT's) within ± 0.06 RRT units of the standard? | | | / | |
| Did compound spectra meet specified EPA "Functional Guidelines" criteria? | | | / | |
| Were chromatogram peaks verified and accounted for? | / | | | |
| XII. Compound quantitation/CRQLs | | | | |
| Were the correct internal standard (IS), quantitation ion and relative response factor (RRF) used to quantitate the compound? | | | / | |
| Were compound quantitation and CRQLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation? | / | | | |
| XIII. Tentatively identified compounds (TICs) | | | | |
| Were the major ions (> 25 percent relative intensity) in the reference spectrum evaluated in sample spectrum? | | | / | |
| Were relative intensities of the major ions within $\pm 20\%$ between the sample and the reference spectra? | | | / | |
| Did the raw data indicate that the laboratory performed a library search for all required peaks in the chromatograms (samples and blanks)? | | | / | |
| XIV. System performance | | | | |
| System performance was found to be acceptable. | / | | | |
| XV. Overall assessment of data | | | | |
| Overall assessment of data was found to be acceptable. | / | | | |
| XVI. Field duplicates | | | | |
| Field duplicate pairs were identified in this SDG. | / | | | |
| Target compounds were detected in the field duplicates. | / | | | |
| XVII. Field blanks | | | | |
| Field blanks were identified in this SDG. | / | | | |
| Target compounds were detected in the field blanks. | | / | | |

TARGET COMPOUND WORKSHEET

METHOD: VOA (EPA Method 524.2)

| A. Chloromethane | Q. 1,2-Dichloropropane | GG. Xylenes, total | WW. Bromobenzene | MMM. Naphthalene |
|------------------------------|-------------------------------|---------------------------------|-----------------------------|--|
| B. Bromomethane | R. cis-1,3-Dichloropropene | HH. Vinyl acetate | XX. 1,2,3-Trichloropropane | NNN. 1,2,3-Trichlorobenzene |
| C. Vinyl chloride | S. Trichloroethane | II. 2-Chloroethylvinyl ether | YY. n-Propylbenzene | OOO. 1,3,5-Trichlorobenzene |
| D. Chloroethane | T. Dibromochloromethane | JJ. Dichlorodifluoromethane | ZZ. 2-Chlorotoluene | PPP. trans-1,2-Dichloroethene |
| E. Methylene chloride | U. 1,1,2-Trichloroethane | KK. Trichlorofluoromethane | AAA. 1,3,5-Trimethylbenzene | QQQ. cis-1,2-Dichloroethene |
| F. Acetone | V. Benzene | LL. Methyl-tert-butyl ether | BBB. 4-Chlorotoluene | RRR. m,p-Xylenes |
| G. Carbon disulfide | W. trans-1,3-Dichloropropene | MM. 1,2-Dibromo-3-chloropropane | CCC. tert-Butylbenzene | SSS. o-Xylene |
| H. 1,1-Dichloroethane | X. Bromoform | NN. Diethyl ether | DDD. 1,2,4-Trimethylbenzene | TTT. 1,1,2-Trichloro-1,2,2-trifluoroethane |
| I. 1,1-Dichloroethane | Y. 4-Methyl-2-pentanone | OO. 2,2-Dichloropropane | EEE. sec-Butylbenzene | UUU. Benzyl chloride |
| J. 1,2-Dichloroethane, total | Z. 2-Hexanone | PP. Bromochloromethane | FFF. 1,3-Dichlorobenzene | VVV. 4-Ethyltoluene |
| K. Chloroform | AA. Tetrachloroethane | QQ. 1,1-Dichloropropene | GGG. p-Isopropyltoluene | WWW. Ethanol |
| L. 1,2-Dichloroethane | BB. 1,1,2,2-Tetrachloroethane | RR. Dibromomethane | HHH. 1,4-Dichlorobenzene | XXX. Ethyl ether |
| M. 2-Butanone | CC. Toluene | SS. 1,3-Dichloropropane | III. n-Butylbenzene | |
| N. 1,1,1-Trichloroethane | DD. Chlorobenzene | TT. 1,2-Dibromoethane | JJJ. 1,2-Dichlorobenzene | |
| O. Carbon tetrachloride | EE. Ethylbenzene | UU. 1,1,1,2-Tetrachloroethane | KKK. 1,2,4-Trichlorobenzene | |
| P. Bromodichloromethane | FF. Styrene | VV. Isopropylbenzene | LLL. Hexachlorobutadiene | |

Notes:

LDC#: 12484C1
SDG#: 04-4096

VALIDATION FINDINGS WORKSHEET

Field Duplicates

Page: 1 of 2
Reviewer: PN
2nd Reviewer: Q

METHOD: GC/MS VOA (EPA Method 524.2)

Y/N/NA Were field duplicate pairs identified in this SDG?
Y/N/NA Were target analytes detected in the field duplicate pairs?

| Compound | Concentration (ug/L) | | RPD | |
|--------------------|----------------------|------|-----|--|
| | 1 | 3 | | |
| 1,1-Dichloroethane | 1 | 0.9 | 11 | |
| Tetrachloroethane | 1.8 | 1.5 | 18 | |
| Trichloroethene | 16.6 | 14.6 | 13 | |
| Chloroform | 1.4 | 1.3 | 7 | |

VALIDATION FINDINGS WORKSHEET
Initial Calibration Calculation Verification

METHOD: GC/MS VOA (EPA Method 524.2)

The Relative Response Factor (RRF), average RRF, and percent relative standard deviation (%RSD) were recalculated for the compounds identified below using the following calculations:

$RRF = (A_x/C_x)/(A_s/C_s)$
 average RRF = sum of the RRFs/number of standards
 $\%RSD = 100 * (S/X)$
 A_x = Area of compound,
 C_x = Concentration of compound,
 S = Standard deviation of the RRFs
 X = Mean of the RRFs
 A_s = Area of associated internal standard
 C_s = Concentration of internal standard

| # | Standard ID | Calibration Date | Compound (Reference Internal Standard) | Reported | | Recalculated | | Reported | | Recalculated | |
|---|---------------|------------------|---|--------------|--------------|-----------------------|-----------------------|----------|------|--------------|------|
| | | | | RRF (10 std) | RRF (10 std) | Average RRF (Initial) | Average RRF (Initial) | %RSD | %RSD | | |
| 1 | 524.2 ICAL | 7/22/04 | Bromoethane Methylene chloride (1st Internal Standard) Toluene Trichloroethene (2nd internal standard) Ethyl Benzene Bromoform (3rd internal standard) | 0.099 | 0.099 | 0.104 | 0.104 | 4.77 | 4.77 | 4.77 | 4.77 |
| 2 | | | Methylene chloride (1st Internal Standard) Trichloroethene (2nd internal standard) Bromoform (3rd internal standard) | 1.497 | 1.497 | 1.516 | 1.516 | 4.06 | 4.06 | 4.06 | 4.06 |
| 3 | | | Methylene chloride (1st Internal Standard) Trichloroethene (2nd internal standard) Bromoform (3rd internal standard) | 3.301 | 3.301 | 3.345 | 3.345 | 3.96 | 3.96 | 3.96 | 3.96 |
| 4 | | | Methylene chloride (1st Internal Standard) Trichloroethene (2nd internal standard) Bromoform (3rd internal standard) | | | | | | | | |

Comments: Refer to Initial Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 1248401
 SDG #: 04-4076

VALIDATION FINDINGS WORKSHEET
Continuing Calibration Results Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent difference (%D) of the initial calibration average Relative Response Factors (RRFs) and the continuing calibration RRFs were recalculated for the compounds identified below using the following calculation:

% Difference = $100 * (\text{ave. RRF} - \text{RRF}) / \text{ave. RRF}$ Where: ave. RRF = initial calibration average RRF
 RRF = $(A_x)(C_s) / (A_s)(C_x)$ RRF = continuing calibration RRF
 A_x = Area of compound, A_s = Area of associated internal standard
 C_x = Concentration of compound, C_s = Concentration of internal standard

| # | Standard ID | Calibration Date | Compound (Reference Internal Standard) | Average RRF (Initial) | Reported | | Recalculated | |
|---|-------------|------------------|--|-----------------------|----------|------|--------------|------|
| | | | | | RRF (CC) | %D | RRF (CC) | %D |
| 1 | 04G3223 | 8/18/04 | <i>500 mg methanol</i> Methylene chloride (1st Internal Standard) | 0.104 | 0.120 | 15.4 | 0.120 | 15.4 |
| | | | Trichloroethene (2nd internal standard) | 1.516 | 1.440 | 5.0 | 1.440 | 5.0 |
| | | | <i>3rd internal standard</i> Bromoform (3rd internal standard) | 3.345 | 3.306 | 1.2 | 3.306 | 1.2 |
| 2 | | | Methylene chloride (1st Internal Standard) | | | | | |
| | | | Trichloroethene (2nd internal standard) | | | | | |
| | | | Bromoform (3rd internal standard) | | | | | |
| 3 | | | Methylene chloride (1st Internal Standard) | | | | | |
| | | | Trichloroethene (2nd internal standard) | | | | | |
| | | | Bromoform (3rd internal standard) | | | | | |
| 4 | | | Methylene chloride (1st Internal Standard) | | | | | |
| | | | Trichloroethene (2nd internal standard) | | | | | |
| | | | Bromoform (3rd internal standard) | | | | | |

Comments: Refer to Continuing Calibration findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12484 C1
 SDG #: 04-4096

VALIDATION FINDINGS WORKSHEET
Surrogate Results Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) of surrogates were recalculated for the compounds identified below using the following calculation:

% Recovery: $SF/SS * 100$

Where: SF = Surrogate Found
 SS = Surrogate Spiked

Sample ID: 2

| | Surrogate Spiked | Surrogate Found | Percent Recovery | Percent Recovery | Percent Difference |
|-----------------------------|------------------|-----------------|------------------|------------------|--------------------|
| | | | Reported | Recalculated | |
| Toluene-d8 | 20 | 19.21 | 96 | 96 | 0 |
| Bromofluorobenzene | ↓ | 20.49 | 103 | 103 | ↓ |
| 1,2-Dichlorobenzene-d4 | ↓ | 21.61 | 108 | 108 | ↓ |
| <i>Dibromofluoromethane</i> | 20 | 22.67 | 114 | 114 | ↓ |

Sample ID: _____

| | Surrogate Spiked | Surrogate Found | Percent Recovery | Percent Recovery | Percent Difference |
|------------------------|------------------|-----------------|------------------|------------------|--------------------|
| | | | Reported | Recalculated | |
| Toluene-d8 | | | | | |
| Bromofluorobenzene | | | | | |
| 1,2-Dichlorobenzene-d4 | | | | | |

Sample ID: _____

| | Surrogate Spiked | Surrogate Found | Percent Recovery | Percent Recovery | Percent Difference |
|------------------------|------------------|-----------------|------------------|------------------|--------------------|
| | | | Reported | Recalculated | |
| Toluene-d8 | | | | | |
| Bromofluorobenzene | | | | | |
| 1,2-Dichlorobenzene-d4 | | | | | |

Sample ID: _____

| | Surrogate Spiked | Surrogate Found | Percent Recovery | Percent Recovery | Percent Difference |
|------------------------|------------------|-----------------|------------------|------------------|--------------------|
| | | | Reported | Recalculated | |
| Toluene-d8 | | | | | |
| Bromofluorobenzene | | | | | |
| 1,2-Dichlorobenzene-d4 | | | | | |

Sample ID: _____

| | Surrogate Spiked | Surrogate Found | Percent Recovery | Percent Recovery | Percent Difference |
|------------------------|------------------|-----------------|------------------|------------------|--------------------|
| | | | Reported | Recalculated | |
| Toluene-d8 | | | | | |
| Bromofluorobenzene | | | | | |
| 1,2-Dichlorobenzene-d4 | | | | | |

LDC #: 12484C1
 SDG #: 04-4096

VALIDATION FINDINGS WORKSHEET
Matrix Spike/Matrix Spike Duplicates Results Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the matrix spike and matrix spike duplicate were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * (SSC - SC) / SA$ Where: SSC = Spiked sample concentration SC = Sample concentration
 SA = Spike added

RPD = $100 * MSC - MSDC / (MSC + MSDC)$ MSC = Matrix spike percent recovery MSDC = Matrix spike duplicate percent recovery

MS/MSD sample: MW-22-3

| Compound | Spike Added (ug/L) | | Sample Concentration (ug/L) | Spiked Sample Concentration (ug/L) | | Matrix Spike Percent Recovery | | Matrix Spike Duplicate Percent Recovery | | MS/MSD RPD | |
|-----------------|--------------------|-----|-----------------------------|------------------------------------|------|-------------------------------|---------|---|---------|------------|--------------|
| | MS | MSD | | MS | MSD | Reported | Recalc. | Reported | Recalc. | Reported | Recalculated |
| | 1,1-Dichloroethene | 20 | | 20 | 0 | 17.2 | 18.9 | 86 | 86 | 95 | 95 |
| Trichloroethene | | | | 18.7 | 20.5 | 94 | 94 | 103 | 103 | 9 | 9 |
| Benzene | | | | 18.5 | 20.4 | 93 | 93 | 102 | 102 | 9 | 10 |
| Toluene | | | | 19.4 | 21.3 | 97 | 97 | 107 | 107 | 10 | 9 |
| Chlorobenzene | | | | 19.7 | 21.4 | 99 | 99 | 107 | 107 | 8 | 8 |

Comments: Refer to Matrix Spike/Matrix Spike Duplicates findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12484 C1
 SDG #: 04-4096

VALIDATION FINDINGS WORKSHEET
Laboratory Control Sample Results Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

The percent recoveries (%R) and Relative Percent Difference (RPD) of the laboratory control sample and laboratory control sample duplicate (if applicable) were recalculated for the compounds identified below using the following calculation:

% Recovery = $100 * SSC/SA$ Where: SSC = Spiked sample concentration
 SA = Spike added

RPD = $100 * (LCS - LCS2) / (LCS + LCS2)$

LCS = Laboratory control sample percent recovery

LCS2 = Laboratory control sample duplicate percent recovery

LCS ID: LCS

| Compound | Spike Added (ug/L) | | Spiked Sample Concentration (ug/L) | | LCS | | LCS2 | | Percent Recovery | | RPD | |
|--------------------|--------------------|------|------------------------------------|------|----------|---------|----------|---------|------------------|---------|----------|---------|
| | LCS | LCS2 | LCS | LCS2 | Reported | Recalc. | Reported | Recalc. | Reported | Recalc. | Reported | Recalc. |
| 1,1-Dichloroethene | 20 | NA | 19.0 | NA | 95 | 95 | | | | | | |
| Trichloroethene | | | 20.2 | | 101 | 101 | | | | | | |
| Benzene | | | 19.8 | | 99 | 99 | | | | | | |
| Toluene | | | 20.6 | | 103 | 103 | | | | | | |
| Chlorobenzene | | | 20.7 | | 104 | 104 | | | NA | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |
| | | | | | | | | | | | | |

Comments: Refer to Laboratory Control Sample findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12484c1
 SDG #: 04-4096

VALIDATION FINDINGS WORKSHEET Sample Calculation Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd reviewer: [Signature]

METHOD: GC/MS VOA (EPA Method 524.2)

- Y N N/A Were all reported results recalculated and verified for all level IV samples?
Y N N/A Were all recalculated results for detected target compounds agree within 10.0% of the reported results?

$$\text{Concentration} = \frac{(A_x)(I_s)(DF)}{(A_s)(RRF)(V_s)(\%S)}$$

- A_x = Area of the characteristic ion (EICP) for the compound to be measured
- A_s = Area of the characteristic ion (EICP) for the specific internal standard
- I_s = Amount of internal standard added in nanograms (ng)
- RRF = Relative response factor of the calibration standard.
- V_s = Volume or weight of sample pruged in milliliters (ml) or grams (g).
- Df = Dilution factor.
- %S = Percent solids, applicable to soils and solid matrices only.

Example:

Sample I.D. _____, _____:
all ND
 Conc. = (_____) (_____) (_____)
 (_____) (_____) (_____) (_____)
 =

| # | Sample ID | Compound | Reported Concentration () () | Calculated Concentration () () | Qualification |
|---|-----------|----------|-----------------------------------|-------------------------------------|---------------|
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

**NASA JPL
Data Validation Reports
LDC# 12484**

Chromium

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 12, 2004
LDC Report Date: September 21, 2004
Matrix: Water
Parameters: Chromium
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-4044

Sample Identification

EB-8-8/12/04
MW-22-1
MW-22-2
MW-22-3
MW-22-3MS
MW-22-3MSD
MW-22-3DUP

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

| Method Blank ID | Analyte | Maximum Concentration | Associated Samples |
|-----------------|----------|-----------------------|----------------------------|
| ICB/CCB | Chromium | 0.244 ug/L | All samples in SDG 04-4044 |

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks with the following exceptions:

| Sample | Analyte | Reported Concentration | Modified Final Concentration |
|--------------|----------|------------------------|------------------------------|
| EB-8-8/12/04 | Chromium | 0.23 ug/L | 0.23U ug/L |

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards (ICP-MS)

ICP-MS was not reviewed in this SDG.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met with the following exceptions:

| Diluted Sample | Analyte | %D (Limits) | Associated Samples | Flag | A or P |
|----------------|----------|--------------------|----------------------------|-----------------|--------|
| MW-22-3L | Chromium | 23.5 (≤ 10) | All samples in SDG 04-4044 | J (all detects) | A |

XI. Sample Result Verification

Raw data were not reviewed for this SDG.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

No field duplicates were identified in this SDG.

XIV. Field Blanks

Sample EB-8-8/12/04 was identified as an equipment blank. No chromium contaminants were found in this blank with the following exceptions:

| Equipment Blank ID | Analyte | Concentration (ug/L) |
|--------------------|----------|----------------------|
| EB-8-8/12/04 | Chromium | 0.23 |

**NASA JPL
Chromium - Data Qualification Summary - SDG 04-4044**

| SDG | Sample | Analyte | Flag | A or P | Reason |
|---------|---|----------|-----------------|--------|--------------------------|
| 04-4044 | EB-8-8/12/04 MW-22-1 MW-22-2 MW-22-3 | Chromium | J (all detects) | A | ICP serial dilution (%D) |

**NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 04-4044**

| SDG | Sample | Analyte | Modified Final Concentration | A or P |
|---------|--------------|----------|------------------------------|--------|
| 04-4044 | EB-8-8/12/04 | Chromium | 0.23U ug/L | A |

Applied P & CH Laboratories
Metal Analysis Results

Client Name: GEOFON, Inc.
Project ID: JPL GW-3Q04
Sample ID: EB-8-8/12/04
Sample Type: Field Sample

Project No: 4-12812
Service ID: 44044
Lab Sample ID: 04-4044-1
Sample Matrix: Water

Collection Date: 08/12/2004
Collected by: JJ/MM
Received Date: 08/12/2004
Moisture %: -

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method |
|--------------|--------|------|-----|----------------|---|----|---|----------|----------|----------|----|--------|
| CHROMIUM | | μg/L | 0.1 | 0.23 <i>UJ</i> | | MS | | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

*er
9/21/04*

Applied P & CH Laboratories
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL GW-3Q04
 Sample ID: **MW-22-1**
 Sample Type: Field Sample

Project No: 4-12812
 Service ID: 44044
 Lab Sample ID: 04-4044-2
 Sample Matrix: Water

Collection Date: 08/12/2004
 Collected by: JJ/MM
 Received Date: 08/12/2004
 Moisture %: -

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method |
|--------------|--------|------|------|--------|---|----|---|----------|----------|----------|------|--------|
| CHROMIUM | | µg/L | 0.13 | 7.3 | J | MS | | 04M1875Q | 08/25/04 | 08/26/04 | 1.25 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

JJ
08/26/04

Applied P & CH Laboratories
Metal Analysis Results

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/12/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44044 | Collected by: JJ/MM |
| Sample ID: MW-22-2 | Lab Sample ID: 04-4044-3 | Received Date: 08/12/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method |
|--------------|--------|-----------------|-----|--------|---|----|---|----------|----------|----------|----|--------|
| CHROMIUM | | $\mu\text{g/L}$ | 0.1 | 9.8 | J | MS | | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor

C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.

Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control

 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control

M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

1/9/2004

Applied P & CH Laboratories
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL GW-3004
 Sample ID: MW-22-3
 Sample Type: Field Sample

Project No: 4-12812
 Service ID: 44044
 Lab Sample ID: 04-4044-4
 Sample Matrix: Water

Collection Date: 08/12/2004
 Collected by: JJ/MM
 Received Date: 08/12/2004
 Moisture %: -

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method |
|--------------|--------|------|-----|--------|---|----|---|----------|----------|----------|----|--------|
| CHROMIUM | | µg/L | 0.1 | 10.0 J | | MS | | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

Handwritten signature/initials

LDC #: 12484A4

VALIDATION COMPLETENESS WORKSHEET

SDG #: 04-4044

Level III

Laboratory: Applied Physics & Chemistry Laboratory

Date: 9/21/04

Page: 1 of 1

Reviewer: *mt*2nd Reviewer: *A***METHOD:** Chromium (EPA Method 200.8)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

| | Validation Area | | Comments |
|-------|--|----|-------------------------|
| I. | Technical holding times | A | Sampling dates: 8/12/04 |
| II. | Calibration | A | |
| III. | Blanks | SW | |
| IV. | ICP Interference Check Sample (ICS) Analysis | N | N.T. required |
| V. | Matrix Spike Analysis | D | |
| VI. | Duplicate Sample Analysis | A | |
| VII. | Laboratory Control Samples (LCS) | A | LCS/LCSD |
| VIII. | Internal Standard (ICP-MS) | N | Not reviewed |
| IX. | Furnace Atomic Absorption QC | N | Not analyzed |
| X. | ICP Serial Dilution | SW | |
| XI. | Sample Result Verification | N | |
| XII. | Overall Assessment of Data | A | |
| XIII. | Field Duplicates | N | |
| XIV. | Field Blanks | SW | EB=1 |

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

AA

| | | | | | | | |
|----|--------------|----|--|----|--|----|--|
| 1 | EB-8-8/12/04 | 11 | | 21 | | 31 | |
| 2 | MW-22-1 | 12 | | 22 | | 32 | |
| 3 | MW-22-2 | 13 | | 23 | | 33 | |
| 4 | MW-22-3 | 14 | | 24 | | 34 | |
| 5 | MW-22-3MS | 15 | | 25 | | 35 | |
| 6 | MW-22-3MSD | 16 | | 26 | | 36 | |
| 7 | MW-22-3DUP | 17 | | 27 | | 37 | |
| 8 | PB | 18 | | 28 | | 38 | |
| 9 | | 19 | | 29 | | 39 | |
| 10 | | 20 | | 30 | | 40 | |

Notes: _____

VALIDATION FINDINGS WORKSHEET
PB/CB/CCB QUALIFIED SAMPLES

LDC #: 12484 A4
 SDG #: 04-4044
 METHOD: Trace Metals (EPA SW 846 Method 6010/7000) Soil preparation factor applied: ug/L
 Sample Concentration units, unless otherwise noted: ug/L Associated Samples: A1

| Sample Identification | | | | | | | Validation Findings | | | | | | | | | | | | | | | | | | | | | | |
|-----------------------|---------------------|--------------------|-------------------------|--------------------|------|--|---------------------|--|--|--|--|--|--|--|--|--|--|--|--|--|--|--|--|--|--|--|--|--|--|
| Analyte | Maximum PB* (mg/kg) | Maximum PB* (ug/L) | Maximum ICB/CCB* (ug/L) | Blank Action Limit | 1 | | | | | | | | | | | | | | | | | | | | | | | | |
| Al | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Sb | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| As | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Ba | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Be | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Cd | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Ca | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Cr | | | 0.244 | 1.22 | 0.23 | | | | | | | | | | | | | | | | | | | | | | | | |
| Cu | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Fe | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Pb | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Mg | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Mn | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Hg | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Ni | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| K | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Se | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Ag | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Na | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Tl | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| V | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Zn | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| B | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Mo | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |
| Sr | | | | | | | | | | | | | | | | | | | | | | | | | | | | | |

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the certifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "N".
 Note: a - The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

LDC #: 12484A4
SDG #: 04-4047

Page: 1 of 1
Reviewer: MH
2nd Reviewer: [Signature]

VALIDATION FINDINGS WORKSHEET

ICP Serial Dilution

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- N N/A If analyte concentrations were > 50X the IDL, was an ICP serial dilution analyzed?
 - Y N N/A Were ICP serial dilution percent differences (%D) ≤ 10%?
 - Y N N/A Is there evidence of negative interference? If yes, professional judgement will be used to qualify the data.
- LEVEL IV ONLY:
 Y N N/A Were recalculated results acceptable? See Level IV Recalculation Worksheet for recalculations.

| # | Diluted Sample ID | Matrix | Analyte | %D | Associated Samples | Qualifications |
|---|-------------------|--------|---------|------|--------------------|----------------|
| 1 | 4 | AD | Cr | 23.5 | A11 | JLT/A |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |

Comments: _____

LDC #: 12484A4
SDG #: 04-4044

VALIDATION FINDINGS WORKSHEET Field Blanks

Page: 1 of 1
Reviewer: [Signature]
2nd reviewer: [Signature]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

N N/A Were field blanks identified in this SDG?
 N N/A Were target analytes detected in the field blanks?

Sample: 1 Field Blank / Trip Blank / Rinsate / Other EB (circle one)

| Analyte | Concentration Units ($\mu\text{g/L}$) |
|---------|---|
| CY | 0.23 |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |

Sample: _____ Field Blank / Trip Blank / Rinsate / Other _____ (circle one)

| Analyte | Concentration Units () |
|---------|-------------------------|
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |
| | |

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 16, 2004
LDC Report Date: September 21, 2004
Matrix: Water
Parameters: Chromium
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-4085

Sample Identification

DUPE-5-3Q04
MW-5
MW-6**
MW-15
MW-16
MW-15MS
MW-15MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

| Method Blank ID | Analyte | Maximum Concentration | Associated Samples |
|-----------------|----------|-----------------------|----------------------------|
| ICB/CCB | Chromium | 0.244 ug/L | All samples in SDG 04-4085 |

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks.

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met with the following exceptions:

| Diluted Sample | Analyte | %D (Limits) | Associated Samples | Flag | A or P |
|----------------|----------|--------------------|----------------------------|-----------------|--------|
| MW-22-3L | Chromium | 23.5 (≤ 10) | All samples in SDG 04-4085 | J (all detects) | A |

XI. Sample Result Verification

All sample result verifications met validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

Samples DUPE-5-3Q04 and MW-5 were identified as field duplicates. No chromium was detected in any of the samples with the following exceptions:

| Analyte | Concentration (ug/L) | | RPD |
|----------|----------------------|------|-----|
| | DUPE-5-3Q04 | MW-5 | |
| Chromium | 11.6 | 10.9 | 6 |

XIV. Field Blanks

No field blanks were identified in this SDG.

NASA JPL
Chromium - Data Qualification Summary - SDG 04-4085

| SDG | Sample | Analyte | Flag | A or P | Reason |
|---------|---|----------|-----------------|--------|--------------------------|
| 04-4085 | DUPE-5-3Q04 MW-5 MW-6** MW-15 MW-16 | Chromium | J (all detects) | A | ICP serial dilution (%D) |

NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 04-4085

No Sample Data Qualified in this SDG

Applied P & CH Laboratories
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL GW-3Q04

Project No: 4-12812
 Service ID: 44085
 Lab Sample ID: 04-4085-1
 Sample Matrix: Water

Collection Date: 08/16/2004
 Collected by: JJ/MM
 Received Date: 08/16/2004
 Moisture %: -

Sample ID: **DUPE-5-3Q04**
 Sample Type: Field Sample

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method |
|--------------|--------|------|-----|--------|---|----|---|----------|----------|----------|----|--------|
| CHROMIUM | | µg/L | 0.1 | 11.6 | J | MS | | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

JJ/MM
 8/16/04

Applied P & CH Laboratories
Metal Analysis Results

Client Name: GEOFON, Inc.

Project No: 4-12812

Collection Date: 08/16/2004

Project ID: JPL GW-3Q04

Service ID: 44085

Collected by: JJ/MM

Lab Sample ID: 04-4085-2

Received Date: 08/16/2004

Sample ID: MW-5

Sample Matrix: Water

Moisture %: -

Sample Type: Field Sample

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method | |
|--------------|--------|------|-----|--------|---|---|---|-------|----------|----------|----------|--------|-------|
| CHROMIUM | | µg/L | 0.1 | 10.9 | J | | | MS | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor

C Qualifier: U - Not Detected or less than IDL

B - Less than RL (PQL, EQL or CRDL), but greater than IDL.

Q Qualifier: N - Spike recovery out of control

* - Duplicate analysis out of control

W - Post digestion spike for GFAA out of control

E - Serial dilution difference out of control

M Qualifier: MS - ICPMS

P - ICP

A - FLAA

F - GFAA CV - Cold Vapor

g/8/04

Applied P & CH Laboratories
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL GW-3Q04
 Sample ID: MW-6
 Sample Type: Field Sample

Project No: 4-12812
 Service ID: 44085
 Lab Sample ID: 04-4085-3
 Sample Matrix: Water

Collection Date: 08/16/2004
 Collected by: JJ/MM
 Received Date: 08/16/2004
 Moisture %: -

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method |
|--------------|--------|------|-----|--------|---|----|---|----------|----------|----------|----|--------|
| CHROMIUM | | µg/L | 0.1 | 28.4 | J | MS | | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

Handwritten signature

Applied P & CH Laboratories
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL GW-3Q04
 Sample ID: MW-15
 Sample Type: Field Sample

Project No: 4-12812
 Service ID: 44085
 Lab Sample ID: 04-4085-4
 Sample Matrix: Water

Collection Date: 08/16/2004
 Collected by: JJ/MM
 Received Date: 08/16/2004
 Moisture %: -

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method |
|--------------|--------|------|-----|--------|---|----|---|----------|----------|----------|----|--------|
| CHROMIUM | | µg/L | 0.1 | 12.6 | J | MS | | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

Handwritten signature/initials
 9/22/04

Applied P & CH Laboratories
Metal Analysis Results

Client Name: GEOFON, Inc.
Project ID: JPL GW-3Q04

Project No: 4-12812
Service ID: 44085
Lab Sample ID: 04-4085-5
Sample Matrix: Water

Collection Date: 08/16/2004
Collected by: JJ/MM
Received Date: 08/16/2004
Moisture %: -

Sample ID: MW-16
Sample Type: Field Sample

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method |
|--------------|--------|-----------------|-----|--------|---|----|---|----------|----------|----------|----|--------|
| CHROMIUM | | $\mu\text{g/L}$ | 0.1 | 9.1 | J | MS | | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

JJ/MM

LDC #: 12484B4 **VALIDATION COMPLETENESS WORKSHEET**

SDG #: 04-4085

Laboratory: Applied Physics & Chemistry Laboratory

Level III/IV

Date: 9/21/04

Page: 1 of 1

Reviewer: MH

2nd Reviewer: *[Signature]*

METHOD: Chromium (EPA Method 200.8)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

| | Validation Area | | Comments |
|-------|--|-----|-------------------------|
| I. | Technical holding times | A | Sampling dates: 8/16/04 |
| II. | Calibration | A | |
| III. | Blanks | SW | |
| IV. | ICP Interference Check Sample (ICS) Analysis | N | N.t. required |
| V. | Matrix Spike Analysis | A | MS/MSD |
| VI. | Duplicate Sample Analysis | A | HW-22-3 |
| VII. | Laboratory Control Samples (LCS) | A | LCS/LCSD |
| VIII. | Internal Standard (ICP-MS) | A | |
| IX. | Furnace Atomic Absorption QC | N | N.t. utilized |
| X. | ICP Serial Dilution | SW | |
| XI. | Sample Result Verification | A N | |
| XII. | Overall Assessment of Data | A | |
| XIII. | Field Duplicates | SW | (1, 2) |
| XIV. | Field Blanks | N | |

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

AS

| | | | | | | | |
|----|--------------------|----|--|----|--|----|--|
| 1 | DUPE-5-3Q04 | 11 | | 21 | | 31 | |
| 2 | MW-5 | 12 | | 22 | | 32 | |
| 3 | MW-6** | 13 | | 23 | | 33 | |
| 4 | MW-15 | 14 | | 24 | | 34 | |
| 5 | MW-16 | 15 | | 25 | | 35 | |
| 6 | MW-15MS | 16 | | 26 | | 36 | |
| 7 | MW-15MSD | 17 | | 27 | | 37 | |
| 8 | MW-15DUP <i>MH</i> | 18 | | 28 | | 38 | |
| 9 | <i>FB</i> | 19 | | 29 | | 39 | |
| 10 | | 20 | | 30 | | 40 | |

Notes: _____

DC #: 12484B4
 SDG #: 04-4085

VALIDATION FINDINGS CHECKLIST

Page: 1 of 1
 Reviewer: WIS
 2nd Reviewer: [Signature]

Method: Metals (EPA SW 826 Method 6010/7000/6020)

| Validation Area | Yes | No | NA | Findings/Comments |
|---|-----|----|----|-------------------|
| I. Technical holding times | | | | |
| All technical holding times were met. | ✓ | | | |
| Cooler temperature criteria was met. | ✓ | | | |
| II. Calibration | | | | |
| Were all instruments calibrated daily, each set-up time? | ✓ | | | |
| Were the proper number of standards used? | ✓ | | | |
| Were all initial and continuing calibration verification %Rs within the 90-110% (80-120% for mercury and 85-115% for cyanide) QC limits? | ✓ | | | |
| Were all initial calibration correlation coefficients ≥ 0.995 ? | ✓ | | | |
| Was a midrange cyanide standard distilled? | | | ✓ | |
| III. Blanks | | | | |
| Was a method blank associated with every sample in this SDG? | ✓ | | | |
| Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet. | ✓ | | | |
| IV. ICP Interference Check Sample | | | | |
| Were ICP interference check samples performed daily? | | | ✓ | |
| Were the AB solution percent recoveries (%R) with the 80-120% QC limits? | | | ✓ | |
| V. Matrix spike/Main spike duplicates | | | | |
| Were a matrix spike (MS) and duplicate (DUP) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD or MS/DUP. Soil / Water. | ✓ | | | |
| Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the 75-125 QC limits? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken. | ✓ | | | |
| Were the MS/MSD or duplicate relative percent differences (RPD) $\leq 20\%$ for waters and $\leq 35\%$ for soil samples? A control limit of $\pm RL$ ($\pm 2X RL$ for soil) was used for samples that were $\leq 5X$ the RL, including when only one of the duplicate sample values were $\leq 5X$ the RL. | ✓ | | | |
| VI. Laboratory control samples | | | | |
| Was an LCS analyzed for this SDG? | ✓ | | | |
| Was an LCS analyzed per extraction batch? | ✓ | | | |
| Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the 80-120% QC limits for water samples and laboratory established QC limits for soils? | ✓ | | | |
| VII. Furnace Atomic Absorption QC | | | | |
| If MSA was performed, was the correlation coefficients ≥ 0.995 ? | | | ✓ | |
| Do all applicable analyses have duplicate injections? | | | ✓ | |

LDC #: 1248484
 SDG #: 04-4085

VALIDATION FINDINGS CHECKLIST

Page: 1 of 1
 Reviewer: WJ
 2nd Reviewer: [Signature]

| Validation Area | Yes | No | NA | Findings/Comments |
|--|-----|----|----|-------------------|
| For sample concentrations > RL, are applicable duplicate injection RSD values < 20%? | | | ✓ | |
| Were analytical spike recoveries within the 85-115% QC limits? | | | ✓ | |
| VII. ICP Serial Dilution | | | | |
| Was an ICP serial dilution analyzed if analyte concentrations were > ^{100X} 50X the IDL? | ✓ | | | |
| Were all percent differences (%Ds) ≤ 10%? | | ✓ | | |
| Was there evidence of negative interference? If yes, professional judgement will be used to qualify the data. | | ✓ | | |
| VIII. Internal Standards (EPA SW 846 Method 8020) | | | | |
| Were all the percent recoveries (%R) within the ⁶⁰⁻¹¹⁵ 30-120% of the intensity of the internal standard in the associated initial calibration? | ✓ | | | |
| If the %Rs were outside the criteria, was a reanalysis performed? | | | ✓ | |
| IX. Regional Quality Assurance and Quality Control | | | | |
| Were performance evaluation (PE) samples performed? | | | ✓ | |
| Were the performance evaluation (PE) samples within the acceptance limits? | | | ✓ | |
| X. Sample Result Verification | | | | |
| Were RLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation? | ✓ | | | |
| XI. Overall assessment of data | | | | |
| Overall assessment of data was found to be acceptable. | ✓ | | | |
| XII. Field duplicates | | | | |
| Field duplicate pairs were identified in this SDG. | ✓ | | | |
| Target analytes were detected in the field duplicates. | ✓ | | | |
| XIII. Field blanks | | | | |
| Field blanks were identified in this SDG. | | ✓ | | |
| Target analytes were detected in the field blanks. | | | ✓ | |

LDC #: 12484B4

SDG #: 04-483

METHOD: Trace Metals (EPA SW 846 Method 6010/7000) / Soil preparation factor applied: 0.1

Sample Concentration units, unless otherwise noted: ug/L

Associated Samples: 67

VALIDATION FINDINGS WORKSHEET
PB/CB/CCB QUALIFIED SAMPLES

Page: 1 of 1
Reviewer: MJK
2nd Reviewer:

| Analyte | Maximum PB* (mg/Kg) | Maximum PB* (ug/L) | Maximum CB/CCB* (ug/L) | Blank Action Limit | Sample Identification |
|---------|---------------------|--------------------|------------------------|--------------------|------------------------|
| Al | | | | | |
| Sb | | | | | |
| As | | | | | |
| Ba | | | | | |
| Be | | | | | |
| Cd | | | | | |
| Ca | | | | | |
| Cr | | | 0.244 | 1.02 | No Sample quality test |
| Cu | | | | | |
| Fe | | | | | |
| Pb | | | | | |
| Mg | | | | | |
| Mn | | | | | |
| Hg | | | | | |
| Ni | | | | | |
| K | | | | | |
| Se | | | | | |
| Ag | | | | | |
| Na | | | | | |
| Tl | | | | | |
| V | | | | | |
| Zn | | | | | |
| B | | | | | |
| Mo | | | | | |
| Sr | | | | | |

Samples with analyte concentrations within five times the associated ICB, CCB or PB concentration are listed above with the identification from the Validation Completeness Worksheet. These sample results were qualified as not detected. "U".
Note: a. The listed analyte concentration is the highest ICB, CCB, or PB detected in the analysis of each element.

LDC #: 12484B4
 SDG #: 04-4.85

VALIDATION FINDINGS WORKSHEET
ICP Serial Dilution

Page: 1 of 1
 Reviewer: MH
 2nd Reviewer: [Signature]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
 N. N/A If analyte concentrations were > 50% the IDL, was an ICP serial dilution analyzed?
 Y. N/A Were ICP serial dilution percent differences (%D) ≤ 10%?
 Y. N/A Is there evidence of negative interference? If yes, professional judgement will be used to qualify the data.

LEVEL IV ONLY:
 Y. N. N/A Were recalculated results acceptable? See Level IV Recalculation Worksheet for recalculations.

| # | Diluted Sample ID | Matrix | Analyte | %D | Associated Samples | Qualifications |
|---|-------------------|--------|---------|------|--------------------|----------------|
| 1 | MW-22-3 | AB | Cy | 23.5 | A1 | Table A |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |

Comments: _____

LDC #: 148434
 SDG #: 04-4-82

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
 Reviewer: MB
 2nd reviewer: [Signature]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

N N/A Were field duplicate pairs identified in this SDG?
 N N/A Were target analytes detected in the field duplicate pairs?

| Analyte | Concentration (<u>ug/L</u>) | | RPD (Limits) | Difference (Limits) | Qualifications |
|------------|-------------------------------|------|--------------|---------------------|----------------|
| | 1 | 2 | | | |
| Aluminum | | | | | |
| Antimony | | | | | |
| Arsenic | | | | | |
| Barium | | | | | |
| Beryllium | | | | | |
| Cadmium | | | | | |
| Calcium | | | | | |
| Chromium | 11.6 | 10.9 | 6 | | |
| Cobalt | | | | | |
| Copper | | | | | |
| Iron | | | | | |
| Lead | | | | | |
| Magnesium | | | | | |
| Manganese | | | | | |
| Mercury | | | | | |
| Nickel | | | | | |
| Potassium | | | | | |
| Selenium | | | | | |
| Silver | | | | | |
| Sodium | | | | | |
| Thallium | | | | | |
| Vanadium | | | | | |
| Zinc | | | | | |
| Cyanide | | | | | |
| Boron | | | | | |
| Molybdenum | | | | | |
| Strontium | | | | | |
| Silicon | | | | | |

Notes: _____

LDC #: 1248484
 SDG #: 04-4085

VALIDATION FINDINGS WORKSHEET
Initial and Continuing Calibration Calculation Verification

Page: 1 of 1
 Reviewer: HJ
 2nd Reviewer: [Signature]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

An initial and continuing calibration verification percent recovery (%R) was recalculated for each type of analysis using the following formula:

$\%R = \frac{\text{Found}}{\text{True}} \times 100$ Where Found = concentration (in ug/L) of each analyte measured in the analysis of the ICV or CCV solution
 True = concentration (in ug/L) of each analyte in the ICV or CCV source

| Standard ID | Type of Analysis | Element | Found (ug/L) | True (ug/L) | Recalculated | | Reported | | Acceptable (Y/N) |
|-------------|----------------------------------|---------|--------------|-------------|--------------|--|----------|--|------------------|
| | | | | | %R | | %R | | |
| ICV | ICP (Initial calibration) | Cr | 50.7 | 50.0 | 101.4 | | 101.4 | | Y |
| | GFAA (Initial calibration) | | | | | | | | |
| | CVAA (Initial calibration) | | | | | | | | |
| CCV | ICP (Continuing calibration) | Cr | 51.94 | 50.0 | 103.9 | | 103.8 | | Y |
| | GFAA (Continuing calibration) | | | | | | | | |
| | CVAA (Continuing calibration) | | | | | | | | |
| | Cyanide (Initial calibration) | | | | | | | | |
| | Cyanide (Continuing calibration) | | | | | | | | |

Comments: Refer to Calibration Verification findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 1248484
 SDG #: 04-4085

VALIDATION FINDINGS WORKSHEET
Level IV Recalculation Worksheet

Page: 1 of 1
 Reviewer: juw
 2nd Reviewer: [Signature]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Percent recoveries (%R) for an ICP interference check sample, a laboratory control sample and a matrix spike sample were recalculated using the following formula:

$$\%R = \frac{\text{Found} \times 100}{\text{True}}$$
 Where, Found = Concentration of each analyte measured in the analysis of the sample. For the matrix spike calculation,
 Found = SSR (spiked sample result) - SR (sample result).
 True = Concentration of each analyte in the source.

A sample and duplicate relative percent difference (RPD) was recalculated using the following formula:

$$RPD = \frac{|S-D|}{(S+D)/2} \times 100$$
 Where, S = Original sample concentration
 D = Duplicate sample concentration

An ICP serial dilution percent difference (%D) was recalculated using the following formula:

$$\%D = \frac{|I-SDR|}{I} \times 100$$
 Where, I = Initial Sample Result (mg/L)
 SDR = Serial Dilution Result (mg/L) (Instrument Reading x 5)

| Sample ID | Type of Analysis | Element | Found / S / I (units) | True / D / SDR (units) | Recalculated | | Reported | | Acceptable (Y/N) |
|---------------|---------------------------|---------|-----------------------|------------------------|---------------|---------------|----------|------|------------------|
| | | | | | %R / RPD / %D | %R / RPD / %D | | | |
| NA | ICP interference check | | | | | | | | |
| LCS | Laboratory control sample | Cr | 50.3 | 50 | 101 | 101 | 94 | 94 | Y |
| b | Matrix spike | | (SSR-SR) 46.9 | 50 | 94 | 94 | 94 | 94 | Y |
| MN-22-3 | Duplicate | | 9.84 | 10.01 | 1.7 | 1.8 | 1.8 | 1.8 | Y |
| MN-22-3 | ICP serial dilution | | 7.69 | 10.01 | 23.2 | 23.5 | 23.5 | 23.5 | Y |

Comments: Refer to appropriate worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12484 B4
 SDG #: 04-4085

VALIDATION FINDINGS WORKSHEET Sample Calculation Verification

Page: 1 of 1
 Reviewer: WJ
 2nd reviewer: _____

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- Y N N/A Have results been reported and calculated correctly?
- Y N N/A Are results within the calibrated range of the instruments and within the linear range of the ICP?
- Y N N/A Are all detection limits below the CRDL?

Detected analyte results for _____ were recalculated and verified using the following equation:

$$\text{Concentration} = \frac{(RD)(FV)(DII)}{(\text{In. Vol.})(\%S)}$$

Recalculation:

From the row data
Cr = 28.4482 ug/L

- RD = Raw data concentration
- FV = Final volume (ml)
- In. Vol. = Initial volume (ml) or weight (G)
- DII = Dilution factor
- %S = Decimal percent solids

| Sample ID | Analyte | Reported Concentration (ug/L) | Calculated Concentration (ug/L) | Acceptable (Y/N) |
|-----------|---------|----------------------------------|------------------------------------|---------------------|
| 3 | Cr | 28.4 | 28.4 | Y |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 17, 2004
LDC Report Date: September 21, 2004
Matrix: Water
Parameters: Chromium
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-4096

Sample Identification

DUPE-3-3Q04
MW-8**
MW-10
MW-13

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 4 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 200.8 for Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the method stated above.

A table summarizing all data qualification flags is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from specified protocols or is of technical advisory nature.

Blanks are summarized in Section III.

Field duplicates are summarized in Section XIII.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

An initial calibration was performed.

The frequency and analysis criteria of the initial calibration verification (ICV) and continuing calibration verification (CCV) were met.

III. Blanks

Method blanks were reviewed for each matrix as applicable.

Data qualification by the initial, continuing and preparation blanks (ICB/CCB/PBs) was based on the maximum contaminant concentration in the ICB/CCB/PBs in the analysis of each analyte. No contaminant concentrations were found in the initial, continuing and preparation blanks with the following exceptions:

| Method Blank ID | Analyte | Maximum Concentration | Associated Samples |
|-----------------|----------|-----------------------|----------------------------|
| ICB/CCB | Chromium | 0.244 ug/L | All samples in SDG 04-4096 |

Sample concentrations were compared to the maximum contaminant concentrations detected in the ICB/CCB/PBs. The sample concentrations were either not detected or were significantly greater (>5X blank contaminants) than the concentrations found in the associated method blanks.

IV. ICP Interference Check Sample (ICS) Analysis

ICP interference check sample analysis was not required.

V. Matrix Spike Analysis

Matrix spike (MS) and matrix spike duplicate (MSD) samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VI. Duplicate Sample Analysis

Duplicate (DUP) sample analyses were reviewed for each matrix as applicable. Results were within QC limits.

VII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VIII. Internal Standards

All internal standard percent recoveries (%R) were within QC limits for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

IX. Furnace Atomic Absorption QC

Graphite furnace atomic absorption was not utilized in this SDG.

X. ICP Serial Dilution

Although ICP serial dilution analysis was not required by the method, it was performed by the laboratory. The analysis criteria were met with the following exceptions:

| Diluted Sample | Analyte | %D (Limits) | Associated Samples | Flag | A or P |
|----------------|----------|--------------------|----------------------------|-----------------|--------|
| MW-22-3L | Chromium | 23.5 (≤ 10) | All samples in SDG 04-4096 | J (all detects) | A |

XI. Sample Result Verification

All sample result verifications met validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

XII. Overall Assessment of Data

Data flags have been summarized at the end of this report.

XIII. Field Duplicates

Samples DUPE-3-3Q04 and MW-10 were identified as field duplicates. No chromium was detected in any of the samples with the following exceptions:

| Analyte | Concentration (ug/L) | | RPD |
|----------|----------------------|-------|-----|
| | DUPE-3-3Q04 | MW-10 | |
| Chromium | 23.8 | 24.2 | 2 |

XIV. Field Blanks

No field blanks were identified in this SDG.

NASA JPL
Chromium - Data Qualification Summary - SDG 04-4096

| SDG | Sample | Analyte | Flag | A or P | Reason |
|------------|---|----------------|-----------------|---------------|--------------------------|
| 04-4096 | DUPE-3-3Q04 MW-8** MW-10 MW-13 | Chromium | J (all detects) | A | ICP serial dilution (%D) |

NASA JPL
Chromium - Laboratory Blank Data Qualification Summary - SDG 04-4096

No Sample Data Qualified in this SDG

Applied P & CH Laboratories
Metal Analysis Results

Client Name: GEOFON, Inc.
Project ID: JPL GW-3Q04

Project No: 4-12812
Service ID: 44096
Lab Sample ID: 04-4096-1
Sample Matrix: Water

Collection Date: 08/17/2004
Collected by: JJ/MM/TM
Received Date: 08/17/2004
Moisture %: -

Sample ID: DUPE-6-3Q04
Sample Type: Field Sample

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method | |
|--------------|--------|------|-----|--------|---|---|---|-------|----------|----------|----------|--------|-------|
| CHROMIUM | | µg/L | 0.1 | 23.8 | J | | | MS | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

Applied P & CH Laboratories
Metal Analysis Results

| | | |
|---------------------------|--------------------------|-----------------------------|
| Client Name: GEOFON, Inc. | Project No: 4-12812 | Collection Date: 08/17/2004 |
| Project ID: JPL GW-3Q04 | Service ID: 44096 | Collected by: JJ/MM/TM |
| Sample ID: MW-8 | Lab Sample ID: 04-4096-2 | Received Date: 08/17/2004 |
| Sample Type: Field Sample | Sample Matrix: Water | Moisture %: - |

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method |
|--------------|--------|------|-----|--------|---|----|---|----------|----------|----------|----|--------|
| .CHROMIUM | | µg/L | 0.1 | 9.8 J | | MS | | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor

| | |
|--|--|
| C Qualifier: U - Not Detected or less than IDL | B - Less than RL (PQL, EQL or CRDL), but greater than IDL. |
| Q Qualifier: N - Spike recovery out of control | * - Duplicate analysis out of control |
| W - Post digestion spike for GFAA out of control | E - Serial dilution difference out of control |
| M Qualifier: MS - ICPMS P - ICP A - FLAA | F - GFAA CV - Cold Vapor |

K. Johnson

Applied P & CH Laboratories
Metal Analysis Results

Client Name: GEOFON, Inc.
 Project ID: JPL GW-3Q04
 Sample ID: MW-10
 Sample Type: Field Sample

Project No: 4-12812
 Service ID: 44096
 Lab Sample ID: 04-4096-3
 Sample Matrix: Water

Collection Date: 08/17/2004
 Collected by: JJ/MM/TM
 Received Date: 08/17/2004
 Moisture %: -

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method | |
|--------------|--------|------|-----|--------|---|---|---|-------|----------|----------|----------|--------|-------|
| .CHROMIUM | | µg/L | 0.1 | 24.2 | J | | | MS | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
 C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
 Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
 M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

Handwritten signature
 9/22/04

Applied P & CH Laboratories
Metal Analysis Results

Client Name: GEOFON, Inc.
Project ID: JPL GW-3Q04

Project No: 4-12812
Service ID: 44096
Lab Sample ID: 04-4096-4
Sample Matrix: Water

Collection Date: 08/17/2004
Collected by: JJ/MM/TM
Received Date: 08/17/2004
Moisture %: -

Sample ID: MW-13
Sample Type: Field Sample

| Element Name | CAS No | Unit | RL | Result | C | M | Q | Batch | D-Date | A-Date | DF | Method |
|--------------|--------|------|-----|--------|---|----|---|----------|----------|----------|----|--------|
| CHROMIUM | | μg/L | 0.1 | 26.1 | J | MS | | 04M1875Q | 08/25/04 | 08/26/04 | 1 | 200.8 |

Note: RL: PQL (EQL) or CRDL D-Date: Digestion Date; A-Date: Analysis Date; DF: Dilution Factor
C Qualifier: U - Not Detected or less than IDL B - Less than RL (PQL, EQL or CRDL), but greater than IDL.
Q Qualifier: N - Spike recovery out of control * - Duplicate analysis out of control
 W - Post digestion spike for GFAA out of control E - Serial dilution difference out of control
M Qualifier: MS - ICPMS P - ICP A - FLAA F - GFAA CV - Cold Vapor

Handwritten signature/initials
9/2/04

LDC #: 12484C4

VALIDATION COMPLETENESS WORKSHEET

Date: 9/21/04

SDG #: 04-4096

Level III / EV

Page: 1 of 1

Laboratory: Applied Physics & Chemistry Laboratory

Reviewer: [Signature]

2nd Reviewer: [Signature]

METHOD: Chromium (EPA Method 200.8)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

| | Validation Area | | Comments |
|-------|--|-----|-------------------------|
| I. | Technical holding times | A | Sampling dates: 8/17/04 |
| II. | Calibration | A | |
| III. | Blanks | SW | |
| IV. | ICP Interference Check Sample (ICS) Analysis | N | M.T. requires |
| V. | Matrix Spike Analysis | A | |
| VI. | Duplicate Sample Analysis | A | |
| VII. | Laboratory Control Samples (LCS) | A | LCS/LCSB |
| VIII. | Internal Standard (ICP-MS) | A | |
| IX. | Furnace Atomic Absorption QC | N | M.T. utilized |
| X. | ICP Serial Dilution | SW | |
| XI. | Sample Result Verification | A ✓ | |
| XII. | Overall Assessment of Data | A | |
| XIII. | Field Duplicates | SW | (1,3) |
| XIV. | Field Blanks | N | |

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples:

| | | | | | | | |
|-----|-------------|----|--|----|--|----|--|
| 1 ✓ | DUPE-3-3Q04 | 11 | | 21 | | 31 | |
| 2 | MW-8** | 12 | | 22 | | 32 | |
| 3 ✓ | MW-10 | 13 | | 23 | | 33 | |
| 4 | MW-13 | 14 | | 24 | | 34 | |
| 5 | PB | 15 | | 25 | | 35 | |
| 6 | | 16 | | 26 | | 36 | |
| 7 | | 17 | | 27 | | 37 | |
| 8 | | 18 | | 28 | | 38 | |
| 9 | | 19 | | 29 | | 39 | |
| 10 | | 20 | | 30 | | 40 | |

Notes: _____

LDC #: 12484 c4
 SDG #: 04-4096

VALIDATION FINDINGS CHECKLIST

Page: 1 of 1
 Reviewer: WJ
 2nd Reviewer: [Signature]

4
Method: Metals (EPA SW 826 Method 6010/7000/6020)

| Validation Area | Yes | No | NA | Findings/Comments |
|--|-----|----|----|-------------------|
| I. Technical holding times | | | | |
| All technical holding times were met. | ✓ | | | |
| Cooler temperature criteria was met. | ✓ | | | |
| II. Calibration | | | | |
| Were all instruments calibrated daily, each set-up time? | ✓ | | | |
| Were the proper number of standards used? | ✓ | | | |
| Were all initial and continuing calibration verification %Rs within the 90-110% (80-120% for mercury and 85-115% for cyanide) QC limits? | ✓ | | | |
| Were all initial calibration correlation coefficients ≥ 0.995 ? | ✓ | | | |
| Was a midrange cyanide standard distilled? | | | ✓ | |
| III. Blanks | | | | |
| Was a method blank associated with every sample in this SDG? | ✓ | | | |
| Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet. | ✓ | | | |
| IV. ICP Interference Check Sample | | | | |
| Were ICP interference check samples performed daily? | | | ✓ | |
| Were the AB solution percent recoveries (%R) with the 80-120% QC limits? | | | ✓ | |
| IV. Matrix spike/Matrix spike duplicates | | | | |
| Were a matrix spike (MS) and duplicate (DUP) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD or MS/DUP. Soil / Water. | ✓ | | | |
| Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the 75-125 QC limits? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken. | ✓ | | | |
| Were the MS/MSD or duplicate relative percent differences (RPD) $\leq 20\%$ for waters and $\leq 35\%$ for soil samples? A control limit of $\pm 1 \text{ RL}$ ($\pm 2 \text{ X RL}$ for soil) was used for samples that were $\leq 5 \text{ X}$ the RL, including when only one of the duplicate sample values were $\leq 5 \text{ X}$ the RL. | ✓ | | | |
| V. Laboratory control samples | | | | |
| Was an LCS analyzed for this SDG? | ✓ | | | |
| Was an LCS analyzed per extraction batch? | ✓ | | | |
| Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the 80-120% QC limits for water samples and laboratory established QC limits for soils? | ✓ | | | |
| VI. Furnace Atomic Absorption QC | | | | |
| If MSA was performed, was the correlation coefficients ≥ 0.995 ? | | | ✓ | |
| Do all applicable analyses have duplicate injections? | | | ✓ | |

LDC #: 1248404
 SDG #: 04-4096

VALIDATION FINDINGS CHECKLIST

Page: 1 of 1
 Reviewer: WV
 2nd Reviewer: [Signature]

| Validation Area | Yes | No | NA | Findings/Comments |
|--|-------------------------------------|-------------------------------------|-------------------------------------|-------------------|
| For sample concentrations > RL, are applicable duplicate injection RSD values < 20%? | | | <input checked="" type="checkbox"/> | |
| Were analytical spike recoveries within the 85-115% QC limits? | | | <input checked="" type="checkbox"/> | |
| VII. ICP Serial Dilution | | | | |
| Was an ICP serial dilution analyzed if analyte concentrations were > ^{100X} 50X the IDL? | <input checked="" type="checkbox"/> | | | |
| Were all percent differences (%Ds) ≤ 10%? | | <input checked="" type="checkbox"/> | | |
| Was there evidence of negative interference? If yes, professional judgement will be used to qualify the data. | | <input checked="" type="checkbox"/> | | |
| VIII. Internal Standards (EPA SW 846 Method 8020) | | | | |
| Were all the percent recoveries (%R) within the ⁶⁻¹³ 30-120% of the intensity of the internal standard in the associated initial calibration? | <input checked="" type="checkbox"/> | | | |
| If the %Rs were outside the criteria, was a reanalysis performed? | | | <input checked="" type="checkbox"/> | |
| IX. Regional Quality Assurance and Quality Control | | | | |
| Were performance evaluation (PE) samples performed? | | | <input checked="" type="checkbox"/> | |
| Were the performance evaluation (PE) samples within the acceptance limits? | | | <input checked="" type="checkbox"/> | |
| X. Sample Result Verification | | | | |
| Were RLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation? | <input checked="" type="checkbox"/> | | | |
| XI. Overall assessment of data | | | | |
| Overall assessment of data was found to be acceptable. | <input checked="" type="checkbox"/> | | | |
| XII. Field duplicates | | | | |
| Field duplicate pairs were identified in this SDG. | <input checked="" type="checkbox"/> | | | |
| Target analytes were detected in the field duplicates. | <input checked="" type="checkbox"/> | | | |
| XIII. Field blanks | | | | |
| Field blanks were identified in this SDG. | | | <input checked="" type="checkbox"/> | |
| Target analytes were detected in the field blanks. | | | <input checked="" type="checkbox"/> | |

LDC #: 1248464
 SDG #: 04-4086

VALIDATION FINDINGS WORKSHEET
PB/CB/CCB QUALIFIED SAMPLES

METHOD: Trace Metals (EPA SW 846 Method 8010/7000) Soil preparation factor applied:

Sample Concentration units, unless otherwise noted: ug/L Associated Samples: AT1

Page: 1 of 1
 Reviewer: MW
 2nd Reviewer: [Signature]

| Analyte | Maximum PB* (mg/kg) | Maximum PB* (ug/L) | Maximum CB/CCB* (ug/L) | Blank Action Limit | Sample Identification |
|---------|---------------------|--------------------|------------------------|--------------------|-----------------------|
| Al | | | | | |
| Sb | | | | | |
| As | | | | | |
| Ba | | | | | |
| Be | | | | | |
| Cd | | | | | |
| Ce | | | | | |
| Cr | | | 0.244 | 1.22 | U6 Sample for test |
| Co | | | | | |
| Cu | | | | | |
| Fe | | | | | |
| Pb | | | | | |
| Mg | | | | | |
| Mn | | | | | |
| Hg | | | | | |
| Ni | | | | | |
| K | | | | | |
| Se | | | | | |
| Ag | | | | | |
| Na | | | | | |
| Ti | | | | | |
| V | | | | | |
| Zn | | | | | |
| B | | | | | |
| Mo | | | | | |
| Sr | | | | | |

Samples with analyte concentrations within five times the associated CB, CCB or PB concentration are listed above with the identifications from the Validation Completeness Worksheet. These sample results were qualified as not detected, "U".
 Note: a - The listed analyte concentration is the highest CB, CCB, or PB detected in the analysis of each element.

VALIDATION FINDINGS WORKSHEET
ICP Serial Dilution

LDC #: 12484c4
 SDG #: 04-4012

Page: 1 of 1
 Reviewer: MM
 2nd Reviewer: A

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".
 Y N N/A If analyte concentrations were $> 50\times$ the IDL, was an ICP serial dilution analyzed?
 Y N N/A Were ICP serial dilution percent differences (%D) $\leq 10\%$?
 Y N N/A Is there evidence of negative interference? If yes, professional judgement will be used to qualify the data.
 LEVEL IV ONLY:
 Y N N/A Were recalculated results acceptable? See Level IV Recalculation Worksheet for recalculations.

| # | Diluted Sample ID | Matrix | Analyte | %D | Associated Samples | Qualifications |
|---|-------------------|--------|---------|------|--------------------|----------------|
| 1 | MM-22-3 | AG | Cr | 23.5 | A11 | JJA/A |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |
| | | | | | | |

Comments:

LDC #: 1248404
 SDG #: 04-4096

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
 Reviewer: MB
 2nd reviewer: [Signature]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Y N N/A Were field duplicate pairs identified in this SDG?
 Y N N/A Were target analytes detected in the field duplicate pairs?

| Analyte | Concentration (<u>ug/L</u>) | | RPD (Limits) | Difference (Limits) | Qualifications |
|------------|-------------------------------|-------------|--------------|---------------------|----------------|
| | 1 | 2 | | | |
| Aluminum | | | | | |
| Antimony | | | | | |
| Arsenic | | | | | |
| Barium | | | | | |
| Beryllium | | | | | |
| Cadmium | | | | | |
| Calcium | | | | | |
| Chromium | <u>23.8</u> | <u>24.2</u> | <u>2</u> | | |
| Cobalt | | | | | |
| Copper | | | | | |
| Iron | | | | | |
| Lead | | | | | |
| Magnesium | | | | | |
| Manganese | | | | | |
| Mercury | | | | | |
| Nickel | | | | | |
| Potassium | | | | | |
| Selenium | | | | | |
| Silver | | | | | |
| Sodium | | | | | |
| Thallium | | | | | |
| Vanadium | | | | | |
| Zinc | | | | | |
| Cyanide | | | | | |
| Boron | | | | | |
| Molybdenum | | | | | |
| Strontium | | | | | |
| Silicon | | | | | |

Notes: _____

LDC #: 12484C4
 SDG #: 04-4096

VALIDATION FINDINGS WORKSHEET
Initial and Continuing Calibration Calculation Verification

Page: 1 of 1
 Reviewer: PMG
 2nd Reviewer: [Signature]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

An initial and continuing calibration verification percent recovery (%R) was recalculated for each type of analysis using the following formula:

$$\%R = \frac{\text{Found}}{\text{True}} \times 100$$

Where, Found = concentration (in ug/L) of each analyte measured in the analysis of the ICV or CCV solution
 True = concentration (in ug/L) of each analyte in the ICV or CCV source

| Standard ID | Type of Analysis | Element | Found (ug/L) | True (ug/L) | Recalculated | | Reported | | Acceptable (Y/N) |
|-------------|---|---------|--------------|-------------|--------------|-------|----------|-------|------------------|
| | | | | | %R | %R | %R | %R | |
| ICV | ICP/ ^{MS} (Initial calibration) | Cr | 50.7 | 50.0 | 101.4 | 101.4 | 101.4 | 101.4 | Y |
| | GFAA (Initial calibration) | | | | | | | | |
| | CVAAs (Initial calibration) | | | | | | | | |
| CCV | ICP/ ^{MS} (Continuing calibration) | Cr | 51.94 | 50.0 | 103.9 | 103.9 | 103.8 | 103.8 | Y |
| | GFAA (Continuing calibration) | | | | | | | | |
| | CVAAs (Continuing calibration) | | | | | | | | |
| | Cyanide (Initial calibration) | | | | | | | | |
| | Cyanide (Continuing calibration) | | | | | | | | |

Comments: Refer to Calibration Verification findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

VALIDATION FINDINGS WORKSHEET
Level IV Recalculation Worksheet

LDC #: 12484c4
 SDG #: 04-4096

Page: 1 of 1
 Reviewer: HW
 2nd Reviewer: [Signature]

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Percent recoveries (%R) for an ICP interference check sample, a laboratory control sample and a matrix spike sample were recalculated using the following formula:

$$\%R = \frac{\text{Found} - \text{True}}{\text{True}} \times 100$$

Where, Found = Concentration of each analyte measured in the analysis of the sample. For the matrix spike calculation,
 Found = SSR (spiked sample result) - SR (sample result),
 True = Concentration of each analyte in the source.

A sample and duplicate relative percent difference (RPD) was recalculated using the following formula:

$$RPD = \frac{|S-D|}{(S+D)/2} \times 100$$

Where, S = Original sample concentration
 D = Duplicate sample concentration

An ICP serial dilution percent difference (%D) was recalculated using the following formula:

$$\%D = \frac{|I-SDR|}{I} \times 100$$

Where, I = Initial Sample Result (mg/L)
 SDR = Serial Dilution Result (mg/L) (Instrument Reading x 5)

| Sample ID | Type of Analysis | Element | Found / S / I (units) | True / D / SDR (units) | Recalculated | | Reported | | Acceptable (Y/N) |
|---------------|---------------------------|---------|-----------------------|------------------------|---------------|---------------|----------|------|------------------|
| | | | | | %R / RPD / %D | %R / RPD / %D | | | |
| NA | ICP interference check | | | | | | | | |
| LC5 | Laboratory control sample | Cu | 50.3 | 50 | 101 | 101 | 101 | 94 | Y |
| MW-22-3 | Matrix spike | | 46.7 (SSR-SR) | 50.0 | 93 | 93 | 94 | 1.8 | Y |
| | Duplicate | | 9.84 | 10.01 | 1.7 | 1.7 | 1.8 | 23-5 | Y |
| | ICP serial dilution | | 9.69 | 10.01 | 23.2 | 23.2 | 23-5 | 23-5 | Y |

Comments: Refer to appropriate worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12484c4
SDG #: 04-4096

VALIDATION FINDINGS WORKSHEET Sample Calculation Verification

Page: 1 of 1
Reviewer: MW
2nd reviewer:

METHOD: Trace Metals (EPA SW 846 Method 6010/7000)

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- N N/A Have results been reported and calculated correctly?
- N N/A Are results within the calibrated range of the instruments and within the linear range of the ICP?
- N N/A Are all detection limits below the CRDL?

Detected analyte results for 2 were recalculated and verified using the following equation:

Concentration = $\frac{(RD)(FV)(DII)}{(In. Vol.)(\%S)}$ Recalculation:

- RD = Raw data concentration
- FV = Final volume (ml)
- In. Vol. = Initial volume (ml) or weight (G)
- DII = Dilution factor
- %S = Decimal percent solids

*From the raw data
Cr = 9.969 ug/L*

| Sample ID | Analyte | Reported Concentration (ug/L) | Calculated Concentration (ug/L) | Acceptable (Y/N) |
|-----------|---------|-------------------------------|---------------------------------|------------------|
| 2 | Cr | 9.8 | 9.8 | Y |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |
| | | | | |

**NASA JPL
Data Validation Reports
LDC# 12484**

Wet Chemistry

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 12, 2004
LDC Report Date: September 21, 2004
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-4044

Sample Identification

EB-8-8/12/04
MW-22-1
MW-22-2
MW-22-3
MW-22-3MS
MW-22-3MSD

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Raw data were not reviewed for this SDG. The review was based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

Raw data were not reviewed for this SDG.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

No field duplicates were identified in this SDG.

X. Field Blanks

Sample EB-8-8/12/04 was identified as an equipment blank. No contaminant concentrations were found in this blank.

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 04-4044

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 04-4044

No Sample Data Qualified in this SDG

Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.
Project ID: JPL GW-3Q04

Project No: 4-12812
Service ID: 44044

Anal. Method: 7196
Collected by: JJ/MM

Component Name: Chromium (VI)
CAS No: 1333-82-0

| Lab ID | Sample ID | Matrix | Coll. Date | Rcv Date | Anal. Date | Batch | Unit | RL | Result | Q |
|---------------|---------------|--------|------------|----------|------------|---------|------|------|--------|---|
| 04-4044-1 | EB-8-8/12/04 | Water | 08/12/04 | 08/12/04 | 08/12/04 | 04W3602 | mg/L | 0.01 | <0.01 | U |
| 04-4044-2 | MW-22-1 | Water | 08/12/04 | 08/12/04 | 08/12/04 | 04W3602 | mg/L | 0.01 | <0.01 | U |
| 04-4044-3 | MW-22-2 | Water | 08/12/04 | 08/12/04 | 08/12/04 | 04W3602 | mg/L | 0.01 | <0.01 | U |
| 04-4044-4 | MW-22-3 | Water | 08/12/04 | 08/12/04 | 08/12/04 | 04W3602 | mg/L | 0.01 | <0.01 | U |
| 04W3602-MB-01 | 04W3602-MB-01 | Water | 08/12/04 | 08/12/04 | 08/12/04 | 04W3602 | mg/L | 0.01 | <0.01 | U |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten signature/initials
9/28/04

Applied P & CH Laboratories
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc.
Project ID: JPL GW-3Q04

Project No: 4-12812
Service ID: 44044

Anal. Method 314.0
Collected by: JJ/MM

Component Name: Perchlorate
CAS No:

| Lab ID | Sample ID | Matrix | Coll. Date | Rcv Date | Anal. Date | Batch | Unit | RL | Result | Q |
|---------------|---------------|--------|------------|----------|------------|---------|------|----|--------|---|
| 04-4044-1 | EB-8-8/12/04 | Water | 08/12/04 | 08/12/04 | 08/23/04 | 04W3687 | µg/L | 4 | <4 | U |
| 04-4044-2 | MW-22-1 | Water | 08/12/04 | 08/12/04 | 08/23/04 | 04W3687 | µg/L | 4 | 2.6 | B |
| 04-4044-3 | MW-22-2 | Water | 08/12/04 | 08/12/04 | 08/23/04 | 04W3687 | µg/L | 4 | 2.8 | B |
| 04-4044-4 | MW-22-3 | Water | 08/12/04 | 08/12/04 | 08/23/04 | 04W3687 | µg/L | 4 | <4 | U |
| 04W3687-MB-01 | 04W3687-MB-01 | Water | 08/23/04 | 08/23/04 | 08/23/04 | 04W3687 | µg/L | 4 | <4 | U |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten signature and date: JJ/MM 9/22/04

LDC #: 12484A6

VALIDATION COMPLETENESS WORKSHEET

Date: 9/21/04

SDG #: 04-4044

Level III

Page: 1 of 1

Laboratory: Applied Physics & Chemistry Laboratory

Reviewer: *my*2nd Reviewer: *[Signature]***METHOD:** Hexavalent Chromium (EPA SW846 Method 7196), Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

| | Validation Area | | Comments |
|-------|--------------------------------------|----|----------------------------------|
| I. | Technical holding times | A | Sampling dates: 8/12/04 |
| IIa. | Initial calibration | A | |
| IIb. | Calibration verification | A | |
| III. | Blanks | A | |
| IV | Matrix Spike/Matrix Spike Duplicates | A | M ₂ /M ₂ b |
| V | Duplicates | N | |
| VI. | Laboratory control samples | A | LCS/LCS ₂ |
| VII. | Sample result verification | N | |
| VIII. | Overall assessment of data | A | |
| IX. | Field duplicates | N | |
| X | Field blanks | ND | EB = 1 |

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: ** Indicates sample underwent Level IV validation

| | | | | | | | |
|----|----------------|----|--|----|--|----|--|
| 1 | EB-8-8/12/04 | 11 | | 21 | | 31 | |
| 2 | MW-22-1 | 12 | | 22 | | 32 | |
| 3 | MW-22-2 | 13 | | 23 | | 33 | |
| 4 | MW-22-3 | 14 | | 24 | | 34 | |
| 5 | MW-22-3MS | 15 | | 25 | | 35 | |
| 6 | MW-22-3MSD | 16 | | 26 | | 36 | |
| 7 | M ₂ | 17 | | 27 | | 37 | |
| 8 | | 18 | | 28 | | 38 | |
| 9 | | 19 | | 29 | | 39 | |
| 10 | | 20 | | 30 | | 40 | |

Notes: _____

LDC #: 12484
 SDG #: 04-40

VALIDATION FINDINGS WORKSHEET
Sample Specific Analysis Reference

Page: 1 of 1
 Reviewer: MI
 2nd reviewer: [Signature]

All circled methods are applicable to each sample.

| Sample ID | Parameter |
|-----------|--|
| 1-4 | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ <u>CR⁶⁺</u> |
| 2-5,6 | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ <u>CR⁶⁺</u> |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |

Comments: _____

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 16, 2004
LDC Report Date: September 21, 2004
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 04-4085

Sample Identification

DUPE-5-3Q04
MW-5
MW-6**
MW-15
MW-16
MW-15MS
MW-15MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196A for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

Samples DUPE-5-3Q04 and MW-5 were identified as field duplicates. No contaminant concentrations were detected in any of the samples.

X. Field Blanks

No field blanks were identified in this SDG.

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 04-4085

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 04-4085

No Sample Data Qualified in this SDG

Applied P & CH Laboratories
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc. Project No: 4-12812 Anal. Method 314.0
 Project ID: JPL GW-3Q04 Service ID: 44085 Collected by: JJ/MM

Component Name: Perchlorate
 CAS No:

| Lab ID | Sample ID | Matrix | Coll. Date | Rcv Date | Anal. Date | Batch | Unit | RL | Result | Q |
|---------------|---------------|--------|------------|----------|------------|---------|------|----|--------|---|
| 04-4085-1 | DUPE-5-3Q04 | Water | 08/16/04 | 08/16/04 | 08/23/04 | 04W3687 | µg/L | 4 | <4 | U |
| 04-4085-2 | MW-5 | Water | 08/16/04 | 08/16/04 | 08/23/04 | 04W3687 | µg/L | 4 | <4 | U |
| 04-4085-3 | MW-6 | Water | 08/16/04 | 08/16/04 | 08/23/04 | 04W3687 | µg/L | 4 | 3.2 | B |
| 04-4085-5 | MW-16 | Water | 08/16/04 | 08/16/04 | 08/24/04 | 04W3698 | µg/L | 40 | 833 | |
| 04W3687-MB-01 | 04W3687-MB-01 | Water | 08/23/04 | 08/23/04 | 08/23/04 | 04W3687 | µg/L | 4 | <4 | U |
| 04W3698-MB-01 | 04W3698-MB-01 | Water | 08/24/04 | 08/24/04 | 08/24/04 | 04W3698 | µg/L | 4 | <4 | U |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

JJ/MM

Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.
Project ID: JPL GW-3Q04

Project No: 4-12812
Service ID: 44085

Anal. Method 7196
Collected by: JJ/MM

Component Name: Chromium (VI)
CAS No: 1333-82-0

| Lab ID | Sample ID | Matrix | Coll. Date | Rcv Date | Anal. Date | Batch | Unit | RL | Result | Q |
|---------------|---------------|--------|------------|----------|------------|---------|------|------|--------|---|
| 04-4085-1 | DUPE-5-3Q04 | Water | 08/16/04 | 08/16/04 | 08/16/04 | 04W3639 | mg/L | 0.01 | <0.01 | U |
| 04-4085-2 | MW-5 | Water | 08/16/04 | 08/16/04 | 08/16/04 | 04W3639 | mg/L | 0.01 | <0.01 | U |
| 04-4085-3 | MW-6 | Water | 08/16/04 | 08/16/04 | 08/16/04 | 04W3639 | mg/L | 0.01 | <0.01 | U |
| 04-4085-4 | MW-15 | Water | 08/16/04 | 08/16/04 | 08/16/04 | 04W3639 | mg/L | 0.01 | <0.01 | U |
| 04-4085-5 | MW-16 | Water | 08/16/04 | 08/16/04 | 08/16/04 | 04W3639 | mg/L | 0.01 | <0.01 | U |
| 04W3639-MB-01 | 04W3639-MB-01 | Water | 08/16/04 | 08/16/04 | 08/16/04 | 04W3639 | mg/L | 0.01 | <0.01 | U |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

JJ
9/22/04

LDC #: 12484B6

VALIDATION COMPLETENESS WORKSHEET

Date: 9/21/04

SDG #: 04-4085

Level III/IV

Page: 1 of 1

Laboratory: Applied Physics & Chemistry Laboratory

Reviewer: WH

2nd Reviewer: [Signature]

METHOD: Hexavalent Chromium (EPA SW846 Method 7196), Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

| | Validation Area | | Comments |
|-------|--------------------------------------|----|--|
| I. | Technical holding times | A | Sampling dates: 8/16/04 |
| IIa. | Initial calibration | A | |
| IIb. | Calibration verification | A | |
| III. | Blanks | A | |
| IV. | Matrix Spike/Matrix Spike Duplicates | A | MS/MSD |
| V. | Duplicates | N | |
| VI. | Laboratory control samples | A | LCG/MSD |
| VII. | Sample result verification | B | Not reviewed for Level III validation. |
| VIII. | Overall assessment of data | A | |
| IX. | Field duplicates | NP | (1,2) |
| X. | Field blanks | N | |

Note: A = Acceptable
 N = Not provided/applicable
 SW = See worksheet

ND = No compounds detected
 R = Rinsate
 FB = Field blank

D = Duplicate
 TB = Trip blank
 EB = Equipment blank

Validated Samples: ** Indicates sample underwent Level IV validation

| | | | | | | | |
|----|-------------|----|--|----|--|----|--|
| 1 | DUPE-5-3Q04 | 11 | | 21 | | 31 | |
| 2 | MW-5 | 12 | | 22 | | 32 | |
| 3 | MW-6** | 13 | | 23 | | 33 | |
| 4 | MW-15 | 14 | | 24 | | 34 | |
| 5 | MW-16 | 15 | | 25 | | 35 | |
| 6 | MW-15MS | 16 | | 26 | | 36 | |
| 7 | MW-15MSD | 17 | | 27 | | 37 | |
| 8 | MB | 18 | | 28 | | 38 | |
| 9 | | 19 | | 29 | | 39 | |
| 10 | | 20 | | 30 | | 40 | |

Notes: _____

LDC #: 2484B6
 SDG #: 04-4085

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2
 Reviewer: MLP
 2nd Reviewer: [Signature]

Method: Inorganics (EPA Method See copy)

| Validation Area | Yes | No | NA | Findings/Comments |
|--|-----|----|----|-------------------|
| I. Technical holding times | | | | |
| All technical holding times were met. | ✓ | | | |
| Cooler temperature criteria was met. | ✓ | | | |
| II. Calibration | | | | |
| Were all instruments calibrated daily, each set-up time? | ✓ | | | |
| Were the proper number of standards used? | ✓ | | | |
| Were all initial calibration correlation coefficients ≥ 0.995 ? | ✓ | | | |
| Were all initial and continuing calibration verification %Rs within the 90-110% QC limits? | ✓ | | | |
| Were titrant checks performed as required? | | | ✓ | |
| Were balance checks performed as required? | | | ✓ | |
| Were all initial and continuing calibration verification %Rs within the 90-110% QC limits? | | | ✓ | |
| III. Blanks | | | | |
| Was a method blank associated with every sample in this SDG? | ✓ | | | |
| Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet. | | ✓ | | |
| IV. Matrix spike/Matrix spike duplicates and Duplicates | | | | |
| Were a matrix spike (MS) and duplicate (DUP) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD or MS/DUP, Soil / Water. | ✓ | | | |
| Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the 75-125 QC limits? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken. | ✓ | | | |
| Were the MS/MSD or duplicate relative percent differences (RPD) $\leq 20\%$ for waters and $\leq 35\%$ for soil samples? A control limit of $\leq 2X$ CRDL ($\leq 2X$ CRDL for soil) was used for samples that were $\leq 5X$ the CRDL, including when only one of the duplicate sample values were $\leq 5X$ the CRDL. | ✓ | | | |
| V. Laboratory control samples | | | | |
| Was an LCS analyzed for this SDG? | ✓ | | | |
| Was an LCS analyzed per extraction batch? | ✓ | | | |
| Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the 80-120% (85-115% for Method 300.0) QC limits? | ✓ | | | |
| VI. Regional Quality Assurance and Quality Control | | | | |
| Were performance evaluation (PE) samples performed? | | | ✓ | |
| Were the performance evaluation (PE) samples within the acceptance limits? | | | ✓ | |

LDC #: 12484B6
 SDG #: 04-4085

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

| Validation Area | Yes | No | NA | Findings/Comments |
|---|-----|----|----|-------------------|
| VII. Sample Result Verification | | | | |
| Were RLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation? | ✓ | | | |
| Were detection limits < RL? | ✓ | | | |
| VIII. Overall assessment of data | | | | |
| Overall assessment of data was found to be acceptable. | ✓ | | | |
| IX. Field duplicates | | | | |
| Field duplicate pairs were identified in this SDG. | ✓ | | | |
| Target analytes were detected in the field duplicates. | ✓ | ✓ | | |
| X. Field blanks | | | | |
| Field blanks were identified in this SDG. | | ✓ | | |
| Target analytes were detected in the field blanks. | | | ✓ | |

LDC #: 12484B6
SDG #: 04-4085

VALIDATION FINDINGS WORKSHEET

Sample Specific Analysis Reference

Page: 1 of 1
Reviewer: MI
2nd reviewer: [Signature]

All circled methods are applicable to each sample.

| Sample ID | Parameter |
|-----------|---|
| 1-5 | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC <u>CR⁶⁺</u> |
| 1-3.5 | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ <u>CO₂</u> |
| RC6,17 | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC <u>CR⁶⁺</u> |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN NH ₃ TKN TOC CR ⁶⁺ |

Comments: _____

LDC #: 12484B6
 SDG #: 04-4085

VALIDATION FINDINGS WORKSHEET
Initial and Continuing Calibration Calculation Verification

Page: 1 of 1
 Reviewer: MH
 2nd Reviewer: [Signature]

METHOD: Inorganics, Method See notes

The correlation coefficient (r) for the calibration of Cu⁶⁺ was recalculated. Calibration date: 7/22/04

An initial or continuing calibration verification percent recovery (%R) was recalculated for each type of analysis using the following formula:

$$\%R = \frac{\text{Found}}{\text{True}} \times 100$$

Where, Found = concentration of each analyte measured in the analysis of the ICV or CCV solution
 True = concentration of each analyte in the ICV or CCV source

| Type of Analysis | Analyte | Conc. (µg/L) (units) | Conc. (µg/L) (units) | Recalculated | | Reported | | Acceptable (Y/N) |
|--------------------------|------------------|----------------------|----------------------|--------------|----|----------|----|------------------|
| | | | | r | %R | r | %R | |
| Initial calibration | | Blank | 0 | | | | | |
| Calibration verification | Cu ⁶⁺ | Standard 1 | 0.050 | 0.043 | | | | |
| | | Standard 2 | 0.100 | 0.084 | | | | |
| | | Standard 3 | 0.200 | 0.170 | | | | |
| | | Standard 4 | 0.300 | 0.254 | | | | |
| | | Standard 5 | 0.500 | 0.412 | | | | |
| | | Standard 6 | | | | | | |
| | | Standard 7 | | | | | | |
| Calibration verification | Cu ⁶⁺ | 0.25 | 0.154 | | | 102 | | Y |
| Calibration verification | Cu ⁶⁺ | 50 | 47.0 | | | 94 | | Y |
| Calibration verification | | | | | | | | |

Comments: Refer to Calibration Verification findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12484 B6
 SDG #: 04-4085

VALIDATION FINDINGS WORKSHEET
Level IV Recalculation Worksheet

Page: 1 of 1
 Reviewer: MH
 2nd Reviewer: A

METHOD: Inorganics, Method see above

Percent recoveries (%R) for a laboratory control sample and a matrix spike sample were recalculated using the following formula:

$$\%R = \frac{\text{Found}}{\text{True}} \times 100$$

Where, Found = concentration of each analyte measured in the analysis of the sample. For the matrix spike calculation, Found = SSR (spiked sample result) - SR (sample result).
 True = concentration of each analyte in the source.

A sample and duplicate relative percent difference (RPD) was recalculated using the following formula:

$$RPD = \frac{|S-D|}{(S+D)/2} \times 100$$

Where, S = Original sample concentration
 D = Duplicate sample concentration

| Sample ID | Type of Analysis | Element | Found / S (units) | True / D (units) | Recalculated | | Acceptable (Y/N) |
|-----------|---------------------------|---------|-------------------|------------------|--------------|-------------------|------------------|
| | | | | | %R / RPD | Reported %R / RPD | |
| LC5 | Laboratory control sample | CO4 | 75.7 | 25 | 104 | 102 | Y |
| 6 | Matrix spike sample | Cr6 | (SSR-SR) 0.238 | 0.25 | 95 | 95 | Y |
| 617 | Duplicate sample | ✓ | 0.236 | 6.238 | 1 | 1 | Y |

Comments: Refer to appropriate worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12484 B6
 SDG #: 04-4085

VALIDATION FINDINGS WORKSHEET
Sample Calculation Verification

Page: 1 of 1
 Reviewer: MIT
 2nd reviewer: [Signature]

METHOD: Inorganics, Method See order

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- Y N N/A Have results been reported and calculated correctly?
- Y N N/A Are results within the calibrated range of the instruments?
- Y N N/A Are all detection limits below the CRQL?

Compound (analyte) results for CO4 3 reported with a positive detect were recalculated and verified using the following equation:

Concentration = _____ Recalculation: _____

$$CO4 = 0.000537 \times Avum$$

$$CO4 = 6006 \times 0.000537$$

$$= 3.22 \mu g/L$$

| # | Sample ID | Analyte | Reported Concentration ($\mu g/L$) | Calculated Concentration ($\mu g/L$) | Acceptable (Y/N) |
|---|-----------|---------|--------------------------------------|--|------------------|
| 1 | 3 | CO4 | 3.2 | 3.2 | Y |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

Note: _____

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: NASA JPL
Collection Date: August 17, 2004
LDC Report Date: September 21, 2004
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level III & IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 04-4096

Sample Identification

DUPE-3-3Q04
MW-8**
MW-10
MW-13
DUPE-3-3Q04MS
DUPE-3-3Q04MSD

**Indicates sample underwent EPA Level IV review

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196A for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section IX.

Samples indicated by a double asterisk on the front cover underwent a EPA Level IV review. A EPA Level III review was performed on all of the other samples. Raw data were not evaluated for the samples reviewed by Level III criteria since this review is based on QC data.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration were met.

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

IV. Matrix Spike/Matrix Spike Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Duplicates

The laboratory has indicated that there were no duplicate (DUP) analyses specified for the samples in this SDG, and therefore duplicate analyses were not performed for this SDG.

VI. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

VII. Sample Result Verification

All sample result verifications were within validation criteria for samples on which a EPA Level IV review was performed. Raw data were not evaluated for the samples reviewed by Level III criteria.

VIII. Overall Assessment of Data

Data flags are summarized at the end of this report.

IX. Field Duplicates

Samples DUPE-3-3Q04 and MW-10 were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

| Analyte | Concentration (ug/L) | | RPD |
|-------------|----------------------|--------|-----|
| | Dupe-7-2Q04 | MW-7** | |
| Perchlorate | 25.5 | 25.3 | 1 |

X. Field Blanks

No field blanks were identified in this SDG.

NASA JPL
Wet Chemistry - Data Qualification Summary - SDG 04-4096

No Sample Data Qualified in this SDG

NASA JPL
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 04-4096

No Sample Data Qualified in this SDG

Wet Analysis Results for Method 7196

Client Name: GEOFON, Inc.
Project ID: JPL GW-3Q04

Project No: 4-12812
Service ID: 44096

Anal. Method 7196
Collected by: JJ/MM/TM

Component Name: Chromium (VI)
CAS No: 1333-82-0

| Lab ID | Sample ID | Matrix | Coll. Date | Rcv Date | Anal. Date | Batch | Unit | RL | Result | Q |
|---------------|---------------|--------|------------|----------|------------|---------|------|------|--------|---|
| 04-4096-1 | DUPE-6-3Q04 | Water | 08/17/04 | 08/17/04 | 08/17/04 | 04W3645 | mg/L | 0.01 | <0.01 | U |
| 04-4096-2 | MW-8 | Water | 08/17/04 | 08/17/04 | 08/17/04 | 04W3645 | mg/L | 0.01 | <0.01 | U |
| 04-4096-3 | MW-10 | Water | 08/17/04 | 08/17/04 | 08/17/04 | 04W3645 | mg/L | 0.01 | <0.01 | U |
| 04-4096-4 | MW-13 | Water | 08/17/04 | 08/17/04 | 08/17/04 | 04W3645 | mg/L | 0.01 | 0.011 | |
| 04W3645-MB-01 | 04W3645-MB-01 | Water | 08/17/04 | 08/17/04 | 08/17/04 | 04W3645 | mg/L | 0.01 | <0.01 | U |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected, or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

Handwritten signature/initials
9/2/04

Applied P & CH Laboratories
Wet Analysis Results for Method 314.0

Client Name: GEOFON, Inc. Project No: 4-12812 Anal. Method 314.0
 Project ID: JPL GW-3Q04 Service ID: 44096 Collected by: JJ/MM/TM

Component Name: Perchlorate
 CAS No:

| Lab ID | Sample ID | Matrix | Coll. Date | Rcv Date | Anal. Date | Batch | Unit | RL | Result | Q |
|---------------|---------------|--------|------------|----------|------------|---------|------|----|--------|---|
| 04-4096-1 | DUPE-6-3Q04 | Water | 08/17/04 | 08/17/04 | 08/24/04 | 04W3698 | µg/L | 4 | 25.5 | |
| 04-4096-2 | MW-8 | Water | 08/17/04 | 08/17/04 | 08/24/04 | 04W3698 | µg/L | 4 | 9.4 | |
| 04-4096-3 | MW-10 | Water | 08/17/04 | 08/17/04 | 08/24/04 | 04W3698 | µg/L | 4 | 25.3 | |
| 04-4096-4 | MW-13 | Water | 08/17/04 | 08/17/04 | 08/24/04 | 04W3698 | µg/L | 24 | 296 | |
| 04W3698-MB-01 | 04W3698-MB-01 | Water | 08/24/04 | 08/24/04 | 08/24/04 | 04W3698 | µg/L | 4 | <4 | U |

Not Detected is shown as PQL, with dilution and moisture corrected if applicable.

Note: Q - Qualifier.

Qualifier: U - Not Detected or less than MDL

B - Less than RL (PQL, EQL or CRDL), but greater than MDL.

JJ
9/22/04

LDC #: 12484C6

VALIDATION COMPLETENESS WORKSHEET

SDG #: 04-4096

Level III/IV

Laboratory: Applied Physics & Chemistry Laboratory

Date: 9/21/04

Page: 1 of 1

Reviewer: WH

2nd Reviewer: *[Signature]***METHOD:** Hexavalent Chromium (EPA SW846 Method 7196), Perchlorate (EPA Method 314.0)

The samples listed below were reviewed for each of the following validation areas. Validation findings are noted in attached validation findings worksheets.

| | Validation Area | | Comments |
|-------|--------------------------------------|----|--|
| I. | Technical holding times | A | Sampling dates: 8/17/04 |
| IIa. | Initial calibration | A | |
| IIb. | Calibration verification | A | |
| III. | Blanks | A | |
| IV | Matrix Spike/Matrix Spike Duplicates | A | MS/MSD |
| V | Duplicates | N | |
| VI. | Laboratory control samples | A | LCS/LCSD |
| VII. | Sample result verification | A | Not reviewed for Level III validation. |
| VIII. | Overall assessment of data | A | |
| IX. | Field duplicates | SW | (1,3) |
| X | Field blanks | N | |

Note: A = Acceptable
N = Not provided/applicable
SW = See worksheet

ND = No compounds detected
R = Rinsate
FB = Field blank

D = Duplicate
TB = Trip blank
EB = Equipment blank

Validated Samples: ** Indicates sample underwent Level IV validation

| | | | | | | | |
|----|----------------|----|--|----|--|----|--|
| 1 | DUPE-3-3Q04 | 11 | | 21 | | 31 | |
| 2 | MW-8** | 12 | | 22 | | 32 | |
| 3 | MW-10 | 13 | | 23 | | 33 | |
| 4 | MW-13 | 14 | | 24 | | 34 | |
| 5 | DUPE-3-3Q04MS | 15 | | 25 | | 35 | |
| 6 | DUPE-3-3Q04MSD | 16 | | 26 | | 36 | |
| 7 | MB | 17 | | 27 | | 37 | |
| 8 | | 18 | | 28 | | 38 | |
| 9 | | 19 | | 29 | | 39 | |
| 10 | | 20 | | 30 | | 40 | |

Notes: _____

LDC #: 248406
 SDG #: 04-4096

VALIDATION FINDINGS CHECKLIST

Page: 1 of 2
 Reviewer: M12
 2nd Reviewer: [Signature]

Method: Inorganics (EPA Method See copy)

| Validation Area | Yes | No | NA | Findings/Comments |
|---|-----|----|----|---------------------|
| I. Technical holding times | | | | |
| All technical holding times were met. | ✓ | | | |
| Cooler temperature criteria was met. | ✓ | | | |
| II. Calibration | | | | |
| Were all instruments calibrated daily, each set-up time? | ✓ | | | |
| Were the proper number of standards used? | ✓ | | | |
| Were all initial calibration correlation coefficients ≥ 0.995 ? | ✓ | | | |
| Were all initial and continuing calibration verification %Rs within the 90-110% QC limits? | ✓ | | | |
| Were titrant checks performed as required? | | | ✓ | |
| Were balance checks performed as required? | | | ✓ | |
| Were all initial and continuing calibration verification %Rs within the 80-110% QC limits? | | | ✓ | |
| III. Blanks | | | | |
| Was a method blank associated with every sample in this SDG? | ✓ | | | |
| Was there contamination in the method blanks? If yes, please see the Blanks validation completeness worksheet. | | ✓ | | |
| IV. Matrix spike/Matrix spike duplicates and Duplicates | | | | |
| Were a matrix spike (MS) and duplicate (DUP) analyzed for each matrix in this SDG? If no, indicate which matrix does not have an associated MS/MSD or MS/DUP. Soil / Water. | | ✓ | | non client for coly |
| Were the MS/MSD percent recoveries (%R) and the relative percent differences (RPD) within the 75-125 QC limits? If the sample concentration exceeded the spike concentration by a factor of 4 or more, no action was taken. | ✓ | | | |
| Were the MS/MSD or duplicate relative percent differences (RPD) $\leq 20\%$ for waters and $\leq 35\%$ for soil samples? A control limit of $\leq \text{CRDL} (\leq 2X \text{ CRDL for soil})$ was used for samples that were $\leq 5X$ the CRDL, including when only one of the duplicate sample values were $\leq 5X$ the CRDL. | ✓ | | | |
| V. Laboratory control samples | | | | |
| Was an LCS analyzed for this SDG? | ✓ | | | |
| Was an LCS analyzed per extraction batch? | ✓ | | | |
| Were the LCS percent recoveries (%R) and relative percent difference (RPD) within the 80-120% (85-115% for Method 300.0) QC limits? | ✓ | | | |
| VI. Regional Quality Assurance and Quality Control | | | | |
| Were performance evaluation (PE) samples performed? | | | ✓ | |
| Were the performance evaluation (PE) samples within the acceptance limits? | | | ✓ | |

LDC #: 12484096
 SDG #: 04-4096

VALIDATION FINDINGS CHECKLIST

Page: 2 of 2
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

| Validation Area | Yes | No | NA | Findings/Comments |
|---|-------------------------------------|-------------------------------------|-------------------------------------|-------------------|
| VI. Sample Result Verification | | | | |
| Were RLs adjusted to reflect all sample dilutions and dry weight factors applicable to level IV validation? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | |
| Were detection limits < RL? | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | |
| VII. Overall assessment of data | | | | |
| Overall assessment of data was found to be acceptable. | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | |
| VIII. Field duplicates | | | | |
| Field duplicate pairs were identified in this SDG. | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | |
| Target analytes were detected in the field duplicates. | <input checked="" type="checkbox"/> | <input type="checkbox"/> | <input type="checkbox"/> | |
| IX. Field blanks | | | | |
| Field blanks were identified in this SDG. | <input type="checkbox"/> | <input checked="" type="checkbox"/> | <input type="checkbox"/> | |
| Target analytes were detected in the field blanks. | <input type="checkbox"/> | <input type="checkbox"/> | <input checked="" type="checkbox"/> | |

LDC #: 12484CB
 SDG #: 04-4096

VALIDATION FINDINGS WORKSHEET
Sample Specific Analysis Reference

Page: 1 of 1
 Reviewer: MI
 2nd reviewer: K

All circled methods are applicable to each sample.

| Sample ID | Parameter |
|-----------|--|
| 1-4 | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ <u>clay</u> |
| or 5,6 | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |
| | pH TDS Cl F NO ₃ NO ₂ SO ₄ PO ₄ ALK CN ⁻ NH ₃ TKN TOC CR ⁰⁺ |

Comments: _____

LDC #: 1248406
 SDG #: 04-4096

VALIDATION FINDINGS WORKSHEET
Field Duplicates

Page: 1 of 1
 Reviewer: MH
 2nd reviewer: ✓

METHOD: Inorganics, Method see work

N N/A Were field duplicate pairs identified in this SDG?
 N N/A Were target analytes detected in the field duplicate pairs?

| Analyte | Concentration (<u>ug/L</u>) | | RPD (Limit) | Difference (Limit) | Qualifier |
|------------|-------------------------------|-------------|-------------|--------------------|-----------|
| | 1 | 3 | | | |
| <u>CO4</u> | <u>25.5</u> | <u>25.3</u> | <u>1</u> | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

| Analyte | Concentration () | | RPD (Limit) | Difference (Limit) | Qualifier |
|---------|-------------------|--|-------------|--------------------|-----------|
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

| Analyte | Concentration () | | RPD (Limit) | Difference (Limit) | Qualifier |
|---------|-------------------|--|-------------|--------------------|-----------|
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

| Analyte | Concentration () | | RPD (Limit) | Difference (Limit) | Qualifier |
|---------|-------------------|--|-------------|--------------------|-----------|
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

LDC #: 1248466
 SDG #: 04-4096

VALIDATION FINDINGS WORKSHEET
Initial and Continuing Calibration Calculation Verification

Page: 1 of 1
 Reviewer: [Signature]
 2nd Reviewer: [Signature]

METHOD: Inorganics, Method See notes

The correlation coefficient (r) for the calibration of ceaf was recalculated. Calibration date: 2/3/04

An initial or continuing calibration verification percent recovery (%R) was recalculated for each type of analysis using the following formula:

$\%R = \frac{\text{Found}}{\text{True}} \times 100$ Where, Found = concentration of each analyte measured in the analysis of the ICV or CCV solution
 True = concentration of each analyte in the ICV or CCV source

| Type of Analysis | Analyte | Conc. (ug/L) | Mean (units) | Recalculated | | Reported | | Acceptable (Y/N) |
|--------------------------|------------|--------------|--------------|--------------|----|----------|----|------------------|
| | | | | r | %R | r | %R | |
| Initial calibration | | | | | | | | |
| Calibration verification | Blank | | | | | | | |
| | Standard 1 | 2 | 3089 | | | | | |
| | Standard 2 | 10 | 16874.3 | | | | | |
| | Standard 3 | 25 | 40814.65 | | | | | |
| | Standard 4 | 50 | 94273.0 | | | | | |
| | Standard 5 | 75 | 140396.3 | | | | | |
| | Standard 6 | 100 | 186528.45 | | | | | |
| Standard 7 | | | | | | | | |
| Calibration verification | ceaf | 0.25 | | | | | | |
| Calibration verification | ceaf | 51.12 | | | | | | |
| Calibration verification | | | | | | | | |

Comments: Refer to Calibration Verification findings worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12484 cb
 SDG #: 04-4096

VALIDATION FINDINGS WORKSHEET
Level IV Recalculation Worksheet

Page: 1 of 1
 Reviewer: Mif
 2nd Reviewer: [Signature]

METHOD: Inorganics, Method see Green

Percent recoveries (%R) for a laboratory control sample and a matrix spike sample were recalculated using the following formula:

$$\%R = \frac{\text{Found}}{\text{True}} \times 100 \quad \text{Where,} \quad \text{Found} = \text{concentration of each analyte measured in the analysis of the sample. For the matrix spike calculation, Found} = \text{SSR (spiked sample result) - SR (sample result).}$$

$$\text{True} = \text{concentration of each analyte in the source.}$$

A sample and duplicate relative percent difference (RPD) was recalculated using the following formula:

$$RPD = \frac{|S-D|}{(S+D)/2} \times 100 \quad \text{Where,} \quad S = \text{Original sample concentration}$$

$$D = \text{Duplicate sample concentration}$$

| Sample ID | Type of Analysis | Element | Found / S (units) | True / D (units) | Recalculated | | Reported | | Acceptable (Y/N) |
|-----------|---------------------------|------------|----------------------|---------------------|--------------|----------|----------|----------|---------------------|
| | | | | | %R / RPD | %R / RPD | %R / RPD | %R / RPD | |
| LC5 | Laboratory control sample | Cr6+ MM | 29.8 | 25 | 119 | 119 | 119 | 119 | Y |
| 5 | Matrix spike sample | Cr6+ ↓ | (SSR-SR) 0.238 | 0.225 | 95 | 95 | 95 | 95 | ↓ |
| 5/6 | Duplicate sample | ↓ | 0.234 | 0.238 | 2 | 2 | 1 | 1 | ↓ |

Comments: Refer to appropriate worksheet for list of qualifications and associated samples when reported results do not agree within 10.0% of the recalculated results.

LDC #: 12484Cb
 SDG #: 04-4596

VALIDATION FINDINGS WORKSHEET
Sample Calculation Verification

Page: 1 of 1
 Reviewer: MIF
 2nd reviewer: [Signature]

METHOD: Inorganics, Method See cover

Please see qualifications below for all questions answered "N". Not applicable questions are identified as "N/A".

- Y N N/A Have results been reported and calculated correctly?
- Y N N/A Are results within the calibrated range of the instruments?
- Y N N/A Are all detection limits below the CRQL?

Compound (analyte) results for 2 reported with a positive detect were recalculated and verified using the following equation:

Concentration = $ClO_4 = 0.000537 \times Area$ Recalculation: $ClO_4 = 17550 \times 0.000537$
 $= 9.42 \text{ ug/L}$

| # | Sample ID | Analyte | Reported Concentration (ug/L) | Calculated Concentration (ug/L) | Acceptable (Y/N) |
|---|-----------|------------------|-------------------------------|---------------------------------|------------------|
| 1 | 2 | ClO ₄ | 9.4 | 9.42 | Y |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |
| | | | | | |

Note: _____